

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082725\
 Data File : BP025599.D
 Acq On : 27 Aug 2025 11:42
 Operator : CG/JU
 Sample : PB169382BL
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 PB169382BL

Quant Time: Aug 27 12:16:37 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081425.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 14 18:14:31 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.778	152	178208	20.000	ng	-0.02
21) Naphthalene-d8	10.549	136	686289	20.000	ng	-0.01
39) Acenaphthene-d10	14.395	164	425674	20.000	ng	-0.01
64) Phenanthrene-d10	17.195	188	961568	20.000	ng	0.00
76) Chrysene-d12	21.642	240	1057440	20.000	ng	0.00
86) Perylene-d12	25.012	264	1170620	20.000	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.402	112	1339736	124.848	ng	0.00
7) Phenol-d6	6.966	99	1604477	116.622	ng	0.00
23) Nitrobenzene-d5	8.919	82	1035647	76.336	ng	-0.02
42) 2,4,6-Tribromophenol	15.913	330	779246	136.003	ng	0.00
45) 2-Fluorobiphenyl	13.007	172	2381137	76.450	ng	-0.01
79) Terphenyl-d14	19.919	244	4485346	77.605	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082725\
Data File : BP025599.D
Acq On : 27 Aug 2025 11:42
Operator : CG/JU
Sample : PB169382BL
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
PB169382BL

Quant Time: Aug 27 12:16:37 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 14 18:14:31 2025
Response via : Initial Calibration

