

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP082820\
 Data File : BP003153.D
 Acq On : 28 Aug 2020 16:33
 Operator : CG/JU
 Sample : L3498-05MS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 HR-06-082620MS

Manual Integrations
 APPROVED

mohammad
 8/31/2020 10:28:05 AM

Quant Time: Aug 28 18:17:24 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP082420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 24 15:32:08 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.66	152	232885	20.00	ng	0.00
21) Naphthalene-d8	10.44	136	911989	20.00	ng	0.00
39) Acenaphthene-d10	14.30	164	560975	20.00	ng	0.00
64) Phenanthrene-d10	17.06	188	1233886	20.00	ng	0.00
76) Chrysene-d12	21.15	240	1273472	20.00	ng	0.00
86) Perylene-d12	23.38	264	993466	20.00	ng	0.00

System Monitoring Compounds

5) 2-Fluorophenol	5.28	112	1614833	122.60	ng	0.01
7) Phenol-d6	6.85	99	1878231	113.75	ng	0.00
23) Nitrobenzene-d5	8.81	82	1121362	88.64	ng	0.00
42) 2,4,6-Tribromophenol	15.80	330	1009831	133.10	ng	0.00
45) 2-Fluorobiphenyl	12.92	172	3171399	82.79	ng	0.00
79) Terphenyl-d14	19.63	244	4718247	81.53	ng	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	3.23	88	172146	34.185	ng	# 58
4) n-Nitrosodimethylamine	3.52	42	153630	44.392	ng	88
8) 2-Chlorophenol	7.23	128	695419	47.357	ng	99
9) Benzaldehyde	6.80	77	220350	28.624	ng	98
10) Phenol	6.88	94	698763	45.664	ng	95
11) bis(2-Chloroethyl)ether	7.09	93	523667	43.237	ng	99
12) 1,3-Dichlorobenzene	7.56	146	662421	37.100	ng	99
13) 1,4-Dichlorobenzene	7.70	146	679018	37.670	ng	99
14) 1,2-Dichlorobenzene	8.01	146	657052	38.448	ng	99
15) Benzyl Alcohol	7.90	79	377499	40.576	ng	98
16) 2,2'-oxybis(1-Chloropropan	8.20	45	410547	40.000	ng	98
17) 2-Methylphenol	8.12	107	523225	47.998	ng	98
18) Hexachloroethane	8.74	117	214218	36.326	ng	94
19) n-Nitroso-di-n-propylamine	8.47	70	340524	45.645	ng	100
20) 3+4-Methylphenols	8.44	107	695760	47.370	ng	99
22) Acetophenone	8.48	105	874087	43.038	ng	100
24) Nitrobenzene	8.85	77	538963	43.376	ng	100
25) Isophorone	9.37	82	1052977	47.175	ng	100
26) 2-Nitrophenol	9.56	139	371710	50.192	ng	97
27) 2,4-Dimethylphenol	9.63	122	606445	54.472	ng	98
28) bis(2-Chloroethoxy)methane	9.86	93	656544	39.492	ng	100
29) 2,4-Dichlorophenol	10.10	162	660114	47.791	ng	99
30) 1,2,4-Trichlorobenzene	10.31	180	636948	38.730	ng	99
31) Naphthalene	10.49	128	1906758	40.695	ng	100
32) Benzoic acid	9.80	122	413950	48.052	ng	96
34) Hexachlorobutadiene	10.79	225	360859	36.561	ng	99
35) Caprolactam	11.37	113	176070m	42.419	ng	
36) 4-Chloro-3-methylphenol	11.74	107	628085	48.702	ng	99
37) 2-Methylnaphthalene	12.11	142	1438091	42.905	ng	99
38) 1-Methylnaphthalene	12.33	142	1360407	42.702	ng	99
40) 1,2,4,5-Tetrachlorobenzene	12.49	216	709307	42.660	ng	99
41) Hexachlorocyclopentadiene	12.47	237	862828	109.778	ng	98
43) 2,4,6-Trichlorophenol	12.73	196	524881	48.010	ng	99
44) 2,4,5-Trichlorophenol	12.80	196	575769	49.789	ng	99

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP082820\
 Data File : BP003153.D
 Acq On : 28 Aug 2020 16:33
 Operator : CG/JU
 Sample : L3498-05MS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 HR-06-082620MS

Manual Integrations
 APPROVED

mohammad
 8/31/2020 10:28:05 AM

Quant Time: Aug 28 18:17:24 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP082420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 24 15:32:08 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) 1,1'-Biphenyl	13.13	154	1839107	43.707	ng	99
47) 2-Chloronaphthalene	13.17	162	1407179	40.855	ng	100
48) 2-Nitroaniline	13.37	65	268585	41.648	ng	95
49) Acenaphthylene	14.02	152	2310619	44.365	ng	100
50) Dimethylphthalate	13.77	163	2127889	49.741	ng	100
51) 2,6-Dinitrotoluene	13.88	165	422836	47.558	ng	97
52) Acenaphthene	14.37	154	1367146	42.728	ng	100
53) 3-Nitroaniline	14.20	138	42243	4.194	ng	93
54) 2,4-Dinitrophenol	14.42	184	501044	100.916	ng	99
55) Dibenzofuran	14.70	168	2176015	43.304	ng	99
56) 4-Nitrophenol	14.53	139	774624	106.889	ng	98
57) 2,4-Dinitrotoluene	14.67	165	586462	49.058	ng	98
58) Fluorene	15.36	166	1687898	43.605	ng	100
59) 2,3,4,6-Tetrachlorophenol	14.94	232	516035	49.321	ng	99
60) Diethylphthalate	15.15	149	1865224	44.441	ng	99
61) 4-Chlorophenyl-phenylether	15.36	204	856760	42.860	ng	99
62) 4-Nitroaniline	15.37	138	267434	25.827	ng	97
63) Azobenzene	15.65	77	1227027	44.049	ng	97
65) 4,6-Dinitro-2-methylphenol	15.44	198	338822	57.419	ng	99
66) n-Nitrosodiphenylamine	15.57	169	960455	26.621	ng	98
67) 4-Bromophenyl-phenylether	16.25	248	584744	42.940	ng	100
68) Hexachlorobenzene	16.37	284	692106	42.688	ng	100
69) Atrazine	16.53	200	280465	23.048	ng	99
70) Pentachlorophenol	16.72	266	917020	118.280	ng	99
71) Phenanthrene	17.10	178	2862198	42.809	ng	100
72) Anthracene	17.19	178	2902076	45.353	ng	100
73) Carbazole	17.46	167	1691906	27.550	ng	99
74) Di-n-butylphthalate	18.03	149	3274106	45.353	ng	100
75) Fluoranthene	19.07	202	3470186	44.907	ng	99
78) Pyrene	19.43	202	3544330	42.883	ng	99
80) Butylbenzylphthalate	20.31	149	1576008	46.311	ng	96
81) Benzo(a)anthracene	21.13	228	3491178	43.766	ng	99
83) Chrysene	21.19	228	3319892	43.468	ng	99
84) Bis(2-ethylhexyl)phthalate	21.08	149	2253492	45.854	ng	98
85) Di-n-octyl phthalate	21.95	149	4013138	47.628	ng	99
87) Indeno(1,2,3-cd)pyrene	25.66	276	4051630	54.687	ng	98
88) Benzo(b)fluoranthene	22.71	252	3852037	62.384	ng	99
89) Benzo(k)fluoranthene	22.76	252	3685507	62.470	ng	99
90) Benzo(a)pyrene	23.29	252	3282365	57.299	ng	99
91) Dibenzo(a,h)anthracene	25.67	278	3497645	57.469	ng	98
92) Benzo(g,h,i)perylene	26.35	276	3310801	54.207	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP082820\
Data File : BP003153.D
Acq On : 28 Aug 2020 16:33
Operator : CG/JU
Sample : L3498-05MS
Misc :
ALS Vial : 8 Sample Multiplier: 1


Instrument :
BNA_P
ClientSampleId :
HR-06-082620MS

Manual Integrations
APPROVED

mohammad
8/31/2020 10:28:05 AM

Quant Time: Aug 28 18:17:24 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP082420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 24 15:32:08 2020
Response via : Initial Calibration





#1





1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.66 min Scan# 812
Delta R.T. -0.01 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Instrument :
BNA_P
ClientSampleId :
HR-06-082620MS

Tot Ion	Ratio	Lower	Upper
152	100		
150	154.3	123.1	184.7
115	54.1	42.8	64.2

Manual Integrations
APPROVED




mohammad
8/31/2020 10:28:05 AM




#2

1,4-Dioxane
Concen: 34.185 ng
RT: 3.23 min Scan# 59
Delta R.T. 0.02 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Tgt Ion	Ratio	Lower	Upper
88	100		
58	55.3	44.3	66.5
43	94.3	14.3	21.5





#4





n-Nitrosodimethylamine
 Concen: 44.392 ng
 RT: 3.52 min Scan# 107
 Delta R.T. 0.01 min
 Lab File: BP003153.D
 Acq: 28 Aug 2020 16:33

Instrument :
 BNA_P
 ClientSampleId :
 HR-06-082620MS

Tot Ion	Ratio	Lower	Upper
42	100		
74	182.5	161.0	241.6
44	6.8	5.6	8.4

Manual Integrations
 APPROVED




mohammad
 8/31/2020 10:28:05 AM




#5

2-Fluorophenol
 Concen: 122.596 ng
 RT: 5.28 min Scan# 407
 Delta R.T. 0.01 min
 Lab File: BP003153.D
 Acq: 28 Aug 2020 16:33

Tgt Ion	Ratio	Lower	Upper
112	100		
64	39.3	33.0	49.6
63	20.5	16.9	25.3





#7

Phenol-d6

Concen: 113.749 ng

RT: 6.85 min Scan# 674

Delta R.T. 0.01 min

Lab File: BP003153.D

Acq: 28 Aug 2020 16:33

Instrument :

BNA_P

ClientSampleId :

HR-06-082620MS

Tot Ion: 99 Resp: 1878231

Ion Ratio Lower Upper

99 100

42 10.8 8.5 12.7





71 29.5 22.8 34.2

Manual Integrations

APPROVED

mohammad

8/31/2020 10:28:05 AM



#8

2-Chlorophenol

Concen: 47.357 ng

RT: 7.23 min Scan# 739

Delta R.T. -0.00 min

Lab File: BP003153.D

Acq: 28 Aug 2020 16:33




Tgt Ion: 128 Resp: 695419


Ion Ratio Lower Upper

128 100

130 32.8 12.2 52.2

64 29.0 9.7 49.7





#9

Benzaldehyde

Concen: 28.624 ng

RT: 6.80 min Scan# 666

Delta R.T. -0.00 min

Lab File: BP003153.D

Acq: 28 Aug 2020 16:33

Instrument :

BNA_P





ClientSampleId :

HR-06-082620MS

Tot Ion	77	Resp	220350
Ion	Ratio	Lower	Upper
77	100		
105	115.8	98.3	138.3
106	112.5	93.9	133.9

Manual Integrations
APPROVED

mohammad
8/31/2020 10:28:05 AM



#10

Phenol

Concen: 45.664 ng




RT: 6.88 min Scan# 678


Delta R.T. -0.00 min

Lab File: BP003153.D

Acq: 28 Aug 2020 16:33

Tgt Ion	94	Resp	698763
Ion	Ratio	Lower	Upper
94	100		
65	22.5	1.0	41.0
66	32.7	9.7	49.7









#11
bis(2-Chloroethyl)ether
Concen: 43.237 ng
RT: 7.09 min Scan# 715
Delta R.T. -0.00 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Instrument :
BNA_P
ClientSampleId :
HR-06-082620MS

Tot Ion	Ratio	Lower	Upper
93	100		
63	57.3	37.8	77.8
95	33.1	12.0	52.0




Manual Integrations
APPROVED


mohammad
8/31/2020 10:28:05 AM



#12
1,3-Dichlorobenzene
Concen: 37.100 ng
RT: 7.56 min Scan# 794
Delta R.T. -0.00 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Tgt Ion	Ratio	Lower	Upper
146	100		
148	65.1	51.8	77.6
75	21.5	17.4	26.2





#13

1,4-Dichlorobenzene

Concen: 37.670 ng

RT: 7.70 min Scan# 818

Delta R.T. -0.00 min

Lab File: BP003153.D

Acq: 28 Aug 2020 16:33

Instrument :

BNA_P





ClientSampleId :

HR-06-082620MS

Tot Ion:	146	Resp:	679018
Ion Ratio	Lower	Upper	
146	100		
148	64.3	50.6	75.8
111	36.2	28.5	42.7

Manual Integrations
APPROVED

mohammad
8/31/2020 10:28:05 AM



#14

1,2-Dichlorobenzene

Concen: 38.448 ng




RT: 8.01 min Scan# 871


Delta R.T. -0.01 min

Lab File: BP003153.D

Acq: 28 Aug 2020 16:33

Tgt Ion:	146	Resp:	657052
Ion Ratio	Lower	Upper	
146	100		
148	63.8	51.4	77.2
111	38.0	31.0	46.4





#15

Benzyl Alcohol

Concen: 40.576 ng

RT: 7.90 min Scan# 853

Delta R.T. -0.00 min

Lab File: BP003153.D

Acq: 28 Aug 2020 16:33

Instrument :

BNA_P





ClientSampleId :

HR-06-082620MS

Tot Ion:	79	Resp:	377499
Ion Ratio	Lower	Upper	
79	100		
108	91.4	75.7	113.5
77	62.0	50.2	75.2

Manual Integrations
APPROVED

mohammad
8/31/2020 10:28:05 AM



#16

2,2'-oxybis(1-Chloropropane)

Concen: 40.000 ng




RT: 8.20 min Scan# 903


Delta R.T. -0.00 min

Lab File: BP003153.D

Acq: 28 Aug 2020 16:33

Tgt Ion:	45	Resp:	410547
Ion Ratio	Lower	Upper	
45	100		
77	22.5	1.6	41.6
79	18.0	0.0	36.7









#17
2-Methylphenol
Concen: 47.998 ng
RT: 8.12 min Scan# 889
Delta R.T. 0.01 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Instrument :
BNA_P
ClientSampleId :
HR-06-082620MS

Tot Ion	Ratio	Resp	Lower	Upper
107	100			
108	111.4	90.8	136.2	
77	38.1	29.4	44.2	
79	37.6	30.0	45.0	




Manual Integrations
APPROVED




mohammad
8/31/2020 10:28:05 AM



#18
Hexachloroethane
Concen: 36.326 ng
RT: 8.74 min Scan# 995
Delta R.T. -0.00 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Tgt Ion	Ratio	Resp	Lower	Upper
117	100			
119	98.4	77.7	116.5	
201	121.2	88.4	132.6	









#19
n-Nitroso-di-n-propylamine
Concen: 45.645 ng
RT: 8.47 min Scan# 949
Delta R.T. -0.00 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Instrument :
BNA_P
ClientSampleId :
HR-06-082620MS

Tot Ion	Ratio	Resp	Lower	Upper
70	100			
42	40.9	32.5	48.7	
101	12.6	10.2	15.2	
130	31.3	25.2	37.8	


Manual Integrations
APPROVED


mohammad
8/31/2020 10:28:05 AM



#20
3+4-Methylphenols
Concen: 47.370 ng
RT: 8.44 min Scan# 944
Delta R.T. -0.00 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Tgt Ion	Ratio	Resp	Lower	Upper
107	100			
108	90.7	69.7	109.7	
77	28.3	7.2	47.2	
79	25.4	4.4	44.4	









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.44 min Scan# 1284
Delta R.T. -0.01 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Instrument :
BNA_P
ClientSampleId :
HR-06-082620MS

Tot Ion	Ion	Ratio	Resp	Lower	Upper
136	100		911989		
137	11.1	8.8	13.2		
54	4.3	3.5	5.3		
68	3.9	3.2	4.8		




Manual Integrations
APPROVED


mohammad
8/31/2020 10:28:05 AM



#22
Acetophenone
Concen: 43.038 ng
RT: 8.48 min Scan# 951
Delta R.T. -0.00 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Tgt Ion	Ion	Ratio	Resp	Lower	Upper
105	100		874087		
71	3.3	3.1	4.7		
51	16.4	13.0	19.6		
120	24.4	19.5	29.3		









#23
 Nitrobenzene-d5
 Concen: 88.641 ng
 RT: 8.81 min Scan# 1007
 Delta R.T. -0.00 min
 Lab File: BP003153.D
 Acq: 28 Aug 2020 16:33

Instrument :
 BNA_P
 ClientSampleId :
 HR-06-082620MS

Tot Ion	Ratio	Lower	Upper
82	100		
128	57.8	44.9	67.3
54	36.6	29.4	44.2




Manual Integrations
 APPROVED


mohammad
 8/31/2020 10:28:05 AM



#24
 Nitrobenzene
 Concen: 43.376 ng
 RT: 8.85 min Scan# 1014
 Delta R.T. -0.00 min
 Lab File: BP003153.D
 Acq: 28 Aug 2020 16:33

Tgt Ion	Ratio	Lower	Upper
77	100		
123	61.2	49.0	73.4
65	12.2	9.6	14.4









#25
Isophorone
Concen: 47.175 ng
RT: 9.37 min Scan# 1103
Delta R.T. -0.01 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Instrument :
BNA_P
ClientSampleId :
HR-06-082620MS

Tot Ion	82	Resp	1052977
Ion	Ratio	Lower	Upper
82	100		
95	8.0	6.0	9.0
138	23.9	19.1	28.7




Manual Integrations
APPROVED


mohammad
8/31/2020 10:28:05 AM



#26
2-Nitrophenol
Concen: 50.192 ng
RT: 9.56 min Scan# 1134
Delta R.T. -0.01 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Tgt Ion	139	Resp	371710
Ion	Ratio	Lower	Upper
139	100		
109	28.3	21.2	31.8
65	32.9	25.4	38.0









#27
2,4-Dimethylphenol
Concen: 54.472 ng
RT: 9.63 min Scan# 1147
Delta R.T. -0.00 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Instrument :
BNA_P
ClientSampleId :
HR-06-082620MS

Tot Ion	Ratio	Resp	Lower	Upper
122	100	606445		
107	107.2	84.3	126.5	
121	58.8	46.0	69.0	




Manual Integrations
APPROVED


mohammad
8/31/2020 10:28:05 AM



#28
bis(2-Chloroethoxy)methane
Concen: 39.492 ng
RT: 9.86 min Scan# 1186
Delta R.T. -0.00 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Tgt Ion	Ratio	Resp	Lower	Upper
93	100	656544		
95	32.4	25.8	38.8	
123	13.5	10.6	15.8	









#29
2,4-Dichlorophenol
Concen: 47.791 ng
RT: 10.10 min Scan# 1226
Delta R.T. -0.00 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Instrument :
BNA_P
ClientSampleId :
HR-06-082620MS

Tot Ion	Ion	Ratio	Resp	Lower	Upper
162	100		660114		
164	64.5	43.6	83.6		
98	29.5	10.1	50.1		




Manual Integrations
APPROVED


mohammad
8/31/2020 10:28:05 AM



#30
1,2,4-Trichlorobenzene
Concen: 38.730 ng
RT: 10.31 min Scan# 1262
Delta R.T. -0.01 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Tgt Ion	Ion	Ratio	Resp	Lower	Upper
180	100		636948		
182	96.0	77.6	116.4		
145	30.1	23.7	35.5		









#31
Naphthalene
Concen: 40.695 ng
RT: 10.49 min Scan# 1293
Delta R.T. -0.01 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Instrument :
BNA_P
ClientSampleId :
HR-06-082620MS

Tot Ion	Ratio	Lower	Upper
128	100		
129	11.0	8.6	13.0
127	13.2	10.5	15.7




Manual Integrations
APPROVED


mohammad
8/31/2020 10:28:05 AM



#32
Benzoic acid
Concen: 48.052 ng
RT: 9.80 min Scan# 1175
Delta R.T. 0.02 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Tgt Ion	Ratio	Lower	Upper
122	100		
105	112.5	98.1	138.1
77	73.3	55.3	95.3





#34

Hexachlorobutadiene

Concen: 36.561 ng

RT: 10.79 min Scan# 1344

Delta R.T. -0.01 min

Lab File: BP003153.D

Acq: 28 Aug 2020 16:33

Instrument :

BNA_P





ClientSampleId :

HR-06-082620MS

Tot Ion:	225	Resp:	360859
Ion Ratio	Lower	Upper	
225	100		
223	63.2	50.6	76.0
227	64.3	52.0	78.0

Manual Integrations
APPROVED

mohammad
8/31/2020 10:28:05 AM



#35

Caprolactam

Concen: 42.419 ng m




RT: 11.37 min Scan# 1442


Delta R.T. 0.01 min

Lab File: BP003153.D

Acq: 28 Aug 2020 16:33

Tgt Ion:	113	Resp:	176070
Ion Ratio	Lower	Upper	
113	100		
55	97.1	81.3	121.3
56	73.7	56.4	96.4









#36
4-Chloro-3-methylphenol
Concen: 48.702 ng
RT: 11.74 min Scan# 1505
Delta R.T. -0.01 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Instrument :
BNA_P
ClientSampleId :
HR-06-082620MS

Tot Ion	Ratio	Lower	Upper
107	100		
144	29.6	24.2	36.4
142	91.0	73.0	109.6




Manual Integrations
APPROVED


mohammad
8/31/2020 10:28:05 AM



#37
2-Methylnaphthalene
Concen: 42.905 ng
RT: 12.11 min Scan# 1568
Delta R.T. -0.01 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Tgt Ion	Ratio	Lower	Upper
142	100		
141	89.2	71.1	106.7
115	32.1	25.2	37.8









#38
 1-Methylnaphthalene
 Concen: 42.702 ng
 RT: 12.33 min Scan# 1606
 Delta R.T. -0.00 min
 Lab File: BP003153.D
 Acq: 28 Aug 2020 16:33

Instrument :
 BNA_P
 ClientSampleId :
 HR-06-082620MS

Tot Ion	Ratio	Lower	Upper
142	100		
141	93.0	73.8	110.8
116	3.7	2.8	4.2




Manual Integrations
 APPROVED


mohammad
 8/31/2020 10:28:05 AM



#39
 Acenaphthene-d10
 Concen: 20.000 ng
 RT: 14.30 min Scan# 1941
 Delta R.T. -0.01 min
 Lab File: BP003153.D
 Acq: 28 Aug 2020 16:33

Tgt Ion	Ratio	Lower	Upper
164	100		
162	100.7	81.2	121.8
160	44.7	36.0	54.0









#40
1,2,4,5-Tetrachlorobenzene
Concen: 42.660 ng
RT: 12.49 min Scan# 1633
Delta R.T. -0.00 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Instrument :
BNA_P
ClientSampleId :
HR-06-082620MS

Tot Ion	Ion	Ratio	Resp	Lower	Upper
216	100				
214	78.4	62.4	93.6		
179	21.1	16.9	25.3		
108	17.4	12.8	19.2		




Manual Integrations
APPROVED


mohammad
8/31/2020 10:28:05 AM



#41
Hexachlorocyclopentadiene
Concen: 109.778 ng
RT: 12.47 min Scan# 1630
Delta R.T. -0.01 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Tgt Ion	Ion	Ratio	Resp	Lower	Upper
237	100				
235	61.6	43.0	83.0		
272	13.6	0.0	33.2		









#42
2,4,6-Tribromophenol
Concen: 133.099 ng
RT: 15.80 min Scan# 2195
Delta R.T. -0.01 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Instrument :
BNA_P
ClientSampleId :
HR-06-082620MS

Tot Ion	Ratio	Lower	Upper
330	100		
332	97.1	77.3	115.9
141	28.7	24.4	36.6




Manual Integrations
APPROVED


mohammad
8/31/2020 10:28:05 AM



#43
2,4,6-Trichlorophenol
Concen: 48.010 ng
RT: 12.73 min Scan# 1673
Delta R.T. -0.01 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Tgt Ion	Ratio	Lower	Upper
196	100		
198	95.1	75.8	113.6
200	30.4	24.6	37.0









#44
2,4,5-Trichlorophenol
Concen: 49.789 ng
RT: 12.80 min Scan# 1686
Delta R.T. -0.01 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Instrument :
BNA_P
ClientSampleId :
HR-06-082620MS

Tot Ion	Ratio	Resp	Lower	Upper
196	100	575769		
198	94.8	76.6	115.0	
97	39.8	30.3	45.5	
132	26.5	21.3	31.9	




Manual Integrations
APPROVED


mohammad
8/31/2020 10:28:05 AM



#45
2-Fluorobiphenyl
Concen: 82.791 ng
RT: 12.92 min Scan# 1706
Delta R.T. -0.01 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Tgt Ion	Ratio	Resp	Lower	Upper
172	100	3171399		
171	35.8	28.7	43.1	
170	23.8	19.0	28.4	









#46
1,1'-Biphenyl
Concen: 43.707 ng
RT: 13.13 min Scan# 1742
Delta R.T. -0.01 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Instrument :
BNA_P
ClientSampleId :
HR-06-082620MS

Tot Ion	Ratio	Lower	Upper
154	100		
153	40.8	20.3	60.3
76	9.0	0.0	29.7




Manual Integrations
APPROVED


mohammad
8/31/2020 10:28:05 AM



#47
2-Chloronaphthalene
Concen: 40.855 ng
RT: 13.17 min Scan# 1748
Delta R.T. -0.01 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Tgt Ion	Ratio	Lower	Upper
162	100		
127	38.4	30.6	45.8
164	33.2	26.3	39.5









#48
2-Nitroaniline
Concen: 41.648 ng
RT: 13.37 min Scan# 1783
Delta R.T. -0.00 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Instrument :
BNA_P
ClientSampleId :
HR-06-082620MS

Tot Ion	Ratio	Lower	Upper
65	100		
92	89.2	70.2	105.4
138	173.3	131.7	197.5




Manual Integrations
APPROVED


mohammad
8/31/2020 10:28:05 AM



#49
Acenaphthylene
Concen: 44.365 ng
RT: 14.02 min Scan# 1893
Delta R.T. -0.01 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Tgt Ion	Ratio	Lower	Upper
152	100		
151	20.3	16.2	24.4
153	13.5	10.5	15.7









#50
Dimethylphthalate
Concen: 49.741 ng
RT: 13.77 min Scan# 1850
Delta R.T. -0.01 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Instrument :
BNA_P
ClientSampleId :
HR-06-082620MS

Tot Ion	Ratio	Lower	Upper
163	100		
194	4.2	3.4	5.0
164	10.1	8.1	12.1




Manual Integrations
APPROVED


mohammad
8/31/2020 10:28:05 AM



#51
2,6-Dinitrotoluene
Concen: 47.558 ng
RT: 13.88 min Scan# 1869
Delta R.T. -0.00 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Tgt Ion	Ratio	Lower	Upper
165	100		
63	31.6	27.5	41.3
89	40.7	33.6	50.4





#52

Acenaphthene

Concen: 42.728 ng

RT: 14.37 min Scan# 1952

Delta R.T. -0.01 min

Lab File: BP003153.D

Acq: 28 Aug 2020 16:33

Instrument :

BNA_P





ClientSampleId :

HR-06-082620MS

Tot Ion:	154	Resp:	1367146
Ion	Ratio	Lower	Upper
154	100		
153	111.9	89.2	133.8
152	53.3	42.5	63.7

Manual Integrations
APPROVED

mohammad
8/31/2020 10:28:05 AM



#53

3-Nitroaniline

Concen: 4.194 ng




RT: 14.20 min Scan# 1924


Delta R.T. -0.01 min

Lab File: BP003153.D

Acq: 28 Aug 2020 16:33

Tgt Ion:	138	Resp:	42243
Ion	Ratio	Lower	Upper
138	100		
108	9.1	7.4	11.0
92	97.1	72.2	108.2









#54
2,4-Dinitrophenol
Concen: 100.916 ng
RT: 14.42 min Scan# 1960
Delta R.T. 0.00 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Instrument :
BNA_P
ClientSampleId :
HR-06-082620MS

Tot Ion	Ratio	Lower	Upper
184	100		
63	39.7	32.0	48.0
154	61.6	48.2	72.2




Manual Integrations
APPROVED


mohammad
8/31/2020 10:28:05 AM



#55
Dibenzofuran
Concen: 43.304 ng
RT: 14.70 min Scan# 2009
Delta R.T. -0.01 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Tgt Ion	Ratio	Lower	Upper
168	100		
139	37.3	29.3	43.9
169	13.5	10.6	15.8









#56
4-Nitrophenol
Concen: 106.889 ng
RT: 14.53 min Scan# 1980
Delta R.T. 0.01 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Instrument :
BNA_P
ClientSampleId :
HR-06-082620MS

Tot Ion	Ratio	Lower	Upper
139	100		
109	66.1	42.8	82.8
65	61.3	41.7	81.7




Manual Integrations
APPROVED


mohammad
8/31/2020 10:28:05 AM



#57
2,4-Dinitrotoluene
Concen: 49.058 ng
RT: 14.67 min Scan# 2003
Delta R.T. -0.01 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Tgt Ion	Ratio	Lower	Upper
165	100		
63	26.9	20.3	30.5
89	54.1	42.0	63.0
182	2.9	2.3	3.5









#58
 Fluorene
 Concen: 43.605 ng
 RT: 15.36 min Scan# 2120
 Delta R.T. -0.01 min
 Lab File: BP003153.D
 Acq: 28 Aug 2020 16:33

Instrument :
 BNA_P
 ClientSampleId :
 HR-06-082620MS

Tot Ion	Ratio	Lower	Upper
166	100		
165	97.6	78.3	117.5
167	13.5	11.0	16.4




Manual Integrations
 APPROVED


mohammad
 8/31/2020 10:28:05 AM



#59
 2,3,4,6-Tetrachlorophenol
 Concen: 49.321 ng
 RT: 14.94 min Scan# 2049
 Delta R.T. -0.00 min
 Lab File: BP003153.D
 Acq: 28 Aug 2020 16:33

Tgt Ion	Ratio	Lower	Upper
232	100		
131	41.9	34.3	51.5
130	2.3	1.8	2.6
166	30.5	24.2	36.2









#60
Diethylphthalate
Concen: 44.441 ng
RT: 15.15 min Scan# 2084
Delta R.T. -0.00 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Instrument :
BNA_P
ClientSampleId :
HR-06-082620MS

Tot Ion	Ratio	Lower	Upper
149	100		
177	23.3	18.1	27.1
150	12.5	10.2	15.4




Manual Integrations
APPROVED


mohammad
8/31/2020 10:28:05 AM



#61
4-Chlorophenyl-phenylether
Concen: 42.860 ng
RT: 15.36 min Scan# 2120
Delta R.T. -0.01 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Tgt Ion	Ratio	Lower	Upper
204	100		
206	32.6	26.6	39.8
141	57.2	45.5	68.3









#62
4-Nitroaniline
Concen: 25.827 ng
RT: 15.37 min Scan# 2123
Delta R.T. -0.00 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Instrument :
BNA_P
ClientSampleId :
HR-06-082620MS

Tot Ion	Ratio	Lower	Upper
138	100		
92	39.0	21.1	61.1
108	68.2	46.6	86.6




Manual Integrations
APPROVED


mohammad
8/31/2020 10:28:05 AM



#63
Azobenzene
Concen: 44.049 ng
RT: 15.65 min Scan# 2170
Delta R.T. -0.00 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Tgt Ion	Ratio	Lower	Upper
77	100		
182	38.6	15.4	55.4
105	25.8	5.1	45.1
51	22.4	1.9	41.9





#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 17.06 min Scan# 2409

Delta R.T. -0.01 min

Lab File: BP003153.D

Acq: 28 Aug 2020 16:33

Instrument :

BNA_P





ClientSampleId :

HR-06-082620MS

Tot Ion:	188	Resp:	1233886
Ion	Ratio	Lower	Upper
188	100		
94	7.4	6.2	9.4
80	7.3	6.4	9.6

Manual Integrations
APPROVED

mohammad
8/31/2020 10:28:05 AM



#65

4,6-Dinitro-2-methylphenol

Concen: 57.419 ng




RT: 15.44 min Scan# 2134


Delta R.T. -0.01 min

Lab File: BP003153.D

Acq: 28 Aug 2020 16:33

Tgt Ion:	198	Resp:	338822
Ion	Ratio	Lower	Upper
198	100		
51	23.8	2.4	42.4
105	36.1	16.1	56.1









#66
n-Nitrosodiphenylamine
Concen: 26.621 ng
RT: 15.57 min Scan# 2156
Delta R.T. -0.01 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Instrument :
BNA_P
ClientSampleId :
HR-06-082620MS

Tot Ion	Ratio	Resp	Lower	Upper
169	100			
168	65.8	54.2	81.4	
167	35.9	29.4	44.2	




Manual Integrations
APPROVED


mohammad
8/31/2020 10:28:05 AM



#67
4-Bromophenyl-phenylether
Concen: 42.940 ng
RT: 16.25 min Scan# 2272
Delta R.T. -0.01 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Tgt Ion	Ratio	Resp	Lower	Upper
248	100			
250	98.4	78.3	117.5	
141	68.9	54.9	82.3	









#68
Hexachlorobenzene
Concen: 42.688 ng
RT: 16.37 min Scan# 2292
Delta R.T. -0.01 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Instrument :
BNA_P
ClientSampleId :
HR-06-082620MS

Tot Ion:	284	Resp:	692106
Ion Ratio		Lower	Upper
284	100		
142	32.0	25.8	38.6
249	30.5	24.3	36.5




Manual Integrations
APPROVED


mohammad
8/31/2020 10:28:05 AM



#69
Atrazine
Concen: 23.048 ng
RT: 16.53 min Scan# 2319
Delta R.T. -0.01 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Tgt Ion:	200	Resp:	280465
Ion Ratio		Lower	Upper
200	100		
173	26.3	6.3	46.3
215	48.6	27.9	67.9









#70
 Pentachlorophenol
 Concen: 118.280 ng
 RT: 16.72 min Scan# 2351
 Delta R.T. -0.00 min
 Lab File: BP003153.D
 Acq: 28 Aug 2020 16:33

Instrument :
 BNA_P
 ClientSampleId :
 HR-06-082620MS

Tot Ion:	266	Resp:	917020
Ion Ratio		Lower	Upper
266	100		
268	64.2	51.3	76.9
264	63.1	51.0	76.6




Manual Integrations
 APPROVED


mohammad
 8/31/2020 10:28:05 AM



#71
 Phenanthrene
 Concen: 42.809 ng
 RT: 17.10 min Scan# 2416
 Delta R.T. -0.01 min
 Lab File: BP003153.D
 Acq: 28 Aug 2020 16:33

Tgt Ion:	178	Resp:	2862198
Ion Ratio		Lower	Upper
178	100		
176	19.5	15.4	23.2
179	15.6	12.4	18.6





#72

Anthracene

Concen: 45.353 ng

RT: 17.19 min Scan# 2431

Delta R.T. -0.01 min

Lab File: BP003153.D

Acq: 28 Aug 2020 16:33

Instrument :

BNA_P

ClientSampleId :

HR-06-082620MS

Tot Ion:178 Resp: 2902076

Ion Ratio Lower Upper

178 100

176 19.0 15.2 22.8





179 15.8 12.6 19.0

Manual Integrations

APPROVED

mohammad

8/31/2020 10:28:05 AM



#73

Carbazole

Concen: 27.550 ng

RT: 17.46 min Scan# 2477

Delta R.T. -0.01 min

Lab File: BP003153.D

Acq: 28 Aug 2020 16:33




Tgt Ion:167 Resp: 1691906


Ion Ratio Lower Upper

167 100

166 21.7 17.6 26.4

139 12.6 10.3 15.5





#74

Di-n-butylphthalate

Concen: 45.353 ng

RT: 18.03 min Scan# 2575

Delta R.T. -0.01 min

Lab File: BP003153.D

Acq: 28 Aug 2020 16:33

Instrument :

BNA_P





ClientSampleId :

HR-06-082620MS

Tot Ion:	149	Resp:	3274106
Ion	Ratio	Lower	Upper
149	100		
150	9.5	7.5	11.3
104	4.3	3.3	4.9

Manual Integrations
APPROVED

mohammad
8/31/2020 10:28:05 AM



#75

Fluoranthene

Concen: 44.907 ng




RT: 19.07 min Scan# 2752


Delta R.T. -0.01 min

Lab File: BP003153.D

Acq: 28 Aug 2020 16:33

Tgt Ion:	202	Resp:	3470186
Ion	Ratio	Lower	Upper
202	100		
101	9.4	0.0	30.0
203	18.0	0.0	37.7









#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.15 min Scan# 3105
Delta R.T. -0.01 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Instrument :
BNA_P
ClientSampleId :
HR-06-082620MS

Tot Ion	Ratio	Lower	Upper
240	100		
120	9.5	8.4	12.6
236	25.9	20.6	31.0




Manual Integrations
APPROVED


mohammad
8/31/2020 10:28:05 AM



#78
Pyrene
Concen: 42.883 ng
RT: 19.43 min Scan# 2812
Delta R.T. -0.01 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Tgt Ion	Ratio	Lower	Upper
202	100		
200	21.8	17.0	25.4
203	17.8	14.0	21.0





#79

Terphenyl-d14

Concen: 81.526 ng

RT: 19.63 min Scan# 2847

Delta R.T. -0.01 min

Lab File: BP003153.D

Acq: 28 Aug 2020 16:33

Instrument :

BNA_P

ClientSampleId :

HR-06-082620MS

Tot Ion:244 Resp: 4718247

Ion Ratio Lower Upper


244 100

212 7.5 5.8 8.8

122 9.2 7.7 11.5

Manual Integrations
APPROVED




mohammad
8/31/2020 10:28:05 AM



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#80

Butylbenzylphthalate

Concen: 46.311 ng

RT: 20.31 min Scan# 2962

Delta R.T. -0.01 min

Lab File: BP003153.D

Acq: 28 Aug 2020 16:33


Tgt Ion:149 Resp: 1576008

Ion Ratio Lower Upper

149 100

91 54.8 46.8 70.2



206 21.4 16.8 25.2




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 91.00 (90.70 to 91.70): BP0

Ion 206.00 (205.70 to 206.70): E









#81
Benzo(a)anthracene
Concen: 43.766 ng
RT: 21.13 min Scan# 3102
Delta R.T. -0.01 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Instrument :
BNA_P
ClientSampleId :
HR-06-082620MS

Tot Ion	Ratio	Lower	Upper
228	100		
226	28.0	22.0	33.0
229	20.5	16.2	24.2




Manual Integrations
APPROVED


mohammad
8/31/2020 10:28:05 AM



#83
Chrysene
Concen: 43.468 ng
RT: 21.19 min Scan# 3111
Delta R.T. -0.01 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Tgt Ion	Ratio	Lower	Upper
228	100		
226	30.9	24.2	36.4
229	20.2	16.0	24.0





#84

Bis(2-ethylhexyl)phthalate

Concen: 45.854 ng

RT: 21.08 min Scan# 3093

Delta R.T. -0.01 min

Lab File: BP003153.D

Acq: 28 Aug 2020 16:33

Instrument :

BNA_P

ClientSampleId :

HR-06-082620MS

Tot Ion:149 Resp: 2253492

Ion Ratio Lower Upper

149 100

167 29.0 22.5 33.7


279 5.6 3.8 5.8

Manual Integrations

APPROVED

mohammad




8/31/2020 10:28:05 AM



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E



#85

Di-n-octyl phthalate

Concen: 47.628 ng

RT: 21.95 min Scan# 3241

Delta R.T. -0.01 min

Lab File: BP003153.D

Acq: 28 Aug 2020 16:33


Tgt Ion:149 Resp: 4013138

Ion Ratio Lower Upper

149 100

167 1.7 1.3 1.9



43 4.5 3.8 5.6




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 43.00 (42.70 to 43.70): BPO









#86
Pervlene-d12
Concen: 20.000 ng
RT: 23.38 min Scan# 3484
Delta R.T. -0.01 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Instrument :
BNA_P
ClientSampleId :
HR-06-082620MS

Tot Ion	Ratio	Lower	Upper
264	100		
260	26.2	19.0	28.6
265	20.7	17.7	26.5




Manual Integrations
APPROVED


mohammad
8/31/2020 10:28:05 AM



#87
Indeno(1,2,3-cd)pyrene
Concen: 54.687 ng
RT: 25.66 min Scan# 3872
Delta R.T. -0.01 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Tgt Ion	Ratio	Lower	Upper
276	100		
138	23.5	20.2	30.4
277	25.6	20.4	30.6









#88
Benzo(b)fluoranthene
Concen: 62.384 ng
RT: 22.71 min Scan# 3370
Delta R.T. -0.01 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Instrument :
BNA_P
ClientSampleId :
HR-06-082620MS

Tot Ion	Ratio	Lower	Upper
252	100		
253	22.6	17.5	26.3
125	9.5	7.6	11.4




Manual Integrations
APPROVED


mohammad
8/31/2020 10:28:05 AM



#89
Benzo(k)fluoranthene
Concen: 62.470 ng
RT: 22.76 min Scan# 3378
Delta R.T. -0.01 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.4	17.8	26.6
125	9.2	8.0	12.0









#90
Benzo(a)pyrene
Concen: 57.299 ng
RT: 23.29 min Scan# 3468
Delta R.T. -0.01 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Instrument :
BNA_P
ClientSampleId :
HR-06-082620MS

Tot Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.7	26.5
125	10.0	8.8	13.2




Manual Integrations
APPROVED


mohammad
8/31/2020 10:28:05 AM



#91
Dibenzo(a,h)anthracene
Concen: 57.469 ng
RT: 25.67 min Scan# 3874
Delta R.T. -0.01 min
Lab File: BP003153.D
Acq: 28 Aug 2020 16:33

Tgt Ion	Ratio	Lower	Upper
278	100		
139	15.0	12.8	19.2
279	24.3	19.0	28.6





#92
 Benzo(a,h,i)perylene
 Concen: 54.207 ng
 RT: 26.35 min Scan# 3989
 Delta R.T. -0.01 min
 Lab File: BP003153.D
 Acq: 28 Aug 2020 16:33

Instrument :
 BNA_P
 ClientSampleId :
 HR-06-082620MS

Tot Ion	Ratio	Lower	Upper
276	100		
277	24.0	19.4	29.2
138	20.9	17.0	25.6

Manual Integrations
 APPROVED

mohammad
 8/31/2020 10:28:05 AM

