

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082823\
 Data File : BP016830.D
 Acq On : 29 Aug 2023 13:55
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020709

Quant Time: Aug 29 22:25:01 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP082423.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 29 01:12:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.028	152	553358	20.000	ng/ul	0.00
20) Naphthalene-d8	10.857	136	2449879	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.681	164	1516278	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.451	188	3144212	20.000	ng/ul	0.00
79) Chrysene-d12	21.545	240	2035261	20.000	ng/ul	0.00
88) Perylene-d12	24.115	264	1971311	20.000	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.381	96	98653	7.368	ng/uL	0.00
4) Pyridine-d5	3.816	84	738595	18.052	ng/ul	0.00
7) Phenol-d5	7.169	99	878955	17.863	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.363	67	567319	18.090	ng/ul	0.00
11) 2-Chlorophenol-d4	7.546	132	697147	18.844	ng/ul	0.00
15) 4-Methylphenol-d8	8.734	113	725890	18.072	ng/ul	0.00
21) Nitrobenzene-d5	9.228	128	349475	18.867	ng/ul	0.00
24) 2-Nitrophenol-d4	9.940	143	378772	19.399	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.463	165	704476	19.353	ng/ul	0.00
31) 4-Chloroaniline-d4	11.016	131	1004933	18.002	ng/ul	0.00
46) Dimethylphthalate-d6	14.092	166	2093434	18.308	ng/ul	0.00
49) Acenaphthylene-d8	14.381	160	2463772	19.060	ng/ul	0.00
54) 4-Nitrophenol-d4	14.887	143	355286	15.032	ng/ul	0.00
60) Fluorene-d10	15.675	176	1774227	18.425	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.804	200	291289	15.556	ng/ul	0.00
73) Anthracene-d10	17.551	188	2605610	18.696	ng/ul	0.00
81) Pyrene-d10	19.780	212	2620017	23.439	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.945	264	1835217	18.585	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.422	88	105871	7.261	ng/uL	94
5) Pyridine	3.834	79	765144	18.155	ng/ul	99
6) Benzaldehyde	7.181	77	564078	20.963	ng/ul	99
8) Phenol	7.199	94	922276	17.910	ng/ul	99
10) Bis(2-Chloroethyl)ether	7.463	93	754104	18.142	ng/ul	98
12) 2-Chlorophenol	7.581	128	704856	18.640	ng/ul	98
13) 2-Methylphenol	8.463	108	697018	18.045	ng/ul	100
14) 2,2'-oxybis(1-Chloropr...	8.540	45	678476	11.190	ng/ul	99
16) Acetophenone	8.875	105	1140478	18.311	ng/ul	99
17) N-Nitroso-di-n-propyla...	8.852	70	605405	18.444	ng/ul	99
18) 4-Methylphenol	8.799	108	757295	18.004	ng/ul	99
19) Hexachloroethane	9.099	117	285751	18.161	ng/ul	97
22) Nitrobenzene	9.269	77	883986	18.602	ng/ul	99
23) Isophorone	9.787	82	1693867	18.353	ng/ul	99
25) 2-Nitrophenol	9.975	139	410368	19.008	ng/ul	99
26) 2,4-Dimethylphenol	10.010	107	807632	18.612	ng/ul	98
27) Bis(2-Chloroethoxy)met...	10.269	93	1045385	18.359	ng/ul	99
29) 2,4-Dichlorophenol	10.493	162	699117	19.120	ng/ul	99
30) Naphthalene	10.910	128	2349453	18.513	ng/ul	100
32) 4-Chloroaniline	11.040	127	978542	17.838	ng/ul	99
33) Hexachlorobutadiene	11.140	225	417156	18.521	ng/ul	99
34) Caprolactam	11.863	113	224853	16.675	ng/ul	95
35) 4-Chloro-3-methylphenol	12.134	107	755619	18.428	ng/ul	99
36) 2-Methylnaphthalene	12.510	142	1629690	18.604	ng/ul	100


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082823\
 Data File : BP016830.D
 Acq On : 29 Aug 2023 13:55
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020709

Quant Time: Aug 29 22:25:01 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP082423.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 29 01:12:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.728	142	1642558	18.595	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.857	216	823043	19.045	ng/ul	99
40) Hexachlorocyclopentadiene	12.810	237	645893	20.323	ng/ul	98
41) 2,4,6-Trichlorophenol	13.110	196	556525	19.284	ng/ul	98
42) 2,4,5-Trichlorophenol	13.175	196	593839	19.016	ng/ul	97
43) 1,1'-Biphenyl	13.516	154	2140674	18.988	ng/ul	99
44) 2-Chloronaphthalene	13.563	162	1680417	18.882	ng/ul	99
45) 2-Nitroaniline	13.792	65	498627	18.475	ng/ul	99
47) Dimethylphthalate	14.139	163	2102942	18.159	ng/ul	100
48) 2,6-Dinitrotoluene	14.287	165	427685	18.847	ng/ul	95
50) Acenaphthylene	14.410	152	2779274	18.785	ng/ul	100
51) 3-Nitroaniline	14.622	138	418396	17.707	ng/ul	97
52) Acenaphthene	14.745	153	1804750	18.438	ng/ul	100
53) 2,4-Dinitrophenol	14.822	184	260535	15.800	ng/ul	98
55) 4-Nitrophenol	14.904	109	266676	14.940	ng/ul	98
56) Dibenzofuran	15.081	168	2469326	18.386	ng/ul	99
57) 2,4-Dinitrotoluene	15.063	165	602845	18.049	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	15.298	232	482344	18.143	ng/ul	96
59) Diethylphthalate	15.492	149	2115977	17.986	ng/ul	100
61) Fluorene	15.734	166	2051703	18.264	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.722	204	1010194	18.316	ng/ul	100
63) 4-Nitroaniline	15.792	138	369930	16.453	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.816	198	319271	15.815	ng/ul	98
67) N-Nitrosodiphenylamine	15.945	169	1701662	19.568	ng/ul	100
68) 4-Bromophenyl-phenylether	16.628	248	582237	19.314	ng/ul	97
69) Hexachlorobenzene	16.710	284	633806	18.778	ng/ul	99
70) Atrazine	16.904	200	610921	18.634	ng/ul	98
71) Pentachlorophenol	17.069	266	432381	17.257	ng/ul	99
72) Phenanthrene	17.492	178	3073699	18.569	ng/ul	100
74) Anthracene	17.586	178	3094967	18.492	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.469	216	868730	20.635	ng/ul	100
76) Pentachlorobenzene	14.975	250	753905	20.038	ng/ul	98
77) Carbazole	17.869	167	2689187	17.840	ng/ul	99
78) Di-n-butylphthalate	18.375	149	3509522	18.662	ng/ul	100
80) Fluoranthene	19.451	202	3211189	23.915	ng/ul	100
82) Pyrene	19.810	202	3239272	23.354	ng/ul	99
83) Butylbenzylphthalate	20.669	149	1314608	21.977	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.468	252	748193	17.199	ng/ul	99
85) Benzo(a)anthracene	21.527	228	2622713	18.708	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.404	149	1859625	21.291	ng/ul	99
87) Chrysene	21.586	228	2390181	18.769	ng/ul	99
89) Di-n-octyl phthalate	22.380	149	2678967	18.046	ng/ul	100
90) Benzo(b)fluoranthene	23.310	252	2243799	17.825	ng/ul	100
91) Benzo(k)fluoranthene	23.363	252	2256645	18.126	ng/ul	99
93) Benzo(a)pyrene	23.998	252	2111967	18.548	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.845	276	2530552	21.749	ng/ul	99
95) Dibenzo(a,h)anthracene	26.868	278	2087269	21.599	ng/ul	99
96) Benzo(g,h,i)perylene	27.703	276	2021615	22.918	ng/ul	98





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
 1,4-Dioxane
 Concen: 7.261 ng/uL
 RT: 3.422 min Scan# 21
 Delta R.T. 0.006 min
 Lab File: BP016830.D
 Acq: 29 Aug 2023 13:55




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020709


Tgt Ion: 88 Resp: 105871
 Ion Ratio Lower Upper
 88 100
 43 28.5 24.0 36.0
 58 68.1 59.8 89.8



#5
 Pyridine
 Concen: 18.155 ng/uL
 RT: 3.834 min Scan# 93
 Delta R.T. 0.000 min
 Lab File: BP016830.D
 Acq: 29 Aug 2023 13:55

Tgt Ion: 79 Resp: 765144
 Ion Ratio Lower Upper
 79 100
 52 76.1 59.6 89.4
 51 35.1 28.2 42.2









#6
Benzaldehyde
Concen: 20.963 ng/ul
RT: 7.181 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP016830.D
Acq: 29 Aug 2023 13:55




Instrument :
BNA_P
ClientSampleId :
SSTD020709


Tgt Ion: 77 Resp: 564078
Ion Ratio Lower Upper
77 100
105 92.2 75.4 113.2
106 90.8 73.1 109.7



#8
Phenol
Concen: 17.910 ng/ul
RT: 7.199 min Scan# 665
Delta R.T. 0.000 min
Lab File: BP016830.D
Acq: 29 Aug 2023 13:55

Tgt Ion: 94 Resp: 922276
Ion Ratio Lower Upper
94 100
65 27.7 22.4 33.6
66 36.5 28.6 42.8









#10
Bis(2-Chloroethyl)ether
Concen: 18.142 ng/ul
RT: 7.463 min Scan# 710
Delta R.T. 0.000 min
Lab File: BP016830.D
Acq: 29 Aug 2023 13:55




Instrument :
BNA_P
ClientSampleId :
SSTD020709


Tgt Ion: 93 Resp: 754104
Ion Ratio Lower Upper
93 100
63 75.9 62.9 94.3
95 32.4 26.3 39.5



#12
2-Chlorophenol
Concen: 18.640 ng/ul
RT: 7.581 min Scan# 730
Delta R.T. 0.000 min
Lab File: BP016830.D
Acq: 29 Aug 2023 13:55

Tgt Ion:128 Resp: 704856
Ion Ratio Lower Upper
128 100
64 48.5 40.0 60.0
130 31.7 26.4 39.6





#13

2-Methylphenol

Concen: 18.045 ng/ul

RT: 8.463 min Scan# 880

Delta R.T. 0.000 min

Lab File: BP016830.D

Acq: 29 Aug 2023 13:55

Instrument :

BNA_P

ClientSampleId :





SSTD020709

Tgt Ion:108 Resp: 697018

Ion Ratio Lower Upper

108 100

107 88.9 71.4 107.2



#14

2,2'-oxybis(1-Chloropropane)

Concen: 11.190 ng/ul

RT: 8.540 min Scan# 893

Delta R.T. -0.006 min

Lab File: BP016830.D

Acq: 29 Aug 2023 13:55




Tgt Ion: 45 Resp: 678476


Ion Ratio Lower Upper

45 100

77 14.2 11.1 16.7

79 10.4 8.2 12.4









#16
Acetophenone
Concen: 18.311 ng/ul
RT: 8.875 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP016830.D
Acq: 29 Aug 2023 13:55




Instrument :
BNA_P
ClientSampleId :
SSTD020709


Tgt Ion:105 Resp: 1140478
Ion Ratio Lower Upper
105 100
77 79.6 64.2 96.4
51 27.3 22.2 33.4



#17
N-Nitroso-di-n-propylamine
Concen: 18.444 ng/ul
RT: 8.852 min Scan# 946
Delta R.T. 0.006 min
Lab File: BP016830.D
Acq: 29 Aug 2023 13:55

Tgt Ion: 70 Resp: 605405
Ion Ratio Lower Upper
70 100
42 56.8 46.0 69.0
101 9.3 7.6 11.4
130 20.6 16.1 24.1









#18
4-Methylphenol
Concen: 18.004 ng/ul
RT: 8.799 min Scan# 911
Delta R.T. 0.000 min
Lab File: BP016830.D
Acq: 29 Aug 2023 13:55




Instrument :
BNA_P
ClientSampleId :
SSTD020709


Tgt Ion:108 Resp: 757295
Ion Ratio Lower Upper
108 100
107 111.5 89.8 134.6



#19
Hexachloroethane
Concen: 18.161 ng/ul
RT: 9.099 min Scan# 988
Delta R.T. -0.006 min
Lab File: BP016830.D
Acq: 29 Aug 2023 13:55

Tgt Ion:117 Resp: 285751
Ion Ratio Lower Upper
117 100
201 98.7 75.4 113.2
199 60.6 47.8 71.6









#22
Nitrobenzene
Concen: 18.602 ng/ul
RT: 9.269 min Scan# 1017
Delta R.T. 0.000 min
Lab File: BP016830.D
Acq: 29 Aug 2023 13:55




Instrument :
BNA_P
ClientSampleId :
SSTD020709


Tgt Ion: 77 Resp: 883986
Ion Ratio Lower Upper
77 100
123 43.3 34.9 52.3
65 12.9 10.6 15.8



#23
Isophorone
Concen: 18.353 ng/ul
RT: 9.787 min Scan# 1105
Delta R.T. 0.000 min
Lab File: BP016830.D
Acq: 29 Aug 2023 13:55

Tgt Ion: 82 Resp: 1693867
Ion Ratio Lower Upper
82 100
95 6.6 5.4 8.0
138 17.5 13.5 20.3









#25
2-Nitrophenol
Concen: 19.008 ng/ul
RT: 9.975 min Scan# 1137
Delta R.T. 0.000 min
Lab File: BP016830.D
Acq: 29 Aug 2023 13:55




Instrument :
BNA_P
ClientSampleId :
SSTD020709


Tgt Ion:139 Resp: 410368
Ion Ratio Lower Upper
139 100
65 51.3 42.0 63.0
109 29.1 23.4 35.2



#26
2,4-Dimethylphenol
Concen: 18.612 ng/ul
RT: 10.010 min Scan# 1143
Delta R.T. 0.000 min
Lab File: BP016830.D
Acq: 29 Aug 2023 13:55

Tgt Ion:107 Resp: 807632
Ion Ratio Lower Upper
107 100
121 49.8 40.7 61.1
122 85.2 69.7 104.5









#27
 Bis(2-Chloroethoxy)methane
 Concen: 18.359 ng/ul
 RT: 10.269 min Scan# 1187
 Delta R.T. 0.000 min
 Lab File: BP016830.D
 Acq: 29 Aug 2023 13:55

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020709

Tgt Ion: 93 Resp: 1045385




Ion	Ratio	Lower	Upper
93	100		
95	32.6	26.3	39.5
123	10.8	8.8	13.2




#29
 2,4-Dichlorophenol
 Concen: 19.120 ng/ul
 RT: 10.493 min Scan# 1225
 Delta R.T. 0.000 min
 Lab File: BP016830.D
 Acq: 29 Aug 2023 13:55

Tgt Ion:162 Resp: 699117

Ion	Ratio	Lower	Upper
162	100		
164	65.4	52.2	78.2
98	36.1	29.3	43.9









#30
Naphthalene
Concen: 18.513 ng/ul
RT: 10.910 min Scan# 1296
Delta R.T. 0.000 min
Lab File: BP016830.D
Acq: 29 Aug 2023 13:55




Instrument :
BNA_P
ClientSampleId :
SSTD020709


Tgt Ion:128 Resp: 2349453
Ion Ratio Lower Upper
128 100
129 10.9 8.8 13.2
127 13.0 10.3 15.5



#32
4-Chloroaniline
Concen: 17.838 ng/ul
RT: 11.040 min Scan# 1318
Delta R.T. 0.000 min
Lab File: BP016830.D
Acq: 29 Aug 2023 13:55

Tgt Ion:127 Resp: 978542
Ion Ratio Lower Upper
127 100
129 31.7 25.9 38.9





#33

Hexachlorobutadiene

Concen: 18.521 ng/ul

RT: 11.140 min Scan# 11

Delta R.T. -0.006 min

Lab File: BP016830.D

Acq: 29 Aug 2023 13:55

Instrument :

BNA_P

ClientSampleId :

SSTD020709





Tgt Ion:225 Resp: 417156

Ion Ratio Lower Upper

225 100

223 62.5 51.0 76.4

227 65.5 52.5 78.7



#34

Caprolactam

Concen: 16.675 ng/ul

RT: 11.863 min Scan# 1458

Delta R.T. 0.006 min

Lab File: BP016830.D

Acq: 29 Aug 2023 13:55




Tgt Ion:113 Resp: 224853


Ion Ratio Lower Upper

113 100

55 168.1 140.6 210.8

56 125.8 104.5 156.7









#35
4-Chloro-3-methylphenol
Concen: 18.428 ng/ul
RT: 12.134 min Scan# 1504
Delta R.T. 0.000 min
Lab File: BP016830.D
Acq: 29 Aug 2023 13:55




Instrument :
BNA_P
ClientSampleId :
SSTD020709


Tgt Ion:107 Resp: 755619
Ion Ratio Lower Upper
107 100
144 26.1 20.6 31.0
142 80.9 63.7 95.5



#36
2-Methylnaphthalene
Concen: 18.604 ng/ul
RT: 12.510 min Scan# 1568
Delta R.T. 0.000 min
Lab File: BP016830.D
Acq: 29 Aug 2023 13:55

Tgt Ion:142 Resp: 1629690
Ion Ratio Lower Upper
142 100
141 89.4 71.4 107.0





#37

1-Methylnaphthalene

Concen: 18.595 ng/ul

RT: 12.728 min Scan# 1605

Delta R.T. -0.006 min

Lab File: BP016830.D

Acq: 29 Aug 2023 13:55

Instrument :

BNA_P

ClientSampleId :

SSTD020709





Tgt Ion:142 Resp: 1642558

Ion Ratio Lower Upper

142 100

141 93.5 73.0 109.4

116 3.9 3.1 4.7



#39

1,2,4,5-Tetrachlorobenzene

Concen: 19.045 ng/ul

RT: 12.857 min Scan# 1627

Delta R.T. -0.006 min

Lab File: BP016830.D

Acq: 29 Aug 2023 13:55

Tgt Ion:216 Resp: 823043




Ion Ratio Lower Upper


216 100

214 78.5 63.4 95.2

179 21.6 17.4 26.2

108 18.3 15.3 22.9





#40

Hexachlorocyclopentadiene

Concen: 20.323 ng/ul

RT: 12.810 min Scan# 1619

Delta R.T. 0.000 min

Lab File: BP016830.D

Acq: 29 Aug 2023 13:55

Instrument :

BNA_P

ClientSampleId :

SSTD020709





Tgt Ion:237 Resp: 645893

Ion Ratio Lower Upper

237 100

235 62.2 51.3 76.9

272 12.4 9.8 14.8



#41

2,4,6-Trichlorophenol

Concen: 19.284 ng/ul

RT: 13.110 min Scan# 1670

Delta R.T. 0.000 min

Lab File: BP016830.D

Acq: 29 Aug 2023 13:55




Tgt Ion:196 Resp: 556525


Ion Ratio Lower Upper

196 100

198 97.3 76.6 115.0

200 30.7 25.2 37.8





#42

2,4,5-Trichlorophenol

Concen: 19.016 ng/ul

RT: 13.175 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP016830.D

Acq: 29 Aug 2023 13:55

Instrument :

BNA_P

ClientSampleId :

SSTD020709





Tgt Ion:196 Resp: 593839

Ion Ratio Lower Upper

196 100

198 94.0 77.8 116.8

200 30.2 24.8 37.2



#43

1,1'-Biphenyl

Concen: 18.988 ng/ul

RT: 13.516 min Scan# 1739

Delta R.T. -0.006 min

Lab File: BP016830.D

Acq: 29 Aug 2023 13:55




Tgt Ion:154 Resp: 2140674


Ion Ratio Lower Upper

154 100

153 40.5 32.7 49.1

76 14.0 11.3 16.9





#44

2-Chloronaphthalene

Concen: 18.882 ng/ul

RT: 13.563 min Scan# 1747

Delta R.T. -0.006 min

Lab File: BP016830.D

Acq: 29 Aug 2023 13:55

Instrument :

BNA_P

ClientSampleId :

SSTD020709





Tgt Ion:162 Resp: 1680417

Ion Ratio Lower Upper

162 100

164 33.8 26.1 39.1

127 38.5 30.8 46.2



#45

2-Nitroaniline

Concen: 18.475 ng/ul

RT: 13.792 min Scan# 1786

Delta R.T. -0.006 min

Lab File: BP016830.D

Acq: 29 Aug 2023 13:55




Tgt Ion: 65 Resp: 498627


Ion Ratio Lower Upper

65 100

92 66.6 53.8 80.8

138 101.2 80.7 121.1









#47
Dimethylphthalate
Concen: 18.159 ng/ul
RT: 14.139 min Scan# 1845
Delta R.T. -0.006 min
Lab File: BP016830.D
Acq: 29 Aug 2023 13:55




Instrument :
BNA_P
ClientSampleId :
SSTD020709


Tgt Ion:163 Resp: 2102942
Ion Ratio Lower Upper
163 100
194 3.9 3.2 4.8
164 10.4 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 18.847 ng/ul
RT: 14.287 min Scan# 1870
Delta R.T. 0.000 min
Lab File: BP016830.D
Acq: 29 Aug 2023 13:55

Tgt Ion:165 Resp: 427685
Ion Ratio Lower Upper
165 100
89 55.5 48.0 72.0
121 20.4 16.9 25.3









#50
Acenaphthylene
Concen: 18.785 ng/ul
RT: 14.410 min Scan# 1891
Delta R.T. -0.006 min
Lab File: BP016830.D
Acq: 29 Aug 2023 13:55




Instrument :
BNA_P
ClientSampleId :
SSTD020709


Tgt Ion:152 Resp: 2779274
Ion Ratio Lower Upper
152 100
151 19.9 16.1 24.1
153 13.4 10.6 15.8



#51
3-Nitroaniline
Concen: 17.707 ng/ul
RT: 14.622 min Scan# 1927
Delta R.T. 0.000 min
Lab File: BP016830.D
Acq: 29 Aug 2023 13:55

Tgt Ion:138 Resp: 418396
Ion Ratio Lower Upper
138 100
108 10.3 8.2 12.4
92 116.0 95.4 143.0









#52
Acenaphthene
Concen: 18.438 ng/ul
RT: 14.745 min Scan# 1948
Delta R.T. -0.006 min
Lab File: BP016830.D
Acq: 29 Aug 2023 13:55




Instrument :
BNA_P
ClientSampleId :
SSTD020709


Tgt Ion:153 Resp: 1804750
Ion Ratio Lower Upper
153 100
152 47.5 38.5 57.7
154 88.0 70.6 105.8



#53
2,4-Dinitrophenol
Concen: 15.800 ng/ul
RT: 14.822 min Scan# 1961
Delta R.T. 0.000 min
Lab File: BP016830.D
Acq: 29 Aug 2023 13:55

Tgt Ion:184 Resp: 260535
Ion Ratio Lower Upper
184 100
63 66.8 55.3 82.9
154 62.7 50.3 75.5









#55
4-Nitrophenol
Concen: 14.940 ng/ul
RT: 14.904 min Scan# 1975
Delta R.T. 0.000 min
Lab File: BP016830.D
Acq: 29 Aug 2023 13:55




Instrument :
BNA_P
ClientSampleId :
SSTD020709


Tgt Ion:109 Resp: 266676
Ion Ratio Lower Upper
109 100
139 134.0 108.6 163.0
65 137.2 107.8 161.6



#56
Dibenzofuran
Concen: 18.386 ng/ul
RT: 15.081 min Scan# 2005
Delta R.T. -0.006 min
Lab File: BP016830.D
Acq: 29 Aug 2023 13:55

Tgt Ion:168 Resp: 2469326
Ion Ratio Lower Upper
168 100
139 37.1 29.2 43.8





#57

2,4-Dinitrotoluene

Concen: 18.049 ng/ul

RT: 15.063 min Scan# 2002

Delta R.T. -0.006 min

Lab File: BP016830.D

Acq: 29 Aug 2023 13:55

Instrument :

BNA_P

ClientSampleId :

SSTD020709





Tgt Ion:165 Resp: 602845

Ion Ratio Lower Upper

165 100

63 48.3 39.0 58.4

182 2.6 2.2 3.2



#58

2,3,4,6-Tetrachlorophenol

Concen: 18.143 ng/ul

RT: 15.298 min Scan# 2042

Delta R.T. 0.000 min

Lab File: BP016830.D

Acq: 29 Aug 2023 13:55

Tgt Ion:232 Resp: 482344




Ion Ratio Lower Upper


232 100

131 44.2 38.1 57.1

130 2.3 1.9 2.9

166 31.6 26.5 39.7









#59
Diethylphthalate
Concen: 17.986 ng/ul
RT: 15.492 min Scan# 2075
Delta R.T. -0.006 min
Lab File: BP016830.D
Acq: 29 Aug 2023 13:55




Instrument :
BNA_P
ClientSampleId :
SSTD020709


Tgt Ion:149 Resp: 2115977
Ion Ratio Lower Upper
149 100
177 21.5 17.3 25.9
150 12.4 10.2 15.2



#61
Fluorene
Concen: 18.264 ng/ul
RT: 15.734 min Scan# 2116
Delta R.T. 0.000 min
Lab File: BP016830.D
Acq: 29 Aug 2023 13:55

Tgt Ion:166 Resp: 2051703
Ion Ratio Lower Upper
166 100
165 98.0 78.8 118.2
167 13.5 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 18.316 ng/ul
RT: 15.722 min Scan# 2114
Delta R.T. -0.006 min
Lab File: BP016830.D
Acq: 29 Aug 2023 13:55




Instrument :
BNA_P
ClientSampleId :
SSTD020709


Tgt Ion:204 Resp: 1010194
Ion Ratio Lower Upper
204 100
206 32.7 26.7 40.1
141 58.4 46.8 70.2



#63
4-Nitroaniline
Concen: 16.453 ng/ul
RT: 15.792 min Scan# 2126
Delta R.T. 0.000 min
Lab File: BP016830.D
Acq: 29 Aug 2023 13:55

Tgt Ion:138 Resp: 369930
Ion Ratio Lower Upper
138 100
92 51.9 41.6 62.4
108 90.4 74.9 112.3









#66
4,6-Dinitro-2-methylphenol
Concen: 15.815 ng/ul
RT: 15.816 min Scan# 2130
Delta R.T. 0.000 min
Lab File: BP016830.D
Acq: 29 Aug 2023 13:55




Instrument :
BNA_P
ClientSampleId :
SSTD020709


Tgt Ion:198 Resp: 319271
Ion Ratio Lower Upper
198 100
182 4.5 3.2 4.8
77 27.0 20.7 31.1



#67
N-Nitrosodiphenylamine
Concen: 19.568 ng/ul
RT: 15.945 min Scan# 2152
Delta R.T. -0.006 min
Lab File: BP016830.D
Acq: 29 Aug 2023 13:55

Tgt Ion:169 Resp: 1701662
Ion Ratio Lower Upper
169 100
168 67.8 54.2 81.2
167 36.9 29.4 44.2









#68
4-Bromophenyl-phenylether
Concen: 19.314 ng/ul
RT: 16.628 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP016830.D
Acq: 29 Aug 2023 13:55




Instrument :
BNA_P
ClientSampleId :
SSTD020709


Tgt Ion:248 Resp: 582237
Ion Ratio Lower Upper
248 100
250 98.2 75.4 113.2
141 74.3 58.7 88.1



#69
Hexachlorobenzene
Concen: 18.778 ng/ul
RT: 16.710 min Scan# 2282
Delta R.T. -0.006 min
Lab File: BP016830.D
Acq: 29 Aug 2023 13:55

Tgt Ion:284 Resp: 633806
Ion Ratio Lower Upper
284 100
142 37.9 29.8 44.8
249 32.0 25.9 38.9





#70

Atrazine

Concen: 18.634 ng/ul

RT: 16.904 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP016830.D

Acq: 29 Aug 2023 13:55

Instrument :

BNA_P

ClientSampleId :

SSTD020709





Tgt Ion:200 Resp: 610921

Ion Ratio Lower Upper

200 100

173 26.2 21.4 32.2

215 47.8 37.3 55.9



#71

Pentachlorophenol

Concen: 17.257 ng/ul

RT: 17.069 min Scan# 2343

Delta R.T. -0.006 min

Lab File: BP016830.D

Acq: 29 Aug 2023 13:55




Tgt Ion:266 Resp: 432381


Ion Ratio Lower Upper

266 100

264 62.6 50.7 76.1

268 62.8 50.8 76.2





#72

Phenanthrene

Concen: 18.569 ng/ul

RT: 17.492 min Scan# 2415

Delta R.T. -0.006 min

Lab File: BP016830.D

Acq: 29 Aug 2023 13:55

Instrument :

BNA_P

ClientSampleId :

SSTD020709





Tgt Ion:178 Resp: 3073699

Ion Ratio Lower Upper

178 100

179 15.7 12.4 18.6

176 19.7 15.6 23.4



#74

Anthracene

Concen: 18.492 ng/ul

RT: 17.586 min Scan# 2431

Delta R.T. -0.006 min

Lab File: BP016830.D

Acq: 29 Aug 2023 13:55




Tgt Ion:178 Resp: 3094967


Ion Ratio Lower Upper

178 100

179 16.0 12.4 18.6

176 19.2 15.3 22.9









#75
1,2,3,4-Tetrachlorobenzene
Concen: 20.635 ng/uL
RT: 13.469 min Scan# 1731
Delta R.T. -0.006 min
Lab File: BP016830.D
Acq: 29 Aug 2023 13:55




Instrument :
BNA_P
ClientSampleId :
SSTD020709


Tgt Ion:216 Resp: 868730
Ion Ratio Lower Upper
216 100
214 79.6 64.0 96.0
218 47.5 38.2 57.2



#76
Pentachlorobenzene
Concen: 20.038 ng/uL
RT: 14.975 min Scan# 1987
Delta R.T. -0.006 min
Lab File: BP016830.D
Acq: 29 Aug 2023 13:55

Tgt Ion:250 Resp: 753905
Ion Ratio Lower Upper
250 100
248 62.9 50.9 76.3
252 63.1 52.2 78.4









#77
 Carbazole
 Concen: 17.840 ng/ul
 RT: 17.869 min Scan# 2479
 Delta R.T. 0.000 min
 Lab File: BP016830.D
 Acq: 29 Aug 2023 13:55




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020709


Tgt Ion:167 Resp: 2689187
 Ion Ratio Lower Upper
 167 100
 166 21.4 17.4 26.2
 139 12.8 10.4 15.6



#78
 Di-n-butylphthalate
 Concen: 18.662 ng/ul
 RT: 18.375 min Scan# 2565
 Delta R.T. -0.006 min
 Lab File: BP016830.D
 Acq: 29 Aug 2023 13:55

Tgt Ion:149 Resp: 3509522
 Ion Ratio Lower Upper
 149 100
 150 9.4 7.5 11.3
 104 5.3 4.1 6.1





#80

Fluoranthene

Concen: 23.915 ng/ul

RT: 19.451 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP016830.D

Acq: 29 Aug 2023 13:55

Instrument :

BNA_P

ClientSampleId :

SSTD020709





Tgt Ion:202 Resp: 3211189

Ion Ratio Lower Upper

202 100

101 11.9 9.7 14.5

100 9.2 7.4 11.2



#82

Pyrene

Concen: 23.354 ng/ul

RT: 19.810 min Scan# 2809

Delta R.T. -0.006 min

Lab File: BP016830.D

Acq: 29 Aug 2023 13:55




Tgt Ion:202 Resp: 3239272


Ion Ratio Lower Upper

202 100

101 13.8 11.4 17.2

100 11.5 9.4 14.2





#83

Butylbenzylphthalate

Concen: 21.977 ng/ul

RT: 20.669 min Scan# 2955

Delta R.T. -0.006 min

Lab File: BP016830.D

Acq: 29 Aug 2023 13:55

Instrument :

BNA_P

ClientSampleId :

SSTD020709





Tgt Ion:149 Resp: 1314608

Ion Ratio Lower Upper

149 100

91 71.5 58.5 87.7

206 20.4 16.1 24.1



#84

3,3'-Dichlorobenzidine

Concen: 17.199 ng/ul

RT: 21.468 min Scan# 3091

Delta R.T. -0.006 min

Lab File: BP016830.D

Acq: 29 Aug 2023 13:55




Tgt Ion:252 Resp: 748193


Ion Ratio Lower Upper

252 100

254 65.7 52.1 78.1

126 11.3 9.6 14.4





#85

Benzo(a)anthracene

Concen: 18.708 ng/ul

RT: 21.527 min Scan# 3101

Delta R.T. -0.006 min

Lab File: BP016830.D

Acq: 29 Aug 2023 13:55

Instrument :

BNA_P

ClientSampleId :

SSTD020709





Tgt Ion:228 Resp: 2622713

Ion Ratio Lower Upper

228 100

229 20.4 16.2 24.4

226 28.1 22.9 34.3



#86

Bis(2-ethylhexyl)phthalate

Concen: 21.291 ng/ul

RT: 21.404 min Scan# 3080

Delta R.T. -0.006 min

Lab File: BP016830.D

Acq: 29 Aug 2023 13:55




Tgt Ion:149 Resp: 1859625


Ion Ratio Lower Upper

149 100

167 27.3 22.1 33.1

279 3.7 3.0 4.4





#87

Chrysene

Concen: 18.769 ng/ul

RT: 21.586 min Scan# 3111

Delta R.T. -0.006 min

Lab File: BP016830.D

Acq: 29 Aug 2023 13:55

Instrument :

BNA_P

ClientSampleId :

SSTD020709





Tgt Ion:228 Resp: 2390181

Ion Ratio Lower Upper

228 100

226 30.5 25.0 37.4

229 19.5 16.0 24.0



#89

Di-n-octyl phthalate

Concen: 18.046 ng/ul




RT: 22.380 min Scan# 3246


Delta R.T. -0.012 min

Lab File: BP016830.D

Acq: 29 Aug 2023 13:55

Tgt Ion:149 Resp: 2678967









#90
Benzo(b)fluoranthene
Concen: 17.825 ng/ul
RT: 23.310 min Scan# 3404
Delta R.T. -0.012 min
Lab File: BP016830.D
Acq: 29 Aug 2023 13:55




Instrument :
BNA_P
ClientSampleId :
SSTD020709


Tgt Ion:252 Resp: 2243799
Ion Ratio Lower Upper
252 100
253 22.2 17.7 26.5
125 11.4 9.0 13.4



#91
Benzo(k)fluoranthene
Concen: 18.126 ng/ul
RT: 23.363 min Scan# 3413
Delta R.T. -0.012 min
Lab File: BP016830.D
Acq: 29 Aug 2023 13:55

Tgt Ion:252 Resp: 2256645
Ion Ratio Lower Upper
252 100
253 21.8 18.1 27.1
125 11.1 8.9 13.3





#93

Benzo(a)pyrene

Concen: 18.548 ng/ul

RT: 23.998 min Scan# 3521

Delta R.T. -0.012 min

Lab File: BP016830.D

Acq: 29 Aug 2023 13:55

Instrument :

BNA_P

ClientSampleId :

SSTD020709





Tgt Ion:252 Resp: 2111967

Ion Ratio Lower Upper

252 100

253 21.9 17.8 26.8

125 11.8 9.9 14.9



#94

Indeno(1,2,3-cd)pyrene

Concen: 21.749 ng/ul

RT: 26.845 min Scan# 4005

Delta R.T. -0.018 min

Lab File: BP016830.D

Acq: 29 Aug 2023 13:55




Tgt Ion:276 Resp: 2530552


Ion Ratio Lower Upper

276 100

138 25.7 20.9 31.3

277 24.9 19.8 29.6









#95
Dibenzo(a,h)anthracene
Concen: 21.599 ng/ul
RT: 26.868 min Scan# 4009
Delta R.T. -0.018 min
Lab File: BP016830.D
Acq: 29 Aug 2023 13:55

Instrument :
BNA_P
ClientSampleId :
SSTD020709

Tgt Ion:278 Resp: 2087269
Ion Ratio Lower Upper
278 100
139 18.2 15.0 22.6
279 23.8 19.4 29.2



#96
Benzo(g,h,i)perylene
Concen: 22.918 ng/ul
RT: 27.703 min Scan# 4151
Delta R.T. -0.018 min
Lab File: BP016830.D
Acq: 29 Aug 2023 13:55

Tgt Ion:276 Resp: 2021615
Ion Ratio Lower Upper
276 100
138 24.5 20.2 30.4
277 23.3 19.3 28.9

