

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082823\
 Data File : BP016813.D
 Acq On : 29 Aug 2023 03:45
 Operator : MA/JU
 Sample : PB155101BL
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK101

Quant Time: Aug 29 05:04:47 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP082423.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 29 01:12:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.028	152	374196	20.000	ng/u1	0.00
20) Naphthalene-d8	10.857	136	1670627	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.681	164	1073806	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.451	188	2391986	20.000	ng/u1	0.00
79) Chrysene-d12	21.545	240	1914421	20.000	ng/u1	0.00
88) Perylene-d12	24.110	264	1686161	20.000	ng/u1	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.381	96	78874	8.711	ng/uL	0.00
4) Pyridine-d5	3.817	84	953308	34.455	ng/u1	0.00
7) Phenol-d5	7.169	99	1129951	33.959	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.364	67	704403	33.216	ng/u1	0.00
11) 2-Chlorophenol-d4	7.546	132	854074	34.139	ng/u1	0.00
15) 4-Methylphenol-d8	8.734	113	901932	33.205	ng/u1	0.00
21) Nitrobenzene-d5	9.228	128	435180	34.452	ng/u1	0.00
24) 2-Nitrophenol-d4	9.940	143	476191	35.765	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.463	165	843068	33.964	ng/u1	0.00
31) 4-Chloroaniline-d4	11.016	131	1251670	32.881	ng/u1	0.00
46) Dimethylphthalate-d6	14.093	166	2774317	34.259	ng/u1	0.00
49) Acenaphthylene-d8	14.381	160	3179413	34.731	ng/u1	0.00
54) 4-Nitrophenol-d4	14.887	143	452510	27.035	ng/u1	0.00
60) Fluorene-d10	15.675	176	2445368	35.859	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.798	200	334846	23.506	ng/u1	0.00
73) Anthracene-d10	17.551	188	3906454	36.844	ng/u1	0.00
81) Pyrene-d10	19.781	212	4394350	41.794	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.945	264	3134690	37.114	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082823\
Data File : BP016813.D
Acq On : 29 Aug 2023 03:45
Operator : MA/JU
Sample : PB155101BL
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK101

Quant Time: Aug 29 05:04:47 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP082423.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Aug 29 01:12:28 2023
Response via : Initial Calibration

