

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP083019\
 Data File : BP000018.D
 Acq On : 30 Aug 2019 23:13
 Operator : HP/JU
 Sample : MDL-S-BL03
 Misc : 4PPM
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 MDL-S-BL03

Quant Time: Aug 31 01:21:24 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP083019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Aug 31 01:09:25 2019
 Response via : Initial Calibration

| Internal Standards | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|-----------------------------|-------|------|----------|--------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4 | 6.89 | 152 | 456542 | 20.00 | ng | 0.00 |
| 21) Naphthalene-d8 | 8.18 | 136 | 1667612 | 20.00 | ng | 0.00 |
| 39) Acenaphthene-d10 | 9.94 | 164 | 895419 | 20.00 | ng | 0.00 |
| 64) Phenanthrene-d10 | 11.43 | 188 | 1633953 | 20.00 | ng | 0.00 |
| 76) Chrysene-d12 | 14.10 | 240 | 1300718 | 20.00 | ng | 0.00 |
| 87) Perylene-d12 | 15.63 | 264 | 1351925 | 20.00 | ng | 0.00 |
| System Monitoring Compounds | | | | | | |
| 5) 2-Fluorophenol | 5.52 | 112 | 3706064 | 129.45 | ng | 0.01 |
| 7) Phenol-d6 | 6.52 | 99 | 4654111 | 128.77 | ng | 0.00 |
| 23) Nitrobenzene-d5 | 7.45 | 82 | 2608412 | 95.98 | ng | 0.00 |
| 42) 2,4,6-Tribromophenol | 10.73 | 330 | 1119137 | 148.28 | ng | 0.00 |
| 45) 2-Fluorobiphenyl | 9.26 | 172 | 4959725 | 92.50 | ng | 0.00 |
| 79) Terphenyl-d14 | 13.05 | 244 | 5625764 | 94.27 | ng | 0.00 |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP083019\
 Data File : BP000018.D
 Acq On : 30 Aug 2019 23:13
 Operator : HP/JU
 Sample : MDL-S-BL03
 Misc : 4PPM
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 MDL-S-BL03

Quant Time: Aug 31 01:21:24 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP083019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Aug 31 01:09:25 2019
 Response via : Initial Calibration

