

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP083123\  
 Data File : BP016898.D  
 Acq On : 31 Aug 2023 15:18  
 Operator : MA/JU  
 Sample : 04113-03  
 Misc :  
 ALS Vial : 12 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 EZYH6

Quant Time: Sep 01 00:55:09 2023  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\SFAM-EPA-BP082423.MA.M  
 Quant Title : SVOA CALIBRATION  
 QLast Update : Wed Aug 30 02:25:00 2023  
 Response via : Initial Calibration


| Compound                      | R.T.   | QIon | Response | Conc   | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|--------|--------|----------|
| Internal Standards            |        |      |          |        |        |          |
| 1) 1,4-Dichlorobenzene-d4     | 8.028  | 152  | 492950   | 20.000 | ng/ul  | 0.00     |
| 20) Naphthalene-d8            | 10.857 | 136  | 1989431  | 20.000 | ng/ul  | 0.00     |
| 38) Acenaphthene-d10          | 14.681 | 164  | 1037169  | 20.000 | ng/ul  | 0.00     |
| 64) Phenanthrene-d10          | 17.451 | 188  | 1719546  | 20.000 | ng/ul  | 0.00     |
| 79) Chrysene-d12              | 21.557 | 240  | 1420970  | 20.000 | ng/ul  | 0.02     |
| 88) Perylene-d12              | 24.133 | 264  | 1667383  | 20.000 | ng/ul  | 0.02     |
| System Monitoring Compounds   |        |      |          |        |        |          |
| 3) 1,4-Dioxane-d8             | 3.381  | 96   | 36325    | 3.046  | ng/uL  | 0.00     |
| 4) Pyridine-d5                | 3.828  | 84   | 33958    | 0.932  | ng/ul  | 0.02     |
| 7) Phenol-d5                  | 7.169  | 99   | 666417   | 15.203 | ng/ul  | 0.00     |
| 9) Bis-(2-Chloroethyl)eth...  | 7.363  | 67   | 449009   | 16.072 | ng/ul  | 0.00     |
| 11) 2-Chlorophenol-d4         | 7.546  | 132  | 536219   | 16.270 | ng/ul  | 0.00     |
| 15) 4-Methylphenol-d8         | 8.734  | 113  | 357600   | 9.994  | ng/ul  | 0.00     |
| 21) Nitrobenzene-d5           | 9.222  | 128  | 263657   | 17.528 | ng/ul  | 0.00     |
| 24) 2-Nitrophenol-d4          | 9.940  | 143  | 266143   | 16.786 | ng/ul  | 0.00     |
| 28) 2,4-Dichlorophenol-d3     | 10.463 | 165  | 446761   | 15.114 | ng/ul  | 0.00     |
| 31) 4-Chloroaniline-d4        | 11.016 | 131  | 144319   | 3.184  | ng/ul  | 0.00     |
| 46) Dimethylphthalate-d6      | 14.092 | 166  | 1369213  | 17.505 | ng/ul  | 0.00     |
| 49) Acenaphthylene-d8         | 14.381 | 160  | 1549135  | 17.520 | ng/ul  | 0.00     |
| 54) 4-Nitrophenol-d4          | 14.898 | 143  | 123802   | 7.658  | ng/ul  | 0.01     |
| 60) Fluorene-d10              | 15.675 | 176  | 1106160  | 16.794 | ng/ul  | 0.00     |
| 65) 4,6-Dinitro-2-methylph... | 15.810 | 200  | 18069    | 1.764  | ng/ul  | 0.01     |
| 73) Anthracene-d10            | 17.551 | 188  | 1410319  | 18.503 | ng/ul  | 0.00     |
| 81) Pyrene-d10                | 19.786 | 212  | 1436285  | 18.404 | ng/ul  | 0.01     |
| 92) Benzo(a)pyrene-d12        | 23.962 | 264  | 1507010  | 18.043 | ng/ul  | 0.02     |
| Target Compounds              |        |      |          |        |        |          |
|                               |        |      |          |        | Qvalue |          |
| 16) Acetophenone              | 8.875  | 105  | 165507   | 2.983  | ng/ul  | 99       |
| 30) Naphthalene               | 10.904 | 128  | 205719   | 1.996  | ng/ul  | 99       |
| 36) 2-Methylnaphthalene       | 12.510 | 142  | 115429   | 1.623  | ng/ul  | 100      |
| 72) Phenanthrene              | 17.492 | 178  | 317890   | 3.512  | ng/ul  | 99       |
| 74) Anthracene                | 17.492 | 178  | 318199   | 3.476  | ng/ul  | 98       |
| 80) Fluoranthene              | 19.457 | 202  | 485858   | 5.183  | ng/ul  | 98       |
| 82) Pyrene                    | 19.816 | 202  | 554689   | 5.728  | ng/ul  | 100      |
| 85) Benzo(a)anthracene        | 21.539 | 228  | 270296   | 2.762  | ng/ul# | 89       |
| 87) Chrysene                  | 21.592 | 228  | 245971   | 2.767  | ng/ul# | 87       |
| 90) Benzo(b)fluoranthene      | 23.327 | 252  | 478738   | 4.496  | ng/ul# | 72       |
| 91) Benzo(k)fluoranthene      | 23.380 | 252  | 126233   | 1.199  | ng/ul# | 47       |
| 93) Benzo(a)pyrene            | 24.015 | 252  | 301091   | 3.126  | ng/ul# | 70       |
| 94) Indeno(1,2,3-cd)pyrene    | 26.868 | 276  | 230257   | 2.340  | ng/ul  | 92       |
| 96) Benzo(g,h,i)perylene      | 27.727 | 276  | 246977   | 3.310  | ng/ul# | 88       |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP083123\  
Data File : BP016898.D  
Acq On : 31 Aug 2023 15:18  
Operator : MA/JU  
Sample : 04113-03  
Misc :  
ALS Vial : 12 Sample Multiplier: 1

Instrument :  
BNA\_P  
ClientSampleId :  
EZYH6

Quant Time: Sep 01 00:55:09 2023  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\SFAM-EPA-BP082423.MA.M  
Quant Title : SVOA CALIBRATION  
QLast Update : Wed Aug 30 02:25:00 2023  
Response via : Initial Calibration









#16  
Acetophenone  
Concen: 2.983 ng/ul  
RT: 8.875 min Scan# 91  
Delta R.T. 0.000 min  
Lab File: BP016898.D  
Acq: 31 Aug 2023 15:18




Instrument :  
BNA\_P  
ClientSampleId :  
EZYH6


Tgt Ion:105 Resp: 165507  
Ion Ratio Lower Upper  
105 100  
77 81.1 64.2 96.4  
51 28.3 22.2 33.4



#30  
Naphthalene  
Concen: 1.996 ng/ul  
RT: 10.904 min Scan# 1295  
Delta R.T. 0.000 min  
Lab File: BP016898.D  
Acq: 31 Aug 2023 15:18

Tgt Ion:128 Resp: 205719  
Ion Ratio Lower Upper  
128 100  
129 11.7 8.8 13.2  
127 12.8 10.3 15.5









#36  
2-Methylnaphthalene  
Concen: 1.623 ng/ul  
RT: 12.510 min Scan# 11  
Delta R.T. 0.006 min  
Lab File: BP016898.D  
Acq: 31 Aug 2023 15:18




Instrument :  
BNA\_P  
ClientSampleId :  
EZYH6


Tgt Ion:142 Resp: 115429  
Ion Ratio Lower Upper  
142 100  
141 89.1 71.4 107.0



#72  
Phenanthrene  
Concen: 3.512 ng/ul  
RT: 17.492 min Scan# 2415  
Delta R.T. 0.006 min  
Lab File: BP016898.D  
Acq: 31 Aug 2023 15:18

Tgt Ion:178 Resp: 317890  
Ion Ratio Lower Upper  
178 100  
179 16.6 12.4 18.6  
176 19.5 15.6 23.4





#74

Anthracene

Concen: 3.476 ng/ul

RT: 17.492 min Scan# 2430

Delta R.T. -0.088 min

Lab File: BP016898.D

Acq: 31 Aug 2023 15:18

Instrument :

BNA\_P

ClientSampleId :

EZYH6





Tgt Ion:178 Resp: 318199

Ion Ratio Lower Upper

178 100

179 16.6 12.4 18.6

176 19.5 15.3 22.9



#80

Fluoranthene

Concen: 5.183 ng/ul

RT: 19.457 min Scan# 2749

Delta R.T. 0.006 min

Lab File: BP016898.D

Acq: 31 Aug 2023 15:18




Tgt Ion:202 Resp: 485858


Ion Ratio Lower Upper

202 100

101 12.9 9.7 14.5

100 9.7 7.4 11.2





#82

Pyrene

Concen: 5.728 ng/ul

RT: 19.816 min Scan# 21

Delta R.T. 0.006 min

Lab File: BP016898.D

Acq: 31 Aug 2023 15:18

Instrument :

BNA\_P

ClientSampleId :

EZYH6





Tgt Ion:202 Resp: 554689

Ion Ratio Lower Upper

202 100

101 14.6 11.4 17.2

100 11.8 9.4 14.2



#85

Benzo(a)anthracene

Concen: 2.762 ng/ul

RT: 21.539 min Scan# 3103

Delta R.T. 0.012 min

Lab File: BP016898.D

Acq: 31 Aug 2023 15:18




Tgt Ion:228 Resp: 270296


Ion Ratio Lower Upper

228 100

229 28.9 16.2 24.4#

226 31.9 22.9 34.3









#87  
Chrysene  
Concen: 2.767 ng/ul  
RT: 21.592 min Scan# 3110  
Delta R.T. 0.012 min  
Lab File: BP016898.D  
Acq: 31 Aug 2023 15:18




Instrument :  
BNA\_P  
ClientSampleId :  
EZYH6


Tgt Ion:228 Resp: 245971  
Ion Ratio Lower Upper  
228 100  
226 34.1 25.0 37.4  
229 31.4 16.0 24.0#



#90  
Benzo(b)fluoranthene  
Concen: 4.496 ng/ul  
RT: 23.327 min Scan# 3407  
Delta R.T. 0.018 min  
Lab File: BP016898.D  
Acq: 31 Aug 2023 15:18

Tgt Ion:252 Resp: 478738  
Ion Ratio Lower Upper  
252 100  
253 28.0 17.7 26.5#  
125 33.7 9.0 13.4#





#91

Benzo(k)fluoranthene

Concen: 1.199 ng/ul

RT: 23.380 min Scan# 3412

Delta R.T. 0.018 min

Lab File: BP016898.D

Acq: 31 Aug 2023 15:18

Instrument :

BNA\_P

ClientSampleId :

EZYH6





Tgt Ion:252 Resp: 126233

Ion Ratio Lower Upper

252 100

253 35.2 18.1 27.1#

125 51.6 8.9 13.3#



#93

Benzo(a)pyrene

Concen: 3.126 ng/ul

RT: 24.015 min Scan# 3524

Delta R.T. 0.024 min

Lab File: BP016898.D

Acq: 31 Aug 2023 15:18




Tgt Ion:252 Resp: 301091


Ion Ratio Lower Upper

252 100

253 29.3 17.8 26.8#

125 35.4 9.9 14.9#





#94

Indeno(1,2,3-cd)pyrene

Concen: 2.340 ng/ul

RT: 26.868 min Scan# 4009

Delta R.T. 0.035 min

Lab File: BP016898.D

Acq: 31 Aug 2023 15:18

Instrument :

BNA\_P

ClientSampleId :

EZYH6





Tgt Ion:276 Resp: 230257

Ion Ratio Lower Upper

276 100

138 29.9 20.9 31.3

277 28.9 19.8 29.6



#96

Benzo(g,h,i)perylene

Concen: 3.310 ng/ul

RT: 27.727 min Scan# 4155

Delta R.T. 0.041 min

Lab File: BP016898.D

Acq: 31 Aug 2023 15:18

Tgt Ion:276 Resp: 246977

Ion Ratio Lower Upper

276 100

138 31.6 20.2 30.4#

277 30.1 19.3 28.9#

