



2-Nitrophenol



Response = 1.644e-001 * Amt - 4.019e-002
 Coef of Det (r^2) = 0.993429 Curve Fit: Linear
 Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP090222.M
 Calibration Table Last Updated: Sat Sep 03 04:07:09 2022


Benzoic acid



Response = $2.108 \times 10^{-1} \times \text{Amt} - 9.062 \times 10^{-2}$
 Coef of Det (r^2) = 0.992205 Curve Fit: Linear
 Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP090222.M
 Calibration Table Last Updated: Sat Sep 03 04:07:09 2022

2-Nitroaniline

Response Ratio



$$\text{Response} = 3.327\text{e-}001 * \text{Amt} - 5.659\text{e-}002$$


Coef of Det (r^2) = 0.998503 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP090222.M

Calibration Table Last Updated: Sat Sep 03 04:07:09 2022

2,4-Dinitrotoluene

Response Ratio




$$\text{Response} = 4.362\text{e-}001 * \text{Amt} - 7.947\text{e-}002$$

Coef of Det (r^2) = 0.998157 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP090222.M

Calibration Table Last Updated: Sat Sep 03 04:07:09 2022

2,4-Dinitrophenol




$$\text{Response} = 1.261\text{e-}001 * \text{Amt} - 4.068\text{e-}002$$

Coef of Det (r^2) = 0.992244 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP090222.M

Calibration Table Last Updated: Sat Sep 03 04:07:09 2022


4,6-Dinitro-2-methylphenol



Response = $1.043 \times 10^{-1} \times \text{Amt} - 4.115 \times 10^{-2}$
 Coef of Det (r^2) = 0.990786 Curve Fit: Linear
 Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP090222.M
 Calibration Table Last Updated: Sat Sep 03 04:07:09 2022

4-Nitroaniline

Response Ratio



$$\text{Response} = 3.704\text{e-}001 * \text{Amt} - 4.898\text{e-}002$$


Coef of Det (r^2) = 0.999304 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP090222.M

Calibration Table Last Updated: Sat Sep 03 04:07:09 2022

Bis(2-ethylhexyl)phthalate

Response Ratio



$$\text{Response} = 8.149\text{e-}001 * \text{Amt} - 7.715\text{e-}002$$


Coef of Det (r^2) = 0.999215 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP090222.M

Calibration Table Last Updated: Sat Sep 03 04:07:09 2022

Butylbenzylphthalate

Response Ratio



$$\text{Response} = 5.765\text{e-}001 * \text{Amt} - 1.070\text{e-}001$$

Coef of Det (r^2) = 0.998410 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP090222.M

Calibration Table Last Updated: Sat Sep 03 04:07:09 2022