

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090523\
 Data File : BP016949.D
 Acq On : 05 Sep 2023 11:32
 Operator : MA/JU
 Sample : SST00537
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST005645

Quant Time: Sep 06 01:17:08 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP090523.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Sep 06 01:15:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.981	152	325960	20.000	ng/ul	0.00
20) Naphthalene-d8	10.810	136	1358554	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.639	164	849195	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.410	188	1803166	20.000	ng/ul	0.00
79) Chrysene-d12	21.504	240	1530520	20.000	ng/ul	0.00
88) Perylene-d12	24.051	264	1482810	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.328	96	16804	1.950	ng/uL	0.00
4) Pyridine-d5	3.769	84	103968	4.222	ng/ul	0.00
7) Phenol-d5	7.128	99	122532	4.115	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.322	67	84373	4.530	ng/ul	0.00
11) 2-Chlorophenol-d4	7.505	132	92567	4.295	ng/ul	0.00
15) 4-Methylphenol-d8	8.687	113	92189	4.063	ng/ul	0.00
21) Nitrobenzene-d5	9.181	128	44890	4.259	ng/ul	0.00
24) 2-Nitrophenol-d4	9.899	143	46344	4.056	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.422	165	90149	4.274	ng/ul	0.00
31) 4-Chloroaniline-d4	10.969	131	138601	4.554	ng/ul	0.00
46) Dimethylphthalate-d6	14.051	166	289663	4.582	ng/ul	0.00
49) Acenaphthylene-d8	14.339	160	317418	4.351	ng/ul	0.00
54) 4-Nitrophenol-d4	14.851	143	44690	3.619	ng/ul	0.00
60) Fluorene-d10	15.634	176	246897	4.562	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.757	200	30582	2.818	ng/ul	0.00
73) Anthracene-d10	17.510	188	371560	4.487	ng/ul	0.00
81) Pyrene-d10	19.745	212	405507	4.350	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.874	264	325314	4.325	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.364	88	16456	1.769	ng/uL#	86
5) Pyridine	3.787	79	115693	4.414	ng/ul#	83
6) Benzaldehyde	7.140	77	35124	2.602	ng/ul	98
8) Phenol	7.157	94	129726	4.193	ng/ul	99
10) Bis(2-Chloroethyl)ether	7.416	93	112365	4.554	ng/ul	98
12) 2-Chlorophenol	7.534	128	99256	4.323	ng/ul	98
13) 2-Methylphenol	8.422	108	93343	4.113	ng/ul	100
14) 2,2'-oxybis(1-Chloropr...	8.493	45	89464	2.574	ng/ul	97
16) Acetophenone	8.828	105	164736	4.500	ng/ul#	92
17) N-Nitroso-di-n-propyla...	8.799	70	86756	4.436	ng/ul	100
18) 4-Methylphenol	8.752	108	104761	4.233	ng/ul	96
19) Hexachloroethane	9.051	117	42158	4.432	ng/ul	97
22) Nitrobenzene	9.222	77	125018	4.442	ng/ul	97
23) Isophorone	9.734	82	230839	4.366	ng/ul	98
25) 2-Nitrophenol	9.928	139	51990	4.133	ng/ul	99
26) 2,4-Dimethylphenol	9.963	107	113685	4.426	ng/ul	97
27) Bis(2-Chloroethoxy)met...	10.222	93	150324	4.588	ng/ul	100
29) 2,4-Dichlorophenol	10.446	162	93777	4.362	ng/ul	98
30) Naphthalene	10.863	128	346904	4.680	ng/ul	100
32) 4-Chloroaniline	10.993	127	148417	4.686	ng/ul	99
33) Hexachlorobutadiene	11.093	225	57619	4.485	ng/ul	97
34) Caprolactam	11.804	113	27279	3.831	ng/ul	95
35) 4-Chloro-3-methylphenol	12.092	107	101000	4.287	ng/ul	98
36) 2-Methylnaphthalene	12.463	142	222561	4.543	ng/ul	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090523\
 Data File : BP016949.D
 Acq On : 05 Sep 2023 11:32
 Operator : MA/JU
 Sample : SST00537
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST005645

Quant Time: Sep 06 01:17:08 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP090523.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Sep 06 01:15:58 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.687	142	227435	4.611	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.816	216	109159	4.363	ng/ul	99
40) Hexachlorocyclopentadiene	12.763	237	56060	3.458	ng/ul	97
41) 2,4,6-Trichlorophenol	13.069	196	71561	4.202	ng/ul	98
42) 2,4,5-Trichlorophenol	13.134	196	77369	4.176	ng/ul	99
43) 1,1'-Biphenyl	13.475	154	310278	4.616	ng/ul	98
44) 2-Chloronaphthalene	13.522	162	240439	4.556	ng/ul	99
45) 2-Nitroaniline	13.751	65	61695	3.925	ng/ul	98
47) Dimethylphthalate	14.098	163	303244	4.572	ng/ul	100
48) 2,6-Dinitrotoluene	14.239	165	49722	3.860	ng/ul	92
50) Acenaphthylene	14.369	152	362891	4.443	ng/ul	98
51) 3-Nitroaniline	14.581	138	42614	3.315	ng/ul	93
52) Acenaphthene	14.704	153	262715	4.610	ng/ul	98
53) 2,4-Dinitrophenol	14.781	184	21166	2.650	ng/ul	98
55) 4-Nitrophenol	14.863	109	36693	3.713	ng/ul	92
56) Dibenzofuran	15.039	168	364577	4.652	ng/ul	98
57) 2,4-Dinitrotoluene	15.022	165	69354	3.751	ng/ul	95
58) 2,3,4,6-Tetrachlorophenol	15.257	232	62258	4.168	ng/ul	98
59) Diethylphthalate	15.451	149	297269	4.485	ng/ul	99
61) Fluorene	15.692	166	296285	4.586	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.681	204	142329	4.559	ng/ul	96
63) 4-Nitroaniline	15.751	138	23526	2.082	ng/ul#	75
66) 4,6-Dinitro-2-methylph...	15.775	198	35073	3.085	ng/ul#	92
67) N-Nitrosodiphenylamine	15.904	169	243209	4.473	ng/ul	100
68) 4-Bromophenyl-phenylether	16.586	248	78626	4.383	ng/ul	98
69) Hexachlorobenzene	16.669	284	87801	4.465	ng/ul	97
70) Atrazine	16.857	200	83075	4.463	ng/ul	97
71) Pentachlorophenol	17.033	266	45630	3.558	ng/ul	98
72) Phenanthrene	17.451	178	460967	4.601	ng/ul	100
74) Anthracene	17.545	178	457421	4.541	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.428	216	113487	4.418	ng/ul	99
76) Pentachlorobenzene	14.934	250	113001	4.567	ng/ul	96
77) Carbazole	17.828	167	409421	4.523	ng/ul	99
78) Di-n-butylphthalate	18.339	149	490099	4.420	ng/ul	99
80) Fluoranthene	19.416	202	509191	4.374	ng/ul	100
82) Pyrene	19.774	202	532974	4.410	ng/ul	99
83) Butylbenzylphthalate	20.633	149	203488	4.071	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.427	252	119901	3.530	ng/ul	99
85) Benzo(a)anthracene	21.486	228	480860	4.461	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.368	149	283079	4.100	ng/ul	100
87) Chrysene	21.545	228	459488	4.557	ng/ul	99
89) Di-n-octyl phthalate	22.339	149	451809	3.635	ng/ul	100
90) Benzo(b)fluoranthene	23.257	252	441050	4.257	ng/ul	99
91) Benzo(k)fluoranthene	23.310	252	430990	4.296	ng/ul	99
93) Benzo(a)pyrene	23.933	252	367643	4.255	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.745	276	428764	4.470	ng/ul	99
95) Dibenzo(a,h)anthracene	26.768	278	353178	4.461	ng/ul	97
96) Benzo(g,h,i)perylene	27.586	276	344423	4.557	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090523\
 Data File : BP016949.D
 Acq On : 05 Sep 2023 11:32
 Operator : MA/JU
 Sample : SSTD00537
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD005645

Quant Time: Sep 06 01:17:08 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP090523.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Sep 06 01:15:58 2023
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 1.769 ng/uL
RT: 3.364 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP016949.D
Acq: 05 Sep 2023 11:32




Instrument :
BNA_P
ClientSampleId :
SSTD005645


Tgt Ion: 88 Resp: 16456
Ion Ratio Lower Upper
88 100
43 30.1 25.0 37.6
58 60.0 60.6 90.8#



#5
Pyridine
Concen: 4.414 ng/uL
RT: 3.787 min Scan# 85
Delta R.T. 0.000 min
Lab File: BP016949.D
Acq: 05 Sep 2023 11:32

Tgt Ion: 79 Resp: 115693
Ion Ratio Lower Upper
79 100
52 80.7 60.1 90.1
51 61.0 30.5 45.7#









#6
Benzaldehyde
Concen: 2.602 ng/ul
RT: 7.140 min Scan# 611
Delta R.T. 0.000 min
Lab File: BP016949.D
Acq: 05 Sep 2023 11:32




Instrument :
BNA_P
ClientSampleId :
SSTD005645


Tgt Ion: 77 Resp: 35124
Ion Ratio Lower Upper
77 100
105 90.9 75.4 113.0
106 88.4 71.4 107.2



#8
Phenol
Concen: 4.193 ng/ul
RT: 7.157 min Scan# 658
Delta R.T. 0.000 min
Lab File: BP016949.D
Acq: 05 Sep 2023 11:32

Tgt Ion: 94 Resp: 129726
Ion Ratio Lower Upper
94 100
65 29.9 23.2 34.8
66 37.2 29.4 44.2









#10
Bis(2-Chloroethyl)ether
Concen: 4.554 ng/ul
RT: 7.416 min Scan# 702
Delta R.T. 0.000 min
Lab File: BP016949.D
Acq: 05 Sep 2023 11:32




Instrument :
BNA_P
ClientSampleId :
SSTD005645


Tgt Ion: 93 Resp: 112365
Ion Ratio Lower Upper
93 100
63 77.4 63.4 95.2
95 33.0 26.1 39.1



#12
2-Chlorophenol
Concen: 4.323 ng/ul
RT: 7.534 min Scan# 722
Delta R.T. 0.000 min
Lab File: BP016949.D
Acq: 05 Sep 2023 11:32

Tgt Ion: 128 Resp: 99256
Ion Ratio Lower Upper
128 100
64 51.2 40.2 60.2
130 33.4 25.2 37.8





#13

2-Methylphenol

Concen: 4.113 ng/ul

RT: 8.422 min Scan# 81

Delta R.T. 0.000 min

Lab File: BP016949.D

Acq: 05 Sep 2023 11:32

Instrument :

BNA_P

ClientSampleId :





SSTD005645

Tgt Ion:108 Resp: 93343

Ion Ratio Lower Upper

108 100

107 88.1 70.6 106.0



#14

2,2'-oxybis(1-Chloropropane)

Concen: 2.574 ng/ul

RT: 8.493 min Scan# 885

Delta R.T. -0.023 min

Lab File: BP016949.D

Acq: 05 Sep 2023 11:32




Tgt Ion: 45 Resp: 89464


Ion Ratio Lower Upper

45 100

77 15.2 11.3 16.9

79 10.0 9.0 13.4









#16
Acetophenone
Concen: 4.500 ng/ul
RT: 8.828 min Scan# 943
Delta R.T. -0.006 min
Lab File: BP016949.D
Acq: 05 Sep 2023 11:32




Instrument :
BNA_P
ClientSampleId :
SSTD005645


Tgt Ion:105 Resp: 164736
Ion Ratio Lower Upper
105 100
77 81.8 63.6 95.4
51 40.5 23.4 35.0#



#17
N-Nitroso-di-n-propylamine
Concen: 4.436 ng/ul
RT: 8.799 min Scan# 937
Delta R.T. -0.006 min
Lab File: BP016949.D
Acq: 05 Sep 2023 11:32

Tgt Ion: 70 Resp: 86756
Ion Ratio Lower Upper
70 100
42 59.3 47.4 71.2
101 9.8 7.3 10.9
130 19.4 15.6 23.4









#18
4-Methylphenol
Concen: 4.233 ng/ul
RT: 8.752 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP016949.D
Acq: 05 Sep 2023 11:32




Instrument :
BNA_P
ClientSampleId :
SSTD005645


Tgt Ion:108 Resp: 104761
Ion Ratio Lower Upper
108 100
107 109.1 90.6 135.8



#19
Hexachloroethane
Concen: 4.432 ng/ul
RT: 9.051 min Scan# 980
Delta R.T. 0.000 min
Lab File: BP016949.D
Acq: 05 Sep 2023 11:32

Tgt Ion:117 Resp: 42158
Ion Ratio Lower Upper
117 100
201 97.0 75.8 113.6
199 62.4 46.8 70.2









#22
Nitrobenzene
Concen: 4.442 ng/ul
RT: 9.222 min Scan# 1009
Delta R.T. 0.000 min
Lab File: BP016949.D
Acq: 05 Sep 2023 11:32




Instrument :
BNA_P
ClientSampleId :
SSTD005645


Tgt Ion: 77 Resp: 125018
Ion Ratio Lower Upper
77 100
123 40.4 34.4 51.6
65 13.3 10.6 15.8



#23
Isophorone
Concen: 4.366 ng/ul
RT: 9.734 min Scan# 1096
Delta R.T. -0.006 min
Lab File: BP016949.D
Acq: 05 Sep 2023 11:32

Tgt Ion: 82 Resp: 230839
Ion Ratio Lower Upper
82 100
95 6.5 5.3 7.9
138 15.9 13.4 20.2









#25
2-Nitrophenol
Concen: 4.133 ng/ul
RT: 9.928 min Scan# 1129
Delta R.T. 0.000 min
Lab File: BP016949.D
Acq: 05 Sep 2023 11:32




Instrument :
BNA_P
ClientSampleId :
SSTD005645


Tgt Ion:139 Resp: 51990
Ion Ratio Lower Upper
139 100
65 54.0 44.0 66.0
109 29.5 24.1 36.1



#26
2,4-Dimethylphenol
Concen: 4.426 ng/ul
RT: 9.963 min Scan# 1135
Delta R.T. -0.006 min
Lab File: BP016949.D
Acq: 05 Sep 2023 11:32

Tgt Ion:107 Resp: 113685
Ion Ratio Lower Upper
107 100
121 49.5 41.0 61.6
122 83.9 69.0 103.4





#27

Bis(2-Chloroethoxy)methane

Concen: 4.588 ng/ul

RT: 10.222 min Scan# 1179

Delta R.T. -0.006 min

Lab File: BP016949.D

Acq: 05 Sep 2023 11:32

Instrument :

BNA_P

ClientSampleId :

SSTD005645





Tgt Ion: 93 Resp: 150324

Ion Ratio Lower Upper

93 100

95 32.5 26.0 39.0

123 10.7 8.6 13.0



#29

2,4-Dichlorophenol

Concen: 4.362 ng/ul

RT: 10.446 min Scan# 1217

Delta R.T. -0.006 min

Lab File: BP016949.D

Acq: 05 Sep 2023 11:32




Tgt Ion:162 Resp: 93777


Ion Ratio Lower Upper

162 100

164 63.1 51.8 77.6

98 36.1 28.2 42.4









#30
Naphthalene
Concen: 4.680 ng/ul
RT: 10.863 min Scan# 1288
Delta R.T. 0.000 min
Lab File: BP016949.D
Acq: 05 Sep 2023 11:32




Instrument :
BNA_P
ClientSampleId :
SSTD005645


Tgt Ion:128 Resp: 346904
Ion Ratio Lower Upper
128 100
129 11.0 8.8 13.2
127 12.8 10.3 15.5



#32
4-Chloroaniline
Concen: 4.686 ng/ul
RT: 10.993 min Scan# 1310
Delta R.T. 0.000 min
Lab File: BP016949.D
Acq: 05 Sep 2023 11:32

Tgt Ion:127 Resp: 148417
Ion Ratio Lower Upper
127 100
129 31.8 26.0 39.0





#33

Hexachlorobutadiene

Concen: 4.485 ng/ul

RT: 11.093 min Scan# 11

Delta R.T. 0.000 min

Lab File: BP016949.D

Acq: 05 Sep 2023 11:32

Instrument :

BNA_P

ClientSampleId :

SSTD005645





Tgt Ion:225 Resp: 57619

Ion Ratio Lower Upper

225 100

223 63.6 50.1 75.1

227 66.6 50.9 76.3



#34

Caprolactam

Concen: 3.831 ng/ul

RT: 11.804 min Scan# 1448

Delta R.T. -0.006 min

Lab File: BP016949.D

Acq: 05 Sep 2023 11:32




Tgt Ion:113 Resp: 27279


Ion Ratio Lower Upper

113 100

55 180.4 142.3 213.5

56 139.5 102.2 153.2









#35
4-Chloro-3-methylphenol
Concen: 4.287 ng/ul
RT: 12.092 min Scan# 1497
Delta R.T. 0.000 min
Lab File: BP016949.D
Acq: 05 Sep 2023 11:32




Instrument :
BNA_P
ClientSampleId :
SSTD005645


Tgt Ion:107 Resp: 101000
Ion Ratio Lower Upper
107 100
144 26.3 20.3 30.5
142 77.8 63.8 95.8



#36
2-Methylnaphthalene
Concen: 4.543 ng/ul
RT: 12.463 min Scan# 1560
Delta R.T. -0.006 min
Lab File: BP016949.D
Acq: 05 Sep 2023 11:32

Tgt Ion:142 Resp: 222561
Ion Ratio Lower Upper
142 100
141 88.2 71.1 106.7





#37

1-Methylnaphthalene

Concen: 4.611 ng/ul

RT: 12.687 min Scan# 1598

Delta R.T. 0.000 min

Lab File: BP016949.D

Acq: 05 Sep 2023 11:32

Instrument :

BNA_P

ClientSampleId :

SSTD005645





Tgt Ion:142 Resp: 227435

Ion Ratio Lower Upper

142 100

141 91.4 74.5 111.7

116 4.4 3.1 4.7



#39

1,2,4,5-Tetrachlorobenzene

Concen: 4.363 ng/ul

RT: 12.816 min Scan# 1620

Delta R.T. 0.000 min

Lab File: BP016949.D

Acq: 05 Sep 2023 11:32

Tgt Ion:216 Resp: 109159




Ion Ratio Lower Upper


216 100

214 79.5 62.8 94.2

179 22.9 17.5 26.3

108 19.0 14.6 21.8









#40
Hexachlorocyclopentadiene
Concen: 3.458 ng/ul
RT: 12.763 min Scan# 1611
Delta R.T. 0.000 min
Lab File: BP016949.D
Acq: 05 Sep 2023 11:32




Instrument :
BNA_P
ClientSampleId :
SSTD005645


Tgt Ion:237	Resp:	56060
Ion Ratio	Lower	Upper
237	100	
235	61.2	50.8 76.2
272	12.8	10.0 15.0



#41
2,4,6-Trichlorophenol
Concen: 4.202 ng/ul
RT: 13.069 min Scan# 1663
Delta R.T. 0.000 min
Lab File: BP016949.D
Acq: 05 Sep 2023 11:32

Tgt Ion:196	Resp:	71561
Ion Ratio	Lower	Upper
196	100	
198	97.4	75.8 113.6
200	31.3	24.6 36.8





#42

2,4,5-Trichlorophenol

Concen: 4.176 ng/ul

RT: 13.134 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP016949.D

Acq: 05 Sep 2023 11:32

Instrument :

BNA_P

ClientSampleId :

SSTD005645





Tgt Ion:196 Resp: 77369

Ion Ratio Lower Upper

196 100

198 95.0 74.9 112.3

200 30.1 24.5 36.7



#43

1,1'-Biphenyl

Concen: 4.616 ng/ul

RT: 13.475 min Scan# 1732

Delta R.T. 0.000 min

Lab File: BP016949.D

Acq: 05 Sep 2023 11:32




Tgt Ion:154 Resp: 310278


Ion Ratio Lower Upper

154 100

153 39.6 32.7 49.1

76 13.8 11.5 17.3









#44
2-Chloronaphthalene
Concen: 4.556 ng/ul
RT: 13.522 min Scan# 1740
Delta R.T. 0.000 min
Lab File: BP016949.D
Acq: 05 Sep 2023 11:32




Instrument :
BNA_P
ClientSampleId :
SSTD005645


Tgt Ion:162 Resp: 240439
Ion Ratio Lower Upper
162 100
164 32.9 26.2 39.4
127 37.2 30.8 46.2



#45
2-Nitroaniline
Concen: 3.925 ng/ul
RT: 13.751 min Scan# 1779
Delta R.T. 0.000 min
Lab File: BP016949.D
Acq: 05 Sep 2023 11:32

Tgt Ion: 65 Resp: 61695
Ion Ratio Lower Upper
65 100
92 68.5 55.4 83.0
138 93.1 76.8 115.2









#47
Dimethylphthalate
Concen: 4.572 ng/ul
RT: 14.098 min Scan# 1838
Delta R.T. -0.006 min
Lab File: BP016949.D
Acq: 05 Sep 2023 11:32




Instrument :
BNA_P
ClientSampleId :
SSTD005645


Tgt Ion:163 Resp: 303244
Ion Ratio Lower Upper
163 100
194 4.1 3.1 4.7
164 9.8 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 3.860 ng/ul
RT: 14.239 min Scan# 1862
Delta R.T. -0.006 min
Lab File: BP016949.D
Acq: 05 Sep 2023 11:32

Tgt Ion:165 Resp: 49722
Ion Ratio Lower Upper
165 100
89 69.6 50.0 75.0
121 24.0 17.8 26.6









#50
Acenaphthylene
Concen: 4.443 ng/ul
RT: 14.369 min Scan# 1884
Delta R.T. 0.000 min
Lab File: BP016949.D
Acq: 05 Sep 2023 11:32




Instrument :
BNA_P
ClientSampleId :
SSTD005645


Tgt Ion:152 Resp: 362891
Ion Ratio Lower Upper
152 100
151 18.7 15.9 23.9
153 12.8 10.5 15.7



#51
3-Nitroaniline
Concen: 3.315 ng/ul
RT: 14.581 min Scan# 1920
Delta R.T. 0.000 min
Lab File: BP016949.D
Acq: 05 Sep 2023 11:32

Tgt Ion:138 Resp: 42614
Ion Ratio Lower Upper
138 100
108 12.5 8.4 12.6
92 129.4 97.2 145.8





#52

Acenaphthene

Concen: 4.610 ng/ul

RT: 14.704 min Scan# 1941

Delta R.T. 0.000 min

Lab File: BP016949.D

Acq: 05 Sep 2023 11:32

Instrument :

BNA_P

ClientSampleId :

SSTD005645





Tgt Ion:153 Resp: 262715

Ion Ratio Lower Upper

153 100

152 48.6 38.6 57.8

154 85.5 70.2 105.4



#53

2,4-Dinitrophenol

Concen: 2.650 ng/ul

RT: 14.781 min Scan# 1954

Delta R.T. 0.000 min

Lab File: BP016949.D

Acq: 05 Sep 2023 11:32




Tgt Ion:184 Resp: 21166


Ion Ratio Lower Upper

184 100

63 75.3 61.6 92.4

154 61.8 50.8 76.2









#55
4-Nitrophenol
Concen: 3.713 ng/ul
RT: 14.863 min Scan# 1969
Delta R.T. -0.006 min
Lab File: BP016949.D
Acq: 05 Sep 2023 11:32




Instrument :
BNA_P
ClientSampleId :
SSTD005645


Tgt Ion:109 Resp: 36693
Ion Ratio Lower Upper
109 100
139 133.9 105.0 157.4
65 152.8 108.4 162.6



#56
Dibenzofuran
Concen: 4.652 ng/ul
RT: 15.039 min Scan# 1998
Delta R.T. 0.000 min
Lab File: BP016949.D
Acq: 05 Sep 2023 11:32

Tgt Ion:168 Resp: 364577
Ion Ratio Lower Upper
168 100
139 38.6 30.1 45.1









#57
2,4-Dinitrotoluene
Concen: 3.751 ng/ul
RT: 15.022 min Scan# 1995
Delta R.T. -0.006 min
Lab File: BP016949.D
Acq: 05 Sep 2023 11:32




Instrument :
BNA_P
ClientSampleId :
SSTD005645


Tgt Ion:165 Resp: 69354
Ion Ratio Lower Upper
165 100
63 56.1 42.1 63.1
182 2.8 2.3 3.5



#58
2,3,4,6-Tetrachlorophenol
Concen: 4.168 ng/ul
RT: 15.257 min Scan# 2035
Delta R.T. 0.000 min
Lab File: BP016949.D
Acq: 05 Sep 2023 11:32

Tgt Ion:232 Resp: 62258
Ion Ratio Lower Upper
232 100
131 49.6 39.2 58.8
130 2.8 2.0 3.0
166 31.7 27.2 40.8









#59
Diethylphthalate
Concen: 4.485 ng/ul
RT: 15.451 min Scan# 2068
Delta R.T. -0.006 min
Lab File: BP016949.D
Acq: 05 Sep 2023 11:32




Instrument :
BNA_P
ClientSampleId :
SSTD005645


Tgt Ion:149 Resp: 297269
Ion Ratio Lower Upper
149 100
177 21.2 17.3 25.9
150 12.8 10.1 15.1



#61
Fluorene
Concen: 4.586 ng/ul
RT: 15.692 min Scan# 2109
Delta R.T. 0.000 min
Lab File: BP016949.D
Acq: 05 Sep 2023 11:32

Tgt Ion:166 Resp: 296285
Ion Ratio Lower Upper
166 100
165 99.0 78.3 117.5
167 13.4 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 4.559 ng/ul
RT: 15.681 min Scan# 2108
Delta R.T. -0.006 min
Lab File: BP016949.D
Acq: 05 Sep 2023 11:32




Instrument :
BNA_P
ClientSampleId :
SSTD005645


Tgt Ion:204 Resp: 142329
Ion Ratio Lower Upper
204 100
206 31.8 26.4 39.6
141 61.2 46.2 69.4



#63
4-Nitroaniline
Concen: 2.082 ng/ul
RT: 15.751 min Scan# 2119
Delta R.T. 0.000 min
Lab File: BP016949.D
Acq: 05 Sep 2023 11:32

Tgt Ion:138 Resp: 23526
Ion Ratio Lower Upper
138 100
92 60.6 41.9 62.9
108 117.8 70.3 105.5#









#66
4,6-Dinitro-2-methylphenol
Concen: 3.085 ng/ul
RT: 15.775 min Scan# 2123
Delta R.T. -0.006 min
Lab File: BP016949.D
Acq: 05 Sep 2023 11:32




Instrument :
BNA_P
ClientSampleId :
SSTD005645


Tgt Ion:198 Resp: 35073
Ion Ratio Lower Upper
198 100
182 7.5 3.6 5.4#
77 30.6 21.4 32.0



#67
N-Nitrosodiphenylamine
Concen: 4.473 ng/ul
RT: 15.904 min Scan# 2145
Delta R.T. -0.006 min
Lab File: BP016949.D
Acq: 05 Sep 2023 11:32

Tgt Ion:169 Resp: 243209
Ion Ratio Lower Upper
169 100
168 67.7 54.2 81.4
167 36.7 29.4 44.2









#68
4-Bromophenyl-phenylether
Concen: 4.383 ng/ul
RT: 16.586 min Scan# 2261
Delta R.T. 0.000 min
Lab File: BP016949.D
Acq: 05 Sep 2023 11:32




Instrument :
BNA_P
ClientSampleId :
SSTD005645


Tgt Ion:	248	Resp:	78626
Ion Ratio	Lower	Upper	
248	100		
250	93.3	76.8	115.2
141	81.3	66.3	99.5



#69
Hexachlorobenzene
Concen: 4.465 ng/ul
RT: 16.669 min Scan# 2275
Delta R.T. 0.000 min
Lab File: BP016949.D
Acq: 05 Sep 2023 11:32

Tgt Ion:	284	Resp:	87801
Ion Ratio	Lower	Upper	
284	100		
142	40.4	31.9	47.9
249	31.1	27.0	40.6





#70

Atrazine

Concen: 4.463 ng/ul

RT: 16.857 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP016949.D

Acq: 05 Sep 2023 11:32

Instrument :

BNA_P

ClientSampleId :

SSTD005645





Tgt Ion:200 Resp: 83075

Ion Ratio Lower Upper

200 100

173 27.1 21.8 32.8

215 44.8 38.3 57.5



#71

Pentachlorophenol

Concen: 3.558 ng/ul

RT: 17.033 min Scan# 2337

Delta R.T. 0.000 min

Lab File: BP016949.D

Acq: 05 Sep 2023 11:32




Tgt Ion:266 Resp: 45630


Ion Ratio Lower Upper

266 100

264 60.1 50.4 75.6

268 63.0 50.4 75.6





#72

Phenanthrene

Concen: 4.601 ng/ul

RT: 17.451 min Scan# 2408

Delta R.T. 0.000 min

Lab File: BP016949.D

Acq: 05 Sep 2023 11:32

Instrument :

BNA_P

ClientSampleId :

SSTD005645





Tgt Ion:178 Resp: 460967

Ion Ratio Lower Upper

178 100

179 15.3 12.5 18.7

176 19.7 15.8 23.8



#74

Anthracene

Concen: 4.541 ng/ul

RT: 17.545 min Scan# 2424

Delta R.T. 0.000 min

Lab File: BP016949.D

Acq: 05 Sep 2023 11:32




Tgt Ion:178 Resp: 457421


Ion Ratio Lower Upper

178 100

179 15.6 12.3 18.5

176 19.5 15.0 22.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 4.418 ng/uL
RT: 13.428 min Scan# 1724
Delta R.T. 0.000 min
Lab File: BP016949.D
Acq: 05 Sep 2023 11:32




Instrument :
BNA_P
ClientSampleId :
SSTD005645


Tgt Ion:216 Resp: 113487
Ion Ratio Lower Upper
216 100
214 80.9 64.5 96.7
218 50.1 38.6 57.8



#76
Pentachlorobenzene
Concen: 4.567 ng/uL
RT: 14.934 min Scan# 1980
Delta R.T. 0.000 min
Lab File: BP016949.D
Acq: 05 Sep 2023 11:32

Tgt Ion:250 Resp: 113001
Ion Ratio Lower Upper
250 100
248 61.2 51.2 76.8
252 60.9 50.9 76.3





#77

Carbazole

Concen: 4.523 ng/ul

RT: 17.828 min Scan# 2472

Delta R.T. 0.000 min

Lab File: BP016949.D

Acq: 05 Sep 2023 11:32

Instrument :

BNA_P

ClientSampleId :

SSTD005645





Tgt Ion:167 Resp: 409421

Ion Ratio Lower Upper

167 100

166 21.7 17.3 25.9

139 13.6 10.3 15.5



#78

Di-n-butylphthalate

Concen: 4.420 ng/ul

RT: 18.339 min Scan# 2559

Delta R.T. 0.000 min

Lab File: BP016949.D

Acq: 05 Sep 2023 11:32




Tgt Ion:149 Resp: 490099


Ion Ratio Lower Upper

149 100

150 8.9 7.4 11.2

104 5.2 4.2 6.4









#80
Fluoranthene
Concen: 4.374 ng/ul
RT: 19.416 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP016949.D
Acq: 05 Sep 2023 11:32




Instrument :
BNA_P
ClientSampleId :
SSTD005645


Tgt Ion:202 Resp: 509191
Ion Ratio Lower Upper
202 100
101 12.6 10.0 15.0
100 9.2 7.5 11.3



#82
Pyrene
Concen: 4.410 ng/ul
RT: 19.774 min Scan# 2803
Delta R.T. 0.000 min
Lab File: BP016949.D
Acq: 05 Sep 2023 11:32

Tgt Ion:202 Resp: 532974
Ion Ratio Lower Upper
202 100
101 13.6 11.3 16.9
100 11.5 9.4 14.0





#83

Butylbenzylphthalate

Concen: 4.071 ng/ul

RT: 20.633 min Scan# 2949

Delta R.T. 0.000 min

Lab File: BP016949.D

Acq: 05 Sep 2023 11:32

Instrument :

BNA_P

ClientSampleId :

SSTD005645





Tgt Ion:149 Resp: 203488

Ion Ratio Lower Upper

149 100

91 73.4 59.1 88.7

206 19.4 15.7 23.5



#84

3,3'-Dichlorobenzidine

Concen: 3.530 ng/ul

RT: 21.427 min Scan# 3084

Delta R.T. 0.000 min

Lab File: BP016949.D

Acq: 05 Sep 2023 11:32




Tgt Ion:252 Resp: 119901


Ion Ratio Lower Upper

252 100

254 64.3 51.8 77.6

126 12.8 10.4 15.6





#85

Benzo(a)anthracene

Concen: 4.461 ng/ul

RT: 21.486 min Scan# 3095

Delta R.T. -0.006 min

Lab File: BP016949.D

Acq: 05 Sep 2023 11:32

Instrument :

BNA_P

ClientSampleId :

SSTD005645





Tgt Ion:228 Resp: 480860

Ion Ratio Lower Upper

228 100

229 19.7 15.8 23.8

226 27.8 22.2 33.2



#86

Bis(2-ethylhexyl)phthalate

Concen: 4.100 ng/ul

RT: 21.368 min Scan# 3074

Delta R.T. 0.000 min

Lab File: BP016949.D

Acq: 05 Sep 2023 11:32




Tgt Ion:149 Resp: 283079


Ion Ratio Lower Upper

149 100

167 27.0 21.4 32.2

279 3.7 3.0 4.4





#87

Chrysene

Concen: 4.557 ng/ul

RT: 21.545 min Scan# 3104

Delta R.T. 0.000 min

Lab File: BP016949.D

Acq: 05 Sep 2023 11:32

Instrument :

BNA_P

ClientSampleId :

SSTD005645





Tgt Ion:228 Resp: 459488

Ion Ratio Lower Upper

228 100

226 29.9 24.6 36.8

229 19.7 15.9 23.9



#89

Di-n-octyl phthalate

Concen: 3.635 ng/ul




RT: 22.339 min Scan# 3239


Delta R.T. 0.000 min

Lab File: BP016949.D

Acq: 05 Sep 2023 11:32

Tgt Ion:149 Resp: 451809





#90

Benzo(b)fluoranthene

Concen: 4.257 ng/ul

RT: 23.257 min Scan# 3395

Delta R.T. 0.000 min

Lab File: BP016949.D

Acq: 05 Sep 2023 11:32

Instrument :

BNA_P

ClientSampleId :

SSTD005645





Tgt Ion:252 Resp: 441050

Ion Ratio Lower Upper

252 100

253 22.1 17.6 26.4

125 11.7 8.8 13.2



#91

Benzo(k)fluoranthene

Concen: 4.296 ng/ul

RT: 23.310 min Scan# 3404

Delta R.T. 0.000 min

Lab File: BP016949.D

Acq: 05 Sep 2023 11:32




Tgt Ion:252 Resp: 430990


Ion Ratio Lower Upper

252 100

253 22.0 17.7 26.5

125 11.5 8.9 13.3





#93

Benzo(a)pyrene

Concen: 4.255 ng/ul

RT: 23.933 min Scan# 3510

Delta R.T. 0.000 min

Lab File: BP016949.D

Acq: 05 Sep 2023 11:32

Instrument :

BNA_P

ClientSampleId :

SSTD005645





Tgt Ion:252 Resp: 367643

Ion Ratio Lower Upper

252 100

253 22.6 17.8 26.6

125 13.3 9.9 14.9



#94

Indeno(1,2,3-cd)pyrene

Concen: 4.470 ng/ul

RT: 26.745 min Scan# 3988

Delta R.T. 0.000 min

Lab File: BP016949.D

Acq: 05 Sep 2023 11:32




Tgt Ion:276 Resp: 428764


Ion Ratio Lower Upper

276 100

138 26.6 21.4 32.0

277 24.3 20.0 30.0





#95

Dibenzo(a,h)anthracene

Concen: 4.461 ng/ul

RT: 26.768 min Scan# 3992

Delta R.T. 0.000 min

Lab File: BP016949.D

Acq: 05 Sep 2023 11:32

Instrument :

BNA_P

ClientSampleId :

SSTD005645





Tgt Ion:278 Resp: 353178

Ion Ratio Lower Upper

278 100

139 20.1 14.9 22.3

279 23.6 19.6 29.4



#96

Benzo(g,h,i)perylene

Concen: 4.557 ng/ul

RT: 27.586 min Scan# 4131

Delta R.T. -0.006 min

Lab File: BP016949.D

Acq: 05 Sep 2023 11:32

Tgt Ion:276 Resp: 344423

Ion Ratio Lower Upper

276 100

138 25.5 19.5 29.3

277 23.6 18.7 28.1

