

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090623\
 Data File : BP016957.D
 Acq On : 06 Sep 2023 16:42
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020718

Quant Time: Sep 06 22:15:42 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP090523.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Sep 06 01:52:05 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.981	152	314182	20.000	ng/u1	0.00
20) Naphthalene-d8	10.810	136	1356389	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.640	164	837535	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.410	188	1739503	20.000	ng/u1	0.00
79) Chrysene-d12	21.510	240	1350322	20.000	ng/u1	0.00
88) Perylene-d12	24.057	264	1183500	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.328	96	70556	8.495	ng/uL	0.00
4) Pyridine-d5	3.764	84	506971	21.361	ng/u1	0.00
7) Phenol-d5	7.128	99	597054	20.801	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.316	67	391105	21.783	ng/u1	0.00
11) 2-Chlorophenol-d4	7.499	132	444772	21.410	ng/u1	0.00
15) 4-Methylphenol-d8	8.693	113	462770	21.163	ng/u1	0.00
21) Nitrobenzene-d5	9.181	128	222949	21.187	ng/u1	0.00
24) 2-Nitrophenol-d4	9.893	143	239722	21.015	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.422	165	441571	20.968	ng/u1	0.00
31) 4-Chloroaniline-d4	10.969	131	633166	20.839	ng/u1	0.00
46) Dimethylphthalate-d6	14.051	166	1291308	20.713	ng/u1	0.00
49) Acenaphthylene-d8	14.340	160	1516794	21.082	ng/u1	0.00
54) 4-Nitrophenol-d4	14.857	143	247340	20.310	ng/u1	0.00
60) Fluorene-d10	15.640	176	1114575	20.881	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.763	200	205438	19.625	ng/u1	0.00
73) Anthracene-d10	17.510	188	1699608	21.276	ng/u1	0.00
81) Pyrene-d10	19.745	212	1771260	21.536	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.886	264	1240743	20.665	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.364	88	78329	8.734	ng/uL	99
5) Pyridine	3.781	79	539960	21.371	ng/u1	98
6) Benzaldehyde	7.134	77	341247	26.231	ng/u1	99
8) Phenol	7.158	94	625987	20.991	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.416	93	511073	21.490	ng/u1	98
12) 2-Chlorophenol	7.534	128	473539	21.397	ng/u1	99
13) 2-Methylphenol	8.422	108	459859	21.022	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.493	45	434649	12.981	ng/u1	98
16) Acetophenone	8.828	105	773767	21.927	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.799	70	414615	21.995	ng/u1	99
18) 4-Methylphenol	8.758	108	512692	21.491	ng/u1	97
19) Hexachloroethane	9.052	117	199189	21.726	ng/u1	98
22) Nitrobenzene	9.222	77	606601	21.589	ng/u1	100
23) Isophorone	9.734	82	1124863	21.309	ng/u1	99
25) 2-Nitrophenol	9.928	139	269174	21.433	ng/u1	99
26) 2,4-Dimethylphenol	9.963	107	546009	21.293	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.222	93	692928	21.181	ng/u1	100
29) 2,4-Dichlorophenol	10.446	162	453117	21.112	ng/u1	97
30) Naphthalene	10.863	128	1566437	21.164	ng/u1	99
32) 4-Chloroaniline	10.993	127	664949	21.029	ng/u1	99
33) Hexachlorobutadiene	11.093	225	261328	20.372	ng/u1	99
34) Caprolactam	11.810	113	144198	20.287	ng/u1	93
35) 4-Chloro-3-methylphenol	12.093	107	499542	21.240	ng/u1	100
36) 2-Methylnaphthalene	12.463	142	1042664	21.318	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090623\
 Data File : BP016957.D
 Acq On : 06 Sep 2023 16:42
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020718


Quant Time: Sep 06 22:15:42 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP090523.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Sep 06 01:52:05 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.687	142	1045625	21.234	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.816	216	507849	20.580	ng/ul	100
40) Hexachlorocyclopentadiene	12.763	237	296899	18.569	ng/ul	99
41) 2,4,6-Trichlorophenol	13.069	196	347929	20.716	ng/ul	99
42) 2,4,5-Trichlorophenol	13.140	196	380767	20.836	ng/ul	98
43) 1,1'-Biphenyl	13.475	154	1405449	21.199	ng/ul	99
44) 2-Chloronaphthalene	13.522	162	1088461	20.911	ng/ul	99
45) 2-Nitroaniline	13.757	65	337518	21.773	ng/ul	96
47) Dimethylphthalate	14.104	163	1367019	20.898	ng/ul	100
48) 2,6-Dinitrotoluene	14.246	165	266408	20.971	ng/ul	97
50) Acenaphthylene	14.369	152	1723801	21.397	ng/ul	100
51) 3-Nitroaniline	14.581	138	264902	20.894	ng/ul	95
52) Acenaphthene	14.704	153	1187931	21.134	ng/ul	100
53) 2,4-Dinitrophenol	14.781	184	146085	18.547	ng/ul	99
55) 4-Nitrophenol	14.869	109	205892	21.124	ng/ul	97
56) Dibenzofuran	15.040	168	1634913	21.154	ng/ul	99
57) 2,4-Dinitrotoluene	15.028	165	392439	21.520	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	15.257	232	303476	20.598	ng/ul	99
59) Diethylphthalate	15.457	149	1372352	20.994	ng/ul	100
61) Fluorene	15.693	166	1345318	21.113	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.687	204	631881	20.520	ng/ul	97
63) 4-Nitroaniline	15.751	138	244096	21.898	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.781	198	218492	19.922	ng/ul	97
67) N-Nitrosodiphenylamine	15.910	169	1127140	21.489	ng/ul	100
68) 4-Bromophenyl-phenylether	16.587	248	353904	20.450	ng/ul	99
69) Hexachlorobenzene	16.669	284	391085	20.615	ng/ul	96
70) Atrazine	16.863	200	374324	20.844	ng/ul	98
71) Pentachlorophenol	17.034	266	235224	19.014	ng/ul	99
72) Phenanthrene	17.457	178	2044987	21.157	ng/ul	99
74) Anthracene	17.545	178	2079094	21.397	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.428	216	513850	20.738	ng/ul	98
76) Pentachlorobenzene	14.934	250	494717	20.728	ng/ul	99
77) Carbazole	17.828	167	1865772	21.368	ng/ul	99
78) Di-n-butylphthalate	18.339	149	2255936	21.089	ng/ul	100
80) Fluoranthene	19.422	202	2225375	21.669	ng/ul	99
82) Pyrene	19.775	202	2340020	21.946	ng/ul	98
83) Butylbenzylphthalate	20.633	149	916317	20.780	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.433	252	615197	20.529	ng/ul	100
85) Benzo(a)anthracene	21.492	228	2006465	21.098	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.375	149	1218962	20.009	ng/ul	100
87) Chrysene	21.551	228	1889167	21.236	ng/ul	100
89) Di-n-octyl phthalate	22.345	149	1910911	19.261	ng/ul	100
90) Benzo(b)fluoranthene	23.263	252	1746937	21.127	ng/ul	99
91) Benzo(k)fluoranthene	23.316	252	1668550	20.838	ng/ul	99
93) Benzo(a)pyrene	23.939	252	1437574	20.844	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.757	276	1564521	20.436	ng/ul	99
95) Dibenzo(a,h)anthracene	26.780	278	1287992	20.383	ng/ul	99
96) Benzo(g,h,i)perylene	27.604	276	1243872	20.617	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Sep 06 22:15:42 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP090523.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Sep 06 01:52:05 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.734 ng/uL
RT: 3.364 min Scan# 11
Delta R.T. 0.001 min
Lab File: BP016957.D
Acq: 06 Sep 2023 16:42




Instrument :
BNA_P
ClientSampleId :
SSTD020718


Tgt Ion: 88 Resp: 78329
Ion Ratio Lower Upper
88 100
43 31.1 25.0 37.6
58 74.0 60.6 90.8



#5
Pyridine
Concen: 21.371 ng/uL
RT: 3.781 min Scan# 84
Delta R.T. -0.005 min
Lab File: BP016957.D
Acq: 06 Sep 2023 16:42

Tgt Ion: 79 Resp: 539960
Ion Ratio Lower Upper
79 100
52 76.7 60.1 90.1
51 36.3 30.5 45.7









#6
Benzaldehyde
Concen: 26.231 ng/ul
RT: 7.134 min Scan# 61
Delta R.T. -0.005 min
Lab File: BP016957.D
Acq: 06 Sep 2023 16:42




Instrument :
BNA_P
ClientSampleId :
SSTD020718


Tgt Ion: 77 Resp: 341247
Ion Ratio Lower Upper
77 100
105 92.7 75.4 113.0
106 88.6 71.4 107.2



#8
Phenol
Concen: 20.991 ng/ul
RT: 7.158 min Scan# 658
Delta R.T. 0.000 min
Lab File: BP016957.D
Acq: 06 Sep 2023 16:42

Tgt Ion: 94 Resp: 625987
Ion Ratio Lower Upper
94 100
65 29.5 23.2 34.8
66 37.2 29.4 44.2









#10
Bis(2-Chloroethyl)ether
Concen: 21.490 ng/ul
RT: 7.416 min Scan# 702
Delta R.T. 0.000 min
Lab File: BP016957.D
Acq: 06 Sep 2023 16:42




Instrument :
BNA_P
ClientSampleId :
SSTD020718


Tgt Ion: 93 Resp: 511073
Ion Ratio Lower Upper
93 100
63 77.9 63.4 95.2
95 31.7 26.1 39.1



#12
2-Chlorophenol
Concen: 21.397 ng/ul
RT: 7.534 min Scan# 722
Delta R.T. 0.000 min
Lab File: BP016957.D
Acq: 06 Sep 2023 16:42

Tgt Ion: 128 Resp: 473539
Ion Ratio Lower Upper
128 100
64 50.6 40.2 60.2
130 32.3 25.2 37.8









#13
2-Methylphenol
Concen: 21.022 ng/ul
RT: 8.422 min Scan# 873
Delta R.T. 0.000 min
Lab File: BP016957.D
Acq: 06 Sep 2023 16:42




Instrument :
BNA_P
ClientSampleId :
SSTD020718


Tgt Ion:108 Resp: 459859
Ion Ratio Lower Upper
108 100
107 88.1 70.6 106.0



#14
2,2'-oxybis(1-Chloropropane)
Concen: 12.981 ng/ul
RT: 8.493 min Scan# 885
Delta R.T. -0.023 min
Lab File: BP016957.D
Acq: 06 Sep 2023 16:42

Tgt Ion: 45 Resp: 434649
Ion Ratio Lower Upper
45 100
77 14.3 11.3 16.9
79 9.9 9.0 13.4









#16
Acetophenone
Concen: 21.927 ng/ul
RT: 8.828 min Scan# 942
Delta R.T. -0.005 min
Lab File: BP016957.D
Acq: 06 Sep 2023 16:42




Instrument :
BNA_P
ClientSampleId :
SSTD020718


Tgt Ion:105 Resp: 773767
Ion Ratio Lower Upper
105 100
77 81.5 63.6 95.4
51 29.4 23.4 35.0



#17
N-Nitroso-di-n-propylamine
Concen: 21.995 ng/ul
RT: 8.799 min Scan# 937
Delta R.T. -0.005 min
Lab File: BP016957.D
Acq: 06 Sep 2023 16:42

Tgt Ion: 70 Resp: 414615
Ion Ratio Lower Upper
70 100
42 60.4 47.4 71.2
101 9.2 7.3 10.9
130 19.2 15.6 23.4









#18
4-Methylphenol
Concen: 21.491 ng/ul
RT: 8.758 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP016957.D
Acq: 06 Sep 2023 16:42




Instrument :
BNA_P
ClientSampleId :
SSTD020718


Tgt Ion:108 Resp: 512692
Ion Ratio Lower Upper
108 100
107 109.8 90.6 135.8



#19
Hexachloroethane
Concen: 21.726 ng/ul
RT: 9.052 min Scan# 980
Delta R.T. 0.000 min
Lab File: BP016957.D
Acq: 06 Sep 2023 16:42

Tgt Ion:117 Resp: 199189
Ion Ratio Lower Upper
117 100
201 92.7 75.8 113.6
199 57.7 46.8 70.2









#22
Nitrobenzene
Concen: 21.589 ng/ul
RT: 9.222 min Scan# 1009
Delta R.T. 0.000 min
Lab File: BP016957.D
Acq: 06 Sep 2023 16:42




Instrument :
BNA_P
ClientSampleId :
SSTD020718


Tgt Ion: 77 Resp: 606601
Ion Ratio Lower Upper
77 100
123 42.8 34.4 51.6
65 13.4 10.6 15.8



#23
Isophorone
Concen: 21.309 ng/ul
RT: 9.734 min Scan# 1096
Delta R.T. -0.005 min
Lab File: BP016957.D
Acq: 06 Sep 2023 16:42

Tgt Ion: 82 Resp: 1124863
Ion Ratio Lower Upper
82 100
95 6.8 5.3 7.9
138 16.4 13.4 20.2









#25
2-Nitrophenol
Concen: 21.433 ng/ul
RT: 9.928 min Scan# 1129
Delta R.T. 0.000 min
Lab File: BP016957.D
Acq: 06 Sep 2023 16:42




Instrument :
BNA_P
ClientSampleId :
SSTD020718


Tgt Ion:139 Resp: 269174
Ion Ratio Lower Upper
139 100
65 54.6 44.0 66.0
109 29.3 24.1 36.1



#26
2,4-Dimethylphenol
Concen: 21.293 ng/ul
RT: 9.963 min Scan# 1135
Delta R.T. -0.005 min
Lab File: BP016957.D
Acq: 06 Sep 2023 16:42

Tgt Ion:107 Resp: 546009
Ion Ratio Lower Upper
107 100
121 49.8 41.0 61.6
122 84.8 69.0 103.4









#27
 Bis(2-Chloroethoxy)methane
 Concen: 21.181 ng/ul
 RT: 10.222 min Scan# 1179
 Delta R.T. -0.005 min
 Lab File: BP016957.D
 Acq: 06 Sep 2023 16:42




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020718


Tgt Ion: 93 Resp: 692928
 Ion Ratio Lower Upper
 93 100
 95 32.3 26.0 39.0
 123 10.6 8.6 13.0



#29
 2,4-Dichlorophenol
 Concen: 21.112 ng/ul
 RT: 10.446 min Scan# 1217
 Delta R.T. -0.005 min
 Lab File: BP016957.D
 Acq: 06 Sep 2023 16:42

Tgt Ion: 162 Resp: 453117
 Ion Ratio Lower Upper
 162 100
 164 65.5 51.8 77.6
 98 38.8 28.2 42.4





#30

Naphthalene

Concen: 21.164 ng/ul

RT: 10.863 min Scan# 1288

Delta R.T. 0.000 min

Lab File: BP016957.D

Acq: 06 Sep 2023 16:42

Instrument :

BNA_P

ClientSampleId :

SSTD020718





Tgt Ion:128 Resp: 1566437

Ion Ratio Lower Upper

128 100

129 10.7 8.8 13.2

127 13.3 10.3 15.5



#32

4-Chloroaniline

Concen: 21.029 ng/ul

RT: 10.993 min Scan# 1310

Delta R.T. 0.000 min

Lab File: BP016957.D




Acq: 06 Sep 2023 16:42


Tgt Ion:127 Resp: 664949

Ion Ratio Lower Upper

127 100

129 32.8 26.0 39.0





#33

Hexachlorobutadiene

Concen: 20.372 ng/ul

RT: 11.093 min Scan# 11

Delta R.T. 0.000 min

Lab File: BP016957.D

Acq: 06 Sep 2023 16:42

Instrument :

BNA_P

ClientSampleId :

SSTD020718





Tgt Ion:225 Resp: 261328

Ion Ratio Lower Upper

225 100

223 62.4 50.1 75.1

227 64.5 50.9 76.3



#34

Caprolactam

Concen: 20.287 ng/ul

RT: 11.810 min Scan# 1449

Delta R.T. 0.000 min

Lab File: BP016957.D

Acq: 06 Sep 2023 16:42




Tgt Ion:113 Resp: 144198


Ion Ratio Lower Upper

113 100

55 185.4 142.3 213.5

56 138.8 102.2 153.2









#35
4-Chloro-3-methylphenol
Concen: 21.240 ng/ul
RT: 12.093 min Scan# 1497
Delta R.T. 0.000 min
Lab File: BP016957.D
Acq: 06 Sep 2023 16:42




Instrument :
BNA_P
ClientSampleId :
SSTD020718


Tgt Ion:107 Resp: 499542
Ion Ratio Lower Upper
107 100
144 25.4 20.3 30.5
142 79.6 63.8 95.8



#36
2-Methylnaphthalene
Concen: 21.318 ng/ul
RT: 12.463 min Scan# 1560
Delta R.T. -0.005 min
Lab File: BP016957.D
Acq: 06 Sep 2023 16:42

Tgt Ion:142 Resp: 1042664
Ion Ratio Lower Upper
142 100
141 88.1 71.1 106.7









#37
1-Methylnaphthalene
Concen: 21.234 ng/ul
RT: 12.687 min Scan# 1598
Delta R.T. 0.000 min
Lab File: BP016957.D
Acq: 06 Sep 2023 16:42




Instrument :
BNA_P
ClientSampleId :
SSTD020718


Tgt Ion:142 Resp: 1045625
Ion Ratio Lower Upper
142 100
141 93.4 74.5 111.7
116 3.8 3.1 4.7



#39
1,2,4,5-Tetrachlorobenzene
Concen: 20.580 ng/ul
RT: 12.816 min Scan# 1620
Delta R.T. 0.000 min
Lab File: BP016957.D
Acq: 06 Sep 2023 16:42

Tgt Ion:216 Resp: 507849
Ion Ratio Lower Upper
216 100
214 78.2 62.8 94.2
179 22.0 17.5 26.3
108 18.7 14.6 21.8









#40
Hexachlorocyclopentadiene
Concen: 18.569 ng/ul
RT: 12.763 min Scan# 1611
Delta R.T. 0.000 min
Lab File: BP016957.D
Acq: 06 Sep 2023 16:42




Instrument :
BNA_P
ClientSampleId :
SSTD020718


Tgt Ion:237 Resp: 296899
Ion Ratio Lower Upper
237 100
235 64.1 50.8 76.2
272 12.5 10.0 15.0



#41
2,4,6-Trichlorophenol
Concen: 20.716 ng/ul
RT: 13.069 min Scan# 1663
Delta R.T. 0.000 min
Lab File: BP016957.D
Acq: 06 Sep 2023 16:42

Tgt Ion:196 Resp: 347929
Ion Ratio Lower Upper
196 100
198 94.3 75.8 113.6
200 30.2 24.6 36.8





#42

2,4,5-Trichlorophenol

Concen: 20.836 ng/ul

RT: 13.140 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP016957.D

Acq: 06 Sep 2023 16:42

Instrument :

BNA_P

ClientSampleId :

SSTD020718





Tgt Ion:196 Resp: 380767

Ion Ratio Lower Upper

196 100

198 95.5 74.9 112.3

200 30.2 24.5 36.7



#43

1,1'-Biphenyl

Concen: 21.199 ng/ul

RT: 13.475 min Scan# 1732

Delta R.T. 0.000 min

Lab File: BP016957.D

Acq: 06 Sep 2023 16:42




Tgt Ion:154 Resp: 1405449


Ion Ratio Lower Upper

154 100

153 40.1 32.7 49.1

76 14.4 11.5 17.3









#44
2-Chloronaphthalene
Concen: 20.911 ng/ul
RT: 13.522 min Scan# 1740
Delta R.T. 0.000 min
Lab File: BP016957.D
Acq: 06 Sep 2023 16:42




Instrument :
BNA_P
ClientSampleId :
SSTD020718


Tgt Ion:162 Resp: 1088461
Ion Ratio Lower Upper
162 100
164 32.6 26.2 39.4
127 39.7 30.8 46.2



#45
2-Nitroaniline
Concen: 21.773 ng/ul
RT: 13.757 min Scan# 1780
Delta R.T. 0.006 min
Lab File: BP016957.D
Acq: 06 Sep 2023 16:42

Tgt Ion: 65 Resp: 337518
Ion Ratio Lower Upper
65 100
92 65.9 55.4 83.0
138 99.4 76.8 115.2









#47
Dimethylphthalate
Concen: 20.898 ng/ul
RT: 14.104 min Scan# 1839
Delta R.T. 0.000 min
Lab File: BP016957.D
Acq: 06 Sep 2023 16:42




Instrument :
BNA_P
ClientSampleId :
SSTD020718


Tgt Ion:163 Resp: 1367019
Ion Ratio Lower Upper
163 100
194 3.6 3.1 4.7
164 9.8 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 20.971 ng/ul
RT: 14.246 min Scan# 1863
Delta R.T. 0.000 min
Lab File: BP016957.D
Acq: 06 Sep 2023 16:42

Tgt Ion:165 Resp: 266408
Ion Ratio Lower Upper
165 100
89 65.1 50.0 75.0
121 22.9 17.8 26.6









#50
Acenaphthylene
Concen: 21.397 ng/ul
RT: 14.369 min Scan# 1884
Delta R.T. 0.000 min
Lab File: BP016957.D
Acq: 06 Sep 2023 16:42




Instrument :
BNA_P
ClientSampleId :
SSTD020718


Tgt Ion:152 Resp: 1723801
Ion Ratio Lower Upper
152 100
151 20.1 15.9 23.9
153 13.3 10.5 15.7



#51
3-Nitroaniline
Concen: 20.894 ng/ul
RT: 14.581 min Scan# 1920
Delta R.T. 0.000 min
Lab File: BP016957.D
Acq: 06 Sep 2023 16:42

Tgt Ion:138 Resp: 264902
Ion Ratio Lower Upper
138 100
108 11.4 8.4 12.6
92 127.2 97.2 145.8





#52

Acenaphthene

Concen: 21.134 ng/ul

RT: 14.704 min Scan# 1941

Delta R.T. 0.000 min

Lab File: BP016957.D

Acq: 06 Sep 2023 16:42

Instrument :

BNA_P

ClientSampleId :

SSTD020718





Tgt Ion:153 Resp: 1187931

Ion Ratio Lower Upper

153 100

152 47.7 38.6 57.8

154 87.9 70.2 105.4



#53

2,4-Dinitrophenol

Concen: 18.547 ng/ul

RT: 14.781 min Scan# 1954

Delta R.T. 0.000 min

Lab File: BP016957.D

Acq: 06 Sep 2023 16:42




Tgt Ion:184 Resp: 146085


Ion Ratio Lower Upper

184 100

63 78.4 61.6 92.4

154 63.3 50.8 76.2









#55
4-Nitrophenol
Concen: 21.124 ng/ul
RT: 14.869 min Scan# 1969
Delta R.T. 0.000 min
Lab File: BP016957.D
Acq: 06 Sep 2023 16:42




Instrument :
BNA_P
ClientSampleId :
SSTD020718


Tgt Ion:109 Resp: 205892
Ion Ratio Lower Upper
109 100
139 130.5 105.0 157.4
65 142.8 108.4 162.6



#56
Dibenzofuran
Concen: 21.154 ng/ul
RT: 15.040 min Scan# 1998
Delta R.T. 0.000 min
Lab File: BP016957.D
Acq: 06 Sep 2023 16:42

Tgt Ion:168 Resp: 1634913
Ion Ratio Lower Upper
168 100
139 37.9 30.1 45.1









#57
2,4-Dinitrotoluene
Concen: 21.520 ng/ul
RT: 15.028 min Scan# 1996
Delta R.T. 0.000 min
Lab File: BP016957.D
Acq: 06 Sep 2023 16:42




Instrument :
BNA_P
ClientSampleId :
SSTD020718


Tgt Ion:165 Resp: 392439
Ion Ratio Lower Upper
165 100
63 53.7 42.1 63.1
182 2.8 2.3 3.5



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.598 ng/ul
RT: 15.257 min Scan# 2035
Delta R.T. 0.000 min
Lab File: BP016957.D
Acq: 06 Sep 2023 16:42

Tgt Ion:232 Resp: 303476
Ion Ratio Lower Upper
232 100
131 49.7 39.2 58.8
130 2.7 2.0 3.0
166 34.7 27.2 40.8









#59
Diethylphthalate
Concen: 20.994 ng/ul
RT: 15.457 min Scan# 2069
Delta R.T. 0.000 min
Lab File: BP016957.D
Acq: 06 Sep 2023 16:42




Instrument :
BNA_P
ClientSampleId :
SSTD020718


Tgt Ion:149 Resp: 1372352
Ion Ratio Lower Upper
149 100
177 21.7 17.3 25.9
150 12.4 10.1 15.1



#61
Fluorene
Concen: 21.113 ng/ul
RT: 15.693 min Scan# 2109
Delta R.T. 0.000 min
Lab File: BP016957.D
Acq: 06 Sep 2023 16:42

Tgt Ion:166 Resp: 1345318
Ion Ratio Lower Upper
166 100
165 97.8 78.3 117.5
167 13.2 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 20.520 ng/ul
RT: 15.687 min Scan# 2108
Delta R.T. 0.000 min
Lab File: BP016957.D
Acq: 06 Sep 2023 16:42




Instrument :
BNA_P
ClientSampleId :
SSTD020718


Tgt Ion:204 Resp: 631881
Ion Ratio Lower Upper
204 100
206 33.3 26.4 39.6
141 60.5 46.2 69.4



#63
4-Nitroaniline
Concen: 21.898 ng/ul
RT: 15.751 min Scan# 2119
Delta R.T. 0.000 min
Lab File: BP016957.D
Acq: 06 Sep 2023 16:42

Tgt Ion:138 Resp: 244096
Ion Ratio Lower Upper
138 100
92 56.4 41.9 62.9
108 90.2 70.3 105.5









#66
4,6-Dinitro-2-methylphenol
Concen: 19.922 ng/ul
RT: 15.781 min Scan# 2124
Delta R.T. 0.000 min
Lab File: BP016957.D
Acq: 06 Sep 2023 16:42




Instrument :
BNA_P
ClientSampleId :
SSTD020718


Tgt Ion:198 Resp: 218492
Ion Ratio Lower Upper
198 100
182 4.6 3.6 5.4
77 28.6 21.4 32.0



#67
N-Nitrosodiphenylamine
Concen: 21.489 ng/ul
RT: 15.910 min Scan# 2146
Delta R.T. 0.000 min
Lab File: BP016957.D
Acq: 06 Sep 2023 16:42

Tgt Ion:169 Resp: 1127140
Ion Ratio Lower Upper
169 100
168 68.1 54.2 81.4
167 37.1 29.4 44.2









#68
4-Bromophenyl-phenylether
Concen: 20.450 ng/ul
RT: 16.587 min Scan# 2261
Delta R.T. 0.000 min
Lab File: BP016957.D
Acq: 06 Sep 2023 16:42




Instrument :
BNA_P
ClientSampleId :
SSTD020718


Tgt Ion:248 Resp: 353904
Ion Ratio Lower Upper
248 100
250 95.9 76.8 115.2
141 84.6 66.3 99.5



#69
Hexachlorobenzene
Concen: 20.615 ng/ul
RT: 16.669 min Scan# 2275
Delta R.T. 0.000 min
Lab File: BP016957.D
Acq: 06 Sep 2023 16:42

Tgt Ion:284 Resp: 391085
Ion Ratio Lower Upper
284 100
142 44.1 31.9 47.9
249 33.3 27.0 40.6





#70

Atrazine

Concen: 20.844 ng/ul

RT: 16.863 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP016957.D

Acq: 06 Sep 2023 16:42

Instrument :

BNA_P

ClientSampleId :

SSTD020718





Tgt Ion:200 Resp: 374324

Ion Ratio Lower Upper

200 100

173 26.5 21.8 32.8

215 46.3 38.3 57.5



#71

Pentachlorophenol

Concen: 19.014 ng/ul

RT: 17.034 min Scan# 2337

Delta R.T. 0.000 min

Lab File: BP016957.D

Acq: 06 Sep 2023 16:42




Tgt Ion:266 Resp: 235224


Ion Ratio Lower Upper

266 100

264 63.4 50.4 75.6

268 63.6 50.4 75.6





#72

Phenanthrene

Concen: 21.157 ng/ul

RT: 17.457 min Scan# 2409

Delta R.T. 0.006 min

Lab File: BP016957.D

Acq: 06 Sep 2023 16:42

Instrument :

BNA_P

ClientSampleId :

SSTD020718





Tgt Ion:178 Resp: 2044987

Ion Ratio Lower Upper

178 100

179 15.1 12.5 18.7

176 19.1 15.8 23.8



#74

Anthracene

Concen: 21.397 ng/ul

RT: 17.545 min Scan# 2424

Delta R.T. 0.000 min

Lab File: BP016957.D

Acq: 06 Sep 2023 16:42




Tgt Ion:178 Resp: 2079094


Ion Ratio Lower Upper

178 100

179 15.6 12.3 18.5

176 18.9 15.0 22.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 20.738 ng/uL

RT: 13.428 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP016957.D

Acq: 06 Sep 2023 16:42

Instrument :

BNA_P

ClientSampleId :

SSTD020718





Tgt Ion:216 Resp: 513850

Ion Ratio Lower Upper

216 100

214 78.1 64.5 96.7

218 47.6 38.6 57.8



#76

Pentachlorobenzene

Concen: 20.728 ng/uL

RT: 14.934 min Scan# 1980

Delta R.T. 0.000 min

Lab File: BP016957.D

Acq: 06 Sep 2023 16:42




Tgt Ion:250 Resp: 494717


Ion Ratio Lower Upper

250 100

248 62.3 51.2 76.8

252 63.4 50.9 76.3









#77
 Carbazole
 Concen: 21.368 ng/ul
 RT: 17.828 min Scan# 2472
 Delta R.T. 0.000 min
 Lab File: BP016957.D
 Acq: 06 Sep 2023 16:42

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020718

Tgt Ion:167 Resp: 1865772




Ion	Ratio	Lower	Upper
167	100		
166	22.1	17.3	25.9
139	13.1	10.3	15.5




#78
 Di-n-butylphthalate
 Concen: 21.089 ng/ul
 RT: 18.339 min Scan# 2559
 Delta R.T. 0.000 min
 Lab File: BP016957.D
 Acq: 06 Sep 2023 16:42

Tgt Ion:149 Resp: 2255936

Ion	Ratio	Lower	Upper
149	100		
150	9.2	7.4	11.2
104	5.5	4.2	6.4





#80

Fluoranthene

Concen: 21.669 ng/ul

RT: 19.422 min Scan# 21

Delta R.T. 0.006 min

Lab File: BP016957.D

Acq: 06 Sep 2023 16:42

Instrument :

BNA_P

ClientSampleId :

SSTD020718





Tgt Ion:202 Resp: 2225375

Ion Ratio Lower Upper

202 100

101 11.8 10.0 15.0

100 9.1 7.5 11.3



#82

Pyrene

Concen: 21.946 ng/ul

RT: 19.775 min Scan# 2803

Delta R.T. 0.000 min

Lab File: BP016957.D

Acq: 06 Sep 2023 16:42




Tgt Ion:202 Resp: 2340020


Ion Ratio Lower Upper

202 100

101 15.1 11.3 16.9

100 12.5 9.4 14.0





#83

Butylbenzylphthalate

Concen: 20.780 ng/ul

RT: 20.633 min Scan# 2949

Delta R.T. 0.000 min

Lab File: BP016957.D

Acq: 06 Sep 2023 16:42

Instrument :

BNA_P

ClientSampleId :

SSTD020718





Tgt Ion:149 Resp: 916317

Ion Ratio Lower Upper

149 100

91 76.8 59.1 88.7

206 18.3 15.7 23.5



#84

3,3'-Dichlorobenzidine

Concen: 20.529 ng/ul

RT: 21.433 min Scan# 3085

Delta R.T. 0.006 min

Lab File: BP016957.D

Acq: 06 Sep 2023 16:42




Tgt Ion:252 Resp: 615197


Ion Ratio Lower Upper

252 100

254 64.5 51.8 77.6

126 13.2 10.4 15.6





#85

Benzo(a)anthracene

Concen: 21.098 ng/ul

RT: 21.492 min Scan# 3095

Delta R.T. 0.000 min

Lab File: BP016957.D

Acq: 06 Sep 2023 16:42

Instrument :

BNA_P

ClientSampleId :

SSTD020718





Tgt Ion:228 Resp: 2006465

Ion Ratio Lower Upper

228 100

229 19.9 15.8 23.8

226 28.0 22.2 33.2



#86

Bis(2-ethylhexyl)phthalate

Concen: 20.009 ng/ul

RT: 21.375 min Scan# 3075

Delta R.T. 0.006 min

Lab File: BP016957.D

Acq: 06 Sep 2023 16:42




Tgt Ion:149 Resp: 1218962


Ion Ratio Lower Upper

149 100

167 27.0 21.4 32.2

279 3.7 3.0 4.4





#87

Chrysene

Concen: 21.236 ng/ul

RT: 21.551 min Scan# 3105

Delta R.T. 0.006 min

Lab File: BP016957.D

Acq: 06 Sep 2023 16:42

Instrument :

BNA_P

ClientSampleId :

SSTD020718





Tgt Ion:228 Resp: 1889167

Ion Ratio Lower Upper

228 100

226 31.0 24.6 36.8

229 19.9 15.9 23.9



#89

Di-n-octyl phthalate

Concen: 19.261 ng/ul




RT: 22.345 min Scan# 3240


Delta R.T. 0.006 min

Lab File: BP016957.D

Acq: 06 Sep 2023 16:42

Tgt Ion:149 Resp: 1910911





#90

Benzo(b)fluoranthene

Concen: 21.127 ng/ul

RT: 23.263 min Scan# 3396

Delta R.T. 0.006 min

Lab File: BP016957.D

Acq: 06 Sep 2023 16:42

Instrument :

BNA_P

ClientSampleId :

SSTD020718





Tgt Ion:252 Resp: 1746937

Ion Ratio Lower Upper

252 100

253 22.2 17.6 26.4

125 11.6 8.8 13.2



#91

Benzo(k)fluoranthene

Concen: 20.838 ng/ul

RT: 23.316 min Scan# 3405

Delta R.T. 0.006 min

Lab File: BP016957.D

Acq: 06 Sep 2023 16:42




Tgt Ion:252 Resp: 1668550


Ion Ratio Lower Upper

252 100

253 21.7 17.7 26.5

125 11.3 8.9 13.3





#93

Benzo(a)pyrene

Concen: 20.844 ng/ul

RT: 23.939 min Scan# 3511

Delta R.T. 0.006 min

Lab File: BP016957.D

Acq: 06 Sep 2023 16:42

Instrument :

BNA_P

ClientSampleId :

SSTD020718





Tgt Ion:252 Resp: 1437574

Ion Ratio Lower Upper

252 100

253 22.1 17.8 26.6

125 12.9 9.9 14.9



#94

Indeno(1,2,3-cd)pyrene

Concen: 20.436 ng/ul

RT: 26.757 min Scan# 3990

Delta R.T. 0.012 min

Lab File: BP016957.D

Acq: 06 Sep 2023 16:42




Tgt Ion:276 Resp: 1564521


Ion Ratio Lower Upper

276 100

138 26.1 21.4 32.0

277 24.5 20.0 30.0





#95

Dibenzo(a,h)anthracene

Concen: 20.383 ng/ul

RT: 26.780 min Scan# 3994

Delta R.T. 0.012 min

Lab File: BP016957.D

Acq: 06 Sep 2023 16:42

Instrument :

BNA_P

ClientSampleId :

SSTD020718





Tgt Ion:278 Resp: 1287992

Ion Ratio Lower Upper

278 100

139 18.8 14.9 22.3

279 23.8 19.6 29.4



#96

Benzo(g,h,i)perylene

Concen: 20.617 ng/ul

RT: 27.604 min Scan# 4134

Delta R.T. 0.012 min

Lab File: BP016957.D

Acq: 06 Sep 2023 16:42

Tgt Ion:276 Resp: 1243872

Ion Ratio Lower Upper

276 100

138 24.4 19.5 29.3

277 23.6 18.7 28.1

