

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090723\
 Data File : BP016990.D
 Acq On : 07 Sep 2023 16:24
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020721

Quant Time: Sep 08 00:39:13 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP090523.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Sep 06 01:52:05 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.981	152	306747	20.000	ng/u1	0.00
20) Naphthalene-d8	10.810	136	1347178	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.639	164	843962	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.404	188	1803630	20.000	ng/u1	0.00
79) Chrysene-d12	21.504	240	1528205	20.000	ng/u1	0.00
88) Perylene-d12	24.045	264	1485658	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.328	96	69253	8.540	ng/uL	0.00
4) Pyridine-d5	3.763	84	493999	21.319	ng/u1	0.00
7) Phenol-d5	7.128	99	594199	21.203	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.316	67	388051	22.137	ng/u1	0.00
11) 2-Chlorophenol-d4	7.498	132	433230	21.360	ng/u1	0.00
15) 4-Methylphenol-d8	8.693	113	458897	21.494	ng/u1	0.00
21) Nitrobenzene-d5	9.175	128	221523	21.195	ng/u1	0.00
24) 2-Nitrophenol-d4	9.892	143	237010	20.920	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.416	165	441432	21.104	ng/u1	0.00
31) 4-Chloroaniline-d4	10.969	131	642636	21.295	ng/u1	0.00
46) Dimethylphthalate-d6	14.051	166	1316359	20.954	ng/u1	0.00
49) Acenaphthylene-d8	14.333	160	1520097	20.967	ng/u1	0.00
54) 4-Nitrophenol-d4	14.851	143	250784	20.436	ng/u1	0.00
60) Fluorene-d10	15.633	176	1145172	21.291	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.763	200	194257	17.897	ng/u1	0.00
73) Anthracene-d10	17.504	188	1750324	21.131	ng/u1	0.00
81) Pyrene-d10	19.745	212	1911077	20.531	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.874	264	1553432	20.611	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.363	88	77546	8.857	ng/uL	100
5) Pyridine	3.781	79	523109	21.206	ng/u1	97
6) Benzaldehyde	7.134	77	340270	26.790	ng/u1	97
8) Phenol	7.157	94	622291	21.372	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.416	93	504767	21.739	ng/u1	99
12) 2-Chlorophenol	7.534	128	460105	21.293	ng/u1	98
13) 2-Methylphenol	8.422	108	454039	21.259	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.493	45	443374	13.562	ng/u1	98
16) Acetophenone	8.828	105	769880	22.346	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.798	70	420213	22.832	ng/u1	98
18) 4-Methylphenol	8.751	108	513872	22.062	ng/u1	97
19) Hexachloroethane	9.045	117	191595	21.404	ng/u1	96
22) Nitrobenzene	9.222	77	609546	21.842	ng/u1	98
23) Isophorone	9.734	82	1126356	21.483	ng/u1	99
25) 2-Nitrophenol	9.928	139	264948	21.240	ng/u1	100
26) 2,4-Dimethylphenol	9.963	107	549327	21.569	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.222	93	701924	21.603	ng/u1	99
29) 2,4-Dichlorophenol	10.445	162	451673	21.189	ng/u1	98
30) Naphthalene	10.857	128	1553879	21.138	ng/u1	100
32) 4-Chloroaniline	10.992	127	678039	21.590	ng/u1	99
33) Hexachlorobutadiene	11.087	225	254809	19.999	ng/u1	99
34) Caprolactam	11.810	113	151215	21.420	ng/u1	95
35) 4-Chloro-3-methylphenol	12.092	107	511839	21.911	ng/u1	98
36) 2-Methylnaphthalene	12.463	142	1041124	21.432	ng/u1	100


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090723\
 Data File : BP016990.D
 Acq On : 07 Sep 2023 16:24
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020721

Quant Time: Sep 08 00:39:13 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP090523.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Sep 06 01:52:05 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.681	142	1050576	21.480	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.810	216	500547	20.130	ng/ul	99
40) Hexachlorocyclopentadiene	12.763	237	264706	16.430	ng/ul	99
41) 2,4,6-Trichlorophenol	13.063	196	349473	20.650	ng/ul	99
42) 2,4,5-Trichlorophenol	13.133	196	385432	20.931	ng/ul	99
43) 1,1'-Biphenyl	13.469	154	1409722	21.102	ng/ul	99
44) 2-Chloronaphthalene	13.516	162	1103530	21.039	ng/ul	99
45) 2-Nitroaniline	13.751	65	350734	22.453	ng/ul	98
47) Dimethylphthalate	14.098	163	1384427	21.003	ng/ul	100
48) 2,6-Dinitrotoluene	14.239	165	269569	21.058	ng/ul	96
50) Acenaphthylene	14.363	152	1729663	21.306	ng/ul	100
51) 3-Nitroaniline	14.580	138	281813	22.059	ng/ul	99
52) Acenaphthene	14.704	153	1205987	21.292	ng/ul	99
53) 2,4-Dinitrophenol	14.780	184	123882	15.608	ng/ul	95
55) 4-Nitrophenol	14.869	109	208048	21.183	ng/ul	99
56) Dibenzofuran	15.039	168	1657647	21.284	ng/ul	100
57) 2,4-Dinitrotoluene	15.022	165	398044	21.661	ng/ul	95
58) 2,3,4,6-Tetrachlorophenol	15.251	232	306132	20.620	ng/ul	96
59) Diethylphthalate	15.451	149	1411058	21.422	ng/ul	100
61) Fluorene	15.692	166	1367329	21.295	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.680	204	638635	20.581	ng/ul	95
63) 4-Nitroaniline	15.751	138	261485	23.280	ng/ul	93
66) 4,6-Dinitro-2-methylph...	15.775	198	207591	18.255	ng/ul	93
67) N-Nitrosodiphenylamine	15.904	169	1148404	21.116	ng/ul	100
68) 4-Bromophenyl-phenylether	16.586	248	360844	20.110	ng/ul	99
69) Hexachlorobenzene	16.663	284	392564	19.958	ng/ul	95
70) Atrazine	16.857	200	381328	20.479	ng/ul	98
71) Pentachlorophenol	17.027	266	239251	18.652	ng/ul	99
72) Phenanthrene	17.451	178	2127032	21.223	ng/ul	100
74) Anthracene	17.539	178	2151613	21.356	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.422	216	511165	19.896	ng/ul	99
76) Pentachlorobenzene	14.928	250	496352	20.057	ng/ul	99
77) Carbazole	17.827	167	1981495	21.886	ng/ul	100
78) Di-n-butylphthalate	18.339	149	2388146	21.531	ng/ul	100
80) Fluoranthene	19.416	202	2377884	20.459	ng/ul	100
82) Pyrene	19.774	202	2496942	20.691	ng/ul	99
83) Butylbenzylphthalate	20.633	149	1064528	21.331	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.427	252	710318	20.944	ng/ul	99
85) Benzo(a)anthracene	21.486	228	2288705	21.265	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.368	149	1517802	22.015	ng/ul	100
87) Chrysene	21.545	228	2133865	21.195	ng/ul	99
89) Di-n-octyl phthalate	22.333	149	2490239	19.996	ng/ul	100
90) Benzo(b)fluoranthene	23.251	252	2130670	20.527	ng/ul	99
91) Benzo(k)fluoranthene	23.309	252	2066432	20.558	ng/ul	99
93) Benzo(a)pyrene	23.933	252	1802781	20.823	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.739	276	2042526	21.253	ng/ul	99
95) Dibenzo(a,h)anthracene	26.762	278	1694021	21.357	ng/ul	99
96) Benzo(g,h,i)perylene	27.586	276	1597863	21.098	ng/ul	100





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 8.857 ng/uL
RT: 3.363 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP016990.D
Acq: 07 Sep 2023 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD020721


Tgt Ion: 88 Resp: 77546
Ion Ratio Lower Upper
88 100
43 30.9 25.0 37.6
58 75.6 60.6 90.8



#5
Pyridine
Concen: 21.206 ng/uL
RT: 3.781 min Scan# 84
Delta R.T. -0.006 min
Lab File: BP016990.D
Acq: 07 Sep 2023 16:24

Tgt Ion: 79 Resp: 523109
Ion Ratio Lower Upper
79 100
52 78.5 60.1 90.1
51 37.6 30.5 45.7









#6
Benzaldehyde
Concen: 26.790 ng/ul
RT: 7.134 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP016990.D
Acq: 07 Sep 2023 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD020721


Tgt Ion: 77 Resp: 340270
Ion Ratio Lower Upper
77 100
105 90.5 75.4 113.0
106 90.6 71.4 107.2



#8
Phenol
Concen: 21.372 ng/ul
RT: 7.157 min Scan# 658
Delta R.T. 0.000 min
Lab File: BP016990.D
Acq: 07 Sep 2023 16:24

Tgt Ion: 94 Resp: 622291
Ion Ratio Lower Upper
94 100
65 28.5 23.2 34.8
66 37.4 29.4 44.2









#10
Bis(2-Chloroethyl)ether
Concen: 21.739 ng/ul
RT: 7.416 min Scan# 701
Delta R.T. 0.000 min
Lab File: BP016990.D
Acq: 07 Sep 2023 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD020721


Tgt Ion: 93 Resp: 504767
Ion Ratio Lower Upper
93 100
63 78.4 63.4 95.2
95 32.5 26.1 39.1



#12
2-Chlorophenol
Concen: 21.293 ng/ul
RT: 7.534 min Scan# 722
Delta R.T. 0.000 min
Lab File: BP016990.D
Acq: 07 Sep 2023 16:24

Tgt Ion:128 Resp: 460105
Ion Ratio Lower Upper
128 100
64 51.2 40.2 60.2
130 33.6 25.2 37.8





#13

2-Methylphenol

Concen: 21.259 ng/ul

RT: 8.422 min Scan# 81

Delta R.T. 0.000 min

Lab File: BP016990.D

Acq: 07 Sep 2023 16:24

Instrument :

BNA_P

ClientSampleId :





SSTD020721

Tgt Ion:108 Resp: 454039

Ion Ratio Lower Upper

108 100

107 90.5 70.6 106.0



#14

2,2'-oxybis(1-Chloropropane)

Concen: 13.562 ng/ul

RT: 8.493 min Scan# 885

Delta R.T. -0.023 min

Lab File: BP016990.D

Acq: 07 Sep 2023 16:24




Tgt Ion: 45 Resp: 443374


Ion Ratio Lower Upper

45 100

77 14.1 11.3 16.9

79 9.8 9.0 13.4









#16
Acetophenone
Concen: 22.346 ng/ul
RT: 8.828 min Scan# 942
Delta R.T. -0.006 min
Lab File: BP016990.D
Acq: 07 Sep 2023 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD020721


Tgt Ion:105 Resp: 769880
Ion Ratio Lower Upper
105 100
77 82.9 63.6 95.4
51 30.2 23.4 35.0



#17
N-Nitroso-di-n-propylamine
Concen: 22.832 ng/ul
RT: 8.798 min Scan# 937
Delta R.T. -0.006 min
Lab File: BP016990.D
Acq: 07 Sep 2023 16:24

Tgt Ion: 70 Resp: 420213
Ion Ratio Lower Upper
70 100
42 60.9 47.4 71.2
101 9.0 7.3 10.9
130 18.6 15.6 23.4









#18
4-Methylphenol
Concen: 22.062 ng/ul
RT: 8.751 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP016990.D
Acq: 07 Sep 2023 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD020721


Tgt Ion:108 Resp: 513872
Ion Ratio Lower Upper
108 100
107 110.3 90.6 135.8



#19
Hexachloroethane
Concen: 21.404 ng/ul
RT: 9.045 min Scan# 979
Delta R.T. -0.006 min
Lab File: BP016990.D
Acq: 07 Sep 2023 16:24

Tgt Ion:117 Resp: 191595
Ion Ratio Lower Upper
117 100
201 88.6 75.8 113.6
199 57.7 46.8 70.2









#22
Nitrobenzene
Concen: 21.842 ng/ul
RT: 9.222 min Scan# 1009
Delta R.T. 0.000 min
Lab File: BP016990.D
Acq: 07 Sep 2023 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD020721


Tgt Ion: 77 Resp: 609546
Ion Ratio Lower Upper
77 100
123 41.5 34.4 51.6
65 13.4 10.6 15.8



#23
Isophorone
Concen: 21.483 ng/ul
RT: 9.734 min Scan# 1096
Delta R.T. -0.006 min
Lab File: BP016990.D
Acq: 07 Sep 2023 16:24

Tgt Ion: 82 Resp: 1126356
Ion Ratio Lower Upper
82 100
95 6.7 5.3 7.9
138 16.2 13.4 20.2









#25
2-Nitrophenol
Concen: 21.240 ng/ul
RT: 9.928 min Scan# 1128
Delta R.T. 0.000 min
Lab File: BP016990.D
Acq: 07 Sep 2023 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD020721


Tgt Ion:139 Resp: 264948
Ion Ratio Lower Upper
139 100
65 55.2 44.0 66.0
109 30.4 24.1 36.1



#26
2,4-Dimethylphenol
Concen: 21.569 ng/ul
RT: 9.963 min Scan# 1135
Delta R.T. -0.006 min
Lab File: BP016990.D
Acq: 07 Sep 2023 16:24

Tgt Ion:107 Resp: 549327
Ion Ratio Lower Upper
107 100
121 49.0 41.0 61.6
122 84.3 69.0 103.4





#27

Bis(2-Chloroethoxy)methane

Concen: 21.603 ng/ul

RT: 10.222 min Scan# 1179

Delta R.T. -0.006 min

Lab File: BP016990.D

Acq: 07 Sep 2023 16:24

Instrument :

BNA_P

ClientSampleId :

SSTD020721





Tgt Ion: 93 Resp: 701924

Ion Ratio Lower Upper

93 100

95 32.1 26.0 39.0

123 10.6 8.6 13.0



#29

2,4-Dichlorophenol

Concen: 21.189 ng/ul

RT: 10.445 min Scan# 1217

Delta R.T. -0.006 min

Lab File: BP016990.D

Acq: 07 Sep 2023 16:24




Tgt Ion:162 Resp: 451673


Ion Ratio Lower Upper

162 100

164 64.2 51.8 77.6

98 38.5 28.2 42.4









#30
Naphthalene
Concen: 21.138 ng/ul
RT: 10.857 min Scan# 1287
Delta R.T. -0.006 min
Lab File: BP016990.D
Acq: 07 Sep 2023 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD020721


Tgt Ion:128 Resp: 1553879
Ion Ratio Lower Upper
128 100
129 11.0 8.8 13.2
127 12.9 10.3 15.5



#32
4-Chloroaniline
Concen: 21.590 ng/ul
RT: 10.992 min Scan# 1310
Delta R.T. 0.000 min
Lab File: BP016990.D
Acq: 07 Sep 2023 16:24

Tgt Ion:127 Resp: 678039
Ion Ratio Lower Upper
127 100
129 31.9 26.0 39.0





#33

Hexachlorobutadiene

Concen: 19.999 ng/ul

RT: 11.087 min Scan# 11

Delta R.T. -0.006 min

Lab File: BP016990.D

Acq: 07 Sep 2023 16:24

Instrument :

BNA_P

ClientSampleId :

SSTD020721





Tgt Ion:225 Resp: 254809

Ion Ratio Lower Upper

225 100

223 63.7 50.1 75.1

227 64.2 50.9 76.3



#34

Caprolactam

Concen: 21.420 ng/ul

RT: 11.810 min Scan# 1449

Delta R.T. 0.000 min

Lab File: BP016990.D

Acq: 07 Sep 2023 16:24




Tgt Ion:113 Resp: 151215


Ion Ratio Lower Upper

113 100

55 184.5 142.3 213.5

56 132.8 102.2 153.2









#35
4-Chloro-3-methylphenol
Concen: 21.911 ng/ul
RT: 12.092 min Scan# 1496
Delta R.T. 0.000 min
Lab File: BP016990.D
Acq: 07 Sep 2023 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD020721


Tgt Ion:107 Resp: 511839
Ion Ratio Lower Upper
107 100
144 26.3 20.3 30.5
142 82.0 63.8 95.8



#36
2-Methylnaphthalene
Concen: 21.432 ng/ul
RT: 12.463 min Scan# 1560
Delta R.T. -0.006 min
Lab File: BP016990.D
Acq: 07 Sep 2023 16:24

Tgt Ion:142 Resp: 1041124
Ion Ratio Lower Upper
142 100
141 89.0 71.1 106.7









#37
1-Methylnaphthalene
Concen: 21.480 ng/ul
RT: 12.681 min Scan# 1597
Delta R.T. -0.006 min
Lab File: BP016990.D
Acq: 07 Sep 2023 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD020721


Tgt Ion:142 Resp: 1050576
Ion Ratio Lower Upper
142 100
141 92.3 74.5 111.7
116 3.8 3.1 4.7



#39
1,2,4,5-Tetrachlorobenzene
Concen: 20.130 ng/ul
RT: 12.810 min Scan# 1619
Delta R.T. -0.006 min
Lab File: BP016990.D
Acq: 07 Sep 2023 16:24

Tgt Ion:216 Resp: 500547
Ion Ratio Lower Upper
216 100
214 78.4 62.8 94.2
179 22.8 17.5 26.3
108 20.1 14.6 21.8









#40
Hexachlorocyclopentadiene
Concen: 16.430 ng/ul
RT: 12.763 min Scan# 1611
Delta R.T. 0.000 min
Lab File: BP016990.D
Acq: 07 Sep 2023 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD020721


Tgt Ion:237 Resp: 264706
Ion Ratio Lower Upper
237 100
235 62.5 50.8 76.2
272 12.5 10.0 15.0



#41
2,4,6-Trichlorophenol
Concen: 20.650 ng/ul
RT: 13.063 min Scan# 1662
Delta R.T. -0.006 min
Lab File: BP016990.D
Acq: 07 Sep 2023 16:24

Tgt Ion:196 Resp: 349473
Ion Ratio Lower Upper
196 100
198 95.2 75.8 113.6
200 30.2 24.6 36.8





#42

2,4,5-Trichlorophenol

Concen: 20.931 ng/ul

RT: 13.133 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP016990.D

Acq: 07 Sep 2023 16:24

Instrument :

BNA_P

ClientSampleId :

SSTD020721





Tgt Ion:196 Resp: 385432

Ion Ratio Lower Upper

196 100

198 94.6 74.9 112.3

200 30.0 24.5 36.7



#43

1,1'-Biphenyl

Concen: 21.102 ng/ul

RT: 13.469 min Scan# 1731

Delta R.T. -0.006 min

Lab File: BP016990.D

Acq: 07 Sep 2023 16:24




Tgt Ion:154 Resp: 1409722


Ion Ratio Lower Upper

154 100

153 40.6 32.7 49.1

76 14.9 11.5 17.3









#44
2-Chloronaphthalene
Concen: 21.039 ng/ul
RT: 13.516 min Scan# 1739
Delta R.T. -0.006 min
Lab File: BP016990.D
Acq: 07 Sep 2023 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD020721


Tgt Ion:162 Resp: 1103530
Ion Ratio Lower Upper
162 100
164 32.4 26.2 39.4
127 39.1 30.8 46.2



#45
2-Nitroaniline
Concen: 22.453 ng/ul
RT: 13.751 min Scan# 1779
Delta R.T. 0.000 min
Lab File: BP016990.D
Acq: 07 Sep 2023 16:24

Tgt Ion: 65 Resp: 350734
Ion Ratio Lower Upper
65 100
92 66.8 55.4 83.0
138 97.3 76.8 115.2









#47
Dimethylphthalate
Concen: 21.003 ng/ul
RT: 14.098 min Scan# 1838
Delta R.T. -0.006 min
Lab File: BP016990.D
Acq: 07 Sep 2023 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD020721


Tgt Ion:163 Resp: 1384427
Ion Ratio Lower Upper
163 100
194 3.9 3.1 4.7
164 10.0 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 21.058 ng/ul
RT: 14.239 min Scan# 1862
Delta R.T. -0.006 min
Lab File: BP016990.D
Acq: 07 Sep 2023 16:24

Tgt Ion:165 Resp: 269569
Ion Ratio Lower Upper
165 100
89 65.8 50.0 75.0
121 23.2 17.8 26.6









#50
Acenaphthylene
Concen: 21.306 ng/ul
RT: 14.363 min Scan# 1884
Delta R.T. -0.006 min
Lab File: BP016990.D
Acq: 07 Sep 2023 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD020721


Tgt Ion:152 Resp: 1729663
Ion Ratio Lower Upper
152 100
151 20.0 15.9 23.9
153 13.2 10.5 15.7



#51
3-Nitroaniline
Concen: 22.059 ng/ul
RT: 14.580 min Scan# 1920
Delta R.T. 0.000 min
Lab File: BP016990.D
Acq: 07 Sep 2023 16:24

Tgt Ion:138 Resp: 281813
Ion Ratio Lower Upper
138 100
108 10.7 8.4 12.6
92 120.6 97.2 145.8





#52

Acenaphthene

Concen: 21.292 ng/ul

RT: 14.704 min Scan# 1941

Delta R.T. 0.000 min

Lab File: BP016990.D

Acq: 07 Sep 2023 16:24

Instrument :

BNA_P

ClientSampleId :

SSTD020721





Tgt Ion:153 Resp: 1205987

Ion Ratio Lower Upper

153 100

152 48.5 38.6 57.8

154 87.0 70.2 105.4



#53

2,4-Dinitrophenol

Concen: 15.608 ng/ul

RT: 14.780 min Scan# 1954

Delta R.T. 0.000 min

Lab File: BP016990.D

Acq: 07 Sep 2023 16:24




Tgt Ion:184 Resp: 123882


Ion Ratio Lower Upper

184 100

63 71.8 61.6 92.4

154 60.0 50.8 76.2









#55
4-Nitrophenol
Concen: 21.183 ng/ul
RT: 14.869 min Scan# 1969
Delta R.T. 0.000 min
Lab File: BP016990.D
Acq: 07 Sep 2023 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD020721


Tgt Ion:109 Resp: 208048
Ion Ratio Lower Upper
109 100
139 132.0 105.0 157.4
65 137.0 108.4 162.6



#56
Dibenzofuran
Concen: 21.284 ng/ul
RT: 15.039 min Scan# 1998
Delta R.T. 0.000 min
Lab File: BP016990.D
Acq: 07 Sep 2023 16:24

Tgt Ion:168 Resp: 1657647
Ion Ratio Lower Upper
168 100
139 37.7 30.1 45.1









#57
2,4-Dinitrotoluene
Concen: 21.661 ng/ul
RT: 15.022 min Scan# 1995
Delta R.T. -0.006 min
Lab File: BP016990.D
Acq: 07 Sep 2023 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD020721


Tgt Ion:165 Resp: 398044
Ion Ratio Lower Upper
165 100
63 56.0 42.1 63.1
182 2.5 2.3 3.5



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.620 ng/ul
RT: 15.251 min Scan# 2034
Delta R.T. -0.006 min
Lab File: BP016990.D
Acq: 07 Sep 2023 16:24

Tgt Ion:232 Resp: 306132
Ion Ratio Lower Upper
232 100
131 52.7 39.2 58.8
130 2.8 2.0 3.0
166 34.6 27.2 40.8









#59
Diethylphthalate
Concen: 21.422 ng/ul
RT: 15.451 min Scan# 2068
Delta R.T. -0.006 min
Lab File: BP016990.D
Acq: 07 Sep 2023 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD020721


Tgt Ion:149 Resp: 1411058
Ion Ratio Lower Upper
149 100
177 21.5 17.3 25.9
150 12.9 10.1 15.1



#61
Fluorene
Concen: 21.295 ng/ul
RT: 15.692 min Scan# 2109
Delta R.T. 0.000 min
Lab File: BP016990.D
Acq: 07 Sep 2023 16:24

Tgt Ion:166 Resp: 1367329
Ion Ratio Lower Upper
166 100
165 97.7 78.3 117.5
167 12.9 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 20.581 ng/ul
RT: 15.680 min Scan# 2107
Delta R.T. -0.006 min
Lab File: BP016990.D
Acq: 07 Sep 2023 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD020721


Tgt Ion:204 Resp: 638635
Ion Ratio Lower Upper
204 100
206 32.8 26.4 39.6
141 62.9 46.2 69.4



#63
4-Nitroaniline
Concen: 23.280 ng/ul
RT: 15.751 min Scan# 2119
Delta R.T. 0.000 min
Lab File: BP016990.D
Acq: 07 Sep 2023 16:24

Tgt Ion:138 Resp: 261485
Ion Ratio Lower Upper
138 100
92 54.6 41.9 62.9
108 96.1 70.3 105.5









#66
4,6-Dinitro-2-methylphenol
Concen: 18.255 ng/ul
RT: 15.775 min Scan# 2123
Delta R.T. -0.006 min
Lab File: BP016990.D
Acq: 07 Sep 2023 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD020721


Tgt Ion:198 Resp: 207591
Ion Ratio Lower Upper
198 100
182 4.5 3.6 5.4
77 31.1 21.4 32.0



#67
N-Nitrosodiphenylamine
Concen: 21.116 ng/ul
RT: 15.904 min Scan# 2145
Delta R.T. -0.006 min
Lab File: BP016990.D
Acq: 07 Sep 2023 16:24

Tgt Ion:169 Resp: 1148404
Ion Ratio Lower Upper
169 100
168 67.8 54.2 81.4
167 37.0 29.4 44.2









#68
4-Bromophenyl-phenylether
Concen: 20.110 ng/ul
RT: 16.586 min Scan# 2261
Delta R.T. 0.000 min
Lab File: BP016990.D
Acq: 07 Sep 2023 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD020721


Tgt Ion:248 Resp: 360844
Ion Ratio Lower Upper
248 100
250 96.2 76.8 115.2
141 80.2 66.3 99.5



#69
Hexachlorobenzene
Concen: 19.958 ng/ul
RT: 16.663 min Scan# 2274
Delta R.T. -0.006 min
Lab File: BP016990.D
Acq: 07 Sep 2023 16:24

Tgt Ion:284 Resp: 392564
Ion Ratio Lower Upper
284 100
142 45.1 31.9 47.9
249 33.5 27.0 40.6





#70

Atrazine

Concen: 20.479 ng/ul

RT: 16.857 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP016990.D

Acq: 07 Sep 2023 16:24

Instrument :

BNA_P

ClientSampleId :

SSTD020721





Tgt Ion:200 Resp: 381328

Ion Ratio Lower Upper

200 100

173 27.9 21.8 32.8

215 46.1 38.3 57.5



#71

Pentachlorophenol

Concen: 18.652 ng/ul

RT: 17.027 min Scan# 2336

Delta R.T. -0.006 min

Lab File: BP016990.D

Acq: 07 Sep 2023 16:24




Tgt Ion:266 Resp: 239251


Ion Ratio Lower Upper

266 100

264 63.2 50.4 75.6

268 61.9 50.4 75.6





#72

Phenanthrene

Concen: 21.223 ng/ul

RT: 17.451 min Scan# 2408

Delta R.T. 0.000 min

Lab File: BP016990.D

Acq: 07 Sep 2023 16:24

Instrument :

BNA_P

ClientSampleId :

SSTD020721





Tgt Ion:178 Resp: 2127032

Ion Ratio Lower Upper

178 100

179 15.4 12.5 18.7

176 19.8 15.8 23.8



#74

Anthracene

Concen: 21.356 ng/ul

RT: 17.539 min Scan# 2423

Delta R.T. -0.006 min

Lab File: BP016990.D

Acq: 07 Sep 2023 16:24




Tgt Ion:178 Resp: 2151613


Ion Ratio Lower Upper

178 100

179 15.5 12.3 18.5

176 19.2 15.0 22.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 19.896 ng/uL

RT: 13.422 min Scan# 1723

Delta R.T. -0.006 min

Lab File: BP016990.D

Acq: 07 Sep 2023 16:24

Instrument :

BNA_P

ClientSampleId :

SSTD020721





Tgt Ion:216 Resp: 511165

Ion Ratio Lower Upper

216 100

214 79.9 64.5 96.7

218 47.9 38.6 57.8



#76

Pentachlorobenzene

Concen: 20.057 ng/uL

RT: 14.928 min Scan# 1979

Delta R.T. -0.006 min

Lab File: BP016990.D

Acq: 07 Sep 2023 16:24




Tgt Ion:250 Resp: 496352


Ion Ratio Lower Upper

250 100

248 64.9 51.2 76.8

252 63.3 50.9 76.3









#77
 Carbazole
 Concen: 21.886 ng/ul
 RT: 17.827 min Scan# 2472
 Delta R.T. 0.000 min
 Lab File: BP016990.D
 Acq: 07 Sep 2023 16:24




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020721


Tgt Ion:167 Resp: 1981495
 Ion Ratio Lower Upper
 167 100
 166 21.7 17.3 25.9
 139 13.2 10.3 15.5



#78
 Di-n-butylphthalate
 Concen: 21.531 ng/ul
 RT: 18.339 min Scan# 2559
 Delta R.T. 0.000 min
 Lab File: BP016990.D
 Acq: 07 Sep 2023 16:24

Tgt Ion:149 Resp: 2388146
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.4 11.2
 104 5.3 4.2 6.4





#80

Fluoranthene

Concen: 20.459 ng/ul

RT: 19.416 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP016990.D

Acq: 07 Sep 2023 16:24

Instrument :

BNA_P

ClientSampleId :

SSTD020721





Tgt Ion:202 Resp: 2377884

Ion Ratio Lower Upper

202 100

101 12.4 10.0 15.0

100 9.4 7.5 11.3



#82

Pyrene

Concen: 20.691 ng/ul

RT: 19.774 min Scan# 2803

Delta R.T. 0.000 min

Lab File: BP016990.D

Acq: 07 Sep 2023 16:24




Tgt Ion:202 Resp: 2496942


Ion Ratio Lower Upper

202 100

101 13.8 11.3 16.9

100 11.5 9.4 14.0









#83
Butylbenzylphthalate
Concen: 21.331 ng/ul
RT: 20.633 min Scan# 2949
Delta R.T. 0.000 min
Lab File: BP016990.D
Acq: 07 Sep 2023 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD020721


Tgt Ion:149 Resp: 1064528
Ion Ratio Lower Upper
149 100
91 74.9 59.1 88.7
206 19.7 15.7 23.5



#84
3,3'-Dichlorobenzidine
Concen: 20.944 ng/ul
RT: 21.427 min Scan# 3084
Delta R.T. 0.000 min
Lab File: BP016990.D
Acq: 07 Sep 2023 16:24

Tgt Ion:252 Resp: 710318
Ion Ratio Lower Upper
252 100
254 64.2 51.8 77.6
126 12.9 10.4 15.6





#85

Benzo(a)anthracene

Concen: 21.265 ng/ul

RT: 21.486 min Scan# 3095

Delta R.T. -0.006 min

Lab File: BP016990.D

Acq: 07 Sep 2023 16:24

Instrument :

BNA_P

ClientSampleId :

SSTD020721





Tgt Ion:228 Resp: 2288705

Ion Ratio Lower Upper

228 100

229 19.9 15.8 23.8

226 28.3 22.2 33.2



#86

Bis(2-ethylhexyl)phthalate

Concen: 22.015 ng/ul

RT: 21.368 min Scan# 3074

Delta R.T. 0.000 min

Lab File: BP016990.D

Acq: 07 Sep 2023 16:24




Tgt Ion:149 Resp: 1517802


Ion Ratio Lower Upper

149 100

167 26.6 21.4 32.2

279 3.5 3.0 4.4





#87

Chrysene

Concen: 21.195 ng/ul

RT: 21.545 min Scan# 3105

Delta R.T. 0.000 min

Lab File: BP016990.D

Acq: 07 Sep 2023 16:24

Instrument :

BNA_P

ClientSampleId :

SSTD020721





Tgt Ion:228 Resp: 2133865

Ion Ratio Lower Upper

228 100

226 31.0 24.6 36.8

229 20.3 15.9 23.9



#89

Di-n-octyl phthalate

Concen: 19.996 ng/ul




RT: 22.333 min Scan# 3238


Delta R.T. -0.006 min

Lab File: BP016990.D

Acq: 07 Sep 2023 16:24

Tgt Ion:149 Resp: 2490239





#90

Benzo(b)fluoranthene

Concen: 20.527 ng/ul

RT: 23.251 min Scan# 31

Delta R.T. -0.006 min

Lab File: BP016990.D

Acq: 07 Sep 2023 16:24

Instrument :

BNA_P

ClientSampleId :

SSTD020721





Tgt Ion:252 Resp: 2130670

Ion Ratio Lower Upper

252 100

253 21.9 17.6 26.4

125 12.2 8.8 13.2



#91

Benzo(k)fluoranthene

Concen: 20.558 ng/ul

RT: 23.309 min Scan# 3404

Delta R.T. 0.000 min

Lab File: BP016990.D

Acq: 07 Sep 2023 16:24




Tgt Ion:252 Resp: 2066432


Ion Ratio Lower Upper

252 100

253 21.5 17.7 26.5

125 11.1 8.9 13.3





#93

Benzo(a)pyrene

Concen: 20.823 ng/ul

RT: 23.933 min Scan# 3511

Delta R.T. 0.000 min

Lab File: BP016990.D

Acq: 07 Sep 2023 16:24

Instrument :

BNA_P

ClientSampleId :

SSTD020721





Tgt Ion:252 Resp: 1802781

Ion Ratio Lower Upper

252 100

253 21.7 17.8 26.6

125 12.5 9.9 14.9



#94

Indeno(1,2,3-cd)pyrene

Concen: 21.253 ng/ul

RT: 26.739 min Scan# 3987

Delta R.T. -0.006 min

Lab File: BP016990.D

Acq: 07 Sep 2023 16:24




Tgt Ion:276 Resp: 2042526


Ion Ratio Lower Upper

276 100

138 27.2 21.4 32.0

277 24.3 20.0 30.0





#95

Dibenzo(a,h)anthracene

Concen: 21.357 ng/ul

RT: 26.762 min Scan# 3991

Delta R.T. -0.006 min

Lab File: BP016990.D

Acq: 07 Sep 2023 16:24

Instrument :

BNA_P

ClientSampleId :

SSTD020721





Tgt Ion:278 Resp: 1694021

Ion Ratio Lower Upper

278 100

139 19.3 14.9 22.3

279 24.1 19.6 29.4



#96

Benzo(g,h,i)perylene

Concen: 21.098 ng/ul

RT: 27.586 min Scan# 4131

Delta R.T. -0.006 min

Lab File: BP016990.D

Acq: 07 Sep 2023 16:24

Tgt Ion:276 Resp: 1597863

Ion Ratio Lower Upper

276 100

138 24.3 19.5 29.3

277 23.2 18.7 28.1

