

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090823\
 Data File : BP017005.D
 Acq On : 08 Sep 2023 11:03
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020722

Quant Time: Sep 09 00:42:21 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP090523.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Sep 06 01:52:05 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.975	152	339010	20.000	ng/u1	0.00
20) Naphthalene-d8	10.804	136	1473452	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.634	164	905007	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.404	188	1860417	20.000	ng/u1	0.00
79) Chrysene-d12	21.504	240	1388117	20.000	ng/u1	0.00
88) Perylene-d12	24.039	264	1240433	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.322	96	79124	8.828	ng/uL	0.00
4) Pyridine-d5	3.764	84	558345	21.803	ng/u1	0.00
7) Phenol-d5	7.122	99	656345	21.192	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.316	67	431495	22.273	ng/u1	0.00
11) 2-Chlorophenol-d4	7.493	132	484833	21.629	ng/u1	-0.01
15) 4-Methylphenol-d8	8.687	113	503262	21.329	ng/u1	0.00
21) Nitrobenzene-d5	9.175	128	238376	20.853	ng/u1	0.00
24) 2-Nitrophenol-d4	9.887	143	258790	20.884	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.410	165	476833	20.843	ng/u1	-0.01
31) 4-Chloroaniline-d4	10.963	131	684214	20.730	ng/u1	0.00
46) Dimethylphthalate-d6	14.045	166	1385714	20.570	ng/u1	0.00
49) Acenaphthylene-d8	14.334	160	1634620	21.025	ng/u1	0.00
54) 4-Nitrophenol-d4	14.845	143	249872	18.988	ng/u1	-0.01
60) Fluorene-d10	15.628	176	1207771	20.940	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.757	200	191786	17.130	ng/u1	0.00
73) Anthracene-d10	17.504	188	1802717	21.100	ng/u1	0.00
81) Pyrene-d10	19.739	212	1839727	21.759	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.874	264	1301048	20.675	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.364	88	83601	8.640	ng/uL	99
5) Pyridine	3.781	79	586385	21.509	ng/u1	99
6) Benzaldehyde	7.134	77	373350	26.597	ng/u1	99
8) Phenol	7.152	94	688475	21.395	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.410	93	557469	21.724	ng/u1	99
12) 2-Chlorophenol	7.528	128	514962	21.564	ng/u1	98
13) 2-Methylphenol	8.416	108	499475	21.161	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.487	45	476784	13.196	ng/u1	98
16) Acetophenone	8.828	105	835493	21.943	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.799	70	450897	22.168	ng/u1	99
18) 4-Methylphenol	8.752	108	554238	21.531	ng/u1	97
19) Hexachloroethane	9.046	117	214465	21.679	ng/u1	95
22) Nitrobenzene	9.216	77	665861	21.815	ng/u1	97
23) Isophorone	9.734	82	1213545	21.163	ng/u1	99
25) 2-Nitrophenol	9.922	139	292347	21.428	ng/u1	98
26) 2,4-Dimethylphenol	9.957	107	597507	21.450	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.216	93	771543	21.710	ng/u1	98
29) 2,4-Dichlorophenol	10.440	162	491098	21.064	ng/u1	98
30) Naphthalene	10.851	128	1700986	21.156	ng/u1	100
32) 4-Chloroaniline	10.987	127	719586	20.949	ng/u1	100
33) Hexachlorobutadiene	11.081	225	272275	19.539	ng/u1	97
34) Caprolactam	11.810	113	152461	19.745	ng/u1	92
35) 4-Chloro-3-methylphenol	12.087	107	553949	21.682	ng/u1	98
36) 2-Methylnaphthalene	12.457	142	1130238	21.272	ng/u1	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090823\
 Data File : BP017005.D
 Acq On : 08 Sep 2023 11:03
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020722

Quant Time: Sep 09 00:42:21 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP090523.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Sep 06 01:52:05 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.675	142	1143085	21.369	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.804	216	534562	20.048	ng/ul	98
40) Hexachlorocyclopentadiene	12.757	237	279403	16.172	ng/ul	98
41) 2,4,6-Trichlorophenol	13.063	196	369966	20.386	ng/ul	99
42) 2,4,5-Trichlorophenol	13.128	196	402320	20.375	ng/ul	99
43) 1,1'-Biphenyl	13.469	154	1526130	21.304	ng/ul	100
44) 2-Chloronaphthalene	13.516	162	1183972	21.050	ng/ul	99
45) 2-Nitroaniline	13.745	65	373189	22.279	ng/ul	95
47) Dimethylphthalate	14.092	163	1465908	20.739	ng/ul	99
48) 2,6-Dinitrotoluene	14.240	165	288722	21.033	ng/ul	98
50) Acenaphthylene	14.363	152	1849654	21.248	ng/ul	100
51) 3-Nitroaniline	14.575	138	284336	20.755	ng/ul	99
52) Acenaphthene	14.698	153	1291179	21.258	ng/ul	100
53) 2,4-Dinitrophenol	14.775	184	114448	13.447	ng/ul	99
55) 4-Nitrophenol	14.863	109	207502	19.702	ng/ul	96
56) Dibenzofuran	15.034	168	1778768	21.299	ng/ul	99
57) 2,4-Dinitrotoluene	15.022	165	417146	21.170	ng/ul	94
58) 2,3,4,6-Tetrachlorophenol	15.251	232	313244	19.676	ng/ul	95
59) Diethylphthalate	15.451	149	1486035	21.038	ng/ul	99
61) Fluorene	15.687	166	1457909	21.174	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.675	204	679589	20.424	ng/ul	95
63) 4-Nitroaniline	15.745	138	250735	20.817	ng/ul	95
66) 4,6-Dinitro-2-methylph...	15.775	198	203113	17.316	ng/ul	98
67) N-Nitrosodiphenylamine	15.904	169	1199624	21.385	ng/ul	99
68) 4-Bromophenyl-phenylether	16.581	248	376006	20.315	ng/ul	98
69) Hexachlorobenzene	16.663	284	402422	19.834	ng/ul	95
70) Atrazine	16.857	200	389329	20.270	ng/ul	98
71) Pentachlorophenol	17.028	266	229411	17.339	ng/ul	99
72) Phenanthrene	17.445	178	2198932	21.271	ng/ul	99
74) Anthracene	17.539	178	2215133	21.316	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.422	216	543996	20.528	ng/uL	98
76) Pentachlorobenzene	14.928	250	519814	20.364	ng/uL	99
77) Carbazole	17.822	167	1972869	21.126	ng/ul	100
78) Di-n-butylphthalate	18.333	149	2411428	21.077	ng/ul	100
80) Fluoranthene	19.410	202	2333010	22.099	ng/ul	98
82) Pyrene	19.769	202	2440815	22.268	ng/ul	98
83) Butylbenzylphthalate	20.627	149	984356	21.715	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.427	252	600664	19.498	ng/ul	100
85) Benzo(a)anthracene	21.486	228	2077101	21.246	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.363	149	1360816	21.730	ng/ul	98
87) Chrysene	21.545	228	1939725	21.211	ng/ul	100
89) Di-n-octyl phthalate	22.333	149	2101318	20.208	ng/ul	100
90) Benzo(b)fluoranthene	23.251	252	1770103	20.424	ng/ul	99
91) Benzo(k)fluoranthene	23.304	252	1759918	20.970	ng/ul	99
93) Benzo(a)pyrene	23.927	252	1503768	20.803	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.727	276	1762338	21.963	ng/ul	99
95) Dibenzo(a,h)anthracene	26.757	278	1454090	21.956	ng/ul	98
96) Benzo(g,h,i)perylene	27.574	276	1413268	22.350	ng/ul	98





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 8.640 ng/uL
RT: 3.364 min Scan# 47
Delta R.T. 0.001 min
Lab File: BP017005.D
Acq: 08 Sep 2023 11:03




Instrument :
BNA_P
ClientSampleId :
SSTD020722


Tgt Ion: 88 Resp: 83601
Ion Ratio Lower Upper
88 100
43 30.6 25.0 37.6
58 74.8 60.6 90.8



#5
Pyridine
Concen: 21.509 ng/uL
RT: 3.781 min Scan# 118
Delta R.T. -0.006 min
Lab File: BP017005.D
Acq: 08 Sep 2023 11:03

Tgt Ion: 79 Resp: 586385
Ion Ratio Lower Upper
79 100
52 76.1 60.1 90.1
51 37.4 30.5 45.7









#6
Benzaldehyde
Concen: 26.597 ng/ul
RT: 7.134 min Scan# 688
Delta R.T. -0.006 min
Lab File: BP017005.D
Acq: 08 Sep 2023 11:03




Instrument :
BNA_P
ClientSampleId :
SSTD020722


Tgt Ion: 77 Resp: 373350
Ion Ratio Lower Upper
77 100
105 92.5 75.4 113.0
106 88.4 71.4 107.2



#8
Phenol
Concen: 21.395 ng/ul
RT: 7.152 min Scan# 691
Delta R.T. -0.006 min
Lab File: BP017005.D
Acq: 08 Sep 2023 11:03

Tgt Ion: 94 Resp: 688475
Ion Ratio Lower Upper
94 100
65 29.4 23.2 34.8
66 37.3 29.4 44.2









#10
 Bis(2-Chloroethyl)ether
 Concen: 21.724 ng/ul
 RT: 7.410 min Scan# 735
 Delta R.T. -0.006 min
 Lab File: BP017005.D
 Acq: 08 Sep 2023 11:03

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020722

Tgt Ion: 93 Resp: 557469




Ion	Ratio	Lower	Upper
93	100		
63	80.5	63.4	95.2
95	32.2	26.1	39.1




#12
 2-Chlorophenol
 Concen: 21.564 ng/ul
 RT: 7.528 min Scan# 755
 Delta R.T. -0.006 min
 Lab File: BP017005.D
 Acq: 08 Sep 2023 11:03

Tgt Ion:128 Resp: 514962

Ion	Ratio	Lower	Upper
128	100		
64	51.7	40.2	60.2
130	32.2	25.2	37.8





#13

2-Methylphenol

Concen: 21.161 ng/ul

RT: 8.416 min Scan# 906

Delta R.T. -0.006 min

Lab File: BP017005.D

Acq: 08 Sep 2023 11:03

Instrument :

BNA_P

ClientSampleId :





SSTD020722

Tgt Ion:108 Resp: 499475

Ion Ratio Lower Upper

108 100

107 87.5 70.6 106.0



#14

2,2'-oxybis(1-Chloropropane)

Concen: 13.196 ng/ul

RT: 8.487 min Scan# 918

Delta R.T. -0.029 min

Lab File: BP017005.D

Acq: 08 Sep 2023 11:03




Tgt Ion: 45 Resp: 476784


Ion Ratio Lower Upper

45 100

77 13.4 11.3 16.9

79 10.1 9.0 13.4









#16
Acetophenone
Concen: 21.943 ng/ul
RT: 8.828 min Scan# 976
Delta R.T. -0.006 min
Lab File: BP017005.D
Acq: 08 Sep 2023 11:03




Instrument :
BNA_P
ClientSampleId :
SSTD020722


Tgt Ion:105 Resp: 835493
Ion Ratio Lower Upper
105 100
77 81.5 63.6 95.4
51 29.0 23.4 35.0



#17
N-Nitroso-di-n-propylamine
Concen: 22.168 ng/ul
RT: 8.799 min Scan# 971
Delta R.T. -0.006 min
Lab File: BP017005.D
Acq: 08 Sep 2023 11:03

Tgt Ion: 70 Resp: 450897
Ion Ratio Lower Upper
70 100
42 60.3 47.4 71.2
101 9.5 7.3 10.9
130 19.1 15.6 23.4









#18
4-Methylphenol
Concen: 21.531 ng/ul
RT: 8.752 min Scan# 963
Delta R.T. -0.006 min
Lab File: BP017005.D
Acq: 08 Sep 2023 11:03




Instrument :
BNA_P
ClientSampleId :
SSTD020722


Tgt Ion:108 Resp: 554238
Ion Ratio Lower Upper
108 100
107 110.2 90.6 135.8



#19
Hexachloroethane
Concen: 21.679 ng/ul
RT: 9.046 min Scan# 1013
Delta R.T. -0.006 min
Lab File: BP017005.D
Acq: 08 Sep 2023 11:03

Tgt Ion:117 Resp: 214465
Ion Ratio Lower Upper
117 100
201 88.9 75.8 113.6
199 55.8 46.8 70.2









#22
Nitrobenzene
Concen: 21.815 ng/ul
RT: 9.216 min Scan# 1042
Delta R.T. -0.006 min
Lab File: BP017005.D
Acq: 08 Sep 2023 11:03




Instrument :
BNA_P
ClientSampleId :
SSTD020722


Tgt Ion: 77 Resp: 665861
Ion Ratio Lower Upper
77 100
123 40.7 34.4 51.6
65 13.6 10.6 15.8



#23
Isophorone
Concen: 21.163 ng/ul
RT: 9.734 min Scan# 1130
Delta R.T. -0.006 min
Lab File: BP017005.D
Acq: 08 Sep 2023 11:03

Tgt Ion: 82 Resp: 1213545
Ion Ratio Lower Upper
82 100
95 6.7 5.3 7.9
138 16.1 13.4 20.2









#25
2-Nitrophenol
Concen: 21.428 ng/ul
RT: 9.922 min Scan# 1162
Delta R.T. -0.006 min
Lab File: BP017005.D
Acq: 08 Sep 2023 11:03




Instrument :
BNA_P
ClientSampleId :
SSTD020722


Tgt Ion:139 Resp: 292347
Ion Ratio Lower Upper
139 100
65 56.9 44.0 66.0
109 29.5 24.1 36.1



#26
2,4-Dimethylphenol
Concen: 21.450 ng/ul
RT: 9.957 min Scan# 1168
Delta R.T. -0.011 min
Lab File: BP017005.D
Acq: 08 Sep 2023 11:03

Tgt Ion:107 Resp: 597507
Ion Ratio Lower Upper
107 100
121 49.5 41.0 61.6
122 83.6 69.0 103.4









#27
 Bis(2-Chloroethoxy)methane
 Concen: 21.710 ng/ul
 RT: 10.216 min Scan# 1212
 Delta R.T. -0.011 min
 Lab File: BP017005.D
 Acq: 08 Sep 2023 11:03




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020722


Tgt Ion: 93 Resp: 771543
 Ion Ratio Lower Upper
 93 100
 95 31.6 26.0 39.0
 123 10.3 8.6 13.0



#29
 2,4-Dichlorophenol
 Concen: 21.064 ng/ul
 RT: 10.440 min Scan# 1250
 Delta R.T. -0.011 min
 Lab File: BP017005.D
 Acq: 08 Sep 2023 11:03

Tgt Ion: 162 Resp: 491098
 Ion Ratio Lower Upper
 162 100
 164 65.3 51.8 77.6
 98 37.8 28.2 42.4









#30
Naphthalene
Concen: 21.156 ng/ul
RT: 10.851 min Scan# 1320
Delta R.T. -0.011 min
Lab File: BP017005.D
Acq: 08 Sep 2023 11:03




Instrument :
BNA_P
ClientSampleId :
SSTD020722


Tgt Ion:128 Resp: 1700986
Ion Ratio Lower Upper
128 100
129 11.1 8.8 13.2
127 13.0 10.3 15.5



#32
4-Chloroaniline
Concen: 20.949 ng/ul
RT: 10.987 min Scan# 1343
Delta R.T. -0.006 min
Lab File: BP017005.D
Acq: 08 Sep 2023 11:03

Tgt Ion:127 Resp: 719586
Ion Ratio Lower Upper
127 100
129 32.7 26.0 39.0





#33

Hexachlorobutadiene

Concen: 19.539 ng/ul

RT: 11.081 min Scan# 11

Delta R.T. -0.011 min

Lab File: BP017005.D

Acq: 08 Sep 2023 11:03

Instrument :

BNA_P

ClientSampleId :

SSTD020722





Tgt Ion:225 Resp: 272275

Ion Ratio Lower Upper

225 100

223 64.2 50.1 75.1

227 65.9 50.9 76.3



#34

Caprolactam

Concen: 19.745 ng/ul

RT: 11.810 min Scan# 1483

Delta R.T. 0.000 min

Lab File: BP017005.D

Acq: 08 Sep 2023 11:03




Tgt Ion:113 Resp: 152461


Ion Ratio Lower Upper

113 100

55 189.6 142.3 213.5

56 135.9 102.2 153.2









#35
4-Chloro-3-methylphenol
Concen: 21.682 ng/ul
RT: 12.087 min Scan# 1530
Delta R.T. -0.006 min
Lab File: BP017005.D
Acq: 08 Sep 2023 11:03




Instrument :
BNA_P
ClientSampleId :
SSTD020722


Tgt Ion:107 Resp: 553949
Ion Ratio Lower Upper
107 100
144 25.3 20.3 30.5
142 77.6 63.8 95.8



#36
2-Methylnaphthalene
Concen: 21.272 ng/ul
RT: 12.457 min Scan# 1593
Delta R.T. -0.011 min
Lab File: BP017005.D
Acq: 08 Sep 2023 11:03

Tgt Ion:142 Resp: 1130238
Ion Ratio Lower Upper
142 100
141 89.5 71.1 106.7









#37
1-Methylnaphthalene
Concen: 21.369 ng/ul
RT: 12.675 min Scan# 1630
Delta R.T. -0.011 min
Lab File: BP017005.D
Acq: 08 Sep 2023 11:03




Instrument :
BNA_P
ClientSampleId :
SSTD020722


Tgt Ion:142 Resp: 1143085
Ion Ratio Lower Upper
142 100
141 92.1 74.5 111.7
116 4.0 3.1 4.7



#39
1,2,4,5-Tetrachlorobenzene
Concen: 20.048 ng/ul
RT: 12.804 min Scan# 1652
Delta R.T. -0.011 min
Lab File: BP017005.D
Acq: 08 Sep 2023 11:03

Tgt Ion:216 Resp: 534562
Ion Ratio Lower Upper
216 100
214 79.1 62.8 94.2
179 24.0 17.5 26.3
108 21.3 14.6 21.8





#40

Hexachlorocyclopentadiene

Concen: 16.172 ng/ul

RT: 12.757 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP017005.D

Acq: 08 Sep 2023 11:03

Instrument :

BNA_P

ClientSampleId :

SSTD020722





Tgt Ion:237 Resp: 279403

Ion Ratio Lower Upper

237 100

235 61.9 50.8 76.2

272 12.3 10.0 15.0



#41

2,4,6-Trichlorophenol

Concen: 20.386 ng/ul

RT: 13.063 min Scan# 1696

Delta R.T. -0.006 min

Lab File: BP017005.D

Acq: 08 Sep 2023 11:03




Tgt Ion:196 Resp: 369966


Ion Ratio Lower Upper

196 100

198 95.2 75.8 113.6

200 31.6 24.6 36.8









#42
2,4,5-Trichlorophenol
Concen: 20.375 ng/ul
RT: 13.128 min Scan# 1707
Delta R.T. -0.011 min
Lab File: BP017005.D
Acq: 08 Sep 2023 11:03




Instrument :
BNA_P
ClientSampleId :
SSTD020722


Tgt Ion:196 Resp: 402320
Ion Ratio Lower Upper
196 100
198 93.3 74.9 112.3
200 29.7 24.5 36.7



#43
1,1'-Biphenyl
Concen: 21.304 ng/ul
RT: 13.469 min Scan# 1765
Delta R.T. -0.006 min
Lab File: BP017005.D
Acq: 08 Sep 2023 11:03

Tgt Ion:154 Resp: 1526130
Ion Ratio Lower Upper
154 100
153 40.9 32.7 49.1
76 14.8 11.5 17.3









#44
2-Chloronaphthalene
Concen: 21.050 ng/ul
RT: 13.516 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP017005.D
Acq: 08 Sep 2023 11:03




Instrument :
BNA_P
ClientSampleId :
SSTD020722


Tgt Ion:162 Resp: 1183972
Ion Ratio Lower Upper
162 100
164 33.1 26.2 39.4
127 39.6 30.8 46.2



#45
2-Nitroaniline
Concen: 22.279 ng/ul
RT: 13.745 min Scan# 1812
Delta R.T. -0.006 min
Lab File: BP017005.D
Acq: 08 Sep 2023 11:03

Tgt Ion: 65 Resp: 373189
Ion Ratio Lower Upper
65 100
92 64.1 55.4 83.0
138 91.4 76.8 115.2









#47
Dimethylphthalate
Concen: 20.739 ng/ul
RT: 14.092 min Scan# 1871
Delta R.T. -0.011 min
Lab File: BP017005.D
Acq: 08 Sep 2023 11:03




Instrument :
BNA_P
ClientSampleId :
SSTD020722


Tgt Ion:163 Resp: 1465908
Ion Ratio Lower Upper
163 100
194 3.8 3.1 4.7
164 10.3 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 21.033 ng/ul
RT: 14.240 min Scan# 1896
Delta R.T. -0.006 min
Lab File: BP017005.D
Acq: 08 Sep 2023 11:03

Tgt Ion:165 Resp: 288722
Ion Ratio Lower Upper
165 100
89 63.9 50.0 75.0
121 21.9 17.8 26.6





#50

Acenaphthylene

Concen: 21.248 ng/ul

RT: 14.363 min Scan# 1917

Delta R.T. -0.006 min

Lab File: BP017005.D

Acq: 08 Sep 2023 11:03

Instrument :

BNA_P

ClientSampleId :

SSTD020722





Tgt Ion:152 Resp: 1849654

Ion Ratio Lower Upper

152 100

151 20.2 15.9 23.9

153 13.0 10.5 15.7



#51

3-Nitroaniline

Concen: 20.755 ng/ul

RT: 14.575 min Scan# 1953

Delta R.T. -0.006 min

Lab File: BP017005.D

Acq: 08 Sep 2023 11:03




Tgt Ion:138 Resp: 284336


Ion Ratio Lower Upper

138 100

108 10.3 8.4 12.6

92 122.7 97.2 145.8









#52
Acenaphthene
Concen: 21.258 ng/ul
RT: 14.698 min Scan# 1974
Delta R.T. -0.006 min
Lab File: BP017005.D
Acq: 08 Sep 2023 11:03




Instrument :
BNA_P
ClientSampleId :
SSTD020722


Tgt Ion:153 Resp: 1291179
Ion Ratio Lower Upper
153 100
152 47.8 38.6 57.8
154 88.1 70.2 105.4



#53
2,4-Dinitrophenol
Concen: 13.447 ng/ul
RT: 14.775 min Scan# 1987
Delta R.T. -0.006 min
Lab File: BP017005.D
Acq: 08 Sep 2023 11:03

Tgt Ion:184 Resp: 114448
Ion Ratio Lower Upper
184 100
63 79.0 61.6 92.4
154 63.6 50.8 76.2









#55
4-Nitrophenol
Concen: 19.702 ng/ul
RT: 14.863 min Scan# 2002
Delta R.T. -0.006 min
Lab File: BP017005.D
Acq: 08 Sep 2023 11:03




Instrument :
BNA_P
ClientSampleId :
SSTD020722


Tgt Ion:109 Resp: 207502
Ion Ratio Lower Upper
109 100
139 131.7 105.0 157.4
65 144.1 108.4 162.6



#56
Dibenzofuran
Concen: 21.299 ng/ul
RT: 15.034 min Scan# 2031
Delta R.T. -0.006 min
Lab File: BP017005.D
Acq: 08 Sep 2023 11:03

Tgt Ion:168 Resp: 1778768
Ion Ratio Lower Upper
168 100
139 38.5 30.1 45.1









#57
2,4-Dinitrotoluene
Concen: 21.170 ng/ul
RT: 15.022 min Scan# 2029
Delta R.T. -0.006 min
Lab File: BP017005.D
Acq: 08 Sep 2023 11:03




Instrument :
BNA_P
ClientSampleId :
SSTD020722


Tgt Ion:165 Resp: 417146
Ion Ratio Lower Upper
165 100
63 57.1 42.1 63.1
182 2.5 2.3 3.5



#58
2,3,4,6-Tetrachlorophenol
Concen: 19.676 ng/ul
RT: 15.251 min Scan# 2068
Delta R.T. -0.006 min
Lab File: BP017005.D
Acq: 08 Sep 2023 11:03

Tgt Ion:232 Resp: 313244
Ion Ratio Lower Upper
232 100
131 52.7 39.2 58.8
130 2.7 2.0 3.0
166 36.4 27.2 40.8









#59
Diethylphthalate
Concen: 21.038 ng/ul
RT: 15.451 min Scan# 2102
Delta R.T. -0.006 min
Lab File: BP017005.D
Acq: 08 Sep 2023 11:03




Instrument :
BNA_P
ClientSampleId :
SSTD020722


Tgt Ion:149 Resp: 1486035
Ion Ratio Lower Upper
149 100
177 21.0 17.3 25.9
150 12.2 10.1 15.1



#61
Fluorene
Concen: 21.174 ng/ul
RT: 15.687 min Scan# 2142
Delta R.T. -0.006 min
Lab File: BP017005.D
Acq: 08 Sep 2023 11:03

Tgt Ion:166 Resp: 1457909
Ion Ratio Lower Upper
166 100
165 99.3 78.3 117.5
167 13.3 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 20.424 ng/ul
RT: 15.675 min Scan# 2140
Delta R.T. -0.011 min
Lab File: BP017005.D
Acq: 08 Sep 2023 11:03




Instrument :
BNA_P
ClientSampleId :
SSTD020722


Tgt Ion:204 Resp: 679589
Ion Ratio Lower Upper
204 100
206 33.2 26.4 39.6
141 63.6 46.2 69.4



#63
4-Nitroaniline
Concen: 20.817 ng/ul
RT: 15.745 min Scan# 2152
Delta R.T. -0.006 min
Lab File: BP017005.D
Acq: 08 Sep 2023 11:03

Tgt Ion:138 Resp: 250735
Ion Ratio Lower Upper
138 100
92 56.9 41.9 62.9
108 91.3 70.3 105.5









#66
4,6-Dinitro-2-methylphenol
Concen: 17.316 ng/ul
RT: 15.775 min Scan# 2157
Delta R.T. -0.006 min
Lab File: BP017005.D
Acq: 08 Sep 2023 11:03




Instrument :
BNA_P
ClientSampleId :
SSTD020722


Tgt Ion:198 Resp: 203113
Ion Ratio Lower Upper
198 100
182 4.7 3.6 5.4
77 27.8 21.4 32.0



#67
N-Nitrosodiphenylamine
Concen: 21.385 ng/ul
RT: 15.904 min Scan# 2179
Delta R.T. -0.006 min
Lab File: BP017005.D
Acq: 08 Sep 2023 11:03

Tgt Ion:169 Resp: 1199624
Ion Ratio Lower Upper
169 100
168 68.3 54.2 81.4
167 37.4 29.4 44.2









#68
4-Bromophenyl-phenylether
Concen: 20.315 ng/ul
RT: 16.581 min Scan# 2294
Delta R.T. -0.006 min
Lab File: BP017005.D
Acq: 08 Sep 2023 11:03




Instrument :
BNA_P
ClientSampleId :
SSTD020722


Tgt Ion:248 Resp: 376006
Ion Ratio Lower Upper
248 100
250 97.1 76.8 115.2
141 85.2 66.3 99.5



#69
Hexachlorobenzene
Concen: 19.834 ng/ul
RT: 16.663 min Scan# 2308
Delta R.T. -0.006 min
Lab File: BP017005.D
Acq: 08 Sep 2023 11:03

Tgt Ion:284 Resp: 402422
Ion Ratio Lower Upper
284 100
142 44.5 31.9 47.9
249 32.7 27.0 40.6





#70

Atrazine

Concen: 20.270 ng/ul

RT: 16.857 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP017005.D

Acq: 08 Sep 2023 11:03

Instrument :

BNA_P

ClientSampleId :

SSTD020722





Tgt Ion:200 Resp: 389329

Ion Ratio Lower Upper

200 100

173 26.9 21.8 32.8

215 46.1 38.3 57.5



#71

Pentachlorophenol

Concen: 17.339 ng/ul

RT: 17.028 min Scan# 2370

Delta R.T. -0.006 min

Lab File: BP017005.D

Acq: 08 Sep 2023 11:03




Tgt Ion:266 Resp: 229411


Ion Ratio Lower Upper

266 100

264 62.3 50.4 75.6

268 62.2 50.4 75.6





#72

Phenanthrene

Concen: 21.271 ng/ul

RT: 17.445 min Scan# 2441

Delta R.T. -0.006 min

Lab File: BP017005.D

Acq: 08 Sep 2023 11:03

Instrument :

BNA_P

ClientSampleId :

SSTD020722





Tgt Ion:178 Resp: 2198932

Ion Ratio Lower Upper

178 100

179 15.7 12.5 18.7

176 19.3 15.8 23.8



#74

Anthracene

Concen: 21.316 ng/ul

RT: 17.539 min Scan# 2457

Delta R.T. -0.006 min

Lab File: BP017005.D

Acq: 08 Sep 2023 11:03




Tgt Ion:178 Resp: 2215133


Ion Ratio Lower Upper

178 100

179 15.5 12.3 18.5

176 19.0 15.0 22.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 20.528 ng/uL

RT: 13.422 min Scan# 1757

Delta R.T. -0.006 min

Lab File: BP017005.D

Acq: 08 Sep 2023 11:03

Instrument :

BNA_P

ClientSampleId :

SSTD020722





Tgt Ion:216 Resp: 543996

Ion Ratio Lower Upper

216 100

214 78.8 64.5 96.7

218 47.1 38.6 57.8



#76

Pentachlorobenzene

Concen: 20.364 ng/uL

RT: 14.928 min Scan# 2013

Delta R.T. -0.006 min

Lab File: BP017005.D

Acq: 08 Sep 2023 11:03




Tgt Ion:250 Resp: 519814


Ion Ratio Lower Upper

250 100

248 63.1 51.2 76.8

252 63.3 50.9 76.3





#77

Carbazole

Concen: 21.126 ng/ul

RT: 17.822 min Scan# 2505

Delta R.T. -0.006 min

Lab File: BP017005.D

Acq: 08 Sep 2023 11:03

Instrument :

BNA_P

ClientSampleId :

SSTD020722





Tgt Ion:167 Resp: 1972869

Ion Ratio Lower Upper

167 100

166 21.8 17.3 25.9

139 13.1 10.3 15.5



#78

Di-n-butylphthalate

Concen: 21.077 ng/ul

RT: 18.333 min Scan# 2592

Delta R.T. -0.006 min

Lab File: BP017005.D

Acq: 08 Sep 2023 11:03




Tgt Ion:149 Resp: 2411428


Ion Ratio Lower Upper

149 100

150 9.3 7.4 11.2

104 5.5 4.2 6.4





#80

Fluoranthene

Concen: 22.099 ng/ul

RT: 19.410 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP017005.D

Acq: 08 Sep 2023 11:03

Instrument :

BNA_P

ClientSampleId :

SSTD020722





Tgt Ion:202 Resp: 2333010

Ion Ratio Lower Upper

202 100

101 13.2 10.0 15.0

100 10.2 7.5 11.3



#82

Pyrene

Concen: 22.268 ng/ul

RT: 19.769 min Scan# 2836

Delta R.T. -0.006 min

Lab File: BP017005.D

Acq: 08 Sep 2023 11:03




Tgt Ion:202 Resp: 2440815


Ion Ratio Lower Upper

202 100

101 15.0 11.3 16.9

100 12.2 9.4 14.0





#83

Butylbenzylphthalate

Concen: 21.715 ng/ul

RT: 20.627 min Scan# 2982

Delta R.T. -0.006 min

Lab File: BP017005.D

Acq: 08 Sep 2023 11:03

Instrument :

BNA_P

ClientSampleId :

SSTD020722





Tgt Ion:149 Resp: 984356

Ion Ratio Lower Upper

149 100

91 79.0 59.1 88.7

206 18.6 15.7 23.5



#84

3,3'-Dichlorobenzidine

Concen: 19.498 ng/ul

RT: 21.427 min Scan# 3118

Delta R.T. 0.000 min

Lab File: BP017005.D

Acq: 08 Sep 2023 11:03




Tgt Ion:252 Resp: 600664


Ion Ratio Lower Upper

252 100

254 64.6 51.8 77.6

126 12.8 10.4 15.6









#85
Benzo(a)anthracene
Concen: 21.246 ng/ul
RT: 21.486 min Scan# 3128
Delta R.T. -0.006 min
Lab File: BP017005.D
Acq: 08 Sep 2023 11:03




Instrument :
BNA_P
ClientSampleId :
SSTD020722


Tgt Ion:228 Resp: 2077101
Ion Ratio Lower Upper
228 100
229 19.8 15.8 23.8
226 27.9 22.2 33.2



#86
Bis(2-ethylhexyl)phthalate
Concen: 21.730 ng/ul
RT: 21.363 min Scan# 3107
Delta R.T. -0.006 min
Lab File: BP017005.D
Acq: 08 Sep 2023 11:03

Tgt Ion:149 Resp: 1360816
Ion Ratio Lower Upper
149 100
167 25.8 21.4 32.2
279 3.1 3.0 4.4





#87

Chrysene

Concen: 21.211 ng/ul

RT: 21.545 min Scan# 3138

Delta R.T. 0.000 min

Lab File: BP017005.D

Acq: 08 Sep 2023 11:03

Instrument :

BNA_P

ClientSampleId :

SSTD020722





Tgt Ion:228 Resp: 1939725

Ion Ratio Lower Upper

228 100

226 30.8 24.6 36.8

229 19.4 15.9 23.9



#89

Di-n-octyl phthalate

Concen: 20.208 ng/ul




RT: 22.333 min Scan# 3272


Delta R.T. -0.006 min

Lab File: BP017005.D

Acq: 08 Sep 2023 11:03

Tgt Ion:149 Resp: 2101318









#90
Benzo(b)fluoranthene
Concen: 20.424 ng/ul
RT: 23.251 min Scan# 3428
Delta R.T. -0.006 min
Lab File: BP017005.D
Acq: 08 Sep 2023 11:03




Instrument :
BNA_P
ClientSampleId :
SSTD020722


Tgt Ion:252 Resp: 1770103
Ion Ratio Lower Upper
252 100
253 21.9 17.6 26.4
125 12.3 8.8 13.2



#91
Benzo(k)fluoranthene
Concen: 20.970 ng/ul
RT: 23.304 min Scan# 3437
Delta R.T. -0.006 min
Lab File: BP017005.D
Acq: 08 Sep 2023 11:03

Tgt Ion:252 Resp: 1759918
Ion Ratio Lower Upper
252 100
253 22.0 17.7 26.5
125 12.1 8.9 13.3





#93

Benzo(a)pyrene

Concen: 20.803 ng/ul

RT: 23.927 min Scan# 3543

Delta R.T. -0.006 min

Lab File: BP017005.D

Acq: 08 Sep 2023 11:03

Instrument :

BNA_P

ClientSampleId :

SSTD020722





Tgt Ion:252 Resp: 1503768

Ion Ratio Lower Upper

252 100

253 21.6 17.8 26.6

125 12.5 9.9 14.9



#94

Indeno(1,2,3-cd)pyrene

Concen: 21.963 ng/ul

RT: 26.727 min Scan# 4019

Delta R.T. -0.017 min

Lab File: BP017005.D

Acq: 08 Sep 2023 11:03




Tgt Ion:276 Resp: 1762338


Ion Ratio Lower Upper

276 100

138 26.9 21.4 32.0

277 23.9 20.0 30.0





#95

Dibenzo(a,h)anthracene

Concen: 21.956 ng/ul

RT: 26.757 min Scan# 4024

Delta R.T. -0.011 min

Lab File: BP017005.D

Acq: 08 Sep 2023 11:03

Instrument :

BNA_P

ClientSampleId :

SSTD020722





Tgt Ion:278 Resp: 1454090

Ion Ratio Lower Upper

278 100

139 19.8 14.9 22.3

279 24.1 19.6 29.4



#96

Benzo(g,h,i)perylene

Concen: 22.350 ng/ul

RT: 27.574 min Scan# 4163

Delta R.T. -0.017 min

Lab File: BP017005.D

Acq: 08 Sep 2023 11:03

Tgt Ion:276 Resp: 1413268

Ion Ratio Lower Upper

276 100

138 26.2 19.5 29.3

277 23.8 18.7 28.1

