

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090823\
 Data File : BP017077.D
 Acq On : 10 Sep 2023 07:17
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 71 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020726

Quant Time: Sep 11 00:32:01 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP090523.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Sep 09 01:41:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.963	152	542720	20.000	ng/u1	-0.01
20) Naphthalene-d8	10.793	136	2273715	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.628	164	1264042	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.392	188	2596572	20.000	ng/u1	0.00
79) Chrysene-d12	21.492	240	1970929	20.000	ng/u1	0.00
88) Perylene-d12	24.021	264	1759258	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.317	96	124866	8.703	ng/uL	0.00
4) Pyridine-d5	3.758	84	934536	22.795	ng/u1	0.00
7) Phenol-d5	7.116	99	1135259	22.896	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.305	67	706990	22.796	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.487	132	819280	22.831	ng/u1	0.00
15) 4-Methylphenol-d8	8.681	113	833746	22.072	ng/u1	0.00
21) Nitrobenzene-d5	9.163	128	398162	22.572	ng/u1	-0.01
24) 2-Nitrophenol-d4	9.881	143	424570	22.204	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.405	165	780575	22.111	ng/u1	0.00
31) 4-Chloroaniline-d4	10.957	131	1074981	21.106	ng/u1	0.00
46) Dimethylphthalate-d6	14.040	166	2010360	21.366	ng/u1	0.00
49) Acenaphthylene-d8	14.328	160	2405564	22.153	ng/u1	0.00
54) 4-Nitrophenol-d4	14.845	143	354673	19.297	ng/u1	0.00
60) Fluorene-d10	15.622	176	1725483	21.419	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.751	200	267717	17.133	ng/u1	0.00
73) Anthracene-d10	17.498	188	2610131	21.889	ng/u1	0.00
81) Pyrene-d10	19.733	212	2737885	22.806	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.857	264	1936746	21.700	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.358	88	135521	8.748	ng/uL	98
5) Pyridine	3.775	79	982560	22.513	ng/u1	97
6) Benzaldehyde	7.122	77	395957	17.620	ng/u1	96
8) Phenol	7.146	94	1182236	22.949	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.405	93	911374	22.185	ng/u1	100
12) 2-Chlorophenol	7.522	128	872595	22.825	ng/u1	98
13) 2-Methylphenol	8.410	108	844638	22.353	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.481	45	669712	11.579	ng/u1	99
16) Acetophenone	8.816	105	1330039	21.820	ng/u1	96
17) N-Nitroso-di-n-propyla...	8.793	70	702189	21.564	ng/u1	97
18) 4-Methylphenol	8.746	108	911893	22.128	ng/u1	99
19) Hexachloroethane	9.034	117	360848	22.785	ng/u1	96
22) Nitrobenzene	9.210	77	1093147	23.209	ng/u1	98
23) Isophorone	9.728	82	1868017	21.110	ng/u1	98
25) 2-Nitrophenol	9.910	139	472110	22.425	ng/u1	97
26) 2,4-Dimethylphenol	9.952	107	959322	22.317	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.210	93	1145549	20.889	ng/u1	100
29) 2,4-Dichlorophenol	10.434	162	785694	21.839	ng/u1	98
30) Naphthalene	10.846	128	2719080	21.916	ng/u1	99
32) 4-Chloroaniline	10.981	127	1132196	21.360	ng/u1	98
33) Hexachlorobutadiene	11.075	225	445662	20.725	ng/u1	99
34) Caprolactam	11.810	113	182776	15.340	ng/u1	93
35) 4-Chloro-3-methylphenol	12.081	107	826853	20.972	ng/u1	99
36) 2-Methylnaphthalene	12.451	142	1716271	20.933	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090823\
 Data File : BP017077.D
 Acq On : 10 Sep 2023 07:17
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 71 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020726


Quant Time: Sep 11 00:32:01 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP090523.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Sep 09 01:41:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.669	142	1723160	20.875	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.798	216	818739	21.984	ng/ul	99
40) Hexachlorocyclopentadiene	12.746	237	481611	19.958	ng/ul	99
41) 2,4,6-Trichlorophenol	13.057	196	559630	22.078	ng/ul	99
42) 2,4,5-Trichlorophenol	13.128	196	605442	21.952	ng/ul	98
43) 1,1'-Biphenyl	13.457	154	2247630	22.463	ng/ul	99
44) 2-Chloronaphthalene	13.504	162	1768307	22.509	ng/ul	100
45) 2-Nitroaniline	13.746	65	573996	24.534	ng/ul	97
47) Dimethylphthalate	14.087	163	2116807	21.441	ng/ul	99
48) 2,6-Dinitrotoluene	14.234	165	427618	22.303	ng/ul	96
50) Acenaphthylene	14.357	152	2713559	22.318	ng/ul	100
51) 3-Nitroaniline	14.569	138	360158	18.823	ng/ul	94
52) Acenaphthene	14.693	153	1874334	22.094	ng/ul	100
53) 2,4-Dinitrophenol	14.769	184	192044	16.155	ng/ul	93
55) 4-Nitrophenol	14.863	109	302045	20.533	ng/ul	94
56) Dibenzofuran	15.028	168	2555186	21.906	ng/ul	98
57) 2,4-Dinitrotoluene	15.016	165	622109	22.604	ng/ul	86
58) 2,3,4,6-Tetrachlorophenol	15.245	232	450584	20.263	ng/ul	99
59) Diethylphthalate	15.445	149	2146994	21.762	ng/ul	100
61) Fluorene	15.681	166	2121213	22.057	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.669	204	983610	21.164	ng/ul	96
63) 4-Nitroaniline	15.740	138	271255	16.124	ng/ul	93
66) 4,6-Dinitro-2-methylph...	15.763	198	280182	17.114	ng/ul#	86
67) N-Nitrosodiphenylamine	15.892	169	1723731	22.016	ng/ul	98
68) 4-Bromophenyl-phenylether	16.575	248	534506	20.691	ng/ul	98
69) Hexachlorobenzene	16.651	284	574003	20.270	ng/ul	93
70) Atrazine	16.851	200	574778	21.442	ng/ul	100
71) Pentachlorophenol	17.022	266	328238	17.775	ng/ul	99
72) Phenanthrene	17.439	178	3157752	21.886	ng/ul	100
74) Anthracene	17.528	178	3211940	22.145	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.410	216	822498	22.238	ng/ul	98
76) Pentachlorobenzene	14.916	250	748821	21.018	ng/ul	99
77) Carbazole	17.816	167	2877047	22.073	ng/ul	99
78) Di-n-butylphthalate	18.322	149	3624928	22.701	ng/ul	99
80) Fluoranthene	19.404	202	3449543	23.013	ng/ul	100
82) Pyrene	19.763	202	3583456	23.025	ng/ul	99
83) Butylbenzylphthalate	20.616	149	1616750	25.119	ng/ul	94
84) 3,3'-Dichlorobenzidine	21.416	252	801062	18.314	ng/ul	100
85) Benzo(a)anthracene	21.474	228	3101476	22.343	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.351	149	2261727	25.436	ng/ul	99
87) Chrysene	21.533	228	2830072	21.796	ng/ul	99
89) Di-n-octyl phthalate	22.316	149	3614678	24.510	ng/ul	100
90) Benzo(b)fluoranthene	23.233	252	2613810	21.265	ng/ul	98
91) Benzo(k)fluoranthene	23.286	252	2595514	21.806	ng/ul	99
93) Benzo(a)pyrene	23.910	252	2229238	21.744	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.704	276	2591722	22.774	ng/ul	97
95) Dibenzo(a,h)anthracene	26.727	278	2148232	22.871	ng/ul	98
96) Benzo(g,h,i)perylene	27.545	276	2017710	22.499	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Sep 11 00:32:01 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP090523.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Sep 09 01:41:21 2023
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 8.748 ng/uL
 RT: 3.358 min Scan# 40
 Delta R.T. -0.006 min
 Lab File: BP017077.D
 Acq: 10 Sep 2023 07:17




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020726


Tgt Ion: 88 Resp: 135521
 Ion Ratio Lower Upper
 88 100
 43 30.7 25.0 37.6
 58 73.9 60.6 90.8



#5
 Pyridine
 Concen: 22.513 ng/uL
 RT: 3.775 min Scan# 117
 Delta R.T. -0.006 min
 Lab File: BP017077.D
 Acq: 10 Sep 2023 07:17

Tgt Ion: 79 Resp: 982560
 Ion Ratio Lower Upper
 79 100
 52 78.0 60.1 90.1
 51 37.2 30.5 45.7









#6
Benzaldehyde
Concen: 17.620 ng/ul
RT: 7.122 min Scan# 61
Delta R.T. -0.012 min
Lab File: BP017077.D
Acq: 10 Sep 2023 07:17




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion: 77 Resp: 395957
Ion Ratio Lower Upper
77 100
105 90.4 75.4 113.0
106 86.3 71.4 107.2



#8
Phenol
Concen: 22.949 ng/ul
RT: 7.146 min Scan# 690
Delta R.T. -0.006 min
Lab File: BP017077.D
Acq: 10 Sep 2023 07:17

Tgt Ion: 94 Resp: 1182236
Ion Ratio Lower Upper
94 100
65 29.9 23.2 34.8
66 38.0 29.4 44.2









#10
Bis(2-Chloroethyl)ether
Concen: 22.185 ng/ul
RT: 7.405 min Scan# 71
Delta R.T. -0.006 min
Lab File: BP017077.D
Acq: 10 Sep 2023 07:17




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion: 93 Resp: 911374
Ion Ratio Lower Upper
93 100
63 79.3 63.4 95.2
95 32.8 26.1 39.1



#12
2-Chlorophenol
Concen: 22.825 ng/ul
RT: 7.522 min Scan# 754
Delta R.T. -0.006 min
Lab File: BP017077.D
Acq: 10 Sep 2023 07:17

Tgt Ion:128 Resp: 872595
Ion Ratio Lower Upper
128 100
64 51.3 40.2 60.2
130 32.9 25.2 37.8





#13

2-Methylphenol

Concen: 22.353 ng/ul

RT: 8.410 min Scan# 906

Delta R.T. -0.006 min

Lab File: BP017077.D

Acq: 10 Sep 2023 07:17

Instrument :

BNA_P

ClientSampleId :





SSTD020726

Tgt Ion:108 Resp: 844638

Ion Ratio Lower Upper

108 100

107 88.6 70.6 106.0



#14

2,2'-oxybis(1-Chloropropane)

Concen: 11.579 ng/ul

RT: 8.481 min Scan# 917

Delta R.T. -0.024 min

Lab File: BP017077.D

Acq: 10 Sep 2023 07:17




Tgt Ion: 45 Resp: 669712


Ion Ratio Lower Upper

45 100

77 14.0 11.3 16.9

79 10.1 9.0 13.4









#16
Acetophenone
Concen: 21.820 ng/ul
RT: 8.816 min Scan# 91
Delta R.T. -0.012 min
Lab File: BP017077.D
Acq: 10 Sep 2023 07:17




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:105 Resp: 1330039
Ion Ratio Lower Upper
105 100
77 83.7 63.6 95.4
51 31.1 23.4 35.0



#17
N-Nitroso-di-n-propylamine
Concen: 21.564 ng/ul
RT: 8.793 min Scan# 970
Delta R.T. -0.006 min
Lab File: BP017077.D
Acq: 10 Sep 2023 07:17

Tgt Ion: 70 Resp: 702189
Ion Ratio Lower Upper
70 100
42 62.6 47.4 71.2
101 9.3 7.3 10.9
130 18.8 15.6 23.4









#18
4-Methylphenol
Concen: 22.128 ng/ul
RT: 8.746 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP017077.D
Acq: 10 Sep 2023 07:17




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:108 Resp: 911893
Ion Ratio Lower Upper
108 100
107 112.5 90.6 135.8



#19
Hexachloroethane
Concen: 22.785 ng/ul
RT: 9.034 min Scan# 1011
Delta R.T. -0.012 min
Lab File: BP017077.D
Acq: 10 Sep 2023 07:17

Tgt Ion:117 Resp: 360848
Ion Ratio Lower Upper
117 100
201 89.9 75.8 113.6
199 56.0 46.8 70.2









#22
Nitrobenzene
Concen: 23.209 ng/ul
RT: 9.210 min Scan# 1109
Delta R.T. -0.006 min
Lab File: BP017077.D
Acq: 10 Sep 2023 07:17




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion: 77 Resp: 1093147
Ion Ratio Lower Upper
77 100
123 41.2 34.4 51.6
65 13.3 10.6 15.8



#23
Isophorone
Concen: 21.110 ng/ul
RT: 9.728 min Scan# 1129
Delta R.T. -0.006 min
Lab File: BP017077.D
Acq: 10 Sep 2023 07:17

Tgt Ion: 82 Resp: 1868017
Ion Ratio Lower Upper
82 100
95 6.8 5.3 7.9
138 16.0 13.4 20.2









#25
2-Nitrophenol
Concen: 22.425 ng/ul
RT: 9.910 min Scan# 1160
Delta R.T. -0.012 min
Lab File: BP017077.D
Acq: 10 Sep 2023 07:17




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:139 Resp: 472110
Ion Ratio Lower Upper
139 100
65 57.7 44.0 66.0
109 30.8 24.1 36.1



#26
2,4-Dimethylphenol
Concen: 22.317 ng/ul
RT: 9.952 min Scan# 1167
Delta R.T. -0.006 min
Lab File: BP017077.D
Acq: 10 Sep 2023 07:17

Tgt Ion:107 Resp: 959322
Ion Ratio Lower Upper
107 100
121 49.2 41.0 61.6
122 83.1 69.0 103.4





#27

Bis(2-Chloroethoxy)methane

Concen: 20.889 ng/ul

RT: 10.210 min Scan# 11

Delta R.T. -0.006 min

Lab File: BP017077.D

Acq: 10 Sep 2023 07:17

Instrument :

BNA_P

ClientSampleId :

SSTD020726





Tgt Ion: 93 Resp: 1145549

Ion Ratio Lower Upper

93 100

95 32.8 26.0 39.0

123 10.8 8.6 13.0



#29

2,4-Dichlorophenol

Concen: 21.839 ng/ul

RT: 10.434 min Scan# 1249

Delta R.T. -0.006 min

Lab File: BP017077.D

Acq: 10 Sep 2023 07:17




Tgt Ion:162 Resp: 785694


Ion Ratio Lower Upper

162 100

164 65.7 51.8 77.6

98 37.9 28.2 42.4









#30
Naphthalene
Concen: 21.916 ng/ul
RT: 10.846 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP017077.D
Acq: 10 Sep 2023 07:17




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:128 Resp: 2719080
Ion Ratio Lower Upper
128 100
129 10.9 8.8 13.2
127 13.3 10.3 15.5



#32
4-Chloroaniline
Concen: 21.360 ng/ul
RT: 10.981 min Scan# 1342
Delta R.T. -0.006 min
Lab File: BP017077.D
Acq: 10 Sep 2023 07:17

Tgt Ion:127 Resp: 1132196
Ion Ratio Lower Upper
127 100
129 31.5 26.0 39.0









#33
Hexachlorobutadiene
Concen: 20.725 ng/ul
RT: 11.075 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP017077.D
Acq: 10 Sep 2023 07:17




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:225 Resp: 445662
Ion Ratio Lower Upper
225 100
223 61.3 50.1 75.1
227 62.8 50.9 76.3



#34
Caprolactam
Concen: 15.340 ng/ul
RT: 11.810 min Scan# 1483
Delta R.T. -0.006 min
Lab File: BP017077.D
Acq: 10 Sep 2023 07:17

Tgt Ion:113 Resp: 182776
Ion Ratio Lower Upper
113 100
55 186.6 142.3 213.5
56 138.0 102.2 153.2









#35
4-Chloro-3-methylphenol
Concen: 20.972 ng/ul
RT: 12.081 min Scan# 1529
Delta R.T. -0.006 min
Lab File: BP017077.D
Acq: 10 Sep 2023 07:17




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:107 Resp: 826853
Ion Ratio Lower Upper
107 100
144 25.3 20.3 30.5
142 78.6 63.8 95.8



#36
2-Methylnaphthalene
Concen: 20.933 ng/ul
RT: 12.451 min Scan# 1592
Delta R.T. -0.006 min
Lab File: BP017077.D
Acq: 10 Sep 2023 07:17

Tgt Ion:142 Resp: 1716271
Ion Ratio Lower Upper
142 100
141 90.2 71.1 106.7









#37
1-Methylnaphthalene
Concen: 20.875 ng/ul
RT: 12.669 min Scan# 1629
Delta R.T. -0.006 min
Lab File: BP017077.D
Acq: 10 Sep 2023 07:17




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:142 Resp: 1723160
Ion Ratio Lower Upper
142 100
141 92.8 74.5 111.7
116 4.1 3.1 4.7



#39
1,2,4,5-Tetrachlorobenzene
Concen: 21.984 ng/ul
RT: 12.798 min Scan# 1651
Delta R.T. -0.006 min
Lab File: BP017077.D
Acq: 10 Sep 2023 07:17

Tgt Ion:216 Resp: 818739
Ion Ratio Lower Upper
216 100
214 78.6 62.8 94.2
179 22.6 17.5 26.3
108 20.5 14.6 21.8





#40

Hexachlorocyclopentadiene

Concen: 19.958 ng/ul

RT: 12.746 min Scan# 1

Delta R.T. -0.012 min

Lab File: BP017077.D

Acq: 10 Sep 2023 07:17

Instrument :

BNA_P

ClientSampleId :

SSTD020726





Tgt Ion:237 Resp: 481611

Ion Ratio Lower Upper

237 100

235 62.5 50.8 76.2

272 11.8 10.0 15.0



#41

2,4,6-Trichlorophenol

Concen: 22.078 ng/ul

RT: 13.057 min Scan# 1695

Delta R.T. -0.000 min

Lab File: BP017077.D

Acq: 10 Sep 2023 07:17




Tgt Ion:196 Resp: 559630


Ion Ratio Lower Upper

196 100

198 93.6 75.8 113.6

200 30.3 24.6 36.8





#42

2,4,5-Trichlorophenol

Concen: 21.952 ng/ul

RT: 13.128 min Scan# 1707

Delta R.T. -0.000 min

Lab File: BP017077.D

Acq: 10 Sep 2023 07:17

Instrument :

BNA_P

ClientSampleId :

SSTD020726





Tgt Ion:196 Resp: 605442

Ion Ratio Lower Upper

196 100

198 95.9 74.9 112.3

200 30.8 24.5 36.7



#43

1,1'-Biphenyl

Concen: 22.463 ng/ul

RT: 13.457 min Scan# 1763

Delta R.T. -0.006 min

Lab File: BP017077.D

Acq: 10 Sep 2023 07:17




Tgt Ion:154 Resp: 2247630


Ion Ratio Lower Upper

154 100

153 40.4 32.7 49.1

76 14.8 11.5 17.3





#44

2-Chloronaphthalene

Concen: 22.509 ng/ul

RT: 13.504 min Scan# 1

Delta R.T. -0.012 min

Lab File: BP017077.D

Acq: 10 Sep 2023 07:17

Instrument :

BNA_P

ClientSampleId :

SSTD020726





Tgt Ion:162 Resp: 1768307

Ion Ratio Lower Upper

162 100

164 32.6 26.2 39.4

127 38.8 30.8 46.2



#45

2-Nitroaniline

Concen: 24.534 ng/ul

RT: 13.746 min Scan# 1812

Delta R.T. -0.000 min

Lab File: BP017077.D

Acq: 10 Sep 2023 07:17




Tgt Ion: 65 Resp: 573996


Ion Ratio Lower Upper

65 100

92 66.3 55.4 83.0

138 98.2 76.8 115.2









#47
Dimethylphthalate
Concen: 21.441 ng/ul
RT: 14.087 min Scan# 1870
Delta R.T. -0.006 min
Lab File: BP017077.D
Acq: 10 Sep 2023 07:17




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:163 Resp: 2116807
Ion Ratio Lower Upper
163 100
194 3.8 3.1 4.7
164 10.3 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 22.303 ng/ul
RT: 14.234 min Scan# 1895
Delta R.T. -0.006 min
Lab File: BP017077.D
Acq: 10 Sep 2023 07:17

Tgt Ion:165 Resp: 427618
Ion Ratio Lower Upper
165 100
89 65.8 50.0 75.0
121 21.4 17.8 26.6





#50

Acenaphthylene

Concen: 22.318 ng/ul

RT: 14.357 min Scan# 1916

Delta R.T. -0.000 min

Lab File: BP017077.D

Acq: 10 Sep 2023 07:17

Instrument :

BNA_P

ClientSampleId :

SSTD020726





Tgt Ion:152 Resp: 2713559

Ion Ratio Lower Upper

152 100

151 20.1 15.9 23.9

153 13.1 10.5 15.7



#51

3-Nitroaniline

Concen: 18.823 ng/ul

RT: 14.569 min Scan# 1952

Delta R.T. -0.006 min

Lab File: BP017077.D

Acq: 10 Sep 2023 07:17




Tgt Ion:138 Resp: 360158


Ion Ratio Lower Upper

138 100

108 10.7 8.4 12.6

92 128.5 97.2 145.8









#52
Acenaphthene
Concen: 22.094 ng/ul
RT: 14.693 min Scan# 1973
Delta R.T. -0.006 min
Lab File: BP017077.D
Acq: 10 Sep 2023 07:17




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:153 Resp: 1874334
Ion Ratio Lower Upper
153 100
152 48.7 38.6 57.8
154 87.9 70.2 105.4



#53
2,4-Dinitrophenol
Concen: 16.155 ng/ul
RT: 14.769 min Scan# 1986
Delta R.T. -0.006 min
Lab File: BP017077.D
Acq: 10 Sep 2023 07:17

Tgt Ion:184 Resp: 192044
Ion Ratio Lower Upper
184 100
63 84.9 61.6 92.4
154 67.2 50.8 76.2









#55
4-Nitrophenol
Concen: 20.533 ng/ul
RT: 14.863 min Scan# 2002
Delta R.T. -0.000 min
Lab File: BP017077.D
Acq: 10 Sep 2023 07:17




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:109 Resp: 302045
Ion Ratio Lower Upper
109 100
139 124.8 105.0 157.4
65 142.5 108.4 162.6



#56
Dibenzofuran
Concen: 21.906 ng/ul
RT: 15.028 min Scan# 2030
Delta R.T. -0.006 min
Lab File: BP017077.D
Acq: 10 Sep 2023 07:17

Tgt Ion:168 Resp: 2555186
Ion Ratio Lower Upper
168 100
139 38.8 30.1 45.1









#57
2,4-Dinitrotoluene
Concen: 22.604 ng/ul
RT: 15.016 min Scan# 2029
Delta R.T. -0.000 min
Lab File: BP017077.D
Acq: 10 Sep 2023 07:17




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:165 Resp: 622109
Ion Ratio Lower Upper
165 100
63 62.9 42.1 63.1
182 2.6 2.3 3.5



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.263 ng/ul
RT: 15.245 min Scan# 2067
Delta R.T. -0.000 min
Lab File: BP017077.D
Acq: 10 Sep 2023 07:17

Tgt Ion:232 Resp: 450584
Ion Ratio Lower Upper
232 100
131 50.6 39.2 58.8
130 2.7 2.0 3.0
166 34.0 27.2 40.8









#59
Diethylphthalate
Concen: 21.762 ng/ul
RT: 15.445 min Scan# 2101
Delta R.T. -0.000 min
Lab File: BP017077.D
Acq: 10 Sep 2023 07:17




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:149 Resp: 2146994
Ion Ratio Lower Upper
149 100
177 21.4 17.3 25.9
150 12.3 10.1 15.1



#61
Fluorene
Concen: 22.057 ng/ul
RT: 15.681 min Scan# 2141
Delta R.T. -0.000 min
Lab File: BP017077.D
Acq: 10 Sep 2023 07:17

Tgt Ion:166 Resp: 2121213
Ion Ratio Lower Upper
166 100
165 99.2 78.3 117.5
167 13.5 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 21.164 ng/ul
RT: 15.669 min Scan# 2139
Delta R.T. -0.006 min
Lab File: BP017077.D
Acq: 10 Sep 2023 07:17




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:204 Resp: 983610
Ion Ratio Lower Upper
204 100
206 32.5 26.4 39.6
141 62.5 46.2 69.4



#63
4-Nitroaniline
Concen: 16.124 ng/ul
RT: 15.740 min Scan# 2151
Delta R.T. -0.006 min
Lab File: BP017077.D
Acq: 10 Sep 2023 07:17

Tgt Ion:138 Resp: 271255
Ion Ratio Lower Upper
138 100
92 54.7 41.9 62.9
108 97.0 70.3 105.5









#66
4,6-Dinitro-2-methylphenol
Concen: 17.114 ng/ul
RT: 15.763 min Scan# 2155
Delta R.T. -0.006 min
Lab File: BP017077.D
Acq: 10 Sep 2023 07:17




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:198 Resp: 280182
Ion Ratio Lower Upper
198 100
182 5.0 3.6 5.4
77 34.7 21.4 32.0#



#67
N-Nitrosodiphenylamine
Concen: 22.016 ng/ul
RT: 15.892 min Scan# 2177
Delta R.T. -0.006 min
Lab File: BP017077.D
Acq: 10 Sep 2023 07:17

Tgt Ion:169 Resp: 1723731
Ion Ratio Lower Upper
169 100
168 69.4 54.2 81.4
167 38.1 29.4 44.2









#68
4-Bromophenyl-phenylether
Concen: 20.691 ng/ul
RT: 16.575 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP017077.D
Acq: 10 Sep 2023 07:17




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:248 Resp: 534506
Ion Ratio Lower Upper
248 100
250 98.6 76.8 115.2
141 82.0 66.3 99.5



#69
Hexachlorobenzene
Concen: 20.270 ng/ul
RT: 16.651 min Scan# 2306
Delta R.T. -0.006 min
Lab File: BP017077.D
Acq: 10 Sep 2023 07:17

Tgt Ion:284 Resp: 574003
Ion Ratio Lower Upper
284 100
142 47.0 31.9 47.9
249 32.7 27.0 40.6





#70

Atrazine

Concen: 21.442 ng/ul

RT: 16.851 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP017077.D

Acq: 10 Sep 2023 07:17

Instrument :

BNA_P

ClientSampleId :

SSTD020726





Tgt Ion:200 Resp: 574778

Ion Ratio Lower Upper

200 100

173 27.6 21.8 32.8

215 47.6 38.3 57.5



#71

Pentachlorophenol

Concen: 17.775 ng/ul

RT: 17.022 min Scan# 2369

Delta R.T. -0.000 min

Lab File: BP017077.D

Acq: 10 Sep 2023 07:17




Tgt Ion:266 Resp: 328238


Ion Ratio Lower Upper

266 100

264 63.3 50.4 75.6

268 62.0 50.4 75.6





#72

Phenanthrene

Concen: 21.886 ng/ul

RT: 17.439 min Scan# 2440

Delta R.T. -0.000 min

Lab File: BP017077.D

Acq: 10 Sep 2023 07:17

Instrument :

BNA_P

ClientSampleId :

SSTD020726





Tgt Ion:178 Resp: 3157752

Ion Ratio Lower Upper

178 100

179 15.4 12.5 18.7

176 19.8 15.8 23.8



#74

Anthracene

Concen: 22.145 ng/ul

RT: 17.528 min Scan# 2455

Delta R.T. -0.006 min

Lab File: BP017077.D

Acq: 10 Sep 2023 07:17




Tgt Ion:178 Resp: 3211940


Ion Ratio Lower Upper

178 100

179 15.8 12.3 18.5

176 19.6 15.0 22.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 22.238 ng/uL

RT: 13.410 min Scan# 11

Delta R.T. -0.006 min

Lab File: BP017077.D

Acq: 10 Sep 2023 07:17

Instrument :

BNA_P

ClientSampleId :

SSTD020726





Tgt Ion:216 Resp: 822498

Ion Ratio Lower Upper

216 100

214 79.0 64.5 96.7

218 47.3 38.6 57.8



#76

Pentachlorobenzene

Concen: 21.018 ng/uL

RT: 14.916 min Scan# 2011

Delta R.T. -0.006 min

Lab File: BP017077.D

Acq: 10 Sep 2023 07:17




Tgt Ion:250 Resp: 748821


Ion Ratio Lower Upper

250 100

248 63.2 51.2 76.8

252 62.4 50.9 76.3





#77

Carbazole

Concen: 22.073 ng/ul

RT: 17.816 min Scan# 2504

Delta R.T. -0.000 min

Lab File: BP017077.D

Acq: 10 Sep 2023 07:17

Instrument :

BNA_P

ClientSampleId :

SSTD020726





Tgt Ion:167 Resp: 2877047

Ion Ratio Lower Upper

167 100

166 21.9 17.3 25.9

139 13.3 10.3 15.5



#78

Di-n-butylphthalate

Concen: 22.701 ng/ul

RT: 18.322 min Scan# 2590

Delta R.T. -0.006 min

Lab File: BP017077.D

Acq: 10 Sep 2023 07:17




Tgt Ion:149 Resp: 3624928


Ion Ratio Lower Upper

149 100

150 9.4 7.4 11.2

104 5.8 4.2 6.4









#80
Fluoranthene
Concen: 23.013 ng/ul
RT: 19.404 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP017077.D
Acq: 10 Sep 2023 07:17




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:202 Resp: 3449543
Ion Ratio Lower Upper
202 100
101 12.3 10.0 15.0
100 9.6 7.5 11.3



#82
Pyrene
Concen: 23.025 ng/ul
RT: 19.763 min Scan# 2835
Delta R.T. -0.000 min
Lab File: BP017077.D
Acq: 10 Sep 2023 07:17

Tgt Ion:202 Resp: 3583456
Ion Ratio Lower Upper
202 100
101 14.7 11.3 16.9
100 12.0 9.4 14.0





#83

Butylbenzylphthalate

Concen: 25.119 ng/ul

RT: 20.616 min Scan# 2982

Delta R.T. -0.006 min

Lab File: BP017077.D

Acq: 10 Sep 2023 07:17

Instrument :

BNA_P

ClientSampleId :

SSTD020726





Tgt Ion:149 Resp: 1616750

Ion Ratio Lower Upper

149 100

91 79.3 59.1 88.7

206 17.7 15.7 23.5



#84

3,3'-Dichlorobenzidine

Concen: 18.314 ng/ul

RT: 21.416 min Scan# 3116

Delta R.T. -0.000 min

Lab File: BP017077.D

Acq: 10 Sep 2023 07:17




Tgt Ion:252 Resp: 801062


Ion Ratio Lower Upper

252 100

254 64.6 51.8 77.6

126 12.5 10.4 15.6









#85
 Benzo(a)anthracene
 Concen: 22.343 ng/ul
 RT: 21.474 min Scan# 3126
 Delta R.T. -0.000 min
 Lab File: BP017077.D
 Acq: 10 Sep 2023 07:17

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020726

Tgt Ion:228 Resp: 3101476




Ion	Ratio	Lower	Upper
228	100		
229	19.6	15.8	23.8
226	27.9	22.2	33.2




#86
 Bis(2-ethylhexyl)phthalate
 Concen: 25.436 ng/ul
 RT: 21.351 min Scan# 3105
 Delta R.T. -0.006 min
 Lab File: BP017077.D
 Acq: 10 Sep 2023 07:17

Tgt Ion:149 Resp: 2261727

Ion	Ratio	Lower	Upper
149	100		
167	26.2	21.4	32.2
279	3.2	3.0	4.4





#87

Chrysene

Concen: 21.796 ng/ul

RT: 21.533 min Scan# 31

Delta R.T. -0.000 min

Lab File: BP017077.D

Acq: 10 Sep 2023 07:17

Instrument :

BNA_P

ClientSampleId :

SSTD020726





Tgt Ion:228 Resp: 2830072

Ion Ratio Lower Upper

228 100

226 31.7 24.6 36.8

229 20.0 15.9 23.9



#89

Di-n-octyl phthalate

Concen: 24.510 ng/ul




RT: 22.316 min Scan# 3269


Delta R.T. -0.006 min

Lab File: BP017077.D

Acq: 10 Sep 2023 07:17

Tgt Ion:149 Resp: 3614678









#90
Benzo(b)fluoranthene
Concen: 21.265 ng/ul
RT: 23.233 min Scan# 3425
Delta R.T. -0.006 min
Lab File: BP017077.D
Acq: 10 Sep 2023 07:17




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:252 Resp: 2613810
Ion Ratio Lower Upper
252 100
253 22.6 17.6 26.4
125 12.1 8.8 13.2



#91
Benzo(k)fluoranthene
Concen: 21.806 ng/ul
RT: 23.286 min Scan# 3434
Delta R.T. -0.006 min
Lab File: BP017077.D
Acq: 10 Sep 2023 07:17

Tgt Ion:252 Resp: 2595514
Ion Ratio Lower Upper
252 100
253 22.5 17.7 26.5
125 12.2 8.9 13.3









#93
Benzo(a)pyrene
Concen: 21.744 ng/ul
RT: 23.910 min Scan# 31
Delta R.T. -0.000 min
Lab File: BP017077.D
Acq: 10 Sep 2023 07:17




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:252 Resp: 2229238
Ion Ratio Lower Upper
252 100
253 21.9 17.8 26.6
125 12.6 9.9 14.9



#94
Indeno(1,2,3-cd)pyrene
Concen: 22.774 ng/ul
RT: 26.704 min Scan# 4015
Delta R.T. -0.006 min
Lab File: BP017077.D
Acq: 10 Sep 2023 07:17

Tgt Ion:276 Resp: 2591722
Ion Ratio Lower Upper
276 100
138 29.1 21.4 32.0
277 24.5 20.0 30.0





#95

Dibenzo(a,h)anthracene

Concen: 22.871 ng/ul

RT: 26.727 min Scan# 40

Delta R.T. -0.012 min

Lab File: BP017077.D

Acq: 10 Sep 2023 07:17

Instrument :

BNA_P

ClientSampleId :

SSTD020726





Tgt Ion:278 Resp: 2148232

Ion Ratio Lower Upper

278 100

139 20.3 14.9 22.3

279 24.2 19.6 29.4



#96

Benzo(g,h,i)perylene

Concen: 22.499 ng/ul

RT: 27.545 min Scan# 4158

Delta R.T. -0.006 min

Lab File: BP017077.D

Acq: 10 Sep 2023 07:17

Tgt Ion:276 Resp: 2017710

Ion Ratio Lower Upper

276 100

138 26.7 19.5 29.3

277 23.9 18.7 28.1

