

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090823\
 Data File : BP017083.D
 Acq On : 10 Sep 2023 11:16
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 97 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020727

Quant Time: Sep 11 00:34:57 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP090523.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Sep 09 01:41:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.963	152	439026	20.000	ng/ul	-0.01
20) Naphthalene-d8	10.793	136	1877340	20.000	ng/ul	-0.01
38) Acenaphthene-d10	14.622	164	1090049	20.000	ng/ul	-0.01
64) Phenanthrene-d10	17.392	188	2101110	20.000	ng/ul	0.00
79) Chrysene-d12	21.486	240	1483516	20.000	ng/ul	0.00
88) Perylene-d12	24.015	264	1334037	20.000	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.322	96	96883	8.347	ng/uL	0.00
4) Pyridine-d5	3.758	84	691905	20.863	ng/ul	0.00
7) Phenol-d5	7.116	99	826205	20.599	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.311	67	539112	21.488	ng/ul	0.00
11) 2-Chlorophenol-d4	7.487	132	604758	20.833	ng/ul	0.00
15) 4-Methylphenol-d8	8.681	113	618317	20.235	ng/ul	0.00
21) Nitrobenzene-d5	9.169	128	295258	20.272	ng/ul	0.00
24) 2-Nitrophenol-d4	9.881	143	314225	19.903	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.404	165	584829	20.064	ng/ul	0.00
31) 4-Chloroaniline-d4	10.951	131	819192	19.480	ng/ul	-0.01
46) Dimethylphthalate-d6	14.040	166	1598531	19.701	ng/ul	0.00
49) Acenaphthylene-d8	14.322	160	1898112	20.270	ng/ul	-0.01
54) 4-Nitrophenol-d4	14.840	143	255306	16.108	ng/ul	0.00
60) Fluorene-d10	15.622	176	1357362	19.538	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.751	200	194390	15.374	ng/ul	0.00
73) Anthracene-d10	17.492	188	1942038	20.126	ng/ul	0.00
81) Pyrene-d10	19.727	212	1903276	21.063	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.851	264	1347513	19.911	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.358	88	103255	8.240	ng/uL	99
5) Pyridine	3.781	79	739118	20.935	ng/ul	100
6) Benzaldehyde	7.122	77	524189	28.835	ng/ul	98
8) Phenol	7.146	94	868087	20.831	ng/ul	97
10) Bis(2-Chloroethyl)ether	7.405	93	698253	21.012	ng/ul	99
12) 2-Chlorophenol	7.522	128	649086	20.988	ng/ul	97
13) 2-Methylphenol	8.410	108	624224	20.421	ng/ul	99
14) 2,2'-oxybis(1-Chloropr...	8.481	45	507836	10.854	ng/ul	98
16) Acetophenone	8.816	105	1033754	20.965	ng/ul	97
17) N-Nitroso-di-n-propyla...	8.793	70	558630	21.208	ng/ul	97
18) 4-Methylphenol	8.746	108	684699	20.539	ng/ul	100
19) Hexachloroethane	9.034	117	269590	21.043	ng/ul	96
22) Nitrobenzene	9.210	77	830054	21.344	ng/ul	98
23) Isophorone	9.722	82	1488510	20.373	ng/ul	98
25) 2-Nitrophenol	9.910	139	352998	20.307	ng/ul	94
26) 2,4-Dimethylphenol	9.952	107	724901	20.424	ng/ul	98
27) Bis(2-Chloroethoxy)met...	10.210	93	923283	20.391	ng/ul	100
29) 2,4-Dichlorophenol	10.434	162	597733	20.122	ng/ul	99
30) Naphthalene	10.846	128	2077059	20.276	ng/ul	99
32) 4-Chloroaniline	10.981	127	856035	19.560	ng/ul	99
33) Hexachlorobutadiene	11.075	225	333923	18.807	ng/ul	99
34) Caprolactam	11.804	113	177506	18.043	ng/ul	91
35) 4-Chloro-3-methylphenol	12.081	107	650638	19.987	ng/ul	97
36) 2-Methylnaphthalene	12.446	142	1354320	20.006	ng/ul	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090823\
 Data File : BP017083.D
 Acq On : 10 Sep 2023 11:16
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 97 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020727

Quant Time: Sep 11 00:34:57 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP090523.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Sep 09 01:41:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.669	142	1355622	19.890	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.798	216	635058	19.774	ng/ul	99
40) Hexachlorocyclopentadiene	12.746	237	387272	18.611	ng/ul	100
41) 2,4,6-Trichlorophenol	13.051	196	434228	19.865	ng/ul	100
42) 2,4,5-Trichlorophenol	13.122	196	463786	19.500	ng/ul	98
43) 1,1'-Biphenyl	13.457	154	1790086	20.746	ng/ul	99
44) 2-Chloronaphthalene	13.504	162	1387358	20.479	ng/ul	99
45) 2-Nitroaniline	13.740	65	434394	21.531	ng/ul	95
47) Dimethylphthalate	14.087	163	1680807	19.742	ng/ul	99
48) 2,6-Dinitrotoluene	14.234	165	330415	19.984	ng/ul	98
50) Acenaphthylene	14.351	152	2154005	20.543	ng/ul	100
51) 3-Nitroaniline	14.569	138	329548	19.972	ng/ul	97
52) Acenaphthene	14.687	153	1495079	20.437	ng/ul	100
53) 2,4-Dinitrophenol	14.769	184	143419	13.990	ng/ul	97
55) 4-Nitrophenol	14.857	109	222917	17.573	ng/ul	94
56) Dibenzofuran	15.028	168	2011913	20.001	ng/ul	99
57) 2,4-Dinitrotoluene	15.010	165	467668	19.705	ng/ul#	85
58) 2,3,4,6-Tetrachlorophenol	15.240	232	342503	17.861	ng/ul	94
59) Diethylphthalate	15.439	149	1685302	19.809	ng/ul	99
61) Fluorene	15.675	166	1658416	19.997	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.669	204	772502	19.275	ng/ul	96
63) 4-Nitroaniline	15.739	138	285861	19.704	ng/ul	90
66) 4,6-Dinitro-2-methylph...	15.763	198	204889	15.466	ng/ul#	90
67) N-Nitrosodiphenylamine	15.892	169	1343274	21.202	ng/ul	100
68) 4-Bromophenyl-phenylether	16.569	248	408986	19.565	ng/ul	97
69) Hexachlorobenzene	16.651	284	437091	19.075	ng/ul	94
70) Atrazine	16.845	200	420657	19.393	ng/ul	98
71) Pentachlorophenol	17.016	266	235591	15.766	ng/ul	99
72) Phenanthrene	17.433	178	2380411	20.389	ng/ul	100
74) Anthracene	17.528	178	2414265	20.570	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.410	216	641435	21.432	ng/ul	99
76) Pentachlorobenzene	14.916	250	591792	20.527	ng/ul	99
77) Carbazole	17.810	167	2086243	19.781	ng/ul	98
78) Di-n-butylphthalate	18.322	149	2675303	20.705	ng/ul	99
80) Fluoranthene	19.398	202	2435114	21.583	ng/ul	98
82) Pyrene	19.757	202	2499903	21.340	ng/ul	98
83) Butylbenzylphthalate	20.616	149	1069022	22.066	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.410	252	604785	18.369	ng/ul	99
85) Benzo(a)anthracene	21.474	228	2130490	20.391	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.351	149	1507288	22.521	ng/ul	98
87) Chrysene	21.527	228	1977156	20.230	ng/ul	100
89) Di-n-octyl phthalate	22.316	149	2334533	20.876	ng/ul	100
90) Benzo(b)fluoranthene	23.227	252	1856367	19.917	ng/ul	98
91) Benzo(k)fluoranthene	23.280	252	1786403	19.792	ng/ul	99
93) Benzo(a)pyrene	23.904	252	1562688	20.101	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.692	276	1871260	21.684	ng/ul	98
95) Dibenzo(a,h)anthracene	26.721	278	1550800	21.773	ng/ul	98
96) Benzo(g,h,i)perylene	27.533	276	1484815	21.834	ng/ul	98





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 8.240 ng/uL
RT: 3.358 min Scan# 40
Delta R.T. -0.006 min
Lab File: BP017083.D
Acq: 10 Sep 2023 11:16




Instrument :
BNA_P
ClientSampleId :
SSTD020727


Tgt Ion: 88 Resp: 103255
Ion Ratio Lower Upper
88 100
43 33.3 25.0 37.6
58 76.1 60.6 90.8



#5
Pyridine
Concen: 20.935 ng/uL
RT: 3.781 min Scan# 118
Delta R.T. -0.000 min
Lab File: BP017083.D
Acq: 10 Sep 2023 11:16

Tgt Ion: 79 Resp: 739118
Ion Ratio Lower Upper
79 100
52 74.9 60.1 90.1
51 37.8 30.5 45.7









#6
Benzaldehyde
Concen: 28.835 ng/ul
RT: 7.122 min Scan# 61
Delta R.T. -0.012 min
Lab File: BP017083.D
Acq: 10 Sep 2023 11:16




Instrument :
BNA_P
ClientSampleId :
SSTD020727


Tgt Ion: 77 Resp: 524189
Ion Ratio Lower Upper
77 100
105 91.1 75.4 113.0
106 88.8 71.4 107.2



#8
Phenol
Concen: 20.831 ng/ul
RT: 7.146 min Scan# 690
Delta R.T. -0.006 min
Lab File: BP017083.D
Acq: 10 Sep 2023 11:16

Tgt Ion: 94 Resp: 868087
Ion Ratio Lower Upper
94 100
65 30.6 23.2 34.8
66 38.5 29.4 44.2









#10
Bis(2-Chloroethyl)ether
Concen: 21.012 ng/ul
RT: 7.405 min Scan# 71
Delta R.T. -0.006 min
Lab File: BP017083.D
Acq: 10 Sep 2023 11:16




Instrument :
BNA_P
ClientSampleId :
SSTD020727


Tgt Ion: 93 Resp: 698253
Ion Ratio Lower Upper
93 100
63 80.0 63.4 95.2
95 32.6 26.1 39.1



#12
2-Chlorophenol
Concen: 20.988 ng/ul
RT: 7.522 min Scan# 754
Delta R.T. -0.006 min
Lab File: BP017083.D
Acq: 10 Sep 2023 11:16

Tgt Ion:128 Resp: 649086
Ion Ratio Lower Upper
128 100
64 52.6 40.2 60.2
130 33.0 25.2 37.8





#13

2-Methylphenol

Concen: 20.421 ng/ul

RT: 8.410 min Scan# 906

Delta R.T. -0.006 min

Lab File: BP017083.D

Acq: 10 Sep 2023 11:16

Instrument :

BNA_P

ClientSampleId :





SSTD020727

Tgt Ion:108 Resp: 624224

Ion Ratio Lower Upper

108 100

107 89.4 70.6 106.0



#14

2,2'-oxybis(1-Chloropropane)

Concen: 10.854 ng/ul

RT: 8.481 min Scan# 917

Delta R.T. -0.024 min

Lab File: BP017083.D

Acq: 10 Sep 2023 11:16




Tgt Ion: 45 Resp: 507836


Ion Ratio Lower Upper

45 100

77 14.5 11.3 16.9

79 10.3 9.0 13.4









#16
Acetophenone
Concen: 20.965 ng/ul
RT: 8.816 min Scan# 91
Delta R.T. -0.012 min
Lab File: BP017083.D
Acq: 10 Sep 2023 11:16




Instrument :
BNA_P
ClientSampleId :
SSTD020727


Tgt Ion:105 Resp: 1033754
Ion Ratio Lower Upper
105 100
77 82.3 63.6 95.4
51 30.6 23.4 35.0



#17
N-Nitroso-di-n-propylamine
Concen: 21.208 ng/ul
RT: 8.793 min Scan# 970
Delta R.T. -0.006 min
Lab File: BP017083.D
Acq: 10 Sep 2023 11:16

Tgt Ion: 70 Resp: 558630
Ion Ratio Lower Upper
70 100
42 62.0 47.4 71.2
101 9.6 7.3 10.9
130 19.0 15.6 23.4









#18
4-Methylphenol
Concen: 20.539 ng/ul
RT: 8.746 min Scan# 90
Delta R.T. -0.006 min
Lab File: BP017083.D
Acq: 10 Sep 2023 11:16




Instrument :
BNA_P
ClientSampleId :
SSTD020727


Tgt Ion:108 Resp: 684699
Ion Ratio Lower Upper
108 100
107 113.2 90.6 135.8



#19
Hexachloroethane
Concen: 21.043 ng/ul
RT: 9.034 min Scan# 1011
Delta R.T. -0.012 min
Lab File: BP017083.D
Acq: 10 Sep 2023 11:16

Tgt Ion:117 Resp: 269590
Ion Ratio Lower Upper
117 100
201 88.8 75.8 113.6
199 57.3 46.8 70.2









#22
Nitrobenzene
Concen: 21.344 ng/ul
RT: 9.210 min Scan# 1108
Delta R.T. -0.006 min
Lab File: BP017083.D
Acq: 10 Sep 2023 11:16




Instrument :
BNA_P
ClientSampleId :
SSTD020727


Tgt Ion: 77 Resp: 830054
Ion Ratio Lower Upper
77 100
123 41.0 34.4 51.6
65 13.2 10.6 15.8



#23
Isophorone
Concen: 20.373 ng/ul
RT: 9.722 min Scan# 1128
Delta R.T. -0.012 min
Lab File: BP017083.D
Acq: 10 Sep 2023 11:16

Tgt Ion: 82 Resp: 1488510
Ion Ratio Lower Upper
82 100
95 6.8 5.3 7.9
138 15.5 13.4 20.2









#25
2-Nitrophenol
Concen: 20.307 ng/ul
RT: 9.910 min Scan# 1160
Delta R.T. -0.012 min
Lab File: BP017083.D
Acq: 10 Sep 2023 11:16




Instrument :
BNA_P
ClientSampleId :
SSTD020727


Tgt Ion:139 Resp: 352998
Ion Ratio Lower Upper
139 100
65 61.4 44.0 66.0
109 31.1 24.1 36.1



#26
2,4-Dimethylphenol
Concen: 20.424 ng/ul
RT: 9.952 min Scan# 1167
Delta R.T. -0.006 min
Lab File: BP017083.D
Acq: 10 Sep 2023 11:16

Tgt Ion:107 Resp: 724901
Ion Ratio Lower Upper
107 100
121 49.8 41.0 61.6
122 84.1 69.0 103.4





#27

Bis(2-Chloroethoxy)methane

Concen: 20.391 ng/ul

RT: 10.210 min Scan# 1212

Delta R.T. -0.006 min

Lab File: BP017083.D

Acq: 10 Sep 2023 11:16

Instrument :

BNA_P

ClientSampleId :

SSTD020727





Tgt Ion: 93 Resp: 923283

Ion Ratio Lower Upper

93 100

95 32.6 26.0 39.0

123 10.8 8.6 13.0



#29

2,4-Dichlorophenol

Concen: 20.122 ng/ul

RT: 10.434 min Scan# 1249

Delta R.T. -0.006 min

Lab File: BP017083.D

Acq: 10 Sep 2023 11:16




Tgt Ion:162 Resp: 597733


Ion Ratio Lower Upper

162 100

164 65.2 51.8 77.6

98 36.6 28.2 42.4









#30
Naphthalene
Concen: 20.276 ng/ul
RT: 10.846 min Scan# 1319
Delta R.T. -0.006 min
Lab File: BP017083.D
Acq: 10 Sep 2023 11:16




Instrument :
BNA_P
ClientSampleId :
SSTD020727


Tgt Ion:128 Resp: 2077059
Ion Ratio Lower Upper
128 100
129 10.7 8.8 13.2
127 13.1 10.3 15.5



#32
4-Chloroaniline
Concen: 19.560 ng/ul
RT: 10.981 min Scan# 1342
Delta R.T. -0.006 min
Lab File: BP017083.D
Acq: 10 Sep 2023 11:16

Tgt Ion:127 Resp: 856035
Ion Ratio Lower Upper
127 100
129 32.2 26.0 39.0









#33
Hexachlorobutadiene
Concen: 18.807 ng/ul
RT: 11.075 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP017083.D
Acq: 10 Sep 2023 11:16




Instrument :
BNA_P
ClientSampleId :
SSTD020727


Tgt Ion:225 Resp: 333923
Ion Ratio Lower Upper
225 100
223 63.8 50.1 75.1
227 64.5 50.9 76.3



#34
Caprolactam
Concen: 18.043 ng/ul
RT: 11.804 min Scan# 1482
Delta R.T. -0.012 min
Lab File: BP017083.D
Acq: 10 Sep 2023 11:16

Tgt Ion:113 Resp: 177506
Ion Ratio Lower Upper
113 100
55 189.3 142.3 213.5
56 139.6 102.2 153.2









#35
4-Chloro-3-methylphenol
Concen: 19.987 ng/ul
RT: 12.081 min Scan# 1529
Delta R.T. -0.006 min
Lab File: BP017083.D
Acq: 10 Sep 2023 11:16




Instrument :
BNA_P
ClientSampleId :
SSTD020727


Tgt Ion:107 Resp: 650638
Ion Ratio Lower Upper
107 100
144 25.3 20.3 30.5
142 76.4 63.8 95.8



#36
2-Methylnaphthalene
Concen: 20.006 ng/ul
RT: 12.446 min Scan# 1591
Delta R.T. -0.012 min
Lab File: BP017083.D
Acq: 10 Sep 2023 11:16

Tgt Ion:142 Resp: 1354320
Ion Ratio Lower Upper
142 100
141 89.6 71.1 106.7





#37

1-Methylnaphthalene

Concen: 19.890 ng/ul

RT: 12.669 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP017083.D

Acq: 10 Sep 2023 11:16

Instrument :

BNA_P

ClientSampleId :

SSTD020727





Tgt Ion:142 Resp: 1355622

Ion Ratio Lower Upper

142 100

141 92.9 74.5 111.7

116 4.0 3.1 4.7



#39

1,2,4,5-Tetrachlorobenzene

Concen: 19.774 ng/ul

RT: 12.798 min Scan# 1651

Delta R.T. -0.006 min

Lab File: BP017083.D

Acq: 10 Sep 2023 11:16

Tgt Ion:216 Resp: 635058




Ion Ratio Lower Upper


216 100

214 79.0 62.8 94.2

179 22.8 17.5 26.3

108 20.1 14.6 21.8









#40
Hexachlorocyclopentadiene
Concen: 18.611 ng/ul
RT: 12.746 min Scan# 1642
Delta R.T. -0.012 min
Lab File: BP017083.D
Acq: 10 Sep 2023 11:16




Instrument :
BNA_P
ClientSampleId :
SSTD020727


Tgt Ion:237 Resp: 387272
Ion Ratio Lower Upper
237 100
235 63.7 50.8 76.2
272 12.0 10.0 15.0



#41
2,4,6-Trichlorophenol
Concen: 19.865 ng/ul
RT: 13.051 min Scan# 1694
Delta R.T. -0.006 min
Lab File: BP017083.D
Acq: 10 Sep 2023 11:16

Tgt Ion:196 Resp: 434228
Ion Ratio Lower Upper
196 100
198 95.2 75.8 113.6
200 30.8 24.6 36.8





#42

2,4,5-Trichlorophenol

Concen: 19.500 ng/ul

RT: 13.122 min Scan# 1706

Delta R.T. -0.006 min

Lab File: BP017083.D

Acq: 10 Sep 2023 11:16

Instrument :

BNA_P

ClientSampleId :

SSTD020727





Tgt Ion:196 Resp: 463786

Ion Ratio Lower Upper

196 100

198 96.3 74.9 112.3

200 31.4 24.5 36.7



#43

1,1'-Biphenyl

Concen: 20.746 ng/ul

RT: 13.457 min Scan# 1763

Delta R.T. -0.006 min

Lab File: BP017083.D

Acq: 10 Sep 2023 11:16




Tgt Ion:154 Resp: 1790086


Ion Ratio Lower Upper

154 100

153 39.9 32.7 49.1

76 14.9 11.5 17.3





#44

2-Chloronaphthalene

Concen: 20.479 ng/ul

RT: 13.504 min Scan# 11

Delta R.T. -0.012 min

Lab File: BP017083.D

Acq: 10 Sep 2023 11:16

Instrument :

BNA_P

ClientSampleId :

SSTD020727





Tgt Ion:162 Resp: 1387358

Ion Ratio Lower Upper

162 100

164 32.4 26.2 39.4

127 39.6 30.8 46.2



#45

2-Nitroaniline

Concen: 21.531 ng/ul

RT: 13.740 min Scan# 1811

Delta R.T. -0.006 min

Lab File: BP017083.D

Acq: 10 Sep 2023 11:16




Tgt Ion: 65 Resp: 434394


Ion Ratio Lower Upper

65 100

92 64.0 55.4 83.0

138 91.5 76.8 115.2









#47
Dimethylphthalate
Concen: 19.742 ng/ul
RT: 14.087 min Scan# 1870
Delta R.T. -0.006 min
Lab File: BP017083.D
Acq: 10 Sep 2023 11:16




Instrument :
BNA_P
ClientSampleId :
SSTD020727


Tgt Ion:163 Resp: 1680807
Ion Ratio Lower Upper
163 100
194 3.7 3.1 4.7
164 10.3 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 19.984 ng/ul
RT: 14.234 min Scan# 1895
Delta R.T. -0.006 min
Lab File: BP017083.D
Acq: 10 Sep 2023 11:16

Tgt Ion:165 Resp: 330415
Ion Ratio Lower Upper
165 100
89 64.1 50.0 75.0
121 21.6 17.8 26.6





#50

Acenaphthylene

Concen: 20.543 ng/ul

RT: 14.351 min Scan# 1915

Delta R.T. -0.006 min

Lab File: BP017083.D

Acq: 10 Sep 2023 11:16

Instrument :

BNA_P

ClientSampleId :

SSTD020727





Tgt Ion:152 Resp: 2154005

Ion Ratio Lower Upper

152 100

151 20.1 15.9 23.9

153 13.3 10.5 15.7



#51

3-Nitroaniline

Concen: 19.972 ng/ul

RT: 14.569 min Scan# 1952

Delta R.T. -0.006 min

Lab File: BP017083.D

Acq: 10 Sep 2023 11:16




Tgt Ion:138 Resp: 329548


Ion Ratio Lower Upper

138 100

108 10.6 8.4 12.6

92 124.7 97.2 145.8









#52
Acenaphthene
Concen: 20.437 ng/ul
RT: 14.687 min Scan# 1972
Delta R.T. -0.012 min
Lab File: BP017083.D
Acq: 10 Sep 2023 11:16




Instrument :
BNA_P
ClientSampleId :
SSTD020727


Tgt Ion:153 Resp: 1495079
Ion Ratio Lower Upper
153 100
152 47.9 38.6 57.8
154 87.4 70.2 105.4



#53
2,4-Dinitrophenol
Concen: 13.990 ng/ul
RT: 14.769 min Scan# 1986
Delta R.T. -0.006 min
Lab File: BP017083.D
Acq: 10 Sep 2023 11:16

Tgt Ion:184 Resp: 143419
Ion Ratio Lower Upper
184 100
63 81.3 61.6 92.4
154 63.8 50.8 76.2









#55
4-Nitrophenol
Concen: 17.573 ng/ul
RT: 14.857 min Scan# 2002
Delta R.T. -0.006 min
Lab File: BP017083.D
Acq: 10 Sep 2023 11:16




Instrument :
BNA_P
ClientSampleId :
SSTD020727


Tgt Ion:109 Resp: 222917
Ion Ratio Lower Upper
109 100
139 127.6 105.0 157.4
65 145.1 108.4 162.6



#56
Dibenzofuran
Concen: 20.001 ng/ul
RT: 15.028 min Scan# 2030
Delta R.T. -0.006 min
Lab File: BP017083.D
Acq: 10 Sep 2023 11:16

Tgt Ion:168 Resp: 2011913
Ion Ratio Lower Upper
168 100
139 38.3 30.1 45.1





#57

2,4-Dinitrotoluene

Concen: 19.705 ng/ul

RT: 15.010 min Scan# 2029

Delta R.T. -0.006 min

Lab File: BP017083.D

Acq: 10 Sep 2023 11:16

Instrument :

BNA_P

ClientSampleId :

SSTD020727





Tgt Ion:165 Resp: 467668

Ion Ratio Lower Upper

165 100

63 63.7 42.1 63.1#

182 2.5 2.3 3.5



#58

2,3,4,6-Tetrachlorophenol

Concen: 17.861 ng/ul

RT: 15.240 min Scan# 2066

Delta R.T. -0.006 min

Lab File: BP017083.D

Acq: 10 Sep 2023 11:16

Tgt Ion:232 Resp: 342503




Ion Ratio Lower Upper


232 100

131 55.1 39.2 58.8

130 3.0 2.0 3.0

166 35.5 27.2 40.8









#59
Diethylphthalate
Concen: 19.809 ng/ul
RT: 15.439 min Scan# 2100
Delta R.T. -0.006 min
Lab File: BP017083.D
Acq: 10 Sep 2023 11:16




Instrument :
BNA_P
ClientSampleId :
SSTD020727


Tgt Ion:149 Resp: 1685302
Ion Ratio Lower Upper
149 100
177 21.1 17.3 25.9
150 12.4 10.1 15.1



#61
Fluorene
Concen: 19.997 ng/ul
RT: 15.675 min Scan# 2140
Delta R.T. -0.006 min
Lab File: BP017083.D
Acq: 10 Sep 2023 11:16

Tgt Ion:166 Resp: 1658416
Ion Ratio Lower Upper
166 100
165 98.0 78.3 117.5
167 13.4 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 19.275 ng/ul
RT: 15.669 min Scan# 2139
Delta R.T. -0.006 min
Lab File: BP017083.D
Acq: 10 Sep 2023 11:16




Instrument :
BNA_P
ClientSampleId :
SSTD020727


Tgt Ion:204 Resp: 772502
Ion Ratio Lower Upper
204 100
206 32.5 26.4 39.6
141 62.6 46.2 69.4



#63
4-Nitroaniline
Concen: 19.704 ng/ul
RT: 15.739 min Scan# 2151
Delta R.T. -0.006 min
Lab File: BP017083.D
Acq: 10 Sep 2023 11:16

Tgt Ion:138 Resp: 285861
Ion Ratio Lower Upper
138 100
92 56.1 41.9 62.9
108 100.3 70.3 105.5









#66
4,6-Dinitro-2-methylphenol
Concen: 15.466 ng/ul
RT: 15.763 min Scan# 2155
Delta R.T. -0.006 min
Lab File: BP017083.D
Acq: 10 Sep 2023 11:16




Instrument :
BNA_P
ClientSampleId :
SSTD020727


Tgt Ion:198 Resp: 204889
Ion Ratio Lower Upper
198 100
182 5.1 3.6 5.4
77 32.7 21.4 32.0#



#67
N-Nitrosodiphenylamine
Concen: 21.202 ng/ul
RT: 15.892 min Scan# 2177
Delta R.T. -0.006 min
Lab File: BP017083.D
Acq: 10 Sep 2023 11:16

Tgt Ion:169 Resp: 1343274
Ion Ratio Lower Upper
169 100
168 68.2 54.2 81.4
167 37.1 29.4 44.2









#68
4-Bromophenyl-phenylether
Concen: 19.565 ng/ul
RT: 16.569 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP017083.D
Acq: 10 Sep 2023 11:16




Instrument :
BNA_P
ClientSampleId :
SSTD020727


Tgt Ion:248 Resp: 408986
Ion Ratio Lower Upper
248 100
250 97.9 76.8 115.2
141 86.9 66.3 99.5



#69
Hexachlorobenzene
Concen: 19.075 ng/ul
RT: 16.651 min Scan# 2306
Delta R.T. -0.006 min
Lab File: BP017083.D
Acq: 10 Sep 2023 11:16

Tgt Ion:284 Resp: 437091
Ion Ratio Lower Upper
284 100
142 46.3 31.9 47.9
249 33.6 27.0 40.6





#70

Atrazine

Concen: 19.393 ng/ul

RT: 16.845 min Scan# 21

Delta R.T. -0.012 min

Lab File: BP017083.D

Acq: 10 Sep 2023 11:16

Instrument :

BNA_P

ClientSampleId :

SSTD020727





Tgt Ion:200 Resp: 420657

Ion Ratio Lower Upper

200 100

173 28.4 21.8 32.8

215 46.2 38.3 57.5



#71

Pentachlorophenol

Concen: 15.766 ng/ul

RT: 17.016 min Scan# 2368

Delta R.T. -0.006 min

Lab File: BP017083.D

Acq: 10 Sep 2023 11:16




Tgt Ion:266 Resp: 235591


Ion Ratio Lower Upper

266 100

264 63.7 50.4 75.6

268 63.2 50.4 75.6





#72

Phenanthrene

Concen: 20.389 ng/ul

RT: 17.433 min Scan# 2441

Delta R.T. -0.006 min

Lab File: BP017083.D

Acq: 10 Sep 2023 11:16

Instrument :

BNA_P

ClientSampleId :

SSTD020727





Tgt Ion:178 Resp: 2380411

Ion Ratio Lower Upper

178 100

179 15.6 12.5 18.7

176 19.6 15.8 23.8



#74

Anthracene

Concen: 20.570 ng/ul

RT: 17.528 min Scan# 2455

Delta R.T. -0.006 min

Lab File: BP017083.D

Acq: 10 Sep 2023 11:16




Tgt Ion:178 Resp: 2414265


Ion Ratio Lower Upper

178 100

179 15.6 12.3 18.5

176 19.4 15.0 22.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 21.432 ng/uL

RT: 13.410 min Scan# 11

Delta R.T. -0.006 min

Lab File: BP017083.D

Acq: 10 Sep 2023 11:16

Instrument :

BNA_P

ClientSampleId :

SSTD020727





Tgt Ion:216 Resp: 641435

Ion Ratio Lower Upper

216 100

214 79.1 64.5 96.7

218 47.8 38.6 57.8



#76

Pentachlorobenzene

Concen: 20.527 ng/uL

RT: 14.916 min Scan# 2011

Delta R.T. -0.006 min

Lab File: BP017083.D

Acq: 10 Sep 2023 11:16




Tgt Ion:250 Resp: 591792


Ion Ratio Lower Upper

250 100

248 63.2 51.2 76.8

252 64.8 50.9 76.3





#77

Carbazole

Concen: 19.781 ng/ul

RT: 17.810 min Scan# 2503

Delta R.T. -0.006 min

Lab File: BP017083.D

Acq: 10 Sep 2023 11:16

Instrument :

BNA_P

ClientSampleId :

SSTD020727





Tgt Ion:167 Resp: 2086243

Ion Ratio Lower Upper

167 100

166 22.2 17.3 25.9

139 13.8 10.3 15.5



#78

Di-n-butylphthalate

Concen: 20.705 ng/ul

RT: 18.322 min Scan# 2590

Delta R.T. -0.006 min

Lab File: BP017083.D

Acq: 10 Sep 2023 11:16




Tgt Ion:149 Resp: 2675303


Ion Ratio Lower Upper

149 100

150 9.1 7.4 11.2

104 5.5 4.2 6.4





#80

Fluoranthene

Concen: 21.583 ng/ul

RT: 19.398 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP017083.D

Acq: 10 Sep 2023 11:16

Instrument :

BNA_P

ClientSampleId :

SSTD020727





Tgt Ion:202 Resp: 2435114

Ion Ratio Lower Upper

202 100

101 12.9 10.0 15.0

100 10.3 7.5 11.3



#82

Pyrene

Concen: 21.340 ng/ul

RT: 19.757 min Scan# 2834

Delta R.T. -0.006 min

Lab File: BP017083.D

Acq: 10 Sep 2023 11:16




Tgt Ion:202 Resp: 2499903


Ion Ratio Lower Upper

202 100

101 15.1 11.3 16.9

100 12.4 9.4 14.0





#83

Butylbenzylphthalate

Concen: 22.066 ng/ul

RT: 20.616 min Scan# 2982

Delta R.T. -0.006 min

Lab File: BP017083.D

Acq: 10 Sep 2023 11:16

Instrument :

BNA_P

ClientSampleId :

SSTD020727





Tgt Ion:149 Resp: 1069022

Ion Ratio Lower Upper

149 100

91 78.0 59.1 88.7

206 18.7 15.7 23.5



#84

3,3'-Dichlorobenzidine

Concen: 18.369 ng/ul

RT: 21.410 min Scan# 3115

Delta R.T. -0.006 min

Lab File: BP017083.D

Acq: 10 Sep 2023 11:16




Tgt Ion:252 Resp: 604785


Ion Ratio Lower Upper

252 100

254 64.2 51.8 77.6

126 14.5 10.4 15.6





#85

Benzo(a)anthracene

Concen: 20.391 ng/ul

RT: 21.474 min Scan# 3128

Delta R.T. -0.000 min

Lab File: BP017083.D

Acq: 10 Sep 2023 11:16

Instrument :

BNA_P

ClientSampleId :

SSTD020727





Tgt Ion:228 Resp: 2130490

Ion Ratio Lower Upper

228 100

229 19.6 15.8 23.8

226 28.2 22.2 33.2



#86

Bis(2-ethylhexyl)phthalate

Concen: 22.521 ng/ul

RT: 21.351 min Scan# 3105

Delta R.T. -0.006 min

Lab File: BP017083.D

Acq: 10 Sep 2023 11:16




Tgt Ion:149 Resp: 1507288


Ion Ratio Lower Upper

149 100

167 25.8 21.4 32.2

279 3.3 3.0 4.4





#87

Chrysene

Concen: 20.230 ng/ul

RT: 21.527 min Scan# 31

Delta R.T. -0.006 min

Lab File: BP017083.D

Acq: 10 Sep 2023 11:16

Instrument :

BNA_P

ClientSampleId :

SSTD020727





Tgt Ion:228 Resp: 1977156

Ion Ratio Lower Upper

228 100

226 30.8 24.6 36.8

229 19.6 15.9 23.9



#89

Di-n-octyl phthalate

Concen: 20.876 ng/ul




RT: 22.316 min Scan# 3269


Delta R.T. -0.006 min

Lab File: BP017083.D

Acq: 10 Sep 2023 11:16

Tgt Ion:149 Resp: 2334533









#90
Benzo(b)fluoranthene
Concen: 19.917 ng/ul
RT: 23.227 min Scan# 3428
Delta R.T. -0.012 min
Lab File: BP017083.D
Acq: 10 Sep 2023 11:16




Instrument :
BNA_P
ClientSampleId :
SSTD020727


Tgt Ion:252 Resp: 1856367
Ion Ratio Lower Upper
252 100
253 22.4 17.6 26.4
125 12.7 8.8 13.2



#91
Benzo(k)fluoranthene
Concen: 19.792 ng/ul
RT: 23.280 min Scan# 3433
Delta R.T. -0.012 min
Lab File: BP017083.D
Acq: 10 Sep 2023 11:16

Tgt Ion:252 Resp: 1786403
Ion Ratio Lower Upper
252 100
253 21.9 17.7 26.5
125 12.4 8.9 13.3





#93

Benzo(a)pyrene

Concen: 20.101 ng/ul

RT: 23.904 min Scan# 3539

Delta R.T. -0.006 min

Lab File: BP017083.D

Acq: 10 Sep 2023 11:16

Instrument :

BNA_P

ClientSampleId :

SSTD020727





Tgt Ion:252 Resp: 1562688

Ion Ratio Lower Upper

252 100

253 21.4 17.8 26.6

125 13.3 9.9 14.9



#94

Indeno(1,2,3-cd)pyrene

Concen: 21.684 ng/ul

RT: 26.692 min Scan# 4013

Delta R.T. -0.018 min

Lab File: BP017083.D

Acq: 10 Sep 2023 11:16




Tgt Ion:276 Resp: 1871260


Ion Ratio Lower Upper

276 100

138 27.4 21.4 32.0

277 23.8 20.0 30.0





#95

Dibenzo(a,h)anthracene

Concen: 21.773 ng/ul

RT: 26.721 min Scan# 4018

Delta R.T. -0.018 min

Lab File: BP017083.D

Acq: 10 Sep 2023 11:16

Instrument :

BNA_P

ClientSampleId :

SSTD020727





Tgt Ion:278 Resp: 1550800

Ion Ratio Lower Upper

278 100

139 20.2 14.9 22.3

279 24.3 19.6 29.4



#96

Benzo(g,h,i)perylene

Concen: 21.834 ng/ul

RT: 27.533 min Scan# 4156

Delta R.T. -0.018 min

Lab File: BP017083.D

Acq: 10 Sep 2023 11:16

Tgt Ion:276 Resp: 1484815

Ion Ratio Lower Upper

276 100

138 26.4 19.5 29.3

277 23.8 18.7 28.1

