

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090823\
 Data File : BP017044.D
 Acq On : 09 Sep 2023 10:34
 Operator : MA/JU
 Sample : 04166-11
 Misc :
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 BBJL4

Quant Time: Sep 09 23:04:31 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP090523.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Sep 09 01:41:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.969	152	316226	20.000	ng/u1	0.00
20) Naphthalene-d8	10.798	136	1387218	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.628	164	828566	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.398	188	1768366	20.000	ng/u1	0.00
79) Chrysene-d12	21.492	240	1438801	20.000	ng/u1	0.00
88) Perylene-d12	24.027	264	1483327	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.322	96	46124	5.517	ng/uL	0.00
4) Pyridine-d5	3.764	84	345203	14.451	ng/u1	0.00
7) Phenol-d5	7.116	99	290557	10.057	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.310	67	839096	46.433	ng/u1	0.00
11) 2-Chlorophenol-d4	7.493	132	745163	35.638	ng/u1	0.00
15) 4-Methylphenol-d8	8.681	113	446117	20.269	ng/u1	0.00
21) Nitrobenzene-d5	9.169	128	476365	44.263	ng/u1	0.00
24) 2-Nitrophenol-d4	9.887	143	495756	42.495	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.410	165	820081	38.076	ng/u1	0.00
31) 4-Chloroaniline-d4	10.957	131	650211	20.924	ng/u1	0.00
46) Dimethylphthalate-d6	14.045	166	3008158	48.773	ng/u1	0.00
49) Acenaphthylene-d8	14.328	160	3264792	45.868	ng/u1	0.00
54) 4-Nitrophenol-d4	14.845	143	94158	7.815	ng/u1	0.00
60) Fluorene-d10	15.628	176	2550244	48.294	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.751	200	403898	37.953	ng/u1	0.00
73) Anthracene-d10	17.498	188	3724685	45.864	ng/u1	0.00
81) Pyrene-d10	19.733	212	4361762	49.771	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.862	264	3706711	49.257	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090823\
Data File : BP017044.D
Acq On : 09 Sep 2023 10:34
Operator : MA/JU
Sample : 04166-11
Misc :
ALS Vial : 55 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
BBJL4

Quant Time: Sep 09 23:04:31 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP090523.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Sep 09 01:41:21 2023
Response via : Initial Calibration

