

Data Path : Z:\SVOASRV\HPCHEM1\BNA_P\DATA\BP090919\
 Data File : BP000156.D
 Acq On : 10 Sep 2019 00:54
 Operator : HP/JU
 Sample : PB122762BL
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Sep 10 03:40:45 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SOM-EPA-BP090519MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Sep 10 02:43:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.89	152	438591	20.00	ng/ul	0.00
18) Naphthalene-d8	8.18	136	1724927	20.00	ng/ul	0.00
36) Acenaphthene-d10	9.95	164	978066	20.00	ng/ul	0.00
62) Phenanthrene-d10	11.45	188	1732535	20.00	ng/ul	0.00
78) Chrysene-d12	14.13	240	1400309	20.00	ng/ul	-0.01
86) Perylene-d12	15.67	264	1393645	20.00	ng/ul	-0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	2.61	96	56350	4.56	ng/uL	0.00
5) Phenol-d5	6.51	99	940962	25.19	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.58	67	486367	24.90	ng/ul	0.00
9) 2-Chlorophenol-d4	6.66	132	761426	25.20	ng/ul	0.00
13) 4-Methylphenol-d8	7.26	113	728517	25.96	ng/ul	0.00
19) Nitrobenzene-d5	7.45	128	361749	27.62	ng/ul	0.00
22) 2-Nitrophenol-d4	7.78	143	379703	29.78	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	8.01	165	673760	24.90	ng/ul	0.00
29) 4-Chloroaniline-d4	8.23	131	974096	29.11	ng/ul	0.00
44) Dimethylphthalate-d6	9.63	166	2028358	28.09	ng/ul	0.00
47) Acenaphthylene-d8	9.79	160	2256156	25.28	ng/ul	0.00
52) 4-Nitrophenol-d4	10.03	143	336694	26.09	ng/ul	0.00
58) Fluorene-d10	10.47	176	1592199	25.96	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	10.52	200	191178	24.01	ng/ul	0.00
71) Anthracene-d10	11.50	188	2375043	28.31	ng/ul	0.00
79) Pyrene-d10	12.89	212	2430999	29.19	ng/ul	0.00
90) Benzo(a)pyrene-d12	15.58	264	2093531	28.98	ng/ul	0.00

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

