

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090921\
 Data File : BP006896.D
 Acq On : 09 Sep 2021 12:41
 Operator : CG/JU
 Sample : PB138843BS
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS843

Manual Integrations
 APPROVED

mohammad
 9/10/2021 3:25:56 PM

Quant Time: Sep 09 14:15:20 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP090821.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 09 11:07:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.940	152	274064	20.000	ng/ul	0.00
20) Naphthalene-d8	10.740	136	1091947	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.563	164	663131	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.310	188	1359218	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.392	240	1384205	20.000	ng/ul	# 0.00
88) Perylene-d12	23.827	264	1444265	20.000	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.370	96	47572	6.908	ng/uL	0.00
4) Pyridine-d5	3.781	84	597378	33.286	ng/ul	0.00
7) Phenol-d5	7.093	99	722276	35.857	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.269	67	417552	36.094	ng/ul	0.00
11) 2-Chlorophenol-d4	7.469	132	583670	35.893	ng/ul	0.00
15) 4-Methylphenol-d8	8.640	113	558519	35.309	ng/ul	0.00
21) Nitrobenzene-d5	9.099	128	257196	38.886	ng/ul	0.00
24) 2-Nitrophenol-d4	9.822	143	252846	38.830	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.357	165	552219	35.966	ng/ul	0.00
31) 4-Chloroaniline-d4	10.869	131	741221	34.909	ng/ul	0.00
46) Dimethylphthalate-d6	13.975	166	1561087	36.347	ng/ul	0.00
49) Acenaphthylene-d8	14.257	160	1980694	36.412	ng/ul	0.00
54) 4-Nitrophenol-d4	14.745	143	285649	38.283	ng/ul	0.00
60) Fluorene-d10	15.551	176	1363645	35.974	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.669	200	227535	37.842	ng/ul	0.00
73) Anthracene-d10	17.410	188	2105364	37.414	ng/ul	0.00
81) Pyrene-d10	19.639	212	2459010	37.933	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.669	264	2508918	36.009	ng/ul	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.405	88	107778	14.064	ng/uL	91
5) Pyridine	3.805	79	689165	38.281	ng/ul	89
6) Benzaldehyde	7.075	77	488151	38.031	ng/ul	89
8) Phenol	7.122	94	857053	41.161	ng/ul	90
10) Bis(2-Chloroethyl)ether	7.363	93	671130	40.259	ng/ul	98
12) 2-Chlorophenol	7.499	128	685120	41.067	ng/ul	94
13) 2-Methylphenol	8.375	108	633572	40.067	ng/ul	100
14) 2,2'-oxybis(1-Chloropr...	8.469	45	810078	42.071	ng/ul	96
16) Acetophenone	8.763	105	979335	40.371	ng/ul	91
17) N-Nitroso-di-n-propyla...	8.752	70	459922	40.692	ng/ul#	97
18) 4-Methylphenol	8.705	108	683400	40.296	ng/ul	100
19) Hexachloroethane	9.022	117	270447	39.809	ng/ul#	80
22) Nitrobenzene	9.140	77	694608	43.399	ng/ul#	88
23) Isophorone	9.669	82	1274829	40.105	ng/ul	95
25) 2-Nitrophenol	9.852	139	328762	45.171	ng/ul#	84
26) 2,4-Dimethylphenol	9.910	107	679774	38.552	ng/ul	93
27) Bis(2-Chloroethoxy)met...	10.152	93	857679	40.760	ng/ul	99
29) 2,4-Dichlorophenol	10.381	162	627875	41.292	ng/ul	94
30) Naphthalene	10.793	128	2082637	39.319	ng/ul	99
32) 4-Chloroaniline	10.899	127	840821	38.880	ng/ul	100
33) Hexachlorobutadiene	11.081	225	393815	37.543	ng/ul	96
34) Caprolactam	11.663	113	176407m	38.922	ng/ul	
35) 4-Chloro-3-methylphenol	12.016	107	637473	41.334	ng/ul	88
36) 2-Methylnaphthalene	12.399	142	1459249	41.091	ng/ul	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090921\
 Data File : BP006896.D
 Acq On : 09 Sep 2021 12:41
 Operator : CG/JU
 Sample : PB138843BS
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS843


Manual Integrations
 APPROVED

mohammad
 9/10/2021 3:25:56 PM

Quant Time: Sep 09 14:15:20 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP090821.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 09 11:07:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.616	142	1400451	38.659	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.763	216	761546	40.106	ng/ul#	95
40) Hexachlorocyclopentadiene	12.746	237	418511	34.150	ng/ul	99
41) 2,4,6-Trichlorophenol	12.998	196	474713	43.192	ng/ul	99
42) 2,4,5-Trichlorophenol	13.063	196	516572	42.162	ng/ul	97
43) 1,1'-Biphenyl	13.404	154	1854101	40.157	ng/ul#	97
44) 2-Chloronaphthalene	13.440	162	1433482	40.269	ng/ul	92
45) 2-Nitroaniline	13.640	65	349638	48.325	ng/ul#	81
47) Dimethylphthalate	14.022	163	1768723	40.833	ng/ul	99
48) 2,6-Dinitrotoluene	14.140	165	339650	46.950	ng/ul#	73
50) Acenaphthylene	14.287	152	2135266	37.832	ng/ul	99
51) 3-Nitroaniline	14.463	138	362128	43.668	ng/ul	83
52) Acenaphthene	14.628	153	1523494	40.765	ng/ul	98
53) 2,4-Dinitrophenol	14.663	184	166891	41.066	ng/ul#	70
55) 4-Nitrophenol	14.763	109	239715	43.707	ng/ul#	85
56) Dibenzofuran	14.963	168	2156116	39.980	ng/ul#	86
57) 2,4-Dinitrotoluene	14.922	165	494535	47.835	ng/ul#	75
58) 2,3,4,6-Tetrachlorophenol	15.181	232	432950	43.854	ng/ul#	79
59) Diethylphthalate	15.387	149	1752279	41.547	ng/ul	95
61) Fluorene	15.610	166	1733599	40.331	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.604	204	868984	39.656	ng/ul#	82
63) 4-Nitroaniline	15.622	138	367150	44.345	ng/ul#	83
66) 4,6-Dinitro-2-methylph...	15.681	198	271501	44.349	ng/ul#	74
67) N-Nitrosodiphenylamine	15.816	169	1515504	42.335	ng/ul	99
68) 4-Bromophenyl-phenylether	16.498	248	528507	41.128	ng/ul#	85
69) Hexachlorobenzene	16.616	284	606248	40.686	ng/ul#	86
70) Atrazine	16.769	200	545685	40.252	ng/ul	95
71) Pentachlorophenol	16.957	266	357081	41.838	ng/ul	98
72) Phenanthrene	17.351	178	2797939	41.323	ng/ul	99
74) Anthracene	17.445	178	2816170	41.844	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.369	216	756876	39.730	ng/uL	98
76) Pentachlorobenzene	14.881	250	716882	38.875	ng/uL	99
77) Carbazole	17.710	167	2569233	42.954	ng/ul	97
78) Di-n-butylphthalate	18.269	149	2927240	43.180	ng/ul#	98
80) Fluoranthene	19.316	202	3340595	41.433	ng/ul	95
82) Pyrene	19.669	202	3458179	41.816	ng/ul	94
83) Butylbenzylphthalate	20.539	149	1268980	44.067	ng/ul	91
84) 3,3'-Dichlorobenzidine	21.310	252	1176613	40.965	ng/ul#	98
85) Benzo(a)anthracene	21.375	228	3403782	42.235	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.304	149	1910758	44.509	ng/ul#	94
87) Chrysene	21.427	228	3324710	41.688	ng/ul	100
89) Di-n-octyl phthalate	22.239	149	3321455	44.498	ng/ul	100
90) Benzo(b)fluoranthene	23.080	252	3557333	42.095	ng/ul	97
91) Benzo(k)fluoranthene	23.133	252	3617758	45.734	ng/ul	98
93) Benzo(a)pyrene	23.721	252	3282272	40.368	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	26.380	276	4237649	44.927	ng/ul	97
95) Dibenzo(a,h)anthracene	26.398	278	3587968	43.965	ng/ul	97
96) Benzo(g,h,i)perylene	27.162	276	3531483	43.655	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed







#2
1,4-Dioxane
Concen: 14.064 ng/uL
RT: 3.405 min Scan# 54
Delta R.T. 0.006 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41

Tgt Ion: 88 Resp: 107778
Ion Ratio Lower Upper
88 100
43 25.7 20.1 30.1
58 69.7 48.4 72.6

Instrument :
BNA_P
ClientSampled :
SLCS843




Manual Integrations
APPROVED


mohammad
9/10/2021 3:25:56 PM



#5
Pyridine
Concen: 38.281 ng/uL
RT: 3.805 min Scan# 122
Delta R.T. 0.006 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41


Tgt Ion: 79 Resp: 689165
Ion Ratio Lower Upper
79 100
52 70.8 49.2 73.8
51 32.5 22.7 34.1





#6
Benzaldehyde
Concen: 38.031 ng/ul
RT: 7.075 min Scan# 678
Delta R.T. 0.006 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Instrument :
BNA_P
ClientSampleId :
SLCS843



Tgt Ion: 77 Resp: 488151
Ion Ratio Lower Upper
77 100
105 102.9 74.4 111.6
106 98.6 70.3 105.5




Manual Integrations
APPROVED


mohammad
9/10/2021 3:25:56 PM



#8
Phenol
Concen: 41.161 ng/ul
RT: 7.122 min Scan# 686
Delta R.T. 0.006 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41


Tgt Ion: 94 Resp: 857053
Ion Ratio Lower Upper
94 100
65 25.6 24.4 36.6
66 33.4 31.8 47.6





#10
Bis(2-Chloroethyl)ether
Concen: 40.259 ng/ul
RT: 7.363 min Scan# 727
Delta R.T. 0.006 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Instrument :
BNA_P
ClientSampled :
SLCS843




Tgt Ion: 93 Resp: 671130
Ion Ratio Lower Upper
93 100
63 73.2 59.7 89.5
95 32.4 25.3 37.9

Manual Integrations
APPROVED



mohammad
9/10/2021 3:25:56 PM




#12
2-Chlorophenol
Concen: 41.067 ng/ul
RT: 7.499 min Scan# 750
Delta R.T. -0.000 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41



Tgt Ion: 128 Resp: 685120
Ion Ratio Lower Upper
128 100
64 45.2 40.6 60.8
130 31.9 26.4 39.6









#13
2-Methylphenol
Concen: 40.067 ng/ul
RT: 8.375 min Scan# 899
Delta R.T. -0.000 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41

Tgt Ion: 108 Resp: 633572
Ion Ratio Lower Upper
108 100
107 90.6 72.2 108.2

Instrument :
BNA_P
ClientSampleId :
SLCS843




Manual Integrations
APPROVED


mohammad
9/10/2021 3:25:56 PM



#14
2,2'-oxybis(1-Chloropropane)
Concen: 42.071 ng/ul
RT: 8.469 min Scan# 915
Delta R.T. -0.000 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41


Tgt Ion: 45 Resp: 810078
Ion Ratio Lower Upper
45 100
77 15.8 13.8 20.8
79 11.8 10.9 16.3





#16
Acetophenone
Concen: 40.371 ng/ul
RT: 8.763 min Scan# 965
Delta R.T. 0.006 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Instrument :
BNA_P
ClientSampled :
SLCS843




Tgt Ion:105 Resp: 979335
Ion Ratio Lower Upper
105 100
77 73.0 67.0 100.4
51 23.4 19.0 28.6

Manual Integrations
APPROVED



mohammad
9/10/2021 3:25:56 PM




#17
N-Nitroso-di-n-propylamine
Concen: 40.692 ng/ul
RT: 8.752 min Scan# 963
Delta R.T. 0.006 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Tgt Ion: 70 Resp: 459922
Ion Ratio Lower Upper
70 100
42 50.7 40.4 60.6
101 10.6 7.2 10.8
130 23.0 14.9 22.3#





#18
4-Methylphenol
Concen: 40.296 ng/ul
RT: 8.705 min Scan# 955
Delta R.T. 0.006 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Instrument :
BNA_P
ClientSampleId :
SLCS843




Tgt Ion:108 Resp: 683400
Ion Ratio Lower Upper
108 100
107 110.4 88.6 133.0

Manual Integrations
APPROVED



mohammad
9/10/2021 3:25:56 PM




#19
Hexachloroethane
Concen: 39.809 ng/ul
RT: 9.022 min Scan# 1009
Delta R.T. -0.000 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Tgt Ion:117 Resp: 270447
Ion Ratio Lower Upper
117 100
201 104.5 68.2 102.2#
199 66.8 43.5 65.3#





#22
Nitrobenzene
Concen: 43.399 ng/ul
RT: 9.140 min Scan# 1029
Delta R.T. 0.006 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Instrument :
BNA_P
ClientSampled :
SLCS843




Tgt Ion: 77 Resp: 694608
Ion Ratio Lower Upper
77 100
123 48.6 31.1 46.7#
65 12.8 10.6 16.0

Manual Integrations
APPROVED



mohammad
9/10/2021 3:25:56 PM




#23
Isophorone
Concen: 40.105 ng/ul
RT: 9.669 min Scan# 1119
Delta R.T. 0.006 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41



Tgt Ion: 82 Resp: 1274829
Ion Ratio Lower Upper
82 100
95 7.1 5.8 8.8
138 18.5 12.5 18.7









#25
2-Nitrophenol
Concen: 45.171 ng/ul
RT: 9.852 min Scan# 1150
Delta R.T. -0.000 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41

Instrument :
BNA_P
ClientSampleId :
SLCS843

Tgt Ion:139 Resp: 328762
Ion Ratio Lower Upper
139 100
65 46.7 48.4 72.6#
109 28.3 28.2 42.2




Manual Integrations
APPROVED


mohammad
9/10/2021 3:25:56 PM



#26
2,4-Dimethylphenol
Concen: 38.552 ng/ul
RT: 9.910 min Scan# 1160
Delta R.T. 0.006 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41

Tgt Ion:107 Resp: 679774
Ion Ratio Lower Upper
107 100
121 51.5 38.1 57.1
122 87.8 64.2 96.4









#27
Bis(2-Chloroethoxy)methane
Concen: 40.760 ng/ul
RT: 10.152 min Scan# 1201
Delta R.T. -0.000 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41

Tgt Ion: 93 Resp: 857679
Ion Ratio Lower Upper
93 100
95 32.4 25.6 38.4
123 11.7 8.2 12.4

Instrument :
BNA_P
ClientSampled :
SLCS843




Manual Integrations
APPROVED


mohammad
9/10/2021 3:25:56 PM



#29
2,4-Dichlorophenol
Concen: 41.292 ng/ul
RT: 10.381 min Scan# 1240
Delta R.T. -0.000 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41


Tgt Ion: 162 Resp: 627875
Ion Ratio Lower Upper
162 100
164 66.2 52.0 78.0
98 35.4 34.3 51.5





#30
Naphthalene
Concen: 39.319 ng/ul
RT: 10.793 min Scan# 1310
Delta R.T. 0.006 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Instrument :
BNA_P
ClientSampleId :
SLCS843




Tgt Ion:128 Resp: 2082637
Ion Ratio Lower Upper
128 100
129 10.7 8.8 13.2
127 13.0 10.6 15.8

Manual Integrations
APPROVED



mohammad
9/10/2021 3:25:56 PM




#32
4-Chloroaniline
Concen: 38.880 ng/ul
RT: 10.899 min Scan# 1328
Delta R.T. 0.006 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Tgt Ion:127 Resp: 840821
Ion Ratio Lower Upper
127 100
129 32.3 25.7 38.5





#33
Hexachlorobutadiene
Concen: 37.543 ng/ul
RT: 11.081 min Scan# 1359
Delta R.T. -0.000 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Instrument :
BNA_P
ClientSampled :
SLCS843




Tgt Ion:225 Resp: 393815
Ion Ratio Lower Upper
225 100
223 64.2 48.2 72.2
227 63.6 49.6 74.4

Manual Integrations
APPROVED



mohammad
9/10/2021 3:25:56 PM




#34
Caprolactam
Concen: 38.922 ng/ul m
RT: 11.663 min Scan# 1458
Delta R.T. -0.000 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Tgt Ion:113 Resp: 176407
Ion Ratio Lower Upper
113 100
55 156.3 125.8 188.6
56 115.0 98.2 147.2





#35
4-Chloro-3-methylphenol
Concen: 41.334 ng/ul
RT: 12.016 min Scan# 1518
Delta R.T. 0.006 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Instrument :
BNA_P
ClientSampled :
SLCS843



Tgt Ion:107 Resp: 637473
Ion Ratio Lower Upper
107 100
144 26.5 18.8 28.2
142 84.9 58.4 87.6




Manual Integrations
APPROVED


mohammad
9/10/2021 3:25:56 PM



#36
2-Methylnaphthalene
Concen: 41.091 ng/ul
RT: 12.399 min Scan# 1583
Delta R.T. -0.000 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41


Tgt Ion:142 Resp: 1459249
Ion Ratio Lower Upper
142 100
141 89.4 72.3 108.5





#37
1-Methylnaphthalene
Concen: 38.659 ng/ul
RT: 12.616 min Scan# 1620
Delta R.T. -0.000 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Instrument :
BNA_P
ClientSampled :
SLCS843




Tgt Ion:142 Resp: 1400451
Ion Ratio Lower Upper
142 100
141 91.3 74.9 112.3
116 3.6 3.3 4.9

Manual Integrations
APPROVED



mohammad
9/10/2021 3:25:56 PM




#39
1,2,4,5-Tetrachlorobenzene
Concen: 40.106 ng/ul
RT: 12.763 min Scan# 1645
Delta R.T. -0.000 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Tgt Ion:216 Resp: 761546
Ion Ratio Lower Upper
216 100
214 79.0 64.3 96.5
179 20.6 19.3 28.9
108 16.5 19.8 29.6#





#40
Hexachlorocyclopentadiene
Concen: 34.150 ng/ul
RT: 12.746 min Scan# 1642
Delta R.T. -0.000 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Instrument :
BNA_P
ClientSampled :
SLCS843




Tgt Ion:237 Resp: 418511
Ion Ratio Lower Upper
237 100
235 62.3 50.4 75.6
272 13.1 9.7 14.5

Manual Integrations
APPROVED



mohammad
9/10/2021 3:25:56 PM




#41
2,4,6-Trichlorophenol
Concen: 43.192 ng/ul
RT: 12.998 min Scan# 1685
Delta R.T. -0.000 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Tgt Ion:196 Resp: 474713
Ion Ratio Lower Upper
196 100
198 94.9 76.2 114.4
200 31.3 24.6 36.8





#42
2,4,5-Trichlorophenol
Concen: 42.162 ng/ul
RT: 13.063 min Scan# 1696
Delta R.T. -0.006 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Instrument :
BNA_P
ClientSampled :
SLCS843




Tgt Ion:196 Resp: 516572
Ion Ratio Lower Upper
196 100
198 95.4 79.3 118.9
200 31.3 25.8 38.8

Manual Integrations
APPROVED



mohammad
9/10/2021 3:25:56 PM




#43
1,1'-Biphenyl
Concen: 40.157 ng/ul
RT: 13.404 min Scan# 1754
Delta R.T. -0.000 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41



Tgt Ion:154 Resp: 1854101
Ion Ratio Lower Upper
154 100
153 40.0 32.4 48.6
76 12.4 12.9 19.3#









#44
2-Chloronaphthalene
Concen: 40.269 ng/ul
RT: 13.440 min Scan# 1760
Delta R.T. -0.000 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41

Instrument :
BNA_P
ClientSampled :
SLCS843

Tgt Ion:162 Resp: 1433482
Ion Ratio Lower Upper
162 100
164 32.9 25.6 38.4
127 37.0 36.2 54.2




Manual Integrations
APPROVED


mohammad
9/10/2021 3:25:56 PM



#45
2-Nitroaniline
Concen: 48.325 ng/ul
RT: 13.640 min Scan# 1794
Delta R.T. -0.000 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41


Tgt Ion: 65 Resp: 349638
Ion Ratio Lower Upper
65 100
92 73.9 53.7 80.5
138 114.8 71.4 107.2#





#47
Dimethylphthalate
Concen: 40.833 ng/ul
RT: 14.022 min Scan# 1859
Delta R.T. -0.000 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Instrument :
BNA_P
ClientSampled :
SLCS843



Tgt Ion:163 Resp: 1768723
Ion Ratio Lower Upper
163 100
194 4.1 3.0 4.4
164 10.2 7.8 11.8




Manual Integrations
APPROVED


mohammad
9/10/2021 3:25:56 PM



#48
2,6-Dinitrotoluene
Concen: 46.950 ng/ul
RT: 14.140 min Scan# 1879
Delta R.T. -0.000 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41


Tgt Ion:165 Resp: 339650
Ion Ratio Lower Upper
165 100
89 50.7 62.2 93.2#
121 20.3 22.1 33.1#





#50
Acenaphthylene
Concen: 37.832 ng/ul
RT: 14.287 min Scan# 1904
Delta R.T. -0.000 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Instrument :
BNA_P
ClientSampled :
SLCS843




Tgt Ion:152 Resp: 2135266
Ion Ratio Lower Upper
152 100
151 20.2 15.9 23.9
153 13.2 10.4 15.6

Manual Integrations
APPROVED



mohammad
9/10/2021 3:25:56 PM




#51
3-Nitroaniline
Concen: 43.668 ng/ul
RT: 14.463 min Scan# 1934
Delta R.T. -0.000 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Tgt Ion:138 Resp: 362128
Ion Ratio Lower Upper
138 100
108 11.8 9.7 14.5
92 112.1 106.7 160.1





#52
Acenaphthene
Concen: 40.765 ng/ul
RT: 14.628 min Scan# 1962
Delta R.T. -0.000 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Instrument :
BNA_P
ClientSampled :
SLCS843




Tgt Ion:153 Resp: 1523494
Ion Ratio Lower Upper
153 100
152 47.5 38.6 57.8
154 88.9 69.0 103.6

Manual Integrations
APPROVED



mohammad
9/10/2021 3:25:56 PM




#53
2,4-Dinitrophenol
Concen: 41.066 ng/ul
RT: 14.663 min Scan# 1968
Delta R.T. -0.006 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Tgt Ion:184 Resp: 166891
Ion Ratio Lower Upper
184 100
63 56.0 73.4 110.2#
154 57.5 59.4 89.0#





#55
4-Nitrophenol
Concen: 43.707 ng/ul
RT: 14.763 min Scan# 1985
Delta R.T. 0.006 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Instrument :
BNA_P
ClientSampleId :
SLCS843



Tgt Ion:109 Resp: 239715
Ion Ratio Lower Upper
109 100
139 146.2 90.3 135.5#
65 128.5 102.2 153.2




Manual Integrations
APPROVED


mohammad
9/10/2021 3:25:56 PM



#56
Dibenzofuran
Concen: 39.980 ng/ul
RT: 14.963 min Scan# 2019
Delta R.T. -0.000 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41


Tgt Ion:168 Resp: 2156116
Ion Ratio Lower Upper
168 100
139 35.2 35.3 52.9#





#57
2,4-Dinitrotoluene
Concen: 47.835 ng/ul
RT: 14.922 min Scan# 2012
Delta R.T. -0.000 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Instrument :
BNA_P
ClientSampled :
SLCS843




Tgt Ion:165 Resp: 494535
Ion Ratio Lower Upper
165 100
63 37.5 45.2 67.8#
182 2.7 1.9 2.9

Manual Integrations
APPROVED



mohammad
9/10/2021 3:25:56 PM




#58
2,3,4,6-Tetrachlorophenol
Concen: 43.854 ng/ul
RT: 15.181 min Scan# 2056
Delta R.T. -0.006 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Tgt Ion:232 Resp: 432950
Ion Ratio Lower Upper
232 100
131 43.4 51.8 77.6#
130 2.2 3.1 4.7#
166 31.0 31.0 46.4





#59
Diethylphthalate
Concen: 41.547 ng/ul
RT: 15.387 min Scan# 2091
Delta R.T. -0.000 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Instrument :
BNA_P
ClientSampled :
SLCS843




Tgt Ion:149 Resp: 1752279
Ion Ratio Lower Upper
149 100
177 22.7 15.5 23.3
150 12.4 9.9 14.9

Manual Integrations
APPROVED



mohammad
9/10/2021 3:25:56 PM




#61
Fluorene
Concen: 40.331 ng/ul
RT: 15.610 min Scan# 2129
Delta R.T. -0.000 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Tgt Ion:166 Resp: 1733599
Ion Ratio Lower Upper
166 100
165 98.5 79.6 119.4
167 13.4 10.7 16.1





#62
4-Chlorophenyl-phenylether
Concen: 39.656 ng/ul
RT: 15.604 min Scan# 2128
Delta R.T. -0.000 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Instrument :
BNA_P
ClientSampled :
SLCS843




Tgt Ion:204 Resp: 868984
Ion Ratio Lower Upper
204 100
206 32.9 26.3 39.5
141 55.2 62.0 93.0#

Manual Integrations
APPROVED



mohammad
9/10/2021 3:25:56 PM




#63
4-Nitroaniline
Concen: 44.345 ng/ul
RT: 15.622 min Scan# 2131
Delta R.T. -0.000 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Tgt Ion:138 Resp: 367150
Ion Ratio Lower Upper
138 100
92 49.5 47.1 70.7
108 73.3 73.8 110.8#





#66
4,6-Dinitro-2-methylphenol
Concen: 44.349 ng/ul
RT: 15.681 min Scan# 2141
Delta R.T. -0.000 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Instrument :
BNA_P
ClientSampled :
SLCS843




Tgt Ion:198 Resp: 271501
Ion Ratio Lower Upper
198 100
182 4.1 4.5 6.7#
77 21.4 31.7 47.5#

Manual Integrations
APPROVED



mohammad
9/10/2021 3:25:56 PM




#67
N-Nitrosodiphenylamine
Concen: 42.335 ng/ul
RT: 15.816 min Scan# 2164
Delta R.T. -0.006 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Tgt Ion:169 Resp: 1515504
Ion Ratio Lower Upper
169 100
168 67.2 53.9 80.9
167 36.6 30.0 45.0





#68
4-Bromophenyl-phenylether
Concen: 41.128 ng/ul
RT: 16.498 min Scan# 2280
Delta R.T. -0.006 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Instrument :
BNA_P
ClientSampled :
SLCS843




Tgt Ion:248 Resp: 528507
Ion Ratio Lower Upper
248 100
250 98.1 78.4 117.6
141 70.5 80.6 121.0#

Manual Integrations
APPROVED



mohammad
9/10/2021 3:25:56 PM




#69
Hexachlorobenzene
Concen: 40.686 ng/ul
RT: 16.616 min Scan# 2300
Delta R.T. -0.006 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41



Tgt Ion:284 Resp: 606248
Ion Ratio Lower Upper
284 100
142 33.5 37.8 56.6#
249 31.0 27.1 40.7









#70
Atrazine
Concen: 40.252 ng/ul
RT: 16.769 min Scan# 2326
Delta R.T. -0.006 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41

Tgt Ion	Ratio	Lower	Upper
200	100		
173	26.5	22.2	33.4
215	48.6	35.4	53.0

Instrument :
BNA_P
ClientSampled :
SLCS843




Manual Integrations
APPROVED


mohammad
9/10/2021 3:25:56 PM



#71
Pentachlorophenol
Concen: 41.838 ng/ul
RT: 16.957 min Scan# 2358
Delta R.T. -0.006 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41


Tgt Ion	Ratio	Lower	Upper
266	100		
264	62.5	51.0	76.4
268	64.1	49.4	74.2





#72
Phenanthrene
Concen: 41.323 ng/ul
RT: 17.351 min Scan# 2425
Delta R.T. -0.006 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Instrument :
BNA_P
ClientSampled :
SLCS843




Tgt Ion:178 Resp: 2797939
Ion Ratio Lower Upper
178 100
179 15.9 12.0 18.0
176 19.1 15.6 23.4

Manual Integrations
APPROVED



mohammad
9/10/2021 3:25:56 PM




#74
Anthracene
Concen: 41.844 ng/ul
RT: 17.445 min Scan# 2441
Delta R.T. -0.006 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41



Tgt Ion:178 Resp: 2816170
Ion Ratio Lower Upper
178 100
179 15.7 12.1 18.1
176 18.7 14.9 22.3









#75
1,2,3,4-Tetrachlorobenzene
Concen: 39.730 ng/uL
RT: 13.369 min Scan# 1748
Delta R.T. -0.000 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41

Tgt Ion	Ratio	Lower	Upper
216	100		
214	78.0	62.9	94.3
218	49.5	37.8	56.8

Instrument :
BNA_P
ClientSampled :
SLCS843




Manual Integrations
APPROVED


mohammad
9/10/2021 3:25:56 PM



#76
Pentachlorobenzene
Concen: 38.875 ng/uL
RT: 14.881 min Scan# 2005
Delta R.T. -0.000 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41


Tgt Ion	Ratio	Lower	Upper
250	100		
248	63.2	51.0	76.6
252	62.7	51.0	76.4





#77
Carbazole
Concen: 42.954 ng/ul
RT: 17.710 min Scan# 2486
Delta R.T. -0.006 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Instrument :
BNA_P
ClientSampled :
SLCS843




Tgt Ion:167 Resp: 2569233
Ion Ratio Lower Upper
167 100
166 21.6 17.6 26.4
139 12.4 11.9 17.9

Manual Integrations
APPROVED



mohammad
9/10/2021 3:25:56 PM




#78
Di-n-butylphthalate
Concen: 43.180 ng/ul
RT: 18.269 min Scan# 2581
Delta R.T. -0.000 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Tgt Ion:149 Resp: 2927240
Ion Ratio Lower Upper
149 100
150 9.2 7.3 10.9
104 4.8 4.8 7.2#





#80
Fluoranthene
Concen: 41.433 ng/ul
RT: 19.316 min Scan# 2759
Delta R.T. -0.006 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Instrument :
BNA_P
ClientSampled :
SLCS843




Tgt Ion:202 Resp: 3340595
Ion Ratio Lower Upper
202 100
101 11.5 11.0 16.4
100 8.6 8.3 12.5

Manual Integrations
APPROVED



mohammad
9/10/2021 3:25:56 PM




#82
Pyrene
Concen: 41.816 ng/ul
RT: 19.669 min Scan# 2819
Delta R.T. -0.006 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Tgt Ion:202 Resp: 3458179
Ion Ratio Lower Upper
202 100
101 13.2 12.6 19.0
100 10.8 10.6 15.8





#83
Butylbenzylphthalate
Concen: 44.067 ng/ul
RT: 20.539 min Scan# 2967
Delta R.T. -0.006 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Instrument :
BNA_P
ClientSampled :
SLCS843




Tgt Ion:149 Resp: 1268980
Ion Ratio Lower Upper
149 100
91 70.5 63.0 94.6
206 20.9 14.1 21.1

Manual Integrations
APPROVED



mohammad
9/10/2021 3:25:56 PM




#84
3,3'-Dichlorobenzidine
Concen: 40.965 ng/ul
RT: 21.310 min Scan# 3098
Delta R.T. -0.006 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Tgt Ion:252 Resp: 1176613
Ion Ratio Lower Upper
252 100
254 65.3 51.6 77.4
126 10.1 10.7 16.1#





#85
Benzo(a)anthracene
Concen: 42.235 ng/ul
RT: 21.375 min Scan# 3109
Delta R.T. -0.006 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Instrument :
BNA_P
ClientSampled :
SLCS843




Tgt Ion:228 Resp: 3403782
Ion Ratio Lower Upper
228 100
229 20.1 15.8 23.8
226 27.5 21.6 32.4

Manual Integrations
APPROVED



mohammad
9/10/2021 3:25:56 PM




#86
Bis(2-ethylhexyl)phthalate
Concen: 44.509 ng/ul
RT: 21.304 min Scan# 3097
Delta R.T. -0.006 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Tgt Ion:149 Resp: 1910758
Ion Ratio Lower Upper
149 100
167 28.3 20.2 30.2
279 4.3 2.7 4.1#





#87
Chrysene
Concen: 41.688 ng/ul
RT: 21.427 min Scan# 3118
Delta R.T. -0.006 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Instrument :
BNA_P
ClientSampleId :
SLCS843



Tgt Ion:228 Resp: 3324710
Ion Ratio Lower Upper
228 100
226 30.3 24.5 36.7
229 19.9 15.9 23.9




Manual Integrations
APPROVED


mohammad
9/10/2021 3:25:56 PM



#89
Di-n-octyl phthalate
Concen: 44.498 ng/ul
RT: 22.239 min Scan# 3256
Delta R.T. -0.012 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41


Tgt Ion:149 Resp: 3321455





#90
Benzo(b)fluoranthene
Concen: 42.095 ng/ul
RT: 23.080 min Scan# 3399
Delta R.T. -0.006 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Instrument :
BNA_P
ClientSampled :
SLCS843




Tgt Ion:252 Resp: 3557333
Ion Ratio Lower Upper
252 100
253 22.7 17.7 26.5
125 10.5 10.1 15.1

Manual Integrations
APPROVED



mohammad
9/10/2021 3:25:56 PM




#91
Benzo(k)fluoranthene
Concen: 45.734 ng/ul
RT: 23.133 min Scan# 3408
Delta R.T. -0.006 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Tgt Ion:252 Resp: 3617758
Ion Ratio Lower Upper
252 100
253 22.1 17.2 25.8
125 10.3 9.4 14.0





#93
Benzo(a)pyrene
Concen: 40.368 ng/ul
RT: 23.721 min Scan# 3508
Delta R.T. -0.006 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Instrument :
BNA_P
ClientSampled :
SLCS843




Tgt Ion:252 Resp: 3282272
Ion Ratio Lower Upper
252 100
253 21.8 17.6 26.4
125 11.3 11.3 16.9

Manual Integrations
APPROVED



mohammad
9/10/2021 3:25:56 PM




#94
Indeno(1,2,3-cd)pyrene
Concen: 44.927 ng/ul
RT: 26.380 min Scan# 3960
Delta R.T. -0.012 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Tgt Ion:276 Resp: 4237649
Ion Ratio Lower Upper
276 100
138 25.2 21.8 32.8
277 25.4 19.9 29.9





#95
Dibenzo(a,h)anthracene
Concen: 43.965 ng/ul
RT: 26.398 min Scan# 3963
Delta R.T. -0.018 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41




Instrument :
BNA_P
ClientSampled :
SLCS843




Tgt Ion:278 Resp: 3587968
Ion Ratio Lower Upper
278 100
139 16.8 15.4 23.2
279 23.9 19.0 28.4

Manual Integrations
APPROVED

mohammad
9/10/2021 3:25:56 PM



#96
Benzo(g,h,i)perylene
Concen: 43.655 ng/ul
RT: 27.162 min Scan# 4093
Delta R.T. -0.012 min
Lab File: BP006896.D
Acq: 09 Sep 2021 12:41



Tgt Ion:276 Resp: 3531483
Ion Ratio Lower Upper
276 100
138 22.4 20.7 31.1
277 24.1 19.0 28.4

