

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090921\
 Data File : BP006924.D
 Acq On : 10 Sep 2021 05:29
 Operator : CG/JU
 Sample : PB138845BS
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS845

Manual Integrations
 APPROVED

mohammad
 9/10/2021 3:26:18 PM

Quant Time: Sep 10 06:42:54 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP090821.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 09 11:07:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.934	152	414035	20.000	ng/ul	0.00
20) Naphthalene-d8	10.740	136	1679326	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.563	164	1029745	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.310	188	2096080	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.392	240	2080653	20.000	ng/ul	0.00
88) Perylene-d12	23.827	264	2101940	20.000	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.369	96	68369	6.571	ng/uL	0.00
4) Pyridine-d5	3.787	84	917564	33.843	ng/ul	0.00
7) Phenol-d5	7.093	99	1137544	37.382	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.269	67	638493	36.534	ng/ul	0.00
11) 2-Chlorophenol-d4	7.469	132	910780	37.074	ng/ul	0.00
15) 4-Methylphenol-d8	8.640	113	869827	36.399	ng/ul	0.00
21) Nitrobenzene-d5	9.099	128	421218	41.410	ng/ul	0.00
24) 2-Nitrophenol-d4	9.816	143	427900	42.729	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.357	165	879404	37.242	ng/ul	0.00
31) 4-Chloroaniline-d4	10.869	131	1160211	35.530	ng/ul	0.00
46) Dimethylphthalate-d6	13.975	166	2423545	36.338	ng/ul	0.00
49) Acenaphthylene-d8	14.257	160	3073728	36.389	ng/ul	0.00
54) 4-Nitrophenol-d4	14.751	143	466237	40.239	ng/ul	0.00
60) Fluorene-d10	15.551	176	2097030	35.626	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.669	200	389839	42.042	ng/ul	0.00
73) Anthracene-d10	17.410	188	3221769	37.127	ng/ul	0.00
81) Pyrene-d10	19.639	212	3759746	38.584	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.668	264	3736140	36.844	ng/ul	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.405	88	158183	13.663	ng/uL	91
5) Pyridine	3.805	79	1033253	37.991	ng/ul	87
6) Benzaldehyde	7.075	77	755160m	38.944	ng/ul	
8) Phenol	7.122	94	1326006	42.154	ng/ul	91
10) Bis(2-Chloroethyl)ether	7.357	93	1031618	40.963	ng/ul	98
12) 2-Chlorophenol	7.499	128	1052109	41.745	ng/ul	94
13) 2-Methylphenol	8.375	108	988294	41.371	ng/ul	99
14) 2,2'-oxybis(1-Chloropr...	8.475	45	1225711	42.137	ng/ul	97
16) Acetophenone	8.763	105	1510549	41.218	ng/ul	91
17) N-Nitroso-di-n-propyla...	8.752	70	712729	41.741	ng/ul#	97
18) 4-Methylphenol	8.704	108	1061786	41.442	ng/ul	99
19) Hexachloroethane	9.022	117	414924	40.428	ng/ul#	78
22) Nitrobenzene	9.140	77	1082616	43.983	ng/ul#	88
23) Isophorone	9.669	82	2010594	41.128	ng/ul#	94
25) 2-Nitrophenol	9.851	139	542754	48.490	ng/ul#	82
26) 2,4-Dimethylphenol	9.910	107	998151	36.808	ng/ul	92
27) Bis(2-Chloroethoxy)met...	10.151	93	1315537	40.652	ng/ul	98
29) 2,4-Dichlorophenol	10.381	162	979645	41.892	ng/ul#	94
30) Naphthalene	10.787	128	3191582	39.180	ng/ul	100
32) 4-Chloroaniline	10.899	127	1308381	39.339	ng/ul	99
33) Hexachlorobutadiene	11.081	225	603069	37.383	ng/ul	96
34) Caprolactam	11.675	113	307365m	44.096	ng/ul	
35) 4-Chloro-3-methylphenol	12.010	107	1000761	42.193	ng/ul	91
36) 2-Methylnaphthalene	12.393	142	2267399	41.516	ng/ul	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090921\
 Data File : BP006924.D
 Acq On : 10 Sep 2021 05:29
 Operator : CG/JU
 Sample : PB138845BS
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS845


Manual Integrations
 APPROVED

mohammad
 9/10/2021 3:26:18 PM


Quant Time: Sep 10 06:42:54 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP090821.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 09 11:07:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.610	142	2175358	39.046	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.763	216	1180217	40.026	ng/ul#	94
40) Hexachlorocyclopentadiene	12.740	237	626400	32.916	ng/ul	98
41) 2,4,6-Trichlorophenol	12.998	196	757628	44.392	ng/ul	98
42) 2,4,5-Trichlorophenol	13.063	196	831722	43.716	ng/ul	97
43) 1,1'-Biphenyl	13.398	154	2856884	39.846	ng/ul#	97
44) 2-Chloronaphthalene	13.440	162	2197500	39.754	ng/ul	93
45) 2-Nitroaniline	13.640	65	579377	51.569	ng/ul#	79
47) Dimethylphthalate	14.022	163	2751733	40.909	ng/ul	99
48) 2,6-Dinitrotoluene	14.140	165	555519	49.451	ng/ul#	73
50) Acenaphthylene	14.281	152	3311837	37.788	ng/ul	99
51) 3-Nitroaniline	14.463	138	605251	47.001	ng/ul	83
52) Acenaphthene	14.628	153	2356352	40.603	ng/ul	98
53) 2,4-Dinitrophenol	14.669	184	301682	47.805	ng/ul#	69
55) 4-Nitrophenol	14.763	109	385719	45.289	ng/ul#	84
56) Dibenzofuran	14.957	168	3318956	39.632	ng/ul#	86
57) 2,4-Dinitrotoluene	14.922	165	795105	49.527	ng/ul#	74
58) 2,3,4,6-Tetrachlorophenol	15.181	232	684588	44.655	ng/ul#	79
59) Diethylphthalate	15.387	149	2724906	41.606	ng/ul	95
61) Fluorene	15.610	166	2670851	40.014	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.604	204	1350740	39.696	ng/ul#	80
63) 4-Nitroaniline	15.628	138	632313	49.182	ng/ul#	81
66) 4,6-Dinitro-2-methylph...	15.681	198	457355	48.445	ng/ul#	75
67) N-Nitrosodiphenylamine	15.816	169	2354009	42.641	ng/ul	99
68) 4-Bromophenyl-phenylether	16.498	248	817595	41.258	ng/ul#	83
69) Hexachlorobenzene	16.616	284	933698	40.633	ng/ul#	85
70) Atrazine	16.769	200	847905	40.558	ng/ul	95
71) Pentachlorophenol	16.957	266	586360	44.550	ng/ul	98
72) Phenanthrene	17.351	178	4283968	41.028	ng/ul	99
74) Anthracene	17.445	178	4316174	41.587	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.363	216	1176029	40.031	ng/uL	99
76) Pentachlorobenzene	14.881	250	1106124	38.896	ng/uL	99
77) Carbazole	17.710	167	3981915	43.169	ng/ul	98
78) Di-n-butylphthalate	18.263	149	4574659	43.759	ng/ul	98
80) Fluoranthene	19.316	202	5121720	42.261	ng/ul	94
82) Pyrene	19.669	202	5251026	42.241	ng/ul	94
83) Butylbenzylphthalate	20.539	149	2015432	46.561	ng/ul#	88
84) 3,3'-Dichlorobenzidine	21.304	252	1705570	39.505	ng/ul	99
85) Benzo(a)anthracene	21.374	228	5118869	42.255	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.304	149	2985457	46.265	ng/ul#	93
87) Chrysene	21.427	228	4925148	41.084	ng/ul	99
89) Di-n-octyl phthalate	22.239	149	5196013	47.831	ng/ul	100
90) Benzo(b)fluoranthene	23.080	252	5277984	42.914	ng/ul	97
91) Benzo(k)fluoranthene	23.133	252	5294075	45.985	ng/ul	98
93) Benzo(a)pyrene	23.721	252	4827603	40.796	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	26.380	276	6152196	44.816	ng/ul	97
95) Dibenzo(a,h)anthracene	26.398	278	5228792	44.024	ng/ul	97
96) Benzo(g,h,i)perylene	27.168	276	5091901	43.250	ng/ul	96




(#) = qualifier out of range (m) = manual integration (+) = signals summed




#2
1,4-Dioxane
Concen: 13.663 ng/uL
RT: 3.405 min Scan# 54
Delta R.T. 0.006 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29





Tgt Ion: 88 Resp: 158183
Ion Ratio Lower Upper
88 100
43 24.9 20.1 30.1
58 70.2 48.4 72.6



#5
Pyridine
Concen: 37.991 ng/uL
RT: 3.805 min Scan# 122
Delta R.T. 0.006 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Tgt Ion: 79 Resp: 1033253
Ion Ratio Lower Upper
79 100
52 72.6 49.2 73.8
51 32.9 22.7 34.1



Instrument :
BNA_P
ClientSampleId :
SLCS845

Manual Integrations
APPROVED

mohammad
9/10/2021 3:26:18 PM







#6
Benzaldehyde
Concen: 38.944 ng/ul m
RT: 7.075 min Scan# 678
Delta R.T. 0.006 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29

Tgt Ion: 77 Resp: 755160
Ion Ratio Lower Upper
77 100
105 102.4 74.4 111.6
106 99.5 70.3 105.5

Instrument :
BNA_P
ClientSampleId :
SLCS845




Manual Integrations
APPROVED


mohammad
9/10/2021 3:26:18 PM



#8
Phenol
Concen: 42.154 ng/ul
RT: 7.122 min Scan# 686
Delta R.T. 0.006 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29


Tgt Ion: 94 Resp: 1326006
Ion Ratio Lower Upper
94 100
65 25.9 24.4 36.6
66 34.0 31.8 47.6





#10
Bis(2-Chloroethyl)ether
Concen: 40.963 ng/ul
RT: 7.357 min Scan# 726
Delta R.T. -0.000 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Instrument :
BNA_P
ClientSampleId :
SLCS845




Tgt Ion: 93 Resp: 1031618
Ion Ratio Lower Upper
93 100
63 73.0 59.7 89.5
95 32.6 25.3 37.9

Manual Integrations
APPROVED



mohammad
9/10/2021 3:26:18 PM




#12
2-Chlorophenol
Concen: 41.745 ng/ul
RT: 7.499 min Scan# 750
Delta R.T. -0.000 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29



Tgt Ion: 128 Resp: 1052109
Ion Ratio Lower Upper
128 100
64 44.4 40.6 60.8
130 32.8 26.4 39.6









#13
2-Methylphenol
Concen: 41.371 ng/ul
RT: 8.375 min Scan# 899
Delta R.T. -0.000 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29

Tgt Ion: 108 Resp: 988294
Ion Ratio Lower Upper
108 100
107 89.0 72.2 108.2

Instrument :
BNA_P
ClientSampleId :
SLCS845




Manual Integrations
APPROVED


mohammad
9/10/2021 3:26:18 PM



#14
2,2'-oxybis(1-Chloropropane)
Concen: 42.137 ng/ul
RT: 8.475 min Scan# 916
Delta R.T. 0.006 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29


Tgt Ion: 45 Resp: 1225711
Ion Ratio Lower Upper
45 100
77 15.6 13.8 20.8
79 12.5 10.9 16.3





#16
Acetophenone
Concen: 41.218 ng/ul
RT: 8.763 min Scan# 965
Delta R.T. 0.006 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Instrument :
BNA_P
ClientSampled :
SLCS845




Tgt Ion:105 Resp: 1510549
Ion Ratio Lower Upper
105 100
77 73.2 67.0 100.4
51 23.6 19.0 28.6

Manual Integrations
APPROVED



mohammad
9/10/2021 3:26:18 PM




#17
N-Nitroso-di-n-propylamine
Concen: 41.741 ng/ul
RT: 8.752 min Scan# 963
Delta R.T. 0.006 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Tgt Ion: 70 Resp: 712729
Ion Ratio Lower Upper
70 100
42 50.3 40.4 60.6
101 10.4 7.2 10.8
130 22.5 14.9 22.3#





#18
4-Methylphenol
Concen: 41.442 ng/ul
RT: 8.704 min Scan# 955
Delta R.T. 0.006 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Instrument :
BNA_P
ClientSampleId :
SLCS845




Tgt Ion:108 Resp: 1061786
Ion Ratio Lower Upper
108 100
107 111.4 88.6 133.0

Manual Integrations
APPROVED



mohammad
9/10/2021 3:26:18 PM




#19
Hexachloroethane
Concen: 40.428 ng/ul
RT: 9.022 min Scan# 1009
Delta R.T. -0.000 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Tgt Ion:117 Resp: 414924
Ion Ratio Lower Upper
117 100
201 107.5 68.2 102.2#
199 67.1 43.5 65.3#





#22
Nitrobenzene
Concen: 43.983 ng/ul
RT: 9.140 min Scan# 1029
Delta R.T. 0.006 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Instrument :
BNA_P
ClientSampleId :
SLCS845




Tgt Ion: 77 Resp: 1082616
Ion Ratio Lower Upper
77 100
123 48.5 31.1 46.7#
65 12.6 10.6 16.0

Manual Integrations
APPROVED



mohammad
9/10/2021 3:26:18 PM




#23
Isophorone
Concen: 41.128 ng/ul
RT: 9.669 min Scan# 1119
Delta R.T. 0.006 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29



Tgt Ion: 82 Resp: 2010594
Ion Ratio Lower Upper
82 100
95 6.9 5.8 8.8
138 18.9 12.5 18.7#









#25
2-Nitrophenol
Concen: 48.490 ng/ul
RT: 9.851 min Scan# 1150
Delta R.T. -0.000 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29

Instrument :
BNA_P
ClientSampled :
SLCS845

Tgt Ion:139 Resp: 542754
Ion Ratio Lower Upper
139 100
65 45.1 48.4 72.6#
109 27.6 28.2 42.2#




Manual Integrations
APPROVED


mohammad
9/10/2021 3:26:18 PM



#26
2,4-Dimethylphenol
Concen: 36.808 ng/ul
RT: 9.910 min Scan# 1160
Delta R.T. 0.006 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29


Tgt Ion:107 Resp: 998151
Ion Ratio Lower Upper
107 100
121 51.9 38.1 57.1
122 88.8 64.2 96.4





#27
Bis(2-Chloroethoxy)methane
Concen: 40.652 ng/ul
RT: 10.151 min Scan# 1201
Delta R.T. -0.000 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Instrument :
BNA_P
ClientSampleId :
SLCS845



Tgt Ion: 93 Resp: 1315537
Ion Ratio Lower Upper
93 100
95 32.7 25.6 38.4
123 11.5 8.2 12.4




Manual Integrations
APPROVED


mohammad
9/10/2021 3:26:18 PM



#29
2,4-Dichlorophenol
Concen: 41.892 ng/ul
RT: 10.381 min Scan# 1240
Delta R.T. -0.000 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29


Tgt Ion: 162 Resp: 979645
Ion Ratio Lower Upper
162 100
164 65.8 52.0 78.0
98 33.7 34.3 51.5#





#30
Naphthalene
Concen: 39.180 ng/ul
RT: 10.787 min Scan# 1309
Delta R.T. -0.000 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Instrument :
BNA_P
ClientSampleId :
SLCS845




Tgt Ion:128 Resp: 3191582
Ion Ratio Lower Upper
128 100
129 10.9 8.8 13.2
127 13.2 10.6 15.8

Manual Integrations
APPROVED



mohammad
9/10/2021 3:26:18 PM




#32
4-Chloroaniline
Concen: 39.339 ng/ul
RT: 10.899 min Scan# 1328
Delta R.T. 0.006 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Tgt Ion:127 Resp: 1308381
Ion Ratio Lower Upper
127 100
129 32.5 25.7 38.5





#33
Hexachlorobutadiene
Concen: 37.383 ng/ul
RT: 11.081 min Scan# 1359
Delta R.T. -0.000 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Instrument :
BNA_P
ClientSampleId :
SLCS845




Tgt Ion:225 Resp: 603069
Ion Ratio Lower Upper
225 100
223 63.2 48.2 72.2
227 64.5 49.6 74.4

Manual Integrations
APPROVED



mohammad
9/10/2021 3:26:18 PM




#34
Caprolactam
Concen: 44.096 ng/ul m
RT: 11.675 min Scan# 1460
Delta R.T. 0.012 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Tgt Ion:113 Resp: 307365
Ion Ratio Lower Upper
113 100
55 153.6 125.8 188.6
56 113.3 98.2 147.2





#35
4-Chloro-3-methylphenol
Concen: 42.193 ng/ul
RT: 12.010 min Scan# 1517
Delta R.T. -0.000 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Instrument :
BNA_P
ClientSampleId :
SLCS845



Tgt Ion:107 Resp: 1000761
Ion Ratio Lower Upper
107 100
144 26.9 18.8 28.2
142 81.4 58.4 87.6




Manual Integrations
APPROVED


mohammad
9/10/2021 3:26:18 PM



#36
2-Methylnaphthalene
Concen: 41.516 ng/ul
RT: 12.393 min Scan# 1582
Delta R.T. -0.006 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29


Tgt Ion:142 Resp: 2267399
Ion Ratio Lower Upper
142 100
141 88.9 72.3 108.5





#37
1-Methylnaphthalene
Concen: 39.046 ng/ul
RT: 12.610 min Scan# 1619
Delta R.T. -0.006 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Instrument :
BNA_P
ClientSampleId :
SLCS845




Tgt Ion:142 Resp: 2175358
Ion Ratio Lower Upper
142 100
141 91.5 74.9 112.3
116 3.6 3.3 4.9

Manual Integrations
APPROVED



mohammad
9/10/2021 3:26:18 PM




#39
1,2,4,5-Tetrachlorobenzene
Concen: 40.026 ng/ul
RT: 12.763 min Scan# 1645
Delta R.T. -0.000 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Tgt Ion:216 Resp: 1180217
Ion Ratio Lower Upper
216 100
214 78.5 64.3 96.5
179 20.1 19.3 28.9
108 15.5 19.8 29.6#









#40
Hexachlorocyclopentadiene
Concen: 32.916 ng/ul
RT: 12.740 min Scan# 1641
Delta R.T. -0.006 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29





Tgt Ion:237 Resp: 626400
Ion Ratio Lower Upper
237 100
235 61.1 50.4 75.6
272 12.3 9.7 14.5



#41
2,4,6-Trichlorophenol
Concen: 44.392 ng/ul
RT: 12.998 min Scan# 1685
Delta R.T. -0.000 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Tgt Ion:196 Resp: 757628
Ion Ratio Lower Upper
196 100
198 98.1 76.2 114.4
200 31.4 24.6 36.8




Instrument :
BNA_P
ClientSampleId :
SLCS845

Manual Integrations
APPROVED




mohammad
9/10/2021 3:26:18 PM




#42
2,4,5-Trichlorophenol
Concen: 43.716 ng/ul
RT: 13.063 min Scan# 1696
Delta R.T. -0.006 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29





Tgt Ion:196 Resp: 831722
Ion Ratio Lower Upper
196 100
198 96.3 79.3 118.9
200 30.8 25.8 38.8



#43
1,1'-Biphenyl
Concen: 39.846 ng/ul
RT: 13.398 min Scan# 1753
Delta R.T. -0.006 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Tgt Ion:154 Resp: 2856884
Ion Ratio Lower Upper
154 100
153 40.3 32.4 48.6
76 12.5 12.9 19.3#



Instrument :
BNA_P
ClientSampleId :
SLCS845


Manual Integrations
APPROVED

mohammad
9/10/2021 3:26:18 PM



#44
2-Chloronaphthalene
Concen: 39.754 ng/ul
RT: 13.440 min Scan# 1760
Delta R.T. -0.000 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Instrument :
BNA_P
ClientSampleId :
SLCS845




Tgt Ion:162 Resp: 2197500
Ion Ratio Lower Upper
162 100
164 32.8 25.6 38.4
127 37.4 36.2 54.2

Manual Integrations
APPROVED



mohammad
9/10/2021 3:26:18 PM




#45
2-Nitroaniline
Concen: 51.569 ng/ul
RT: 13.640 min Scan# 1794
Delta R.T. -0.000 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Tgt Ion: 65 Resp: 579377
Ion Ratio Lower Upper
65 100
92 73.9 53.7 80.5
138 117.6 71.4 107.2#





#47
Dimethylphthalate
Concen: 40.909 ng/ul
RT: 14.022 min Scan# 1859
Delta R.T. -0.000 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Instrument :
BNA_P
ClientSampleId :
SLCS845




Tgt Ion:163 Resp: 2751733
Ion Ratio Lower Upper
163 100
194 4.3 3.0 4.4
164 10.2 7.8 11.8

Manual Integrations
APPROVED



mohammad
9/10/2021 3:26:18 PM




#48
2,6-Dinitrotoluene
Concen: 49.451 ng/ul
RT: 14.140 min Scan# 1879
Delta R.T. -0.000 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Tgt Ion:165 Resp: 555519
Ion Ratio Lower Upper
165 100
89 51.0 62.2 93.2#
121 19.3 22.1 33.1#





#50
Acenaphthylene
Concen: 37.788 ng/ul
RT: 14.281 min Scan# 1903
Delta R.T. -0.006 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Instrument :
BNA_P
ClientSampleId :
SLCS845




Tgt Ion:152 Resp: 3311837
Ion Ratio Lower Upper
152 100
151 20.4 15.9 23.9
153 13.5 10.4 15.6

Manual Integrations
APPROVED



mohammad
9/10/2021 3:26:18 PM




#51
3-Nitroaniline
Concen: 47.001 ng/ul
RT: 14.463 min Scan# 1934
Delta R.T. -0.000 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Tgt Ion:138 Resp: 605251
Ion Ratio Lower Upper
138 100
108 11.0 9.7 14.5
92 112.3 106.7 160.1





#52
Acenaphthene
Concen: 40.603 ng/ul
RT: 14.628 min Scan# 1962
Delta R.T. -0.000 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Instrument :
BNA_P
ClientSampleId :
SLCS845




Tgt Ion:153 Resp: 2356352
Ion Ratio Lower Upper
153 100
152 46.8 38.6 57.8
154 88.4 69.0 103.6

Manual Integrations
APPROVED



mohammad
9/10/2021 3:26:18 PM




#53
2,4-Dinitrophenol
Concen: 47.805 ng/ul
RT: 14.669 min Scan# 1969
Delta R.T. -0.000 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Tgt Ion:184 Resp: 301682
Ion Ratio Lower Upper
184 100
63 53.6 73.4 110.2#
154 57.0 59.4 89.0#





#55
4-Nitrophenol
Concen: 45.289 ng/ul
RT: 14.763 min Scan# 1985
Delta R.T. 0.006 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Instrument :
BNA_P
ClientSampleId :
SLCS845



Tgt Ion:109 Resp: 385719
Ion Ratio Lower Upper
109 100
139 148.3 90.3 135.5#
65 129.7 102.2 153.2




Manual Integrations
APPROVED


mohammad
9/10/2021 3:26:18 PM



#56
Dibenzofuran
Concen: 39.632 ng/ul
RT: 14.957 min Scan# 2018
Delta R.T. -0.006 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29


Tgt Ion:168 Resp: 3318956
Ion Ratio Lower Upper
168 100
139 35.0 35.3 52.9#





#57
2,4-Dinitrotoluene
Concen: 49.527 ng/ul
RT: 14.922 min Scan# 2012
Delta R.T. -0.000 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Instrument :
BNA_P
ClientSampleId :
SLCS845




Tgt Ion:165 Resp: 795105
Ion Ratio Lower Upper
165 100
63 36.5 45.2 67.8#
182 2.9 1.9 2.9#

Manual Integrations
APPROVED



mohammad
9/10/2021 3:26:18 PM




#58
2,3,4,6-Tetrachlorophenol
Concen: 44.655 ng/ul
RT: 15.181 min Scan# 2056
Delta R.T. -0.006 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Tgt Ion:232 Resp: 684588
Ion Ratio Lower Upper
232 100
131 43.4 51.8 77.6#
130 2.2 3.1 4.7#
166 31.0 31.0 46.4





#59
Diethylphthalate
Concen: 41.606 ng/ul
RT: 15.387 min Scan# 2091
Delta R.T. -0.000 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Instrument :
BNA_P
ClientSampleId :
SLCS845




Tgt Ion:149 Resp: 2724906
Ion Ratio Lower Upper
149 100
177 23.0 15.5 23.3
150 12.4 9.9 14.9

Manual Integrations
APPROVED



mohammad
9/10/2021 3:26:18 PM




#61
Fluorene
Concen: 40.014 ng/ul
RT: 15.610 min Scan# 2129
Delta R.T. -0.000 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Tgt Ion:166 Resp: 2670851
Ion Ratio Lower Upper
166 100
165 98.1 79.6 119.4
167 13.3 10.7 16.1





#62
4-Chlorophenyl-phenylether
Concen: 39.696 ng/ul
RT: 15.604 min Scan# 2128
Delta R.T. -0.000 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Instrument :
BNA_P
ClientSampleId :
SLCS845




Tgt Ion:204 Resp: 1350740
Ion Ratio Lower Upper
204 100
206 33.0 26.3 39.5
141 52.8 62.0 93.0#

Manual Integrations
APPROVED



mohammad
9/10/2021 3:26:18 PM




#63
4-Nitroaniline
Concen: 49.182 ng/ul
RT: 15.628 min Scan# 2132
Delta R.T. 0.006 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Tgt Ion:138 Resp: 632313
Ion Ratio Lower Upper
138 100
92 48.9 47.1 70.7
108 71.3 73.8 110.8#





#66
4,6-Dinitro-2-methylphenol
Concen: 48.445 ng/ul
RT: 15.681 min Scan# 2141
Delta R.T. -0.000 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Instrument :
BNA_P
ClientSampleId :
SLCS845




Tgt Ion:198 Resp: 457355
Ion Ratio Lower Upper
198 100
182 3.6 4.5 6.7#
77 22.4 31.7 47.5#

Manual Integrations
APPROVED



mohammad
9/10/2021 3:26:18 PM




#67
N-Nitrosodiphenylamine
Concen: 42.641 ng/ul
RT: 15.816 min Scan# 2164
Delta R.T. -0.006 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Tgt Ion:169 Resp: 2354009
Ion Ratio Lower Upper
169 100
168 66.9 53.9 80.9
167 36.0 30.0 45.0





#68
4-Bromophenyl-phenylether
Concen: 41.258 ng/ul
RT: 16.498 min Scan# 2280
Delta R.T. -0.006 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Instrument :
BNA_P
ClientSampleId :
SLCS845




Tgt Ion:248 Resp: 817595
Ion Ratio Lower Upper
248 100
250 96.2 78.4 117.6
141 68.6 80.6 121.0#

Manual Integrations
APPROVED



mohammad
9/10/2021 3:26:18 PM




#69
Hexachlorobenzene
Concen: 40.633 ng/ul
RT: 16.616 min Scan# 2300
Delta R.T. -0.006 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29



Tgt Ion:284 Resp: 933698
Ion Ratio Lower Upper
284 100
142 32.5 37.8 56.6#
249 30.9 27.1 40.7









#70
Atrazine
Concen: 40.558 ng/ul
RT: 16.769 min Scan# 2326
Delta R.T. -0.006 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29

Tgt Ion:200 Resp: 847905
Ion Ratio Lower Upper
200 100
173 26.3 22.2 33.4
215 48.5 35.4 53.0

Instrument :
BNA_P
ClientSampleId :
SLCS845




Manual Integrations
APPROVED


mohammad
9/10/2021 3:26:18 PM



#71
Pentachlorophenol
Concen: 44.550 ng/ul
RT: 16.957 min Scan# 2358
Delta R.T. -0.006 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29


Tgt Ion:266 Resp: 586360
Ion Ratio Lower Upper
266 100
264 62.5 51.0 76.4
268 64.2 49.4 74.2





#72
Phenanthrene
Concen: 41.028 ng/ul
RT: 17.351 min Scan# 2425
Delta R.T. -0.006 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Instrument :
BNA_P
ClientSampleId :
SLCS845



Tgt Ion:178 Resp: 4283968
Ion Ratio Lower Upper
178 100
179 15.7 12.0 18.0
176 19.6 15.6 23.4




Manual Integrations
APPROVED


mohammad
9/10/2021 3:26:18 PM



#74
Anthracene
Concen: 41.587 ng/ul
RT: 17.445 min Scan# 2441
Delta R.T. -0.006 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29

Tgt Ion:178 Resp: 4316174
Ion Ratio Lower Upper
178 100
179 15.8 12.1 18.1
176 19.1 14.9 22.3









#75
1,2,3,4-Tetrachlorobenzene
Concen: 40.031 ng/uL
RT: 13.363 min Scan# 1747
Delta R.T. -0.006 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29

Tgt Ion:216 Resp: 1176029
Ion Ratio Lower Upper
216 100
214 79.2 62.9 94.3
218 49.1 37.8 56.8

Instrument :
BNA_P
ClientSampleId :
SLCS845




Manual Integrations
APPROVED


mohammad
9/10/2021 3:26:18 PM



#76
Pentachlorobenzene
Concen: 38.896 ng/uL
RT: 14.881 min Scan# 2005
Delta R.T. -0.000 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29


Tgt Ion:250 Resp: 1106124
Ion Ratio Lower Upper
250 100
248 62.3 51.0 76.6
252 63.7 51.0 76.4





#77
Carbazole
Concen: 43.169 ng/ul
RT: 17.710 min Scan# 2486
Delta R.T. -0.006 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Instrument :
BNA_P
ClientSampleId :
SLCS845




Tgt Ion:167 Resp: 3981915
Ion Ratio Lower Upper
167 100
166 21.9 17.6 26.4
139 12.6 11.9 17.9

Manual Integrations
APPROVED



mohammad
9/10/2021 3:26:18 PM




#78
Di-n-butylphthalate
Concen: 43.759 ng/ul
RT: 18.263 min Scan# 2580
Delta R.T. -0.006 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29



Tgt Ion:149 Resp: 4574659
Ion Ratio Lower Upper
149 100
150 9.5 7.3 10.9
104 4.9 4.8 7.2









#80
Fluoranthene
Concen: 42.261 ng/ul
RT: 19.316 min Scan# 2759
Delta R.T. -0.006 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29

Tgt Ion:202 Resp: 5121720
Ion Ratio Lower Upper
202 100
101 11.2 11.0 16.4
100 8.6 8.3 12.5

Instrument :
BNA_P
ClientSampleId :
SLCS845




Manual Integrations
APPROVED


mohammad
9/10/2021 3:26:18 PM



#82
Pyrene
Concen: 42.241 ng/ul
RT: 19.669 min Scan# 2819
Delta R.T. -0.006 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29


Tgt Ion:202 Resp: 5251026
Ion Ratio Lower Upper
202 100
101 13.0 12.6 19.0
100 10.8 10.6 15.8





#83
Butylbenzylphthalate
Concen: 46.561 ng/ul
RT: 20.539 min Scan# 2967
Delta R.T. -0.006 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Instrument :
BNA_P
ClientSampleId :
SLCS845




Tgt Ion:149 Resp: 2015432
Ion Ratio Lower Upper
149 100
91 67.6 63.0 94.6
206 22.2 14.1 21.1#

Manual Integrations
APPROVED



mohammad
9/10/2021 3:26:18 PM




#84
3,3'-Dichlorobenzidine
Concen: 39.505 ng/ul
RT: 21.304 min Scan# 3097
Delta R.T. -0.012 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Tgt Ion:252 Resp: 1705570
Ion Ratio Lower Upper
252 100
254 64.4 51.6 77.4
126 11.1 10.7 16.1





#85
Benzo(a)anthracene
Concen: 42.255 ng/ul
RT: 21.374 min Scan# 3109
Delta R.T. -0.006 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Instrument :
BNA_P
ClientSampleId :
SLCS845




Tgt Ion:228 Resp: 5118869
Ion Ratio Lower Upper
228 100
229 20.3 15.8 23.8
226 28.0 21.6 32.4

Manual Integrations
APPROVED



mohammad
9/10/2021 3:26:18 PM




#86
Bis(2-ethylhexyl)phthalate
Concen: 46.265 ng/ul
RT: 21.304 min Scan# 3097
Delta R.T. -0.006 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29



Tgt Ion:149 Resp: 2985457
Ion Ratio Lower Upper
149 100
167 29.0 20.2 30.2
279 4.9 2.7 4.1#









#87
Chrysene
Concen: 41.084 ng/ul
RT: 21.427 min Scan# 3118
Delta R.T. -0.006 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29

Tgt Ion:228 Resp: 4925148
Ion Ratio Lower Upper
228 100
226 30.9 24.5 36.7
229 20.5 15.9 23.9

Instrument :
BNA_P
ClientSampleId :
SLCS845




Manual Integrations
APPROVED


mohammad
9/10/2021 3:26:18 PM



#89
Di-n-octyl phthalate
Concen: 47.831 ng/ul
RT: 22.239 min Scan# 3256
Delta R.T. -0.012 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29


Tgt Ion:149 Resp: 5196013





#90
Benzo(b)fluoranthene
Concen: 42.914 ng/ul
RT: 23.080 min Scan# 3399
Delta R.T. -0.006 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Instrument :
BNA_P
ClientSampled :
SLCS845




Tgt Ion:252 Resp: 5277984
Ion Ratio Lower Upper
252 100
253 22.5 17.7 26.5
125 10.4 10.1 15.1

Manual Integrations
APPROVED



mohammad
9/10/2021 3:26:18 PM




#91
Benzo(k)fluoranthene
Concen: 45.985 ng/ul
RT: 23.133 min Scan# 3408
Delta R.T. -0.006 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Tgt Ion:252 Resp: 5294075
Ion Ratio Lower Upper
252 100
253 22.2 17.2 25.8
125 10.1 9.4 14.0





#93
Benzo(a)pyrene
Concen: 40.796 ng/ul
RT: 23.721 min Scan# 3508
Delta R.T. -0.006 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Instrument :
BNA_P
ClientSampleId :
SLCS845




Tgt Ion:252 Resp: 4827603
Ion Ratio Lower Upper
252 100
253 22.0 17.6 26.4
125 11.3 11.3 16.9

Manual Integrations
APPROVED



mohammad
9/10/2021 3:26:18 PM




#94
Indeno(1,2,3-cd)pyrene
Concen: 44.816 ng/ul
RT: 26.380 min Scan# 3960
Delta R.T. -0.012 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Tgt Ion:276 Resp: 6152196
Ion Ratio Lower Upper
276 100
138 24.3 21.8 32.8
277 25.0 19.9 29.9





#95
Dibenzo(a,h)anthracene
Concen: 44.024 ng/ul
RT: 26.398 min Scan# 3963
Delta R.T. -0.018 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29




Instrument :
BNA_P
ClientSampled :
SLCS845




Tgt Ion:278 Resp: 5228792
Ion Ratio Lower Upper
278 100
139 16.8 15.4 23.2
279 24.2 19.0 28.4

Manual Integrations
APPROVED

mohammad
9/10/2021 3:26:18 PM



#96
Benzo(g,h,i)perylene
Concen: 43.250 ng/ul
RT: 27.168 min Scan# 4094
Delta R.T. -0.006 min
Lab File: BP006924.D
Acq: 10 Sep 2021 05:29



Tgt Ion:276 Resp: 5091901
Ion Ratio Lower Upper
276 100
138 22.3 20.7 31.1
277 24.1 19.0 28.4

