

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090921\
 Data File : BP006946.D
 Acq On : 10 Sep 2021 18:02
 Operator : CG/JU
 Sample : PB138976BS
 Misc :
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS976

Manual Integrations
 APPROVED

mohammad
 9/13/2021 10:22:06 AM

Quant Time: Sep 10 23:35:46 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP090821.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 09 11:07:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.934	152	307104	20.000	ng/ul	0.00
20) Naphthalene-d8	10.740	136	1274453	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.557	164	801607	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.310	188	1637230	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.392	240	1622976	20.000	ng/ul	# 0.00
88) Perylene-d12	23.821	264	1586314	20.000	ng/ul	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.370	96	50606	6.558	ng/uL	0.00
4) Pyridine-d5	3.781	84	673109	33.471	ng/ul	0.00
7) Phenol-d5	7.093	99	818564	36.265	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.263	67	469559	36.222	ng/ul	0.00
11) 2-Chlorophenol-d4	7.463	132	653918	35.887	ng/ul	0.00
15) 4-Methylphenol-d8	8.640	113	640586	36.140	ng/ul	0.00
21) Nitrobenzene-d5	9.093	128	300595	38.940	ng/ul	0.00
24) 2-Nitrophenol-d4	9.816	143	302080	39.748	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.352	165	642750	35.867	ng/ul	0.00
31) 4-Chloroaniline-d4	10.869	131	812913	32.803	ng/ul	0.00
46) Dimethylphthalate-d6	13.969	166	1805877	34.783	ng/ul	0.00
49) Acenaphthylene-d8	14.251	160	2287128	34.783	ng/ul	0.00
54) 4-Nitrophenol-d4	14.745	143	335935	37.245	ng/ul	0.00
60) Fluorene-d10	15.551	176	1581697	34.519	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.669	200	263764	36.418	ng/ul	0.00
73) Anthracene-d10	17.410	188	2421811	35.730	ng/ul	0.00
81) Pyrene-d10	19.639	212	2775979	36.522	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.663	264	2635647	34.440	ng/ul	-0.02
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.399	88	108067	12.584	ng/uL	90
5) Pyridine	3.805	79	688825	34.146	ng/ul	89
6) Benzaldehyde	7.075	77	527641m	36.685	ng/ul	
8) Phenol	7.122	94	877715	37.618	ng/ul	90
10) Bis(2-Chloroethyl)ether	7.358	93	672679	36.011	ng/ul	98
12) 2-Chlorophenol	7.499	128	690307	36.926	ng/ul	94
13) 2-Methylphenol	8.375	108	651406	36.763	ng/ul	100
14) 2,2'-oxybis(1-Chloropr...	8.469	45	815923	37.816	ng/ul	97
16) Acetophenone	8.758	105	1005676	36.997	ng/ul	91
17) N-Nitroso-di-n-propyla...	8.746	70	478196	37.757	ng/ul#	97
18) 4-Methylphenol	8.705	108	704227	37.057	ng/ul	100
19) Hexachloroethane	9.016	117	270910	35.587	ng/ul	85
22) Nitrobenzene	9.134	77	721460	38.622	ng/ul#	89
23) Isophorone	9.663	82	1331837	35.898	ng/ul	96
25) 2-Nitrophenol	9.846	139	349207	41.110	ng/ul	87
26) 2,4-Dimethylphenol	9.905	107	730710	35.506	ng/ul	92
27) Bis(2-Chloroethoxy)met...	10.146	93	878073	35.754	ng/ul	99
29) 2,4-Dichlorophenol	10.381	162	646410	36.423	ng/ul	94
30) Naphthalene	10.787	128	2129088	34.440	ng/ul	100
32) 4-Chloroaniline	10.893	127	826752	32.755	ng/ul	99
33) Hexachlorobutadiene	11.075	225	398273	32.531	ng/ul	96
34) Caprolactam	11.663	113	197106m	37.261	ng/ul	
35) 4-Chloro-3-methylphenol	12.010	107	664337	36.907	ng/ul	90
36) 2-Methylnaphthalene	12.393	142	1511298	36.463	ng/ul	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090921\
 Data File : BP006946.D
 Acq On : 10 Sep 2021 18:02
 Operator : CG/JU
 Sample : PB138976BS
 Misc :
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS976

Manual Integrations
 APPROVED

mohammad
 9/13/2021 10:22:06 AM

Quant Time: Sep 10 23:35:46 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP090821.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 09 11:07:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.610	142	1442170	34.109	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.757	216	787382	34.303	ng/ul#	93
40) Hexachlorocyclopentadiene	12.740	237	417455	28.179	ng/ul	99
41) 2,4,6-Trichlorophenol	12.993	196	498150	37.495	ng/ul	99
42) 2,4,5-Trichlorophenol	13.063	196	549903	37.129	ng/ul	97
43) 1,1'-Biphenyl	13.398	154	1925878	34.506	ng/ul#	97
44) 2-Chloronaphthalene	13.440	162	1475339	34.285	ng/ul	92
45) 2-Nitroaniline	13.640	65	374934	42.869	ng/ul#	78
47) Dimethylphthalate	14.016	163	1828604	34.922	ng/ul	98
48) 2,6-Dinitrotoluene	14.134	165	351157	40.156	ng/ul#	78
50) Acenaphthylene	14.281	152	2216881	32.493	ng/ul	100
51) 3-Nitroaniline	14.463	138	369089	36.819	ng/ul	80
52) Acenaphthene	14.622	153	1584207	35.067	ng/ul	97
53) 2,4-Dinitrophenol	14.663	184	155357	31.624	ng/ul#	76
55) 4-Nitrophenol	14.763	109	251168	37.884	ng/ul#	83
56) Dibenzofuran	14.957	168	2238294	34.334	ng/ul	87
57) 2,4-Dinitrotoluene	14.916	165	515852	41.278	ng/ul#	79
58) 2,3,4,6-Tetrachlorophenol	15.181	232	436424	36.570	ng/ul#	78
59) Diethylphthalate	15.381	149	1806733	35.438	ng/ul	95
61) Fluorene	15.610	166	1825795	35.139	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.604	204	909426	34.333	ng/ul#	80
63) 4-Nitroaniline	15.622	138	409837	40.950	ng/ul	85
66) 4,6-Dinitro-2-methylph...	15.681	198	278942	37.827	ng/ul#	75
67) N-Nitrosodiphenylamine	15.816	169	1567382	36.349	ng/ul	99
68) 4-Bromophenyl-phenylether	16.498	248	540866	34.943	ng/ul#	83
69) Hexachlorobenzene	16.610	284	613514	34.182	ng/ul#	88
70) Atrazine	16.769	200	551906	33.798	ng/ul	94
71) Pentachlorophenol	16.957	266	332690	32.361	ng/ul	99
72) Phenanthrene	17.351	178	2894867	35.494	ng/ul	99
74) Anthracene	17.439	178	2922601	36.052	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.363	216	805508	35.103	ng/uL	99
76) Pentachlorobenzene	14.881	250	731217	32.919	ng/uL	100
77) Carbazole	17.710	167	2645239	36.715	ng/ul	97
78) Di-n-butylphthalate	18.263	149	2982530	36.525	ng/ul	98
80) Fluoranthene	19.316	202	3392924	35.891	ng/ul#	93
82) Pyrene	19.669	202	3517323	36.274	ng/ul#	92
83) Butylbenzylphthalate	20.539	149	1285025	38.059	ng/ul#	88
84) 3,3'-Dichlorobenzidine	21.304	252	1090598	32.384	ng/ul	99
85) Benzo(a)anthracene	21.375	228	3341061	35.357	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.304	149	1955581	38.851	ng/ul#	94
87) Chrysene	21.427	228	3274122	35.014	ng/ul	100
89) Di-n-octyl phthalate	22.233	149	3311378	40.390	ng/ul	100
90) Benzo(b)fluoranthene	23.080	252	3474619	37.434	ng/ul	98
91) Benzo(k)fluoranthene	23.127	252	3306379	38.055	ng/ul	98
93) Benzo(a)pyrene	23.716	252	3073158	34.411	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	26.374	276	3834042	37.008	ng/ul	98
95) Dibenzo(a,h)anthracene	26.392	278	3242830	36.178	ng/ul	97
96) Benzo(g,h,i)perylene	27.157	276	3171667	35.696	ng/ul	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090921\
 Data File : BP006946.D
 Acq On : 10 Sep 2021 18:02
 Operator : CG/JU
 Sample : PB138976BS
 Misc :
 ALS Vial : 46 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampled :
 SLCS976

Manual Integrations
 APPROVED

mohammad
 9/13/2021 10:22:06 AM


Quant Time: Sep 10 23:35:46 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP090821.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 09 11:07:24 2021
 Response via : Initial Calibration





#2
1,4-Dioxane
Concen: 12.584 ng/uL
RT: 3.399 min Scan# 53
Delta R.T. -0.000 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Instrument :
BNA_P
ClientSampleId :
SLCS976




Tgt Ion: 88 Resp: 108067
Ion Ratio Lower Upper
88 100
43 25.8 20.1 30.1
58 71.0 48.4 72.6

Manual Integrations
APPROVED



mohammad
9/13/2021 10:22:06 AM




#5
Pyridine
Concen: 34.146 ng/uL
RT: 3.805 min Scan# 122
Delta R.T. 0.006 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Tgt Ion: 79 Resp: 688825
Ion Ratio Lower Upper
79 100
52 70.7 49.2 73.8
51 32.4 22.7 34.1





#6
Benzaldehyde
Concen: 36.685 ng/ul m
RT: 7.075 min Scan# 678
Delta R.T. 0.006 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Instrument :
BNA_P
ClientSampleId :
SLCS976




Tgt Ion: 77 Resp: 527641
Ion Ratio Lower Upper
77 100
105 103.9 74.4 111.6
106 100.9 70.3 105.5

Manual Integrations
APPROVED



mohammad
9/13/2021 10:22:06 AM




#8
Phenol
Concen: 37.618 ng/ul
RT: 7.122 min Scan# 686
Delta R.T. 0.006 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Tgt Ion: 94 Resp: 877715
Ion Ratio Lower Upper
94 100
65 25.1 24.4 36.6
66 33.1 31.8 47.6





#10
Bis(2-Chloroethyl)ether
Concen: 36.011 ng/ul
RT: 7.358 min Scan# 726
Delta R.T. -0.000 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Instrument :
BNA_P
ClientSampleId :
SLCS976




Tgt Ion: 93 Resp: 672679
Ion Ratio Lower Upper
93 100
63 72.7 59.7 89.5
95 32.3 25.3 37.9

Manual Integrations
APPROVED



mohammad
9/13/2021 10:22:06 AM




#12
2-Chlorophenol
Concen: 36.926 ng/ul
RT: 7.499 min Scan# 750
Delta R.T. -0.000 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02



Tgt Ion: 128 Resp: 690307
Ion Ratio Lower Upper
128 100
64 43.9 40.6 60.8
130 32.6 26.4 39.6









#13
2-Methylphenol
Concen: 36.763 ng/ul
RT: 8.375 min Scan# 899
Delta R.T. -0.000 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02

Instrument :
BNA_P
ClientSampleId :
SLCS976

Tgt Ion:108 Resp: 651406
Ion Ratio Lower Upper
108 100
107 90.0 72.2 108.2




Manual Integrations
APPROVED


mohammad
9/13/2021 10:22:06 AM



#14
2,2'-oxybis(1-Chloropropane)
Concen: 37.816 ng/ul
RT: 8.469 min Scan# 915
Delta R.T. -0.000 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02


Tgt Ion: 45 Resp: 815923
Ion Ratio Lower Upper
45 100
77 16.2 13.8 20.8
79 12.1 10.9 16.3





#16
Acetophenone
Concen: 36.997 ng/ul
RT: 8.758 min Scan# 964
Delta R.T. -0.000 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Instrument :
BNA_P
ClientSampleId :
SLCS976




Tgt Ion:105 Resp: 1005676
Ion Ratio Lower Upper
105 100
77 73.2 67.0 100.4
51 23.8 19.0 28.6

Manual Integrations
APPROVED



mohammad
9/13/2021 10:22:06 AM




#17
N-Nitroso-di-n-propylamine
Concen: 37.757 ng/ul
RT: 8.746 min Scan# 962
Delta R.T. -0.000 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Tgt Ion: 70 Resp: 478196
Ion Ratio Lower Upper
70 100
42 50.7 40.4 60.6
101 10.1 7.2 10.8
130 22.6 14.9 22.3#





#18
4-Methylphenol
Concen: 37.057 ng/ul
RT: 8.705 min Scan# 955
Delta R.T. 0.006 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Instrument :
BNA_P
ClientSampleId :
SLCS976




Tgt Ion:108 Resp: 704227
Ion Ratio Lower Upper
108 100
107 111.3 88.6 133.0

Manual Integrations
APPROVED



mohammad
9/13/2021 10:22:06 AM




#19
Hexachloroethane
Concen: 35.587 ng/ul
RT: 9.016 min Scan# 1008
Delta R.T. -0.006 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Tgt Ion:117 Resp: 270910
Ion Ratio Lower Upper
117 100
201 101.6 68.2 102.2
199 62.4 43.5 65.3






#22
Nitrobenzene
Concen: 38.622 ng/ul
RT: 9.134 min Scan# 1028
Delta R.T. -0.000 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02





Tgt Ion: 77 Resp: 721460
Ion Ratio Lower Upper
77 100
123 47.5 31.1 46.7#
65 13.2 10.6 16.0




Instrument :
BNA_P
ClientSampled :
SLCS976

Manual Integrations
APPROVED



mohammad
9/13/2021 10:22:06 AM




#23
Isophorone
Concen: 35.898 ng/ul
RT: 9.663 min Scan# 1118
Delta R.T. -0.000 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Tgt Ion: 82 Resp: 1331837
Ion Ratio Lower Upper
82 100
95 6.7 5.8 8.8
138 18.0 12.5 18.7





#25
2-Nitrophenol
Concen: 41.110 ng/ul
RT: 9.846 min Scan# 1149
Delta R.T. -0.006 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Instrument :
BNA_P
ClientSampleId :
SLCS976



Tgt Ion:139 Resp: 349207
Ion Ratio Lower Upper
139 100
65 50.0 48.4 72.6
109 28.8 28.2 42.2




Manual Integrations
APPROVED


mohammad
9/13/2021 10:22:06 AM



#26
2,4-Dimethylphenol
Concen: 35.506 ng/ul
RT: 9.905 min Scan# 1159
Delta R.T. -0.000 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02


Tgt Ion:107 Resp: 730710
Ion Ratio Lower Upper
107 100
121 51.7 38.1 57.1
122 88.1 64.2 96.4





#27
Bis(2-Chloroethoxy)methane
Concen: 35.754 ng/ul
RT: 10.146 min Scan# 1200
Delta R.T. -0.006 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Instrument :
BNA_P
ClientSampled :
SLCS976



Tgt Ion: 93 Resp: 878073
Ion Ratio Lower Upper
93 100
95 32.6 25.6 38.4
123 11.3 8.2 12.4




Manual Integrations
APPROVED


mohammad
9/13/2021 10:22:06 AM



#29
2,4-Dichlorophenol
Concen: 36.423 ng/ul
RT: 10.381 min Scan# 1240
Delta R.T. -0.000 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02


Tgt Ion: 162 Resp: 646410
Ion Ratio Lower Upper
162 100
164 64.4 52.0 78.0
98 34.4 34.3 51.5





#30
Naphthalene
Concen: 34.440 ng/ul
RT: 10.787 min Scan# 1309
Delta R.T. -0.000 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Instrument :
BNA_P
ClientSampleId :
SLCS976




Tgt Ion:128 Resp: 2129088
Ion Ratio Lower Upper
128 100
129 10.9 8.8 13.2
127 13.2 10.6 15.8

Manual Integrations
APPROVED



mohammad
9/13/2021 10:22:06 AM




#32
4-Chloroaniline
Concen: 32.755 ng/ul
RT: 10.893 min Scan# 1327
Delta R.T. -0.000 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Tgt Ion:127 Resp: 826752
Ion Ratio Lower Upper
127 100
129 32.8 25.7 38.5





#33
Hexachlorobutadiene
Concen: 32.531 ng/ul
RT: 11.075 min Scan# 1358
Delta R.T. -0.006 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Instrument :
BNA_P
ClientSampled :
SLCS976




Tgt Ion:225 Resp: 398273
Ion Ratio Lower Upper
225 100
223 64.6 48.2 72.2
227 64.2 49.6 74.4

Manual Integrations
APPROVED



mohammad
9/13/2021 10:22:06 AM




#34
Caprolactam
Concen: 37.261 ng/ul m
RT: 11.663 min Scan# 1458
Delta R.T. -0.000 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Tgt Ion:113 Resp: 197106
Ion Ratio Lower Upper
113 100
55 167.0 125.8 188.6
56 121.9 98.2 147.2





#35
4-Chloro-3-methylphenol
Concen: 36.907 ng/ul
RT: 12.010 min Scan# 1517
Delta R.T. -0.000 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Instrument :
BNA_P
ClientSampleId :
SLCS976




Tgt Ion:107 Resp: 664337
Ion Ratio Lower Upper
107 100
144 27.0 18.8 28.2
142 81.9 58.4 87.6

Manual Integrations
APPROVED



mohammad
9/13/2021 10:22:06 AM




#36
2-Methylnaphthalene
Concen: 36.463 ng/ul
RT: 12.393 min Scan# 1582
Delta R.T. -0.006 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Tgt Ion:142 Resp: 1511298
Ion Ratio Lower Upper
142 100
141 87.3 72.3 108.5





#37
1-Methylnaphthalene
Concen: 34.109 ng/ul
RT: 12.610 min Scan# 1619
Delta R.T. -0.006 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Instrument :
BNA_P
ClientSampleId :
SLCS976




Tgt Ion:142 Resp: 1442170
Ion Ratio Lower Upper
142 100
141 91.9 74.9 112.3
116 3.6 3.3 4.9

Manual Integrations
APPROVED



mohammad
9/13/2021 10:22:06 AM




#39
1,2,4,5-Tetrachlorobenzene
Concen: 34.303 ng/ul
RT: 12.757 min Scan# 1644
Delta R.T. -0.006 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Tgt Ion:216 Resp: 787382
Ion Ratio Lower Upper
216 100
214 77.5 64.3 96.5
179 20.1 19.3 28.9
108 16.5 19.8 29.6#





#40
Hexachlorocyclopentadiene
Concen: 28.179 ng/ul
RT: 12.740 min Scan# 1641
Delta R.T. -0.006 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Instrument :
BNA_P
ClientSampled :
SLCS976




Tgt Ion:237 Resp: 417455
Ion Ratio Lower Upper
237 100
235 62.8 50.4 75.6
272 13.1 9.7 14.5

Manual Integrations
APPROVED



mohammad
9/13/2021 10:22:06 AM




#41
2,4,6-Trichlorophenol
Concen: 37.495 ng/ul
RT: 12.993 min Scan# 1684
Delta R.T. -0.006 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Tgt Ion:196 Resp: 498150
Ion Ratio Lower Upper
196 100
198 96.7 76.2 114.4
200 30.5 24.6 36.8







#42
2,4,5-Trichlorophenol
Concen: 37.129 ng/ul
RT: 13.063 min Scan# 1696
Delta R.T. -0.006 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Tgt Ion:196 Resp: 549903
Ion Ratio Lower Upper
196 100
198 95.4 79.3 118.9
200 31.2 25.8 38.8




Instrument :
BNA_P
ClientSampleId :
SLCS976

Manual Integrations
APPROVED



mohammad
9/13/2021 10:22:06 AM




#43
1,1'-Biphenyl
Concen: 34.506 ng/ul
RT: 13.398 min Scan# 1753
Delta R.T. -0.006 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02



Tgt Ion:154 Resp: 1925878
Ion Ratio Lower Upper
154 100
153 40.8 32.4 48.6
76 12.4 12.9 19.3#






#44
2-Chloronaphthalene
Concen: 34.285 ng/ul
RT: 13.440 min Scan# 1760
Delta R.T. -0.000 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Instrument :
BNA_P
ClientSampleId :
SLCS976

Manual Integrations
APPROVED

mohammad
9/13/2021 10:22:06 AM







Tgt Ion:162 Resp: 1475339
Ion Ratio Lower Upper
162 100
164 33.3 25.6 38.4
127 37.3 36.2 54.2



#45
2-Nitroaniline
Concen: 42.869 ng/ul
RT: 13.640 min Scan# 1794
Delta R.T. -0.000 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02


Tgt Ion: 65 Resp: 374934
Ion Ratio Lower Upper
65 100
92 73.9 53.7 80.5
138 119.2 71.4 107.2#





#47
Dimethylphthalate
Concen: 34.922 ng/ul
RT: 14.016 min Scan# 1858
Delta R.T. -0.006 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Instrument :
BNA_P
ClientSampled :
SLCS976




Tgt Ion:163 Resp: 1828604
Ion Ratio Lower Upper
163 100
194 4.3 3.0 4.4
164 10.4 7.8 11.8

Manual Integrations
APPROVED



mohammad
9/13/2021 10:22:06 AM




#48
2,6-Dinitrotoluene
Concen: 40.156 ng/ul
RT: 14.134 min Scan# 1878
Delta R.T. -0.006 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Tgt Ion:165 Resp: 351157
Ion Ratio Lower Upper
165 100
89 55.2 62.2 93.2#
121 21.5 22.1 33.1#





#50
Acenaphthylene
Concen: 32.493 ng/ul
RT: 14.281 min Scan# 1903
Delta R.T. -0.006 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Instrument :
BNA_P
ClientSampleId :
SLCS976




Tgt Ion:152 Resp: 2216881
Ion Ratio Lower Upper
152 100
151 20.1 15.9 23.9
153 13.0 10.4 15.6

Manual Integrations
APPROVED



mohammad
9/13/2021 10:22:06 AM




#51
3-Nitroaniline
Concen: 36.819 ng/ul
RT: 14.463 min Scan# 1934
Delta R.T. -0.000 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Tgt Ion:138 Resp: 369089
Ion Ratio Lower Upper
138 100
108 11.6 9.7 14.5
92 107.7 106.7 160.1





#52
Acenaphthene
Concen: 35.067 ng/ul
RT: 14.622 min Scan# 1961
Delta R.T. -0.006 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Instrument :
BNA_P
ClientSampleId :
SLCS976




Tgt Ion:153 Resp: 1584207
Ion Ratio Lower Upper
153 100
152 48.0 38.6 57.8
154 89.8 69.0 103.6

Manual Integrations
APPROVED



mohammad
9/13/2021 10:22:06 AM




#53
2,4-Dinitrophenol
Concen: 31.624 ng/ul
RT: 14.663 min Scan# 1968
Delta R.T. -0.006 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Tgt Ion:184 Resp: 155357
Ion Ratio Lower Upper
184 100
63 62.4 73.4 110.2#
154 60.3 59.4 89.0





#55
4-Nitrophenol
Concen: 37.884 ng/ul
RT: 14.763 min Scan# 1985
Delta R.T. 0.006 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Instrument :
BNA_P
ClientSampleId :
SLCS976




Tgt Ion:109 Resp: 251168
Ion Ratio Lower Upper
109 100
139 148.4 90.3 135.5#
65 129.9 102.2 153.2

Manual Integrations
APPROVED



mohammad
9/13/2021 10:22:06 AM




#56
Dibenzofuran
Concen: 34.334 ng/ul
RT: 14.957 min Scan# 2018
Delta R.T. -0.006 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02



Tgt Ion:168 Resp: 2238294
Ion Ratio Lower Upper
168 100
139 35.5 35.3 52.9









#57
2,4-Dinitrotoluene
Concen: 41.278 ng/ul
RT: 14.916 min Scan# 2011
Delta R.T. -0.006 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02

Tgt Ion:165 Resp: 515852
Ion Ratio Lower Upper
165 100
63 40.2 45.2 67.8#
182 2.8 1.9 2.9

Instrument :
BNA_P
ClientSampled :
SLCS976




Manual Integrations
APPROVED


mohammad
9/13/2021 10:22:06 AM



#58
2,3,4,6-Tetrachlorophenol
Concen: 36.570 ng/ul
RT: 15.181 min Scan# 2056
Delta R.T. -0.006 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02


Tgt Ion:232 Resp: 436424
Ion Ratio Lower Upper
232 100
131 42.5 51.8 77.6#
130 2.3 3.1 4.7#
166 30.5 31.0 46.4#





#59
Diethylphthalate
Concen: 35.438 ng/ul
RT: 15.381 min Scan# 2090
Delta R.T. -0.006 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Instrument :
BNA_P
ClientSampleId :
SLCS976



Tgt Ion:149 Resp: 1806733
Ion Ratio Lower Upper
149 100
177 22.9 15.5 23.3
150 12.6 9.9 14.9




Manual Integrations
APPROVED


mohammad
9/13/2021 10:22:06 AM



#61
Fluorene
Concen: 35.139 ng/ul
RT: 15.610 min Scan# 2129
Delta R.T. -0.000 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02


Tgt Ion:166 Resp: 1825795
Ion Ratio Lower Upper
166 100
165 98.0 79.6 119.4
167 13.5 10.7 16.1





#62
4-Chlorophenyl-phenylether
Concen: 34.333 ng/ul
RT: 15.604 min Scan# 2128
Delta R.T. -0.000 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Instrument :
BNA_P
ClientSampled :
SLCS976




Tgt Ion:204 Resp: 909426
Ion Ratio Lower Upper
204 100
206 32.9 26.3 39.5
141 53.0 62.0 93.0#

Manual Integrations
APPROVED



mohammad
9/13/2021 10:22:06 AM




#63
4-Nitroaniline
Concen: 40.950 ng/ul
RT: 15.622 min Scan# 2131
Delta R.T. -0.000 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Tgt Ion:138 Resp: 409837
Ion Ratio Lower Upper
138 100
92 50.5 47.1 70.7
108 75.3 73.8 110.8





#66
4,6-Dinitro-2-methylphenol
Concen: 37.827 ng/ul
RT: 15.681 min Scan# 2141
Delta R.T. -0.000 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Instrument :
BNA_P
ClientSampled :
SLCS976




Tgt Ion:198 Resp: 278942
Ion Ratio Lower Upper
198 100
182 4.3 4.5 6.7#
77 22.4 31.7 47.5#

Manual Integrations
APPROVED



mohammad
9/13/2021 10:22:06 AM




#67
N-Nitrosodiphenylamine
Concen: 36.349 ng/ul
RT: 15.816 min Scan# 2164
Delta R.T. -0.006 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Tgt Ion:169 Resp: 1567382
Ion Ratio Lower Upper
169 100
168 66.6 53.9 80.9
167 36.4 30.0 45.0





#68
4-Bromophenyl-phenylether
Concen: 34.943 ng/ul
RT: 16.498 min Scan# 2280
Delta R.T. -0.006 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Instrument :
BNA_P
ClientSampled :
SLCS976




Tgt Ion:248 Resp: 540866
Ion Ratio Lower Upper
248 100
250 98.5 78.4 117.6
141 68.3 80.6 121.0#

Manual Integrations
APPROVED



mohammad
9/13/2021 10:22:06 AM




#69
Hexachlorobenzene
Concen: 34.182 ng/ul
RT: 16.610 min Scan# 2299
Delta R.T. -0.012 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Tgt Ion:284 Resp: 613514
Ion Ratio Lower Upper
284 100
142 35.4 37.8 56.6#
249 31.4 27.1 40.7





#70
Atrazine
Concen: 33.798 ng/ul
RT: 16.769 min Scan# 2326
Delta R.T. -0.006 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Instrument :
BNA_P
ClientSampled :
SLCS976




Tgt Ion:200 Resp: 551906
Ion Ratio Lower Upper
200 100
173 26.3 22.2 33.4
215 49.9 35.4 53.0

Manual Integrations
APPROVED



mohammad
9/13/2021 10:22:06 AM




#71
Pentachlorophenol
Concen: 32.361 ng/ul
RT: 16.957 min Scan# 2358
Delta R.T. -0.006 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Tgt Ion:266 Resp: 332690
Ion Ratio Lower Upper
266 100
264 63.2 51.0 76.4
268 62.4 49.4 74.2





#72
Phenanthrene
Concen: 35.494 ng/ul
RT: 17.351 min Scan# 2425
Delta R.T. -0.006 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Instrument :
BNA_P
ClientSampled :
SLCS976




Tgt Ion:178 Resp: 2894867
Ion Ratio Lower Upper
178 100
179 15.5 12.0 18.0
176 19.3 15.6 23.4

Manual Integrations
APPROVED



mohammad
9/13/2021 10:22:06 AM




#74
Anthracene
Concen: 36.052 ng/ul
RT: 17.439 min Scan# 2440
Delta R.T. -0.012 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02



Tgt Ion:178 Resp: 2922601
Ion Ratio Lower Upper
178 100
179 15.5 12.1 18.1
176 18.7 14.9 22.3









#75
1,2,3,4-Tetrachlorobenzene
Concen: 35.103 ng/uL
RT: 13.363 min Scan# 1747
Delta R.T. -0.006 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02

Tgt Ion:216 Resp: 805508
Ion Ratio Lower Upper
216 100
214 77.4 62.9 94.3
218 48.3 37.8 56.8

Instrument :
BNA_P
ClientSampled :
SLCS976




Manual Integrations
APPROVED


mohammad
9/13/2021 10:22:06 AM



#76
Pentachlorobenzene
Concen: 32.919 ng/uL
RT: 14.881 min Scan# 2005
Delta R.T. -0.000 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02


Tgt Ion:250 Resp: 731217
Ion Ratio Lower Upper
250 100
248 63.7 51.0 76.6
252 63.4 51.0 76.4





#77
 Carbazole
 Concen: 36.715 ng/ul
 RT: 17.710 min Scan# 2486
 Delta R.T. -0.006 min
 Lab File: BP006946.D
 Acq: 10 Sep 2021 18:02

Instrument :
 BNA_P
 ClientSampleId :
 SLCS976






Tgt Ion:167 Resp: 2645239


Ion	Ratio	Lower	Upper
167	100		
166	21.7	17.6	26.4
139	12.6	11.9	17.9

Manual Integrations
 APPROVED

mohammad
 9/13/2021 10:22:06 AM






#78
 Di-n-butylphthalate
 Concen: 36.525 ng/ul
 RT: 18.263 min Scan# 2580
 Delta R.T. -0.006 min
 Lab File: BP006946.D
 Acq: 10 Sep 2021 18:02



Tgt Ion:149 Resp: 2982530


Ion	Ratio	Lower	Upper
149	100		
150	9.3	7.3	10.9
104	4.9	4.8	7.2





#80
Fluoranthene
Concen: 35.891 ng/ul
RT: 19.316 min Scan# 2759
Delta R.T. -0.006 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Instrument :
BNA_P
ClientSampleId :
SLCS976




Tgt Ion:202 Resp: 3392924
Ion Ratio Lower Upper
202 100
101 10.8 11.0 16.4#
100 8.1 8.3 12.5#

Manual Integrations
APPROVED



mohammad
9/13/2021 10:22:06 AM




#82
Pyrene
Concen: 36.274 ng/ul
RT: 19.669 min Scan# 2819
Delta R.T. -0.006 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Tgt Ion:202 Resp: 3517323
Ion Ratio Lower Upper
202 100
101 12.3 12.6 19.0#
100 10.2 10.6 15.8#





#83
Butylbenzylphthalate
Concen: 38.059 ng/ul
RT: 20.539 min Scan# 2967
Delta R.T. -0.006 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Instrument :
BNA_P
ClientSampled :
SLCS976




Tgt Ion:149 Resp: 1285025
Ion Ratio Lower Upper
149 100
91 68.0 63.0 94.6
206 21.4 14.1 21.1#

Manual Integrations
APPROVED



mohammad
9/13/2021 10:22:06 AM




#84
3,3'-Dichlorobenzidine
Concen: 32.384 ng/ul
RT: 21.304 min Scan# 3097
Delta R.T. -0.012 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Tgt Ion:252 Resp: 1090598
Ion Ratio Lower Upper
252 100
254 64.8 51.6 77.4
126 11.5 10.7 16.1





#85
Benzo(a)anthracene
Concen: 35.357 ng/ul
RT: 21.375 min Scan# 3109
Delta R.T. -0.006 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Instrument :
BNA_P
ClientSampled :
SLCS976




Tgt Ion:228 Resp: 3341061
Ion Ratio Lower Upper
228 100
229 19.7 15.8 23.8
226 27.5 21.6 32.4

Manual Integrations
APPROVED



mohammad
9/13/2021 10:22:06 AM




#86
Bis(2-ethylhexyl)phthalate
Concen: 38.851 ng/ul
RT: 21.304 min Scan# 3097
Delta R.T. -0.006 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Tgt Ion:149 Resp: 1955581
Ion Ratio Lower Upper
149 100
167 28.1 20.2 30.2
279 4.7 2.7 4.1#





#87
Chrysene
Concen: 35.014 ng/ul
RT: 21.427 min Scan# 3118
Delta R.T. -0.006 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Instrument :
BNA_P
ClientSampleId :
SLCS976



Tgt Ion:228 Resp: 3274122
Ion Ratio Lower Upper
228 100
226 30.2 24.5 36.7
229 19.9 15.9 23.9




Manual Integrations
APPROVED


mohammad
9/13/2021 10:22:06 AM



#89
Di-n-octyl phthalate
Concen: 40.390 ng/ul
RT: 22.233 min Scan# 3255
Delta R.T. -0.018 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02


Tgt Ion:149 Resp: 3311378





#90
Benzo(b)fluoranthene
Concen: 37.434 ng/ul
RT: 23.080 min Scan# 3399
Delta R.T. -0.006 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Instrument :
BNA_P
ClientSampled :
SLCS976



Tgt Ion:252 Resp: 3474619
Ion Ratio Lower Upper
252 100
253 22.3 17.7 26.5
125 10.3 10.1 15.1




Manual Integrations
APPROVED


mohammad
9/13/2021 10:22:06 AM



#91
Benzo(k)fluoranthene
Concen: 38.055 ng/ul
RT: 23.127 min Scan# 3407
Delta R.T. -0.012 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02


Tgt Ion:252 Resp: 3306379
Ion Ratio Lower Upper
252 100
253 22.4 17.2 25.8
125 10.4 9.4 14.0





#93
Benzo(a)pyrene
Concen: 34.411 ng/ul
RT: 23.716 min Scan# 3507
Delta R.T. -0.012 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Instrument :
BNA_P
ClientSampled :
SLCS976



Tgt Ion:252 Resp: 3073158
Ion Ratio Lower Upper
252 100
253 22.6 17.6 26.4
125 11.7 11.3 16.9




Manual Integrations
APPROVED


mohammad
9/13/2021 10:22:06 AM



#94
Indeno(1,2,3-cd)pyrene
Concen: 37.008 ng/ul
RT: 26.374 min Scan# 3959
Delta R.T. -0.018 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02


Tgt Ion:276 Resp: 3834042
Ion Ratio Lower Upper
276 100
138 25.2 21.8 32.8
277 25.2 19.9 29.9





#95
Dibenzo(a,h)anthracene
Concen: 36.178 ng/ul
RT: 26.392 min Scan# 3962
Delta R.T. -0.024 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02




Instrument :
BNA_P
ClientSampleId :
SLCS976




Tgt Ion:278 Resp: 3242830
Ion Ratio Lower Upper
278 100
139 16.6 15.4 23.2
279 23.8 19.0 28.4

Manual Integrations
APPROVED

mohammad
9/13/2021 10:22:06 AM



#96
Benzo(g,h,i)perylene
Concen: 35.696 ng/ul
RT: 27.157 min Scan# 4092
Delta R.T. -0.018 min
Lab File: BP006946.D
Acq: 10 Sep 2021 18:02



Tgt Ion:276 Resp: 3171667
Ion Ratio Lower Upper
276 100
138 21.6 20.7 31.1
277 23.2 19.0 28.4

