

Data Path : Z:\SVOASRV\HPCHEM1\BNA_P\DATA\BP091019\
 Data File : BP000196.D
 Acq On : 11 Sep 2019 12:16
 Operator : HP/JU
 Sample : K4817-01
 Misc : GCMS CONFIRMATION
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Sep 12 02:13:39 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SOM-EPA-BP090519MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Sep 10 02:43:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.89	152	296033	20.00	ng/ul	0.00
18) Naphthalene-d8	8.18	136	1131279	20.00	ng/ul	0.00
36) Acenaphthene-d10	9.95	164	622100	20.00	ng/ul	0.00
62) Phenanthrene-d10	11.45	188	1188053	20.00	ng/ul	0.00
78) Chrysene-d12	14.15	240	1004129	20.00	ng/ul	0.00
86) Perylene-d12	15.70	264	1036201	20.00	ng/ul	0.01

System Monitoring Compounds

3) 1,4-Dioxane-d8	2.62	96	92	0.01	ng/uL	0.01
5) Phenol-d5	6.51	99	14	0.00	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.58	67	336	0.03	ng/ul	0.00
9) 2-Chlorophenol-d4	6.66	132	34	0.00	ng/ul	0.00
13) 4-Methylphenol-d8	7.25	113	21	0.00	ng/ul	0.00
19) Nitrobenzene-d5	7.46	128	59	0.01	ng/ul	0.01
22) 2-Nitrophenol-d4	7.77	143	34	0.00	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	8.00	165	34	0.00	ng/ul	-0.02
29) 4-Chloroaniline-d4	8.23	131	180	0.01	ng/ul	0.00
44) Dimethylphthalate-d6	9.63	166	90	0.00	ng/ul	0.00
47) Acenaphthylene-d8	9.95	160	283814	5.00	ng/ul	0.15
52) 4-Nitrophenol-d4	10.04	143	14	0.00	ng/ul	0.00
58) Fluorene-d10	10.47	176	1297	0.03	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	10.52	200	26	0.00	ng/ul	0.00
71) Anthracene-d10	11.50	188	6450	0.11	ng/ul	0.00
79) Pyrene-d10	12.90	212	256	0.00	ng/ul	0.00
90) Benzo(a)pyrene-d12	15.70	264	1035690	19.28	ng/ul	0.11

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP091019\
Data File : BP000196.D
Acq On : 11 Sep 2019 12:16
Operator : HP/JU
Sample : K4817-01
Misc : GCMS CONFIRMATION
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :

Quant Time: Sep 12 02:13:39 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SOM-EPA-BP090519MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Sep 10 02:43:18 2019
Response via : Initial Calibration

