

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091023\
 Data File : BP017088.D
 Acq On : 11 Sep 2023 04:12
 Operator : MA/JU
 Sample : SST04046
 Misc :
 ALS Vial : 104 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST040606

Quant Time: Sep 11 05:46:49 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP091023.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Sep 11 05:32:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.963	152	251105	20.000	ng/u1	0.00
20) Naphthalene-d8	10.787	136	1126129	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.622	164	691328	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.386	188	1520804	20.000	ng/u1	0.00
79) Chrysene-d12	21.486	240	1477422	20.000	ng/u1	0.00
88) Perylene-d12	24.021	264	1587141	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.316	96	110876	16.702	ng/uL	0.00
4) Pyridine-d5	3.752	84	847137	44.660	ng/u1	0.00
7) Phenol-d5	7.110	99	975372	42.517	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.305	67	649583	45.268	ng/u1	0.00
11) 2-Chlorophenol-d4	7.481	132	730255	43.982	ng/u1	0.00
15) 4-Methylphenol-d8	8.675	113	775962	44.399	ng/u1	0.00
21) Nitrobenzene-d5	9.157	128	362047	41.440	ng/u1	0.00
24) 2-Nitrophenol-d4	9.875	143	385476	40.702	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.398	165	714178	40.846	ng/u1	0.00
31) 4-Chloroaniline-d4	10.951	131	1094172	43.375	ng/u1	0.00
46) Dimethylphthalate-d6	14.034	166	2194467	42.644	ng/u1	0.00
49) Acenaphthylene-d8	14.322	160	2560090	43.107	ng/u1	0.00
54) 4-Nitrophenol-d4	14.839	143	458760	45.638	ng/u1	0.00
60) Fluorene-d10	15.616	176	1861255	42.244	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.745	200	362796	39.641	ng/u1	0.00
73) Anthracene-d10	17.492	188	2911478	41.687	ng/u1	0.00
81) Pyrene-d10	19.727	212	3404029	37.827	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.857	264	3422572	42.507	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.352	88	120319	16.787	ng/uL	99
5) Pyridine	3.769	79	892589	44.203	ng/u1	96
6) Benzaldehyde	7.122	77	614386	59.089	ng/u1	97
8) Phenol	7.140	94	1032715	43.328	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.399	93	819074	43.093	ng/u1	99
12) 2-Chlorophenol	7.516	128	749412	42.368	ng/u1	99
13) 2-Methylphenol	8.404	108	742416	42.464	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.475	45	821117	30.683	ng/u1	99
16) Acetophenone	8.810	105	1259927	44.673	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.787	70	688183	45.678	ng/u1	99
18) 4-Methylphenol	8.740	108	818151	42.910	ng/u1	100
19) Hexachloroethane	9.028	117	309461	42.232	ng/u1	99
22) Nitrobenzene	9.204	77	1006128	43.130	ng/u1	99
23) Isophorone	9.722	82	1883871	42.985	ng/u1	99
25) 2-Nitrophenol	9.904	139	424503	40.712	ng/u1	98
26) 2,4-Dimethylphenol	9.946	107	866275	40.690	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.204	93	1120869	41.268	ng/u1	99
29) 2,4-Dichlorophenol	10.428	162	712624	39.993	ng/u1	98
30) Naphthalene	10.840	128	2471995	40.228	ng/u1	100
32) 4-Chloroaniline	10.975	127	1083765	41.282	ng/u1	99
33) Hexachlorobutadiene	11.069	225	380836	35.758	ng/u1	98
34) Caprolactam	11.804	113	162472	27.532	ng/u1	96
35) 4-Chloro-3-methylphenol	12.075	107	824257	42.212	ng/u1	99
36) 2-Methylnaphthalene	12.445	142	1687541	41.557	ng/u1	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091023\
 Data File : BP017088.D
 Acq On : 11 Sep 2023 04:12
 Operator : MA/JU
 Sample : SST04046
 Misc :
 ALS Vial : 104 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST040606

Quant Time: Sep 11 05:46:49 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP091023.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Sep 11 05:32:49 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.663	142	1705677	41.720	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.792	216	779698	38.279	ng/ul	99
40) Hexachlorocyclopentadiene	12.740	237	432537	32.774	ng/ul	99
41) 2,4,6-Trichlorophenol	13.045	196	550826	39.733	ng/ul	100
42) 2,4,5-Trichlorophenol	13.116	196	591055	39.184	ng/ul	98
43) 1,1'-Biphenyl	13.451	154	2211980	40.421	ng/ul	100
44) 2-Chloronaphthalene	13.504	162	1726778	40.190	ng/ul	99
45) 2-Nitroaniline	13.739	65	614751	48.044	ng/ul	97
47) Dimethylphthalate	14.086	163	2207878	40.890	ng/ul	100
48) 2,6-Dinitrotoluene	14.228	165	437838	41.754	ng/ul	98
50) Acenaphthylene	14.351	152	2933716	44.117	ng/ul	99
51) 3-Nitroaniline	14.569	138	467905	44.712	ng/ul	95
52) Acenaphthene	14.686	153	1921209	41.408	ng/ul	99
53) 2,4-Dinitrophenol	14.763	184	239162	36.785	ng/ul	99
55) 4-Nitrophenol	14.857	109	380594	47.307	ng/ul	98
56) Dibenzofuran	15.022	168	2622630	41.110	ng/ul	100
57) 2,4-Dinitrotoluene	15.010	165	673883	44.769	ng/ul	91
58) 2,3,4,6-Tetrachlorophenol	15.239	232	468006	38.483	ng/ul	94
59) Diethylphthalate	15.439	149	2280400	42.263	ng/ul	99
61) Fluorene	15.675	166	2225386	42.311	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.663	204	1031476	40.580	ng/ul	98
63) 4-Nitroaniline	15.739	138	466276	50.677	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.763	198	390188	40.693	ng/ul#	92
67) N-Nitrosodiphenylamine	15.892	169	1804642	39.354	ng/ul	99
68) 4-Bromophenyl-phenylether	16.569	248	552121	36.491	ng/ul	99
69) Hexachlorobenzene	16.645	284	611234	36.854	ng/ul	99
70) Atrazine	16.845	200	639342	40.721	ng/ul	98
71) Pentachlorophenol	17.016	266	402409	37.206	ng/ul	99
72) Phenanthrene	17.433	178	3452336	40.853	ng/ul	100
74) Anthracene	17.527	178	3553249	41.827	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.404	216	824070	38.040	ng/uL	99
76) Pentachlorobenzene	14.916	250	711467	34.095	ng/uL	100
77) Carbazole	17.810	167	3290066	43.098	ng/ul	100
78) Di-n-butylphthalate	18.322	149	3986025	42.620	ng/ul	99
80) Fluoranthene	19.398	202	4021758	35.792	ng/ul	99
82) Pyrene	19.757	202	4300234	36.860	ng/ul	99
83) Butylbenzylphthalate	20.616	149	1931798	40.040	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.410	252	1321516	40.304	ng/ul	99
85) Benzo(a)anthracene	21.474	228	4387100	42.163	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.351	149	2847706	42.724	ng/ul	99
87) Chrysene	21.527	228	4034687	41.453	ng/ul	98
89) Di-n-octyl phthalate	22.315	149	4966774	37.331	ng/ul	100
90) Benzo(b)fluoranthene	23.233	252	4217698	38.035	ng/ul	99
91) Benzo(k)fluoranthene	23.286	252	4315747	40.190	ng/ul	99
93) Benzo(a)pyrene	23.910	252	4033992	43.615	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.709	276	4839282	47.135	ng/ul	99
95) Dibenzo(a,h)anthracene	26.733	278	4017564	47.411	ng/ul	99
96) Benzo(g,h,i)perylene	27.550	276	3647300	45.080	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed






#2
1,4-Dioxane
Concen: 16.787 ng/uL
RT: 3.352 min Scan# 41
Delta R.T. -0.000 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12

Instrument :
BNA_P
ClientSampleId :
SSTD040606







Tgt Ion: 88 Resp: 120319
Ion Ratio Lower Upper
88 100
43 35.5 27.3 40.9
58 78.7 63.7 95.5



#5
Pyridine
Concen: 44.203 ng/uL
RT: 3.769 min Scan# 116
Delta R.T. -0.006 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12

Tgt Ion: 79 Resp: 892589
Ion Ratio Lower Upper
79 100
52 81.8 61.6 92.4
51 38.9 30.9 46.3









#6
Benzaldehyde
Concen: 59.089 ng/ul
RT: 7.122 min Scan# 61
Delta R.T. 0.006 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion: 77 Resp: 614386
Ion Ratio Lower Upper
77 100
105 91.4 70.3 105.5
106 87.8 69.0 103.6



#8
Phenol
Concen: 43.328 ng/ul
RT: 7.140 min Scan# 689
Delta R.T. -0.000 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12

Tgt Ion: 94 Resp: 1032715
Ion Ratio Lower Upper
94 100
65 30.6 25.4 38.0
66 38.8 30.6 46.0









#10
Bis(2-Chloroethyl)ether
Concen: 43.093 ng/ul
RT: 7.399 min Scan# 710
Delta R.T. 0.006 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion: 93 Resp: 819074
Ion Ratio Lower Upper
93 100
63 80.9 65.4 98.0
95 32.2 25.1 37.7



#12
2-Chlorophenol
Concen: 42.368 ng/ul
RT: 7.516 min Scan# 753
Delta R.T. -0.000 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12

Tgt Ion:128 Resp: 749412
Ion Ratio Lower Upper
128 100
64 51.8 42.2 63.4
130 31.4 25.3 37.9





#13

2-Methylphenol

Concen: 42.464 ng/ul

RT: 8.404 min Scan# 904

Delta R.T. 0.006 min

Lab File: BP017088.D

Acq: 11 Sep 2023 04:12

Instrument :

BNA_P

ClientSampleId :





SSTD040606

Tgt Ion:108 Resp: 742416

Ion Ratio Lower Upper

108 100

107 91.0 72.3 108.5



#14

2,2'-oxybis(1-Chloropropane)

Concen: 30.683 ng/ul

RT: 8.475 min Scan# 916

Delta R.T. -0.000 min

Lab File: BP017088.D

Acq: 11 Sep 2023 04:12




Tgt Ion: 45 Resp: 821117


Ion Ratio Lower Upper

45 100

77 13.2 11.0 16.4

79 9.7 7.8 11.6









#16
Acetophenone
Concen: 44.673 ng/ul
RT: 8.810 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion:105 Resp: 1259927
Ion Ratio Lower Upper
105 100
77 82.9 67.0 100.4
51 31.3 25.4 38.0



#17
N-Nitroso-di-n-propylamine
Concen: 45.678 ng/ul
RT: 8.787 min Scan# 969
Delta R.T. 0.006 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12

Tgt Ion: 70 Resp: 688183
Ion Ratio Lower Upper
70 100
42 64.1 50.9 76.3
101 9.0 6.7 10.1
130 18.6 14.8 22.2









#18
4-Methylphenol
Concen: 42.910 ng/ul
RT: 8.740 min Scan# 960
Delta R.T. 0.006 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion:108 Resp: 818151
Ion Ratio Lower Upper
108 100
107 112.8 90.2 135.4



#19
Hexachloroethane
Concen: 42.232 ng/ul
RT: 9.028 min Scan# 1010
Delta R.T. -0.000 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12

Tgt Ion:117 Resp: 309461
Ion Ratio Lower Upper
117 100
201 88.1 69.9 104.9
199 53.5 43.9 65.9









#22
Nitrobenzene
Concen: 43.130 ng/ul
RT: 9.204 min Scan# 1040
Delta R.T. -0.000 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion: 77 Resp: 1006128
Ion Ratio Lower Upper
77 100
123 39.9 32.8 49.2
65 13.8 11.0 16.4



#23
Isophorone
Concen: 42.985 ng/ul
RT: 9.722 min Scan# 1128
Delta R.T. 0.006 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12

Tgt Ion: 82 Resp: 1883871
Ion Ratio Lower Upper
82 100
95 6.6 5.2 7.8
138 16.3 12.6 19.0









#25
2-Nitrophenol
Concen: 40.712 ng/ul
RT: 9.904 min Scan# 1159
Delta R.T. -0.000 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion:139 Resp: 424503
Ion Ratio Lower Upper
139 100
65 60.9 49.9 74.9
109 32.0 25.3 37.9



#26
2,4-Dimethylphenol
Concen: 40.690 ng/ul
RT: 9.946 min Scan# 1166
Delta R.T. -0.000 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12

Tgt Ion:107 Resp: 866275
Ion Ratio Lower Upper
107 100
121 49.3 39.1 58.7
122 82.4 66.6 100.0









#27
Bis(2-Chloroethoxy)methane
Concen: 41.268 ng/ul
RT: 10.204 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion: 93 Resp: 1120869
Ion Ratio Lower Upper
93 100
95 32.6 26.6 39.8
123 11.0 8.8 13.2



#29
2,4-Dichlorophenol
Concen: 39.993 ng/ul
RT: 10.428 min Scan# 1248
Delta R.T. 0.006 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12

Tgt Ion:162 Resp: 712624
Ion Ratio Lower Upper
162 100
164 65.5 51.0 76.6
98 39.0 30.7 46.1









#30
Naphthalene
Concen: 40.228 ng/ul
RT: 10.840 min Scan# 1318
Delta R.T. -0.000 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion:128 Resp: 2471995
Ion Ratio Lower Upper
128 100
129 11.1 8.9 13.3
127 13.0 10.4 15.6



#32
4-Chloroaniline
Concen: 41.282 ng/ul
RT: 10.975 min Scan# 1341
Delta R.T. -0.000 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12

Tgt Ion:127 Resp: 1083765
Ion Ratio Lower Upper
127 100
129 31.8 25.9 38.9









#33
Hexachlorobutadiene
Concen: 35.758 ng/ul
RT: 11.069 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion:225 Resp: 380836
Ion Ratio Lower Upper
225 100
223 62.5 51.6 77.4
227 64.4 51.8 77.8



#34
Caprolactam
Concen: 27.532 ng/ul
RT: 11.804 min Scan# 1482
Delta R.T. 0.012 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12

Tgt Ion:113 Resp: 162472
Ion Ratio Lower Upper
113 100
55 194.4 160.4 240.6
56 142.8 118.5 177.7









#35
4-Chloro-3-methylphenol
Concen: 42.212 ng/ul
RT: 12.075 min Scan# 1528
Delta R.T. -0.000 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion:107 Resp: 824257
Ion Ratio Lower Upper
107 100
144 25.7 20.9 31.3
142 78.4 63.2 94.8



#36
2-Methylnaphthalene
Concen: 41.557 ng/ul
RT: 12.445 min Scan# 1591
Delta R.T. 0.006 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12

Tgt Ion:142 Resp: 1687541
Ion Ratio Lower Upper
142 100
141 90.2 71.4 107.0





#37

1-Methylnaphthalene

Concen: 41.720 ng/ul

RT: 12.663 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP017088.D

Acq: 11 Sep 2023 04:12

Instrument :

BNA_P

ClientSampleId :

SSTD040606





Tgt Ion:142 Resp: 1705677

Ion Ratio Lower Upper

142 100

141 92.4 73.1 109.7

116 4.0 3.2 4.8



#39

1,2,4,5-Tetrachlorobenzene

Concen: 38.279 ng/ul

RT: 12.792 min Scan# 1650

Delta R.T. -0.000 min

Lab File: BP017088.D

Acq: 11 Sep 2023 04:12

Tgt Ion:216 Resp: 779698




Ion Ratio Lower Upper


216 100

214 79.2 62.2 93.2

179 22.9 18.2 27.4

108 20.8 17.0 25.6









#40
Hexachlorocyclopentadiene
Concen: 32.774 ng/ul
RT: 12.740 min Scan# 1641
Delta R.T. -0.000 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion:237 Resp: 432537
Ion Ratio Lower Upper
237 100
235 64.0 50.9 76.3
272 12.3 8.8 13.2



#41
2,4,6-Trichlorophenol
Concen: 39.733 ng/ul
RT: 13.045 min Scan# 1693
Delta R.T. -0.000 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12

Tgt Ion:196 Resp: 550826
Ion Ratio Lower Upper
196 100
198 95.9 76.7 115.1
200 30.5 24.4 36.6





#42

2,4,5-Trichlorophenol

Concen: 39.184 ng/ul

RT: 13.116 min Scan# 1705

Delta R.T. -0.000 min

Lab File: BP017088.D

Acq: 11 Sep 2023 04:12

Instrument :

BNA_P

ClientSampleId :

SSTD040606





Tgt Ion:196 Resp: 591055

Ion Ratio Lower Upper

196 100

198 93.4 76.6 115.0

200 29.8 23.8 35.8



#43

1,1'-Biphenyl

Concen: 40.421 ng/ul

RT: 13.451 min Scan# 1762

Delta R.T. -0.000 min

Lab File: BP017088.D

Acq: 11 Sep 2023 04:12




Tgt Ion:154 Resp: 2211980


Ion Ratio Lower Upper

154 100

153 40.9 32.6 49.0

76 15.5 12.2 18.4





#44

2-Chloronaphthalene

Concen: 40.190 ng/ul

RT: 13.504 min Scan# 11

Delta R.T. 0.006 min

Lab File: BP017088.D

Acq: 11 Sep 2023 04:12

Instrument :

BNA_P

ClientSampleId :

SSTD040606





Tgt Ion:162 Resp: 1726778

Ion Ratio Lower Upper

162 100

164 33.0 26.0 39.0

127 39.5 31.8 47.8



#45

2-Nitroaniline

Concen: 48.044 ng/ul

RT: 13.739 min Scan# 1811

Delta R.T. 0.006 min

Lab File: BP017088.D

Acq: 11 Sep 2023 04:12




Tgt Ion: 65 Resp: 614751


Ion Ratio Lower Upper

65 100

92 64.4 52.5 78.7

138 94.3 71.8 107.6









#47
Dimethylphthalate
Concen: 40.890 ng/ul
RT: 14.086 min Scan# 1869
Delta R.T. 0.006 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion:163 Resp: 2207878
Ion Ratio Lower Upper
163 100
194 3.8 2.9 4.3
164 10.0 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 41.754 ng/ul
RT: 14.228 min Scan# 1894
Delta R.T. -0.000 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12

Tgt Ion:165 Resp: 437838
Ion Ratio Lower Upper
165 100
89 68.2 55.6 83.4
121 22.9 20.0 30.0









#50
Acenaphthylene
Concen: 44.117 ng/ul
RT: 14.351 min Scan# 1915
Delta R.T. -0.000 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion:152 Resp: 2933716
Ion Ratio Lower Upper
152 100
151 20.2 15.5 23.3
153 13.2 10.2 15.4



#51
3-Nitroaniline
Concen: 44.712 ng/ul
RT: 14.569 min Scan# 1952
Delta R.T. 0.006 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12

Tgt Ion:138 Resp: 467905
Ion Ratio Lower Upper
138 100
108 10.4 8.5 12.7
92 123.3 103.3 154.9









#52
Acenaphthene
Concen: 41.408 ng/ul
RT: 14.686 min Scan# 1972
Delta R.T. -0.000 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion:153 Resp: 1921209
Ion Ratio Lower Upper
153 100
152 48.2 38.9 58.3
154 87.5 69.0 103.6



#53
2,4-Dinitrophenol
Concen: 36.785 ng/ul
RT: 14.763 min Scan# 1985
Delta R.T. -0.000 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12

Tgt Ion:184 Resp: 239162
Ion Ratio Lower Upper
184 100
63 91.6 74.5 111.7
154 66.8 53.8 80.8









#55
4-Nitrophenol
Concen: 47.307 ng/ul
RT: 14.857 min Scan# 2000
Delta R.T. 0.006 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion:109 Resp: 380594
Ion Ratio Lower Upper
109 100
139 127.6 98.6 147.8
65 145.3 117.0 175.4



#56
Dibenzofuran
Concen: 41.110 ng/ul
RT: 15.022 min Scan# 2029
Delta R.T. -0.000 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12

Tgt Ion:168 Resp: 2622630
Ion Ratio Lower Upper
168 100
139 38.4 30.6 45.8









#57
2,4-Dinitrotoluene
Concen: 44.769 ng/ul
RT: 15.010 min Scan# 2027
Delta R.T. -0.000 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion:165 Resp: 673883
Ion Ratio Lower Upper
165 100
63 64.3 57.7 86.5
182 2.5 2.2 3.4



#58
2,3,4,6-Tetrachlorophenol
Concen: 38.483 ng/ul
RT: 15.239 min Scan# 2066
Delta R.T. 0.006 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12

Tgt Ion:232 Resp: 468006
Ion Ratio Lower Upper
232 100
131 53.0 46.9 70.3
130 2.7 2.4 3.6
166 35.3 29.6 44.4









#59
Diethylphthalate
Concen: 42.263 ng/ul
RT: 15.439 min Scan# 2100
Delta R.T. -0.000 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion:149 Resp: 2280400
Ion Ratio Lower Upper
149 100
177 21.1 17.4 26.0
150 12.3 9.5 14.3



#61
Fluorene
Concen: 42.311 ng/ul
RT: 15.675 min Scan# 2140
Delta R.T. 0.006 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12

Tgt Ion:166 Resp: 2225386
Ion Ratio Lower Upper
166 100
165 98.3 78.8 118.2
167 13.5 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 40.580 ng/ul
RT: 15.663 min Scan# 2138
Delta R.T. -0.000 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion:204 Resp: 1031476
Ion Ratio Lower Upper
204 100
206 32.1 26.6 40.0
141 61.3 50.6 76.0



#63
4-Nitroaniline
Concen: 50.677 ng/ul
RT: 15.739 min Scan# 2151
Delta R.T. 0.006 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12

Tgt Ion:138 Resp: 466276
Ion Ratio Lower Upper
138 100
92 54.6 45.0 67.6
108 99.0 77.2 115.8









#66
4,6-Dinitro-2-methylphenol
Concen: 40.693 ng/ul
RT: 15.763 min Scan# 2155
Delta R.T. 0.006 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion:198 Resp: 390188
Ion Ratio Lower Upper
198 100
182 4.2 4.7 7.1#
77 31.7 29.6 44.4



#67
N-Nitrosodiphenylamine
Concen: 39.354 ng/ul
RT: 15.892 min Scan# 2177
Delta R.T. 0.006 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12

Tgt Ion:169 Resp: 1804642
Ion Ratio Lower Upper
169 100
168 67.5 54.5 81.7
167 37.8 30.4 45.6









#68
4-Bromophenyl-phenylether
Concen: 36.491 ng/ul
RT: 16.569 min Scan# 2292
Delta R.T. -0.000 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion:248 Resp: 552121
Ion Ratio Lower Upper
248 100
250 97.3 78.9 118.3
141 83.4 67.9 101.9



#69
Hexachlorobenzene
Concen: 36.854 ng/ul
RT: 16.645 min Scan# 2305
Delta R.T. -0.000 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12

Tgt Ion:284 Resp: 611234
Ion Ratio Lower Upper
284 100
142 48.9 38.9 58.3
249 33.8 26.1 39.1





#70

Atrazine

Concen: 40.721 ng/ul

RT: 16.845 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP017088.D

Acq: 11 Sep 2023 04:12

Instrument :

BNA_P

ClientSampleId :

SSTD040606





Tgt Ion:200 Resp: 639342

Ion Ratio Lower Upper

200 100

173 27.2 21.0 31.4

215 46.1 35.9 53.9



#71

Pentachlorophenol

Concen: 37.206 ng/ul

RT: 17.016 min Scan# 2368

Delta R.T. 0.006 min

Lab File: BP017088.D

Acq: 11 Sep 2023 04:12




Tgt Ion:266 Resp: 402409


Ion Ratio Lower Upper

266 100

264 62.9 49.8 74.8

268 64.3 50.9 76.3





#72

Phenanthrene

Concen: 40.853 ng/ul

RT: 17.433 min Scan# 2439

Delta R.T. -0.000 min

Lab File: BP017088.D

Acq: 11 Sep 2023 04:12

Instrument :

BNA_P

ClientSampleId :

SSTD040606





Tgt Ion:178 Resp: 3452336

Ion Ratio Lower Upper

178 100

179 15.5 12.3 18.5

176 19.8 15.8 23.8



#74

Anthracene

Concen: 41.827 ng/ul

RT: 17.527 min Scan# 2455

Delta R.T. -0.000 min

Lab File: BP017088.D

Acq: 11 Sep 2023 04:12




Tgt Ion:178 Resp: 3553249


Ion Ratio Lower Upper

178 100

179 15.7 12.2 18.2

176 19.5 15.4 23.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 38.040 ng/uL
RT: 13.404 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion:216 Resp: 824070
Ion Ratio Lower Upper
216 100
214 78.6 63.8 95.6
218 47.4 38.1 57.1



#76
Pentachlorobenzene
Concen: 34.095 ng/uL
RT: 14.916 min Scan# 2011
Delta R.T. 0.006 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12

Tgt Ion:250 Resp: 711467
Ion Ratio Lower Upper
250 100
248 62.3 49.9 74.9
252 63.7 51.4 77.0









#77
 Carbazole
 Concen: 43.098 ng/ul
 RT: 17.810 min Scan# 2503
 Delta R.T. -0.000 min
 Lab File: BP017088.D
 Acq: 11 Sep 2023 04:12




Instrument :
 BNA_P
 ClientSampleId :
 SSTD040606


Tgt Ion:167 Resp: 3290066
 Ion Ratio Lower Upper
 167 100
 166 22.0 17.7 26.5
 139 13.6 10.8 16.2



#78
 Di-n-butylphthalate
 Concen: 42.620 ng/ul
 RT: 18.322 min Scan# 2590
 Delta R.T. -0.000 min
 Lab File: BP017088.D
 Acq: 11 Sep 2023 04:12

Tgt Ion:149 Resp: 3986025
 Ion Ratio Lower Upper
 149 100
 150 9.4 7.3 10.9
 104 5.8 4.7 7.1









#80
Fluoranthene
Concen: 35.792 ng/ul
RT: 19.398 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion:202 Resp: 4021758
Ion Ratio Lower Upper
202 100
101 13.1 9.9 14.9
100 10.1 7.8 11.8



#82
Pyrene
Concen: 36.860 ng/ul
RT: 19.757 min Scan# 2834
Delta R.T. -0.000 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12

Tgt Ion:202 Resp: 4300234
Ion Ratio Lower Upper
202 100
101 15.3 11.9 17.9
100 12.5 9.8 14.6









#83
Butylbenzylphthalate
Concen: 40.040 ng/ul
RT: 20.616 min Scan# 2980
Delta R.T. -0.000 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion:149 Resp: 1931798
Ion Ratio Lower Upper
149 100
91 76.4 64.6 96.8
206 18.9 14.6 22.0



#84
3,3'-Dichlorobenzidine
Concen: 40.304 ng/ul
RT: 21.410 min Scan# 3115
Delta R.T. -0.000 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12

Tgt Ion:252 Resp: 1321516
Ion Ratio Lower Upper
252 100
254 65.2 51.7 77.5
126 14.0 11.5 17.3





#85

Benzo(a)anthracene

Concen: 42.163 ng/ul

RT: 21.474 min Scan# 3125

Delta R.T. 0.006 min

Lab File: BP017088.D

Acq: 11 Sep 2023 04:12

Instrument :

BNA_P

ClientSampleId :

SSTD040606





Tgt Ion:228 Resp: 4387100

Ion Ratio Lower Upper

228 100

229 20.0 16.0 24.0

226 29.1 22.9 34.3



#86

Bis(2-ethylhexyl)phthalate

Concen: 42.724 ng/ul

RT: 21.351 min Scan# 3105

Delta R.T. -0.000 min

Lab File: BP017088.D

Acq: 11 Sep 2023 04:12




Tgt Ion:149 Resp: 2847706


Ion Ratio Lower Upper

149 100

167 26.6 20.8 31.2

279 3.5 2.6 3.8









#87
Chrysene
Concen: 41.453 ng/ul
RT: 21.527 min Scan# 3135
Delta R.T. -0.000 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion:228 Resp: 4034687
Ion Ratio Lower Upper
228 100
226 31.7 24.6 36.8
229 20.2 15.6 23.4



#89
Di-n-octyl phthalate
Concen: 37.331 ng/ul
RT: 22.315 min Scan# 3269
Delta R.T. -0.000 min
Lab File: BP017088.D
Acq: 11 Sep 2023 04:12

Tgt Ion:149 Resp: 4966774





#90

Benzo(b)fluoranthene

Concen: 38.035 ng/ul

RT: 23.233 min Scan# 3424

Delta R.T. 0.006 min

Lab File: BP017088.D

Acq: 11 Sep 2023 04:12

Instrument :

BNA_P

ClientSampleId :

SSTD040606





Tgt Ion:252 Resp: 4217698

Ion Ratio Lower Upper

252 100

253 22.6 17.9 26.9

125 12.3 10.3 15.5



#91

Benzo(k)fluoranthene

Concen: 40.190 ng/ul

RT: 23.286 min Scan# 3434

Delta R.T. 0.006 min

Lab File: BP017088.D

Acq: 11 Sep 2023 04:12




Tgt Ion:252 Resp: 4315747


Ion Ratio Lower Upper

252 100

253 22.5 17.7 26.5

125 12.4 10.0 15.0





#93

Benzo(a)pyrene

Concen: 43.615 ng/ul

RT: 23.910 min Scan# 3

Delta R.T. 0.006 min

Lab File: BP017088.D

Acq: 11 Sep 2023 04:12

Instrument :

BNA_P

ClientSampleId :

SSTD040606





Tgt Ion:252 Resp: 4033992

Ion Ratio Lower Upper

252 100

253 22.0 17.6 26.4

125 13.5 11.3 16.9



#94

Indeno(1,2,3-cd)pyrene

Concen: 47.135 ng/ul

RT: 26.709 min Scan# 4016

Delta R.T. 0.012 min

Lab File: BP017088.D

Acq: 11 Sep 2023 04:12




Tgt Ion:276 Resp: 4839282


Ion Ratio Lower Upper

276 100

138 30.0 23.4 35.2

277 25.4 20.1 30.1





#95

Dibenzo(a,h)anthracene

Concen: 47.411 ng/ul

RT: 26.733 min Scan# 4018

Delta R.T. 0.012 min

Lab File: BP017088.D

Acq: 11 Sep 2023 04:12

Instrument :

BNA_P

ClientSampleId :

SSTD040606





Tgt Ion:278 Resp: 4017564

Ion Ratio Lower Upper

278 100

139 20.5 16.5 24.7

279 24.4 19.2 28.8



#96

Benzo(g,h,i)perylene

Concen: 45.080 ng/ul

RT: 27.550 min Scan# 4159

Delta R.T. 0.012 min

Lab File: BP017088.D

Acq: 11 Sep 2023 04:12

Tgt Ion:276 Resp: 3647300

Ion Ratio Lower Upper

276 100

138 27.3 22.0 33.0

277 23.8 18.7 28.1

