

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091023\
 Data File : BP017109.D
 Acq On : 11 Sep 2023 18:13
 Operator : MA/JU
 Sample : PB155340BS
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS340

Quant Time: Sep 12 02:30:18 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP091023.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Sep 11 06:13:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.958	152	354405	20.000	ng/u1	0.00
20) Naphthalene-d8	10.787	136	1596031	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.616	164	974084	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.387	188	2038882	20.000	ng/u1	0.00
79) Chrysene-d12	21.492	240	1526640	20.000	ng/u1	0.00
88) Perylene-d12	24.021	264	1155291	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.317	96	63013	6.786	ng/uL	0.00
4) Pyridine-d5	3.746	84	924234	33.008	ng/u1	0.00
7) Phenol-d5	7.111	99	1213693	37.674	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.299	67	768569	35.571	ng/u1	0.00
11) 2-Chlorophenol-d4	7.481	132	864339	37.122	ng/u1	0.00
15) 4-Methylphenol-d8	8.675	113	910225	35.717	ng/u1	0.00
21) Nitrobenzene-d5	9.158	128	421747	37.116	ng/u1	0.00
24) 2-Nitrophenol-d4	9.869	143	463650	39.535	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.399	165	835933	37.591	ng/u1	0.00
31) 4-Chloroaniline-d4	10.946	131	1152348	32.316	ng/u1	0.00
46) Dimethylphthalate-d6	14.034	166	2500691	36.079	ng/u1	0.00
49) Acenaphthylene-d8	14.316	160	2894839	36.679	ng/u1	0.00
54) 4-Nitrophenol-d4	14.840	143	502767	35.160	ng/u1	0.00
60) Fluorene-d10	15.616	176	2120411	36.051	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.745	200	420933	36.584	ng/u1	0.00
73) Anthracene-d10	17.492	188	3197273	36.585	ng/u1	0.00
81) Pyrene-d10	19.728	212	3460455	43.043	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.851	264	2125754	38.329	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.352	88	140446	13.935	ng/uL	99
5) Pyridine	3.770	79	964065	32.779	ng/u1	99
6) Benzaldehyde	7.116	77	653236	34.464	ng/u1	99
8) Phenol	7.140	94	1275552	37.258	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.393	93	986649	36.137	ng/u1	99
12) 2-Chlorophenol	7.511	128	915527	37.946	ng/u1	99
13) 2-Methylphenol	8.399	108	895994	36.494	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.469	45	833422	19.489	ng/u1	100
16) Acetophenone	8.811	105	1506464	36.803	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.787	70	834048	38.084	ng/u1	99
18) 4-Methylphenol	8.740	108	998440	37.205	ng/u1	98
19) Hexachloroethane	9.022	117	374193	36.805	ng/u1	97
22) Nitrobenzene	9.205	77	1215254	37.311	ng/u1	99
23) Isophorone	9.716	82	2246051	37.964	ng/u1	99
25) 2-Nitrophenol	9.905	139	516661	39.213	ng/u1	97
26) 2,4-Dimethylphenol	9.946	107	926555	32.922	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.205	93	1352407	36.454	ng/u1	98
29) 2,4-Dichlorophenol	10.422	162	854207	37.396	ng/u1	97
30) Naphthalene	10.834	128	2963186	36.036	ng/u1	99
32) 4-Chloroaniline	10.969	127	1008961	28.492	ng/u1	99
33) Hexachlorobutadiene	11.063	225	442226	34.763	ng/u1	100
34) Caprolactam	11.804	113	204132	25.526	ng/u1	99
35) 4-Chloro-3-methylphenol	12.075	107	1001582	39.007	ng/u1	98
36) 2-Methylnaphthalene	12.440	142	1986301	36.177	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091023\
 Data File : BP017109.D
 Acq On : 11 Sep 2023 18:13
 Operator : MA/JU
 Sample : PB155340BS
 Misc :
 ALS Vial : 31 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SLCS340


Quant Time: Sep 12 02:30:18 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP091023.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Sep 11 06:13:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.663	142	1940956	34.769	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.793	216	884499	35.049	ng/ul	99
40) Hexachlorocyclopentadiene	12.740	237	442291	29.581	ng/ul	99
41) 2,4,6-Trichlorophenol	13.046	196	612742	36.453	ng/ul	99
42) 2,4,5-Trichlorophenol	13.116	196	688634	37.638	ng/ul	99
43) 1,1'-Biphenyl	13.451	154	2664501	36.985	ng/ul	100
44) 2-Chloronaphthalene	13.499	162	2062738	36.518	ng/ul	99
45) 2-Nitroaniline	13.734	65	747086	41.739	ng/ul	99
47) Dimethylphthalate	14.081	163	2634221	37.171	ng/ul	99
48) 2,6-Dinitrotoluene	14.228	165	532844	41.803	ng/ul	97
50) Acenaphthylene	14.346	152	3298817	35.463	ng/ul	98
51) 3-Nitroaniline	14.569	138	574854	39.649	ng/ul	95
52) Acenaphthene	14.687	153	2311051	36.927	ng/ul	99
53) 2,4-Dinitrophenol	14.763	184	300354	37.377	ng/ul	98
55) 4-Nitrophenol	14.857	109	439459	37.272	ng/ul	97
56) Dibenzofuran	15.022	168	3125326	36.764	ng/ul	99
57) 2,4-Dinitrotoluene	15.010	165	799945	40.938	ng/ul	91
58) 2,3,4,6-Tetrachlorophenol	15.234	232	504875	35.398	ng/ul	99
59) Diethylphthalate	15.440	149	2704730	37.820	ng/ul	99
61) Fluorene	15.675	166	2654912	37.418	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.663	204	1206310	36.734	ng/ul	100
63) 4-Nitroaniline	15.740	138	600200	42.901	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.763	198	444775	35.505	ng/ul#	90
67) N-Nitrosodiphenylamine	15.887	169	2142760	38.210	ng/ul	99
68) 4-Bromophenyl-phenylether	16.569	248	648668	38.044	ng/ul	97
69) Hexachlorobenzene	16.645	284	697815	36.780	ng/ul	98
70) Atrazine	16.845	200	335912	17.313	ng/ul	98
71) Pentachlorophenol	17.010	266	385622	29.896	ng/ul	99
72) Phenanthrene	17.434	178	4011624	37.550	ng/ul	100
74) Anthracene	17.528	178	3998569	37.020	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.404	216	85398	3.296	ng/uL	99
76) Pentachlorobenzene	14.910	250	873854	39.143	ng/uL	99
77) Carbazole	17.810	167	3754104	38.133	ng/ul	100
78) Di-n-butylphthalate	18.322	149	4541262	40.008	ng/ul	99
80) Fluoranthene	19.398	202	4392779	45.294	ng/ul	98
82) Pyrene	19.757	202	4569408	44.113	ng/ul	97
83) Butylbenzylphthalate	20.616	149	1908254	44.775	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.410	252	1119529	35.909	ng/ul	100
85) Benzo(a)anthracene	21.475	228	3849833	37.617	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.351	149	2528284	41.452	ng/ul	99
87) Chrysene	21.527	228	3516407	37.466	ng/ul	98
89) Di-n-octyl phthalate	22.316	149	3725782	43.246	ng/ul	100
90) Benzo(b)fluoranthene	23.233	252	2915090	41.322	ng/ul	99
91) Benzo(k)fluoranthene	23.286	252	2903502	40.933	ng/ul	98
93) Benzo(a)pyrene	23.910	252	2494124	38.032	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.704	276	2910159	39.507	ng/ul	100
95) Dibenzo(a,h)anthracene	26.727	278	2408144	39.210	ng/ul	100
96) Benzo(g,h,i)perylene	27.539	276	2314082	41.218	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Sep 12 02:30:18 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP091023.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Sep 11 06:13:29 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 13.935 ng/uL
RT: 3.352 min Scan# 41
Delta R.T. -0.000 min
Lab File: BP017109.D
Acq: 11 Sep 2023 18:13




Instrument :
BNA_P
ClientSampleId :
SLCS340


Tgt Ion: 88 Resp: 140446
Ion Ratio Lower Upper
88 100
43 34.4 27.3 40.9
58 78.7 63.7 95.5



#5
Pyridine
Concen: 32.779 ng/uL
RT: 3.770 min Scan# 116
Delta R.T. -0.006 min
Lab File: BP017109.D
Acq: 11 Sep 2023 18:13

Tgt Ion: 79 Resp: 964065
Ion Ratio Lower Upper
79 100
52 77.2 61.6 92.4
51 38.0 30.9 46.3









#6
Benzaldehyde
Concen: 34.464 ng/ul
RT: 7.116 min Scan# 685
Delta R.T. 0.000 min
Lab File: BP017109.D
Acq: 11 Sep 2023 18:13




Instrument :
BNA_P
ClientSampleId :
SLCS340


Tgt Ion: 77 Resp: 653236
Ion Ratio Lower Upper
77 100
105 89.5 70.3 105.5
106 86.7 69.0 103.6



#8
Phenol
Concen: 37.258 ng/ul
RT: 7.140 min Scan# 689
Delta R.T. 0.000 min
Lab File: BP017109.D
Acq: 11 Sep 2023 18:13

Tgt Ion: 94 Resp: 1275552
Ion Ratio Lower Upper
94 100
65 30.6 25.4 38.0
66 37.8 30.6 46.0









#10
Bis(2-Chloroethyl)ether
Concen: 36.137 ng/ul
RT: 7.393 min Scan# 732
Delta R.T. 0.000 min
Lab File: BP017109.D
Acq: 11 Sep 2023 18:13




Instrument :
BNA_P
ClientSampleId :
SLCS340


Tgt Ion: 93 Resp: 986649
Ion Ratio Lower Upper
93 100
63 82.9 65.4 98.0
95 32.0 25.1 37.7



#12
2-Chlorophenol
Concen: 37.946 ng/ul
RT: 7.511 min Scan# 752
Delta R.T. -0.006 min
Lab File: BP017109.D
Acq: 11 Sep 2023 18:13

Tgt Ion:128 Resp: 915527
Ion Ratio Lower Upper
128 100
64 53.8 42.2 63.4
130 32.2 25.3 37.9





#13

2-Methylphenol

Concen: 36.494 ng/ul

RT: 8.399 min Scan# 903

Delta R.T. 0.000 min

Lab File: BP017109.D

Acq: 11 Sep 2023 18:13

Instrument :

BNA_P

ClientSampleId :





SLCS340

Tgt Ion:108 Resp: 895994

Ion Ratio Lower Upper

108 100

107 88.9 72.3 108.5



#14

2,2'-oxybis(1-Chloropropane)

Concen: 19.489 ng/ul

RT: 8.469 min Scan# 915

Delta R.T. -0.006 min

Lab File: BP017109.D

Acq: 11 Sep 2023 18:13




Tgt Ion: 45 Resp: 833422


Ion Ratio Lower Upper

45 100

77 13.7 11.0 16.4

79 9.6 7.8 11.6









#16
Acetophenone
Concen: 36.803 ng/ul
RT: 8.811 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP017109.D
Acq: 11 Sep 2023 18:13




Instrument :
BNA_P
ClientSampleId :
SLCS340


Tgt Ion:105 Resp: 1506464
Ion Ratio Lower Upper
105 100
77 82.4 67.0 100.4
51 31.8 25.4 38.0



#17
N-Nitroso-di-n-propylamine
Concen: 38.084 ng/ul
RT: 8.787 min Scan# 969
Delta R.T. 0.006 min
Lab File: BP017109.D
Acq: 11 Sep 2023 18:13

Tgt Ion: 70 Resp: 834048
Ion Ratio Lower Upper
70 100
42 64.0 50.9 76.3
101 8.8 6.7 10.1
130 19.0 14.8 22.2





#18

4-Methylphenol

Concen: 37.205 ng/ul

RT: 8.740 min Scan# 9

Delta R.T. 0.006 min

Lab File: BP017109.D

Acq: 11 Sep 2023 18:13

Instrument :

BNA_P

ClientSampleId :





SLCS340

Tgt Ion:108 Resp: 998440

Ion Ratio Lower Upper

108 100

107 110.9 90.2 135.4



#19

Hexachloroethane

Concen: 36.805 ng/ul

RT: 9.022 min Scan# 1009

Delta R.T. -0.006 min

Lab File: BP017109.D

Acq: 11 Sep 2023 18:13




Tgt Ion:117 Resp: 374193


Ion Ratio Lower Upper

117 100

201 84.7 69.9 104.9

199 53.0 43.9 65.9









#22
Nitrobenzene
Concen: 37.311 ng/ul
RT: 9.205 min Scan# 1040
Delta R.T. 0.000 min
Lab File: BP017109.D
Acq: 11 Sep 2023 18:13




Instrument :
BNA_P
ClientSampleId :
SLCS340


Tgt Ion: 77 Resp: 1215254
Ion Ratio Lower Upper
77 100
123 39.9 32.8 49.2
65 13.6 11.0 16.4



#23
Isophorone
Concen: 37.964 ng/ul
RT: 9.716 min Scan# 1127
Delta R.T. 0.000 min
Lab File: BP017109.D
Acq: 11 Sep 2023 18:13

Tgt Ion: 82 Resp: 2246051
Ion Ratio Lower Upper
82 100
95 6.8 5.2 7.8
138 15.6 12.6 19.0









#25
2-Nitrophenol
Concen: 39.213 ng/ul
RT: 9.905 min Scan# 1159
Delta R.T. 0.000 min
Lab File: BP017109.D
Acq: 11 Sep 2023 18:13




Instrument :
BNA_P
ClientSampleId :
SLCS340


Tgt Ion:139 Resp: 516661
Ion Ratio Lower Upper
139 100
65 59.2 49.9 74.9
109 30.8 25.3 37.9



#26
2,4-Dimethylphenol
Concen: 32.922 ng/ul
RT: 9.946 min Scan# 1166
Delta R.T. 0.000 min
Lab File: BP017109.D
Acq: 11 Sep 2023 18:13

Tgt Ion:107 Resp: 926555
Ion Ratio Lower Upper
107 100
121 49.9 39.1 58.7
122 83.0 66.6 100.0









#27
 Bis(2-Chloroethoxy)methane
 Concen: 36.454 ng/ul
 RT: 10.205 min Scan# 1210
 Delta R.T. 0.000 min
 Lab File: BP017109.D
 Acq: 11 Sep 2023 18:13




Instrument :
 BNA_P
 ClientSampleId :
 SLCS340


Tgt Ion: 93 Resp: 1352407
 Ion Ratio Lower Upper
 93 100
 95 32.1 26.6 39.8
 123 10.7 8.8 13.2



#29
 2,4-Dichlorophenol
 Concen: 37.396 ng/ul
 RT: 10.422 min Scan# 1247
 Delta R.T. 0.000 min
 Lab File: BP017109.D
 Acq: 11 Sep 2023 18:13

Tgt Ion: 162 Resp: 854207
 Ion Ratio Lower Upper
 162 100
 164 66.7 51.0 76.6
 98 39.9 30.7 46.1









#30
Naphthalene
Concen: 36.036 ng/ul
RT: 10.834 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP017109.D
Acq: 11 Sep 2023 18:13




Instrument :
BNA_P
ClientSampleId :
SLCS340


Tgt Ion:128 Resp: 2963186
Ion Ratio Lower Upper
128 100
129 11.2 8.9 13.3
127 13.5 10.4 15.6



#32
4-Chloroaniline
Concen: 28.492 ng/ul
RT: 10.969 min Scan# 1340
Delta R.T. -0.006 min
Lab File: BP017109.D
Acq: 11 Sep 2023 18:13

Tgt Ion:127 Resp: 1008961
Ion Ratio Lower Upper
127 100
129 33.1 25.9 38.9





#33

Hexachlorobutadiene

Concen: 34.763 ng/ul

RT: 11.063 min Scan# 11

Delta R.T. -0.006 min

Lab File: BP017109.D

Acq: 11 Sep 2023 18:13

Instrument :

BNA_P

ClientSampleId :

SLCS340





Tgt Ion:225 Resp: 442226

Ion Ratio Lower Upper

225 100

223 63.8 51.6 77.4

227 64.8 51.8 77.8



#34

Caprolactam

Concen: 25.526 ng/ul

RT: 11.804 min Scan# 1482

Delta R.T. 0.012 min

Lab File: BP017109.D

Acq: 11 Sep 2023 18:13




Tgt Ion:113 Resp: 204132


Ion Ratio Lower Upper

113 100

55 201.1 160.4 240.6

56 146.1 118.5 177.7









#35
4-Chloro-3-methylphenol
Concen: 39.007 ng/ul
RT: 12.075 min Scan# 1527
Delta R.T. 0.000 min
Lab File: BP017109.D
Acq: 11 Sep 2023 18:13




Instrument :
BNA_P
ClientSampleId :
SLCS340


Tgt Ion:107 Resp: 1001582
Ion Ratio Lower Upper
107 100
144 25.1 20.9 31.3
142 77.8 63.2 94.8



#36
2-Methylnaphthalene
Concen: 36.177 ng/ul
RT: 12.440 min Scan# 1590
Delta R.T. 0.000 min
Lab File: BP017109.D
Acq: 11 Sep 2023 18:13

Tgt Ion:142 Resp: 1986301
Ion Ratio Lower Upper
142 100
141 89.1 71.4 107.0





#37

1-Methylnaphthalene

Concen: 34.769 ng/ul

RT: 12.663 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP017109.D

Acq: 11 Sep 2023 18:13

Instrument :

BNA_P

ClientSampleId :

SLCS340





Tgt Ion:142 Resp: 1940956

Ion Ratio Lower Upper

142 100

141 93.2 73.1 109.7

116 3.9 3.2 4.8



#39

1,2,4,5-Tetrachlorobenzene

Concen: 35.049 ng/ul

RT: 12.793 min Scan# 1650

Delta R.T. 0.000 min

Lab File: BP017109.D

Acq: 11 Sep 2023 18:13

Tgt Ion:216 Resp: 884499




Ion Ratio Lower Upper


216 100

214 79.3 62.2 93.2

179 22.6 18.2 27.4

108 20.8 17.0 25.6









#40
Hexachlorocyclopentadiene
Concen: 29.581 ng/ul
RT: 12.740 min Scan# 1641
Delta R.T. 0.000 min
Lab File: BP017109.D
Acq: 11 Sep 2023 18:13




Instrument :
BNA_P
ClientSampleId :
SLCS340


Tgt Ion:237 Resp: 442291
Ion Ratio Lower Upper
237 100
235 62.8 50.9 76.3
272 11.8 8.8 13.2



#41
2,4,6-Trichlorophenol
Concen: 36.453 ng/ul
RT: 13.046 min Scan# 1693
Delta R.T. 0.000 min
Lab File: BP017109.D
Acq: 11 Sep 2023 18:13

Tgt Ion:196 Resp: 612742
Ion Ratio Lower Upper
196 100
198 94.5 76.7 115.1
200 30.5 24.4 36.6









#42
2,4,5-Trichlorophenol
Concen: 37.638 ng/ul
RT: 13.116 min Scan# 1705
Delta R.T. 0.000 min
Lab File: BP017109.D
Acq: 11 Sep 2023 18:13




Instrument :
BNA_P
ClientSampleId :
SLCS340


Tgt Ion:196 Resp: 688634
Ion Ratio Lower Upper
196 100
198 95.8 76.6 115.0
200 31.4 23.8 35.8



#43
1,1'-Biphenyl
Concen: 36.985 ng/ul
RT: 13.451 min Scan# 1762
Delta R.T. 0.000 min
Lab File: BP017109.D
Acq: 11 Sep 2023 18:13

Tgt Ion:154 Resp: 2664501
Ion Ratio Lower Upper
154 100
153 40.9 32.6 49.0
76 15.8 12.2 18.4





#44

2-Chloronaphthalene

Concen: 36.518 ng/ul

RT: 13.499 min Scan# 1770

Delta R.T. 0.000 min

Lab File: BP017109.D

Acq: 11 Sep 2023 18:13

Instrument :

BNA_P

ClientSampleId :

SLCS340





Tgt Ion:162 Resp: 2062738

Ion Ratio Lower Upper

162 100

164 32.9 26.0 39.0

127 39.4 31.8 47.8



#45

2-Nitroaniline

Concen: 41.739 ng/ul

RT: 13.734 min Scan# 1810

Delta R.T. 0.000 min

Lab File: BP017109.D

Acq: 11 Sep 2023 18:13




Tgt Ion: 65 Resp: 747086


Ion Ratio Lower Upper

65 100

92 64.3 52.5 78.7

138 90.3 71.8 107.6









#47
Dimethylphthalate
Concen: 37.171 ng/ul
RT: 14.081 min Scan# 1869
Delta R.T. 0.000 min
Lab File: BP017109.D
Acq: 11 Sep 2023 18:13




Instrument :
BNA_P
ClientSampleId :
SLCS340


Tgt Ion:163 Resp: 2634221
Ion Ratio Lower Upper
163 100
194 3.9 2.9 4.3
164 10.3 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 41.803 ng/ul
RT: 14.228 min Scan# 1894
Delta R.T. 0.000 min
Lab File: BP017109.D
Acq: 11 Sep 2023 18:13

Tgt Ion:165 Resp: 532844
Ion Ratio Lower Upper
165 100
89 67.6 55.6 83.4
121 22.1 20.0 30.0









#50
Acenaphthylene
Concen: 35.463 ng/ul
RT: 14.346 min Scan# 1914
Delta R.T. -0.006 min
Lab File: BP017109.D
Acq: 11 Sep 2023 18:13




Instrument :
BNA_P
ClientSampleId :
SLCS340


Tgt Ion:152 Resp: 3298817
Ion Ratio Lower Upper
152 100
151 20.3 15.5 23.3
153 13.3 10.2 15.4



#51
3-Nitroaniline
Concen: 39.649 ng/ul
RT: 14.569 min Scan# 1952
Delta R.T. 0.006 min
Lab File: BP017109.D
Acq: 11 Sep 2023 18:13

Tgt Ion:138 Resp: 574854
Ion Ratio Lower Upper
138 100
108 10.0 8.5 12.7
92 122.9 103.3 154.9





#52

Acenaphthene

Concen: 36.927 ng/ul

RT: 14.687 min Scan# 1972

Delta R.T. 0.000 min

Lab File: BP017109.D

Acq: 11 Sep 2023 18:13

Instrument :

BNA_P

ClientSampleId :

SLCS340





Tgt Ion:153 Resp: 2311051

Ion Ratio Lower Upper

153 100

152 47.4 38.9 58.3

154 86.2 69.0 103.6



#53

2,4-Dinitrophenol

Concen: 37.377 ng/ul

RT: 14.763 min Scan# 1985

Delta R.T. 0.000 min

Lab File: BP017109.D

Acq: 11 Sep 2023 18:13




Tgt Ion:184 Resp: 300354


Ion Ratio Lower Upper

184 100

63 91.7 74.5 111.7

154 66.0 53.8 80.8









#55
4-Nitrophenol
Concen: 37.272 ng/ul
RT: 14.857 min Scan# 2000
Delta R.T. 0.006 min
Lab File: BP017109.D
Acq: 11 Sep 2023 18:13




Instrument :
BNA_P
ClientSampleId :
SLCS340


Tgt Ion:109 Resp: 439459
Ion Ratio Lower Upper
109 100
139 126.1 98.6 147.8
65 143.0 117.0 175.4



#56
Dibenzofuran
Concen: 36.764 ng/ul
RT: 15.022 min Scan# 2029
Delta R.T. 0.000 min
Lab File: BP017109.D
Acq: 11 Sep 2023 18:13

Tgt Ion:168 Resp: 3125326
Ion Ratio Lower Upper
168 100
139 38.8 30.6 45.8





#57

2,4-Dinitrotoluene

Concen: 40.938 ng/ul

RT: 15.010 min Scan# 2026

Delta R.T. 0.000 min

Lab File: BP017109.D

Acq: 11 Sep 2023 18:13

Instrument :

BNA_P

ClientSampleId :

SLCS340





Tgt Ion:165 Resp: 799945

Ion Ratio Lower Upper

165 100

63 64.4 57.7 86.5

182 2.7 2.2 3.4



#58

2,3,4,6-Tetrachlorophenol

Concen: 35.398 ng/ul

RT: 15.234 min Scan# 2065

Delta R.T. 0.000 min

Lab File: BP017109.D

Acq: 11 Sep 2023 18:13

Tgt Ion:232 Resp: 504875




Ion Ratio Lower Upper


232 100

131 58.2 46.9 70.3

130 3.3 2.4 3.6

166 37.7 29.6 44.4









#59
Diethylphthalate
Concen: 37.820 ng/ul
RT: 15.440 min Scan# 2100
Delta R.T. 0.000 min
Lab File: BP017109.D
Acq: 11 Sep 2023 18:13




Instrument :
BNA_P
ClientSampleId :
SLCS340


Tgt Ion:149 Resp: 2704730
Ion Ratio Lower Upper
149 100
177 21.5 17.4 26.0
150 12.5 9.5 14.3



#61
Fluorene
Concen: 37.418 ng/ul
RT: 15.675 min Scan# 2140
Delta R.T. 0.006 min
Lab File: BP017109.D
Acq: 11 Sep 2023 18:13

Tgt Ion:166 Resp: 2654912
Ion Ratio Lower Upper
166 100
165 98.7 78.8 118.2
167 13.4 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 36.734 ng/ul
RT: 15.663 min Scan# 2138
Delta R.T. 0.000 min
Lab File: BP017109.D
Acq: 11 Sep 2023 18:13




Instrument :
BNA_P
ClientSampleId :
SLCS340


Tgt Ion:204 Resp: 1206310
Ion Ratio Lower Upper
204 100
206 32.9 26.6 40.0
141 63.1 50.6 76.0



#63
4-Nitroaniline
Concen: 42.901 ng/ul
RT: 15.740 min Scan# 2151
Delta R.T. 0.006 min
Lab File: BP017109.D
Acq: 11 Sep 2023 18:13

Tgt Ion:138 Resp: 600200
Ion Ratio Lower Upper
138 100
92 57.5 45.0 67.6
108 101.2 77.2 115.8









#66
4,6-Dinitro-2-methylphenol
Concen: 35.505 ng/ul
RT: 15.763 min Scan# 2154
Delta R.T. 0.006 min
Lab File: BP017109.D
Acq: 11 Sep 2023 18:13




Instrument :
BNA_P
ClientSampleId :
SLCS340


Tgt Ion:198 Resp: 444775
Ion Ratio Lower Upper
198 100
182 4.2 4.7 7.1#
77 30.3 29.6 44.4



#67
N-Nitrosodiphenylamine
Concen: 38.210 ng/ul
RT: 15.887 min Scan# 2176
Delta R.T. 0.000 min
Lab File: BP017109.D
Acq: 11 Sep 2023 18:13

Tgt Ion:169 Resp: 2142760
Ion Ratio Lower Upper
169 100
168 68.4 54.5 81.7
167 37.4 30.4 45.6









#68
4-Bromophenyl-phenylether
Concen: 38.044 ng/ul
RT: 16.569 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP017109.D
Acq: 11 Sep 2023 18:13




Instrument :
BNA_P
ClientSampleId :
SLCS340


Tgt Ion:248 Resp: 648668
Ion Ratio Lower Upper
248 100
250 96.8 78.9 118.3
141 81.3 67.9 101.9



#69
Hexachlorobenzene
Concen: 36.780 ng/ul
RT: 16.645 min Scan# 2305
Delta R.T. 0.000 min
Lab File: BP017109.D
Acq: 11 Sep 2023 18:13

Tgt Ion:284 Resp: 697815
Ion Ratio Lower Upper
284 100
142 49.8 38.9 58.3
249 33.8 26.1 39.1





#70

Atrazine

Concen: 17.313 ng/ul

RT: 16.845 min Scan# 2339

Delta R.T. 0.000 min

Lab File: BP017109.D

Acq: 11 Sep 2023 18:13

Instrument :

BNA_P

ClientSampleId :

SLCS340





Tgt Ion:200 Resp: 335912

Ion Ratio Lower Upper

200 100

173 27.2 21.0 31.4

215 46.1 35.9 53.9



#71

Pentachlorophenol

Concen: 29.896 ng/ul

RT: 17.010 min Scan# 2367

Delta R.T. 0.000 min

Lab File: BP017109.D

Acq: 11 Sep 2023 18:13




Tgt Ion:266 Resp: 385622


Ion Ratio Lower Upper

266 100

264 62.0 49.8 74.8

268 62.8 50.9 76.3





#72

Phenanthrene

Concen: 37.550 ng/ul

RT: 17.434 min Scan# 2439

Delta R.T. 0.000 min

Lab File: BP017109.D

Acq: 11 Sep 2023 18:13

Instrument :

BNA_P

ClientSampleId :

SLCS340





Tgt Ion:178 Resp: 4011624

Ion Ratio Lower Upper

178 100

179 15.6 12.3 18.5

176 19.9 15.8 23.8



#74

Anthracene

Concen: 37.020 ng/ul

RT: 17.528 min Scan# 2455

Delta R.T. 0.000 min

Lab File: BP017109.D

Acq: 11 Sep 2023 18:13




Tgt Ion:178 Resp: 3998569


Ion Ratio Lower Upper

178 100

179 15.5 12.2 18.2

176 19.5 15.4 23.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 3.296 ng/uL
RT: 13.404 min Scan# 1754
Delta R.T. 0.000 min
Lab File: BP017109.D
Acq: 11 Sep 2023 18:13




Instrument :
BNA_P
ClientSampleId :
SLCS340


Tgt Ion:216 Resp: 85398
Ion Ratio Lower Upper
216 100
214 80.1 63.8 95.6
218 47.0 38.1 57.1



#76
Pentachlorobenzene
Concen: 39.143 ng/uL
RT: 14.910 min Scan# 2010
Delta R.T. 0.000 min
Lab File: BP017109.D
Acq: 11 Sep 2023 18:13

Tgt Ion:250 Resp: 873854
Ion Ratio Lower Upper
250 100
248 62.9 49.9 74.9
252 62.9 51.4 77.0









#77
 Carbazole
 Concen: 38.133 ng/ul
 RT: 17.810 min Scan# 2503
 Delta R.T. 0.000 min
 Lab File: BP017109.D
 Acq: 11 Sep 2023 18:13




Instrument :
 BNA_P
 ClientSampleId :
 SLCS340


Tgt Ion:167 Resp: 3754104
 Ion Ratio Lower Upper
 167 100
 166 22.2 17.7 26.5
 139 13.8 10.8 16.2



#78
 Di-n-butylphthalate
 Concen: 40.008 ng/ul
 RT: 18.322 min Scan# 2590
 Delta R.T. 0.000 min
 Lab File: BP017109.D
 Acq: 11 Sep 2023 18:13

Tgt Ion:149 Resp: 4541262
 Ion Ratio Lower Upper
 149 100
 150 9.5 7.3 10.9
 104 5.7 4.7 7.1





#80

Fluoranthene

Concen: 45.294 ng/ul

RT: 19.398 min Scan# 2773

Delta R.T. 0.000 min

Lab File: BP017109.D

Acq: 11 Sep 2023 18:13

Instrument :

BNA_P

ClientSampleId :

SLCS340





Tgt Ion:202 Resp: 4392779

Ion Ratio Lower Upper

202 100

101 13.2 9.9 14.9

100 10.1 7.8 11.8



#82

Pyrene

Concen: 44.113 ng/ul

RT: 19.757 min Scan# 2834

Delta R.T. 0.000 min

Lab File: BP017109.D

Acq: 11 Sep 2023 18:13




Tgt Ion:202 Resp: 4569408


Ion Ratio Lower Upper

202 100

101 15.9 11.9 17.9

100 13.2 9.8 14.6









#83
Butylbenzylphthalate
Concen: 44.775 ng/ul
RT: 20.616 min Scan# 2980
Delta R.T. 0.000 min
Lab File: BP017109.D
Acq: 11 Sep 2023 18:13




Instrument :
BNA_P
ClientSampleId :
SLCS340


Tgt Ion:149 Resp: 1908254
Ion Ratio Lower Upper
149 100
91 79.7 64.6 96.8
206 18.1 14.6 22.0



#84
3,3'-Dichlorobenzidine
Concen: 35.909 ng/ul
RT: 21.410 min Scan# 3115
Delta R.T. 0.000 min
Lab File: BP017109.D
Acq: 11 Sep 2023 18:13

Tgt Ion:252 Resp: 1119529
Ion Ratio Lower Upper
252 100
254 64.8 51.7 77.5
126 14.6 11.5 17.3





#85

Benzo(a)anthracene

Concen: 37.617 ng/ul

RT: 21.475 min Scan# 3125

Delta R.T. 0.006 min

Lab File: BP017109.D

Acq: 11 Sep 2023 18:13

Instrument :

BNA_P

ClientSampleId :

SLCS340





Tgt Ion:228 Resp: 3849833

Ion Ratio Lower Upper

228 100

229 20.2 16.0 24.0

226 28.6 22.9 34.3



#86

Bis(2-ethylhexyl)phthalate

Concen: 41.452 ng/ul

RT: 21.351 min Scan# 3105

Delta R.T. 0.000 min

Lab File: BP017109.D

Acq: 11 Sep 2023 18:13




Tgt Ion:149 Resp: 2528284


Ion Ratio Lower Upper

149 100

167 26.5 20.8 31.2

279 3.4 2.6 3.8









#87
 Chrysene
 Concen: 37.466 ng/ul
 RT: 21.527 min Scan# 3135
 Delta R.T. 0.000 min
 Lab File: BP017109.D
 Acq: 11 Sep 2023 18:13

Instrument :
 BNA_P
 ClientSampleId :
 SLCS340




Tgt Ion:228 Resp: 3516407


Ion	Ratio	Lower	Upper
228	100		
226	31.7	24.6	36.8
229	20.1	15.6	23.4



#89
 Di-n-octyl phthalate
 Concen: 43.246 ng/ul
 RT: 22.316 min Scan# 3269
 Delta R.T. 0.000 min
 Lab File: BP017109.D
 Acq: 11 Sep 2023 18:13

Tgt Ion:149 Resp: 3725782









#90
Benzo(b)fluoranthene
Concen: 41.322 ng/ul
RT: 23.233 min Scan# 3424
Delta R.T. 0.006 min
Lab File: BP017109.D
Acq: 11 Sep 2023 18:13




Instrument :
BNA_P
ClientSampleId :
SLCS340


Tgt Ion:252 Resp: 2915090
Ion Ratio Lower Upper
252 100
253 21.9 17.9 26.9
125 12.7 10.3 15.5



#91
Benzo(k)fluoranthene
Concen: 40.933 ng/ul
RT: 23.286 min Scan# 3434
Delta R.T. 0.006 min
Lab File: BP017109.D
Acq: 11 Sep 2023 18:13

Tgt Ion:252 Resp: 2903502
Ion Ratio Lower Upper
252 100
253 23.0 17.7 26.5
125 12.2 10.0 15.0





#93

Benzo(a)pyrene

Concen: 38.032 ng/ul

RT: 23.910 min Scan# 3

Delta R.T. 0.006 min

Lab File: BP017109.D

Acq: 11 Sep 2023 18:13

Instrument :

BNA_P

ClientSampleId :

SLCS340





Tgt Ion:252 Resp: 2494124

Ion Ratio Lower Upper

252 100

253 21.8 17.6 26.4

125 13.5 11.3 16.9



#94

Indeno(1,2,3-cd)pyrene

Concen: 39.507 ng/ul

RT: 26.704 min Scan# 4015

Delta R.T. 0.006 min

Lab File: BP017109.D

Acq: 11 Sep 2023 18:13




Tgt Ion:276 Resp: 2910159


Ion Ratio Lower Upper

276 100

138 29.6 23.4 35.2

277 25.0 20.1 30.1





#95

Dibenzo(a,h)anthracene

Concen: 39.210 ng/ul

RT: 26.727 min Scan# 4018

Delta R.T. 0.006 min

Lab File: BP017109.D

Acq: 11 Sep 2023 18:13

Instrument :

BNA_P

ClientSampleId :

SLCS340





Tgt Ion:278 Resp: 2408144

Ion Ratio Lower Upper

278 100

139 20.7 16.5 24.7

279 23.6 19.2 28.8



#96

Benzo(g,h,i)perylene

Concen: 41.218 ng/ul

RT: 27.539 min Scan# 4157

Delta R.T. 0.000 min

Lab File: BP017109.D

Acq: 11 Sep 2023 18:13

Tgt Ion:276 Resp: 2314082

Ion Ratio Lower Upper

276 100

138 26.9 22.0 33.0

277 23.9 18.7 28.1

