

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091023\
 Data File : BP017110.D
 Acq On : 11 Sep 2023 18:47
 Operator : MA/JU
 Sample : PB155343BS
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS343

Quant Time: Sep 12 02:30:59 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP091023.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Sep 11 06:13:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.958	152	246731	20.000	ng/u1	0.00
20) Naphthalene-d8	10.787	136	1147185	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.616	164	716171	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.386	188	1538238	20.000	ng/u1	0.00
79) Chrysene-d12	21.486	240	1391469	20.000	ng/u1	0.00
88) Perylene-d12	24.016	264	1362233	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.311	96	47991	7.424	ng/uL	0.00
4) Pyridine-d5	3.746	84	693132	35.557	ng/u1	0.00
7) Phenol-d5	7.111	99	927617	41.360	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.299	67	593680	39.468	ng/u1	0.00
11) 2-Chlorophenol-d4	7.481	132	656980	40.529	ng/u1	0.00
15) 4-Methylphenol-d8	8.669	113	711813	40.121	ng/u1	0.00
21) Nitrobenzene-d5	9.157	128	327573	40.108	ng/u1	0.00
24) 2-Nitrophenol-d4	9.875	143	358128	42.486	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.399	165	662250	41.433	ng/u1	0.00
31) 4-Chloroaniline-d4	10.946	131	909798	35.497	ng/u1	0.00
46) Dimethylphthalate-d6	14.034	166	2026711	39.771	ng/u1	0.00
49) Acenaphthylene-d8	14.316	160	2306306	39.746	ng/u1	0.00
54) 4-Nitrophenol-d4	14.840	143	412651	39.250	ng/u1	0.00
60) Fluorene-d10	15.616	176	1697816	39.262	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.745	200	335452	38.643	ng/u1	0.00
73) Anthracene-d10	17.486	188	2631311	39.908	ng/u1	0.00
81) Pyrene-d10	19.727	212	3028494	41.329	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.851	264	2782267	42.545	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.352	88	101140	14.414	ng/uL	98
5) Pyridine	3.770	79	717954	35.064	ng/u1	98
6) Benzaldehyde	7.116	77	505237	38.288	ng/u1	99
8) Phenol	7.134	94	990919	41.576	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.393	93	758121	39.884	ng/u1	98
12) 2-Chlorophenol	7.511	128	701459	41.761	ng/u1	98
13) 2-Methylphenol	8.399	108	698766	40.881	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.475	45	634428	21.310	ng/u1	100
16) Acetophenone	8.810	105	1180786	41.436	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.781	70	651133	42.707	ng/u1	99
18) 4-Methylphenol	8.734	108	788688	42.214	ng/u1	100
19) Hexachloroethane	9.028	117	282557	39.920	ng/u1	99
22) Nitrobenzene	9.205	77	947917	40.490	ng/u1	100
23) Isophorone	9.716	82	1774207	41.722	ng/u1	100
25) 2-Nitrophenol	9.905	139	404264	42.687	ng/u1	96
26) 2,4-Dimethylphenol	9.946	107	721540	35.669	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.204	93	1063259	39.873	ng/u1	99
29) 2,4-Dichlorophenol	10.422	162	676569	41.208	ng/u1	99
30) Naphthalene	10.834	128	2333051	39.474	ng/u1	100
32) 4-Chloroaniline	10.975	127	805375	31.641	ng/u1	99
33) Hexachlorobutadiene	11.063	225	349093	38.179	ng/u1	98
34) Caprolactam	11.799	113	166980	29.050	ng/u1	98
35) 4-Chloro-3-methylphenol	12.075	107	800986	43.400	ng/u1	99
36) 2-Methylnaphthalene	12.440	142	1584518	40.150	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091023\
 Data File : BP017110.D
 Acq On : 11 Sep 2023 18:47
 Operator : MA/JU
 Sample : PB155343BS
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS343


Quant Time: Sep 12 02:30:59 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP091023.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Sep 11 06:13:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.663	142	1536186	38.285	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.793	216	706969	38.103	ng/ul	99
40) Hexachlorocyclopentadiene	12.740	237	335871	30.553	ng/ul	99
41) 2,4,6-Trichlorophenol	13.046	196	488854	39.556	ng/ul	99
42) 2,4,5-Trichlorophenol	13.116	196	554636	41.231	ng/ul	99
43) 1,1'-Biphenyl	13.451	154	2115446	39.939	ng/ul	100
44) 2-Chloronaphthalene	13.498	162	1658372	39.933	ng/ul	100
45) 2-Nitroaniline	13.734	65	595651	45.263	ng/ul	99
47) Dimethylphthalate	14.081	163	2120831	40.704	ng/ul	99
48) 2,6-Dinitrotoluene	14.228	165	422736	45.109	ng/ul	96
50) Acenaphthylene	14.345	152	2668847	39.023	ng/ul	98
51) 3-Nitroaniline	14.563	138	465755	43.693	ng/ul	97
52) Acenaphthene	14.681	153	1839783	39.983	ng/ul	99
53) 2,4-Dinitrophenol	14.763	184	228034	38.597	ng/ul	97
55) 4-Nitrophenol	14.851	109	363924	41.982	ng/ul	98
56) Dibenzofuran	15.022	168	2515562	40.248	ng/ul	99
57) 2,4-Dinitrotoluene	15.010	165	646690	45.013	ng/ul	93
58) 2,3,4,6-Tetrachlorophenol	15.234	232	408148	38.922	ng/ul	98
59) Diethylphthalate	15.440	149	2176413	41.392	ng/ul	99
61) Fluorene	15.669	166	2129859	40.829	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.663	204	972183	40.266	ng/ul	98
63) 4-Nitroaniline	15.739	138	498135	48.428	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.757	198	357040	37.778	ng/ul	98
67) N-Nitrosodiphenylamine	15.887	169	1739802	41.122	ng/ul	99
68) 4-Bromophenyl-phenylether	16.563	248	522433	40.612	ng/ul	96
69) Hexachlorobenzene	16.645	284	572105	39.969	ng/ul	98
70) Atrazine	16.839	200	267213	18.255	ng/ul	98
71) Pentachlorophenol	17.010	266	317339	32.610	ng/ul	99
72) Phenanthrene	17.428	178	3301586	40.962	ng/ul	99
74) Anthracene	17.522	178	3276545	40.208	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.404	216	68164	3.487	ng/uL	97
76) Pentachlorobenzene	14.910	250	701418	41.645	ng/uL	98
77) Carbazole	17.810	167	3173281	42.724	ng/ul	100
78) Di-n-butylphthalate	18.316	149	3729196	43.546	ng/ul	99
80) Fluoranthene	19.398	202	3787218	42.844	ng/ul	99
82) Pyrene	19.757	202	3980983	42.166	ng/ul	99
83) Butylbenzylphthalate	20.616	149	1760166	45.313	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.410	252	1215343	42.769	ng/ul	99
85) Benzo(a)anthracene	21.469	228	3868501	41.472	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.345	149	2512223	45.190	ng/ul	99
87) Chrysene	21.527	228	3587256	41.934	ng/ul	98
89) Di-n-octyl phthalate	22.316	149	4357354	42.894	ng/ul	100
90) Benzo(b)fluoranthene	23.227	252	3712130	44.626	ng/ul	100
91) Benzo(k)fluoranthene	23.286	252	3608812	43.147	ng/ul	99
93) Benzo(a)pyrene	23.904	252	3248020	42.004	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.704	276	3947325	45.447	ng/ul	98
95) Dibenzo(a,h)anthracene	26.721	278	3301427	45.589	ng/ul	100
96) Benzo(g,h,i)perylene	27.545	276	3081775	46.554	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Sep 12 02:30:59 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP091023.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Sep 11 06:13:29 2023
Response via : Initial Calibration





#2

1,4-Dioxane

Concen: 14.414 ng/uL

RT: 3.352 min Scan# 41

Delta R.T. -0.000 min

Lab File: BP017110.D

Acq: 11 Sep 2023 18:47

Instrument :

BNA_P

ClientSampleId :

SLCS343





Tgt Ion: 88 Resp: 101140

Ion Ratio Lower Upper

88 100

43 34.1 27.3 40.9

58 77.6 63.7 95.5



#5

Pyridine

Concen: 35.064 ng/uL

RT: 3.770 min Scan# 116

Delta R.T. -0.006 min

Lab File: BP017110.D

Acq: 11 Sep 2023 18:47




Tgt Ion: 79 Resp: 717954


Ion Ratio Lower Upper

79 100

52 79.8 61.6 92.4

51 38.7 30.9 46.3









#6
Benzaldehyde
Concen: 38.288 ng/ul
RT: 7.116 min Scan# 685
Delta R.T. 0.000 min
Lab File: BP017110.D
Acq: 11 Sep 2023 18:47




Instrument :
BNA_P
ClientSampleId :
SLCS343


Tgt Ion: 77 Resp: 505237
Ion Ratio Lower Upper
77 100
105 88.5 70.3 105.5
106 86.7 69.0 103.6



#8
Phenol
Concen: 41.576 ng/ul
RT: 7.134 min Scan# 688
Delta R.T. -0.006 min
Lab File: BP017110.D
Acq: 11 Sep 2023 18:47

Tgt Ion: 94 Resp: 990919
Ion Ratio Lower Upper
94 100
65 31.1 25.4 38.0
66 39.3 30.6 46.0









#10
Bis(2-Chloroethyl)ether
Concen: 39.884 ng/ul
RT: 7.393 min Scan# 732
Delta R.T. 0.000 min
Lab File: BP017110.D
Acq: 11 Sep 2023 18:47




Instrument :
BNA_P
ClientSampleId :
SLCS343


Tgt Ion: 93 Resp: 758121
Ion Ratio Lower Upper
93 100
63 83.5 65.4 98.0
95 31.7 25.1 37.7



#12
2-Chlorophenol
Concen: 41.761 ng/ul
RT: 7.511 min Scan# 752
Delta R.T. -0.006 min
Lab File: BP017110.D
Acq: 11 Sep 2023 18:47

Tgt Ion:128 Resp: 701459
Ion Ratio Lower Upper
128 100
64 54.3 42.2 63.4
130 32.4 25.3 37.9









#13
2-Methylphenol
Concen: 40.881 ng/ul
RT: 8.399 min Scan# 903
Delta R.T. 0.000 min
Lab File: BP017110.D
Acq: 11 Sep 2023 18:47




Instrument :
BNA_P
ClientSampleId :
SLCS343


Tgt Ion:108 Resp: 698766
Ion Ratio Lower Upper
108 100
107 89.4 72.3 108.5



#14
2,2'-oxybis(1-Chloropropane)
Concen: 21.310 ng/ul
RT: 8.475 min Scan# 916
Delta R.T. 0.000 min
Lab File: BP017110.D
Acq: 11 Sep 2023 18:47

Tgt Ion: 45 Resp: 634428
Ion Ratio Lower Upper
45 100
77 13.6 11.0 16.4
79 9.8 7.8 11.6









#16
Acetophenone
Concen: 41.436 ng/ul
RT: 8.810 min Scan# 911
Delta R.T. 0.000 min
Lab File: BP017110.D
Acq: 11 Sep 2023 18:47




Instrument :
BNA_P
ClientSampleId :
SLCS343


Tgt Ion:105 Resp: 1180786
Ion Ratio Lower Upper
105 100
77 84.2 67.0 100.4
51 32.3 25.4 38.0



#17
N-Nitroso-di-n-propylamine
Concen: 42.707 ng/ul
RT: 8.781 min Scan# 968
Delta R.T. 0.000 min
Lab File: BP017110.D
Acq: 11 Sep 2023 18:47

Tgt Ion: 70 Resp: 651133
Ion Ratio Lower Upper
70 100
42 64.3 50.9 76.3
101 8.8 6.7 10.1
130 17.8 14.8 22.2









#18
4-Methylphenol
Concen: 42.214 ng/ul
RT: 8.734 min Scan# 960
Delta R.T. 0.000 min
Lab File: BP017110.D
Acq: 11 Sep 2023 18:47




Instrument :
BNA_P
ClientSampleId :
SLCS343


Tgt Ion:108 Resp: 788688
Ion Ratio Lower Upper
108 100
107 112.3 90.2 135.4



#19
Hexachloroethane
Concen: 39.920 ng/ul
RT: 9.028 min Scan# 1010
Delta R.T. 0.000 min
Lab File: BP017110.D
Acq: 11 Sep 2023 18:47

Tgt Ion:117 Resp: 282557
Ion Ratio Lower Upper
117 100
201 88.3 69.9 104.9
199 54.7 43.9 65.9









#22
Nitrobenzene
Concen: 40.490 ng/ul
RT: 9.205 min Scan# 1040
Delta R.T. 0.000 min
Lab File: BP017110.D
Acq: 11 Sep 2023 18:47




Instrument :
BNA_P
ClientSampleId :
SLCS343


Tgt Ion: 77 Resp: 947917
Ion Ratio Lower Upper
77 100
123 40.8 32.8 49.2
65 13.7 11.0 16.4



#23
Isophorone
Concen: 41.722 ng/ul
RT: 9.716 min Scan# 1127
Delta R.T. 0.000 min
Lab File: BP017110.D
Acq: 11 Sep 2023 18:47

Tgt Ion: 82 Resp: 1774207
Ion Ratio Lower Upper
82 100
95 6.6 5.2 7.8
138 15.6 12.6 19.0









#25
2-Nitrophenol
Concen: 42.687 ng/ul
RT: 9.905 min Scan# 1159
Delta R.T. 0.000 min
Lab File: BP017110.D
Acq: 11 Sep 2023 18:47




Instrument :
BNA_P
ClientSampleId :
SLCS343


Tgt Ion:139 Resp: 404264
Ion Ratio Lower Upper
139 100
65 58.4 49.9 74.9
109 31.5 25.3 37.9



#26
2,4-Dimethylphenol
Concen: 35.669 ng/ul
RT: 9.946 min Scan# 1166
Delta R.T. 0.000 min
Lab File: BP017110.D
Acq: 11 Sep 2023 18:47

Tgt Ion:107 Resp: 721540
Ion Ratio Lower Upper
107 100
121 48.1 39.1 58.7
122 84.7 66.6 100.0









#27
Bis(2-Chloroethoxy)methane
Concen: 39.873 ng/ul
RT: 10.204 min Scan# 1210
Delta R.T. 0.000 min
Lab File: BP017110.D
Acq: 11 Sep 2023 18:47




Instrument :
BNA_P
ClientSampleId :
SLCS343


Tgt Ion: 93 Resp: 1063259
Ion Ratio Lower Upper
93 100
95 32.3 26.6 39.8
123 10.6 8.8 13.2



#29
2,4-Dichlorophenol
Concen: 41.208 ng/ul
RT: 10.422 min Scan# 1247
Delta R.T. 0.000 min
Lab File: BP017110.D
Acq: 11 Sep 2023 18:47

Tgt Ion: 162 Resp: 676569
Ion Ratio Lower Upper
162 100
164 64.4 51.0 76.6
98 38.5 30.7 46.1





#30

Naphthalene

Concen: 39.474 ng/ul

RT: 10.834 min Scan# 11

Delta R.T. -0.006 min

Lab File: BP017110.D

Acq: 11 Sep 2023 18:47

Instrument :

BNA_P

ClientSampleId :

SLCS343





Tgt Ion:128 Resp: 2333051

Ion Ratio Lower Upper

128 100

129 11.0 8.9 13.3

127 13.0 10.4 15.6



#32

4-Chloroaniline

Concen: 31.641 ng/ul

RT: 10.975 min Scan# 1341

Delta R.T. 0.000 min

Lab File: BP017110.D




Acq: 11 Sep 2023 18:47


Tgt Ion:127 Resp: 805375

Ion Ratio Lower Upper

127 100

129 32.1 25.9 38.9









#33
Hexachlorobutadiene
Concen: 38.179 ng/ul
RT: 11.063 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP017110.D
Acq: 11 Sep 2023 18:47




Instrument :
BNA_P
ClientSampleId :
SLCS343


Tgt Ion:225 Resp: 349093
Ion Ratio Lower Upper
225 100
223 62.1 51.6 77.4
227 63.4 51.8 77.8



#34
Caprolactam
Concen: 29.050 ng/ul
RT: 11.799 min Scan# 1481
Delta R.T. 0.006 min
Lab File: BP017110.D
Acq: 11 Sep 2023 18:47

Tgt Ion:113 Resp: 166980
Ion Ratio Lower Upper
113 100
55 197.4 160.4 240.6
56 146.4 118.5 177.7









#35
4-Chloro-3-methylphenol
Concen: 43.400 ng/ul
RT: 12.075 min Scan# 1528
Delta R.T. 0.000 min
Lab File: BP017110.D
Acq: 11 Sep 2023 18:47




Instrument :
BNA_P
ClientSampleId :
SLCS343


Tgt Ion:107 Resp: 800986
Ion Ratio Lower Upper
107 100
144 25.4 20.9 31.3
142 79.4 63.2 94.8



#36
2-Methylnaphthalene
Concen: 40.150 ng/ul
RT: 12.440 min Scan# 1590
Delta R.T. 0.000 min
Lab File: BP017110.D
Acq: 11 Sep 2023 18:47

Tgt Ion:142 Resp: 1584518
Ion Ratio Lower Upper
142 100
141 89.5 71.4 107.0





#37

1-Methylnaphthalene

Concen: 38.285 ng/ul

RT: 12.663 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP017110.D

Acq: 11 Sep 2023 18:47

Instrument :

BNA_P

ClientSampleId :

SLCS343





Tgt Ion:142 Resp: 1536186

Ion Ratio Lower Upper

142 100

141 92.9 73.1 109.7

116 3.9 3.2 4.8



#39

1,2,4,5-Tetrachlorobenzene

Concen: 38.103 ng/ul

RT: 12.793 min Scan# 1650

Delta R.T. 0.000 min

Lab File: BP017110.D

Acq: 11 Sep 2023 18:47

Tgt Ion:216 Resp: 706969




Ion Ratio Lower Upper


216 100

214 78.1 62.2 93.2

179 22.6 18.2 27.4

108 20.7 17.0 25.6









#40
Hexachlorocyclopentadiene
Concen: 30.553 ng/ul
RT: 12.740 min Scan# 1641
Delta R.T. 0.000 min
Lab File: BP017110.D
Acq: 11 Sep 2023 18:47




Instrument :
BNA_P
ClientSampleId :
SLCS343


Tgt Ion:237 Resp: 335871
Ion Ratio Lower Upper
237 100
235 63.1 50.9 76.3
272 11.9 8.8 13.2



#41
2,4,6-Trichlorophenol
Concen: 39.556 ng/ul
RT: 13.046 min Scan# 1693
Delta R.T. 0.000 min
Lab File: BP017110.D
Acq: 11 Sep 2023 18:47

Tgt Ion:196 Resp: 488854
Ion Ratio Lower Upper
196 100
198 96.5 76.7 115.1
200 31.1 24.4 36.6





#42

2,4,5-Trichlorophenol

Concen: 41.231 ng/ul

RT: 13.116 min Scan# 1705

Delta R.T. 0.000 min

Lab File: BP017110.D

Acq: 11 Sep 2023 18:47

Instrument :

BNA_P

ClientSampleId :

SLCS343





Tgt Ion:196 Resp: 554636

Ion Ratio Lower Upper

196 100

198 94.8 76.6 115.0

200 31.2 23.8 35.8



#43

1,1'-Biphenyl

Concen: 39.939 ng/ul

RT: 13.451 min Scan# 1762

Delta R.T. 0.000 min

Lab File: BP017110.D

Acq: 11 Sep 2023 18:47




Tgt Ion:154 Resp: 2115446


Ion Ratio Lower Upper

154 100

153 40.6 32.6 49.0

76 15.4 12.2 18.4





#44

2-Chloronaphthalene

Concen: 39.933 ng/ul

RT: 13.498 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP017110.D

Acq: 11 Sep 2023 18:47

Instrument :

BNA_P

ClientSampleId :

SLCS343





Tgt Ion:162 Resp: 1658372

Ion Ratio Lower Upper

162 100

164 32.7 26.0 39.0

127 39.5 31.8 47.8



#45

2-Nitroaniline

Concen: 45.263 ng/ul

RT: 13.734 min Scan# 1810

Delta R.T. 0.000 min

Lab File: BP017110.D

Acq: 11 Sep 2023 18:47




Tgt Ion: 65 Resp: 595651


Ion Ratio Lower Upper

65 100

92 63.9 52.5 78.7

138 89.6 71.8 107.6









#47
Dimethylphthalate
Concen: 40.704 ng/ul
RT: 14.081 min Scan# 1869
Delta R.T. 0.000 min
Lab File: BP017110.D
Acq: 11 Sep 2023 18:47




Instrument :
BNA_P
ClientSampleId :
SLCS343


Tgt Ion:163 Resp: 2120831
Ion Ratio Lower Upper
163 100
194 3.9 2.9 4.3
164 10.1 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 45.109 ng/ul
RT: 14.228 min Scan# 1894
Delta R.T. 0.000 min
Lab File: BP017110.D
Acq: 11 Sep 2023 18:47

Tgt Ion:165 Resp: 422736
Ion Ratio Lower Upper
165 100
89 66.9 55.6 83.4
121 22.8 20.0 30.0









#50
Acenaphthylene
Concen: 39.023 ng/ul
RT: 14.345 min Scan# 1914
Delta R.T. -0.006 min
Lab File: BP017110.D
Acq: 11 Sep 2023 18:47




Instrument :
BNA_P
ClientSampleId :
SLCS343


Tgt Ion:152 Resp: 2668847
Ion Ratio Lower Upper
152 100
151 20.3 15.5 23.3
153 13.4 10.2 15.4



#51
3-Nitroaniline
Concen: 43.693 ng/ul
RT: 14.563 min Scan# 1951
Delta R.T. 0.000 min
Lab File: BP017110.D
Acq: 11 Sep 2023 18:47

Tgt Ion:138 Resp: 465755
Ion Ratio Lower Upper
138 100
108 10.2 8.5 12.7
92 125.5 103.3 154.9









#52
Acenaphthene
Concen: 39.983 ng/ul
RT: 14.681 min Scan# 1971
Delta R.T. -0.006 min
Lab File: BP017110.D
Acq: 11 Sep 2023 18:47




Instrument :
BNA_P
ClientSampleId :
SLCS343


Tgt Ion:153 Resp: 1839783
Ion Ratio Lower Upper
153 100
152 48.5 38.9 58.3
154 87.8 69.0 103.6



#53
2,4-Dinitrophenol
Concen: 38.597 ng/ul
RT: 14.763 min Scan# 1985
Delta R.T. 0.000 min
Lab File: BP017110.D
Acq: 11 Sep 2023 18:47

Tgt Ion:184 Resp: 228034
Ion Ratio Lower Upper
184 100
63 89.7 74.5 111.7
154 65.4 53.8 80.8









#55
4-Nitrophenol
Concen: 41.982 ng/ul
RT: 14.851 min Scan# 2000
Delta R.T. 0.000 min
Lab File: BP017110.D
Acq: 11 Sep 2023 18:47




Instrument :
BNA_P
ClientSampleId :
SLCS343


Tgt Ion:109 Resp: 363924
Ion Ratio Lower Upper
109 100
139 121.0 98.6 147.8
65 144.2 117.0 175.4



#56
Dibenzofuran
Concen: 40.248 ng/ul
RT: 15.022 min Scan# 2029
Delta R.T. 0.000 min
Lab File: BP017110.D
Acq: 11 Sep 2023 18:47

Tgt Ion:168 Resp: 2515562
Ion Ratio Lower Upper
168 100
139 39.0 30.6 45.8









#57
2,4-Dinitrotoluene
Concen: 45.013 ng/ul
RT: 15.010 min Scan# 2027
Delta R.T. 0.000 min
Lab File: BP017110.D
Acq: 11 Sep 2023 18:47




Instrument :
BNA_P
ClientSampleId :
SLCS343


Tgt Ion:165 Resp: 646690
Ion Ratio Lower Upper
165 100
63 66.2 57.7 86.5
182 2.8 2.2 3.4



#58
2,3,4,6-Tetrachlorophenol
Concen: 38.922 ng/ul
RT: 15.234 min Scan# 2065
Delta R.T. 0.000 min
Lab File: BP017110.D
Acq: 11 Sep 2023 18:47

Tgt Ion:232 Resp: 408148
Ion Ratio Lower Upper
232 100
131 56.4 46.9 70.3
130 3.0 2.4 3.6
166 36.3 29.6 44.4









#59
Diethylphthalate
Concen: 41.392 ng/ul
RT: 15.440 min Scan# 2100
Delta R.T. 0.000 min
Lab File: BP017110.D
Acq: 11 Sep 2023 18:47




Instrument :
BNA_P
ClientSampleId :
SLCS343


Tgt Ion:149 Resp: 2176413
Ion Ratio Lower Upper
149 100
177 21.0 17.4 26.0
150 12.2 9.5 14.3



#61
Fluorene
Concen: 40.829 ng/ul
RT: 15.669 min Scan# 2139
Delta R.T. 0.000 min
Lab File: BP017110.D
Acq: 11 Sep 2023 18:47

Tgt Ion:166 Resp: 2129859
Ion Ratio Lower Upper
166 100
165 98.5 78.8 118.2
167 13.7 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 40.266 ng/ul
RT: 15.663 min Scan# 2138
Delta R.T. 0.000 min
Lab File: BP017110.D
Acq: 11 Sep 2023 18:47




Instrument :
BNA_P
ClientSampleId :
SLCS343


Tgt Ion:204 Resp: 972183
Ion Ratio Lower Upper
204 100
206 32.4 26.6 40.0
141 61.1 50.6 76.0



#63
4-Nitroaniline
Concen: 48.428 ng/ul
RT: 15.739 min Scan# 2151
Delta R.T. 0.006 min
Lab File: BP017110.D
Acq: 11 Sep 2023 18:47

Tgt Ion:138 Resp: 498135
Ion Ratio Lower Upper
138 100
92 55.9 45.0 67.6
108 102.5 77.2 115.8









#66
4,6-Dinitro-2-methylphenol
Concen: 37.778 ng/ul
RT: 15.757 min Scan# 2154
Delta R.T. 0.000 min
Lab File: BP017110.D
Acq: 11 Sep 2023 18:47




Instrument :
BNA_P
ClientSampleId :
SLCS343


Tgt Ion:198 Resp: 357040
Ion Ratio Lower Upper
198 100
182 4.9 4.7 7.1
77 35.6 29.6 44.4



#67
N-Nitrosodiphenylamine
Concen: 41.122 ng/ul
RT: 15.887 min Scan# 2176
Delta R.T. 0.000 min
Lab File: BP017110.D
Acq: 11 Sep 2023 18:47

Tgt Ion:169 Resp: 1739802
Ion Ratio Lower Upper
169 100
168 69.5 54.5 81.7
167 37.6 30.4 45.6









#68
4-Bromophenyl-phenylether
Concen: 40.612 ng/ul
RT: 16.563 min Scan# 2291
Delta R.T. -0.006 min
Lab File: BP017110.D
Acq: 11 Sep 2023 18:47




Instrument :
BNA_P
ClientSampleId :
SLCS343


Tgt Ion:248 Resp: 522433
Ion Ratio Lower Upper
248 100
250 95.5 78.9 118.3
141 89.9 67.9 101.9



#69
Hexachlorobenzene
Concen: 39.969 ng/ul
RT: 16.645 min Scan# 2305
Delta R.T. 0.000 min
Lab File: BP017110.D
Acq: 11 Sep 2023 18:47

Tgt Ion:284 Resp: 572105
Ion Ratio Lower Upper
284 100
142 46.7 38.9 58.3
249 32.6 26.1 39.1





#70

Atrazine

Concen: 18.255 ng/ul

RT: 16.839 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP017110.D

Acq: 11 Sep 2023 18:47

Instrument :

BNA_P

ClientSampleId :

SLCS343





Tgt Ion:200 Resp: 267213

Ion Ratio Lower Upper

200 100

173 27.5 21.0 31.4

215 45.5 35.9 53.9



#71

Pentachlorophenol

Concen: 32.610 ng/ul

RT: 17.010 min Scan# 2367

Delta R.T. 0.000 min

Lab File: BP017110.D

Acq: 11 Sep 2023 18:47




Tgt Ion:266 Resp: 317339


Ion Ratio Lower Upper

266 100

264 63.4 49.8 74.8

268 64.5 50.9 76.3





#72

Phenanthrene

Concen: 40.962 ng/ul

RT: 17.428 min Scan# 2439

Delta R.T. -0.006 min

Lab File: BP017110.D

Acq: 11 Sep 2023 18:47

Instrument :

BNA_P

ClientSampleId :

SLCS343





Tgt Ion:178 Resp: 3301586

Ion Ratio Lower Upper

178 100

179 15.7 12.3 18.5

176 20.0 15.8 23.8



#74

Anthracene

Concen: 40.208 ng/ul

RT: 17.522 min Scan# 2454

Delta R.T. -0.006 min

Lab File: BP017110.D

Acq: 11 Sep 2023 18:47




Tgt Ion:178 Resp: 3276545


Ion Ratio Lower Upper

178 100

179 16.1 12.2 18.2

176 19.5 15.4 23.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 3.487 ng/uL
RT: 13.404 min Scan# 1754
Delta R.T. 0.000 min
Lab File: BP017110.D
Acq: 11 Sep 2023 18:47




Instrument :
BNA_P
ClientSampleId :
SLCS343


Tgt Ion:216 Resp: 68164
Ion Ratio Lower Upper
216 100
214 77.6 63.8 95.6
218 49.9 38.1 57.1



#76
Pentachlorobenzene
Concen: 41.645 ng/uL
RT: 14.910 min Scan# 2010
Delta R.T. 0.000 min
Lab File: BP017110.D
Acq: 11 Sep 2023 18:47

Tgt Ion:250 Resp: 701418
Ion Ratio Lower Upper
250 100
248 63.6 49.9 74.9
252 63.0 51.4 77.0









#77
 Carbazole
 Concen: 42.724 ng/ul
 RT: 17.810 min Scan# 2503
 Delta R.T. 0.000 min
 Lab File: BP017110.D
 Acq: 11 Sep 2023 18:47




Instrument :
 BNA_P
 ClientSampleId :
 SLCS343


Tgt Ion:167 Resp: 3173281
 Ion Ratio Lower Upper
 167 100
 166 22.2 17.7 26.5
 139 13.4 10.8 16.2



#78
 Di-n-butylphthalate
 Concen: 43.546 ng/ul
 RT: 18.316 min Scan# 2589
 Delta R.T. -0.006 min
 Lab File: BP017110.D
 Acq: 11 Sep 2023 18:47

Tgt Ion:149 Resp: 3729196
 Ion Ratio Lower Upper
 149 100
 150 9.5 7.3 10.9
 104 5.8 4.7 7.1





#80

Fluoranthene

Concen: 42.844 ng/ul

RT: 19.398 min Scan# 2773

Delta R.T. 0.000 min

Lab File: BP017110.D

Acq: 11 Sep 2023 18:47

Instrument :

BNA_P

ClientSampleId :

SLCS343





Tgt Ion:202 Resp: 3787218

Ion Ratio Lower Upper

202 100

101 12.9 9.9 14.9

100 10.1 7.8 11.8



#82

Pyrene

Concen: 42.166 ng/ul

RT: 19.757 min Scan# 2834

Delta R.T. 0.000 min

Lab File: BP017110.D

Acq: 11 Sep 2023 18:47




Tgt Ion:202 Resp: 3980983


Ion Ratio Lower Upper

202 100

101 15.0 11.9 17.9

100 12.7 9.8 14.6





#83

Butylbenzylphthalate

Concen: 45.313 ng/ul

RT: 20.616 min Scan# 2980

Delta R.T. 0.000 min

Lab File: BP017110.D

Acq: 11 Sep 2023 18:47

Instrument :

BNA_P

ClientSampleId :

SLCS343





Tgt Ion:149 Resp: 1760166

Ion Ratio Lower Upper

149 100

91 76.6 64.6 96.8

206 18.7 14.6 22.0



#84

3,3'-Dichlorobenzidine

Concen: 42.769 ng/ul

RT: 21.410 min Scan# 3115

Delta R.T. 0.000 min

Lab File: BP017110.D

Acq: 11 Sep 2023 18:47




Tgt Ion:252 Resp: 1215343


Ion Ratio Lower Upper

252 100

254 63.8 51.7 77.5

126 13.3 11.5 17.3





#85

Benzo(a)anthracene

Concen: 41.472 ng/ul

RT: 21.469 min Scan# 3125

Delta R.T. 0.000 min

Lab File: BP017110.D

Acq: 11 Sep 2023 18:47

Instrument :

BNA_P

ClientSampleId :

SLCS343





Tgt Ion:228 Resp: 3868501

Ion Ratio Lower Upper

228 100

229 20.2 16.0 24.0

226 28.7 22.9 34.3



#86

Bis(2-ethylhexyl)phthalate

Concen: 45.190 ng/ul

RT: 21.345 min Scan# 3104

Delta R.T. -0.006 min

Lab File: BP017110.D

Acq: 11 Sep 2023 18:47




Tgt Ion:149 Resp: 2512223


Ion Ratio Lower Upper

149 100

167 25.7 20.8 31.2

279 2.9 2.6 3.8





#87

Chrysene

Concen: 41.934 ng/ul

RT: 21.527 min Scan# 3135

Delta R.T. 0.000 min

Lab File: BP017110.D

Acq: 11 Sep 2023 18:47

Instrument :

BNA_P

ClientSampleId :

SLCS343





Tgt Ion:228 Resp: 3587256

Ion Ratio Lower Upper

228 100

226 31.9 24.6 36.8

229 20.1 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 42.894 ng/ul




RT: 22.316 min Scan# 3269


Delta R.T. 0.000 min

Lab File: BP017110.D

Acq: 11 Sep 2023 18:47

Tgt Ion:149 Resp: 4357354









#90
Benzo(b)fluoranthene
Concen: 44.626 ng/ul
RT: 23.227 min Scan# 3424
Delta R.T. 0.000 min
Lab File: BP017110.D
Acq: 11 Sep 2023 18:47




Instrument :
BNA_P
ClientSampleId :
SLCS343


Tgt Ion:252 Resp: 3712130
Ion Ratio Lower Upper
252 100
253 22.3 17.9 26.9
125 12.6 10.3 15.5



#91
Benzo(k)fluoranthene
Concen: 43.147 ng/ul
RT: 23.286 min Scan# 3434
Delta R.T. 0.006 min
Lab File: BP017110.D
Acq: 11 Sep 2023 18:47

Tgt Ion:252 Resp: 3608812
Ion Ratio Lower Upper
252 100
253 22.4 17.7 26.5
125 11.9 10.0 15.0





#93

Benzo(a)pyrene

Concen: 42.004 ng/ul

RT: 23.904 min Scan# 3

Delta R.T. 0.000 min

Lab File: BP017110.D

Acq: 11 Sep 2023 18:47

Instrument :

BNA_P

ClientSampleId :

SLCS343





Tgt Ion:252 Resp: 3248020

Ion Ratio Lower Upper

252 100

253 22.0 17.6 26.4

125 14.1 11.3 16.9



#94

Indeno(1,2,3-cd)pyrene

Concen: 45.447 ng/ul

RT: 26.704 min Scan# 4015

Delta R.T. 0.006 min

Lab File: BP017110.D

Acq: 11 Sep 2023 18:47




Tgt Ion:276 Resp: 3947325


Ion Ratio Lower Upper

276 100

138 30.9 23.4 35.2

277 25.2 20.1 30.1





#95

Dibenzo(a,h)anthracene

Concen: 45.589 ng/ul

RT: 26.721 min Scan# 4018

Delta R.T. 0.000 min

Lab File: BP017110.D

Acq: 11 Sep 2023 18:47

Instrument :

BNA_P

ClientSampleId :

SLCS343





Tgt Ion:278 Resp: 3301427

Ion Ratio Lower Upper

278 100

139 20.7 16.5 24.7

279 24.2 19.2 28.8



#96

Benzo(g,h,i)perylene

Concen: 46.554 ng/ul

RT: 27.545 min Scan# 4158

Delta R.T. 0.006 min

Lab File: BP017110.D

Acq: 11 Sep 2023 18:47

Tgt Ion:276 Resp: 3081775

Ion Ratio Lower Upper

276 100

138 27.4 22.0 33.0

277 24.1 18.7 28.1

