

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091023\
 Data File : BP017103.D
 Acq On : 11 Sep 2023 14:14
 Operator : MA/JU
 Sample : 04268-05
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 H0GH3

Quant Time: Sep 12 02:46:05 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP091023.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Sep 11 06:13:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.957	152	320885	20.000	ng/u1	0.00
20) Naphthalene-d8	10.787	136	1434717	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.616	164	883523	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.386	188	1919205	20.000	ng/u1	0.00
79) Chrysene-d12	21.486	240	1666337	20.000	ng/u1	0.00
88) Perylene-d12	24.015	264	1739948	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.316	96	40027	4.761	ng/uL	0.00
4) Pyridine-d5	3.757	84	250729	9.890	ng/u1	0.00
7) Phenol-d5	7.110	99	190637	6.536	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.298	67	633237	32.369	ng/u1	0.00
11) 2-Chlorophenol-d4	7.481	132	518302	24.585	ng/u1	0.00
15) 4-Methylphenol-d8	8.669	113	357865	15.509	ng/u1	0.00
21) Nitrobenzene-d5	9.157	128	340251	33.311	ng/u1	0.00
24) 2-Nitrophenol-d4	9.869	143	347365	32.950	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.392	165	574033	28.716	ng/u1	0.00
31) 4-Chloroaniline-d4	10.945	131	790206	24.652	ng/u1	0.00
46) Dimethylphthalate-d6	14.033	166	2268441	36.083	ng/u1	0.00
49) Acenaphthylene-d8	14.316	160	2438268	34.061	ng/u1	0.00
54) 4-Nitrophenol-d4	14.833	143	70545	5.439	ng/u1	0.00
60) Fluorene-d10	15.616	176	1935911	36.288	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.739	200	318985	29.452	ng/u1	0.00
73) Anthracene-d10	17.486	188	3221230	39.157	ng/u1	0.00
81) Pyrene-d10	19.727	212	3809528	43.412	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.851	264	3394519	40.639	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091023\
Data File : BP017103.D
Acq On : 11 Sep 2023 14:14
Operator : MA/JU
Sample : 04268-05
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
H0GH3

Quant Time: Sep 12 02:46:05 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP091023.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Sep 11 06:13:29 2023
Response via : Initial Calibration

