

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091223\
 Data File : BP017183.D
 Acq On : 14 Sep 2023 18:40
 Operator : MA/JU
 Sample : AR1242-50PPM
 Misc : GCMS CONFIRMATION
 ALS Vial : 72 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 AR1242-50PPM

Quant Time: Sep 14 22:49:37 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP091223.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Sep 12 23:27:27 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.946	152	242221	20.000	ng/ul	-0.01
20) Naphthalene-d8	10.775	136	922784	20.000	ng/ul	-0.01
38) Acenaphthene-d10	14.610	164	545063	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.381	188	1257629	20.000	ng/ul	0.00
79) Chrysene-d12	21.480	240	899031	20.000	ng/ul	0.00
88) Perylene-d12	24.004	264	853950	20.000	ng/ul	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.317	96	340	0.059	ng/uL	0.00
4) Pyridine-d5	3.764	84	150	0.009	ng/ul	0.01
7) Phenol-d5	7.105	99	6	0.000	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.287	67	217	0.018	ng/ul	-0.01
11) 2-Chlorophenol-d4	7.458	132	58	0.004	ng/ul	-0.02
15) 4-Methylphenol-d8	8.652	113	13	0.001	ng/ul	-0.02
21) Nitrobenzene-d5	9.152	128	44	0.007	ng/ul	0.00
24) 2-Nitrophenol-d4	9.875	143	81	0.012	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.393	165	161	0.012	ng/ul	0.00
31) 4-Chloroaniline-d4	10.934	131	13	0.001	ng/ul	-0.01
46) Dimethylphthalate-d6	14.034	166	159	0.004	ng/ul	0.00
49) Acenaphthylene-d8	14.304	160	172	0.004	ng/ul	-0.01
54) 4-Nitrophenol-d4	14.828	143	18	0.002	ng/ul	0.00
60) Fluorene-d10	15.604	176	817	0.024	ng/ul	-0.01
65) 4,6-Dinitro-2-methylph...	15.734	200	24	0.003	ng/ul	-0.01
73) Anthracene-d10	17.481	188	3947	0.074	ng/ul	0.00
81) Pyrene-d10	19.716	212	1356	0.029	ng/ul	-0.01
92) Benzo(a)pyrene-d12	23.833	264	1816	0.045	ng/ul	-0.02

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091223\
Data File : BP017183.D
Acq On : 14 Sep 2023 18:40
Operator : MA/JU
Sample : AR1242-50PPM
Misc : GCMS CONFIRMATION
ALS Vial : 72 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
AR1242-50PPM

Quant Time: Sep 14 22:49:37 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP091223.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Sep 12 23:27:27 2023
Response via : Initial Calibration

