

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091223\
 Data File : BP017133.D
 Acq On : 13 Sep 2023 01:18
 Operator : MA/JU
 Sample : 04247-15
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 BBJ55

Quant Time: Sep 13 02:09:20 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP091223.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Sep 12 23:27:27 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.957	152	416124	20.000	ng/u1	0.00
20) Naphthalene-d8	10.781	136	1951991	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.616	164	1337245	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.381	188	3063456	20.000	ng/u1	0.00
79) Chrysene-d12	21.480	240	2749596	20.000	ng/u1	0.00
88) Perylene-d12	24.010	264	2723801	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.317	96	44689	4.513	ng/uL	0.00
4) Pyridine-d5	3.752	84	176334	6.225	ng/u1	0.00
7) Phenol-d5	7.105	99	233412	6.898	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.299	67	589297	27.865	ng/u1	0.00
11) 2-Chlorophenol-d4	7.475	132	625371	23.168	ng/u1	0.00
15) 4-Methylphenol-d8	8.663	113	437136	15.588	ng/u1	0.00
21) Nitrobenzene-d5	9.157	128	382091	27.954	ng/u1	0.00
24) 2-Nitrophenol-d4	9.869	143	425674	29.172	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.393	165	778875	27.600	ng/u1	0.00
31) 4-Chloroaniline-d4	10.940	131	815385	19.170	ng/u1	0.00
46) Dimethylphthalate-d6	14.028	166	2759481	28.998	ng/u1	0.00
49) Acenaphthylene-d8	14.316	160	3015114	27.923	ng/u1	0.00
54) 4-Nitrophenol-d4	14.834	143	90529	4.922	ng/u1	0.00
60) Fluorene-d10	15.610	176	2356059	28.634	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.739	200	438191	25.602	ng/u1	0.00
73) Anthracene-d10	17.486	188	3704334	28.460	ng/u1	0.00
81) Pyrene-d10	19.722	212	4552811	31.572	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.845	264	3911283	30.424	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091223\
Data File : BP017133.D
Acq On : 13 Sep 2023 01:18
Operator : MA/JU
Sample : 04247-15
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
BBJ55

Quant Time: Sep 13 02:09:20 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP091223.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Sep 12 23:27:27 2023
Response via : Initial Calibration

