

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091223\
 Data File : BP017171.D
 Acq On : 14 Sep 2023 11:42
 Operator : MA/JU
 Sample : 04324-20
 Misc :
 ALS Vial : 60 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 BBJA8

Quant Time: Sep 14 23:11:01 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP091223.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Sep 12 23:27:27 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.951	152	196240	20.000	ng/u1	0.00
20) Naphthalene-d8	10.775	136	788952	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.610	164	456214	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.380	188	954407	20.000	ng/u1	0.00
79) Chrysene-d12	21.480	240	836723	20.000	ng/u1	0.00
88) Perylene-d12	24.009	264	966098	20.000	ng/u1	-0.01

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.310	96	24975	5.348	ng/uL	0.00
4) Pyridine-d5	3.752	84	167875	12.567	ng/u1	0.00
7) Phenol-d5	7.104	99	123563	7.744	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.293	67	391327	39.237	ng/u1	0.00
11) 2-Chlorophenol-d4	7.469	132	374648	29.431	ng/u1	0.00
15) 4-Methylphenol-d8	8.657	113	238442	18.030	ng/u1	-0.01
21) Nitrobenzene-d5	9.151	128	233099	42.194	ng/u1	0.00
24) 2-Nitrophenol-d4	9.863	143	239931	40.681	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.387	165	410753	36.012	ng/u1	0.00
31) 4-Chloroaniline-d4	10.939	131	577180	33.573	ng/u1	0.00
46) Dimethylphthalate-d6	14.028	166	1413656	43.544	ng/u1	0.00
49) Acenaphthylene-d8	14.310	160	1565946	42.509	ng/u1	0.00
54) 4-Nitrophenol-d4	14.833	143	27866	4.441	ng/u1	0.00
60) Fluorene-d10	15.610	176	1223135	43.573	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.733	200	165583	31.053	ng/u1	-0.01
73) Anthracene-d10	17.480	188	1870562	46.129	ng/u1	0.00
81) Pyrene-d10	19.721	212	2130932	48.561	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.845	264	2097777	46.006	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091223\
Data File : BP017171.D
Acq On : 14 Sep 2023 11:42
Operator : MA/JU
Sample : 04324-20
Misc :
ALS Vial : 60 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
BBJA8

Quant Time: Sep 14 23:11:01 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP091223.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Sep 12 23:27:27 2023
Response via : Initial Calibration

