




Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091225\
Data File : BP025736.D
Acq On : 12 Sep 2025 09:03
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 12 11:24:47 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 11 16:45:04 2025
Response via : Initial Calibration



TIC: BP025736.D\data.ms

(70) Pentachlorophenol (C)




16.830min (-0.000) 27880.92 ng

response 1551454

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.30	64.05
264.00	62.70	62.95
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091225\
Data File : BP025736.D
Acq On : 12 Sep 2025 09:03
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 12 11:24:47 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 11 16:45:04 2025
Response via : Initial Calibration



TIC: BP025736.D\data.ms

(77) Benzidine

19.519min (+ 0.012) 6627.59 ng

response 8052586

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.40	14.47
183.00	11.50	12.21
0.00	0.00	0.00