

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091225\
 Data File : BP025737.D
 Acq On : 12 Sep 2025 10:24
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 12 11:25:26 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 11 16:45:04 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	106	0.00
2	1,4-Dioxane	40.000	42.942	-7.4	108	0.00
3	Pyridine	40.000	43.122	-7.8	104	0.00
4	n-Nitrosodimethylamine	40.000	42.612	-6.5	106	0.00
5 S	2-Fluorophenol	80.000	87.844	-9.8	106	0.00
6	Aniline	40.000	43.873	-9.7	105	0.00
7 S	Phenol-d6	80.000	89.501	-11.9	105	0.00
8	2-Chlorophenol	40.000	44.106	-10.3	105	0.00
9	Benzaldehyde	40.000	35.372	11.6	105	0.00
10 C	Phenol	40.000	44.572	-11.4	104	0.00
11	bis(2-Chloroethyl)ether	40.000	43.403	-8.5	103	0.00
12	1,3-Dichlorobenzene	40.000	43.026	-7.6	105	0.00
13 C	1,4-Dichlorobenzene	40.000	42.711	-6.8	105	0.00
14	1,2-Dichlorobenzene	40.000	43.016	-7.5	105	0.00
15	Benzyl Alcohol	40.000	43.673	-9.2	104	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	43.612	-9.0	106	0.00
17	2-Methylphenol	40.000	45.600	-14.0	107	0.00
18	Hexachloroethane	40.000	42.297	-5.7	104	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	45.455	-13.6	105	0.00
20	3+4-Methylphenols	40.000	43.906	-9.8	105	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	103	0.00
22	Acetophenone	40.000	43.894	-9.7	103	0.00
23 S	Nitrobenzene-d5	80.000	87.241	-9.1	103	0.00
24	Nitrobenzene	40.000	44.086	-10.2	103	0.00
25	Isophorone	40.000	43.506	-8.8	104	0.00
26 C	2-Nitrophenol	40.000	45.163	-12.9	104	0.00
27	2,4-Dimethylphenol	40.000	45.186	-13.0	104	0.00
28	bis(2-Chloroethoxy)methane	40.000	44.058	-10.1	105	0.00
29 C	2,4-Dichlorophenol	40.000	45.130	-12.8	106	0.00
30	1,2,4-Trichlorobenzene	40.000	43.201	-8.0	105	0.00
31	Naphthalene	40.000	43.234	-8.1	105	0.00
32	Benzoic acid	40.000	44.242	-10.6	103	0.00
33	4-Chloroaniline	40.000	45.998	-15.0	105	0.00
34 C	Hexachlorobutadiene	40.000	42.887	-7.2	105	0.00
35	Caprolactam	40.000	43.264	-8.2	103	0.00
36 C	4-Chloro-3-methylphenol	40.000	44.229	-10.6	104	0.00
37	2-Methylnaphthalene	40.000	43.039	-7.6	103	0.00
38	1-Methylnaphthalene	40.000	42.903	-7.3	102	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	104	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	43.497	-8.7	104	0.00
41 P	Hexachlorocyclopentadiene	40.000	44.712	-11.8	101	0.00
42 S	2,4,6-Tribromophenol	80.000	87.671	-9.6	103	-0.01
43 C	2,4,6-Trichlorophenol	40.000	45.065	-12.7	105	0.00
44	2,4,5-Trichlorophenol	40.000	44.427	-11.1	103	0.00
45 S	2-Fluorobiphenyl	80.000	85.723	-7.2	105	0.00
46	1,1'-Biphenyl	40.000	43.304	-8.3	104	0.00
47	2-Chloronaphthalene	40.000	43.638	-9.1	106	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091225\
 Data File : BP025737.D
 Acq On : 12 Sep 2025 10:24
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 12 11:25:26 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 11 16:45:04 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	45.310	-13.3	104	0.00
49	Acenaphthylene	40.000	43.541	-8.9	106	0.00
50	Dimethylphthalate	40.000	42.893	-7.2	103	0.00
51	2,6-Dinitrotoluene	40.000	44.614	-11.5	105	-0.01
52 C	Acenaphthene	40.000	42.862	-7.2	103	-0.01
53	3-Nitroaniline	40.000	46.836	-17.1	106	0.00
54 P	2,4-Dinitrophenol	40.000	43.026	-7.6	107	0.00
55	Dibenzofuran	40.000	42.798	-7.0	105	0.00
56 P	4-Nitrophenol	40.000	44.163	-10.4	105	0.00
57	2,4-Dinitrotoluene	40.000	45.490	-13.7	105	0.00
58	Fluorene	40.000	43.084	-7.7	104	-0.01
59	2,3,4,6-Tetrachlorophenol	40.000	44.512	-11.3	104	0.00
60	Diethylphthalate	40.000	43.138	-7.8	106	0.00
61	4-Chlorophenyl-phenylether	40.000	42.760	-6.9	105	-0.01
62	4-Nitroaniline	40.000	46.803	-17.0	105	-0.01
63	Azobenzene	40.000	43.637	-9.1	104	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	105	0.00
65	4,6-Dinitro-2-methylphenol	40.000	46.254	-15.6	108	0.00
66 c	n-Nitrosodiphenylamine	40.000	42.435	-6.1	103	0.00
67	4-Bromophenyl-phenylether	40.000	42.771	-6.9	104	0.00
68	Hexachlorobenzene	40.000	43.167	-7.9	107	0.00
69	Atrazine	40.000	43.353	-8.4	103	0.00
70 C	Pentachlorophenol	40.000	44.247	-10.6	105	0.00
71	Phenanthrene	40.000	42.464	-6.2	105	0.00
72	Anthracene	40.000	43.065	-7.7	106	0.00
73	Carbazole	40.000	43.543	-8.9	105	-0.01
74	Di-n-butylphthalate	40.000	43.638	-9.1	105	0.00
75 C	Fluoranthene	40.000	42.908	-7.3	106	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	106	-0.02
77	Benzydine	40.000	45.453	-13.6	106	-0.01
78	Pyrene	40.000	42.472	-6.2	105	0.00
79 S	Terphenyl-d14	80.000	82.959	-3.7	103	-0.02
80	Butylbenzylphthalate	40.000	44.756	-11.9	107	-0.02
81	Benzo(a)anthracene	40.000	42.589	-6.5	106	-0.02
82	3,3'-Dichlorobenzidine	40.000	44.439	-11.1	107	-0.02
83	Chrysene	40.000	42.169	-5.4	106	-0.02
84	Bis(2-ethylhexyl)phthalate	40.000	43.888	-9.7	105	-0.02
85 c	Di-n-octyl phthalate	40.000	43.636	-9.1	105	-0.02
86 I	Perylene-d12	20.000	20.000	0.0	104	-0.02
87	Indeno(1,2,3-cd)pyrene	40.000	43.950	-9.9	104	-0.04
88	Benzo(b)fluoranthene	40.000	44.923	-12.3	107	-0.01
89	Benzo(k)fluoranthene	40.000	43.318	-8.3	105	-0.01
90 C	Benzo(a)pyrene	40.000	44.027	-10.1	105	-0.01
91	Dibenzo(a,h)anthracene	40.000	43.854	-9.6	105	-0.04
92	Benzo(g,h,i)perylene	40.000	43.863	-9.7	106	-0.03

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091225\
Data File : BP025737.D
Acq On : 12 Sep 2025 10:24
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Sep 12 11:25:26 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 11 16:45:04 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0