

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091520\
 Data File : BP003287.D
 Acq On : 15 Sep 2020 13:36
 Operator : CG/JU
 Sample : SSTDICC05
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC005

Quant Time: Sep 15 16:27:12 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP091520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 15 16:17:32 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.628	152	70414	20.00	ng	0.00
21) Naphthalene-d8	10.404	136	275811	20.00	ng	0.00
39) Acenaphthene-d10	14.269	164	186194	20.00	ng	0.00
64) Phenanthrene-d10	17.028	188	422767	20.00	ng	0.00
76) Chrysene-d12	21.122	240	516379	20.00	ng	0.00
86) Perylene-d12	23.351	264	564746	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.240	112	42375	10.64	ng	0.00
7) Phenol-d6	6.822	99	55389	11.09	ng	0.00
23) Nitrobenzene-d5	8.775	82	48189	12.60	ng	0.00
42) 2,4,6-Tribromophenol	15.775	330	27220	10.81	ng	0.00
45) 2-Fluorobiphenyl	12.887	172	140455	11.05	ng	0.00
79) Terphenyl-d14	19.604	244	265406	11.31	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.164	88	9597	6.30	ng	93
3) Pyridine	3.558	79	23315	5.82	ng	91
4) n-Nitrosodimethylamine	3.470	42	6622	6.33	ng	# 95
6) Aniline	6.958	93	31831	5.86	ng	96
8) 2-Chlorophenol	7.199	128	23579	5.31	ng	91
9) Benzaldehyde	6.769	77	16394	7.04	ng	91
10) Phenol	6.852	94	26757	5.78	ng	88
11) bis(2-Chloroethyl)ether	7.058	93	22338	6.10	ng	98
12) 1,3-Dichlorobenzene	7.663	146	29515	5.47	ng	91
13) 1,4-Dichlorobenzene	7.663	146	29515	5.42	ng	93
14) 1,2-Dichlorobenzene	7.975	146	28466	5.51	ng	96
15) Benzyl Alcohol	7.875	79	16420	5.84	ng	89
16) 2,2'-oxybis(1-Chloropr...	8.163	45	16064	5.18	ng	81
17) 2-Methylphenol	8.093	107	19642	5.96	ng	95
18) Hexachloroethane	8.693	117	11148	6.25	ng	95
19) n-Nitroso-di-n-propyla...	8.434	70	15664	6.94	ng	95
20) 3+4-Methylphenols	8.428	107	24613	5.54	ng	# 74
22) Acetophenone	8.446	105	35136	5.72	ng	# 93
24) Nitrobenzene	8.816	77	23893	6.36	ng	89
25) Isophorone	9.340	82	44277	6.56	ng	# 91
26) 2-Nitrophenol	9.534	139	11947	5.33	ng	# 81
27) 2,4-Dimethylphenol	9.616	122	18086	5.37	ng	91
28) bis(2-Chloroethoxy)met...	9.834	93	29684	5.90	ng	97
29) 2,4-Dichlorophenol	10.099	162	22592	5.41	ng	94
30) 1,2,4-Trichlorobenzene	10.275	180	28253	5.68	ng	99
31) Naphthalene	10.452	128	77678	5.48	ng	100
32) Benzoic acid	9.781	122	880	0.40	ng	92
33) 4-Chloroaniline	10.581	127	30050	5.08	ng	# 93
34) Hexachlorobutadiene	10.752	225	19452	6.52	ng	97
35) Caprolactam	11.340	113	7316	5.83	ng	95
36) 4-Chloro-3-methylphenol	11.734	107	24775	6.35	ng	86
37) 2-Methylnaphthalene	12.081	142	56670	5.59	ng	95
38) 1-Methylnaphthalene	12.304	142	53939	5.60	ng	99
40) 1,2,4,5-Tetrachloroben...	12.457	216	31713	5.75	ng	98
41) Hexachlorocyclopentadiene	12.440	237	938	0.36	ng	88
43) 2,4,6-Trichlorophenol	12.716	196	19275	5.31	ng	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091520\
 Data File : BP003287.D
 Acq On : 15 Sep 2020 13:36
 Operator : CG/JU
 Sample : SSTDICC05
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC005


Quant Time: Sep 15 16:27:12 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP091520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 15 16:17:32 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.804	196	18633	4.85	ng	# 96
46) 1,1'-Biphenyl	13.104	154	74657	5.35	ng	99
47) 2-Chloronaphthalene	13.140	162	61156	5.35	ng	96
48) 2-Nitroaniline	13.357	65	13620	6.36	ng	# 80
49) Acenaphthylene	13.993	152	92130	5.33	ng	99
50) Dimethylphthalate	13.734	163	79412	5.59	ng	99
51) 2,6-Dinitrotoluene	13.851	165	16205	5.49	ng	# 89
52) Acenaphthene	14.334	154	57587	5.42	ng	97
53) 3-Nitroaniline	14.193	138	15804	4.73	ng	# 77
54) 2,4-Dinitrophenol	14.451	184	1666	1.26	ng	# 76
55) Dibenzofuran	14.675	168	92711	5.56	ng	95
56) 4-Nitrophenol	14.587	139	4835	2.01	ng	# 68
57) 2,4-Dinitrotoluene	14.657	165	20411	5.14	ng	# 72
58) Fluorene	15.328	166	74402	5.79	ng	99
59) 2,3,4,6-Tetrachlorophenol	14.840	232	14354	4.13	ng	93
60) Diethylphthalate	15.110	149	81612	5.86	ng	99
61) 4-Chlorophenyl-phenyle...	15.328	204	41101	6.19	ng	98
62) 4-Nitroaniline	15.363	138	16359	4.76	ng	# 72
63) Azobenzene	15.616	77	61266	6.63	ng	93
65) 4,6-Dinitro-2-methylph...	15.440	198	6162	3.05	ng	91
66) n-Nitrosodiphenylamine	15.540	169	66307	5.36	ng	98
67) 4-Bromophenyl-phenylether	16.222	248	26987	5.78	ng	98
68) Hexachlorobenzene	16.345	284	32158	5.79	ng	99
69) Atrazine	16.504	200	25274	6.06	ng	97
70) Pentachlorophenol	16.710	266	5555	2.09	ng	94
71) Phenanthrene	17.069	178	125374	5.47	ng	98
72) Anthracene	17.163	178	121280	5.53	ng	99
73) Carbazole	17.445	167	115067	5.47	ng	98
74) Di-n-butylphthalate	18.004	149	152010	6.15	ng	# 99
75) Fluoranthene	19.051	202	157115	5.93	ng	100
77) Benzidine	19.245	184	66703	5.65	ng	97
78) Pyrene	19.404	202	166615	4.97	ng	99
80) Butylbenzylphthalate	20.280	149	75354	5.46	ng	88
81) Benzo(a)anthracene	21.110	228	175881	5.44	ng	98
82) 3,3'-Dichlorobenzidine	21.051	252	65489	5.50	ng	99
83) Chrysene	21.163	228	171261	5.53	ng	98
84) Bis(2-ethylhexyl)phtha...	21.051	149	115260	5.78	ng	98
85) Di-n-octyl phthalate	21.916	149	198836	5.82	ng	98
87) Indeno(1,2,3-cd)pyrene	25.615	276	226349	5.37	ng	97
88) Benzo(b)fluoranthene	22.680	252	185833	5.29	ng	99
89) Benzo(k)fluoranthene	22.727	252	182279	5.44	ng	100
90) Benzo(a)pyrene	23.251	252	176187	5.41	ng	99
91) Dibenzo(a,h)anthracene	25.621	278	190007	5.49	ng	99
92) Benzo(g,h,i)perylene	26.310	276	184513	5.31	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Sep 15 16:27:12 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP091520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 15 16:17:32 2020
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 20.00 ng
 RT: 7.628 min Scan# 806
 Delta R.T. 0.006 min
 Lab File: BP003287.D
 Acq: 15 Sep 2020 13:36




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC005


Tgt Ion:152 Resp: 70414
 Ion Ratio Lower Upper
 152 100
 150 153.8 123.8 185.6
 115 59.9 43.8 65.6



#2
 1,4-Dioxane
 Concen: 6.30 ng
 RT: 3.164 min Scan# 47
 Delta R.T. -0.000 min
 Lab File: BP003287.D
 Acq: 15 Sep 2020 13:36

Tgt Ion: 88 Resp: 9597
 Ion Ratio Lower Upper
 88 100
 58 49.9 44.2 66.4
 43 21.5 15.2 22.8









#3
Pyridine
Concen: 5.82 ng
RT: 3.558 min Scan# 114
Delta R.T. 0.006 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Instrument :
BNA_P
ClientSampleId :
SSTDIC005

Tgt Ion: 79 Resp: 23315




Ion	Ratio	Lower	Upper
79	100		
52	43.2	41.3	61.9
51	25.5	21.2	31.8




#4
n-Nitrosodimethylamine
Concen: 6.33 ng
RT: 3.470 min Scan# 99
Delta R.T. 0.006 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Tgt Ion: 42 Resp: 6622

Ion	Ratio	Lower	Upper
42	100		
74	179.6	148.6	223.0
44	12.2	5.3	7.9









#5
2-Fluorophenol
Concen: 10.64 ng
RT: 5.240 min Scan# 400
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:	112	Resp:	42375
Ion	Ratio	Lower	Upper
112	100		
64	47.2	32.8	49.2
63	25.2	17.0	25.4



#6
Aniline
Concen: 5.86 ng
RT: 6.958 min Scan# 692
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Tgt Ion:	93	Resp:	31831
Ion	Ratio	Lower	Upper
93	100		
66	34.3	25.4	38.0
65	16.4	12.5	18.7









#7
Phenol-d6
Concen: 11.09 ng
RT: 6.822 min Scan# 669
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion: 99	Resp: 55389
Ion Ratio	Lower Upper
99 100	
42 12.9	8.6 13.0
71 34.4	23.4 35.2



#8
2-Chlorophenol
Concen: 5.31 ng
RT: 7.199 min Scan# 733
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Tgt Ion: 128	Resp: 23579
Ion Ratio	Lower Upper
128 100	
130 34.2	13.0 53.0
64 39.5	10.4 50.4









#9
Benzaldehyde
Concen: 7.04 ng
RT: 6.769 min Scan# 660
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:	77	Resp:	16394
Ion	Ratio	Lower	Upper
77	100		
105	104.3	97.1	137.1
106	105.9	93.7	133.7



#10
Phenol
Concen: 5.78 ng
RT: 6.852 min Scan# 674
Delta R.T. 0.006 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Tgt Ion:	94	Resp:	26757
Ion	Ratio	Lower	Upper
94	100		
65	25.8	2.3	42.3
66	39.7	10.9	50.9









#11
bis(2-Chloroethyl)ether
Concen: 6.10 ng
RT: 7.058 min Scan# 709
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion: 93 Resp: 22338
Ion Ratio Lower Upper
93 100
63 60.6 39.0 79.0
95 31.3 12.8 52.8



#12
1,3-Dichlorobenzene
Concen: 5.47 ng
RT: 7.663 min Scan# 812
Delta R.T. 0.147 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Tgt Ion: 146 Resp: 29515
Ion Ratio Lower Upper
146 100
148 58.6 52.5 78.7
75 26.8 18.1 27.1









#13
1,4-Dichlorobenzene
Concen: 5.42 ng
RT: 7.663 min Scan# 812
Delta R.T. 0.006 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:	146	Resp:	29515
Ion	Ratio	Lower	Upper
146	100		
148	58.6	51.3	76.9
111	39.9	28.9	43.3



#14
1,2-Dichlorobenzene
Concen: 5.51 ng
RT: 7.975 min Scan# 865
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Tgt Ion:	146	Resp:	28466
Ion	Ratio	Lower	Upper
146	100		
148	66.0	51.6	77.4
111	43.2	31.0	46.6









#15
Benzyl Alcohol
Concen: 5.84 ng
RT: 7.875 min Scan# 848
Delta R.T. 0.006 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion: 79	Resp: 16420
Ion Ratio	Lower Upper
79 100	
108 75.8	73.5 110.3
77 64.8	50.3 75.5



#16
2,2'-oxybis(1-Chloropropane)
Concen: 5.18 ng
RT: 8.163 min Scan# 897
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Tgt Ion: 45	Resp: 16064
Ion Ratio	Lower Upper
45 100	
77 29.7	2.3 42.3
79 26.9	0.0 36.6









#17
2-Methylphenol
Concen: 5.96 ng
RT: 8.093 min Scan# 885
Delta R.T. 0.012 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:	107	Resp:	19642
Ion	Ratio	Lower	Upper
107	100		
108	109.2	89.6	134.4
77	42.6	31.8	47.6
79	45.5	31.1	46.7



#18
Hexachloroethane
Concen: 6.25 ng
RT: 8.693 min Scan# 987
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Tgt Ion:	117	Resp:	11148
Ion	Ratio	Lower	Upper
117	100		
119	97.9	78.8	118.2
201	105.9	92.3	138.5









#19
n-Nitroso-di-n-propylamine
Concen: 6.94 ng
RT: 8.434 min Scan# 943
Delta R.T. 0.006 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:	70	Resp:	15664
Ion	Ratio	Lower	Upper
70	100		
42	40.0	32.8	49.2
101	10.4	9.6	14.4
130	25.0	24.3	36.5



#20
3+4-Methylphenols
Concen: 5.54 ng
RT: 8.428 min Scan# 942
Delta R.T. 0.018 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Tgt Ion:	107	Resp:	24613
Ion	Ratio	Lower	Upper
107	100		
108	97.9	69.0	109.0
77	74.8	8.5	48.5#
79	33.3	4.8	44.8









#21
Naphthalene-d8
Concen: 20.00 ng
RT: 10.404 min Scan# 1278
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Instrument :
BNA_P
ClientSampleId :
SSTDIC005

Tgt Ion:136 Resp: 275811




Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.6	13.0
54	4.8	3.6	5.4
68	4.9	3.3	4.9




#22
Acetophenone
Concen: 5.72 ng
RT: 8.446 min Scan# 945
Delta R.T. 0.006 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Tgt Ion:105 Resp: 35136

Ion	Ratio	Lower	Upper
105	100		
71	4.4	2.6	3.8#
51	21.6	13.5	20.3#
120	22.5	19.6	29.4









#23
Nitrobenzene-d5
Concen: 12.60 ng
RT: 8.775 min Scan# 1001
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:	82	Resp:	48189
Ion	Ratio	Lower	Upper
82	100		
128	44.0	45.6	68.4#
54	36.5	30.1	45.1



#24
Nitrobenzene
Concen: 6.36 ng
RT: 8.816 min Scan# 1008
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Tgt Ion:	77	Resp:	23893
Ion	Ratio	Lower	Upper
77	100		
123	50.5	48.6	72.8
65	12.7	10.0	15.0









#25
Isophorone
Concen: 6.56 ng
RT: 9.340 min Scan# 1097
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion: 82 Resp: 44277
Ion Ratio Lower Upper
82 100
95 7.6 6.4 9.6
138 17.5 18.7 28.1#



#26
2-Nitrophenol
Concen: 5.33 ng
RT: 9.534 min Scan# 1130
Delta R.T. 0.006 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Tgt Ion: 139 Resp: 11947
Ion Ratio Lower Upper
139 100
109 35.6 22.1 33.1#
65 45.9 26.5 39.7#









#27
2,4-Dimethylphenol
Concen: 5.37 ng
RT: 9.616 min Scan# 1144
Delta R.T. 0.012 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:	122	Resp:	18086
Ion	Ratio	Lower	Upper
122	100		
107	121.3	87.3	130.9
121	55.9	47.3	70.9



#28
bis(2-Chloroethoxy)methane
Concen: 5.90 ng
RT: 9.834 min Scan# 1181
Delta R.T. 0.006 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Tgt Ion:	93	Resp:	29684
Ion	Ratio	Lower	Upper
93	100		
95	32.1	26.7	40.1
123	11.3	10.5	15.7









#29
2,4-Dichlorophenol
Concen: 5.41 ng
RT: 10.099 min Scan# 1226
Delta R.T. 0.030 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:	162	Resp:	22592
Ion	Ratio	Lower	Upper
162	100		
164	70.1	45.8	85.8
98	33.3	10.2	50.2



#30
1,2,4-Trichlorobenzene
Concen: 5.68 ng
RT: 10.275 min Scan# 1256
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Tgt Ion:	180	Resp:	28253
Ion	Ratio	Lower	Upper
180	100		
182	94.1	76.2	114.2
145	29.4	23.8	35.8









#31
Naphthalene
Concen: 5.48 ng
RT: 10.452 min Scan# 1286
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:	128	Resp:	77678
Ion	Ratio	Lower	Upper
128	100		
129	11.1	8.8	13.2
127	13.4	10.6	15.8



#32
Benzoic acid
Concen: 0.40 ng
RT: 9.781 min Scan# 1172
Delta R.T. 0.006 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Tgt Ion:	122	Resp:	880
Ion	Ratio	Lower	Upper
122	100		
105	112.4	94.1	134.1
77	86.2	50.1	90.1









#33
4-Chloroaniline
Concen: 5.08 ng
RT: 10.581 min Scan# 1308
Delta R.T. 0.012 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:	127	Resp:	30050
Ion	Ratio	Lower	Upper
127	100		
129	32.1	25.7	38.5
65	27.0	17.3	25.9#
92	21.7	12.9	19.3#



#34
Hexachlorobutadiene
Concen: 6.52 ng
RT: 10.752 min Scan# 1337
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Tgt Ion:	225	Resp:	19452
Ion	Ratio	Lower	Upper
225	100		
223	66.4	50.1	75.1
227	65.0	51.4	77.2









#35
Caprolactam
Concen: 5.83 ng
RT: 11.340 min Scan# 1437
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:	113	Resp:	7316
Ion	Ratio	Lower	Upper
113	100		
55	102.4	77.3	117.3
56	80.7	55.9	95.9



#36
4-Chloro-3-methylphenol
Concen: 6.35 ng
RT: 11.734 min Scan# 1504
Delta R.T. 0.012 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Tgt Ion:	107	Resp:	24775
Ion	Ratio	Lower	Upper
107	100		
144	23.7	22.7	34.1
142	74.3	70.7	106.1









#37
2-Methylnaphthalene
Concen: 5.59 ng
RT: 12.081 min Scan# 1563
Delta R.T. 0.006 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36




Instrument :
BNA_P
ClientSampleId :
SSTDICC005


Tgt Ion:	142	Resp:	56670
Ion	Ratio	Lower	Upper
142	100		
141	85.9	71.4	107.2
115	37.6	25.8	38.6



#38
1-Methylnaphthalene
Concen: 5.60 ng
RT: 12.304 min Scan# 1601
Delta R.T. 0.006 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Tgt Ion:	142	Resp:	53939
Ion	Ratio	Lower	Upper
142	100		
141	92.3	74.4	111.6
116	3.9	2.9	4.3









#39
Acenaphthene-d10
Concen: 20.00 ng
RT: 14.269 min Scan# 1935
Delta R.T. -0.006 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:164 Resp: 186194
Ion Ratio Lower Upper
164 100
162 100.3 79.4 119.2
160 44.3 36.1 54.1



#40
1,2,4,5-Tetrachlorobenzene
Concen: 5.75 ng
RT: 12.457 min Scan# 1627
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Tgt Ion:216 Resp: 31713
Ion Ratio Lower Upper
216 100
214 80.1 63.0 94.6
179 22.2 16.9 25.3
108 15.6 12.6 19.0









#41
Hexachlorocyclopentadiene
Concen: 0.36 ng
RT: 12.440 min Scan# 1624
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:	237	Resp:	938
Ion	Ratio	Lower	Upper
237	100		
235	70.9	43.2	83.2
272	23.2	0.0	33.3



#42
2,4,6-Tribromophenol
Concen: 10.81 ng
RT: 15.775 min Scan# 2191
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Tgt Ion:	330	Resp:	27220
Ion	Ratio	Lower	Upper
330	100		
332	100.1	77.1	115.7
141	32.3	23.4	35.2









#43
2,4,6-Trichlorophenol
Concen: 5.31 ng
RT: 12.716 min Scan# 1671
Delta R.T. 0.012 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:196 Resp: 19275
Ion Ratio Lower Upper
196 100
198 94.9 77.0 115.6
200 28.0 25.1 37.7



#44
2,4,5-Trichlorophenol
Concen: 4.85 ng
RT: 12.804 min Scan# 1686
Delta R.T. 0.018 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Tgt Ion:196 Resp: 18633
Ion Ratio Lower Upper
196 100
198 95.6 77.3 115.9
97 48.3 31.0 46.4#
132 26.2 20.9 31.3









#45
2-Fluorobiphenyl
Concen: 11.05 ng
RT: 12.887 min Scan# 1700
Delta R.T. -0.006 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:	172	Resp:	140455
Ion	Ratio	Lower	Upper
172	100		
171	35.3	28.5	42.7
170	22.0	19.2	28.8



#46
1,1'-Biphenyl
Concen: 5.35 ng
RT: 13.104 min Scan# 1737
Delta R.T. 0.006 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Tgt Ion:	154	Resp:	74657
Ion	Ratio	Lower	Upper
154	100		
153	40.9	20.4	60.4
76	10.4	0.0	29.3









#47
2-Chloronaphthalene
Concen: 5.35 ng
RT: 13.140 min Scan# 1743
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:	162	Resp:	61156
Ion	Ratio	Lower	Upper
162	100		
127	40.6	29.9	44.9
164	34.2	26.1	39.1



#48
2-Nitroaniline
Concen: 6.36 ng
RT: 13.357 min Scan# 1780
Delta R.T. 0.006 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Tgt Ion:	65	Resp:	13620
Ion	Ratio	Lower	Upper
65	100		
92	85.5	70.6	106.0
138	128.4	133.6	200.4









#49
Acenaphthylene
Concen: 5.33 ng
RT: 13.993 min Scan# 1888
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:152	Resp:	92130
Ion Ratio	Lower	Upper
152	100	
151	20.3	15.9 23.9
153	13.7	10.5 15.7



#50
Dimethylphthalate
Concen: 5.59 ng
RT: 13.734 min Scan# 1844
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Tgt Ion:163	Resp:	79412
Ion Ratio	Lower	Upper
163	100	
194	4.3	3.3 4.9
164	10.4	8.2 12.2









#51
2,6-Dinitrotoluene
Concen: 5.49 ng
RT: 13.851 min Scan# 1864
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:	165	Resp:	16205
Ion Ratio	Lower	Upper	
165	100		
63	43.8	27.8	41.6#
89	48.1	34.6	52.0



#52
Acenaphthene
Concen: 5.42 ng
RT: 14.334 min Scan# 1946
Delta R.T. -0.006 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Tgt Ion:	154	Resp:	57587
Ion Ratio	Lower	Upper	
154	100		
153	111.1	91.0	136.6
152	51.3	43.0	64.6









#53
3-Nitroaniline
Concen: 4.73 ng
RT: 14.193 min Scan# 1922
Delta R.T. 0.012 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:138 Resp: 15804
Ion Ratio Lower Upper
138 100
108 12.9 7.5 11.3#
92 117.5 75.0 112.6#



#54
2,4-Dinitrophenol
Concen: 1.26 ng
RT: 14.451 min Scan# 1966
Delta R.T. 0.047 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Tgt Ion:184 Resp: 1666
Ion Ratio Lower Upper
184 100
63 45.9 32.3 48.5
154 87.9 49.1 73.7#





#55

Dibenzofuran

Concen: 5.56 ng

RT: 14.675 min Scan# 2004

Delta R.T. 0.000 min

Lab File: BP003287.D

Acq: 15 Sep 2020 13:36





Instrument :

BNA_P

ClientSampleId :

SSTDIC005

Tgt Ion:168	Resp:	92711
Ion	Ratio	Lower Upper
168	100	
139	39.7	29.1 43.7
169	14.4	10.4 15.6



#56

4-Nitrophenol

Concen: 2.01 ng




RT: 14.587 min Scan# 1989


Delta R.T. 0.059 min

Lab File: BP003287.D

Acq: 15 Sep 2020 13:36

Tgt Ion:139	Resp:	4835
Ion	Ratio	Lower Upper
139	100	
109	88.1	44.6 84.6#
65	88.1	41.6 81.6#









#57
2,4-Dinitrotoluene
Concen: 5.14 ng
RT: 14.657 min Scan# 2001
Delta R.T. 0.006 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:	165	Resp:	20411
Ion	Ratio	Lower	Upper
165	100		
63	43.2	20.6	31.0#
89	72.4	42.6	63.8#
182	3.4	2.4	3.6



#58
Fluorene
Concen: 5.79 ng
RT: 15.328 min Scan# 2115
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Tgt Ion:	166	Resp:	74402
Ion	Ratio	Lower	Upper
166	100		
165	97.8	77.8	116.8
167	13.3	11.0	16.4









#59
2,3,4,6-Tetrachlorophenol
Concen: 4.13 ng
RT: 14.840 min Scan# 2032
Delta R.T. -0.076 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:232	Resp:	14354
Ion Ratio	Lower	Upper
232	100	
131	39.2	34.1 51.1
130	2.4	1.8 2.8
166	25.3	24.6 36.8



#60
Diethylphthalate
Concen: 5.86 ng
RT: 15.110 min Scan# 2078
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Tgt Ion:149	Resp:	81612
Ion Ratio	Lower	Upper
149	100	
177	23.5	18.0 27.0
150	12.4	10.0 15.0









#61
4-Chlorophenyl-phenylether
Concen: 6.19 ng
RT: 15.328 min Scan# 2115
Delta R.T. 0.006 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:	204	Resp:	41101
Ion	Ratio	Lower	Upper
204	100		
206	33.5	25.9	38.9
141	56.9	46.5	69.7



#62
4-Nitroaniline
Concen: 4.76 ng
RT: 15.363 min Scan# 2121
Delta R.T. 0.012 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Tgt Ion:	138	Resp:	16359
Ion	Ratio	Lower	Upper
138	100		
92	57.2	20.7	60.7
108	91.3	47.8	87.8









#63
Azobenzene
Concen: 6.63 ng
RT: 15.616 min Scan# 2164
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Instrument :
BNA_P
ClientSampleId :
SSTDIC005

Tgt Ion: 77 Resp: 61266




Ion	Ratio	Lower	Upper
77	100		
182	29.3	16.9	56.9
105	22.8	5.6	45.6
51	22.8	2.9	42.9




#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 17.028 min Scan# 2404
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Tgt Ion: 188 Resp: 422767

Ion	Ratio	Lower	Upper
188	100		
94	7.6	6.2	9.2
80	8.3	6.2	9.2









#65
4,6-Dinitro-2-methylphenol
Concen: 3.05 ng
RT: 15.440 min Scan# 2134
Delta R.T. 0.018 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:198	Resp:	6162
Ion	Ratio	Lower Upper
198	100	
51	30.5	6.4 46.4
105	32.5	18.9 58.9



#66
n-Nitrosodiphenylamine
Concen: 5.36 ng
RT: 15.540 min Scan# 2151
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Tgt Ion:169	Resp:	66307
Ion	Ratio	Lower Upper
169	100	
168	65.4	53.8 80.6
167	36.0	28.9 43.3









#67
4-Bromophenyl-phenylether
Concen: 5.78 ng
RT: 16.222 min Scan# 2267
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:	248	Resp:	26987
Ion	Ratio	Lower	Upper
248	100		
250	96.5	76.7	115.1
141	69.9	53.0	79.6



#68
Hexachlorobenzene
Concen: 5.79 ng
RT: 16.345 min Scan# 2288
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Tgt Ion:	284	Resp:	32158
Ion	Ratio	Lower	Upper
284	100		
142	30.6	24.2	36.4
249	29.5	23.8	35.8









#69
Atrazine
Concen: 6.06 ng
RT: 16.504 min Scan# 2315
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:	200	Resp:	25274
Ion	Ratio	Lower	Upper
200	100		
173	22.9	5.9	45.9
215	47.7	28.4	68.4



#70
Pentachlorophenol
Concen: 2.09 ng
RT: 16.710 min Scan# 2350
Delta R.T. 0.018 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Tgt Ion:	266	Resp:	5555
Ion	Ratio	Lower	Upper
266	100		
268	56.9	51.1	76.7
264	61.5	50.6	76.0









#71
Phenanthrene
Concen: 5.47 ng
RT: 17.069 min Scan# 2411
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:178 Resp: 125374
Ion Ratio Lower Upper
178 100
176 20.1 15.3 22.9
179 14.4 12.5 18.7



#72
Anthracene
Concen: 5.53 ng
RT: 17.163 min Scan# 2427
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Tgt Ion:178 Resp: 121280
Ion Ratio Lower Upper
178 100
176 19.2 15.0 22.4
179 15.0 12.5 18.7









#73
 Carbazole
 Concen: 5.47 ng
 RT: 17.445 min Scan# 2475
 Delta R.T. 0.012 min
 Lab File: BP003287.D
 Acq: 15 Sep 2020 13:36




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC005


Tgt Ion:	167	Resp:	115067
Ion	Ratio	Lower	Upper
167	100		
166	21.2	17.6	26.4
139	14.0	10.2	15.4



#74
 Di-n-butylphthalate
 Concen: 6.15 ng
 RT: 18.004 min Scan# 2570
 Delta R.T. 0.000 min
 Lab File: BP003287.D
 Acq: 15 Sep 2020 13:36

Tgt Ion:	149	Resp:	152010
Ion	Ratio	Lower	Upper
149	100		
150	9.2	7.3	10.9
104	5.1	3.4	5.0#









#75
Fluoranthene
Concen: 5.93 ng
RT: 19.051 min Scan# 2748
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:	202	Resp:	157115
Ion	Ratio	Lower	Upper
202	100		
101	9.6	0.0	29.4
203	17.9	0.0	37.8



#76
Chrysene-d12
Concen: 20.00 ng
RT: 21.122 min Scan# 3100
Delta R.T. -0.006 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Tgt Ion:	240	Resp:	516379
Ion	Ratio	Lower	Upper
240	100		
120	10.2	7.8	11.6
236	25.7	20.8	31.2









#77
Benzidine
Concen: 5.65 ng
RT: 19.245 min Scan# 2781
Delta R.T. 0.012 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:184 Resp: 66703
Ion Ratio Lower Upper
184 100
185 12.6 11.5 17.3
183 11.9 9.6 14.4



#78
Pyrene
Concen: 4.97 ng
RT: 19.404 min Scan# 2808
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Tgt Ion:202 Resp: 166615
Ion Ratio Lower Upper
202 100
200 21.2 17.1 25.7
203 17.0 14.2 21.2









#79
 Terphenyl-d14
 Concen: 11.31 ng
 RT: 19.604 min Scan# 2842
 Delta R.T. 0.000 min
 Lab File: BP003287.D
 Acq: 15 Sep 2020 13:36




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC005


Tgt Ion:	244	Resp:	265406
Ion	Ratio	Lower	Upper
244	100		
212	7.2	5.8	8.6
122	9.7	7.8	11.8



#80
 Butylbenzylphthalate
 Concen: 5.46 ng
 RT: 20.280 min Scan# 2957
 Delta R.T. 0.000 min
 Lab File: BP003287.D
 Acq: 15 Sep 2020 13:36

Tgt Ion:	149	Resp:	75354
Ion	Ratio	Lower	Upper
149	100		
91	67.8	45.4	68.0
206	22.4	17.1	25.7









#81
Benzo(a)anthracene
Concen: 5.44 ng
RT: 21.110 min Scan# 3098
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Instrument :
BNA_P
ClientSampleId :
SSTDIC005

Tgt Ion:228 Resp: 175881




Ion	Ratio	Lower	Upper
228	100		
226	26.1	22.1	33.1
229	19.6	15.9	23.9




#82
3,3'-Dichlorobenzidine
Concen: 5.50 ng
RT: 21.051 min Scan# 3088
Delta R.T. 0.006 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Tgt Ion:252 Resp: 65489

Ion	Ratio	Lower	Upper
252	100		
254	64.5	51.0	76.4
126	9.2	7.9	11.9









#83
Chrysene
Concen: 5.53 ng
RT: 21.163 min Scan# 3107
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:228	Resp: 171261
Ion Ratio	Lower Upper
228 100	
226 29.7	24.4 36.6
229 18.9	15.8 23.6



#84
Bis(2-ethylhexyl)phthalate
Concen: 5.78 ng
RT: 21.051 min Scan# 3088
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Tgt Ion:149	Resp: 115260
Ion Ratio	Lower Upper
149 100	
167 29.7	22.6 34.0
279 5.5	4.2 6.2









#85
Di-n-octyl phthalate
Concen: 5.82 ng
RT: 21.916 min Scan# 3235
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Instrument :
BNA_P
ClientSampleId :
SSTDIC005

Tgt Ion:149 Resp: 198836




Ion	Ratio	Lower	Upper
149	100		
167	1.9	1.3	1.9
43	5.2	3.7	5.5




#86
Perylene-d12
Concen: 20.00 ng
RT: 23.351 min Scan# 3479
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Tgt Ion:264 Resp: 564746

Ion	Ratio	Lower	Upper
264	100		
260	24.1	18.8	28.2
265	21.3	17.4	26.2









#87
Indeno(1,2,3-cd)pyrene
Concen: 5.37 ng
RT: 25.615 min Scan# 3864
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion: 276 Resp: 226349
Ion Ratio Lower Upper
276 100
138 21.0 18.7 28.1
277 25.7 20.4 30.6



#88
Benzo(b)fluoranthene
Concen: 5.29 ng
RT: 22.680 min Scan# 3365
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Tgt Ion: 252 Resp: 185833
Ion Ratio Lower Upper
252 100
253 22.6 17.6 26.4
125 9.7 7.3 10.9









#89
Benzo(k)fluoranthene
Concen: 5.44 ng
RT: 22.727 min Scan# 3373
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:252 Resp: 182279
Ion Ratio Lower Upper
252 100
253 22.5 17.8 26.8
125 9.5 7.4 11.2



#90
Benzo(a)pyrene
Concen: 5.41 ng
RT: 23.251 min Scan# 3462
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Tgt Ion:252 Resp: 176187
Ion Ratio Lower Upper
252 100
253 21.7 17.4 26.0
125 10.8 8.2 12.4









#91
Dibenzo(a,h)anthracene
Concen: 5.49 ng
RT: 25.621 min Scan# 3865
Delta R.T. -0.006 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Instrument :
BNA_P
ClientSampleId :
SSTDIC005

Tgt Ion:	278	Resp:	190007
Ion	Ratio	Lower	Upper
278	100		
139	14.3	12.4	18.6
279	23.5	19.0	28.6



#92
Benzo(g,h,i)perylene
Concen: 5.31 ng
RT: 26.310 min Scan# 3982
Delta R.T. 0.000 min
Lab File: BP003287.D
Acq: 15 Sep 2020 13:36

Tgt Ion:	276	Resp:	184513
Ion	Ratio	Lower	Upper
276	100		
277	24.5	19.0	28.6
138	18.4	16.6	24.8

