

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP091619\
 Data File : BP000241.D
 Acq On : 16 Sep 2019 16:11
 Operator : HP/JU
 Sample : SSTDICC020
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020

Quant Time: Sep 16 17:24:48 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP091619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 16 17:15:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.79	152	356474	20.00	ng	0.00
21) Naphthalene-d8	8.07	136	1350824	20.00	ng	0.00
39) Acenaphthene-d10	9.83	164	706740	20.00	ng	0.00
64) Phenanthrene-d10	11.33	188	1244160	20.00	ng	0.00
76) Chrysene-d12	14.00	240	948008	20.00	ng	0.00
87) Perylene-d12	15.49	264	1090623	20.00	ng	0.00

System Monitoring Compounds

5) 2-Fluorophenol	5.40	112	889022	39.77	ng	0.00
7) Phenol-d6	6.42	99	1106409	39.21	ng	0.00
23) Nitrobenzene-d5	7.35	82	903450	41.04	ng	0.00
42) 2,4,6-Tribromophenol	10.63	330	295162	49.55	ng	0.00
45) 2-Fluorobiphenyl	9.15	172	1857293	43.89	ng	0.00
79) Terphenyl-d14	12.93	244	2092094	48.10	ng	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	2.51	88	188108	18.294	ng	99
3) Pyridine	3.26	79	500970	18.083	ng	99
4) n-Nitrosodimethylamine	3.19	42	142435	15.922	ng	98
6) Aniline	6.45	93	767203	18.785	ng	98
8) 2-Chlorophenol	6.57	128	480584	20.791	ng	99
9) Benzaldehyde	6.33	77	332348	19.098	ng	99
10) Phenol	6.43	94	637349	20.327	ng	98
11) bis(2-Chloroethyl)ether	6.52	93	466747	18.844	ng	97
12) 1,3-Dichlorobenzene	6.73	146	571646	21.202	ng	100
13) 1,4-Dichlorobenzene	6.80	146	571687	21.294	ng	99
14) 1,2-Dichlorobenzene	6.96	146	535974	21.728	ng	99
15) Benzyl Alcohol	6.93	79	401088	19.385	ng	98
16) 2,2'-oxybis(1-Chloropropan	7.07	45	457516	17.177	ng	99
17) 2-Methylphenol	7.04	107	401406	20.220	ng	99
18) Hexachloroethane	7.30	117	208488	20.938	ng	99
19) n-Nitroso-di-n-propylamine	7.20	70	291909	18.175	ng	97
20) 3+4-Methylphenols	7.19	107	509782	20.590	ng	# 89
22) Acetophenone	7.19	105	657635	20.325	ng	96
24) Nitrobenzene	7.36	77	469268	19.676	ng	95
25) Isophorone	7.60	82	877375	18.862	ng	97
26) 2-Nitrophenol	7.69	139	230641	24.833	ng	99
27) 2,4-Dimethylphenol	7.72	122	394311	21.039	ng	100
28) bis(2-Chloroethoxy)methane	7.82	93	596233	19.712	ng	99
29) 2,4-Dichlorophenol	7.93	162	418477	21.715	ng	99
30) 1,2,4-Trichlorobenzene	8.02	180	483496	22.007	ng	100
31) Naphthalene	8.09	128	1398799	22.570	ng	99
32) Benzoic acid	7.81	122	250769	21.793	ng	96
33) 4-Chloroaniline	8.14	127	617361	21.120	ng	99
34) Hexachlorobutadiene	8.22	225	285060	23.027	ng	99
35) Caprolactam	8.49	113	136882	19.972	ng	99
36) 4-Chloro-3-methylphenol	8.62	107	422853	20.110	ng	97
37) 2-Methylnaphthalene	8.79	142	935825	21.607	ng	99
38) 1-Methylnaphthalene	8.89	142	862077	21.124	ng	99
40) 1,2,4,5-Tetrachlorobenzene	8.96	216	438401	21.915	ng	99


Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP091619\
 Data File : BP000241.D
 Acq On : 16 Sep 2019 16:11
 Operator : HP/JU
 Sample : SSTDICC020
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020

Quant Time: Sep 16 17:24:48 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP091619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 16 17:15:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Hexachlorocyclopentadiene	8.95	237	239504	21.744	ng	100
43) 2,4,6-Trichlorophenol	9.06	196	293830	21.299	ng	99
44) 2,4,5-Trichlorophenol	9.10	196	301071	20.859	ng	100
46) 1,1'-Biphenyl	9.25	154	1101293	22.073	ng	100
47) 2-Chloronaphthalene	9.27	162	903245	21.809	ng	99
48) 2-Nitroaniline	9.37	65	221403	19.791	ng	99
49) Acenaphthylene	9.69	152	1396445	22.723	ng	99
50) Dimethylphthalate	9.55	163	1024462	21.272	ng	99
51) 2,6-Dinitrotoluene	9.61	165	217486	21.275	ng	99
52) Acenaphthene	9.87	154	858893	22.273	ng	100
53) 3-Nitroaniline	9.77	138	254258	20.820	ng	91
54) 2,4-Dinitrophenol	9.89	184	83298	25.806	ng	99
55) Dibenzofuran	10.04	168	1230881	22.479	ng	99
56) 4-Nitrophenol	9.93	139	186808	21.401	ng	98
57) 2,4-Dinitrotoluene	10.02	165	283028	21.687	ng	# 86
58) Fluorene	10.39	166	938756	22.779	ng	99
59) 2,3,4,6-Tetrachlorophenol	10.16	232	257832	23.592	ng	100
60) Diethylphthalate	10.26	149	1012328	21.000	ng	99
61) 4-Chlorophenyl-phenylether	10.37	204	485089	22.646	ng	98
62) 4-Nitroaniline	10.39	138	252691	21.693	ng	98
63) Azobenzene	10.54	77	883959	19.149	ng	98
65) 4,6-Dinitro-2-methylphenol	10.43	198	110107	24.928	ng	96
66) n-Nitrosodiphenylamine	10.49	169	839056	22.700	ng	99
67) 4-Bromophenyl-phenylether	10.87	248	301648	23.857	ng	99
68) Hexachlorobenzene	10.95	284	329783	23.937	ng	99
69) Atrazine	11.03	200	252021	23.434	ng	97
70) Pentachlorophenol	11.13	266	181215	24.275	ng	99
71) Phenanthrene	11.35	178	1423012	22.948	ng	100
72) Anthracene	11.40	178	1408773	22.998	ng	99
73) Carbazole	11.56	167	1319076	22.813	ng	99
74) Di-n-butylphthalate	11.90	149	1617885	23.076	ng	100
75) Fluoranthene	12.56	202	1502033	22.290	ng	97
77) Benzidine	12.67	184	758056	23.176	ng	98
78) Pyrene	12.79	202	1545081	22.154	ng	100
80) Butylbenzylphthalate	13.42	149	683097	21.983	ng	94
81) Benzo(a)anthracene	13.99	228	1301375	21.184	ng	99
82) 3,3'-Dichlorobenzidine	13.95	252	515569	22.750	ng	100
83) Chrysene	14.03	228	1280566	22.039	ng	98
84) Bis(2-ethylhexyl)phthalate	13.99	149	861076	21.293	ng	99
85) Di-n-octyl phthalate	14.61	149	1514954	21.259	ng	99
86) Indeno(1,2,3-cd)pyrene	16.97	276	1343915	18.458	ng	99
88) Benzo(b)fluoranthene	15.05	252	1326760	19.916	ng	99
89) Benzo(k)fluoranthene	15.08	252	1177555	19.316	ng	99
90) Benzo(a)pyrene	15.42	252	1238164	20.590	ng	100
91) Dibenzo(a,h)anthracene	16.99	278	1091504	18.150	ng	99
92) Benzo(g,h,i)perylene	17.42	276	1054360	17.714	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed




#1




1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.79 min Scan# 799
Delta R.T. -0.00 min
Lab File: BP000241.D
Acq: 16 Sep 2019 16:11

Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tot Ion	Ratio	Lower	Upper
152	100		
150	157.8	125.2	187.8
115	63.8	49.7	74.5




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#2


1,4-Dioxane
Concen: 18.294 ng
RT: 2.51 min Scan# 72
Delta R.T. 0.01 min
Lab File: BP000241.D
Acq: 16 Sep 2019 16:11

Tgt Ion	Ratio	Lower	Upper
88	100		
58	57.5	45.4	68.2
43	19.3	15.0	22.6



Abundance Ion 88.00 (87.70 to 88.70): BP0
Ion 58.00 (57.70 to 58.70): BP0
Ion 43.00 (42.70 to 43.70): BP0





#3

Pvridine

Concen: 18.083 ng

RT: 3.26 min Scan# 199

Delta R.T. 0.01 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11

Instrument :

BNA_P

ClientSampleId :

SSTDIC020


Tot Ion: 79 Resp: 500970

Ion Ratio Lower Upper

79 100

52 51.7 40.9 61.3




51 25.6 20.2 30.2



Abundance Ion 79.00 (78.70 to 79.70): BP0

Ion 52.00 (51.70 to 52.70): BP0

Ion 51.00 (50.70 to 51.70): BP0



#4

n-Nitrosodimethylamine

Concen: 15.922 ng

RT: 3.19 min Scan# 187

Delta R.T. -0.00 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11


Tgt Ion: 42 Resp: 142435

Ion Ratio Lower Upper

42 100

74 189.6 154.5 231.7



44 8.3 5.9 8.9




Abundance Ion 42.00 (41.70 to 42.70): BP0

Ion 74.00 (73.70 to 74.70): BP0

Ion 44.00 (43.70 to 44.70): BP0





#5

2-Fluorophenol

Concen: 39.769 ng

RT: 5.40 min Scan# 564

Delta R.T. -0.00 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11

Instrument :

BNA_P

ClientSampleId :

SSTDICC020


Tot Ion: 112 Resp: 889022

Ion Ratio Lower Upper

112 100

64 44.8 36.2 54.4




63 22.7 18.3 27.5



Abundance Ion 112.00 (111.70 to 112.70): B

Ion 64.00 (63.70 to 64.70): BP0

Ion 63.00 (62.70 to 63.70): BP0



#6

Aniline

Concen: 18.785 ng

RT: 6.45 min Scan# 741

Delta R.T. -0.01 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11


Tgt Ion: 93 Resp: 767203

Ion Ratio Lower Upper

93 100

66 33.9 26.2 39.2



65 16.6 12.7 19.1




Abundance Ion 93.00 (92.70 to 93.70): BP0

Ion 66.00 (65.70 to 66.70): BP0

Ion 65.00 (64.70 to 65.70): BP0





#7

Phenol-d6

Concen: 39.205 ng

RT: 6.42 min Scan# 736

Delta R.T. -0.01 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11

Instrument :

BNA_P

ClientSampleId :

SSTDIC020


Tot Ion: 99 Resp: 1106409

Ion Ratio Lower Upper

99 100

42 11.0 8.5 12.7




71 28.4 22.6 34.0



Abundance Ion 99.00 (98.70 to 99.70): BP0

Ion 42.00 (41.70 to 42.70): BP0

Ion 71.00 (70.70 to 71.70): BP0



#8

2-Chlorophenol

Concen: 20.791 ng

RT: 6.57 min Scan# 763

Delta R.T. -0.00 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11


Tgt Ion: 128 Resp: 480584

Ion Ratio Lower Upper

128 100

130 32.4 12.5 52.5



64 35.1 16.0 56.0




Abundance Ion 128.00 (127.70 to 128.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 64.00 (63.70 to 64.70): BP0





#9

Benzaldehyde

Concen: 19.098 ng

RT: 6.33 min Scan# 722

Delta R.T. -0.00 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11

Instrument :

BNA_P

ClientSampleId :

SSTDICC020





Tot Ion: 77 Resp: 332348

Ion Ratio Lower Upper

77 100

105 107.5 89.6 129.6

106 103.9 84.6 124.6



#10

Phenol

Concen: 20.327 ng

RT: 6.43 min Scan# 738

Delta R.T. -0.01 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11




Tgt Ion: 94 Resp: 637349


Ion Ratio Lower Upper

94 100

65 23.0 3.3 43.3

66 31.1 12.7 52.7









#11
bis(2-Chloroethyl)ether
Concen: 18.844 ng
RT: 6.52 min Scan# 754
Delta R.T. -0.00 min
Lab File: BP000241.D
Acq: 16 Sep 2019 16:11




Instrument :
BNA_P
ClientSampleId :
SSTDICC020


Tot Ion	Ratio	Lower	Upper
93	100		
63	58.2	41.0	81.0
95	32.4	12.7	52.7



#12
1,3-Dichlorobenzene
Concen: 21.202 ng
RT: 6.73 min Scan# 790
Delta R.T. -0.00 min
Lab File: BP000241.D
Acq: 16 Sep 2019 16:11

Tgt Ion	Ratio	Lower	Upper
146	100		
148	65.0	52.1	78.1
75	24.6	20.2	30.4





#13

1,4-Dichlorobenzene

Concen: 21.294 ng

RT: 6.80 min Scan# 802





Delta R.T. -0.00 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11

Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tot Ion	Ratio	Lower	Upper
146	100		
148	63.8	50.7	76.1
111	42.2	34.1	51.1



#14

1,2-Dichlorobenzene

Concen: 21.728 ng




RT: 6.96 min Scan# 829


Delta R.T. -0.00 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11

Tgt Ion	Ratio	Lower	Upper
146	100		
148	65.4	51.7	77.5
111	41.2	33.4	50.2









#15
Benzyl Alcohol
Concen: 19.385 ng
RT: 6.93 min Scan# 823
Delta R.T. -0.00 min
Lab File: BP000241.D
Acq: 16 Sep 2019 16:11




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tot Ion	Ratio	Lower	Upper
79	100		
108	89.8	69.2	103.8
77	61.5	48.7	73.1



#16
2,2'-oxybis(1-Chloropropane)
Concen: 17.177 ng
RT: 7.07 min Scan# 847
Delta R.T. -0.00 min
Lab File: BP000241.D
Acq: 16 Sep 2019 16:11

Tgt Ion	Ratio	Lower	Upper
45	100		
77	20.3	0.7	40.7
79	16.1	0.0	36.4









#17
2-Methylphenol
Concen: 20.220 ng
RT: 7.04 min Scan# 842
Delta R.T. -0.00 min
Lab File: BP000241.D
Acq: 16 Sep 2019 16:11




Instrument :
BNA_P
ClientSampleId :
SSTDICC020


Tot Ion:	107	Resp:	401406
Ion	Ratio	Lower	Upper
107	100		
108	113.2	90.1	135.1
77	40.1	31.5	47.3
79	40.1	32.2	48.2



#18
Hexachloroethane
Concen: 20.938 ng
RT: 7.30 min Scan# 887
Delta R.T. -0.00 min
Lab File: BP000241.D
Acq: 16 Sep 2019 16:11

Tgt Ion:	117	Resp:	208488
Ion	Ratio	Lower	Upper
117	100		
119	95.3	75.6	113.4
201	98.5	79.4	119.2









#19
n-Nitroso-di-n-propylamine
Concen: 18.175 ng
RT: 7.20 min Scan# 869
Delta R.T. -0.01 min
Lab File: BP000241.D
Acq: 16 Sep 2019 16:11




Instrument :
BNA_P
ClientSampleId :
SSTDICC020


Tot Ion:	70	Resp:	291909
Ion	Ratio	Lower	Upper
70	100		
42	41.1	30.7	46.1
101	11.7	9.6	14.4
130	26.5	22.0	33.0



#20
3+4-Methylphenols
Concen: 20.590 ng
RT: 7.19 min Scan# 868
Delta R.T. -0.00 min
Lab File: BP000241.D
Acq: 16 Sep 2019 16:11

Tgt Ion:	107	Resp:	509782
Ion	Ratio	Lower	Upper
107	100		
108	90.6	68.8	108.8
77	116.6	75.7	115.7#
79	26.7	6.0	46.0





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.07 min Scan# 1018

Delta R.T. -0.00 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11

Instrument :

BNA_P

ClientSampleId :

SSTDIC020

Tot Ion:136 Resp: 1350824



Ion Ratio Lower Upper

136 100

137 11.0 8.6 13.0

54 5.7 4.5 6.7

68 5.8 4.5 6.7





Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BP0

Ion 68.00 (67.70 to 68.70): BP0



#22

Acetophenone

Concen: 20.325 ng

RT: 7.19 min Scan# 868

Delta R.T. -0.01 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11

Tgt Ion:105 Resp: 657635



Ion Ratio Lower Upper

105 100

71 5.7 5.2 7.8

51 24.3 16.6 24.8

120 23.0 18.5 27.7





Abundance Ion 105.00 (104.70 to 105.70): E

Ion 71.00 (70.70 to 71.70): BP0

Ion 51.00 (50.70 to 51.70): BP0

Ion 120.00 (119.70 to 120.70): E









#23
Nitrobenzene-d5
Concen: 41.038 ng
RT: 7.35 min Scan# 894
Delta R.T. -0.01 min
Lab File: BP000241.D
Acq: 16 Sep 2019 16:11




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tot Ion	82	Resp	903450
Ion Ratio	100	Lower	Upper
128	47.5	40.7	61.1
54	38.9	30.1	45.1



#24
Nitrobenzene
Concen: 19.676 ng
RT: 7.36 min Scan# 897
Delta R.T. -0.01 min
Lab File: BP000241.D
Acq: 16 Sep 2019 16:11

Tgt Ion	77	Resp	469268
Ion Ratio	100	Lower	Upper
123	49.3	43.1	64.7
65	12.1	9.4	14.2





#25

Isophorone

Concen: 18.862 ng

RT: 7.60 min Scan# 938

Delta R.T. -0.01 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11

Instrument :

BNA_P

ClientSampleId :

SSTDIC020





Tot Ion: 82 Resp: 877375

Ion Ratio Lower Upper

82 100

95 7.2 5.9 8.9

138 18.1 15.9 23.9



#26

2-Nitrophenol

Concen: 24.833 ng

RT: 7.69 min Scan# 952

Delta R.T. -0.00 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11




Tgt Ion: 139 Resp: 230641


Ion Ratio Lower Upper

139 100

109 27.9 21.8 32.6

65 38.3 30.5 45.7





#27

2,4-Dimethylphenol

Concen: 21.039 ng

RT: 7.72 min Scan# 958

Delta R.T. -0.00 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11

Instrument :

BNA_P

ClientSampleId :

SSTDIC020


Tot Ion:122 Resp: 394311

Ion Ratio Lower Upper

122 100

107 114.0 91.5 137.3




121 58.9 46.9 70.3



Abundance Ion 122.00 (121.70 to 122.70): E

Ion 107.00 (106.70 to 107.70): E

Ion 121.00 (120.70 to 121.70): E



#28

bis(2-Chloroethoxy)methane

Concen: 19.712 ng

RT: 7.82 min Scan# 974

Delta R.T. -0.01 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11


Tgt Ion: 93 Resp: 596233

Ion Ratio Lower Upper

93 100

95 32.3 26.2 39.4



123 11.6 10.2 15.2




Abundance Ion 93.00 (92.70 to 93.70): BP0

Ion 95.00 (94.70 to 95.70): BP0

Ion 123.00 (122.70 to 123.70): E





#29

2,4-Dichlorophenol

Concen: 21.715 ng

RT: 7.93 min Scan# 993

Delta R.T. -0.00 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11

Instrument :

BNA_P

ClientSampleId :

SSTDIC020


Tot Ion:162 Resp: 418477

Ion Ratio Lower Upper

162 100

164 65.5 44.5 84.5




98 34.2 15.2 55.2



Abundance Ion 162.00 (161.70 to 162.70): E

Ion 164.00 (163.70 to 164.70): E

Ion 98.00 (97.70 to 98.70): B



#30

1,2,4-Trichlorobenzene

Concen: 22.007 ng

RT: 8.02 min Scan# 1008

Delta R.T. -0.00 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11


Tgt Ion:180 Resp: 483496

Ion Ratio Lower Upper

180 100

182 96.4 77.2 115.8



145 30.0 24.3 36.5




Abundance Ion 180.00 (179.70 to 180.70): E

Ion 182.00 (181.70 to 182.70): E

Ion 145.00 (144.70 to 145.70): E









#31
Naphthalene
Concen: 22.570 ng
RT: 8.09 min Scan# 1021
Delta R.T. -0.00 min
Lab File: BP000241.D
Acq: 16 Sep 2019 16:11

Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tot Ion:128 Resp: 1398799




Ion	Ratio	Lower	Upper
128	100		
129	11.2	9.2	13.8
127	13.6	11.0	16.6




#32
Benzoic acid
Concen: 21.793 ng
RT: 7.81 min Scan# 973
Delta R.T. -0.01 min
Lab File: BP000241.D
Acq: 16 Sep 2019 16:11

Tgt Ion:122 Resp: 250769

Ion	Ratio	Lower	Upper
122	100		
105	122.5	105.4	145.4
77	80.8	65.3	105.3









#33
4-Chloroaniline
Concen: 21.120 ng
RT: 8.14 min Scan# 1029
Delta R.T. -0.00 min
Lab File: BP000241.D
Acq: 16 Sep 2019 16:11




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tot Ion:	127	Resp:	617361
Ion	Ratio	Lower	Upper
127	100		
129	32.0	25.7	38.5
65	23.8	19.9	29.9
92	18.7	15.4	23.0



#34
Hexachlorobutadiene
Concen: 23.027 ng
RT: 8.22 min Scan# 1043
Delta R.T. -0.00 min
Lab File: BP000241.D
Acq: 16 Sep 2019 16:11

Tgt Ion:	225	Resp:	285060
Ion	Ratio	Lower	Upper
225	100		
223	62.7	50.2	75.2
227	65.9	51.8	77.6









#35
Caprolactam
Concen: 19.972 ng
RT: 8.49 min Scan# 1088
Delta R.T. -0.01 min
Lab File: BP000241.D
Acq: 16 Sep 2019 16:11




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tot Ion:113 Resp: 136882
Ion Ratio Lower Upper
113 100
55 116.3 96.8 136.8
56 90.1 67.5 107.5



#36
4-Chloro-3-methylphenol
Concen: 20.110 ng
RT: 8.62 min Scan# 1110
Delta R.T. -0.01 min
Lab File: BP000241.D
Acq: 16 Sep 2019 16:11

Tgt Ion:107 Resp: 422853
Ion Ratio Lower Upper
107 100
144 25.1 21.5 32.3
142 79.3 65.7 98.5









#37
2-Methylnaphthalene
Concen: 21.607 ng
RT: 8.79 min Scan# 1139
Delta R.T. -0.00 min
Lab File: BP000241.D
Acq: 16 Sep 2019 16:11




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tot Ion	Ratio	Lower	Upper
142	100		
141	88.5	71.3	106.9
115	33.7	27.2	40.8



#38
1-Methylnaphthalene
Concen: 21.124 ng
RT: 8.89 min Scan# 1156
Delta R.T. -0.00 min
Lab File: BP000241.D
Acq: 16 Sep 2019 16:11

Tgt Ion	Ratio	Lower	Upper
142	100		
141	91.7	73.9	110.9
116	3.7	3.0	4.6





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.83 min Scan# 1317

Delta R.T. -0.00 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11

Instrument :

BNA_P

ClientSampleId :

SSTDICC020


Tot Ion:164 Resp: 706740

Ion Ratio Lower Upper

164 100

162 99.5 79.9 119.9

160 44.4 35.6 53.4






Abundance

Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#40

1,2,4,5-Tetrachlorobenzene

Concen: 21.915 ng

RT: 8.96 min Scan# 1168

Delta R.T. -0.00 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11

Tgt Ion:216 Resp: 438401


Ion Ratio Lower Upper

216 100

214 78.6 63.6 95.4

179 21.3 17.0 25.4

108 19.1 15.1 22.7





Abundance


Ion 216.00 (215.70 to 216.70): E

Ion 214.00 (213.70 to 214.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 108.00 (107.70 to 108.70): E





#41

Hexachlorocyclopentadiene

Concen: 21.744 ng

RT: 8.95 min Scan# 1166

Delta R.T. -0.00 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11

Instrument :

BNA_P

ClientSampleId :

SSTDIC020


Tot Ion:237 Resp: 239504

Ion Ratio Lower Upper

237 100

235 63.2 43.5 83.5




272 12.9 0.0 33.0



Abundance Ion 236.80 (236.50 to 237.50): E

Ion 234.90 (234.60 to 235.60): E

Ion 271.75 (271.45 to 272.45): E



#42

2,4,6-Tribromophenol

Concen: 49.548 ng

RT: 10.63 min Scan# 1452

Delta R.T. -0.00 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11


Tgt Ion:330 Resp: 295162

Ion Ratio Lower Upper

330 100

332 96.2 76.6 115.0



141 34.1 26.5 39.7




Abundance Ion 329.80 (329.50 to 330.50): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E









#43
2,4,6-Trichlorophenol
Concen: 21.299 ng
RT: 9.06 min Scan# 1186
Delta R.T. -0.00 min
Lab File: BP000241.D
Acq: 16 Sep 2019 16:11




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tot Ion	Ratio	Lower	Upper
196	100		
198	93.8	76.0	114.0
200	29.9	24.6	37.0



#44
2,4,5-Trichlorophenol
Concen: 20.859 ng
RT: 9.10 min Scan# 1193
Delta R.T. -0.00 min
Lab File: BP000241.D
Acq: 16 Sep 2019 16:11

Tgt Ion	Ratio	Lower	Upper
196	100		
198	97.9	78.1	117.1
97	41.9	33.3	49.9
132	25.3	20.0	30.0









#45
2-Fluorobiphenyl
Concen: 43.887 ng
RT: 9.15 min Scan# 1201
Delta R.T. -0.00 min
Lab File: BP000241.D
Acq: 16 Sep 2019 16:11




Instrument :
BNA_P
ClientSampleId :
SSTDICC020


Tot Ion:172 Resp: 1857293
Ion Ratio Lower Upper
172 100
171 36.5 29.8 44.6
170 23.9 19.8 29.6



#46
1,1'-Biphenyl
Concen: 22.073 ng
RT: 9.25 min Scan# 1218
Delta R.T. -0.00 min
Lab File: BP000241.D
Acq: 16 Sep 2019 16:11

Tgt Ion:154 Resp: 1101293
Ion Ratio Lower Upper
154 100
153 41.3 21.4 61.4
76 13.0 0.0 33.4









#47
2-Chloronaphthalene
Concen: 21.809 ng
RT: 9.27 min Scan# 1222
Delta R.T. -0.00 min
Lab File: BP000241.D
Acq: 16 Sep 2019 16:11




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tot Ion	Ratio	Lower	Upper
162	100		
127	39.8	32.4	48.6
164	32.8	26.3	39.5



#48
2-Nitroaniline
Concen: 19.791 ng
RT: 9.37 min Scan# 1238
Delta R.T. -0.00 min
Lab File: BP000241.D
Acq: 16 Sep 2019 16:11

Tgt Ion	Ratio	Lower	Upper
65	100		
92	86.7	68.7	103.1
138	144.5	114.4	171.6





#49

Acenaphthylene

Concen: 22.723 ng

RT: 9.69 min Scan# 1293

Delta R.T. -0.00 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11

Instrument :

BNA_P

ClientSampleId :

SSTDICC020





Tot Ion:152 Resp: 1396445

Ion	Ratio	Lower	Upper
152	100		
151	20.4	16.4	24.6
153	13.3	11.0	16.4

152 100

151 20.4 16.4 24.6

153 13.3 11.0 16.4



#50

Dimethylphthalate

Concen: 21.272 ng

RT: 9.55 min Scan# 1269

Delta R.T. -0.00 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11




Tgt Ion:163 Resp: 1024462


Ion	Ratio	Lower	Upper
163	100		
194	4.2	3.4	5.0
164	10.2	8.5	12.7

163 100

194 4.2 3.4 5.0

164 10.2 8.5 12.7









#51
2,6-Dinitrotoluene
Concen: 21.275 ng
RT: 9.61 min Scan# 1279
Delta R.T. -0.00 min
Lab File: BP000241.D
Acq: 16 Sep 2019 16:11

Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tot Ion	Ratio	Lower	Upper
165	100		
63	43.6	35.8	53.8
89	52.6	42.6	63.8




Abundance Ion 165.00 (164.70 to 165.70): E
Ion 63.00 (62.70 to 63.70): BP0
Ion 89.00 (88.70 to 89.70): BP0






#52
Acenaphthene
Concen: 22.273 ng
RT: 9.87 min Scan# 1323
Delta R.T. -0.00 min
Lab File: BP000241.D
Acq: 16 Sep 2019 16:11

Tgt Ion	Ratio	Lower	Upper
154	100		
153	112.4	89.8	134.6
152	53.5	43.4	65.0



Abundance Ion 154.00 (153.70 to 154.70): E
Ion 153.00 (152.70 to 153.70): E
Ion 152.00 (151.70 to 152.70): E









#53
3-Nitroaniline
Concen: 20.820 ng
RT: 9.77 min Scan# 1307
Delta R.T. -0.01 min
Lab File: BP000241.D
Acq: 16 Sep 2019 16:11




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tot Ion:138 Resp: 254258
Ion Ratio Lower Upper
138 100
108 12.4 9.1 13.7
92 118.0 86.7 130.1



#54
2,4-Dinitrophenol
Concen: 25.806 ng
RT: 9.89 min Scan# 1326
Delta R.T. -0.00 min
Lab File: BP000241.D
Acq: 16 Sep 2019 16:11

Tgt Ion:184 Resp: 83298
Ion Ratio Lower Upper
184 100
63 50.7 39.5 59.3
154 60.8 48.8 73.2





#55

Dibenzofuran

Concen: 22.479 ng

RT: 10.04 min Scan# 1352

Delta R.T. -0.00 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11

Instrument :

BNA_P

ClientSampleId :

SSTDIC020


Tot Ion:168 Resp: 1230881

Ion	Ratio	Lower	Upper
168	100		
139	36.9	30.0	45.0
169	13.4	11.0	16.6

168 100

139 36.9 30.0 45.0

169 13.4 11.0 16.6






Abundance

Ion 168.00 (167.70 to 168.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 169.00 (168.70 to 169.70): E



#56

4-Nitrophenol

Concen: 21.401 ng

RT: 9.93 min Scan# 1334

Delta R.T. -0.00 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11


Tgt Ion:139 Resp: 186808

Ion	Ratio	Lower	Upper
139	100		
109	67.0	46.2	86.2
65	77.2	59.4	99.4

139 100

109 67.0 46.2 86.2

65 77.2 59.4 99.4






Abundance

Ion 139.00 (138.70 to 139.70): E

Ion 109.00 (108.70 to 109.70): E

Ion 65.00 (64.70 to 65.70): BPO









#57
2,4-Dinitrotoluene
Concen: 21.687 ng
RT: 10.02 min Scan# 1348
Delta R.T. -0.01 min
Lab File: BP000241.D
Acq: 16 Sep 2019 16:11




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tot Ion:165 Resp: 283028
Ion Ratio Lower Upper
165 100
63 33.6 21.7 32.5#
89 67.3 45.0 67.6
182 3.0 2.4 3.6



#58
Fluorene
Concen: 22.779 ng
RT: 10.39 min Scan# 1411
Delta R.T. -0.00 min
Lab File: BP000241.D
Acq: 16 Sep 2019 16:11

Tgt Ion:166 Resp: 938756
Ion Ratio Lower Upper
166 100
165 97.0 78.2 117.2
167 13.2 10.7 16.1









#59
2,3,4,6-Tetrachlorophenol
Concen: 23.592 ng
RT: 10.16 min Scan# 1372
Delta R.T. -0.00 min
Lab File: BP000241.D
Acq: 16 Sep 2019 16:11




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tot Ion	Ratio	Lower	Upper
232	100		
131	41.8	33.6	50.4
130	2.2	1.8	2.6
166	30.1	24.2	36.2



#60
Diethylphthalate
Concen: 21.000 ng
RT: 10.26 min Scan# 1389
Delta R.T. -0.00 min
Lab File: BP000241.D
Acq: 16 Sep 2019 16:11

Tgt Ion	Ratio	Lower	Upper
149	100		
177	23.0	18.1	27.1
150	12.5	10.3	15.5






#61
4-Chlorophenyl-phenylether
Concen: 22.646 ng
RT: 10.37 min Scan# 1409
Delta R.T. -0.00 min
Lab File: BP000241.D
Acq: 16 Sep 2019 16:11

Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tot Ion	Ratio	Lower	Upper
204	100		
206	33.2	26.3	39.5
141	62.3	48.6	72.8






Abundance

Ion 204.00 (203.70 to 204.70): E


Ion 206.00 (205.70 to 206.70): E

Ion 141.00 (140.70 to 141.70): E



#62
4-Nitroaniline
Concen: 21.693 ng
RT: 10.39 min Scan# 1412
Delta R.T. -0.00 min
Lab File: BP000241.D
Acq: 16 Sep 2019 16:11

Tgt Ion	Ratio	Lower	Upper
138	100		
92	47.6	30.1	70.1
108	68.6	49.5	89.5






Abundance

Ion 138.00 (137.70 to 138.70): E

Ion 92.00 (91.70 to 92.70): BP0

Ion 108.00 (107.70 to 108.70): E









#63
Azobenzene
Concen: 19.149 ng
RT: 10.54 min Scan# 1437
Delta R.T. -0.00 min
Lab File: BP000241.D
Acq: 16 Sep 2019 16:11




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tot Ion	Ratio	Lower	Upper
77	100		
182	29.5	11.1	51.1
105	24.0	4.7	44.7
51	22.3	2.2	42.2



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.33 min Scan# 1571
Delta R.T. -0.00 min
Lab File: BP000241.D
Acq: 16 Sep 2019 16:11

Tgt Ion	Ratio	Lower	Upper
188	100		
94	11.0	8.5	12.7
80	11.5	8.6	13.0





#65

4,6-Dinitro-2-methylphenol

Concen: 24.928 ng

RT: 10.43 min Scan# 1419

Delta R.T. -0.00 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11

Instrument :

BNA_P

ClientSampleId :

SSTDIC020


Tot Ion:198 Resp: 110107

Ion Ratio Lower Upper

198 100

51 24.9 7.4 47.4




105 37.8 20.2 60.2



Abundance Ion 198.00 (197.70 to 198.70): E

Ion 51.00 (50.70 to 51.70): BP0

Ion 105.00 (104.70 to 105.70): E



#66

n-Nitrosodiphenylamine

Concen: 22.700 ng

RT: 10.49 min Scan# 1429

Delta R.T. -0.00 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11


Tgt Ion:169 Resp: 839056

Ion Ratio Lower Upper

169 100

168 66.7 52.9 79.3



167 35.8 29.0 43.6




Abundance Ion 169.00 (168.70 to 169.70): E

Ion 168.00 (167.70 to 168.70): E

Ion 167.00 (166.70 to 167.70): E





#67

4-Bromophenyl-phenylether

Concen: 23.857 ng

RT: 10.87 min Scan# 1493

Delta R.T. -0.00 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11

Instrument :

BNA_P

ClientSampleId :

SSTDICC020


Tot Ion:248 Resp: 301648

Ion Ratio Lower Upper

248 100

250 97.2 77.0 115.4

141 74.7 58.5 87.7






Abundance

Ion 248.00 (247.70 to 248.70): E

Ion 250.00 (249.70 to 250.70): E

Ion 141.00 (140.70 to 141.70): E



#68

Hexachlorobenzene

Concen: 23.937 ng

RT: 10.95 min Scan# 1506

Delta R.T. -0.00 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11


Tgt Ion:284 Resp: 329783

Ion Ratio Lower Upper

284 100

142 30.4 23.6 35.4

249 30.4 24.3 36.5






Abundance

Ion 283.80 (283.50 to 284.50): E

Ion 142.00 (141.70 to 142.70): E

Ion 249.00 (248.70 to 249.70): E





#69

Atrazine

Concen: 23.434 ng

RT: 11.03 min Scan# 1520

Delta R.T. 0.01 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11

Instrument :

BNA_P

ClientSampleId :

SSTDICC020


Tot Ion:200 Resp: 252021

Ion Ratio Lower Upper

200 100

173 25.8 7.6 47.6




215 50.5 28.6 68.6



Abundance Ion 200.00 (199.70 to 200.70): E

Ion 173.00 (172.70 to 173.70): E

Ion 215.00 (214.70 to 215.70): E



#70

Pentachlorophenol

Concen: 24.275 ng

RT: 11.13 min Scan# 1538

Delta R.T. -0.00 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11


Tgt Ion:266 Resp: 181215

Ion Ratio Lower Upper

266 100

268 63.1 51.6 77.4



264 64.4 50.8 76.2




Abundance Ion 265.70 (265.40 to 266.40): E

Ion 268.00 (267.70 to 268.70): E

Ion 264.00 (263.70 to 264.70): E









#71
Phenanthrene
Concen: 22.948 ng
RT: 11.35 min Scan# 1575
Delta R.T. -0.00 min
Lab File: BP000241.D
Acq: 16 Sep 2019 16:11




Instrument :
BNA_P
ClientSampleId :
SSTDICC020


Tot Ion:178 Resp: 1423012
Ion Ratio Lower Upper
178 100
176 19.8 15.8 23.8
179 15.8 12.7 19.1



#72
Anthracene
Concen: 22.998 ng
RT: 11.40 min Scan# 1584
Delta R.T. -0.00 min
Lab File: BP000241.D
Acq: 16 Sep 2019 16:11

Tgt Ion:178 Resp: 1408773
Ion Ratio Lower Upper
178 100
176 18.7 15.4 23.0
179 15.8 12.7 19.1









#73
 Carbazole
 Concen: 22.813 ng
 RT: 11.56 min Scan# 1610
 Delta R.T. -0.00 min
 Lab File: BP000241.D
 Acq: 16 Sep 2019 16:11

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020

Tot Ion:167 Resp: 1319076




Ion	Ratio	Lower	Upper
167	100		
166	21.9	17.8	26.8
139	13.0	10.8	16.2




#74
 Di-n-butylphthalate
 Concen: 23.076 ng
 RT: 11.90 min Scan# 1669
 Delta R.T. -0.00 min
 Lab File: BP000241.D
 Acq: 16 Sep 2019 16:11

Tgt Ion:149 Resp: 1617885

Ion	Ratio	Lower	Upper
149	100		
150	9.2	7.4	11.2
104	4.6	3.5	5.3





#75

Fluoranthene

Concen: 22.290 ng

RT: 12.56 min Scan# 1780

Delta R.T. 0.01 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11

Instrument :

BNA_P

ClientSampleId :

SSTDICC020


Tot Ion:202 Resp: 1502033

Ion Ratio Lower Upper

202 100

101 11.6 0.0 33.9




203 17.5 0.0 38.0



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E



#76

Chrysene-d12

Concen: 20.000 ng

RT: 14.00 min Scan# 2026

Delta R.T. 0.01 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11


Tgt Ion:240 Resp: 948008

Ion Ratio Lower Upper

240 100

120 11.2 10.4 15.6



236 26.7 21.0 31.6




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E





#77

Benzidine

Concen: 23.176 ng

RT: 12.67 min Scan# 1800

Delta R.T. 0.01 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11

Instrument :

BNA_P

ClientSampleId :

SSTDIC020


Tot Ion:184 Resp: 758056

Ion Ratio Lower Upper

184 100

185 14.7 12.7 19.1




183 12.0 9.5 14.3



Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E



#78

Pyrene

Concen: 22.154 ng

RT: 12.79 min Scan# 1819

Delta R.T. -0.00 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11


Tgt Ion:202 Resp: 1545081

Ion Ratio Lower Upper

202 100

200 21.8 17.5 26.3



203 17.6 14.3 21.5




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E





#79

Terphenyl-d14

Concen: 48.100 ng

RT: 12.93 min Scan# 1844

Delta R.T. -0.00 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11

Instrument :

BNA_P

ClientSampleId :

SSTDIC020


Tot Ion:244 Resp: 2092094

Ion Ratio Lower Upper

244 100

212 7.2 5.9 8.9




122 11.4 9.0 13.4



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#80

Butylbenzylphthalate

Concen: 21.983 ng

RT: 13.42 min Scan# 1926

Delta R.T. -0.00 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11


Tgt Ion:149 Resp: 683097

Ion Ratio Lower Upper

149 100

91 68.5 50.6 76.0



206 21.2 17.8 26.6




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 91.00 (90.70 to 91.70): BP0

Ion 206.00 (205.70 to 206.70): E





#81

Benzo(a)anthracene

Concen: 21.184 ng

RT: 13.99 min Scan# 2023

Delta R.T. -0.00 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11

Instrument :

BNA_P

ClientSampleId :

SSTDIC020


Tot Ion:228 Resp: 1301375

Ion Ratio Lower Upper

228 100

226 26.9 22.2 33.4

229 19.9 16.0 24.0






Abundance

Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#82

3,3'-Dichlorobenzidine

Concen: 22.750 ng

RT: 13.95 min Scan# 2017

Delta R.T. -0.00 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11


Tgt Ion:252 Resp: 515569

Ion Ratio Lower Upper

252 100

254 64.4 51.5 77.3

126 11.3 8.4 12.6






Abundance

Ion 252.00 (251.70 to 252.70): E

Ion 254.00 (253.70 to 254.70): E

Ion 126.00 (125.70 to 126.70): E





#83

Chrysene

Concen: 22.039 ng

RT: 14.03 min Scan# 2030

Delta R.T. -0.00 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11

Instrument :

BNA_P

ClientSampleId :

SSTDIC020


Tot Ion:228 Resp: 1280566

Ion Ratio Lower Upper

228 100

226 30.0 24.9 37.3




229 19.9 16.2 24.2



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#84

Bis(2-ethylhexyl)phthalate

Concen: 21.293 ng

RT: 13.99 min Scan# 2024

Delta R.T. 0.01 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11


Tgt Ion:149 Resp: 861076

Ion Ratio Lower Upper

149 100

167 28.0 23.0 34.4



279 4.9 3.7 5.5




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E





#85

Di-n-octyl phthalate

Concen: 21.259 ng

RT: 14.61 min Scan# 2129

Delta R.T. -0.00 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11

Instrument :

BNA_P

ClientSampleId :

SSTDIC020


Tot Ion:149 Resp: 1514954

Ion Ratio Lower Upper

149 100

167 1.6 1.4 2.0




43 6.0 5.0 7.6



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 43.00 (42.70 to 43.70): BP0



#86

Indeno(1,2,3-cd)pyrene

Concen: 18.458 ng

RT: 16.97 min Scan# 2530

Delta R.T. -0.00 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11


Tgt Ion:276 Resp: 1343915

Ion Ratio Lower Upper

276 100

138 27.8 22.6 33.8



277 25.6 20.4 30.6




Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E





#87

Pervlene-d12

Concen: 20.000 ng

RT: 15.49 min Scan# 2278

Delta R.T. 0.01 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11

Instrument :

BNA_P

ClientSampleId :

SSTDIC020


Tot Ion:264 Resp: 1090623

Ion Ratio Lower Upper

264 100

260 23.9 18.9 28.3

265 21.5 17.2 25.8






Abundance

Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E



#88

Benzo(b)fluoranthene

Concen: 19.916 ng

RT: 15.05 min Scan# 2204

Delta R.T. -0.00 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11


Tgt Ion:252 Resp: 1326760

Ion Ratio Lower Upper

252 100

253 22.0 17.8 26.6

125 11.5 9.5 14.3






Abundance

Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E









#89
Benzo(k)fluoranthene
Concen: 19.316 ng
RT: 15.08 min Scan# 2209
Delta R.T. -0.00 min
Lab File: BP000241.D
Acq: 16 Sep 2019 16:11




Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tot Ion:252 Resp: 1177555
Ion Ratio Lower Upper
252 100
253 22.0 17.8 26.8
125 11.8 9.6 14.4



#90
Benzo(a)pyrene
Concen: 20.590 ng
RT: 15.42 min Scan# 2267
Delta R.T. -0.00 min
Lab File: BP000241.D
Acq: 16 Sep 2019 16:11

Tgt Ion:252 Resp: 1238164
Ion Ratio Lower Upper
252 100
253 21.8 17.7 26.5
125 12.6 10.0 15.0





#91

Dibenzo(a,h)anthracene

Concen: 18.150 ng

RT: 16.99 min Scan# 2533

Delta R.T. -0.00 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11

Instrument :


BNA_P

ClientSampleId :

SSTDIC020

Tot Ion:278 Resp: 1091504




Ion	Ratio	Lower	Upper
278	100		
139	18.4	14.5	21.7
279	23.5	19.0	28.6



Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E



#92

Benzo(g,h,i)perylene

Concen: 17.714 ng

RT: 17.42 min Scan# 2606


Delta R.T. -0.00 min

Lab File: BP000241.D

Acq: 16 Sep 2019 16:11

Tgt Ion:276 Resp: 1054360

Ion	Ratio	Lower	Upper
276	100		
277	24.0	19.1	28.7
138	22.9	19.5	29.3



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

