


Data Path : Z:\SVOASRV\HPCHEM1\BNA_P\DATA\BP091619\
 Data File : BP000239.D
 Acq On : 16 Sep 2019 15:23
 Operator : HP/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP091619.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Sep 16 18:59:05 2019



AutoFind: Scans 1617, 1618, 1619; Background Corrected with Scan 1609

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	24.1	52665	PASS
68	69	0.00	2	1.5	961	PASS
69	198	0.00	100	29.3	64016	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	44.6	97412	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	218506	PASS
199	198	5	9	6.8	14947	PASS
275	198	10	60	26.4	57718	PASS
365	198	1	100	3.2	6938	PASS
441	198	0.01	100	13.4	29270	PASS
442	198	50	100	85.6	187040	PASS
443	442	15	24	19.3	36043	PASS