

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091622\
 Data File : BP011723.D
 Acq On : 16 Sep 2022 09:45
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 16 15:25:39 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 14 19:12:50 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	108	-0.01
2	1,4-Dioxane	40.000	39.464	1.3	110	0.00
3	Pyridine	40.000	41.023	-2.6	112	0.00
4	n-Nitrosodimethylamine	40.000	41.346	-3.4	113	0.00
5 S	2-Fluorophenol	80.000	80.868	-1.1	111	0.00
6	Aniline	40.000	40.767	-1.9	110	0.00
7 S	Phenol-d6	80.000	81.884	-2.4	110	0.00
8	2-Chlorophenol	40.000	40.261	-0.7	109	0.00
9	Benzaldehyde	40.000	39.617	1.0	112	-0.01
10 C	Phenol	40.000	40.705	-1.8	110	0.00
11	bis(2-Chloroethyl)ether	40.000	40.352	-0.9	110	0.00
12	1,3-Dichlorobenzene	40.000	39.339	1.7	108	-0.01
13 C	1,4-Dichlorobenzene	40.000	39.457	1.4	108	-0.01
14	1,2-Dichlorobenzene	40.000	39.316	1.7	108	0.00
15	Benzyl Alcohol	40.000	41.280	-3.2	111	-0.01
16	2,2'-oxybis(1-Chloropropane	40.000	40.879	-2.2	111	-0.01
17	2-Methylphenol	40.000	40.777	-1.9	110	-0.01
18	Hexachloroethane	40.000	40.853	-2.1	111	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.921	-2.3	108	0.00
20	3+4-Methylphenols	40.000	40.562	-1.4	109	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	109	-0.01
22	Acetophenone	40.000	40.156	-0.4	109	0.00
23 S	Nitrobenzene-d5	80.000	82.063	-2.6	110	0.00
24	Nitrobenzene	40.000	40.756	-1.9	110	0.00
25	Isophorone	40.000	40.577	-1.4	108	0.00
26 C	2-Nitrophenol	40.000	36.605	8.5	105	0.00
27	2,4-Dimethylphenol	40.000	40.676	-1.7	108	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.265	-0.7	109	0.00
29 C	2,4-Dichlorophenol	40.000	40.392	-1.0	108	0.00
30	1,2,4-Trichlorobenzene	40.000	39.086	2.3	107	0.00
31	Naphthalene	40.000	39.457	1.4	108	0.00
32	Benzoic acid	40.000	35.724	10.7	102	0.00
33	4-Chloroaniline	40.000	40.179	-0.4	108	0.00
34 C	Hexachlorobutadiene	40.000	38.509	3.7	106	0.00
35	Caprolactam	40.000	38.914	2.7	104	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.286	-0.7	107	0.00
37	2-Methylnaphthalene	40.000	39.653	0.9	108	0.00
38	1-Methylnaphthalene	40.000	39.583	1.0	108	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	107	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.406	1.5	106	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.842	0.4	105	0.00
42 S	2,4,6-Tribromophenol	80.000	78.314	2.1	102	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.684	-1.7	106	0.00
44	2,4,5-Trichlorophenol	40.000	40.288	-0.7	106	0.00
45 S	2-Fluorobiphenyl	80.000	80.868	-1.1	108	0.00
46	1,1'-Biphenyl	40.000	40.237	-0.6	108	-0.01
47	2-Chloronaphthalene	40.000	39.945	0.1	107	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091622\
 Data File : BP011723.D
 Acq On : 16 Sep 2022 09:45
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 16 15:25:39 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 14 19:12:50 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.974	-7.4	109	0.00
49	Acenaphthylene	40.000	41.057	-2.6	108	0.00
50	Dimethylphthalate	40.000	40.271	-0.7	108	0.00
51	2,6-Dinitrotoluene	40.000	41.426	-3.6	106	0.00
52 C	Acenaphthene	40.000	40.344	-0.9	107	-0.01
53	3-Nitroaniline	40.000	39.067	2.3	108	0.00
54 P	2,4-Dinitrophenol	40.000	35.332	11.7	100	0.00
55	Dibenzofuran	40.000	39.997	0.0	108	0.00
56 P	4-Nitrophenol	40.000	40.528	-1.3	106	0.00
57	2,4-Dinitrotoluene	40.000	38.733	3.2	107	0.00
58	Fluorene	40.000	40.931	-2.3	108	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.495	1.3	104	0.00
60	Diethylphthalate	40.000	41.045	-2.6	109	0.00
61	4-Chlorophenyl-phenylether	40.000	40.211	-0.5	108	0.00
62	4-Nitroaniline	40.000	39.879	0.3	110	0.00
63	Azobenzene	40.000	42.595	-6.5	110	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	107	0.00
65	4,6-Dinitro-2-methylphenol	40.000	36.435	8.9	103	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.415	-1.0	107	0.00
67	4-Bromophenyl-phenylether	40.000	38.763	3.1	104	0.00
68	Hexachlorobenzene	40.000	38.268	4.3	104	0.00
69	Atrazine	40.000	39.961	0.1	104	0.00
70 C	Pentachlorophenol	40.000	38.183	4.5	102	0.00
71	Phenanthrene	40.000	39.807	0.5	107	0.00
72	Anthracene	40.000	40.703	-1.8	107	0.00
73	Carbazole	40.000	40.847	-2.1	107	0.00
74	Di-n-butylphthalate	40.000	41.657	-4.1	107	0.00
75 C	Fluoranthene	40.000	40.466	-1.2	107	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	106	0.00
77	Benzydine	40.000	46.804	-17.0	107	0.00
78	Pyrene	40.000	40.949	-2.4	108	0.00
79 S	Terphenyl-d14	80.000	81.548	-1.9	107	0.00
80	Butylbenzylphthalate	40.000	42.543	-6.4	107	0.00
81	Benzo(a)anthracene	40.000	40.486	-1.2	107	0.00
82	3,3'-Dichlorobenzidine	40.000	41.678	-4.2	105	0.00
83	Chrysene	40.000	40.060	-0.2	106	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	43.402	-8.5	107	0.00
85 c	Di-n-octyl phthalate	40.000	40.736	-1.8	105	0.00
86 I	Perylene-d12	20.000	20.000	0.0	104	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.978	-2.4	105	0.00
88	Benzo(b)fluoranthene	40.000	41.888	-4.7	107	0.00
89	Benzo(k)fluoranthene	40.000	40.608	-1.5	104	0.00
90 C	Benzo(a)pyrene	40.000	41.411	-3.5	105	0.00
91	Dibenzo(a,h)anthracene	40.000	41.081	-2.7	105	0.00
92	Benzo(g,h,i)perylene	40.000	40.338	-0.8	104	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091622\
Data File : BP011723.D
Acq On : 16 Sep 2022 09:45
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Sep 16 15:25:39 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 14 19:12:50 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0