

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091622\
 Data File : BP011727.D
 Acq On : 16 Sep 2022 13:37
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 16 15:26:56 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 14 19:12:50 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	128	0.00
2	1,4-Dioxane	40.000	39.070	2.3	130	0.00
3	Pyridine	40.000	40.445	-1.1	131	0.00
4	n-Nitrosodimethylamine	40.000	41.524	-3.8	136	0.00
5 S	2-Fluorophenol	80.000	81.275	-1.6	133	0.00
6	Aniline	40.000	41.136	-2.8	132	0.00
7 S	Phenol-d6	80.000	82.685	-3.4	132	0.00
8	2-Chlorophenol	40.000	40.694	-1.7	131	0.00
9	Benzaldehyde	40.000	40.015	-0.0	135	0.00
10 C	Phenol	40.000	41.220	-3.0	133	0.00
11	bis(2-Chloroethyl)ether	40.000	40.454	-1.1	132	0.00
12	1,3-Dichlorobenzene	40.000	39.325	1.7	129	0.00
13 C	1,4-Dichlorobenzene	40.000	39.229	1.9	128	0.00
14	1,2-Dichlorobenzene	40.000	39.130	2.2	128	0.00
15	Benzyl Alcohol	40.000	41.663	-4.2	133	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.580	-1.4	132	0.00
17	2-Methylphenol	40.000	41.022	-2.6	132	0.00
18	Hexachloroethane	40.000	40.231	-0.6	131	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.411	-3.5	131	0.00
20	3+4-Methylphenols	40.000	41.152	-2.9	132	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	130	0.00
22	Acetophenone	40.000	40.404	-1.0	131	0.00
23 S	Nitrobenzene-d5	80.000	82.658	-3.3	132	0.00
24	Nitrobenzene	40.000	40.983	-2.5	132	0.00
25	Isophorone	40.000	41.621	-4.1	132	0.00
26 C	2-Nitrophenol	40.000	38.301	4.2	131	0.00
27	2,4-Dimethylphenol	40.000	41.450	-3.6	131	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.455	-1.1	131	0.00
29 C	2,4-Dichlorophenol	40.000	40.873	-2.2	130	0.00
30	1,2,4-Trichlorobenzene	40.000	39.358	1.6	129	0.00
31	Naphthalene	40.000	39.524	1.2	129	0.00
32	Benzoic acid	40.000	37.857	5.4	130	0.01
33	4-Chloroaniline	40.000	40.566	-1.4	130	0.00
34 C	Hexachlorobutadiene	40.000	38.834	2.9	127	0.00
35	Caprolactam	40.000	40.706	-1.8	129	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.913	-2.3	130	0.00
37	2-Methylnaphthalene	40.000	40.003	-0.0	130	0.00
38	1-Methylnaphthalene	40.000	39.891	0.3	129	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	128	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.600	1.0	128	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.271	-0.7	127	0.00
42 S	2,4,6-Tribromophenol	80.000	78.779	1.5	124	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.956	-2.4	129	0.00
44	2,4,5-Trichlorophenol	40.000	40.640	-1.6	129	0.00
45 S	2-Fluorobiphenyl	80.000	79.521	0.6	128	0.00
46	1,1'-Biphenyl	40.000	39.676	0.8	129	0.00
47	2-Chloronaphthalene	40.000	39.320	1.7	127	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091622\
 Data File : BP011727.D
 Acq On : 16 Sep 2022 13:37
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 16 15:26:56 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 14 19:12:50 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	43.574	-8.9	133	0.00
49	Acenaphthylene	40.000	40.797	-2.0	129	0.00
50	Dimethylphthalate	40.000	39.780	0.5	128	0.00
51	2,6-Dinitrotoluene	40.000	41.819	-4.5	129	0.00
52 C	Acenaphthene	40.000	39.953	0.1	128	0.00
53	3-Nitroaniline	40.000	39.211	2.0	130	0.00
54 P	2,4-Dinitrophenol	40.000	36.335	9.2	125	0.00
55	Dibenzofuran	40.000	39.439	1.4	128	0.00
56 P	4-Nitrophenol	40.000	40.422	-1.1	128	0.00
57	2,4-Dinitrotoluene	40.000	38.825	2.9	130	0.00
58	Fluorene	40.000	40.311	-0.8	128	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.985	0.0	127	0.00
60	Diethylphthalate	40.000	40.676	-1.7	130	0.00
61	4-Chlorophenyl-phenylether	40.000	39.603	1.0	128	0.00
62	4-Nitroaniline	40.000	39.573	1.1	131	0.00
63	Azobenzene	40.000	41.839	-4.6	130	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	127	0.00
65	4,6-Dinitro-2-methylphenol	40.000	37.469	6.3	126	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.707	-1.8	128	0.00
67	4-Bromophenyl-phenylether	40.000	39.230	1.9	124	0.00
68	Hexachlorobenzene	40.000	39.017	2.5	125	0.00
69	Atrazine	40.000	40.384	-1.0	125	0.00
70 C	Pentachlorophenol	40.000	39.298	1.8	125	0.00
71	Phenanthrene	40.000	39.875	0.3	127	0.00
72	Anthracene	40.000	40.666	-1.7	127	0.00
73	Carbazole	40.000	40.596	-1.5	126	0.00
74	Di-n-butylphthalate	40.000	41.254	-3.1	126	0.00
75 C	Fluoranthene	40.000	39.987	0.0	125	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	122	0.00
77	Benzydine	40.000	45.511	-13.8	120	0.00
78	Pyrene	40.000	41.076	-2.7	125	0.00
79 S	Terphenyl-d14	80.000	80.556	-0.7	122	0.00
80	Butylbenzylphthalate	40.000	43.178	-7.9	126	0.00
81	Benzo(a)anthracene	40.000	40.184	-0.5	123	0.00
82	3,3'-Dichlorobenzidine	40.000	41.662	-4.2	121	0.00
83	Chrysene	40.000	39.569	1.1	121	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	43.485	-8.7	123	0.00
85 c	Di-n-octyl phthalate	40.000	41.243	-3.1	122	0.00
86 I	Perylene-d12	20.000	20.000	0.0	120	0.01
87	Indeno(1,2,3-cd)pyrene	40.000	40.492	-1.2	120	0.02
88	Benzo(b)fluoranthene	40.000	40.470	-1.2	120	0.00
89	Benzo(k)fluoranthene	40.000	41.226	-3.1	123	0.00
90 C	Benzo(a)pyrene	40.000	41.161	-2.9	121	0.00
91	Dibenzo(a,h)anthracene	40.000	40.462	-1.2	120	0.02
92	Benzo(g,h,i)perylene	40.000	39.976	0.1	119	0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091622\
Data File : BP011727.D
Acq On : 16 Sep 2022 13:37
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Sep 16 15:26:56 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 14 19:12:50 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0