

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091823\
 Data File : BP017210.D
 Acq On : 18 Sep 2023 17:59
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020739

Quant Time: Sep 18 22:33:26 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP091223.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Sep 16 03:18:48 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.946	152	254442	20.000	ng/u1	0.00
20) Naphthalene-d8	10.769	136	1101986	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.604	164	694689	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.375	188	1537453	20.000	ng/u1	0.00
79) Chrysene-d12	21.480	240	1374566	20.000	ng/u1	0.00
88) Perylene-d12	23.998	264	1256989	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.311	96	50595	8.356	ng/uL	0.00
4) Pyridine-d5	3.740	84	367329	21.207	ng/u1	0.00
7) Phenol-d5	7.099	99	432304	20.895	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.287	67	280362	21.681	ng/u1	0.00
11) 2-Chlorophenol-d4	7.463	132	350952	21.263	ng/u1	0.00
15) 4-Methylphenol-d8	8.658	113	358475	20.906	ng/u1	0.00
21) Nitrobenzene-d5	9.146	128	169796	22.005	ng/u1	0.00
24) 2-Nitrophenol-d4	9.857	143	184536	22.401	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.381	165	351213	22.045	ng/u1	0.00
31) 4-Chloroaniline-d4	10.934	131	503095	20.951	ng/u1	0.00
46) Dimethylphthalate-d6	14.022	166	1061141	21.465	ng/u1	0.00
49) Acenaphthylene-d8	14.304	160	1235986	22.034	ng/u1	0.00
54) 4-Nitrophenol-d4	14.828	143	183892	19.247	ng/u1	0.00
60) Fluorene-d10	15.598	176	921580	21.560	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.734	200	169446	19.727	ng/u1	0.00
73) Anthracene-d10	17.475	188	1440165	22.047	ng/u1	0.00
81) Pyrene-d10	19.716	212	1645032	22.819	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.833	264	1307254	22.034	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.346	88	53997	8.293	ng/uL	94
5) Pyridine	3.764	79	378615	21.103	ng/u1	99
6) Benzaldehyde	7.105	77	285513	24.787	ng/u1	98
8) Phenol	7.122	94	455478	20.874	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.381	93	369569	21.235	ng/u1	100
12) 2-Chlorophenol	7.499	128	354266	21.146	ng/u1	98
13) 2-Methylphenol	8.387	108	340956	20.960	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.458	45	322117	13.657	ng/u1	99
16) Acetophenone	8.793	105	570903	21.272	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.763	70	299630	21.726	ng/u1	97
18) 4-Methylphenol	8.722	108	370947	20.611	ng/u1	96
19) Hexachloroethane	9.010	117	148477	21.549	ng/u1	99
22) Nitrobenzene	9.187	77	441075	21.949	ng/u1	99
23) Isophorone	9.699	82	818175	21.727	ng/u1	99
25) 2-Nitrophenol	9.893	139	202393	22.105	ng/u1	98
26) 2,4-Dimethylphenol	9.928	107	399611	21.416	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.187	93	500062	21.187	ng/u1	99
29) 2,4-Dichlorophenol	10.410	162	348634	21.619	ng/u1	98
30) Naphthalene	10.822	128	1193056	21.395	ng/u1	100
32) 4-Chloroaniline	10.957	127	495946	21.165	ng/u1	97
33) Hexachlorobutadiene	11.046	225	236512	21.598	ng/u1	98
34) Caprolactam	11.781	113	114115	22.511	ng/u1	97
35) 4-Chloro-3-methylphenol	12.057	107	375935	21.545	ng/u1	98
36) 2-Methylnaphthalene	12.428	142	806368	21.430	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091823\
 Data File : BP017210.D
 Acq On : 18 Sep 2023 17:59
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020739

Quant Time: Sep 18 22:33:26 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP091223.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Sep 16 03:18:48 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.646	142	819785	21.524	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.775	216	454077	21.946	ng/ul	100
40) Hexachlorocyclopentadiene	12.722	237	292608	22.698	ng/ul	99
41) 2,4,6-Trichlorophenol	13.028	196	292819	22.278	ng/ul	99
42) 2,4,5-Trichlorophenol	13.099	196	313952	22.467	ng/ul	99
43) 1,1'-Biphenyl	13.440	154	1072699	21.534	ng/ul	100
44) 2-Chloronaphthalene	13.487	162	858303	21.697	ng/ul	99
45) 2-Nitroaniline	13.722	65	247652	23.112	ng/ul	99
47) Dimethylphthalate	14.069	163	1072157	21.364	ng/ul	100
48) 2,6-Dinitrotoluene	14.210	165	204462	22.285	ng/ul	95
50) Acenaphthylene	14.334	152	1416119	21.853	ng/ul	99
51) 3-Nitroaniline	14.551	138	209847	21.380	ng/ul	99
52) Acenaphthene	14.669	153	926682	21.610	ng/ul	99
53) 2,4-Dinitrophenol	14.751	184	148802	23.155	ng/ul	99
55) 4-Nitrophenol	14.840	109	151540	19.817	ng/ul	96
56) Dibenzofuran	15.004	168	1277367	21.556	ng/ul	100
57) 2,4-Dinitrotoluene	14.998	165	307319	22.495	ng/ul	94
58) 2,3,4,6-Tetrachlorophenol	15.222	232	271507	22.148	ng/ul	98
59) Diethylphthalate	15.422	149	1064632	21.487	ng/ul	100
61) Fluorene	15.657	166	1051529	21.463	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.651	204	545886	21.915	ng/ul	98
63) 4-Nitroaniline	15.722	138	196610	21.547	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.745	198	185624	20.136	ng/ul	97
67) N-Nitrosodiphenylamine	15.875	169	869565	21.752	ng/ul	99
68) 4-Bromophenyl-phenylether	16.551	248	325030	22.178	ng/ul	97
69) Hexachlorobenzene	16.634	284	375537	21.649	ng/ul	98
70) Atrazine	16.834	200	327435	22.627	ng/ul	98
71) Pentachlorophenol	16.998	266	227975	20.223	ng/ul	99
72) Phenanthrene	17.422	178	1667644	21.584	ng/ul	99
74) Anthracene	17.510	178	1685016	21.533	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.393	216	478646	22.517	ng/ul	98
76) Pentachlorobenzene	14.898	250	433324	22.465	ng/ul	98
77) Carbazole	17.798	167	1474669	21.461	ng/ul	99
78) Di-n-butylphthalate	18.310	149	1787738	22.832	ng/ul	99
80) Fluoranthene	19.386	202	1934677	22.961	ng/ul	100
82) Pyrene	19.745	202	2037621	22.887	ng/ul	100
83) Butylbenzylphthalate	20.604	149	757026	23.841	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.404	252	557389	19.779	ng/ul	98
85) Benzo(a)anthracene	21.463	228	1940131	21.557	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.339	149	1070573	24.018	ng/ul	100
87) Chrysene	21.516	228	1787517	21.351	ng/ul	99
89) Di-n-octyl phthalate	22.304	149	1732476	24.857	ng/ul	100
90) Benzo(b)fluoranthene	23.216	252	1648671	22.252	ng/ul	99
91) Benzo(k)fluoranthene	23.269	252	1661743	22.560	ng/ul	99
93) Benzo(a)pyrene	23.886	252	1488404	21.653	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.668	276	1572711	20.661	ng/ul	99
95) Dibenzo(a,h)anthracene	26.692	278	1313577	20.543	ng/ul	99
96) Benzo(g,h,i)perylene	27.504	276	1227866	20.861	ng/ul	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091823\
Data File : BP017210.D
Acq On : 18 Sep 2023 17:59
Operator : MA/JU
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTD020739

Quant Time: Sep 18 22:33:26 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP091223.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Sep 16 03:18:48 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.293 ng/uL
RT: 3.346 min Scan# 41
Delta R.T. -0.000 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59




Instrument :
BNA_P
ClientSampleId :
SSTD020739


Tgt Ion: 88 Resp: 53997
Ion Ratio Lower Upper
88 100
43 27.8 21.9 32.9
58 69.9 50.8 76.2



#5
Pyridine
Concen: 21.103 ng/uL
RT: 3.764 min Scan# 115
Delta R.T. -0.000 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59

Tgt Ion: 79 Resp: 378615
Ion Ratio Lower Upper
79 100
52 67.1 54.5 81.7
51 33.8 27.0 40.6









#6
Benzaldehyde
Concen: 24.787 ng/ul
RT: 7.105 min Scan# 683
Delta R.T. -0.000 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59




Instrument :
BNA_P
ClientSampleId :
SSTD020739


Tgt Ion: 77 Resp: 285513
Ion Ratio Lower Upper
77 100
105 95.1 75.3 112.9
106 90.9 74.6 111.8



#8
Phenol
Concen: 20.874 ng/ul
RT: 7.122 min Scan# 686
Delta R.T. -0.006 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59

Tgt Ion: 94 Resp: 455478
Ion Ratio Lower Upper
94 100
65 28.2 22.7 34.1
66 37.7 29.6 44.4









#10
Bis(2-Chloroethyl)ether
Concen: 21.235 ng/ul
RT: 7.381 min Scan# 710
Delta R.T. -0.006 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59




Instrument :
BNA_P
ClientSampleId :
SSTD020739


Tgt Ion: 93 Resp: 369569
Ion Ratio Lower Upper
93 100
63 74.5 59.5 89.3
95 32.3 25.8 38.6



#12
2-Chlorophenol
Concen: 21.146 ng/ul
RT: 7.499 min Scan# 750
Delta R.T. -0.000 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59

Tgt Ion:128 Resp: 354266
Ion Ratio Lower Upper
128 100
64 47.6 36.8 55.2
130 31.7 26.3 39.5





#13

2-Methylphenol

Concen: 20.960 ng/ul

RT: 8.387 min Scan# 901

Delta R.T. -0.000 min

Lab File: BP017210.D

Acq: 18 Sep 2023 17:59

Instrument :

BNA_P

ClientSampleId :





SSTD020739

Tgt Ion:108 Resp: 340956

Ion Ratio Lower Upper

108 100

107 91.4 72.2 108.4



#14

2,2'-oxybis(1-Chloropropane)

Concen: 13.657 ng/ul

RT: 8.458 min Scan# 913

Delta R.T. -0.000 min

Lab File: BP017210.D

Acq: 18 Sep 2023 17:59




Tgt Ion: 45 Resp: 322117


Ion Ratio Lower Upper

45 100

77 15.3 12.5 18.7

79 11.7 8.8 13.2









#16
Acetophenone
Concen: 21.272 ng/ul
RT: 8.793 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59




Instrument :
BNA_P
ClientSampleId :
SSTD020739


Tgt Ion:105 Resp: 570903
Ion Ratio Lower Upper
105 100
77 80.9 63.0 94.4
51 26.6 20.5 30.7



#17
N-Nitroso-di-n-propylamine
Concen: 21.726 ng/ul
RT: 8.763 min Scan# 965
Delta R.T. -0.006 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59

Tgt Ion: 70 Resp: 299630
Ion Ratio Lower Upper
70 100
42 58.4 43.9 65.9
101 9.1 7.0 10.4
130 19.8 15.8 23.6









#18
4-Methylphenol
Concen: 20.611 ng/ul
RT: 8.722 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59




Instrument :
BNA_P
ClientSampleId :
SSTD020739


Tgt Ion:108 Resp: 370947
Ion Ratio Lower Upper
108 100
107 110.0 91.4 137.0



#19
Hexachloroethane
Concen: 21.549 ng/ul
RT: 9.010 min Scan# 1007
Delta R.T. -0.006 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59

Tgt Ion:117 Resp: 148477
Ion Ratio Lower Upper
117 100
201 100.5 79.9 119.9
199 63.1 49.4 74.2









#22
Nitrobenzene
Concen: 21.949 ng/ul
RT: 9.187 min Scan# 1037
Delta R.T. -0.000 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59




Instrument :
BNA_P
ClientSampleId :
SSTD020739


Tgt Ion: 77 Resp: 441075
Ion Ratio Lower Upper
77 100
123 42.3 34.6 51.8
65 12.9 10.6 16.0



#23
Isophorone
Concen: 21.727 ng/ul
RT: 9.699 min Scan# 1124
Delta R.T. -0.006 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59

Tgt Ion: 82 Resp: 818175
Ion Ratio Lower Upper
82 100
95 6.9 5.6 8.4
138 16.9 14.0 21.0









#25
2-Nitrophenol
Concen: 22.105 ng/ul
RT: 9.893 min Scan# 1157
Delta R.T. -0.000 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59




Instrument :
BNA_P
ClientSampleId :
SSTD020739


Tgt Ion:139 Resp: 202393
Ion Ratio Lower Upper
139 100
65 51.2 41.9 62.9
109 30.6 23.5 35.3



#26
2,4-Dimethylphenol
Concen: 21.416 ng/ul
RT: 9.928 min Scan# 1163
Delta R.T. -0.006 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59

Tgt Ion:107 Resp: 399611
Ion Ratio Lower Upper
107 100
121 50.2 41.7 62.5
122 85.1 70.4 105.6









#27
 Bis(2-Chloroethoxy)methane
 Concen: 21.187 ng/ul
 RT: 10.187 min Scan# 11
 Delta R.T. -0.006 min
 Lab File: BP017210.D
 Acq: 18 Sep 2023 17:59




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020739


Tgt Ion: 93 Resp: 500062
 Ion Ratio Lower Upper
 93 100
 95 31.9 26.2 39.4
 123 11.2 9.3 13.9



#29
 2,4-Dichlorophenol
 Concen: 21.619 ng/ul
 RT: 10.410 min Scan# 1245
 Delta R.T. -0.000 min
 Lab File: BP017210.D
 Acq: 18 Sep 2023 17:59

Tgt Ion:162 Resp: 348634
 Ion Ratio Lower Upper
 162 100
 164 66.3 53.2 79.8
 98 35.0 30.0 45.0









#30
Naphthalene
Concen: 21.395 ng/ul
RT: 10.822 min Scan# 1315
Delta R.T. -0.000 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59




Instrument :
BNA_P
ClientSampleId :
SSTD020739


Tgt Ion:128 Resp: 1193056
Ion Ratio Lower Upper
128 100
129 10.8 8.6 13.0
127 12.8 10.4 15.6



#32
4-Chloroaniline
Concen: 21.165 ng/ul
RT: 10.957 min Scan# 1338
Delta R.T. -0.006 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59

Tgt Ion:127 Resp: 495946
Ion Ratio Lower Upper
127 100
129 32.5 24.8 37.2









#33
Hexachlorobutadiene
Concen: 21.598 ng/ul
RT: 11.046 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59




Instrument :
BNA_P
ClientSampleId :
SSTD020739


Tgt Ion:225 Resp: 236512
Ion Ratio Lower Upper
225 100
223 63.2 49.0 73.4
227 63.7 50.2 75.4



#34
Caprolactam
Concen: 22.511 ng/ul
RT: 11.781 min Scan# 1478
Delta R.T. -0.000 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59

Tgt Ion:113 Resp: 114115
Ion Ratio Lower Upper
113 100
55 160.2 130.7 196.1
56 117.7 96.8 145.2









#35
4-Chloro-3-methylphenol
Concen: 21.545 ng/ul
RT: 12.057 min Scan# 1525
Delta R.T. -0.000 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59




Instrument :
BNA_P
ClientSampleId :
SSTD020739


Tgt Ion:107 Resp: 375935
Ion Ratio Lower Upper
107 100
144 26.0 20.6 31.0
142 81.4 63.4 95.0



#36
2-Methylnaphthalene
Concen: 21.430 ng/ul
RT: 12.428 min Scan# 1588
Delta R.T. -0.000 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59

Tgt Ion:142 Resp: 806368
Ion Ratio Lower Upper
142 100
141 89.4 70.1 105.1





#37

1-Methylnaphthalene

Concen: 21.524 ng/ul

RT: 12.646 min Scan# 1625

Delta R.T. -0.006 min

Lab File: BP017210.D

Acq: 18 Sep 2023 17:59

Instrument :

BNA_P

ClientSampleId :

SSTD020739





Tgt Ion:142 Resp: 819785

Ion Ratio Lower Upper

142 100

141 92.6 74.1 111.1

116 4.3 3.1 4.7



#39

1,2,4,5-Tetrachlorobenzene

Concen: 21.946 ng/ul

RT: 12.775 min Scan# 1647

Delta R.T. -0.006 min

Lab File: BP017210.D

Acq: 18 Sep 2023 17:59

Tgt Ion:216 Resp: 454077




Ion Ratio Lower Upper


216 100

214 77.7 62.1 93.1

179 20.1 15.6 23.4

108 16.2 12.8 19.2





#40

Hexachlorocyclopentadiene

Concen: 22.698 ng/ul

RT: 12.722 min Scan# 1638

Delta R.T. -0.006 min

Lab File: BP017210.D

Acq: 18 Sep 2023 17:59

Instrument :

BNA_P

ClientSampleId :

SSTD020739





Tgt Ion:237 Resp: 292608

Ion Ratio Lower Upper

237 100

235 62.1 50.4 75.6

272 12.3 10.2 15.4



#41

2,4,6-Trichlorophenol

Concen: 22.278 ng/ul

RT: 13.028 min Scan# 1690

Delta R.T. -0.006 min

Lab File: BP017210.D

Acq: 18 Sep 2023 17:59




Tgt Ion:196 Resp: 292819


Ion Ratio Lower Upper

196 100

198 96.1 76.0 114.0

200 29.9 25.0 37.4









#42
2,4,5-Trichlorophenol
Concen: 22.467 ng/ul
RT: 13.099 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59




Instrument :
BNA_P
ClientSampleId :
SSTD020739


Tgt Ion:196 Resp: 313952
Ion Ratio Lower Upper
196 100
198 95.4 76.9 115.3
200 30.0 25.0 37.6



#43
1,1'-Biphenyl
Concen: 21.534 ng/ul
RT: 13.440 min Scan# 1760
Delta R.T. -0.000 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59

Tgt Ion:154 Resp: 1072699
Ion Ratio Lower Upper
154 100
153 39.7 31.8 47.8
76 13.1 10.8 16.2









#44
2-Chloronaphthalene
Concen: 21.697 ng/ul
RT: 13.487 min Scan# 1768
Delta R.T. -0.000 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59




Instrument :
BNA_P
ClientSampleId :
SSTD020739


Tgt Ion:162 Resp: 858303
Ion Ratio Lower Upper
162 100
164 33.0 26.6 39.8
127 37.7 31.0 46.4



#45
2-Nitroaniline
Concen: 23.112 ng/ul
RT: 13.722 min Scan# 1808
Delta R.T. -0.006 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59

Tgt Ion: 65 Resp: 247652
Ion Ratio Lower Upper
65 100
92 67.5 55.6 83.4
138 102.7 82.2 123.2









#47
Dimethylphthalate
Concen: 21.364 ng/ul
RT: 14.069 min Scan# 1867
Delta R.T. -0.000 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59




Instrument :
BNA_P
ClientSampleId :
SSTD020739


Tgt Ion:163 Resp: 1072157
Ion Ratio Lower Upper
163 100
194 4.0 3.3 4.9
164 10.0 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 22.285 ng/ul
RT: 14.210 min Scan# 1891
Delta R.T. -0.006 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59

Tgt Ion:165 Resp: 204462
Ion Ratio Lower Upper
165 100
89 65.3 47.8 71.8
121 21.6 17.4 26.2









#50
Acenaphthylene
Concen: 21.853 ng/ul
RT: 14.334 min Scan# 1912
Delta R.T. -0.000 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59




Instrument :
BNA_P
ClientSampleId :
SSTD020739


Tgt Ion:152 Resp: 1416119
Ion Ratio Lower Upper
152 100
151 20.1 15.8 23.8
153 13.1 10.6 15.8



#51
3-Nitroaniline
Concen: 21.380 ng/ul
RT: 14.551 min Scan# 1949
Delta R.T. -0.006 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59

Tgt Ion:138 Resp: 209847
Ion Ratio Lower Upper
138 100
108 11.6 8.5 12.7
92 118.7 94.2 141.4









#52
Acenaphthene
Concen: 21.610 ng/ul
RT: 14.669 min Scan# 1969
Delta R.T. -0.006 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59




Instrument :
BNA_P
ClientSampleId :
SSTD020739


Tgt Ion:153 Resp: 926682
Ion Ratio Lower Upper
153 100
152 47.4 38.6 57.8
154 88.5 70.4 105.6



#53
2,4-Dinitrophenol
Concen: 23.155 ng/ul
RT: 14.751 min Scan# 1983
Delta R.T. -0.006 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59

Tgt Ion:184 Resp: 148802
Ion Ratio Lower Upper
184 100
63 70.8 55.2 82.8
154 65.6 53.0 79.4









#55
4-Nitrophenol
Concen: 19.817 ng/ul
RT: 14.840 min Scan# 1998
Delta R.T. -0.000 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59




Instrument :
BNA_P
ClientSampleId :
SSTD020739


Tgt Ion:109 Resp: 151540
Ion Ratio Lower Upper
109 100
139 121.4 101.1 151.7
65 127.9 105.0 157.4



#56
Dibenzofuran
Concen: 21.556 ng/ul
RT: 15.004 min Scan# 2026
Delta R.T. -0.006 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59

Tgt Ion:168 Resp: 1277367
Ion Ratio Lower Upper
168 100
139 37.4 29.9 44.9









#57
2,4-Dinitrotoluene
Concen: 22.495 ng/ul
RT: 14.998 min Scan# 2024
Delta R.T. -0.000 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59




Instrument :
BNA_P
ClientSampleId :
SSTD020739


Tgt Ion:165 Resp: 307319
Ion Ratio Lower Upper
165 100
63 57.6 42.3 63.5
182 2.9 2.2 3.4



#58
2,3,4,6-Tetrachlorophenol
Concen: 22.148 ng/ul
RT: 15.222 min Scan# 2063
Delta R.T. -0.006 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59

Tgt Ion:232 Resp: 271507
Ion Ratio Lower Upper
232 100
131 44.2 37.0 55.6
130 2.6 1.8 2.8
166 31.0 24.3 36.5





#59

Diethylphthalate

Concen: 21.487 ng/ul

RT: 15.422 min Scan# 2097

Delta R.T. -0.006 min

Lab File: BP017210.D

Acq: 18 Sep 2023 17:59

Instrument :

BNA_P

ClientSampleId :

SSTD020739





Tgt Ion:149 Resp: 1064632

Ion Ratio Lower Upper

149 100

177 21.9 17.5 26.3

150 12.6 10.0 15.0



#61

Fluorene

Concen: 21.463 ng/ul

RT: 15.657 min Scan# 2137

Delta R.T. -0.006 min

Lab File: BP017210.D

Acq: 18 Sep 2023 17:59




Tgt Ion:166 Resp: 1051529


Ion Ratio Lower Upper

166 100

165 98.0 78.2 117.2

167 13.5 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 21.915 ng/ul
RT: 15.651 min Scan# 2136
Delta R.T. -0.000 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59




Instrument :
BNA_P
ClientSampleId :
SSTD020739


Tgt Ion:204 Resp: 545886
Ion Ratio Lower Upper
204 100
206 33.0 27.3 40.9
141 54.7 45.0 67.6



#63
4-Nitroaniline
Concen: 21.547 ng/ul
RT: 15.722 min Scan# 2148
Delta R.T. -0.000 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59

Tgt Ion:138 Resp: 196610
Ion Ratio Lower Upper
138 100
92 54.3 43.3 64.9
108 94.9 75.3 112.9









#66
4,6-Dinitro-2-methylphenol
Concen: 20.136 ng/ul
RT: 15.745 min Scan# 2152
Delta R.T. -0.006 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59




Instrument :
BNA_P
ClientSampleId :
SSTD020739


Tgt Ion:198 Resp: 185624
Ion Ratio Lower Upper
198 100
182 5.2 4.2 6.2
77 26.4 22.5 33.7



#67
N-Nitrosodiphenylamine
Concen: 21.752 ng/ul
RT: 15.875 min Scan# 2174
Delta R.T. -0.006 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59

Tgt Ion:169 Resp: 869565
Ion Ratio Lower Upper
169 100
168 66.6 52.7 79.1
167 36.8 29.0 43.4









#68
4-Bromophenyl-phenylether
Concen: 22.178 ng/ul
RT: 16.551 min Scan# 211
Delta R.T. -0.006 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59




Instrument :
BNA_P
ClientSampleId :
SSTD020739


Tgt Ion:248 Resp: 325030
Ion Ratio Lower Upper
248 100
250 96.0 78.6 118.0
141 71.2 60.4 90.6



#69
Hexachlorobenzene
Concen: 21.649 ng/ul
RT: 16.634 min Scan# 2303
Delta R.T. -0.000 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59

Tgt Ion:284 Resp: 375537
Ion Ratio Lower Upper
284 100
142 33.5 28.6 42.8
249 31.4 25.0 37.6





#70

Atrazine

Concen: 22.627 ng/ul

RT: 16.834 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP017210.D

Acq: 18 Sep 2023 17:59

Instrument :

BNA_P

ClientSampleId :

SSTD020739





Tgt Ion:200 Resp: 327435

Ion Ratio Lower Upper

200 100

173 26.4 20.7 31.1

215 50.2 38.8 58.2



#71

Pentachlorophenol

Concen: 20.223 ng/ul

RT: 16.998 min Scan# 2365

Delta R.T. -0.006 min

Lab File: BP017210.D

Acq: 18 Sep 2023 17:59




Tgt Ion:266 Resp: 227975


Ion Ratio Lower Upper

266 100

264 63.7 50.7 76.1

268 65.5 51.1 76.7





#72

Phenanthrene

Concen: 21.584 ng/ul

RT: 17.422 min Scan# 2436

Delta R.T. -0.000 min

Lab File: BP017210.D

Acq: 18 Sep 2023 17:59

Instrument :

BNA_P

ClientSampleId :

SSTD020739





Tgt Ion:178 Resp: 1667644

Ion Ratio Lower Upper

178 100

179 15.3 12.2 18.2

176 18.8 15.6 23.4



#74

Anthracene

Concen: 21.533 ng/ul

RT: 17.510 min Scan# 2452

Delta R.T. -0.006 min

Lab File: BP017210.D

Acq: 18 Sep 2023 17:59




Tgt Ion:178 Resp: 1685016


Ion Ratio Lower Upper

178 100

179 15.6 12.4 18.6

176 18.7 14.9 22.3









#75
1,2,3,4-Tetrachlorobenzene
Concen: 22.517 ng/uL
RT: 13.393 min Scan# 1752
Delta R.T. -0.000 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59




Instrument :
BNA_P
ClientSampleId :
SSTD020739


Tgt Ion:216 Resp: 478646
Ion Ratio Lower Upper
216 100
214 78.2 61.3 91.9
218 47.2 38.3 57.5



#76
Pentachlorobenzene
Concen: 22.465 ng/uL
RT: 14.898 min Scan# 2008
Delta R.T. -0.000 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59

Tgt Ion:250 Resp: 433324
Ion Ratio Lower Upper
250 100
248 61.9 51.1 76.7
252 63.9 50.7 76.1





#77

Carbazole

Concen: 21.461 ng/ul

RT: 17.798 min Scan# 2501

Delta R.T. -0.000 min

Lab File: BP017210.D

Acq: 18 Sep 2023 17:59

Instrument :

BNA_P

ClientSampleId :

SSTD020739





Tgt Ion:167 Resp: 1474669

Ion Ratio Lower Upper

167 100

166 21.8 17.0 25.6

139 13.1 10.4 15.6



#78

Di-n-butylphthalate

Concen: 22.832 ng/ul

RT: 18.310 min Scan# 2588

Delta R.T. -0.000 min

Lab File: BP017210.D

Acq: 18 Sep 2023 17:59




Tgt Ion:149 Resp: 1787738


Ion Ratio Lower Upper

149 100

150 9.3 7.2 10.8

104 5.4 4.2 6.4









#80
Fluoranthene
Concen: 22.961 ng/ul
RT: 19.386 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59




Instrument :
BNA_P
ClientSampleId :
SSTD020739


Tgt Ion:202 Resp: 1934677
Ion Ratio Lower Upper
202 100
101 10.4 8.3 12.5
100 8.2 6.4 9.6



#82
Pyrene
Concen: 22.887 ng/ul
RT: 19.745 min Scan# 2832
Delta R.T. -0.000 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59

Tgt Ion:202 Resp: 2037621
Ion Ratio Lower Upper
202 100
101 12.1 9.7 14.5
100 10.1 7.8 11.8





#83

Butylbenzylphthalate

Concen: 23.841 ng/ul

RT: 20.604 min Scan# 2978

Delta R.T. -0.006 min

Lab File: BP017210.D

Acq: 18 Sep 2023 17:59

Instrument :

BNA_P

ClientSampleId :

SSTD020739





Tgt Ion:149 Resp: 757026

Ion Ratio Lower Upper

149 100

91 75.0 57.4 86.0

206 20.6 16.7 25.1



#84

3,3'-Dichlorobenzidine

Concen: 19.779 ng/ul

RT: 21.404 min Scan# 3114

Delta R.T. -0.000 min

Lab File: BP017210.D

Acq: 18 Sep 2023 17:59




Tgt Ion:252 Resp: 557389


Ion Ratio Lower Upper

252 100

254 65.0 51.4 77.2

126 9.5 9.4 14.2









#85
Benzo(a)anthracene
Concen: 21.557 ng/ul
RT: 21.463 min Scan# 3124
Delta R.T. -0.000 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59




Instrument :
BNA_P
ClientSampleId :
SSTD020739


Tgt Ion:228 Resp: 1940131
Ion Ratio Lower Upper
228 100
229 19.4 16.1 24.1
226 26.8 22.2 33.4



#86
Bis(2-ethylhexyl)phthalate
Concen: 24.018 ng/ul
RT: 21.339 min Scan# 3103
Delta R.T. -0.006 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59

Tgt Ion:149 Resp: 1070573
Ion Ratio Lower Upper
149 100
167 26.8 21.3 31.9
279 4.1 3.4 5.2





#87

Chrysene

Concen: 21.351 ng/ul

RT: 21.516 min Scan# 3133

Delta R.T. -0.006 min

Lab File: BP017210.D

Acq: 18 Sep 2023 17:59

Instrument :

BNA_P

ClientSampleId :

SSTD020739





Tgt Ion:228 Resp: 1787517

Ion Ratio Lower Upper

228 100

226 30.6 23.9 35.9

229 19.9 15.3 22.9



#89

Di-n-octyl phthalate

Concen: 24.857 ng/ul




RT: 22.304 min Scan# 3267


Delta R.T. -0.000 min

Lab File: BP017210.D

Acq: 18 Sep 2023 17:59

Tgt Ion:149 Resp: 1732476









#90
Benzo(b)fluoranthene
Concen: 22.252 ng/ul
RT: 23.216 min Scan# 3422
Delta R.T. -0.000 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59




Instrument :
BNA_P
ClientSampleId :
SSTD020739


Tgt Ion:252 Resp: 1648671
Ion Ratio Lower Upper
252 100
253 22.0 18.2 27.2
125 9.3 7.5 11.3



#91
Benzo(k)fluoranthene
Concen: 22.560 ng/ul
RT: 23.269 min Scan# 3431
Delta R.T. -0.000 min
Lab File: BP017210.D
Acq: 18 Sep 2023 17:59

Tgt Ion:252 Resp: 1661743
Ion Ratio Lower Upper
252 100
253 21.9 17.8 26.6
125 9.7 7.4 11.2





#93

Benzo(a)pyrene

Concen: 21.653 ng/ul

RT: 23.886 min Scan# 3

Delta R.T. -0.006 min

Lab File: BP017210.D

Acq: 18 Sep 2023 17:59

Instrument :

BNA_P

ClientSampleId :

SSTD020739





Tgt Ion:252 Resp: 1488404

Ion Ratio Lower Upper

252 100

253 21.8 17.4 26.2

125 10.5 8.1 12.1



#94

Indeno(1,2,3-cd)pyrene

Concen: 20.661 ng/ul

RT: 26.668 min Scan# 4009

Delta R.T. -0.006 min

Lab File: BP017210.D

Acq: 18 Sep 2023 17:59




Tgt Ion:276 Resp: 1572711


Ion Ratio Lower Upper

276 100

138 21.9 17.5 26.3

277 25.3 19.6 29.4





#95

Dibenzo(a,h)anthracene

Concen: 20.543 ng/ul

RT: 26.692 min Scan# 4013

Delta R.T. -0.006 min

Lab File: BP017210.D

Acq: 18 Sep 2023 17:59

Instrument :

BNA_P

ClientSampleId :

SSTD020739





Tgt Ion:278 Resp: 1313577

Ion Ratio Lower Upper

278 100

139 15.3 11.9 17.9

279 23.9 19.0 28.6



#96

Benzo(g,h,i)perylene

Concen: 20.861 ng/ul

RT: 27.504 min Scan# 4151

Delta R.T. -0.006 min

Lab File: BP017210.D

Acq: 18 Sep 2023 17:59

Tgt Ion:276 Resp: 1227866

Ion Ratio Lower Upper

276 100

138 20.7 16.2 24.4

277 23.1 19.6 29.4

