

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091823\
 Data File : BP017198.D
 Acq On : 18 Sep 2023 11:06
 Operator : MA/JU
 Sample : PB155449BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK449

Quant Time: Sep 18 22:39:31 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP091223.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Sep 16 03:18:48 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.940	152	179593	20.000	ng/u1	0.00
20) Naphthalene-d8	10.769	136	753777	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.604	164	468526	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.375	188	1038119	20.000	ng/u1	0.00
79) Chrysene-d12	21.475	240	755714	20.000	ng/u1	0.00
88) Perylene-d12	23.998	264	729197	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.305	96	25545	5.977	ng/uL	0.00
4) Pyridine-d5	3.740	84	376786	30.820	ng/u1	0.00
7) Phenol-d5	7.093	99	440029	30.132	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.281	67	298558	32.710	ng/u1	0.00
11) 2-Chlorophenol-d4	7.464	132	347100	29.794	ng/u1	0.00
15) 4-Methylphenol-d8	8.652	113	349971	28.917	ng/u1	0.00
21) Nitrobenzene-d5	9.140	128	165477	31.351	ng/u1	0.00
24) 2-Nitrophenol-d4	9.858	143	177256	31.457	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.381	165	328121	30.110	ng/u1	0.00
31) 4-Chloroaniline-d4	10.928	131	499778	30.427	ng/u1	0.00
46) Dimethylphthalate-d6	14.016	166	1148154	34.436	ng/u1	0.00
49) Acenaphthylene-d8	14.304	160	1206265	31.885	ng/u1	0.00
54) 4-Nitrophenol-d4	14.822	143	131189	20.359	ng/u1	0.00
60) Fluorene-d10	15.598	176	973960	33.784	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.728	200	109643	18.904	ng/u1	0.00
73) Anthracene-d10	17.475	188	1561578	35.404	ng/u1	0.00
81) Pyrene-d10	19.716	212	1737374	43.836	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.833	264	1225020	35.594	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091823\
Data File : BP017198.D
Acq On : 18 Sep 2023 11:06
Operator : MA/JU
Sample : PB155449BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK449

Quant Time: Sep 18 22:39:31 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP091223.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Sep 16 03:18:48 2023
Response via : Initial Calibration

