




Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091925\
Data File : BP025791.D
Acq On : 19 Sep 2025 09:31
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 19 11:12:57 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 11 16:45:04 2025
Response via : Initial Calibration



TIC: BP025791.D\data.ms

(70) Pentachlorophenol (C)




16.825min (-0.006) 12969.34 ng

response 772758

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.30	64.28
264.00	62.70	63.32
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP091925\
Data File : BP025791.D
Acq On : 19 Sep 2025 09:31
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 19 11:12:57 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP091225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 11 16:45:04 2025
Response via : Initial Calibration



TIC: BP025791.D\data.ms

(77) Benzidine

19.501min (-0.006) 14332.79 ng

response 4314810

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.40	14.17
183.00	11.50	11.63
0.00	0.00	0.00