

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092023\
 Data File : BP017243.D
 Acq On : 20 Sep 2023 17:26
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020741

Quant Time: Sep 20 22:36:41 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP092023.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Sep 20 14:34:55 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.940	152	175253	20.000	ng/u1	0.00
20) Naphthalene-d8	10.763	136	722618	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.598	164	433510	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.374	188	925878	20.000	ng/u1	0.00
79) Chrysene-d12	21.474	240	904249	20.000	ng/u1	-0.01
88) Perylene-d12	23.998	264	1027148	20.000	ng/u1	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.305	96	34055	7.982	ng/uL	0.00
4) Pyridine-d5	3.740	84	244937	20.534	ng/u1	0.00
7) Phenol-d5	7.093	99	295715	20.545	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.281	67	181216	20.862	ng/u1	0.00
11) 2-Chlorophenol-d4	7.457	132	235419	20.622	ng/u1	0.00
15) 4-Methylphenol-d8	8.651	113	230095	20.562	ng/u1	0.00
21) Nitrobenzene-d5	9.140	128	110348	21.123	ng/u1	0.00
24) 2-Nitrophenol-d4	9.851	143	118814	20.946	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.375	165	229081	21.376	ng/u1	0.00
31) 4-Chloroaniline-d4	10.928	131	325455	20.964	ng/u1	0.00
46) Dimethylphthalate-d6	14.016	166	650859	20.958	ng/u1	0.00
49) Acenaphthylene-d8	14.298	160	756710	21.186	ng/u1	0.00
54) 4-Nitrophenol-d4	14.822	143	108734	20.097	ng/u1	0.00
60) Fluorene-d10	15.598	176	559303	20.920	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.727	200	101440	19.439	ng/u1	0.00
73) Anthracene-d10	17.474	188	873969	21.283	ng/u1	0.00
81) Pyrene-d10	19.715	212	1024090	20.943	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.833	264	1033456	20.897	ng/u1	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.340	88	34505	7.897	ng/uL	90
5) Pyridine	3.757	79	254810	20.577	ng/u1	99
6) Benzaldehyde	7.098	77	177994	24.232	ng/u1	97
8) Phenol	7.122	94	299018	20.655	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.375	93	243491	20.807	ng/u1	96
12) 2-Chlorophenol	7.493	128	240084	20.709	ng/u1	98
13) 2-Methylphenol	8.381	108	217733	20.325	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.451	45	188460	12.734	ng/u1	99
16) Acetophenone	8.793	105	369537	21.370	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.757	70	183332	21.244	ng/u1	97
18) 4-Methylphenol	8.716	108	242833	21.157	ng/u1	100
19) Hexachloroethane	9.004	117	99219	20.616	ng/u1	97
22) Nitrobenzene	9.181	77	276992	20.896	ng/u1	97
23) Isophorone	9.692	82	500863	21.086	ng/u1	99
25) 2-Nitrophenol	9.887	139	131933	21.464	ng/u1	98
26) 2,4-Dimethylphenol	9.922	107	259838	21.216	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.181	93	320872	20.926	ng/u1	97
29) 2,4-Dichlorophenol	10.404	162	225310	21.324	ng/u1	99
30) Naphthalene	10.816	128	787838	20.939	ng/u1	99
32) 4-Chloroaniline	10.951	127	316269	21.259	ng/u1	100
33) Hexachlorobutadiene	11.039	225	153643	20.893	ng/u1	98
34) Caprolactam	11.769	113	61538	20.156	ng/u1	99
35) 4-Chloro-3-methylphenol	12.051	107	228230	21.489	ng/u1	99
36) 2-Methylnaphthalene	12.422	142	518370	21.108	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092023\
 Data File : BP017243.D
 Acq On : 20 Sep 2023 17:26
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020741

Quant Time: Sep 20 22:36:41 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP092023.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Sep 20 14:34:55 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.639	142	533060	21.234	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.769	216	287295	20.650	ng/ul	99
40) Hexachlorocyclopentadiene	12.716	237	145929	18.535	ng/ul	99
41) 2,4,6-Trichlorophenol	13.028	196	175901	20.844	ng/ul	99
42) 2,4,5-Trichlorophenol	13.098	196	187907	21.253	ng/ul	100
43) 1,1'-Biphenyl	13.433	154	681741	20.727	ng/ul	99
44) 2-Chloronaphthalene	13.480	162	544328	20.875	ng/ul	99
45) 2-Nitroaniline	13.716	65	136283	21.275	ng/ul	99
47) Dimethylphthalate	14.063	163	650082	20.781	ng/ul	99
48) 2,6-Dinitrotoluene	14.210	165	116415	21.425	ng/ul	93
50) Acenaphthylene	14.328	152	876386	20.983	ng/ul	100
51) 3-Nitroaniline	14.545	138	122748	20.992	ng/ul	99
52) Acenaphthene	14.663	153	572811	20.767	ng/ul	100
53) 2,4-Dinitrophenol	14.751	184	60868	17.908	ng/ul	97
55) 4-Nitrophenol	14.839	109	86171	20.387	ng/ul	97
56) Dibenzofuran	15.004	168	800360	21.188	ng/ul	99
57) 2,4-Dinitrotoluene	14.992	165	175813	21.967	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	15.216	232	160932	21.452	ng/ul	97
59) Diethylphthalate	15.416	149	637255	21.071	ng/ul	99
61) Fluorene	15.651	166	649467	21.162	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.645	204	330280	21.234	ng/ul	99
63) 4-Nitroaniline	15.716	138	116756	22.372	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.745	198	112051	19.904	ng/ul	99
67) N-Nitrosodiphenylamine	15.869	169	530918	20.847	ng/ul	99
68) 4-Bromophenyl-phenylether	16.551	248	192156	20.658	ng/ul	100
69) Hexachlorobenzene	16.633	284	225288	20.873	ng/ul	96
70) Atrazine	16.827	200	186325	20.421	ng/ul	98
71) Pentachlorophenol	16.998	266	131619	19.741	ng/ul	99
72) Phenanthrene	17.416	178	1013874	20.907	ng/ul	99
74) Anthracene	17.510	178	1054101	21.447	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.386	216	289032	20.340	ng/uL	97
76) Pentachlorobenzene	14.892	250	277432	20.668	ng/uL	98
77) Carbazole	17.798	167	919664	21.037	ng/ul	99
78) Di-n-butylphthalate	18.304	149	1030738	21.442	ng/ul	99
80) Fluoranthene	19.386	202	1197532	20.999	ng/ul	98
82) Pyrene	19.739	202	1262039	20.817	ng/ul	98
83) Butylbenzylphthalate	20.604	149	455688	21.199	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.404	252	406741	20.703	ng/ul	99
85) Benzo(a)anthracene	21.462	228	1281364	20.996	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.339	149	665675	21.444	ng/ul	99
87) Chrysene	21.515	228	1206721	21.010	ng/ul	99
89) Di-n-octyl phthalate	22.298	149	1161406	20.556	ng/ul	100
90) Benzo(b)fluoranthene	23.215	252	1269472	20.963	ng/ul	100
91) Benzo(k)fluoranthene	23.268	252	1268575	20.763	ng/ul	99
93) Benzo(a)pyrene	23.880	252	1215089	20.971	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.668	276	1494773	20.582	ng/ul	100
95) Dibenzo(a,h)anthracene	26.692	278	1249846	20.662	ng/ul	98
96) Benzo(g,h,i)perylene	27.503	276	1197196	20.441	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092023\
 Data File : BP017243.D
 Acq On : 20 Sep 2023 17:26
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020741

Quant Time: Sep 20 22:36:41 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP092023.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Sep 20 14:34:55 2023
 Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 7.897 ng/uL
 RT: 3.340 min Scan# 41
 Delta R.T. -0.000 min
 Lab File: BP017243.D
 Acq: 20 Sep 2023 17:26




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020741


Tgt Ion:	88	Resp:	34505
Ion	Ratio	Lower	Upper
88	100		
43	26.9	25.0	37.4
58	61.9	56.6	84.8



#5
 Pyridine
 Concen: 20.577 ng/uL
 RT: 3.757 min Scan# 114
 Delta R.T. -0.000 min
 Lab File: BP017243.D
 Acq: 20 Sep 2023 17:26

Tgt Ion:	79	Resp:	254810
Ion	Ratio	Lower	Upper
79	100		
52	63.3	51.4	77.2
51	30.9	25.4	38.2









#6
Benzaldehyde
Concen: 24.232 ng/ul
RT: 7.098 min Scan# 61
Delta R.T. -0.000 min
Lab File: BP017243.D
Acq: 20 Sep 2023 17:26




Instrument :
BNA_P
ClientSampleId :
SSTD020741


Tgt Ion: 77 Resp: 177994
Ion Ratio Lower Upper
77 100
105 95.7 78.0 117.0
106 95.0 73.2 109.8



#8
Phenol
Concen: 20.655 ng/ul
RT: 7.122 min Scan# 686
Delta R.T. -0.000 min
Lab File: BP017243.D
Acq: 20 Sep 2023 17:26

Tgt Ion: 94 Resp: 299018
Ion Ratio Lower Upper
94 100
65 27.9 21.4 32.2
66 35.1 28.2 42.2









#10
Bis(2-Chloroethyl)ether
Concen: 20.807 ng/ul
RT: 7.375 min Scan# 729
Delta R.T. -0.006 min
Lab File: BP017243.D
Acq: 20 Sep 2023 17:26




Instrument :
BNA_P
ClientSampleId :
SSTD020741


Tgt Ion: 93 Resp: 243491
Ion Ratio Lower Upper
93 100
63 73.9 55.4 83.0
95 32.1 25.1 37.7



#12
2-Chlorophenol
Concen: 20.709 ng/ul
RT: 7.493 min Scan# 749
Delta R.T. -0.000 min
Lab File: BP017243.D
Acq: 20 Sep 2023 17:26

Tgt Ion:128 Resp: 240084
Ion Ratio Lower Upper
128 100
64 45.1 35.8 53.8
130 33.7 25.6 38.4





#13

2-Methylphenol

Concen: 20.325 ng/ul

RT: 8.381 min Scan# 901

Delta R.T. -0.006 min

Lab File: BP017243.D

Acq: 20 Sep 2023 17:26

Instrument :

BNA_P

ClientSampleId :





SSTD020741

Tgt Ion:108 Resp: 217733

Ion Ratio Lower Upper

108 100

107 91.6 73.5 110.3



#14

2,2'-oxybis(1-Chloropropane)

Concen: 12.734 ng/ul

RT: 8.451 min Scan# 912

Delta R.T. -0.000 min

Lab File: BP017243.D

Acq: 20 Sep 2023 17:26




Tgt Ion: 45 Resp: 188460


Ion Ratio Lower Upper

45 100

77 16.8 13.3 19.9

79 12.4 9.4 14.2









#16
Acetophenone
Concen: 21.370 ng/ul
RT: 8.793 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP017243.D
Acq: 20 Sep 2023 17:26




Instrument :
BNA_P
ClientSampleId :
SSTD020741


Tgt Ion:105 Resp: 369537
Ion Ratio Lower Upper
105 100
77 76.4 63.6 95.4
51 22.5 19.2 28.8



#17
N-Nitroso-di-n-propylamine
Concen: 21.244 ng/ul
RT: 8.757 min Scan# 964
Delta R.T. -0.000 min
Lab File: BP017243.D
Acq: 20 Sep 2023 17:26

Tgt Ion: 70 Resp: 183332
Ion Ratio Lower Upper
70 100
42 55.9 42.3 63.5
101 10.2 7.5 11.3
130 22.1 17.3 25.9









#18
4-Methylphenol
Concen: 21.157 ng/ul
RT: 8.716 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP017243.D
Acq: 20 Sep 2023 17:26




Instrument :
BNA_P
ClientSampleId :
SSTD020741


Tgt Ion:108 Resp: 242833
Ion Ratio Lower Upper
108 100
107 112.5 90.0 135.0



#19
Hexachloroethane
Concen: 20.616 ng/ul
RT: 9.004 min Scan# 1006
Delta R.T. -0.000 min
Lab File: BP017243.D
Acq: 20 Sep 2023 17:26

Tgt Ion:117 Resp: 99219
Ion Ratio Lower Upper
117 100
201 104.5 81.4 122.0
199 69.0 52.4 78.6









#22
Nitrobenzene
Concen: 20.896 ng/ul
RT: 9.181 min Scan# 1036
Delta R.T. -0.006 min
Lab File: BP017243.D
Acq: 20 Sep 2023 17:26




Instrument :
BNA_P
ClientSampleId :
SSTD020741


Tgt Ion: 77 Resp: 276992
Ion Ratio Lower Upper
77 100
123 44.3 37.4 56.0
65 12.6 9.8 14.6



#23
Isophorone
Concen: 21.086 ng/ul
RT: 9.692 min Scan# 1123
Delta R.T. -0.006 min
Lab File: BP017243.D
Acq: 20 Sep 2023 17:26

Tgt Ion: 82 Resp: 500863
Ion Ratio Lower Upper
82 100
95 6.7 5.5 8.3
138 17.6 14.2 21.4









#25
2-Nitrophenol
Concen: 21.464 ng/ul
RT: 9.887 min Scan# 1156
Delta R.T. -0.000 min
Lab File: BP017243.D
Acq: 20 Sep 2023 17:26




Instrument :
BNA_P
ClientSampleId :
SSTD020741


Tgt Ion:139 Resp: 131933
Ion Ratio Lower Upper
139 100
65 49.0 40.6 60.8
109 29.7 24.2 36.4



#26
2,4-Dimethylphenol
Concen: 21.216 ng/ul
RT: 9.922 min Scan# 1162
Delta R.T. -0.006 min
Lab File: BP017243.D
Acq: 20 Sep 2023 17:26

Tgt Ion:107 Resp: 259838
Ion Ratio Lower Upper
107 100
121 52.2 41.1 61.7
122 85.2 69.4 104.0





#27

Bis(2-Chloroethoxy)methane

Concen: 20.926 ng/ul

RT: 10.181 min Scan# 1206

Delta R.T. -0.006 min

Lab File: BP017243.D

Acq: 20 Sep 2023 17:26

Instrument :

BNA_P

ClientSampleId :

SSTD020741





Tgt Ion: 93 Resp: 320872

Ion Ratio Lower Upper

93 100

95 32.8 24.9 37.3

123 11.2 9.4 14.2



#29

2,4-Dichlorophenol

Concen: 21.324 ng/ul

RT: 10.404 min Scan# 1244

Delta R.T. -0.006 min

Lab File: BP017243.D

Acq: 20 Sep 2023 17:26




Tgt Ion: 162 Resp: 225310


Ion Ratio Lower Upper

162 100

164 64.6 52.6 78.8

98 33.9 27.0 40.4









#30
Naphthalene
Concen: 20.939 ng/ul
RT: 10.816 min Scan# 1314
Delta R.T. -0.000 min
Lab File: BP017243.D
Acq: 20 Sep 2023 17:26




Instrument :
BNA_P
ClientSampleId :
SSTD020741


Tgt Ion:128 Resp: 787838
Ion Ratio Lower Upper
128 100
129 10.5 8.6 13.0
127 12.9 10.6 15.8



#32
4-Chloroaniline
Concen: 21.259 ng/ul
RT: 10.951 min Scan# 1337
Delta R.T. -0.006 min
Lab File: BP017243.D
Acq: 20 Sep 2023 17:26

Tgt Ion:127 Resp: 316269
Ion Ratio Lower Upper
127 100
129 31.4 25.1 37.7









#33
Hexachlorobutadiene
Concen: 20.893 ng/ul
RT: 11.039 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP017243.D
Acq: 20 Sep 2023 17:26




Instrument :
BNA_P
ClientSampleId :
SSTD020741


Tgt Ion:225 Resp: 153643
Ion Ratio Lower Upper
225 100
223 63.1 48.2 72.2
227 63.1 50.2 75.4



#34
Caprolactam
Concen: 20.156 ng/ul
RT: 11.769 min Scan# 1476
Delta R.T. -0.006 min
Lab File: BP017243.D
Acq: 20 Sep 2023 17:26

Tgt Ion:113 Resp: 61538
Ion Ratio Lower Upper
113 100
55 161.3 127.4 191.0
56 117.8 94.5 141.7









#35
4-Chloro-3-methylphenol
Concen: 21.489 ng/ul
RT: 12.051 min Scan# 1524
Delta R.T. -0.006 min
Lab File: BP017243.D
Acq: 20 Sep 2023 17:26




Instrument :
BNA_P
ClientSampleId :
SSTD020741


Tgt Ion:107 Resp: 228230
Ion Ratio Lower Upper
107 100
144 26.7 20.6 31.0
142 78.2 62.9 94.3



#36
2-Methylnaphthalene
Concen: 21.108 ng/ul
RT: 12.422 min Scan# 1587
Delta R.T. -0.006 min
Lab File: BP017243.D
Acq: 20 Sep 2023 17:26

Tgt Ion:142 Resp: 518370
Ion Ratio Lower Upper
142 100
141 89.2 71.9 107.9





#37

1-Methylnaphthalene

Concen: 21.234 ng/ul

RT: 12.639 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP017243.D

Acq: 20 Sep 2023 17:26

Instrument :

BNA_P

ClientSampleId :

SSTD020741





Tgt Ion:142 Resp: 533060

Ion Ratio Lower Upper

142 100

141 92.0 73.4 110.2

116 3.8 3.1 4.7



#39

1,2,4,5-Tetrachlorobenzene

Concen: 20.650 ng/ul

RT: 12.769 min Scan# 1646

Delta R.T. -0.006 min

Lab File: BP017243.D

Acq: 20 Sep 2023 17:26

Tgt Ion:216 Resp: 287295




Ion Ratio Lower Upper


216 100

214 79.2 62.3 93.5

179 21.4 16.4 24.6

108 17.5 13.8 20.8





#40

Hexachlorocyclopentadiene

Concen: 18.535 ng/ul

RT: 12.716 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP017243.D

Acq: 20 Sep 2023 17:26

Instrument :

BNA_P

ClientSampleId :

SSTD020741





Tgt Ion:237 Resp: 145929

Ion Ratio Lower Upper

237 100

235 64.9 52.8 79.2

272 12.5 9.8 14.6



#41

2,4,6-Trichlorophenol

Concen: 20.844 ng/ul

RT: 13.028 min Scan# 1690

Delta R.T. -0.006 min

Lab File: BP017243.D

Acq: 20 Sep 2023 17:26




Tgt Ion:196 Resp: 175901


Ion Ratio Lower Upper

196 100

198 96.3 76.2 114.2

200 30.6 25.1 37.7





#42

2,4,5-Trichlorophenol

Concen: 21.253 ng/ul

RT: 13.098 min Scan# 1702

Delta R.T. -0.006 min

Lab File: BP017243.D

Acq: 20 Sep 2023 17:26

Instrument :

BNA_P

ClientSampleId :

SSTD020741





Tgt Ion:196 Resp: 187907

Ion Ratio Lower Upper

196 100

198 96.8 77.4 116.0

200 31.2 24.3 36.5



#43

1,1'-Biphenyl

Concen: 20.727 ng/ul

RT: 13.433 min Scan# 1759

Delta R.T. -0.000 min

Lab File: BP017243.D

Acq: 20 Sep 2023 17:26




Tgt Ion:154 Resp: 681741


Ion Ratio Lower Upper

154 100

153 39.4 31.5 47.3

76 12.9 11.0 16.6





#44

2-Chloronaphthalene

Concen: 20.875 ng/ul

RT: 13.480 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP017243.D

Acq: 20 Sep 2023 17:26

Instrument :

BNA_P

ClientSampleId :

SSTD020741





Tgt Ion:162 Resp: 544328

Ion Ratio Lower Upper

162 100

164 32.2 26.8 40.2

127 38.4 30.7 46.1



#45

2-Nitroaniline

Concen: 21.275 ng/ul

RT: 13.716 min Scan# 1807

Delta R.T. -0.006 min

Lab File: BP017243.D

Acq: 20 Sep 2023 17:26




Tgt Ion: 65 Resp: 136283


Ion Ratio Lower Upper

65 100

92 70.2 57.7 86.5

138 105.3 83.5 125.3





#47

Dimethylphthalate

Concen: 20.781 ng/ul

RT: 14.063 min Scan# 1866

Delta R.T. -0.006 min

Lab File: BP017243.D

Acq: 20 Sep 2023 17:26

Instrument :

BNA_P

ClientSampleId :

SSTD020741





Tgt Ion:163 Resp: 650082

Ion Ratio Lower Upper

163 100

194 4.1 3.2 4.8

164 9.8 7.5 11.3



#48

2,6-Dinitrotoluene

Concen: 21.425 ng/ul

RT: 14.210 min Scan# 1891

Delta R.T. -0.000 min

Lab File: BP017243.D

Acq: 20 Sep 2023 17:26




Tgt Ion:165 Resp: 116415


Ion Ratio Lower Upper

165 100

89 58.6 51.7 77.5

121 21.3 19.7 29.5









#50
Acenaphthylene
Concen: 20.983 ng/ul
RT: 14.328 min Scan# 1911
Delta R.T. -0.006 min
Lab File: BP017243.D
Acq: 20 Sep 2023 17:26




Instrument :
BNA_P
ClientSampleId :
SSTD020741


Tgt Ion:152 Resp: 876386
Ion Ratio Lower Upper
152 100
151 20.2 16.0 24.0
153 13.0 10.4 15.6



#51
3-Nitroaniline
Concen: 20.992 ng/ul
RT: 14.545 min Scan# 1948
Delta R.T. -0.006 min
Lab File: BP017243.D
Acq: 20 Sep 2023 17:26

Tgt Ion:138 Resp: 122748
Ion Ratio Lower Upper
138 100
108 11.3 8.2 12.4
92 122.1 97.4 146.0









#52
Acenaphthene
Concen: 20.767 ng/ul
RT: 14.663 min Scan# 1968
Delta R.T. -0.006 min
Lab File: BP017243.D
Acq: 20 Sep 2023 17:26




Instrument :
BNA_P
ClientSampleId :
SSTD020741


Tgt Ion:153 Resp: 572811
Ion Ratio Lower Upper
153 100
152 48.3 38.0 57.0
154 87.6 70.1 105.1



#53
2,4-Dinitrophenol
Concen: 17.908 ng/ul
RT: 14.751 min Scan# 1983
Delta R.T. -0.006 min
Lab File: BP017243.D
Acq: 20 Sep 2023 17:26

Tgt Ion:184 Resp: 60868
Ion Ratio Lower Upper
184 100
63 66.4 53.2 79.8
154 66.3 48.9 73.3









#55
4-Nitrophenol
Concen: 20.387 ng/ul
RT: 14.839 min Scan# 1998
Delta R.T. -0.000 min
Lab File: BP017243.D
Acq: 20 Sep 2023 17:26




Instrument :
BNA_P
ClientSampleId :
SSTD020741


Tgt Ion:109 Resp: 86171
Ion Ratio Lower Upper
109 100
139 132.9 105.7 158.5
65 126.1 105.4 158.0



#56
Dibenzofuran
Concen: 21.188 ng/ul
RT: 15.004 min Scan# 2026
Delta R.T. -0.000 min
Lab File: BP017243.D
Acq: 20 Sep 2023 17:26

Tgt Ion:168 Resp: 800360
Ion Ratio Lower Upper
168 100
139 37.4 30.6 45.8





#57

2,4-Dinitrotoluene

Concen: 21.967 ng/ul

RT: 14.992 min Scan# 20

Delta R.T. -0.006 min

Lab File: BP017243.D

Acq: 20 Sep 2023 17:26

Instrument :

BNA_P

ClientSampleId :

SSTD020741





Tgt Ion:165 Resp: 175813

Ion Ratio Lower Upper

165 100

63 57.6 45.6 68.4

182 2.7 2.0 3.0



#58

2,3,4,6-Tetrachlorophenol

Concen: 21.452 ng/ul

RT: 15.216 min Scan# 2062

Delta R.T. -0.006 min

Lab File: BP017243.D

Acq: 20 Sep 2023 17:26

Tgt Ion:232 Resp: 160932




Ion Ratio Lower Upper


232 100

131 45.3 38.4 57.6

130 2.1 2.0 3.0

166 31.3 24.6 36.8









#59
Diethylphthalate
Concen: 21.071 ng/ul
RT: 15.416 min Scan# 2096
Delta R.T. -0.006 min
Lab File: BP017243.D
Acq: 20 Sep 2023 17:26




Instrument :
BNA_P
ClientSampleId :
SSTD020741


Tgt Ion:149 Resp: 637255
Ion Ratio Lower Upper
149 100
177 21.6 17.0 25.6
150 12.8 9.8 14.6



#61
Fluorene
Concen: 21.162 ng/ul
RT: 15.651 min Scan# 2136
Delta R.T. -0.006 min
Lab File: BP017243.D
Acq: 20 Sep 2023 17:26

Tgt Ion:166 Resp: 649467
Ion Ratio Lower Upper
166 100
165 96.7 77.8 116.6
167 13.5 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 21.234 ng/ul
RT: 15.645 min Scan# 2135
Delta R.T. -0.006 min
Lab File: BP017243.D
Acq: 20 Sep 2023 17:26




Instrument :
BNA_P
ClientSampleId :
SSTD020741


Tgt Ion:204 Resp: 330280
Ion Ratio Lower Upper
204 100
206 33.2 26.2 39.4
141 56.3 45.9 68.9



#63
4-Nitroaniline
Concen: 22.372 ng/ul
RT: 15.716 min Scan# 2147
Delta R.T. -0.006 min
Lab File: BP017243.D
Acq: 20 Sep 2023 17:26

Tgt Ion:138 Resp: 116756
Ion Ratio Lower Upper
138 100
92 52.9 42.4 63.6
108 89.5 73.0 109.6









#66
4,6-Dinitro-2-methylphenol
Concen: 19.904 ng/ul
RT: 15.745 min Scan# 2152
Delta R.T. -0.006 min
Lab File: BP017243.D
Acq: 20 Sep 2023 17:26




Instrument :
BNA_P
ClientSampleId :
SSTD020741


Tgt Ion:198 Resp: 112051
Ion Ratio Lower Upper
198 100
182 5.2 4.3 6.5
77 25.1 20.4 30.6



#67
N-Nitrosodiphenylamine
Concen: 20.847 ng/ul
RT: 15.869 min Scan# 2173
Delta R.T. -0.006 min
Lab File: BP017243.D
Acq: 20 Sep 2023 17:26

Tgt Ion:169 Resp: 530918
Ion Ratio Lower Upper
169 100
168 66.8 53.5 80.3
167 36.6 30.0 45.0









#68
4-Bromophenyl-phenylether
Concen: 20.658 ng/ul
RT: 16.551 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP017243.D
Acq: 20 Sep 2023 17:26




Instrument :
BNA_P
ClientSampleId :
SSTD020741


Tgt Ion:248 Resp: 192156
Ion Ratio Lower Upper
248 100
250 98.7 78.6 117.8
141 70.5 56.3 84.5



#69
Hexachlorobenzene
Concen: 20.873 ng/ul
RT: 16.633 min Scan# 2303
Delta R.T. -0.000 min
Lab File: BP017243.D
Acq: 20 Sep 2023 17:26

Tgt Ion:284 Resp: 225288
Ion Ratio Lower Upper
284 100
142 32.9 29.5 44.3
249 32.3 25.7 38.5





#70

Atrazine

Concen: 20.421 ng/ul

RT: 16.827 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP017243.D

Acq: 20 Sep 2023 17:26

Instrument :

BNA_P

ClientSampleId :

SSTD020741





Tgt Ion:200 Resp: 186325

Ion Ratio Lower Upper

200 100

173 27.1 20.3 30.5

215 47.0 38.2 57.4



#71

Pentachlorophenol

Concen: 19.741 ng/ul

RT: 16.998 min Scan# 2365

Delta R.T. -0.006 min

Lab File: BP017243.D

Acq: 20 Sep 2023 17:26




Tgt Ion:266 Resp: 131619


Ion Ratio Lower Upper

266 100

264 62.5 48.9 73.3

268 64.4 51.6 77.4





#72

Phenanthrene

Concen: 20.907 ng/ul

RT: 17.416 min Scan# 2436

Delta R.T. -0.006 min

Lab File: BP017243.D

Acq: 20 Sep 2023 17:26

Instrument :

BNA_P

ClientSampleId :

SSTD020741





Tgt Ion:178 Resp: 1013874

Ion Ratio Lower Upper

178 100

179 15.3 12.7 19.1

176 19.5 15.8 23.6



#74

Anthracene

Concen: 21.447 ng/ul

RT: 17.510 min Scan# 2452

Delta R.T. -0.006 min

Lab File: BP017243.D

Acq: 20 Sep 2023 17:26




Tgt Ion:178 Resp: 1054101


Ion Ratio Lower Upper

178 100

179 15.5 12.5 18.7

176 18.8 15.1 22.7





#75

1,2,3,4-Tetrachlorobenzene

Concen: 20.340 ng/uL

RT: 13.386 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP017243.D

Acq: 20 Sep 2023 17:26

Instrument :

BNA_P

ClientSampleId :

SSTD020741





Tgt Ion:216 Resp: 289032

Ion Ratio Lower Upper

216 100

214 78.2 61.2 91.8

218 48.3 36.8 55.2



#76

Pentachlorobenzene

Concen: 20.668 ng/uL

RT: 14.892 min Scan# 2007

Delta R.T. -0.006 min

Lab File: BP017243.D

Acq: 20 Sep 2023 17:26




Tgt Ion:250 Resp: 277432


Ion Ratio Lower Upper

250 100

248 64.4 50.1 75.1

252 63.4 52.2 78.2





#77

Carbazole

Concen: 21.037 ng/ul

RT: 17.798 min Scan# 2501

Delta R.T. -0.000 min

Lab File: BP017243.D

Acq: 20 Sep 2023 17:26

Instrument :

BNA_P

ClientSampleId :

SSTD020741





Tgt Ion:167 Resp: 919664

Ion Ratio Lower Upper

167 100

166 21.2 16.5 24.7

139 12.4 10.5 15.7



#78

Di-n-butylphthalate

Concen: 21.442 ng/ul

RT: 18.304 min Scan# 2587

Delta R.T. -0.006 min

Lab File: BP017243.D

Acq: 20 Sep 2023 17:26




Tgt Ion:149 Resp: 1030738


Ion Ratio Lower Upper

149 100

150 9.2 7.5 11.3

104 5.4 4.5 6.7





#80

Fluoranthene

Concen: 20.999 ng/ul

RT: 19.386 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP017243.D

Acq: 20 Sep 2023 17:26

Instrument :

BNA_P

ClientSampleId :

SSTD020741





Tgt Ion:202 Resp: 1197532

Ion Ratio Lower Upper

202 100

101 10.0 8.4 12.6

100 7.7 6.6 10.0



#82

Pyrene

Concen: 20.817 ng/ul

RT: 19.739 min Scan# 2831

Delta R.T. -0.012 min

Lab File: BP017243.D

Acq: 20 Sep 2023 17:26




Tgt Ion:202 Resp: 1262039


Ion Ratio Lower Upper

202 100

101 13.0 9.8 14.8

100 10.9 8.1 12.1





#83

Butylbenzylphthalate

Concen: 21.199 ng/ul

RT: 20.604 min Scan# 2978

Delta R.T. -0.006 min

Lab File: BP017243.D

Acq: 20 Sep 2023 17:26

Instrument :

BNA_P

ClientSampleId :

SSTD020741





Tgt Ion:149 Resp: 455688

Ion Ratio Lower Upper

149 100

91 69.8 59.8 89.6

206 21.0 15.7 23.5



#84

3,3'-Dichlorobenzidine

Concen: 20.703 ng/ul

RT: 21.404 min Scan# 3114

Delta R.T. -0.006 min

Lab File: BP017243.D

Acq: 20 Sep 2023 17:26




Tgt Ion:252 Resp: 406741


Ion Ratio Lower Upper

252 100

254 64.9 51.9 77.9

126 10.0 9.1 13.7









#85
Benzo(a)anthracene
Concen: 20.996 ng/ul
RT: 21.462 min Scan# 3125
Delta R.T. -0.006 min
Lab File: BP017243.D
Acq: 20 Sep 2023 17:26




Instrument :
BNA_P
ClientSampleId :
SSTD020741


Tgt Ion:228 Resp: 1281364
Ion Ratio Lower Upper
228 100
229 19.3 15.8 23.6
226 27.2 21.8 32.8



#86
Bis(2-ethylhexyl)phthalate
Concen: 21.444 ng/ul
RT: 21.339 min Scan# 3103
Delta R.T. -0.006 min
Lab File: BP017243.D
Acq: 20 Sep 2023 17:26

Tgt Ion:149 Resp: 665675
Ion Ratio Lower Upper
149 100
167 26.6 21.0 31.6
279 4.2 3.0 4.4





#87

Chrysene

Concen: 21.010 ng/ul

RT: 21.515 min Scan# 3135

Delta R.T. -0.012 min

Lab File: BP017243.D

Acq: 20 Sep 2023 17:26

Instrument :

BNA_P

ClientSampleId :

SSTD020741





Tgt Ion:228 Resp: 1206721

Ion Ratio Lower Upper

228 100

226 30.1 24.1 36.1

229 19.8 15.0 22.6



#89

Di-n-octyl phthalate

Concen: 20.556 ng/ul




RT: 22.298 min Scan# 3266


Delta R.T. -0.018 min

Lab File: BP017243.D

Acq: 20 Sep 2023 17:26

Tgt Ion:149 Resp: 1161406









#90
Benzo(b)fluoranthene
Concen: 20.963 ng/ul
RT: 23.215 min Scan# 3422
Delta R.T. -0.012 min
Lab File: BP017243.D
Acq: 20 Sep 2023 17:26




Instrument :
BNA_P
ClientSampleId :
SSTD020741


Tgt Ion:252 Resp: 1269472
Ion Ratio Lower Upper
252 100
253 21.8 17.4 26.0
125 9.4 7.8 11.8



#91
Benzo(k)fluoranthene
Concen: 20.763 ng/ul
RT: 23.268 min Scan# 3431
Delta R.T. -0.012 min
Lab File: BP017243.D
Acq: 20 Sep 2023 17:26

Tgt Ion:252 Resp: 1268575
Ion Ratio Lower Upper
252 100
253 21.7 17.9 26.9
125 9.3 7.8 11.6





#93

Benzo(a)pyrene

Concen: 20.971 ng/ul

RT: 23.880 min Scan# 31

Delta R.T. -0.018 min

Lab File: BP017243.D

Acq: 20 Sep 2023 17:26

Instrument :

BNA_P

ClientSampleId :

SSTD020741





Tgt Ion:252 Resp: 1215089

Ion Ratio Lower Upper

252 100

253 22.3 17.4 26.0

125 10.4 8.9 13.3



#94

Indeno(1,2,3-cd)pyrene

Concen: 20.582 ng/ul

RT: 26.668 min Scan# 4009

Delta R.T. -0.024 min

Lab File: BP017243.D

Acq: 20 Sep 2023 17:26




Tgt Ion:276 Resp: 1494773


Ion Ratio Lower Upper

276 100

138 22.7 17.9 26.9

277 24.5 19.5 29.3





#95

Dibenzo(a,h)anthracene

Concen: 20.662 ng/ul

RT: 26.692 min Scan# 4013

Delta R.T. -0.024 min

Lab File: BP017243.D

Acq: 20 Sep 2023 17:26

Instrument :

BNA_P

ClientSampleId :

SSTD020741





Tgt Ion:278 Resp: 1249846

Ion Ratio Lower Upper

278 100

139 15.7 13.4 20.0

279 23.8 19.5 29.3



#96

Benzo(g,h,i)perylene

Concen: 20.441 ng/ul

RT: 27.503 min Scan# 4151

Delta R.T. -0.030 min

Lab File: BP017243.D

Acq: 20 Sep 2023 17:26

Tgt Ion:276 Resp: 1197196

Ion Ratio Lower Upper

276 100

138 21.8 17.3 25.9

277 23.9 18.8 28.2

