

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092122\
 Data File : BP011776.D
 Acq On : 21 Sep 2022 20:18
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP092122

Quant Time: Sep 22 03:44:02 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP092122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 22 02:45:14 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	149	0.00
2	1,4-Dioxane	40.000	40.016	-0.0	148	0.00
3	Pyridine	40.000	40.578	-1.4	141	0.00
4	n-Nitrosodimethylamine	40.000	37.184	7.0	133	0.00
5 S	2-Fluorophenol	80.000	81.775	-2.2	149	0.00
6	Aniline	40.000	39.515	1.2	140	0.00
7 S	Phenol-d6	80.000	79.892	0.1	143	0.00
8	2-Chlorophenol	40.000	40.467	-1.2	146	0.00
9	Benzaldehyde	40.000	39.135	2.2	145	0.00
10 C	Phenol	40.000	39.667	0.8	142	0.00
11	bis(2-Chloroethyl)ether	40.000	38.988	2.5	143	0.00
12	1,3-Dichlorobenzene	40.000	40.196	-0.5	148	0.00
13 C	1,4-Dichlorobenzene	40.000	40.227	-0.6	148	0.00
14	1,2-Dichlorobenzene	40.000	39.444	1.4	145	0.00
15	Benzyl Alcohol	40.000	40.918	-2.3	143	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.808	5.5	137	0.00
17	2-Methylphenol	40.000	40.452	-1.1	144	0.00
18	Hexachloroethane	40.000	39.741	0.6	148	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.046	-0.1	141	0.00
20	3+4-Methylphenols	40.000	40.745	-1.9	144	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	143	0.00
22	Acetophenone	40.000	40.231	-0.6	139	0.00
23 S	Nitrobenzene-d5	80.000	80.366	-0.5	139	0.00
24	Nitrobenzene	40.000	39.837	0.4	138	0.00
25	Isophorone	40.000	41.164	-2.9	141	0.00
26 C	2-Nitrophenol	40.000	41.653	-4.1	152	0.00
27	2,4-Dimethylphenol	40.000	42.823	-7.1	146	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.758	-1.9	142	0.00
29 C	2,4-Dichlorophenol	40.000	43.451	-8.6	147	0.00
30	1,2,4-Trichlorobenzene	40.000	41.925	-4.8	146	0.00
31	Naphthalene	40.000	40.519	-1.3	142	0.00
32	Benzoic acid	40.000	42.285	-5.7	151	0.01
33	4-Chloroaniline	40.000	41.362	-3.4	142	0.00
34 C	Hexachlorobutadiene	40.000	42.779	-6.9	149	0.00
35	Caprolactam	40.000	42.371	-5.9	142	0.01
36 C	4-Chloro-3-methylphenol	40.000	42.356	-5.9	142	0.00
37	2-Methylnaphthalene	40.000	41.196	-3.0	143	0.00
38	1-Methylnaphthalene	40.000	40.759	-1.9	141	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	141	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	42.999	-7.5	147	0.00
41 P	Hexachlorocyclopentadiene	40.000	45.518	-13.8	151	0.00
42 S	2,4,6-Tribromophenol	80.000	88.606	-10.8	146	0.00
43 C	2,4,6-Trichlorophenol	40.000	45.111	-12.8	147	0.00
44	2,4,5-Trichlorophenol	40.000	43.541	-8.9	145	0.00
45 S	2-Fluorobiphenyl	80.000	81.235	-1.5	139	0.00
46	1,1'-Biphenyl	40.000	40.840	-2.1	140	0.00
47	2-Chloronaphthalene	40.000	40.901	-2.3	140	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092122\
 Data File : BP011776.D
 Acq On : 21 Sep 2022 20:18
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP092122

Quant Time: Sep 22 03:44:02 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP092122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 22 02:45:14 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.848	-7.1	139	0.00
49	Acenaphthylene	40.000	40.968	-2.4	139	0.00
50	Dimethylphthalate	40.000	40.900	-2.2	139	0.00
51	2,6-Dinitrotoluene	40.000	42.725	-6.8	140	0.00
52 C	Acenaphthene	40.000	40.337	-0.8	138	0.00
53	3-Nitroaniline	40.000	42.683	-6.7	138	0.00
54 P	2,4-Dinitrophenol	40.000	41.434	-3.6	153	0.00
55	Dibenzofuran	40.000	39.974	0.1	137	0.00
56 P	4-Nitrophenol	40.000	42.146	-5.4	141	0.00
57	2,4-Dinitrotoluene	40.000	40.085	-0.2	142	0.00
58	Fluorene	40.000	39.867	0.3	135	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	43.335	-8.3	144	0.00
60	Diethylphthalate	40.000	40.921	-2.3	139	0.00
61	4-Chlorophenyl-phenylether	40.000	41.294	-3.2	139	0.00
62	4-Nitroaniline	40.000	39.562	1.1	138	0.00
63	Azobenzene	40.000	39.147	2.1	134	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	139	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.579	-6.4	153	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.866	-4.7	139	0.00
67	4-Bromophenyl-phenylether	40.000	43.403	-8.5	144	0.00
68	Hexachlorobenzene	40.000	42.523	-6.3	141	0.00
69	Atrazine	40.000	45.193	-13.0	145	0.00
70 C	Pentachlorophenol	40.000	42.260	-5.6	148	0.00
71	Phenanthrene	40.000	40.307	-0.8	136	0.00
72	Anthracene	40.000	41.361	-3.4	137	0.00
73	Carbazole	40.000	40.862	-2.2	135	0.00
74	Di-n-butylphthalate	40.000	43.027	-7.6	139	0.00
75 C	Fluoranthene	40.000	41.913	-4.8	137	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	131	0.00
77	Benzydine	40.000	41.746	-4.4	143	0.00
78	Pyrene	40.000	43.279	-8.2	136	0.00
79 S	Terphenyl-d14	80.000	85.068	-6.3	132	0.00
80	Butylbenzylphthalate	40.000	43.457	-8.6	144	0.00
81	Benzo(a)anthracene	40.000	42.197	-5.5	132	0.00
82	3,3'-Dichlorobenzidine	40.000	43.255	-8.1	135	0.00
83	Chrysene	40.000	41.741	-4.4	132	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	43.265	-8.2	137	0.00
85 c	Di-n-octyl phthalate	40.000	44.578	-11.4	137	0.00
86 I	Perylene-d12	20.000	20.000	0.0	124	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	37.086	7.3	108	0.00
88	Benzo(b)fluoranthene	40.000	43.117	-7.8	128	0.00
89	Benzo(k)fluoranthene	40.000	43.461	-8.7	126	0.00
90 C	Benzo(a)pyrene	40.000	42.736	-6.8	125	0.00
91	Dibenzo(a,h)anthracene	40.000	37.322	6.7	109	0.00
92	Benzo(g,h,i)perylene	40.000	35.544	11.1	105	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092122\
Data File : BP011776.D
Acq On : 21 Sep 2022 20:18
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP092122

Quant Time: Sep 22 03:44:02 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP092122.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 22 02:45:14 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0