

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092223\
 Data File : BP017291.D
 Acq On : 22 Sep 2023 17:43
 Operator : MA/JU
 Sample : PB155619BL
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK619

Quant Time: Sep 22 23:40:23 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP092023.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 21 23:51:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.934	152	141465	20.000	ng/ul	0.00
20) Naphthalene-d8	10.758	136	566564	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.593	164	332709	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.363	188	762276	20.000	ng/ul	0.00
79) Chrysene-d12	21.469	240	734841	20.000	ng/ul	0.00
88) Perylene-d12	23.992	264	820518	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.305	96	29364	8.526	ng/uL	0.00
4) Pyridine-d5	3.734	84	409226	42.502	ng/ul	0.00
7) Phenol-d5	7.093	99	487321	41.944	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.275	67	285104	40.661	ng/ul	0.00
11) 2-Chlorophenol-d4	7.458	132	395696	42.941	ng/ul	0.00
15) 4-Methylphenol-d8	8.646	113	379386	42.001	ng/ul	-0.01
21) Nitrobenzene-d5	9.134	128	180663	44.108	ng/ul	0.00
24) 2-Nitrophenol-d4	9.846	143	198093	44.541	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.375	165	358312	42.644	ng/ul	0.00
31) 4-Chloroaniline-d4	10.922	131	514848	42.299	ng/ul	0.00
46) Dimethylphthalate-d6	14.010	166	1059458	44.451	ng/ul	0.00
49) Acenaphthylene-d8	14.293	160	1180032	43.047	ng/ul	0.00
54) 4-Nitrophenol-d4	14.828	143	164749	39.676	ng/ul	0.00
60) Fluorene-d10	15.593	176	922687	44.967	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.722	200	115225	26.819	ng/ul	0.00
73) Anthracene-d10	17.469	188	1577002	46.647	ng/ul	0.00
81) Pyrene-d10	19.710	212	1936800	48.739	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.827	264	1919224	48.581	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092223\
Data File : BP017291.D
Acq On : 22 Sep 2023 17:43
Operator : MA/JU
Sample : PB155619BL
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK619

Quant Time: Sep 22 23:40:23 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP092023.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Sep 21 23:51:07 2023
Response via : Initial Calibration

