

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092424\
 Data File : BP022071.D
 Acq On : 26 Sep 2024 20:38
 Operator : RC/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020772

Quant Time: Sep 27 01:20:27 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP092424.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Sep 24 15:23:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.699	152	98929	20.000	ng/u1	-0.01
20) Naphthalene-d8	10.463	136	396438	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.316	164	329414	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.116	188	890082	20.000	ng/u1	0.00
79) Chrysene-d12	21.551	240	1102504	20.000	ng/u1	0.01
88) Perylene-d12	24.833	264	1338133	20.000	ng/u1	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.217	96	18937	7.498	ng/uL	0.00
4) Pyridine-d5	3.634	84	136657	18.929	ng/u1	0.00
7) Phenol-d5	6.881	99	151429	18.712	ng/u1	-0.01
9) Bis-(2-Chloroethyl)eth...	7.040	67	95201	19.105	ng/u1	0.00
11) 2-Chlorophenol-d4	7.246	132	116936	19.972	ng/u1	0.00
15) 4-Methylphenol-d8	8.399	113	130874	20.409	ng/u1	0.00
21) Nitrobenzene-d5	8.846	128	61391	20.690	ng/u1	0.00
24) 2-Nitrophenol-d4	9.552	143	70442	20.938	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.093	165	150550	20.606	ng/u1	0.00
31) 4-Chloroaniline-d4	10.599	131	172900	19.834	ng/u1	0.00
46) Dimethylphthalate-d6	13.722	166	508252	19.960	ng/u1	0.00
49) Acenaphthylene-d8	14.010	160	547830	20.320	ng/u1	0.00
54) 4-Nitrophenol-d4	14.528	143	83041	19.065	ng/u1	0.00
60) Fluorene-d10	15.328	176	471744	20.944	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.445	200	93335	16.890	ng/u1	0.00
73) Anthracene-d10	17.216	188	813780	19.584	ng/u1	0.00
81) Pyrene-d10	19.592	212	1102074	19.295	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.604	264	1405637	19.972	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.252	88	20688	7.562	ng/uL	96
5) Pyridine	3.652	79	134740	18.370	ng/u1	99
6) Benzaldehyde	6.852	77	109275	23.870	ng/u1	99
8) Phenol	6.910	94	156872	18.575	ng/u1	96
10) Bis(2-Chloroethyl)ether	7.134	93	119521	18.491	ng/u1	93
12) 2-Chlorophenol	7.275	128	119598	19.704	ng/u1	99
13) 2-Methylphenol	8.140	108	121295	19.803	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.216	45	113425	18.916	ng/u1	95
16) Acetophenone	8.510	105	217190	20.362	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.499	70	115200	20.833	ng/u1	94
18) 4-Methylphenol	8.463	108	135208	20.243	ng/u1	100
19) Hexachloroethane	8.769	117	59588	20.796	ng/u1	98
22) Nitrobenzene	8.881	77	185722	20.472	ng/u1	98
23) Isophorone	9.410	82	311706	19.837	ng/u1	97
25) 2-Nitrophenol	9.587	139	74863	20.797	ng/u1	89
26) 2,4-Dimethylphenol	9.652	107	165280	20.268	ng/u1	96
27) Bis(2-Chloroethoxy)met...	9.875	93	174802	19.298	ng/u1	99
29) 2,4-Dichlorophenol	10.122	162	149427	20.586	ng/u1	98
30) Naphthalene	10.510	128	410880	19.697	ng/u1	99
32) 4-Chloroaniline	10.622	127	171874	19.907	ng/u1	100
33) Hexachlorobutadiene	10.793	225	136412	21.166	ng/u1	96
34) Caprolactam	11.393	113	41210	18.715	ng/u1	95
35) 4-Chloro-3-methylphenol	11.746	107	162564	20.701	ng/u1	98
36) 2-Methylnaphthalene	12.116	142	313001	20.191	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092424\
 Data File : BP022071.D
 Acq On : 26 Sep 2024 20:38
 Operator : RC/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020772

Quant Time: Sep 27 01:20:27 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP092424.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Sep 24 15:23:13 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.340	142	316455	20.241	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.498	216	242920	19.890	ng/ul#	97
40) Hexachlorocyclopentadiene	12.475	237	148238	19.296	ng/ul	98
41) 2,4,6-Trichlorophenol	12.740	196	146805	20.107	ng/ul	98
42) 2,4,5-Trichlorophenol	12.810	196	161841	20.470	ng/ul	98
43) 1,1'-Biphenyl	13.140	154	452734	19.219	ng/ul	98
44) 2-Chloronaphthalene	13.181	162	369581	19.486	ng/ul	98
45) 2-Nitroaniline	13.387	65	114186	20.821	ng/ul	99
47) Dimethylphthalate	13.769	163	496808	19.425	ng/ul	99
48) 2,6-Dinitrotoluene	13.887	165	99066	20.205	ng/ul	94
50) Acenaphthylene	14.034	152	558368	19.860	ng/ul	99
51) 3-Nitroaniline	14.222	138	91275	20.447	ng/ul	99
52) Acenaphthene	14.375	153	401838	19.671	ng/ul	97
53) 2,4-Dinitrophenol	14.428	184	42260	12.667	ng/ul	95
55) 4-Nitrophenol	14.545	109	104484	21.203	ng/ul	89
56) Dibenzofuran	14.722	168	615329	20.260	ng/ul	99
57) 2,4-Dinitrotoluene	14.687	165	158702	21.004	ng/ul#	97
58) 2,3,4,6-Tetrachlorophenol	14.957	232	162707	21.379	ng/ul	98
59) Diethylphthalate	15.145	149	521572	20.606	ng/ul	99
61) Fluorene	15.381	166	519335	20.562	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.381	204	299425	20.591	ng/ul	99
63) 4-Nitroaniline	15.398	138	102427	22.545	ng/ul	88
66) 4,6-Dinitro-2-methylph...	15.457	198	102473	17.033	ng/ul	94
67) N-Nitrosodiphenylamine	15.598	169	451575	19.263	ng/ul	99
68) 4-Bromophenyl-phenylether	16.292	248	214202	20.040	ng/ul	98
69) Hexachlorobenzene	16.410	284	259934	20.028	ng/ul	99
70) Atrazine	16.575	200	207882	19.878	ng/ul	100
71) Pentachlorophenol	16.763	266	149453	17.259	ng/ul	99
72) Phenanthrene	17.157	178	903186	19.353	ng/ul	99
74) Anthracene	17.251	178	929190	19.565	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.104	216	249254	18.661	ng/ul	99
76) Pentachlorobenzene	14.639	250	281550	19.168	ng/ul	99
77) Carbazole	17.533	167	810863	19.372	ng/ul	99
78) Di-n-butylphthalate	18.128	149	984157	20.394	ng/ul	99
80) Fluoranthene	19.245	202	1255913	19.237	ng/ul	99
82) Pyrene	19.622	202	1363899	19.409	ng/ul#	95
83) Butylbenzylphthalate	20.580	149	493596	20.155	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.451	252	499336	19.781	ng/ul	97
85) Benzo(a)anthracene	21.533	228	1490181	19.613	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.469	149	749700	20.864	ng/ul	99
87) Chrysene	21.598	228	1396717	19.587	ng/ul	99
89) Di-n-octyl phthalate	22.721	149	1282938	19.408	ng/ul	100
90) Benzo(b)fluoranthene	23.792	252	1612884	19.416	ng/ul	99
91) Benzo(k)fluoranthene	23.863	252	1600059	19.327	ng/ul	99
93) Benzo(a)pyrene	24.680	252	1504844	19.660	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	28.545	276	935111	9.370	ng/ul	98
95) Dibenzo(a,h)anthracene	28.627	278	1534297	19.084	ng/ul	97
96) Benzo(g,h,i)perylene	29.698	276	1457285	18.818	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

```
Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092424\  
Data File : BP022071.D  
Acq On    : 26 Sep 2024   20:38  
Operator  : RC/JU  
Sample    : SSTDCCC020  
Misc      :  
ALS Vial  : 48    Sample Multiplier: 1
```

Instrument :
BNA_P
ClientSampleId :
SSTD020772

Quant Time: Sep 27 01:20:27 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP092424.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Sep 24 15:23:13 2024
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.562 ng/uL
RT: 3.252 min Scan# 41
Delta R.T. -0.006 min
Lab File: BP022071.D
Acq: 26 Sep 2024 20:38




Instrument :
BNA_P
ClientSampleId :
SSTD020772


Tgt Ion: 88 Resp: 20688
Ion Ratio Lower Upper
88 100
43 27.5 22.1 33.1
58 56.7 49.3 73.9



#5
Pyridine
Concen: 18.370 ng/uL
RT: 3.652 min Scan# 113
Delta R.T. 0.000 min
Lab File: BP022071.D
Acq: 26 Sep 2024 20:38

Tgt Ion: 79 Resp: 134740
Ion Ratio Lower Upper
79 100
52 60.3 47.3 70.9
51 32.1 25.3 37.9









#6
Benzaldehyde
Concen: 23.870 ng/ul
RT: 6.852 min Scan# 61
Delta R.T. -0.012 min
Lab File: BP022071.D
Acq: 26 Sep 2024 20:38




Instrument :
BNA_P
ClientSampleId :
SSTD020772


Tgt Ion: 77 Resp: 109275
Ion Ratio Lower Upper
77 100
105 86.3 70.6 106.0
106 83.9 67.4 101.0



#8
Phenol
Concen: 18.575 ng/ul
RT: 6.910 min Scan# 667
Delta R.T. -0.006 min
Lab File: BP022071.D
Acq: 26 Sep 2024 20:38

Tgt Ion: 94 Resp: 156872
Ion Ratio Lower Upper
94 100
65 34.4 26.6 40.0
66 52.9 39.6 59.4









#10
 Bis(2-Chloroethyl)ether
 Concen: 18.491 ng/ul
 RT: 7.134 min Scan# 705
 Delta R.T. -0.006 min
 Lab File: BP022071.D
 Acq: 26 Sep 2024 20:38

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020772

Tgt Ion: 93 Resp: 119521




Ion	Ratio	Lower	Upper
93	100		
63	78.8	57.7	86.5
95	34.1	24.9	37.3




#12
 2-Chlorophenol
 Concen: 19.704 ng/ul
 RT: 7.275 min Scan# 729
 Delta R.T. -0.006 min
 Lab File: BP022071.D
 Acq: 26 Sep 2024 20:38

Tgt Ion: 128 Resp: 119598

Ion	Ratio	Lower	Upper
128	100		
64	50.8	41.3	61.9
130	32.1	25.8	38.8









#13
2-Methylphenol
Concen: 19.803 ng/ul
RT: 8.140 min Scan# 81
Delta R.T. -0.006 min
Lab File: BP022071.D
Acq: 26 Sep 2024 20:38




Instrument :
BNA_P
ClientSampleId :
SSTD020772


Tgt Ion:108 Resp: 121295
Ion Ratio Lower Upper
108 100
107 88.5 71.3 106.9



#14
2,2'-oxybis(1-Chloropropane)
Concen: 18.916 ng/ul
RT: 8.216 min Scan# 889
Delta R.T. -0.006 min
Lab File: BP022071.D
Acq: 26 Sep 2024 20:38

Tgt Ion: 45 Resp: 113425
Ion Ratio Lower Upper
45 100
77 25.6 18.6 27.8
79 18.3 13.2 19.8









#16
Acetophenone
Concen: 20.362 ng/ul
RT: 8.510 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP022071.D
Acq: 26 Sep 2024 20:38




Instrument :
BNA_P
ClientSampleId :
SSTD020772


Tgt Ion:105 Resp: 217190
Ion Ratio Lower Upper
105 100
77 84.7 67.9 101.9
51 26.7 19.8 29.6



#17
N-Nitroso-di-n-propylamine
Concen: 20.833 ng/ul
RT: 8.499 min Scan# 937
Delta R.T. -0.006 min
Lab File: BP022071.D
Acq: 26 Sep 2024 20:38

Tgt Ion: 70 Resp: 115200
Ion Ratio Lower Upper
70 100
42 49.4 35.5 53.3
101 10.5 7.3 10.9
130 19.2 13.9 20.9









#18
4-Methylphenol
Concen: 20.243 ng/ul
RT: 8.463 min Scan# 911
Delta R.T. -0.006 min
Lab File: BP022071.D
Acq: 26 Sep 2024 20:38




Instrument :
BNA_P
ClientSampleId :
SSTD020772


Tgt Ion:108 Resp: 135208
Ion Ratio Lower Upper
108 100
107 114.7 92.0 138.0



#19
Hexachloroethane
Concen: 20.796 ng/ul
RT: 8.769 min Scan# 983
Delta R.T. -0.006 min
Lab File: BP022071.D
Acq: 26 Sep 2024 20:38

Tgt Ion:117 Resp: 59588
Ion Ratio Lower Upper
117 100
201 135.7 110.2 165.4
199 82.5 67.9 101.9









#22
Nitrobenzene
Concen: 20.472 ng/ul
RT: 8.881 min Scan# 1003
Delta R.T. -0.006 min
Lab File: BP022071.D
Acq: 26 Sep 2024 20:38




Instrument :
BNA_P
ClientSampleId :
SSTD020772


Tgt Ion: 77 Resp: 185722
Ion Ratio Lower Upper
77 100
123 35.6 29.8 44.8
65 14.2 11.3 16.9



#23
Isophorone
Concen: 19.837 ng/ul
RT: 9.410 min Scan# 1092
Delta R.T. 0.000 min
Lab File: BP022071.D
Acq: 26 Sep 2024 20:38

Tgt Ion: 82 Resp: 311706
Ion Ratio Lower Upper
82 100
95 8.8 7.0 10.4
138 16.6 11.9 17.9









#25
2-Nitrophenol
Concen: 20.797 ng/ul
RT: 9.587 min Scan# 1122
Delta R.T. -0.012 min
Lab File: BP022071.D
Acq: 26 Sep 2024 20:38




Instrument :
BNA_P
ClientSampleId :
SSTD020772


Tgt Ion:139 Resp: 74863
Ion Ratio Lower Upper
139 100
65 66.8 47.8 71.6
109 52.1 35.3 52.9



#26
2,4-Dimethylphenol
Concen: 20.268 ng/ul
RT: 9.652 min Scan# 1133
Delta R.T. -0.006 min
Lab File: BP022071.D
Acq: 26 Sep 2024 20:38

Tgt Ion:107 Resp: 165280
Ion Ratio Lower Upper
107 100
121 41.7 37.0 55.4
122 69.5 57.6 86.4









#27
 Bis(2-Chloroethoxy)methane
 Concen: 19.298 ng/ul
 RT: 9.875 min Scan# 1171
 Delta R.T. -0.012 min
 Lab File: BP022071.D
 Acq: 26 Sep 2024 20:38




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020772


Tgt Ion: 93 Resp: 174802
 Ion Ratio Lower Upper
 93 100
 95 32.5 25.3 37.9
 123 9.9 8.1 12.1



#29
 2,4-Dichlorophenol
 Concen: 20.586 ng/ul
 RT: 10.122 min Scan# 1213
 Delta R.T. 0.000 min
 Lab File: BP022071.D
 Acq: 26 Sep 2024 20:38

Tgt Ion:162 Resp: 149427
 Ion Ratio Lower Upper
 162 100
 164 66.8 52.4 78.6
 98 35.4 29.8 44.8









#30
Naphthalene
Concen: 19.697 ng/ul
RT: 10.510 min Scan# 1280
Delta R.T. -0.012 min
Lab File: BP022071.D
Acq: 26 Sep 2024 20:38




Instrument :
BNA_P
ClientSampleId :
SSTD020772


Tgt Ion:128 Resp: 410880
Ion Ratio Lower Upper
128 100
129 10.8 8.9 13.3
127 13.3 10.9 16.3



#32
4-Chloroaniline
Concen: 19.907 ng/ul
RT: 10.622 min Scan# 1298
Delta R.T. -0.006 min
Lab File: BP022071.D
Acq: 26 Sep 2024 20:38

Tgt Ion:127 Resp: 171874
Ion Ratio Lower Upper
127 100
129 31.4 25.1 37.7





#33

Hexachlorobutadiene

Concen: 21.166 ng/ul

RT: 10.793 min Scan# 1

Delta R.T. -0.012 min

Lab File: BP022071.D

Acq: 26 Sep 2024 20:38

Instrument :

BNA_P

ClientSampleId :

SSTD020772





Tgt Ion:225 Resp: 136412

Ion Ratio Lower Upper

225 100

223 64.7 47.1 70.7

227 64.3 51.6 77.4



#34

Caprolactam

Concen: 18.715 ng/ul

RT: 11.393 min Scan# 1429

Delta R.T. 0.000 min

Lab File: BP022071.D

Acq: 26 Sep 2024 20:38




Tgt Ion:113 Resp: 41210


Ion Ratio Lower Upper

113 100

55 139.2 108.9 163.3

56 118.4 101.8 152.8









#35
4-Chloro-3-methylphenol
Concen: 20.701 ng/ul
RT: 11.746 min Scan# 1489
Delta R.T. -0.012 min
Lab File: BP022071.D
Acq: 26 Sep 2024 20:38




Instrument :
BNA_P
ClientSampleId :
SSTD020772


Tgt Ion:107 Resp: 162564
Ion Ratio Lower Upper
107 100
144 24.1 19.9 29.9
142 73.2 57.4 86.0



#36
2-Methylnaphthalene
Concen: 20.191 ng/ul
RT: 12.116 min Scan# 1552
Delta R.T. -0.012 min
Lab File: BP022071.D
Acq: 26 Sep 2024 20:38

Tgt Ion:142 Resp: 313001
Ion Ratio Lower Upper
142 100
141 91.7 72.6 109.0





#37

1-Methylnaphthalene

Concen: 20.241 ng/ul

RT: 12.340 min Scan# 1590

Delta R.T. -0.006 min

Lab File: BP022071.D

Acq: 26 Sep 2024 20:38

Instrument :

BNA_P

ClientSampleId :

SSTD020772





Tgt Ion:142 Resp: 316455

Ion Ratio Lower Upper

142 100

141 93.4 74.6 111.8

116 4.9 3.9 5.9



#39

1,2,4,5-Tetrachlorobenzene

Concen: 19.890 ng/ul

RT: 12.498 min Scan# 1617

Delta R.T. 0.000 min

Lab File: BP022071.D

Acq: 26 Sep 2024 20:38

Tgt Ion:216 Resp: 242920




Ion Ratio Lower Upper


216 100

214 77.0 62.8 94.2

179 20.8 17.9 26.9

108 13.8 14.1 21.1#









#40
Hexachlorocyclopentadiene
Concen: 19.296 ng/ul
RT: 12.475 min Scan# 1613
Delta R.T. -0.012 min
Lab File: BP022071.D
Acq: 26 Sep 2024 20:38




Instrument :
BNA_P
ClientSampleId :
SSTD020772


Tgt Ion:237 Resp: 148238
Ion Ratio Lower Upper
237 100
235 63.8 51.8 77.6
272 11.4 10.5 15.7



#41
2,4,6-Trichlorophenol
Concen: 20.107 ng/ul
RT: 12.740 min Scan# 1658
Delta R.T. -0.006 min
Lab File: BP022071.D
Acq: 26 Sep 2024 20:38

Tgt Ion:196 Resp: 146805
Ion Ratio Lower Upper
196 100
198 97.3 77.6 116.4
200 33.2 24.2 36.4





#42

2,4,5-Trichlorophenol

Concen: 20.470 ng/ul

RT: 12.810 min Scan# 1669

Delta R.T. -0.006 min

Lab File: BP022071.D

Acq: 26 Sep 2024 20:38

Instrument :

BNA_P

ClientSampleId :

SSTD020772





Tgt Ion:196 Resp: 161841

Ion Ratio Lower Upper

196 100

198 96.4 78.2 117.2

200 31.3 26.1 39.1



#43

1,1'-Biphenyl

Concen: 19.219 ng/ul

RT: 13.140 min Scan# 1726

Delta R.T. -0.012 min

Lab File: BP022071.D

Acq: 26 Sep 2024 20:38




Tgt Ion:154 Resp: 452734


Ion Ratio Lower Upper

154 100

153 40.4 32.6 49.0

76 10.8 10.1 15.1





#44

2-Chloronaphthalene

Concen: 19.486 ng/ul

RT: 13.181 min Scan# 1733

Delta R.T. -0.006 min

Lab File: BP022071.D

Acq: 26 Sep 2024 20:38

Instrument :

BNA_P

ClientSampleId :

SSTD020772





Tgt Ion:162 Resp: 369581

Ion Ratio Lower Upper

162 100

164 34.0 26.2 39.2

127 39.6 30.8 46.2



#45

2-Nitroaniline

Concen: 20.821 ng/ul

RT: 13.387 min Scan# 1768

Delta R.T. -0.006 min

Lab File: BP022071.D

Acq: 26 Sep 2024 20:38




Tgt Ion: 65 Resp: 114186


Ion Ratio Lower Upper

65 100

92 65.9 53.2 79.8

138 88.4 69.8 104.8









#47
Dimethylphthalate
Concen: 19.425 ng/ul
RT: 13.769 min Scan# 1833
Delta R.T. -0.006 min
Lab File: BP022071.D
Acq: 26 Sep 2024 20:38




Instrument :
BNA_P
ClientSampleId :
SSTD020772


Tgt Ion:163 Resp: 496808
Ion Ratio Lower Upper
163 100
194 4.5 3.5 5.3
164 10.3 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 20.205 ng/ul
RT: 13.887 min Scan# 1853
Delta R.T. -0.006 min
Lab File: BP022071.D
Acq: 26 Sep 2024 20:38

Tgt Ion:165 Resp: 99066
Ion Ratio Lower Upper
165 100
89 69.0 50.9 76.3
121 21.5 16.1 24.1





#50

Acenaphthylene

Concen: 19.860 ng/ul

RT: 14.034 min Scan# 11

Delta R.T. 0.000 min

Lab File: BP022071.D

Acq: 26 Sep 2024 20:38

Instrument :

BNA_P

ClientSampleId :

SSTD020772





Tgt Ion:152 Resp: 558368

Ion Ratio Lower Upper

152 100

151 19.8 15.6 23.4

153 13.5 10.6 15.8



#51

3-Nitroaniline

Concen: 20.447 ng/ul

RT: 14.222 min Scan# 1910

Delta R.T. -0.006 min

Lab File: BP022071.D

Acq: 26 Sep 2024 20:38




Tgt Ion:138 Resp: 91275


Ion Ratio Lower Upper

138 100

108 13.8 10.6 16.0

92 132.4 107.2 160.8









#52
Acenaphthene
Concen: 19.671 ng/ul
RT: 14.375 min Scan# 1936
Delta R.T. -0.012 min
Lab File: BP022071.D
Acq: 26 Sep 2024 20:38




Instrument :
BNA_P
ClientSampleId :
SSTD020772


Tgt Ion:153 Resp: 401838
Ion Ratio Lower Upper
153 100
152 47.7 37.1 55.7
154 88.0 67.4 101.0



#53
2,4-Dinitrophenol
Concen: 12.667 ng/ul
RT: 14.428 min Scan# 1945
Delta R.T. 0.000 min
Lab File: BP022071.D
Acq: 26 Sep 2024 20:38

Tgt Ion:184 Resp: 42260
Ion Ratio Lower Upper
184 100
63 60.9 44.7 67.1
154 72.3 55.0 82.6









#55
4-Nitrophenol
Concen: 21.203 ng/ul
RT: 14.545 min Scan# 1965
Delta R.T. 0.006 min
Lab File: BP022071.D
Acq: 26 Sep 2024 20:38




Instrument :
BNA_P
ClientSampleId :
SSTD020772


Tgt Ion:109 Resp: 104484
Ion Ratio Lower Upper
109 100
139 84.6 76.8 115.2
65 103.2 90.8 136.2



#56
Dibenzofuran
Concen: 20.260 ng/ul
RT: 14.722 min Scan# 1995
Delta R.T. 0.000 min
Lab File: BP022071.D
Acq: 26 Sep 2024 20:38

Tgt Ion:168 Resp: 615329
Ion Ratio Lower Upper
168 100
139 39.2 31.0 46.6









#57
2,4-Dinitrotoluene
Concen: 21.004 ng/ul
RT: 14.687 min Scan# 1989
Delta R.T. 0.006 min
Lab File: BP022071.D
Acq: 26 Sep 2024 20:38




Instrument :
BNA_P
ClientSampleId :
SSTD020772


Tgt Ion:165 Resp: 158702
Ion Ratio Lower Upper
165 100
63 46.1 38.6 57.8
182 2.6 2.6 4.0#



#58
2,3,4,6-Tetrachlorophenol
Concen: 21.379 ng/ul
RT: 14.957 min Scan# 2035
Delta R.T. -0.006 min
Lab File: BP022071.D
Acq: 26 Sep 2024 20:38

Tgt Ion:232 Resp: 162707
Ion Ratio Lower Upper
232 100
131 36.9 28.2 42.2
130 2.1 1.8 2.6
166 26.9 21.8 32.6









#59
Diethylphthalate
Concen: 20.606 ng/ul
RT: 15.145 min Scan# 2067
Delta R.T. -0.023 min
Lab File: BP022071.D
Acq: 26 Sep 2024 20:38




Instrument :
BNA_P
ClientSampleId :
SSTD020772


Tgt Ion:149 Resp: 521572
Ion Ratio Lower Upper
149 100
177 23.8 18.6 27.8
150 12.8 9.8 14.6



#61
Fluorene
Concen: 20.562 ng/ul
RT: 15.381 min Scan# 2107
Delta R.T. 0.000 min
Lab File: BP022071.D
Acq: 26 Sep 2024 20:38

Tgt Ion:166 Resp: 519335
Ion Ratio Lower Upper
166 100
165 97.4 77.9 116.9
167 13.5 10.5 15.7









#62
4-Chlorophenyl-phenylether
Concen: 20.591 ng/ul
RT: 15.381 min Scan# 2107
Delta R.T. 0.000 min
Lab File: BP022071.D
Acq: 26 Sep 2024 20:38




Instrument :
BNA_P
ClientSampleId :
SSTD020772


Tgt Ion:204 Resp: 299425
Ion Ratio Lower Upper
204 100
206 33.4 26.7 40.1
141 51.7 42.6 64.0



#63
4-Nitroaniline
Concen: 22.545 ng/ul
RT: 15.398 min Scan# 2110
Delta R.T. 0.000 min
Lab File: BP022071.D
Acq: 26 Sep 2024 20:38

Tgt Ion:138 Resp: 102427
Ion Ratio Lower Upper
138 100
92 64.1 48.4 72.6
108 136.7 95.7 143.5









#66
4,6-Dinitro-2-methylphenol
Concen: 17.033 ng/ul
RT: 15.457 min Scan# 2120
Delta R.T. -0.006 min
Lab File: BP022071.D
Acq: 26 Sep 2024 20:38




Instrument :
BNA_P
ClientSampleId :
SSTD020772


Tgt Ion:198 Resp: 102473
Ion Ratio Lower Upper
198 100
182 3.9 3.4 5.2
77 28.2 19.8 29.8



#67
N-Nitrosodiphenylamine
Concen: 19.263 ng/ul
RT: 15.598 min Scan# 2144
Delta R.T. -0.006 min
Lab File: BP022071.D
Acq: 26 Sep 2024 20:38

Tgt Ion:169 Resp: 451575
Ion Ratio Lower Upper
169 100
168 67.0 54.1 81.1
167 37.2 28.7 43.1









#68
4-Bromophenyl-phenylether
Concen: 20.040 ng/ul
RT: 16.292 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP022071.D
Acq: 26 Sep 2024 20:38




Instrument :
BNA_P
ClientSampleId :
SSTD020772


Tgt Ion:248 Resp: 214202
Ion Ratio Lower Upper
248 100
250 99.0 78.4 117.6
141 67.4 51.4 77.0



#69
Hexachlorobenzene
Concen: 20.028 ng/ul
RT: 16.410 min Scan# 2282
Delta R.T. 0.006 min
Lab File: BP022071.D
Acq: 26 Sep 2024 20:38

Tgt Ion:284 Resp: 259934
Ion Ratio Lower Upper
284 100
142 29.3 23.4 35.0
249 32.6 25.4 38.0





#70

Atrazine

Concen: 19.878 ng/ul

RT: 16.575 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP022071.D

Acq: 26 Sep 2024 20:38

Instrument :

BNA_P

ClientSampleId :

SSTD020772





Tgt Ion:200 Resp: 207882

Ion Ratio Lower Upper

200 100

173 23.8 19.0 28.6

215 47.2 38.2 57.2



#71

Pentachlorophenol

Concen: 17.259 ng/ul

RT: 16.763 min Scan# 2342

Delta R.T. 0.006 min

Lab File: BP022071.D

Acq: 26 Sep 2024 20:38




Tgt Ion:266 Resp: 149453


Ion Ratio Lower Upper

266 100

264 63.3 50.2 75.4

268 63.4 51.4 77.0





#72

Phenanthrene

Concen: 19.353 ng/ul

RT: 17.157 min Scan# 2410

Delta R.T. 0.000 min

Lab File: BP022071.D

Acq: 26 Sep 2024 20:38

Instrument :

BNA_P

ClientSampleId :

SSTD020772





Tgt Ion:178 Resp: 903186

Ion Ratio Lower Upper

178 100

179 15.6 12.4 18.6

176 19.2 15.0 22.6



#74

Anthracene

Concen: 19.565 ng/ul

RT: 17.251 min Scan# 2425

Delta R.T. -0.006 min

Lab File: BP022071.D

Acq: 26 Sep 2024 20:38




Tgt Ion:178 Resp: 929190


Ion Ratio Lower Upper

178 100

179 14.8 12.3 18.5

176 18.0 14.9 22.3





#75

1,2,3,4-Tetrachlorobenzene

Concen: 18.661 ng/uL

RT: 13.104 min Scan# 1

Delta R.T. -0.012 min

Lab File: BP022071.D

Acq: 26 Sep 2024 20:38

Instrument :

BNA_P

ClientSampleId :

SSTD020772





Tgt Ion:216 Resp: 249254

Ion Ratio Lower Upper

216 100

214 75.7 60.6 90.8

218 45.4 38.1 57.1



#76

Pentachlorobenzene

Concen: 19.168 ng/uL

RT: 14.639 min Scan# 1981

Delta R.T. -0.006 min

Lab File: BP022071.D

Acq: 26 Sep 2024 20:38




Tgt Ion:250 Resp: 281550


Ion Ratio Lower Upper

250 100

248 63.2 50.3 75.5

252 64.4 50.1 75.1





#77

Carbazole

Concen: 19.372 ng/ul

RT: 17.533 min Scan# 2474

Delta R.T. 0.012 min

Lab File: BP022071.D

Acq: 26 Sep 2024 20:38

Instrument :

BNA_P

ClientSampleId :

SSTD020772





Tgt Ion:167 Resp: 810863

Ion Ratio Lower Upper

167 100

166 21.8 17.0 25.6

139 13.0 9.9 14.9



#78

Di-n-butylphthalate

Concen: 20.394 ng/ul

RT: 18.128 min Scan# 2574

Delta R.T. 0.006 min

Lab File: BP022071.D

Acq: 26 Sep 2024 20:38




Tgt Ion:149 Resp: 984157


Ion Ratio Lower Upper

149 100

150 9.3 7.2 10.8

104 6.3 5.0 7.6





#80

Fluoranthene

Concen: 19.237 ng/ul

RT: 19.245 min Scan# 21

Delta R.T. 0.006 min

Lab File: BP022071.D

Acq: 26 Sep 2024 20:38

Instrument :

BNA_P

ClientSampleId :

SSTD020772





Tgt Ion:202 Resp: 1255913

Ion Ratio Lower Upper

202 100

101 6.7 5.8 8.6

100 5.4 4.6 7.0



#82

Pyrene

Concen: 19.409 ng/ul

RT: 19.622 min Scan# 2828

Delta R.T. 0.006 min

Lab File: BP022071.D

Acq: 26 Sep 2024 20:38




Tgt Ion:202 Resp: 1363899


Ion Ratio Lower Upper

202 100

101 7.2 7.4 11.0#

100 6.1 6.2 9.2#





#83

Butylbenzylphthalate

Concen: 20.155 ng/ul

RT: 20.580 min Scan# 2989

Delta R.T. 0.012 min

Lab File: BP022071.D

Acq: 26 Sep 2024 20:38

Instrument :

BNA_P

ClientSampleId :

SSTD020772





Tgt Ion:149 Resp: 493596

Ion Ratio Lower Upper

149 100

91 75.0 60.2 90.2

206 23.9 19.8 29.6



#84

3,3'-Dichlorobenzidine

Concen: 19.781 ng/ul

RT: 21.451 min Scan# 3139

Delta R.T. 0.006 min

Lab File: BP022071.D

Acq: 26 Sep 2024 20:38




Tgt Ion:252 Resp: 499336


Ion Ratio Lower Upper

252 100

254 66.2 51.2 76.8

126 7.6 5.8 8.8





#85

Benzo(a)anthracene

Concen: 19.613 ng/ul

RT: 21.533 min Scan# 3153

Delta R.T. 0.006 min

Lab File: BP022071.D

Acq: 26 Sep 2024 20:38

Instrument :

BNA_P

ClientSampleId :

SSTD020772





Tgt Ion:228 Resp: 1490181

Ion Ratio Lower Upper

228 100

229 19.7 15.7 23.5

226 26.9 21.4 32.2



#86

Bis(2-ethylhexyl)phthalate

Concen: 20.864 ng/ul

RT: 21.469 min Scan# 3142

Delta R.T. 0.006 min

Lab File: BP022071.D

Acq: 26 Sep 2024 20:38




Tgt Ion:149 Resp: 749700


Ion Ratio Lower Upper

149 100

167 29.2 23.0 34.4

279 6.0 4.5 6.7





#87

Chrysene

Concen: 19.587 ng/ul

RT: 21.598 min Scan# 3164

Delta R.T. 0.012 min

Lab File: BP022071.D

Acq: 26 Sep 2024 20:38

Instrument :

BNA_P

ClientSampleId :

SSTD020772





Tgt Ion:228 Resp: 1396717

Ion Ratio Lower Upper

228 100

226 30.4 23.6 35.4

229 19.8 15.7 23.5



#89

Di-n-octyl phthalate

Concen: 19.408 ng/ul




RT: 22.721 min Scan# 3355


Delta R.T. 0.006 min

Lab File: BP022071.D

Acq: 26 Sep 2024 20:38

Tgt Ion:149 Resp: 1282938





#90

Benzo(b)fluoranthene

Concen: 19.416 ng/ul

RT: 23.792 min Scan# 3537

Delta R.T. 0.006 min

Lab File: BP022071.D

Acq: 26 Sep 2024 20:38

Instrument :

BNA_P

ClientSampleId :

SSTD020772





Tgt Ion:252 Resp: 1612884

Ion Ratio Lower Upper

252 100

253 22.1 17.8 26.6

125 5.4 4.8 7.2



#91

Benzo(k)fluoranthene

Concen: 19.327 ng/ul

RT: 23.863 min Scan# 3549

Delta R.T. 0.012 min

Lab File: BP022071.D

Acq: 26 Sep 2024 20:38




Tgt Ion:252 Resp: 1600059


Ion Ratio Lower Upper

252 100

253 21.8 17.4 26.2

125 5.1 5.0 7.6





#93

Benzo(a)pyrene

Concen: 19.660 ng/ul

RT: 24.680 min Scan# 3

Delta R.T. 0.006 min

Lab File: BP022071.D

Acq: 26 Sep 2024 20:38

Instrument :

BNA_P

ClientSampleId :

SSTD020772





Tgt Ion:252 Resp: 1504844

Ion Ratio Lower Upper

252 100

253 21.7 17.5 26.3

125 5.7 4.8 7.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 9.370 ng/ul

RT: 28.545 min Scan# 4345

Delta R.T. -0.006 min

Lab File: BP022071.D

Acq: 26 Sep 2024 20:38




Tgt Ion:276 Resp: 935111


Ion Ratio Lower Upper

276 100

138 9.9 9.2 13.8

277 23.5 19.4 29.0





#95

Dibenzo(a,h)anthracene

Concen: 19.084 ng/ul

RT: 28.627 min Scan# 41

Delta R.T. 0.006 min

Lab File: BP022071.D

Acq: 26 Sep 2024 20:38

Instrument :

BNA_P

ClientSampleId :

SSTD020772





Tgt Ion:278 Resp: 1534297

Ion Ratio Lower Upper

278 100

139 8.3 8.0 12.0

279 24.0 18.6 27.8



#96

Benzo(g,h,i)perylene

Concen: 18.818 ng/ul

RT: 29.698 min Scan# 4541

Delta R.T. -0.006 min

Lab File: BP022071.D

Acq: 26 Sep 2024 20:38

Tgt Ion:276 Resp: 1457285

Ion Ratio Lower Upper

276 100

138 11.8 9.0 13.6

277 23.6 19.1 28.7

