

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092424\
 Data File : BP022028.D
 Acq On : 25 Sep 2024 14:43
 Operator : RC/JU
 Sample : PB163543BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK543

Quant Time: Sep 26 02:43:06 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP092424.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Sep 24 15:23:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.710	152	114245	20.000	ng/u1	0.00
20) Naphthalene-d8	10.469	136	437754	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.310	164	347310	20.000	ng/u1	-0.01
64) Phenanthrene-d10	17.116	188	835644	20.000	ng/u1	0.00
79) Chrysene-d12	21.545	240	921173	20.000	ng/u1	0.00
88) Perylene-d12	24.833	264	989840	20.000	ng/u1	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.222	96	16083	5.514	ng/uL	0.00
4) Pyridine-d5	3.634	84	221109	26.521	ng/u1	0.00
7) Phenol-d5	6.893	99	282133	30.190	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.046	67	170067	29.554	ng/u1	0.00
11) 2-Chlorophenol-d4	7.246	132	207450	30.681	ng/u1	0.00
15) 4-Methylphenol-d8	8.410	113	226353	30.566	ng/u1	0.00
21) Nitrobenzene-d5	8.846	128	104943	32.030	ng/u1	0.00
24) 2-Nitrophenol-d4	9.563	143	122190	32.891	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.099	165	251469	31.170	ng/u1	0.00
31) 4-Chloroaniline-d4	10.599	131	303956	31.576	ng/u1	0.00
46) Dimethylphthalate-d6	13.716	166	899118	33.490	ng/u1	0.00
49) Acenaphthylene-d8	13.998	160	902841	31.762	ng/u1	0.00
54) 4-Nitrophenol-d4	14.522	143	128327	27.944	ng/u1	0.00
60) Fluorene-d10	15.322	176	756782	31.867	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.445	200	124426	23.983	ng/u1	0.00
73) Anthracene-d10	17.216	188	1302963	33.400	ng/u1	0.00
81) Pyrene-d10	19.592	212	1676112	35.121	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.610	264	1719946	33.036	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092424\
Data File : BP022028.D
Acq On : 25 Sep 2024 14:43
Operator : RC/JU
Sample : PB163543BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK543

Quant Time: Sep 26 02:43:06 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP092424.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Sep 24 15:23:13 2024
Response via : Initial Calibration

