

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092424\
 Data File : BP022034.D
 Acq On : 25 Sep 2024 18:48
 Operator : RC/JU
 Sample : P4013-02
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DDCG9

Quant Time: Sep 26 02:08:25 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP092424.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Sep 24 15:23:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.704	152	79042	20.000	ng/u1	0.00
20) Naphthalene-d8	10.469	136	285565	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.322	164	219510	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.110	188	559580	20.000	ng/u1	0.00
79) Chrysene-d12	21.545	240	669739	20.000	ng/u1	# 0.00
88) Perylene-d12	24.827	264	774867	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.222	96	10243	5.076	ng/uL	0.00
4) Pyridine-d5	3.640	84	37704	6.537	ng/u1	0.00
7) Phenol-d5	6.887	99	52562	8.129	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.045	67	116000	29.136	ng/u1	0.00
11) 2-Chlorophenol-d4	7.245	132	128312	27.429	ng/u1	0.00
15) 4-Methylphenol-d8	8.404	113	94750	18.493	ng/u1	0.00
21) Nitrobenzene-d5	8.845	128	65726	30.751	ng/u1	0.00
24) 2-Nitrophenol-d4	9.557	143	77342	31.914	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.098	165	161269	30.643	ng/u1	0.00
31) 4-Chloroaniline-d4	10.598	131	22245	3.542	ng/u1	0.00
46) Dimethylphthalate-d6	13.722	166	557332	32.845	ng/u1	0.00
49) Acenaphthylene-d8	14.010	160	564809	31.439	ng/u1	0.00
54) 4-Nitrophenol-d4	14.545	143	20289	6.990	ng/u1	0.02
60) Fluorene-d10	15.327	176	483304	32.200	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.439	200	82972	23.883	ng/u1	0.00
73) Anthracene-d10	17.221	188	865274	33.123	ng/u1	0.00
81) Pyrene-d10	19.592	212	1158907	33.400	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.609	264	1362277	33.426	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092424\
Data File : BP022034.D
Acq On : 25 Sep 2024 18:48
Operator : RC/JU
Sample : P4013-02
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DDCG9

Quant Time: Sep 26 02:08:25 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP092424.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Sep 24 15:23:13 2024
Response via : Initial Calibration

