

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092424\
 Data File : BP022036.D
 Acq On : 25 Sep 2024 20:10
 Operator : RC/JU
 Sample : P4013-04
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DDCH1

Quant Time: Sep 26 02:08:52 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP092424.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Sep 24 15:23:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.705	152	69696	20.000	ng/u1	0.00
20) Naphthalene-d8	10.469	136	260111	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.322	164	196956	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.110	188	490657	20.000	ng/u1	0.00
79) Chrysene-d12	21.539	240	582722	20.000	ng/u1	0.00
88) Perylene-d12	24.816	264	725455	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.223	96	11641	6.542	ng/uL	0.00
4) Pyridine-d5	3.640	84	35080	6.897	ng/u1	0.00
7) Phenol-d5	6.887	99	63184	11.083	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.046	67	138641	39.492	ng/u1	0.00
11) 2-Chlorophenol-d4	7.246	132	155326	37.656	ng/u1	0.00
15) 4-Methylphenol-d8	8.405	113	110946	24.558	ng/u1	0.00
21) Nitrobenzene-d5	8.846	128	81276	41.748	ng/u1	0.00
24) 2-Nitrophenol-d4	9.557	143	94988	43.031	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.099	165	187325	39.077	ng/u1	0.00
31) 4-Chloroaniline-d4	10.605	131	14228	2.488	ng/u1	0.00
46) Dimethylphthalate-d6	13.722	166	651863	42.815	ng/u1	0.00
49) Acenaphthylene-d8	14.010	160	551246	34.198	ng/u1	0.00
54) 4-Nitrophenol-d4	14.534	143	24063	9.240	ng/u1	0.00
60) Fluorene-d10	15.322	176	563769	41.862	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.440	200	103616	34.015	ng/u1	0.00
73) Anthracene-d10	17.216	188	916527	40.013	ng/u1	0.00
81) Pyrene-d10	19.592	212	1232784	40.835	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.586	264	1439943	37.738	ng/u1	-0.02

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092424\
Data File : BP022036.D
Acq On : 25 Sep 2024 20:10
Operator : RC/JU
Sample : P4013-04
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DDCH1

Quant Time: Sep 26 02:08:52 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP092424.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Sep 24 15:23:13 2024
Response via : Initial Calibration

