

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092521\
 Data File : BP007182.D
 Acq On : 25 Sep 2021 12:27
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Sep 26 02:12:52 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP092121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 23 11:17:25 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.881	152	198371	20.000	ng	-0.01
21) Naphthalene-d8	10.681	136	757043	20.000	ng	#-0.01
39) Acenaphthene-d10	14.516	164	495398	20.000	ng	-0.01
64) Phenanthrene-d10	17.269	188	1043900	20.000	ng	0.00
76) Chrysene-d12	21.351	240	1057588	20.000	ng	0.00
86) Perylene-d12	23.757	264	1134220	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.458	112	894969	75.521	ng	0.00
7) Phenol-d6	7.052	99	1211149	74.777	ng	-0.01
23) Nitrobenzene-d5	9.046	82	1015656	78.130	ng	-0.01
42) 2,4,6-Tribromophenol	16.004	330	572014	89.063	ng	-0.01
45) 2-Fluorobiphenyl	13.145	172	2675795	77.380	ng	0.00
79) Terphenyl-d14	19.822	244	4279368	77.539	ng	-0.01
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.346	88	174945	36.506	ng	# 87
3) Pyridine	3.758	79	484480	36.944	ng	# 87
4) n-Nitrosodimethylamine	3.664	42	168109	37.393	ng	# 77
6) Aniline	7.210	93	664585	37.225	ng	100
8) 2-Chlorophenol	7.446	128	491175	38.137	ng	99
9) Benzaldehyde	7.022	77	324116	35.525	ng	93
10) Phenol	7.081	94	576536	36.920	ng	91
11) bis(2-Chloroethyl)ether	7.310	93	448530	36.388	ng	96
12) 1,3-Dichlorobenzene	7.769	146	556273	38.290	ng	97
13) 1,4-Dichlorobenzene	7.916	146	569896	38.479	ng	98
14) 1,2-Dichlorobenzene	8.234	146	549362	38.278	ng	98
15) Benzyl Alcohol	8.128	79	404687	39.011	ng	94
16) 2,2'-oxybis(1-Chloropr...	8.416	45	507396	35.950	ng	90
17) 2-Methylphenol	8.328	107	401428	38.133	ng	94
18) Hexachloroethane	8.957	117	193287	38.871	ng	94
19) n-Nitroso-di-n-propyla...	8.693	70	324958	36.889	ng	98
20) 3+4-Methylphenols	8.657	107	551722	37.904	ng	93
22) Acetophenone	8.710	105	699279	38.335	ng	98
24) Nitrobenzene	9.087	77	481440	38.845	ng	95
25) Isophorone	9.616	82	879189	38.786	ng	97
26) 2-Nitrophenol	9.799	139	274249	42.435	ng	87
27) 2,4-Dimethylphenol	9.863	122	397267	39.041	ng	96
28) bis(2-Chloroethoxy)met...	10.098	93	596881	37.323	ng	98
29) 2,4-Dichlorophenol	10.334	162	481940	40.601	ng	98
30) 1,2,4-Trichlorobenzene	10.546	180	543111	40.163	ng	98
31) Naphthalene	10.734	128	1501335	38.858	ng	99
32) Benzoic acid	10.010	122	331840	42.430	ng	94
33) 4-Chloroaniline	10.845	127	629346	39.427	ng	98
34) Hexachlorobutadiene	11.016	225	340279	42.027	ng	98
35) Caprolactam	11.645	113	156860	42.954	ng	# 82
36) 4-Chloro-3-methylphenol	11.975	107	483483	40.072	ng	99
37) 2-Methylnaphthalene	12.345	142	1050857	38.843	ng	97
38) 1-Methylnaphthalene	12.563	142	1023078	38.703	ng	100
40) 1,2,4,5-Tetrachloroben...	12.710	216	609911	40.203	ng	98
41) Hexachlorocyclopentadiene	12.692	237	315137	37.480	ng	98
43) 2,4,6-Trichlorophenol	12.951	196	423605	40.921	ng	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092521\
 Data File : BP007182.D
 Acq On : 25 Sep 2021 12:27
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :


Quant Time: Sep 26 02:12:52 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP092121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 23 11:17:25 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	13.028	196	447136	40.112	ng	96
46) 1,1'-Biphenyl	13.351	154	1407585	38.199	ng	100
47) 2-Chloronaphthalene	13.392	162	1102716	38.632	ng	98
48) 2-Nitroaniline	13.598	65	285389	41.008	ng	93
49) Acenaphthylene	14.239	152	1725942	39.258	ng	100
50) Dimethylphthalate	13.981	163	1398716	39.256	ng	99
51) 2,6-Dinitrotoluene	14.092	165	299775	41.188	ng	89
52) Acenaphthene	14.581	154	1019982	38.293	ng	99
53) 3-Nitroaniline	14.422	138	326545	41.495	ng	98
54) 2,4-Dinitrophenol	14.634	184	182162	43.448	ng	91
55) Dibenzofuran	14.916	168	1717984	38.658	ng	95
56) 4-Nitrophenol	14.734	139	269111	42.123	ng	# 82
57) 2,4-Dinitrotoluene	14.886	165	411333	42.827	ng	87
58) Fluorene	15.563	166	1333642	38.845	ng	99
59) 2,3,4,6-Tetrachlorophenol	15.145	232	405144	41.389	ng	95
60) Diethylphthalate	15.339	149	1370209	39.687	ng	98
61) 4-Chlorophenyl-phenyle...	15.557	204	716555	39.496	ng	92
62) 4-Nitroaniline	15.586	138	338077	42.717	ng	# 81
63) Azobenzene	15.851	77	1150332	38.449	ng	93
65) 4,6-Dinitro-2-methylph...	15.651	198	275576	43.990	ng	94
66) n-Nitrosodiphenylamine	15.775	169	1197021	38.538	ng	99
67) 4-Bromophenyl-phenylether	16.451	248	456718	39.919	ng	95
68) Hexachlorobenzene	16.575	284	516948	40.883	ng	96
69) Atrazine	16.728	200	445878	40.725	ng	97
70) Pentachlorophenol	16.916	266	317583	40.443	ng	99
71) Phenanthrene	17.310	178	2168725	38.746	ng	99
72) Anthracene	17.398	178	2151684	39.296	ng	99
73) Carbazole	17.669	167	2120563	39.522	ng	99
74) Di-n-butylphthalate	18.222	149	2387762	40.156	ng	99
75) Fluoranthene	19.274	202	2592648	40.204	ng	98
77) Benzidine	19.451	184	1195057	36.769	ng	99
78) Pyrene	19.627	202	2673711	38.533	ng	99
80) Butylbenzylphthalate	20.498	149	1125291	40.496	ng	92
81) Benzo(a)anthracene	21.333	228	2646204	38.548	ng	100
82) 3,3'-Dichlorobenzidine	21.268	252	964651	40.917	ng	98
83) Chrysene	21.386	228	2510736	38.268	ng	99
84) Bis(2-ethylhexyl)phtha...	21.257	149	1564487	39.161	ng	98
85) Di-n-octyl phthalate	22.180	149	2720884	40.003	ng	99
87) Indeno(1,2,3-cd)pyrene	26.286	276	3190548	39.143	ng	97
88) Benzo(b)fluoranthene	23.021	252	2594035	38.270	ng	99
89) Benzo(k)fluoranthene	23.021	252	2594035	37.798	ng	99
90) Benzo(a)pyrene	23.651	252	2231110	38.743	ng	# 98
91) Dibenzo(a,h)anthracene	26.298	278	2655263	38.871	ng	# 96
92) Benzo(g,h,i)perylene	27.062	276	2630528	39.461	ng	# 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Sep 26 02:12:52 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP092121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 23 11:17:25 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.881 min Scan# 815
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:152 Resp: 198371
Ion Ratio Lower Upper
152 100
150 158.4 125.6 188.4
115 56.7 48.5 72.7



#2
1,4-Dioxane
Concen: 36.506 ng
RT: 3.346 min Scan# 44
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Tgt Ion: 88 Resp: 174945
Ion Ratio Lower Upper
88 100
58 71.6 46.7 70.1#
43 24.3 20.3 30.5









#3
Pyridine
Concen: 36.944 ng
RT: 3.758 min Scan# 114
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 79 Resp: 484480
Ion Ratio Lower Upper
79 100
52 64.9 41.7 62.5#
51 32.9 24.1 36.1



#4
n-Nitrosodimethylamine
Concen: 37.393 ng
RT: 3.664 min Scan# 98
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Tgt Ion: 42 Resp: 168109
Ion Ratio Lower Upper
42 100
74 160.1 105.4 158.0#
44 7.4 4.2 6.4#









#5
2-Fluorophenol
Concen: 75.521 ng
RT: 5.458 min Scan# 403
Delta R.T. -0.006 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:112 Resp: 894969
Ion Ratio Lower Upper
112 100
64 50.8 41.4 62.2
63 26.1 23.8 35.8



#6
Aniline
Concen: 37.225 ng
RT: 7.210 min Scan# 701
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Tgt Ion: 93 Resp: 664585
Ion Ratio Lower Upper
93 100
66 37.0 29.7 44.5
65 18.9 15.3 22.9









#7
Phenol-d6
Concen: 74.777 ng
RT: 7.052 min Scan# 674
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 99 Resp: 1211149
Ion Ratio Lower Upper
99 100
42 13.9 13.4 20.0
71 31.3 30.6 46.0



#8
2-Chlorophenol
Concen: 38.137 ng
RT: 7.446 min Scan# 741
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Tgt Ion: 128 Resp: 491175
Ion Ratio Lower Upper
128 100
130 32.0 12.4 52.4
64 41.6 20.3 60.3









#9
Benzaldehyde
Concen: 35.525 ng
RT: 7.022 min Scan# 669
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 77 Resp: 324116
Ion Ratio Lower Upper
77 100
105 106.2 76.7 116.7
106 101.4 76.2 116.2



#10
Phenol
Concen: 36.920 ng
RT: 7.081 min Scan# 679
Delta R.T. -0.006 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27


Tgt Ion: 94 Resp: 576536
Ion Ratio Lower Upper
94 100
65 25.3 8.8 48.8
66 34.3 20.7 60.7








#11
bis(2-Chloroethyl)ether
Concen: 36.388 ng
RT: 7.310 min Scan# 718
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Instrument :
BNA_P
ClientSampleId :







Tgt Ion: 93 Resp: 448530
Ion Ratio Lower Upper
93 100
63 71.1 46.2 86.2
95 32.0 12.2 52.2



#12
1,3-Dichlorobenzene
Concen: 38.290 ng
RT: 7.769 min Scan# 796
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Tgt Ion:146 Resp: 556273
Ion Ratio Lower Upper
146 100
148 64.8 51.5 77.3
75 27.3 25.7 38.5









#13
1,4-Dichlorobenzene
Concen: 38.479 ng
RT: 7.916 min Scan# 821
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:146 Resp: 569896
Ion Ratio Lower Upper
146 100
148 64.5 51.3 76.9
111 38.5 33.0 49.6



#14
1,2-Dichlorobenzene
Concen: 38.278 ng
RT: 8.234 min Scan# 875
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Tgt Ion:146 Resp: 549362
Ion Ratio Lower Upper
146 100
148 64.3 50.3 75.5
111 40.3 33.7 50.5





#15

Benzyl Alcohol

Concen: 39.011 ng

RT: 8.128 min Scan# 857

Delta R.T. -0.006 min

Lab File: BP007182.D

Acq: 25 Sep 2021 12:27





Tgt Ion: 79 Resp: 404687

Ion Ratio Lower Upper

79 100

108 85.1 62.5 93.7

77 63.6 52.9 79.3



#16

2,2'-oxybis(1-Chloropropane)

Concen: 35.950 ng

RT: 8.416 min Scan# 906

Delta R.T. -0.006 min

Lab File: BP007182.D

Acq: 25 Sep 2021 12:27




Tgt Ion: 45 Resp: 507396


Ion Ratio Lower Upper

45 100

77 16.6 2.0 42.0

79 12.8 0.0 36.7









#17
2-Methylphenol
Concen: 38.133 ng
RT: 8.328 min Scan# 891
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:	107	Resp:	401428
Ion	Ratio	Lower	Upper
107	100		
108	110.7	87.7	131.5
77	41.7	40.2	60.2
79	42.5	39.0	58.4



#18
Hexachloroethane
Concen: 38.871 ng
RT: 8.957 min Scan# 998
Delta R.T. -0.018 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Tgt Ion:	117	Resp:	193287
Ion	Ratio	Lower	Upper
117	100		
119	96.6	76.2	114.2
201	113.6	83.4	125.0









#19
n-Nitroso-di-n-propylamine
Concen: 36.889 ng
RT: 8.693 min Scan# 953
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 70	Resp: 324958
Ion Ratio	Lower Upper
70 100	
42 49.2	40.8 61.2
101 11.0	8.8 13.2
130 24.4	18.9 28.3



#20
3+4-Methylphenols
Concen: 37.904 ng
RT: 8.657 min Scan# 947
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Tgt Ion: 107	Resp: 551722
Ion Ratio	Lower Upper
107 100	
108 87.9	71.0 111.0
77 29.4	16.6 56.6
79 25.9	12.7 52.7









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.681 min Scan# 1291
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:	136	Resp:	757043
Ion Ratio	Lower	Upper	
136	100		
137	11.3	8.6	12.8
54	7.1	4.6	6.8#
68	5.3	3.6	5.4



#22
Acetophenone
Concen: 38.335 ng
RT: 8.710 min Scan# 956
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Tgt Ion:	105	Resp:	699279
Ion Ratio	Lower	Upper	
105	100		
71	2.2	1.9	2.9
51	21.8	18.4	27.6
120	23.1	17.9	26.9









#23
Nitrobenzene-d5
Concen: 78.130 ng
RT: 9.046 min Scan# 1013
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 82 Resp: 1015656
Ion Ratio Lower Upper
82 100
128 48.0 35.7 53.5
54 44.9 35.0 52.4



#24
Nitrobenzene
Concen: 38.845 ng
RT: 9.087 min Scan# 1020
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Tgt Ion: 77 Resp: 481440
Ion Ratio Lower Upper
77 100
123 51.2 38.2 57.2
65 12.8 11.4 17.2









#25
Isophorone
Concen: 38.786 ng
RT: 9.616 min Scan# 1110
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 82 Resp: 879189
Ion Ratio Lower Upper
82 100
95 7.1 6.5 9.7
138 20.3 15.0 22.4



#26
2-Nitrophenol
Concen: 42.435 ng
RT: 9.799 min Scan# 1141
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Tgt Ion:139 Resp: 274249
Ion Ratio Lower Upper
139 100
109 29.4 28.8 43.2
65 41.8 41.0 61.6









#27
2,4-Dimethylphenol
Concen: 39.041 ng
RT: 9.863 min Scan# 1152
Delta R.T. -0.006 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:122 Resp: 397267
Ion Ratio Lower Upper
122 100
107 112.6 94.5 141.7
121 58.6 47.4 71.2



#28
bis(2-Chloroethoxy)methane
Concen: 37.323 ng
RT: 10.098 min Scan# 1192
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Tgt Ion: 93 Resp: 596881
Ion Ratio Lower Upper
93 100
95 33.3 25.9 38.9
123 11.8 10.2 15.2









#29
2,4-Dichlorophenol
Concen: 40.601 ng
RT: 10.334 min Scan# 1232
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:162 Resp: 481940
Ion Ratio Lower Upper
162 100
164 64.7 44.9 84.9
98 32.1 14.3 54.3



#30
1,2,4-Trichlorobenzene
Concen: 40.163 ng
RT: 10.546 min Scan# 1268
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Tgt Ion:180 Resp: 543111
Ion Ratio Lower Upper
180 100
182 96.1 77.9 116.9
145 29.0 24.7 37.1









#31
Naphthalene
Concen: 38.858 ng
RT: 10.734 min Scan# 1300
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27




Instrument :
BNA_P
ClientSampled :


Tgt Ion:128 Resp: 1501335
Ion Ratio Lower Upper
128 100
129 10.7 9.0 13.6
127 13.3 10.7 16.1



#32
Benzoic acid
Concen: 42.430 ng
RT: 10.010 min Scan# 1177
Delta R.T. 0.006 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Tgt Ion:122 Resp: 331840
Ion Ratio Lower Upper
122 100
105 124.7 103.2 143.2
77 85.0 76.2 116.2









#33
4-Chloroaniline
Concen: 39.427 ng
RT: 10.845 min Scan# 1319
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:127	Resp: 629346
Ion Ratio	Lower Upper
127 100	
129 32.0	26.1 39.1
65 26.1	22.5 33.7
92 17.8	15.0 22.6



#34
Hexachlorobutadiene
Concen: 42.027 ng
RT: 11.016 min Scan# 1348
Delta R.T. -0.018 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Tgt Ion:225	Resp: 340279		
Ion Ratio	Lower	Upper	
225 100			
223 62.4	51.0	76.6	
227 62.6	51.4	77.2	









#35
Caprolactam
Concen: 42.954 ng
RT: 11.645 min Scan# 1455
Delta R.T. -0.000 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:113 Resp: 156860
Ion Ratio Lower Upper
113 100
55 141.3 98.9 138.9#
56 105.3 71.7 111.7



#36
4-Chloro-3-methylphenol
Concen: 40.072 ng
RT: 11.975 min Scan# 1511
Delta R.T. -0.006 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Tgt Ion:107 Resp: 483483
Ion Ratio Lower Upper
107 100
144 27.5 21.1 31.7
142 83.0 66.0 99.0









#37
2-Methylnaphthalene
Concen: 38.843 ng
RT: 12.345 min Scan# 1574
Delta R.T. -0.006 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:142 Resp: 1050857
Ion Ratio Lower Upper
142 100
141 89.5 71.8 107.6
115 33.4 31.6 47.4



#38
1-Methylnaphthalene
Concen: 38.703 ng
RT: 12.563 min Scan# 1611
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Tgt Ion:142 Resp: 1023078
Ion Ratio Lower Upper
142 100
141 91.8 73.3 109.9
116 3.6 3.5 5.3









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.516 min Scan# 1943
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:164 Resp: 495398
Ion Ratio Lower Upper
164 100
162 100.1 80.2 120.2
160 44.5 34.5 51.7



#40
1,2,4,5-Tetrachlorobenzene
Concen: 40.203 ng
RT: 12.710 min Scan# 1636
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Tgt Ion:216 Resp: 609911
Ion Ratio Lower Upper
216 100
214 78.0 61.9 92.9
179 19.9 16.8 25.2
108 15.6 14.9 22.3









#41
Hexachlorocyclopentadiene
Concen: 37.480 ng
RT: 12.692 min Scan# 1633
Delta R.T. -0.006 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:	237	Resp:	315137
Ion Ratio	Lower	Upper	
237	100		
235	62.6	44.5	84.5
272	13.2	0.0	32.5



#42
2,4,6-Tribromophenol
Concen: 89.063 ng
RT: 16.004 min Scan# 2196
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Tgt Ion:	330	Resp:	572014
Ion Ratio	Lower	Upper	
330	100		
332	96.9	77.8	116.8
141	28.7	25.4	38.2









#43
2,4,6-Trichlorophenol
Concen: 40.921 ng
RT: 12.951 min Scan# 1677
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27




Instrument :
BNA_P
ClientSampled :


Tgt Ion:196 Resp: 423605
Ion Ratio Lower Upper
196 100
198 95.6 79.5 119.3
200 30.1 25.0 37.6



#44
2,4,5-Trichlorophenol
Concen: 40.112 ng
RT: 13.028 min Scan# 1690
Delta R.T. -0.000 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Tgt Ion:196 Resp: 447136
Ion Ratio Lower Upper
196 100
198 96.2 74.7 112.1
97 40.4 37.4 56.0
132 24.8 20.0 30.0









#45
2-Fluorobiphenyl
Concen: 77.380 ng
RT: 13.145 min Scan# 1710
Delta R.T. -0.006 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:172 Resp: 2675795
Ion Ratio Lower Upper
172 100
171 35.5 27.9 41.9
170 23.4 18.6 27.8



#46
1,1'-Biphenyl
Concen: 38.199 ng
RT: 13.351 min Scan# 1745
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Tgt Ion:154 Resp: 1407585
Ion Ratio Lower Upper
154 100
153 40.3 20.4 60.4
76 11.3 0.0 30.8









#47
2-Chloronaphthalene
Concen: 38.632 ng
RT: 13.392 min Scan# 1752
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27




Instrument :
BNA_P
ClientSampled :


Tgt Ion:162 Resp: 1102716
Ion Ratio Lower Upper
162 100
127 37.0 31.4 47.0
164 33.3 26.6 39.8



#48
2-Nitroaniline
Concen: 41.008 ng
RT: 13.598 min Scan# 1787
Delta R.T. -0.006 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Tgt Ion: 65 Resp: 285389
Ion Ratio Lower Upper
65 100
92 75.2 57.7 86.5
138 125.5 92.1 138.1









#49
Acenaphthylene
Concen: 39.258 ng
RT: 14.239 min Scan# 1896
Delta R.T. -0.006 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:152 Resp: 1725942
Ion Ratio Lower Upper
152 100
151 19.9 16.2 24.2
153 13.4 10.6 16.0



#50
Dimethylphthalate
Concen: 39.256 ng
RT: 13.981 min Scan# 1852
Delta R.T. -0.006 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Tgt Ion:163 Resp: 1398716
Ion Ratio Lower Upper
163 100
194 4.3 3.4 5.2
164 9.9 8.2 12.2









#51
2,6-Dinitrotoluene
Concen: 41.188 ng
RT: 14.092 min Scan# 1871
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27




Instrument :
BNA_P
ClientSampled :


Tgt Ion:165 Resp: 299775
Ion Ratio Lower Upper
165 100
63 44.8 41.3 61.9
89 47.5 44.8 67.2



#52
Acenaphthene
Concen: 38.293 ng
RT: 14.581 min Scan# 1954
Delta R.T. -0.006 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Tgt Ion:154 Resp: 1019982
Ion Ratio Lower Upper
154 100
153 113.2 89.4 134.2
152 53.4 43.0 64.4









#53
3-Nitroaniline
Concen: 41.495 ng
RT: 14.422 min Scan# 1927
Delta R.T. -0.006 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:138 Resp: 326545
Ion Ratio Lower Upper
138 100
108 9.6 8.4 12.6
92 108.0 87.8 131.8



#54
2,4-Dinitrophenol
Concen: 43.448 ng
RT: 14.634 min Scan# 1963
Delta R.T. -0.006 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Tgt Ion:184 Resp: 182162
Ion Ratio Lower Upper
184 100
63 51.0 47.0 70.6
154 59.3 52.3 78.5









#55
Dibenzenofuran
Concen: 38.658 ng
RT: 14.916 min Scan# 2011
Delta R.T. -0.006 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:168 Resp: 1717984
Ion Ratio Lower Upper
168 100
139 35.1 31.0 46.6
169 13.4 10.8 16.2



#56
4-Nitrophenol
Concen: 42.123 ng
RT: 14.734 min Scan# 1980
Delta R.T. -0.000 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Tgt Ion:139 Resp: 269111
Ion Ratio Lower Upper
139 100
109 67.7 72.3 112.3#
65 81.9 72.5 112.5









#57
2,4-Dinitrotoluene
Concen: 42.827 ng
RT: 14.886 min Scan# 2006
Delta R.T. -0.000 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:165	Resp: 411333		
Ion	Ratio	Lower	Upper
165	100		
63	33.3	32.3	48.5
89	58.7	55.9	83.9
182	2.8	2.2	3.4



#58
Fluorene
Concen: 38.845 ng
RT: 15.563 min Scan# 2121
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Tgt Ion:166	Resp: 1333642		
Ion	Ratio	Lower	Upper
166	100		
165	96.8	78.0	117.0
167	13.4	10.5	15.7









#59
2,3,4,6-Tetrachlorophenol
Concen: 41.389 ng
RT: 15.145 min Scan# 2050
Delta R.T. -0.006 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:	232	Resp:	405144
Ion Ratio	Lower	Upper	
232	100		
131	38.5	32.2	48.4
130	2.1	2.0	3.0
166	28.5	19.4	29.2



#60
Diethylphthalate
Concen: 39.687 ng
RT: 15.339 min Scan# 2083
Delta R.T. -0.006 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Tgt Ion:	149	Resp:	1370209
Ion Ratio	Lower	Upper	
149	100		
177	23.5	17.5	26.3
150	12.5	9.9	14.9









#61
4-Chlorophenyl-phenylether
Concen: 39.496 ng
RT: 15.557 min Scan# 2120
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:204 Resp: 716555
Ion Ratio Lower Upper
204 100
206 33.2 26.6 40.0
141 51.7 48.6 73.0



#62
4-Nitroaniline
Concen: 42.717 ng
RT: 15.586 min Scan# 2125
Delta R.T. -0.006 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Tgt Ion:138 Resp: 338077
Ion Ratio Lower Upper
138 100
92 46.8 31.2 71.2
108 72.5 78.6 118.6#









#63
Azobenzene
Concen: 38.449 ng
RT: 15.851 min Scan# 2170
Delta R.T. -0.006 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Instrument :
BNA_P
ClientSampleId :

Tgt Ion: 77 Resp: 1150332




Ion	Ratio	Lower	Upper
77	100		
182	30.4	6.0	46.0
105	22.6	0.0	38.8
51	26.1	8.8	48.8




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.269 min Scan# 2411
Delta R.T. -0.006 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Tgt Ion: 188 Resp: 1043900

Ion	Ratio	Lower	Upper
188	100		
94	8.1	5.6	8.4
80	8.6	6.3	9.5









#65
4,6-Dinitro-2-methylphenol
Concen: 43.990 ng
RT: 15.651 min Scan# 2136
Delta R.T. -0.000 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:198 Resp: 275576
Ion Ratio Lower Upper
198 100
51 31.0 14.0 54.0
105 35.4 19.5 59.5



#66
n-Nitrosodiphenylamine
Concen: 38.538 ng
RT: 15.775 min Scan# 2157
Delta R.T. -0.006 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Tgt Ion:169 Resp: 1197021
Ion Ratio Lower Upper
169 100
168 66.7 54.4 81.6
167 36.1 29.0 43.6









#67
4-Bromophenyl-phenylether
Concen: 39.919 ng
RT: 16.451 min Scan# 2272
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27




Instrument :
BNA_P
ClientSampled :


Tgt Ion:248 Resp: 456718
Ion Ratio Lower Upper
248 100
250 97.2 79.8 119.8
141 65.6 57.8 86.6



#68
Hexachlorobenzene
Concen: 40.883 ng
RT: 16.575 min Scan# 2293
Delta R.T. -0.006 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Tgt Ion:284 Resp: 516948
Ion Ratio Lower Upper
284 100
142 29.0 26.9 40.3
249 30.4 24.5 36.7









#69
Atrazine
Concen: 40.725 ng
RT: 16.728 min Scan# 2319
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:200 Resp: 445878
Ion Ratio Lower Upper
200 100
173 26.5 4.1 44.1
215 49.4 27.7 67.7



#70
Pentachlorophenol
Concen: 40.443 ng
RT: 16.916 min Scan# 2351
Delta R.T. -0.006 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Tgt Ion:266 Resp: 317583
Ion Ratio Lower Upper
266 100
268 64.3 51.6 77.4
264 62.2 50.3 75.5









#71
Phenanthrene
Concen: 38.746 ng
RT: 17.310 min Scan# 2418
Delta R.T. -0.006 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:178 Resp: 2168725
Ion Ratio Lower Upper
178 100
176 19.4 15.1 22.7
179 15.5 12.4 18.6



#72
Anthracene
Concen: 39.296 ng
RT: 17.398 min Scan# 2433
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Tgt Ion:178 Resp: 2151684
Ion Ratio Lower Upper
178 100
176 18.8 14.9 22.3
179 15.5 12.2 18.2









#73
 Carbazole
 Concen: 39.522 ng
 RT: 17.669 min Scan# 2479
 Delta R.T. -0.012 min
 Lab File: BP007182.D
 Acq: 25 Sep 2021 12:27




Instrument :
 BNA_P
 ClientSampleId :


Tgt Ion:167 Resp: 2120563
 Ion Ratio Lower Upper
 167 100
 166 21.3 17.1 25.7
 139 12.1 10.7 16.1



#74
 Di-n-butylphthalate
 Concen: 40.156 ng
 RT: 18.222 min Scan# 2573
 Delta R.T. -0.006 min
 Lab File: BP007182.D
 Acq: 25 Sep 2021 12:27

Tgt Ion:149 Resp: 2387762
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.3 10.9
 104 4.6 4.5 6.7









#75
Fluoranthene
Concen: 40.204 ng
RT: 19.274 min Scan# 2752
Delta R.T. -0.006 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:202 Resp: 2592648
Ion Ratio Lower Upper
202 100
101 9.8 0.0 27.6
203 17.5 0.0 37.8



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.351 min Scan# 3105
Delta R.T. -0.006 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Tgt Ion:240 Resp: 1057588
Ion Ratio Lower Upper
240 100
120 9.4 6.5 9.7
236 25.7 20.6 31.0









#77
Benzidine
Concen: 36.769 ng
RT: 19.451 min Scan# 2782
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:184 Resp: 1195057
Ion Ratio Lower Upper
184 100
185 14.3 11.0 16.4
183 11.9 9.5 14.3



#78
Pyrene
Concen: 38.533 ng
RT: 19.627 min Scan# 2812
Delta R.T. -0.006 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Tgt Ion:202 Resp: 2673711
Ion Ratio Lower Upper
202 100
200 20.9 16.3 24.5
203 17.5 14.2 21.2









#79
 Terphenyl-d14
 Concen: 77.539 ng
 RT: 19.822 min Scan# 2845
 Delta R.T. -0.012 min
 Lab File: BP007182.D
 Acq: 25 Sep 2021 12:27




Instrument :
 BNA_P
 ClientSampleId :


Tgt Ion:244 Resp: 4279368
 Ion Ratio Lower Upper
 244 100
 212 7.1 6.0 9.0
 122 9.1 7.1 10.7



#80
 Butylbenzylphthalate
 Concen: 40.496 ng
 RT: 20.498 min Scan# 2960
 Delta R.T. -0.012 min
 Lab File: BP007182.D
 Acq: 25 Sep 2021 12:27

Tgt Ion:149 Resp: 1125291
 Ion Ratio Lower Upper
 149 100
 91 61.3 54.6 81.8
 206 23.4 16.7 25.1









#81
Benzo(a)anthracene
Concen: 38.548 ng
RT: 21.333 min Scan# 3102
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27




Instrument :
BNA_P
ClientSampled :


Tgt Ion:228 Resp: 2646204
Ion Ratio Lower Upper
228 100
226 27.1 21.6 32.4
229 20.1 15.8 23.6



#82
3,3'-Dichlorobenzidine
Concen: 40.917 ng
RT: 21.268 min Scan# 3091
Delta R.T. -0.006 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Tgt Ion:252 Resp: 964651
Ion Ratio Lower Upper
252 100
254 64.9 50.6 76.0
126 9.1 7.0 10.6









#83
Chrysene
Concen: 38.268 ng
RT: 21.386 min Scan# 3111
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:228 Resp: 2510736
Ion Ratio Lower Upper
228 100
226 30.4 23.8 35.6
229 20.1 15.4 23.2



#84
Bis(2-ethylhexyl)phthalate
Concen: 39.161 ng
RT: 21.257 min Scan# 3089
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Tgt Ion:149 Resp: 1564487
Ion Ratio Lower Upper
149 100
167 28.5 21.9 32.9
279 4.9 4.3 6.5









#85
Di-n-octyl phthalate
Concen: 40.003 ng
RT: 22.180 min Scan# 3246
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:149 Resp: 2720884
Ion Ratio Lower Upper
149 100
167 1.7 1.4 2.0
43 6.7 5.6 8.4



#86
Perylene-d12
Concen: 20.000 ng
RT: 23.757 min Scan# 3514
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Tgt Ion:264 Resp: 1134220
Ion Ratio Lower Upper
264 100
260 23.8 18.6 28.0
265 22.0 17.6 26.4









#87
Indeno(1,2,3-cd)pyrene
Concen: 39.143 ng
RT: 26.286 min Scan# 3944
Delta R.T. -0.018 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27




Instrument :
BNA_P
ClientSampled :


Tgt Ion:276 Resp: 3190548
Ion Ratio Lower Upper
276 100
138 22.4 15.3 22.9
277 25.4 20.4 30.6



#88
Benzo(b)fluoranthene
Concen: 38.270 ng
RT: 23.021 min Scan# 3389
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Tgt Ion:252 Resp: 2594035
Ion Ratio Lower Upper
252 100
253 22.0 17.5 26.3
125 8.9 6.2 9.4









#89
Benzo(k)fluoranthene
Concen: 37.798 ng
RT: 23.021 min Scan# 3389
Delta R.T. -0.065 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:252 Resp: 2594035
Ion Ratio Lower Upper
252 100
253 22.0 17.6 26.4
125 8.9 6.0 9.0



#90
Benzo(a)pyrene
Concen: 38.743 ng
RT: 23.651 min Scan# 3496
Delta R.T. -0.018 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Tgt Ion:252 Resp: 2231110
Ion Ratio Lower Upper
252 100
253 22.2 17.9 26.9
125 9.9 6.5 9.7#









#91
Dibenzo(a,h)anthracene
Concen: 38.871 ng
RT: 26.298 min Scan# 3946
Delta R.T. -0.018 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Instrument :
BNA_P
ClientSampled :

Tgt Ion:278 Resp: 2655263
Ion Ratio Lower Upper
278 100
139 14.7 9.3 13.9#
279 24.3 18.8 28.2



#92
Benzo(g,h,i)perylene
Concen: 39.461 ng
RT: 27.062 min Scan# 4076
Delta R.T. -0.012 min
Lab File: BP007182.D
Acq: 25 Sep 2021 12:27

Tgt Ion:276 Resp: 2630528
Ion Ratio Lower Upper
276 100
277 23.4 18.9 28.3
138 19.2 12.6 18.8#

