

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092523\
 Data File : BP017358.D
 Acq On : 26 Sep 2023 13:03
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020752

Quant Time: Sep 27 00:27:22 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP092023.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Sep 26 04:52:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.916	152	173800	20.000	ng/ul	-0.01
20) Naphthalene-d8	10.746	136	725510	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.587	164	445361	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.363	188	990470	20.000	ng/ul	0.00
79) Chrysene-d12	21.469	240	966412	20.000	ng/ul	0.00
88) Perylene-d12	23.986	264	1062410	20.000	ng/ul	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.281	96	33132	7.830	ng/uL	-0.02
4) Pyridine-d5	3.717	84	235765	19.931	ng/ul	-0.01
7) Phenol-d5	7.069	99	295481	20.700	ng/ul	-0.01
9) Bis-(2-Chloroethyl)eth...	7.258	67	179165	20.798	ng/ul	-0.01
11) 2-Chlorophenol-d4	7.440	132	240626	21.254	ng/ul	0.00
15) 4-Methylphenol-d8	8.634	113	234032	21.089	ng/ul	0.00
21) Nitrobenzene-d5	9.116	128	114996	21.925	ng/ul	-0.01
24) 2-Nitrophenol-d4	9.834	143	127412	22.372	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.363	165	242455	22.534	ng/ul	0.00
31) 4-Chloroaniline-d4	10.910	131	335097	21.499	ng/ul	0.00
46) Dimethylphthalate-d6	14.004	166	692159	21.695	ng/ul	0.00
49) Acenaphthylene-d8	14.287	160	789811	21.524	ng/ul	0.00
54) 4-Nitrophenol-d4	14.816	143	107706	19.378	ng/ul	0.00
60) Fluorene-d10	15.587	176	594687	21.651	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.722	200	103031	18.456	ng/ul	0.00
73) Anthracene-d10	17.463	188	943950	21.489	ng/ul	0.00
81) Pyrene-d10	19.704	212	1120297	21.437	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.821	264	1091644	21.341	ng/ul	0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.323	88	34338	7.924	ng/uL	88
5) Pyridine	3.740	79	245896	20.023	ng/ul	98
6) Benzaldehyde	7.075	77	177627	24.385	ng/ul	98
8) Phenol	7.099	94	295949	20.614	ng/ul	97
10) Bis(2-Chloroethyl)ether	7.358	93	244750	21.090	ng/ul	98
12) 2-Chlorophenol	7.469	128	242527	21.095	ng/ul	100
13) 2-Methylphenol	8.363	108	222121	20.908	ng/ul	98
14) 2,2'-oxybis(1-Chloropr...	8.434	45	207764	14.155	ng/ul	99
16) Acetophenone	8.769	105	368915	21.513	ng/ul	98
17) N-Nitroso-di-n-propyla...	8.740	70	186541	21.796	ng/ul	97
18) 4-Methylphenol	8.699	108	241713	21.235	ng/ul	99
19) Hexachloroethane	8.981	117	100350	21.026	ng/ul	97
22) Nitrobenzene	9.163	77	285232	21.432	ng/ul	97
23) Isophorone	9.675	82	513789	21.544	ng/ul	99
25) 2-Nitrophenol	9.869	139	137993	22.360	ng/ul	99
26) 2,4-Dimethylphenol	9.910	107	266860	21.703	ng/ul	97
27) Bis(2-Chloroethoxy)met...	10.163	93	323669	21.024	ng/ul	97
29) 2,4-Dichlorophenol	10.387	162	236323	22.277	ng/ul	99
30) Naphthalene	10.799	128	793553	21.006	ng/ul	99
32) 4-Chloroaniline	10.934	127	323499	21.659	ng/ul	99
33) Hexachlorobutadiene	11.022	225	163858	22.194	ng/ul	97
34) Caprolactam	11.757	113	64408	21.012	ng/ul	95
35) 4-Chloro-3-methylphenol	12.040	107	240191	22.525	ng/ul	98
36) 2-Methylnaphthalene	12.404	142	531333	21.550	ng/ul	100


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092523\
 Data File : BP017358.D
 Acq On : 26 Sep 2023 13:03
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020752

Quant Time: Sep 27 00:27:22 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP092023.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Sep 26 04:52:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.628	142	541869	21.499	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.757	216	304627	21.314	ng/ul	99
40) Hexachlorocyclopentadiene	12.699	237	129998	16.072	ng/ul	96
41) 2,4,6-Trichlorophenol	13.016	196	192734	22.231	ng/ul	99
42) 2,4,5-Trichlorophenol	13.087	196	203133	22.364	ng/ul	99
43) 1,1'-Biphenyl	13.422	154	706197	20.900	ng/ul	98
44) 2-Chloronaphthalene	13.469	162	564488	21.072	ng/ul	99
45) 2-Nitroaniline	13.704	65	149820	22.766	ng/ul	96
47) Dimethylphthalate	14.051	163	687879	21.404	ng/ul	98
48) 2,6-Dinitrotoluene	14.198	165	132300	23.701	ng/ul	97
50) Acenaphthylene	14.316	152	913018	21.279	ng/ul	100
51) 3-Nitroaniline	14.540	138	135877	22.619	ng/ul	94
52) Acenaphthene	14.651	153	598535	21.122	ng/ul	99
53) 2,4-Dinitrophenol	14.740	184	51348	14.705	ng/ul	94
55) 4-Nitrophenol	14.834	109	90189	20.770	ng/ul	92
56) Dibenzofuran	14.993	168	835180	21.522	ng/ul	98
57) 2,4-Dinitrotoluene	14.981	165	196648	23.916	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	15.210	232	175749	22.804	ng/ul	94
59) Diethylphthalate	15.410	149	678748	21.846	ng/ul	98
61) Fluorene	15.645	166	681373	21.611	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.634	204	351672	22.008	ng/ul	99
63) 4-Nitroaniline	15.710	138	128152	23.903	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.740	198	110359	18.325	ng/ul	97
67) N-Nitrosodiphenylamine	15.863	169	569205	20.893	ng/ul	98
68) 4-Bromophenyl-phenylether	16.539	248	215269	21.634	ng/ul	98
69) Hexachlorobenzene	16.622	284	252113	21.835	ng/ul	95
70) Atrazine	16.822	200	205488	21.053	ng/ul	99
71) Pentachlorophenol	16.987	266	145776	20.438	ng/ul	98
72) Phenanthrene	17.404	178	1090332	21.018	ng/ul	99
74) Anthracene	17.498	178	1115975	21.226	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.375	216	307700	20.242	ng/uL	98
76) Pentachlorobenzene	14.881	250	301150	20.972	ng/uL	98
77) Carbazole	17.786	167	982401	21.007	ng/ul	99
78) Di-n-butylphthalate	18.298	149	1168670	22.726	ng/ul	99
80) Fluoranthene	19.375	202	1301651	21.356	ng/ul	100
82) Pyrene	19.733	202	1373832	21.204	ng/ul	99
83) Butylbenzylphthalate	20.598	149	547448	23.830	ng/ul	94
84) 3,3'-Dichlorobenzidine	21.392	252	439326	20.923	ng/ul	100
85) Benzo(a)anthracene	21.451	228	1394314	21.377	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.327	149	818974	24.685	ng/ul	99
87) Chrysene	21.510	228	1300978	21.194	ng/ul	100
89) Di-n-octyl phthalate	22.292	149	1420717	24.311	ng/ul	100
90) Benzo(b)fluoranthene	23.204	252	1323621	21.131	ng/ul	99
91) Benzo(k)fluoranthene	23.257	252	1345360	21.289	ng/ul	98
93) Benzo(a)pyrene	23.874	252	1258990	21.007	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.651	276	1525860	20.312	ng/ul	98
95) Dibenzo(a,h)anthracene	26.674	278	1265749	20.230	ng/ul	98
96) Benzo(g,h,i)perylene	27.486	276	1208772	19.954	ng/ul	98





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
 1,4-Dioxane
 Concen: 7.924 ng/uL
 RT: 3.323 min Scan# 40
 Delta R.T. -0.012 min
 Lab File: BP017358.D
 Acq: 26 Sep 2023 13:03




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020752


Tgt Ion: 88 Resp: 34338
 Ion Ratio Lower Upper
 88 100
 43 28.0 25.0 37.4
 58 59.0 56.6 84.8



#5
 Pyridine
 Concen: 20.023 ng/uL
 RT: 3.740 min Scan# 111
 Delta R.T. -0.012 min
 Lab File: BP017358.D
 Acq: 26 Sep 2023 13:03

Tgt Ion: 79 Resp: 245896
 Ion Ratio Lower Upper
 79 100
 52 63.1 51.4 77.2
 51 32.7 25.4 38.2









#6
Benzaldehyde
Concen: 24.385 ng/ul
RT: 7.075 min Scan# 61
Delta R.T. -0.012 min
Lab File: BP017358.D
Acq: 26 Sep 2023 13:03




Instrument :
BNA_P
ClientSampleId :
SSTD020752


Tgt Ion: 77 Resp: 177627
Ion Ratio Lower Upper
77 100
105 95.2 78.0 117.0
106 90.7 73.2 109.8



#8
Phenol
Concen: 20.614 ng/ul
RT: 7.099 min Scan# 682
Delta R.T. -0.012 min
Lab File: BP017358.D
Acq: 26 Sep 2023 13:03

Tgt Ion: 94 Resp: 295949
Ion Ratio Lower Upper
94 100
65 27.6 21.4 32.2
66 37.7 28.2 42.2









#10
 Bis(2-Chloroethyl)ether
 Concen: 21.090 ng/ul
 RT: 7.358 min Scan# 71
 Delta R.T. -0.006 min
 Lab File: BP017358.D
 Acq: 26 Sep 2023 13:03

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020752

Tgt Ion: 93 Resp: 244750




Ion	Ratio	Lower	Upper
93	100		
63	71.2	55.4	83.0
95	32.5	25.1	37.7




#12
 2-Chlorophenol
 Concen: 21.095 ng/ul
 RT: 7.469 min Scan# 745
 Delta R.T. -0.012 min
 Lab File: BP017358.D
 Acq: 26 Sep 2023 13:03

Tgt Ion: 128 Resp: 242527

Ion	Ratio	Lower	Upper
128	100		
64	45.0	35.8	53.8
130	32.2	25.6	38.4





#13

2-Methylphenol

Concen: 20.908 ng/ul

RT: 8.363 min Scan# 898

Delta R.T. -0.006 min

Lab File: BP017358.D

Acq: 26 Sep 2023 13:03

Instrument :

BNA_P

ClientSampleId :





SSTD020752

Tgt Ion:108 Resp: 222121

Ion Ratio Lower Upper

108 100

107 90.1 73.5 110.3



#14

2,2'-oxybis(1-Chloropropane)

Concen: 14.155 ng/ul

RT: 8.434 min Scan# 909

Delta R.T. -0.006 min

Lab File: BP017358.D

Acq: 26 Sep 2023 13:03




Tgt Ion: 45 Resp: 207764


Ion Ratio Lower Upper

45 100

77 17.5 13.3 19.9

79 11.8 9.4 14.2









#16
Acetophenone
Concen: 21.513 ng/ul
RT: 8.769 min Scan# 90
Delta R.T. -0.006 min
Lab File: BP017358.D
Acq: 26 Sep 2023 13:03




Instrument :
BNA_P
ClientSampleId :
SSTD020752


Tgt Ion:105 Resp: 368915
Ion Ratio Lower Upper
105 100
77 77.2 63.6 95.4
51 24.0 19.2 28.8



#17
N-Nitroso-di-n-propylamine
Concen: 21.796 ng/ul
RT: 8.740 min Scan# 961
Delta R.T. -0.006 min
Lab File: BP017358.D
Acq: 26 Sep 2023 13:03

Tgt Ion: 70 Resp: 186541
Ion Ratio Lower Upper
70 100
42 55.6 42.3 63.5
101 9.7 7.5 11.3
130 20.9 17.3 25.9









#18
4-Methylphenol
Concen: 21.235 ng/ul
RT: 8.699 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP017358.D
Acq: 26 Sep 2023 13:03




Instrument :
BNA_P
ClientSampleId :
SSTD020752


Tgt Ion:108 Resp: 241713
Ion Ratio Lower Upper
108 100
107 111.8 90.0 135.0



#19
Hexachloroethane
Concen: 21.026 ng/ul
RT: 8.981 min Scan# 1002
Delta R.T. -0.006 min
Lab File: BP017358.D
Acq: 26 Sep 2023 13:03

Tgt Ion:117 Resp: 100350
Ion Ratio Lower Upper
117 100
201 107.5 81.4 122.0
199 65.5 52.4 78.6









#22
Nitrobenzene
Concen: 21.432 ng/ul
RT: 9.163 min Scan# 110
Delta R.T. -0.006 min
Lab File: BP017358.D
Acq: 26 Sep 2023 13:03




Instrument :
BNA_P
ClientSampleId :
SSTD020752


Tgt Ion: 77 Resp: 285232
Ion Ratio Lower Upper
77 100
123 44.8 37.4 56.0
65 13.2 9.8 14.6



#23
Isophorone
Concen: 21.544 ng/ul
RT: 9.675 min Scan# 1120
Delta R.T. -0.006 min
Lab File: BP017358.D
Acq: 26 Sep 2023 13:03

Tgt Ion: 82 Resp: 513789
Ion Ratio Lower Upper
82 100
95 7.2 5.5 8.3
138 18.3 14.2 21.4









#25
2-Nitrophenol
Concen: 22.360 ng/ul
RT: 9.869 min Scan# 1153
Delta R.T. 0.000 min
Lab File: BP017358.D
Acq: 26 Sep 2023 13:03




Instrument :
BNA_P
ClientSampleId :
SSTD020752


Tgt Ion:139 Resp: 137993
Ion Ratio Lower Upper
139 100
65 50.6 40.6 60.8
109 30.9 24.2 36.4



#26
2,4-Dimethylphenol
Concen: 21.703 ng/ul
RT: 9.910 min Scan# 1160
Delta R.T. 0.000 min
Lab File: BP017358.D
Acq: 26 Sep 2023 13:03

Tgt Ion:107 Resp: 266860
Ion Ratio Lower Upper
107 100
121 49.5 41.1 61.7
122 84.1 69.4 104.0





#27

Bis(2-Chloroethoxy)methane

Concen: 21.024 ng/ul

RT: 10.163 min Scan# 1203

Delta R.T. -0.006 min

Lab File: BP017358.D

Acq: 26 Sep 2023 13:03

Instrument :

BNA_P

ClientSampleId :

SSTD020752





Tgt Ion: 93 Resp: 323669

Ion Ratio Lower Upper

93 100

95 33.0 24.9 37.3

123 12.0 9.4 14.2



#29

2,4-Dichlorophenol

Concen: 22.277 ng/ul

RT: 10.387 min Scan# 1241

Delta R.T. -0.006 min

Lab File: BP017358.D

Acq: 26 Sep 2023 13:03




Tgt Ion:162 Resp: 236323


Ion Ratio Lower Upper

162 100

164 65.4 52.6 78.8

98 34.8 27.0 40.4









#30
Naphthalene
Concen: 21.006 ng/ul
RT: 10.799 min Scan# 1311
Delta R.T. 0.000 min
Lab File: BP017358.D
Acq: 26 Sep 2023 13:03




Instrument :
BNA_P
ClientSampleId :
SSTD020752


Tgt Ion:128 Resp: 793553
Ion Ratio Lower Upper
128 100
129 10.7 8.6 13.0
127 12.9 10.6 15.8



#32
4-Chloroaniline
Concen: 21.659 ng/ul
RT: 10.934 min Scan# 1334
Delta R.T. -0.006 min
Lab File: BP017358.D
Acq: 26 Sep 2023 13:03

Tgt Ion:127 Resp: 323499
Ion Ratio Lower Upper
127 100
129 32.0 25.1 37.7





#33

Hexachlorobutadiene

Concen: 22.194 ng/ul

RT: 11.022 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP017358.D

Acq: 26 Sep 2023 13:03

Instrument :

BNA_P

ClientSampleId :

SSTD020752





Tgt Ion:225 Resp: 163858

Ion Ratio Lower Upper

225 100

223 64.4 48.2 72.2

227 63.6 50.2 75.4



#34

Caprolactam

Concen: 21.012 ng/ul

RT: 11.757 min Scan# 1474

Delta R.T. -0.006 min

Lab File: BP017358.D

Acq: 26 Sep 2023 13:03




Tgt Ion:113 Resp: 64408


Ion Ratio Lower Upper

113 100

55 152.1 127.4 191.0

56 112.5 94.5 141.7









#35
4-Chloro-3-methylphenol
Concen: 22.525 ng/ul
RT: 12.040 min Scan# 1523
Delta R.T. -0.006 min
Lab File: BP017358.D
Acq: 26 Sep 2023 13:03




Instrument :
BNA_P
ClientSampleId :
SSTD020752


Tgt Ion:107 Resp: 240191
Ion Ratio Lower Upper
107 100
144 26.1 20.6 31.0
142 80.3 62.9 94.3



#36
2-Methylnaphthalene
Concen: 21.550 ng/ul
RT: 12.404 min Scan# 1584
Delta R.T. 0.000 min
Lab File: BP017358.D
Acq: 26 Sep 2023 13:03

Tgt Ion:142 Resp: 531333
Ion Ratio Lower Upper
142 100
141 89.5 71.9 107.9





#37

1-Methylnaphthalene

Concen: 21.499 ng/ul

RT: 12.628 min Scan# 1622

Delta R.T. 0.000 min

Lab File: BP017358.D

Acq: 26 Sep 2023 13:03

Instrument :

BNA_P

ClientSampleId :

SSTD020752





Tgt Ion:142 Resp: 541869

Ion Ratio Lower Upper

142 100

141 92.6 73.4 110.2

116 3.9 3.1 4.7



#39

1,2,4,5-Tetrachlorobenzene

Concen: 21.314 ng/ul

RT: 12.757 min Scan# 1644

Delta R.T. 0.000 min

Lab File: BP017358.D

Acq: 26 Sep 2023 13:03

Tgt Ion:216 Resp: 304627




Ion Ratio Lower Upper


216 100

214 78.6 62.3 93.5

179 21.5 16.4 24.6

108 16.5 13.8 20.8





#40

Hexachlorocyclopentadiene

Concen: 16.072 ng/ul

RT: 12.699 min Scan# 1634

Delta R.T. -0.006 min

Lab File: BP017358.D

Acq: 26 Sep 2023 13:03

Instrument :

BNA_P

ClientSampleId :

SSTD020752





Tgt Ion:237 Resp: 129998

Ion Ratio Lower Upper

237 100

235 62.7 52.8 79.2

272 12.6 9.8 14.6



#41

2,4,6-Trichlorophenol

Concen: 22.231 ng/ul

RT: 13.016 min Scan# 1688

Delta R.T. 0.000 min

Lab File: BP017358.D

Acq: 26 Sep 2023 13:03




Tgt Ion:196 Resp: 192734


Ion Ratio Lower Upper

196 100

198 94.2 76.2 114.2

200 29.9 25.1 37.7









#42
2,4,5-Trichlorophenol
Concen: 22.364 ng/ul
RT: 13.087 min Scan# 1700
Delta R.T. 0.000 min
Lab File: BP017358.D
Acq: 26 Sep 2023 13:03




Instrument :
BNA_P
ClientSampleId :
SSTD020752


Tgt Ion:196 Resp: 203133
Ion Ratio Lower Upper
196 100
198 97.9 77.4 116.0
200 30.1 24.3 36.5



#43
1,1'-Biphenyl
Concen: 20.900 ng/ul
RT: 13.422 min Scan# 1757
Delta R.T. 0.006 min
Lab File: BP017358.D
Acq: 26 Sep 2023 13:03

Tgt Ion:154 Resp: 706197
Ion Ratio Lower Upper
154 100
153 40.1 31.5 47.3
76 12.4 11.0 16.6









#44
2-Chloronaphthalene
Concen: 21.072 ng/ul
RT: 13.469 min Scan# 1765
Delta R.T. 0.000 min
Lab File: BP017358.D
Acq: 26 Sep 2023 13:03




Instrument :
BNA_P
ClientSampleId :
SSTD020752


Tgt Ion:162 Resp: 564488
Ion Ratio Lower Upper
162 100
164 33.0 26.8 40.2
127 38.5 30.7 46.1



#45
2-Nitroaniline
Concen: 22.766 ng/ul
RT: 13.704 min Scan# 1805
Delta R.T. 0.000 min
Lab File: BP017358.D
Acq: 26 Sep 2023 13:03

Tgt Ion: 65 Resp: 149820
Ion Ratio Lower Upper
65 100
92 67.7 57.7 86.5
138 101.5 83.5 125.3









#47
Dimethylphthalate
Concen: 21.404 ng/ul
RT: 14.051 min Scan# 1864
Delta R.T. 0.000 min
Lab File: BP017358.D
Acq: 26 Sep 2023 13:03




Instrument :
BNA_P
ClientSampleId :
SSTD020752


Tgt Ion:163 Resp: 687879
Ion Ratio Lower Upper
163 100
194 3.9 3.2 4.8
164 10.3 7.5 11.3



#48
2,6-Dinitrotoluene
Concen: 23.701 ng/ul
RT: 14.198 min Scan# 1889
Delta R.T. 0.000 min
Lab File: BP017358.D
Acq: 26 Sep 2023 13:03

Tgt Ion:165 Resp: 132300
Ion Ratio Lower Upper
165 100
89 62.5 51.7 77.5
121 21.9 19.7 29.5





#50

Acenaphthylene

Concen: 21.279 ng/ul

RT: 14.316 min Scan# 1909

Delta R.T. 0.000 min

Lab File: BP017358.D

Acq: 26 Sep 2023 13:03

Instrument :

BNA_P

ClientSampleId :

SSTD020752





Tgt Ion:152 Resp: 913018

Ion Ratio Lower Upper

152 100

151 19.7 16.0 24.0

153 13.1 10.4 15.6



#51

3-Nitroaniline

Concen: 22.619 ng/ul

RT: 14.540 min Scan# 1947

Delta R.T. 0.000 min

Lab File: BP017358.D

Acq: 26 Sep 2023 13:03




Tgt Ion:138 Resp: 135877


Ion Ratio Lower Upper

138 100

108 11.1 8.2 12.4

92 114.9 97.4 146.0









#52
Acenaphthene
Concen: 21.122 ng/ul
RT: 14.651 min Scan# 1966
Delta R.T. 0.000 min
Lab File: BP017358.D
Acq: 26 Sep 2023 13:03




Instrument :
BNA_P
ClientSampleId :
SSTD020752


Tgt Ion:153 Resp: 598535
Ion Ratio Lower Upper
153 100
152 47.1 38.0 57.0
154 87.1 70.1 105.1



#53
2,4-Dinitrophenol
Concen: 14.705 ng/ul
RT: 14.740 min Scan# 1981
Delta R.T. 0.000 min
Lab File: BP017358.D
Acq: 26 Sep 2023 13:03

Tgt Ion:184 Resp: 51348
Ion Ratio Lower Upper
184 100
63 68.8 53.2 79.8
154 67.8 48.9 73.3









#55
4-Nitrophenol
Concen: 20.770 ng/ul
RT: 14.834 min Scan# 1996
Delta R.T. 0.006 min
Lab File: BP017358.D
Acq: 26 Sep 2023 13:03




Instrument :
BNA_P
ClientSampleId :
SSTD020752


Tgt Ion:109 Resp: 90189
Ion Ratio Lower Upper
109 100
139 126.1 105.7 158.5
65 118.8 105.4 158.0



#56
Dibenzofuran
Concen: 21.522 ng/ul
RT: 14.993 min Scan# 2024
Delta R.T. 0.006 min
Lab File: BP017358.D
Acq: 26 Sep 2023 13:03

Tgt Ion:168 Resp: 835180
Ion Ratio Lower Upper
168 100
139 36.8 30.6 45.8





#57

2,4-Dinitrotoluene

Concen: 23.916 ng/ul

RT: 14.981 min Scan# 2022

Delta R.T. 0.000 min

Lab File: BP017358.D

Acq: 26 Sep 2023 13:03

Instrument :

BNA_P

ClientSampleId :

SSTD020752





Tgt Ion:165 Resp: 196648

Ion Ratio Lower Upper

165 100

63 58.4 45.6 68.4

182 2.8 2.0 3.0



#58

2,3,4,6-Tetrachlorophenol

Concen: 22.804 ng/ul

RT: 15.210 min Scan# 2061

Delta R.T. 0.006 min

Lab File: BP017358.D

Acq: 26 Sep 2023 13:03

Tgt Ion:232 Resp: 175749




Ion Ratio Lower Upper


232 100

131 42.7 38.4 57.6

130 2.5 2.0 3.0

166 28.8 24.6 36.8









#59
Diethylphthalate
Concen: 21.846 ng/ul
RT: 15.410 min Scan# 2095
Delta R.T. 0.006 min
Lab File: BP017358.D
Acq: 26 Sep 2023 13:03




Instrument :
BNA_P
ClientSampleId :
SSTD020752


Tgt Ion:149 Resp: 678748
Ion Ratio Lower Upper
149 100
177 22.7 17.0 25.6
150 12.2 9.8 14.6



#61
Fluorene
Concen: 21.611 ng/ul
RT: 15.645 min Scan# 2135
Delta R.T. 0.006 min
Lab File: BP017358.D
Acq: 26 Sep 2023 13:03

Tgt Ion:166 Resp: 681373
Ion Ratio Lower Upper
166 100
165 99.8 77.8 116.6
167 13.2 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 22.008 ng/ul
RT: 15.634 min Scan# 2133
Delta R.T. 0.000 min
Lab File: BP017358.D
Acq: 26 Sep 2023 13:03




Instrument :
BNA_P
ClientSampleId :
SSTD020752


Tgt Ion:204 Resp: 351672
Ion Ratio Lower Upper
204 100
206 32.7 26.2 39.4
141 56.8 45.9 68.9



#63
4-Nitroaniline
Concen: 23.903 ng/ul
RT: 15.710 min Scan# 2146
Delta R.T. 0.006 min
Lab File: BP017358.D
Acq: 26 Sep 2023 13:03

Tgt Ion:138 Resp: 128152
Ion Ratio Lower Upper
138 100
92 52.5 42.4 63.6
108 94.6 73.0 109.6









#66
4,6-Dinitro-2-methylphenol
Concen: 18.325 ng/ul
RT: 15.740 min Scan# 2150
Delta R.T. 0.006 min
Lab File: BP017358.D
Acq: 26 Sep 2023 13:03




Instrument :
BNA_P
ClientSampleId :
SSTD020752


Tgt Ion:198 Resp: 110359
Ion Ratio Lower Upper
198 100
182 4.8 4.3 6.5
77 23.6 20.4 30.6



#67
N-Nitrosodiphenylamine
Concen: 20.893 ng/ul
RT: 15.863 min Scan# 2172
Delta R.T. 0.006 min
Lab File: BP017358.D
Acq: 26 Sep 2023 13:03

Tgt Ion:169 Resp: 569205
Ion Ratio Lower Upper
169 100
168 65.5 53.5 80.3
167 35.6 30.0 45.0









#68
4-Bromophenyl-phenylether
Concen: 21.634 ng/ul
RT: 16.539 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP017358.D
Acq: 26 Sep 2023 13:03




Instrument :
BNA_P
ClientSampleId :
SSTD020752


Tgt Ion:248 Resp: 215269
Ion Ratio Lower Upper
248 100
250 95.1 78.6 117.8
141 71.5 56.3 84.5



#69
Hexachlorobenzene
Concen: 21.835 ng/ul
RT: 16.622 min Scan# 2301
Delta R.T. 0.006 min
Lab File: BP017358.D
Acq: 26 Sep 2023 13:03

Tgt Ion:284 Resp: 252113
Ion Ratio Lower Upper
284 100
142 32.6 29.5 44.3
249 30.9 25.7 38.5





#70

Atrazine

Concen: 21.053 ng/ul

RT: 16.822 min Scan# 21

Delta R.T. 0.006 min

Lab File: BP017358.D

Acq: 26 Sep 2023 13:03

Instrument :

BNA_P

ClientSampleId :

SSTD020752





Tgt Ion:200 Resp: 205488

Ion Ratio Lower Upper

200 100

173 25.5 20.3 30.5

215 46.9 38.2 57.4



#71

Pentachlorophenol

Concen: 20.438 ng/ul

RT: 16.987 min Scan# 2363

Delta R.T. 0.000 min

Lab File: BP017358.D

Acq: 26 Sep 2023 13:03




Tgt Ion:266 Resp: 145776


Ion Ratio Lower Upper

266 100

264 62.3 48.9 73.3

268 62.6 51.6 77.4





#72

Phenanthrene

Concen: 21.018 ng/ul

RT: 17.404 min Scan# 2434

Delta R.T. 0.000 min

Lab File: BP017358.D

Acq: 26 Sep 2023 13:03

Instrument :

BNA_P

ClientSampleId :

SSTD020752





Tgt Ion:178 Resp: 1090332

Ion Ratio Lower Upper

178 100

179 15.7 12.7 19.1

176 19.1 15.8 23.6



#74

Anthracene

Concen: 21.226 ng/ul

RT: 17.498 min Scan# 2450

Delta R.T. 0.006 min

Lab File: BP017358.D

Acq: 26 Sep 2023 13:03




Tgt Ion:178 Resp: 1115975


Ion Ratio Lower Upper

178 100

179 15.4 12.5 18.7

176 18.8 15.1 22.7





#75

1,2,3,4-Tetrachlorobenzene

Concen: 20.242 ng/uL

RT: 13.375 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP017358.D

Acq: 26 Sep 2023 13:03

Instrument :

BNA_P

ClientSampleId :

SSTD020752





Tgt Ion:216 Resp: 307700

Ion Ratio Lower Upper

216 100

214 75.9 61.2 91.8

218 48.6 36.8 55.2



#76

Pentachlorobenzene

Concen: 20.972 ng/uL

RT: 14.881 min Scan# 2005

Delta R.T. 0.000 min

Lab File: BP017358.D

Acq: 26 Sep 2023 13:03




Tgt Ion:250 Resp: 301150


Ion Ratio Lower Upper

250 100

248 64.4 50.1 75.1

252 64.1 52.2 78.2









#77
Carbazole
Concen: 21.007 ng/ul
RT: 17.786 min Scan# 2499
Delta R.T. 0.006 min
Lab File: BP017358.D
Acq: 26 Sep 2023 13:03




Instrument :
BNA_P
ClientSampleId :
SSTD020752


Tgt Ion:167 Resp: 982401
Ion Ratio Lower Upper
167 100
166 21.5 16.5 24.7
139 13.0 10.5 15.7



#78
Di-n-butylphthalate
Concen: 22.726 ng/ul
RT: 18.298 min Scan# 2586
Delta R.T. 0.006 min
Lab File: BP017358.D
Acq: 26 Sep 2023 13:03

Tgt Ion:149 Resp: 1168670
Ion Ratio Lower Upper
149 100
150 8.9 7.5 11.3
104 5.4 4.5 6.7





#80

Fluoranthene

Concen: 21.356 ng/ul

RT: 19.375 min Scan# 21

Delta R.T. 0.006 min

Lab File: BP017358.D

Acq: 26 Sep 2023 13:03

Instrument :

BNA_P

ClientSampleId :

SSTD020752





Tgt Ion:202 Resp: 1301651

Ion Ratio Lower Upper

202 100

101 10.7 8.4 12.6

100 8.4 6.6 10.0



#82

Pyrene

Concen: 21.204 ng/ul

RT: 19.733 min Scan# 2830

Delta R.T. 0.006 min

Lab File: BP017358.D

Acq: 26 Sep 2023 13:03




Tgt Ion:202 Resp: 1373832


Ion Ratio Lower Upper

202 100

101 11.9 9.8 14.8

100 10.1 8.1 12.1





#83

Butylbenzylphthalate

Concen: 23.830 ng/ul

RT: 20.598 min Scan# 2975

Delta R.T. 0.012 min

Lab File: BP017358.D

Acq: 26 Sep 2023 13:03

Instrument :

BNA_P

ClientSampleId :

SSTD020752





Tgt Ion:149 Resp: 547448

Ion Ratio Lower Upper

149 100

91 69.3 59.8 89.6

206 21.6 15.7 23.5



#84

3,3'-Dichlorobenzidine

Concen: 20.923 ng/ul

RT: 21.392 min Scan# 3112

Delta R.T. 0.006 min

Lab File: BP017358.D

Acq: 26 Sep 2023 13:03




Tgt Ion:252 Resp: 439326


Ion Ratio Lower Upper

252 100

254 64.7 51.9 77.9

126 11.9 9.1 13.7





#85

Benzo(a)anthracene

Concen: 21.377 ng/ul

RT: 21.451 min Scan# 3121

Delta R.T. 0.006 min

Lab File: BP017358.D

Acq: 26 Sep 2023 13:03

Instrument :

BNA_P

ClientSampleId :

SSTD020752





Tgt Ion:228 Resp: 1394314

Ion Ratio Lower Upper

228 100

229 19.8 15.8 23.6

226 27.6 21.8 32.8



#86

Bis(2-ethylhexyl)phthalate

Concen: 24.685 ng/ul

RT: 21.327 min Scan# 3101

Delta R.T. 0.006 min

Lab File: BP017358.D

Acq: 26 Sep 2023 13:03




Tgt Ion:149 Resp: 818974


Ion Ratio Lower Upper

149 100

167 26.8 21.0 31.6

279 3.8 3.0 4.4





#87

Chrysene

Concen: 21.194 ng/ul

RT: 21.510 min Scan# 3131

Delta R.T. 0.006 min

Lab File: BP017358.D

Acq: 26 Sep 2023 13:03

Instrument :

BNA_P

ClientSampleId :

SSTD020752





Tgt Ion:228 Resp: 1300978

Ion Ratio Lower Upper

228 100

226 29.9 24.1 36.1

229 19.0 15.0 22.6



#89

Di-n-octyl phthalate

Concen: 24.311 ng/ul




RT: 22.292 min Scan# 3265


Delta R.T. 0.012 min

Lab File: BP017358.D

Acq: 26 Sep 2023 13:03

Tgt Ion:149 Resp: 1420717





#90

Benzo(b)fluoranthene

Concen: 21.131 ng/ul

RT: 23.204 min Scan# 3418

Delta R.T. 0.012 min

Lab File: BP017358.D

Acq: 26 Sep 2023 13:03

Instrument :

BNA_P

ClientSampleId :

SSTD020752





Tgt Ion:252 Resp: 1323621

Ion Ratio Lower Upper

252 100

253 22.1 17.4 26.0

125 9.5 7.8 11.8



#91

Benzo(k)fluoranthene

Concen: 21.289 ng/ul

RT: 23.257 min Scan# 3429

Delta R.T. 0.012 min

Lab File: BP017358.D

Acq: 26 Sep 2023 13:03




Tgt Ion:252 Resp: 1345360


Ion Ratio Lower Upper

252 100

253 21.4 17.9 26.9

125 9.3 7.8 11.6





#93

Benzo(a)pyrene

Concen: 21.007 ng/ul

RT: 23.874 min Scan# 3

Delta R.T. 0.012 min

Lab File: BP017358.D

Acq: 26 Sep 2023 13:03

Instrument :

BNA_P

ClientSampleId :

SSTD020752





Tgt Ion:252 Resp: 1258990

Ion Ratio Lower Upper

252 100

253 22.6 17.4 26.0

125 10.6 8.9 13.3



#94

Indeno(1,2,3-cd)pyrene

Concen: 20.312 ng/ul

RT: 26.651 min Scan# 4006

Delta R.T. 0.018 min

Lab File: BP017358.D

Acq: 26 Sep 2023 13:03




Tgt Ion:276 Resp: 1525860


Ion Ratio Lower Upper

276 100

138 20.9 17.9 26.9

277 24.6 19.5 29.3





#95

Dibenzo(a,h)anthracene

Concen: 20.230 ng/ul

RT: 26.674 min Scan# 4007

Delta R.T. 0.018 min

Lab File: BP017358.D

Acq: 26 Sep 2023 13:03

Instrument :

BNA_P

ClientSampleId :

SSTD020752





Tgt Ion:278 Resp: 1265749

Ion Ratio Lower Upper

278 100

139 15.1 13.4 20.0

279 24.7 19.5 29.3



#96

Benzo(g,h,i)perylene

Concen: 19.954 ng/ul

RT: 27.486 min Scan# 4148

Delta R.T. 0.024 min

Lab File: BP017358.D

Acq: 26 Sep 2023 13:03

Tgt Ion:276 Resp: 1208772

Ion Ratio Lower Upper

276 100

138 20.2 17.3 25.9

277 23.7 18.8 28.2

