

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092523\
 Data File : BP017369.D
 Acq On : 26 Sep 2023 19:18
 Operator : MA/JU
 Sample : 04450-13
 Misc :
 ALS Vial : 61 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 BBJ46

Quant Time: Sep 27 00:30:00 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP092023.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Sep 26 04:52:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.922	152	156779	20.000	ng/ul	0.00
20) Naphthalene-d8	10.746	136	683529	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.581	164	414449	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.357	188	841915	20.000	ng/ul	0.00
79) Chrysene-d12	21.457	240	773571	20.000	ng/ul	0.00
88) Perylene-d12	23.968	264	868700	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.293	96	18082	4.737	ng/uL	0.00
4) Pyridine-d5	3.734	84	153452	14.381	ng/ul	0.00
7) Phenol-d5	7.081	99	110896	8.612	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.263	67	325681	41.911	ng/ul	0.00
11) 2-Chlorophenol-d4	7.446	132	321882	31.518	ng/ul	0.00
15) 4-Methylphenol-d8	8.634	113	193678	19.347	ng/ul	0.00
21) Nitrobenzene-d5	9.122	128	208454	42.184	ng/ul	0.00
24) 2-Nitrophenol-d4	9.834	143	213753	39.838	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.357	165	364406	35.948	ng/ul	0.00
31) 4-Chloroaniline-d4	10.910	131	378382	25.768	ng/ul	0.00
46) Dimethylphthalate-d6	13.998	166	1332051	44.866	ng/ul	0.00
49) Acenaphthylene-d8	14.281	160	1419872	41.581	ng/ul	0.00
54) 4-Nitrophenol-d4	14.822	143	23566	4.556	ng/ul	0.00
60) Fluorene-d10	15.581	176	1115040	43.624	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.716	200	134632	28.372	ng/ul	0.00
73) Anthracene-d10	17.457	188	1710708	45.815	ng/ul	0.00
81) Pyrene-d10	19.698	212	2032041	48.575	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.804	264	1924273	46.007	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092523\
Data File : BP017369.D
Acq On : 26 Sep 2023 19:18
Operator : MA/JU
Sample : 04450-13
Misc :
ALS Vial : 61 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
BBJ46

Quant Time: Sep 27 00:30:00 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP092023.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Sep 26 04:52:29 2023
Response via : Initial Calibration

