

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092721\
 Data File : BP007207.D
 Acq On : 27 Sep 2021 16:07
 Operator : CG/JU
 Sample : M3944-01
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 BOILER-PIT-3-5

Quant Time: Sep 27 16:51:49 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP092121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 23 11:17:25 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.881	152	69994	20.000	ng	-0.01
21) Naphthalene-d8	10.681	136	358389	20.000	ng	#-0.01
39) Acenaphthene-d10	14.510	164	330625	20.000	ng	-0.02
64) Phenanthrene-d10	17.263	188	909100	20.000	ng	#-0.01
76) Chrysene-d12	21.369	240	1310425	20.000	ng	# 0.01
86) Perylene-d12	23.821	264	1410151	20.000	ng	0.05
System Monitoring Compounds						
5) 2-Fluorophenol	5.458	112	469702	112.332	ng	0.00
7) Phenol-d6	7.052	99	656789	114.925	ng	-0.01
23) Nitrobenzene-d5	9.040	82	374943	60.926	ng	-0.02
42) 2,4,6-Tribromophenol	16.004	330	637745	148.784	ng	-0.01
45) 2-Fluorobiphenyl	13.140	172	1267769	54.933	ng	-0.01
79) Terphenyl-d14	19.833	244	3818452	55.838	ng	0.00
Target Compounds						
84) Bis(2-ethylhexyl)phtha...	21.274	149	299618	6.053	ng	# 75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092721\
Data File : BP007207.D
Acq On : 27 Sep 2021 16:07
Operator : CG/JU
Sample : M3944-01
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
BOILER-PIT-3-5

Quant Time: Sep 27 16:51:49 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP092121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 23 11:17:25 2021
Response via : Initial Calibration

