

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092823\
 Data File : BP017411.D
 Acq On : 28 Sep 2023 13:34
 Operator : MA/JU
 Sample : PB155751BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK751

Quant Time: Sep 28 22:29:58 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP092023.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 28 15:31:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.963	152	156464	20.000	ng/ul	0.00
20) Naphthalene-d8	10.798	136	736889	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.651	164	457936	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.439	188	925207	20.000	ng/ul	0.00
79) Chrysene-d12	21.574	240	798474	20.000	ng/ul	0.00
88) Perylene-d12	24.157	264	861179	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.299	96	26382	6.926	ng/uL	0.00
4) Pyridine-d5	3.734	84	377518	35.450	ng/ul	0.00
7) Phenol-d5	7.110	99	462621	36.001	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.299	67	285614	36.829	ng/ul	0.00
11) 2-Chlorophenol-d4	7.481	132	374386	36.733	ng/ul	0.00
15) 4-Methylphenol-d8	8.681	113	376545	37.690	ng/ul	0.00
21) Nitrobenzene-d5	9.169	128	182091	34.181	ng/ul	0.00
24) 2-Nitrophenol-d4	9.887	143	196742	34.012	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.410	165	361760	33.103	ng/ul	0.00
31) 4-Chloroaniline-d4	10.969	131	528191	33.365	ng/ul	0.00
46) Dimethylphthalate-d6	14.069	166	1157310	35.278	ng/ul	0.00
49) Acenaphthylene-d8	14.351	160	1245234	33.004	ng/ul	0.00
54) 4-Nitrophenol-d4	14.886	143	111629	19.532	ng/ul	0.00
60) Fluorene-d10	15.657	176	943981	33.425	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.792	200	99750	19.129	ng/ul	0.00
73) Anthracene-d10	17.545	188	1437243	35.026	ng/ul	0.00
81) Pyrene-d10	19.792	212	1641570	38.017	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.980	264	1497580	36.118	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092823\
Data File : BP017411.D
Acq On : 28 Sep 2023 13:34
Operator : MA/JU
Sample : PB155751BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK751

Quant Time: Sep 28 22:29:58 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP092023.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Sep 28 15:31:53 2023
Response via : Initial Calibration

