

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092923\
 Data File : BP017454.D
 Acq On : 29 Sep 2023 20:09
 Operator : MA/JU
 Sample : 04497-14
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 BBJ73

Quant Time: Sep 29 22:37:14 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP092023.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 29 22:31:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.952	152	256638	20.000	ng/ul	-0.01
20) Naphthalene-d8	10.793	136	1046546	20.000	ng/ul	-0.01
38) Acenaphthene-d10	14.645	164	645955	20.000	ng/ul	-0.01
64) Phenanthrene-d10	17.433	188	1459979	20.000	ng/ul	-0.01
79) Chrysene-d12	21.563	240	1481947	20.000	ng/ul	-0.01
88) Perylene-d12	24.139	264	1740152	20.000	ng/ul	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.293	96	13365	2.139	ng/uL	0.00
4) Pyridine-d5	3.734	84	184518	10.564	ng/ul	0.00
7) Phenol-d5	7.105	99	114760	5.445	ng/ul	-0.01
9) Bis-(2-Chloroethyl)eth...	7.293	67	355092	27.916	ng/ul	-0.01
11) 2-Chlorophenol-d4	7.469	132	354636	21.214	ng/ul	-0.01
15) 4-Methylphenol-d8	8.669	113	183391	11.191	ng/ul	-0.02
21) Nitrobenzene-d5	9.157	128	246482	32.578	ng/ul	-0.01
24) 2-Nitrophenol-d4	9.875	143	268994	32.744	ng/ul	-0.01
28) 2,4-Dichlorophenol-d3	10.404	165	404893	26.087	ng/ul	-0.01
31) 4-Chloroaniline-d4	10.957	131	291848	12.981	ng/ul	-0.01
46) Dimethylphthalate-d6	14.063	166	1436707	31.048	ng/ul	0.00
49) Acenaphthylene-d8	14.345	160	1687354	31.705	ng/ul	-0.01
54) 4-Nitrophenol-d4	14.887	143	51588	6.399	ng/ul	0.00
60) Fluorene-d10	15.651	176	1309681	32.875	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.792	200	203765	24.763	ng/ul	-0.01
73) Anthracene-d10	17.533	188	2039649	31.500	ng/ul	-0.01
81) Pyrene-d10	19.786	212	2506051	31.271	ng/ul	-0.01
92) Benzo(a)pyrene-d12	23.974	264	2707439	32.315	ng/ul	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092923\
Data File : BP017454.D
Acq On : 29 Sep 2023 20:09
Operator : MA/JU
Sample : 04497-14
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
BBJ73

Quant Time: Sep 29 22:37:14 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP092023.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Sep 29 22:31:03 2023
Response via : Initial Calibration

