

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092923\
 Data File : BP017475.D
 Acq On : 30 Sep 2023 09:10
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020763

Quant Time: Oct 02 00:57:39 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP092023.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Sep 30 04:02:20 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.946	152	162155	20.000	ng/u1	0.00
20) Naphthalene-d8	10.787	136	648399	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.640	164	398690	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.428	188	895153	20.000	ng/u1	0.00
79) Chrysene-d12	21.557	240	870225	20.000	ng/u1	0.00
88) Perylene-d12	24.127	264	958427	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.293	96	33905	8.588	ng/uL	0.00
4) Pyridine-d5	3.728	84	240033	21.749	ng/u1	0.00
7) Phenol-d5	7.105	99	298054	22.380	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.287	67	178998	22.271	ng/u1	0.00
11) 2-Chlorophenol-d4	7.463	132	243607	23.063	ng/u1	0.00
15) 4-Methylphenol-d8	8.669	113	239566	23.138	ng/u1	0.00
21) Nitrobenzene-d5	9.157	128	117398	25.045	ng/u1	0.00
24) 2-Nitrophenol-d4	9.875	143	130723	25.683	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.399	165	245247	25.504	ng/u1	0.00
31) 4-Chloroaniline-d4	10.957	131	338484	24.299	ng/u1	0.00
46) Dimethylphthalate-d6	14.057	166	689746	24.150	ng/u1	0.00
49) Acenaphthylene-d8	14.340	160	806019	24.537	ng/u1	0.00
54) 4-Nitrophenol-d4	14.875	143	105767	21.256	ng/u1	0.00
60) Fluorene-d10	15.645	176	607444	24.705	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.787	200	103082	20.431	ng/u1	0.00
73) Anthracene-d10	17.528	188	977480	24.621	ng/u1	0.00
81) Pyrene-d10	19.780	212	1158423	24.616	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.957	264	1102367	23.889	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.334	88	35654	8.819	ng/uL	91
5) Pyridine	3.752	79	250962	21.903	ng/u1	99
6) Benzaldehyde	7.105	77	168108	24.735	ng/u1	98
8) Phenol	7.128	94	303255	22.640	ng/u1	96
10) Bis(2-Chloroethyl)ether	7.387	93	240085	22.173	ng/u1	97
12) 2-Chlorophenol	7.499	128	247774	23.099	ng/u1	98
13) 2-Methylphenol	8.399	108	227146	22.917	ng/u1	96
14) 2,2'-oxybis(1-Chloropr...	8.463	45	191312	13.970	ng/u1	99
16) Acetophenone	8.805	105	368253	23.016	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.775	70	183608	22.994	ng/u1	99
18) 4-Methylphenol	8.734	108	242566	22.840	ng/u1	99
19) Hexachloroethane	9.016	117	104491	23.466	ng/u1	95
22) Nitrobenzene	9.199	77	289153	24.310	ng/u1	97
23) Isophorone	9.710	82	504671	23.678	ng/u1	98
25) 2-Nitrophenol	9.904	139	139062	25.213	ng/u1	99
26) 2,4-Dimethylphenol	9.946	107	268242	24.410	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.204	93	314328	22.845	ng/u1	98
29) 2,4-Dichlorophenol	10.428	162	242054	25.530	ng/u1	97
30) Naphthalene	10.840	128	792312	23.468	ng/u1	99
32) 4-Chloroaniline	10.981	127	327736	24.552	ng/u1	100
33) Hexachlorobutadiene	11.063	225	166816	25.281	ng/u1	99
34) Caprolactam	11.804	113	68834	25.127	ng/u1	96
35) 4-Chloro-3-methylphenol	12.087	107	246050	25.819	ng/u1	99
36) 2-Methylnaphthalene	12.451	142	529693	24.038	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092923\
 Data File : BP017475.D
 Acq On : 30 Sep 2023 09:10
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020763

Quant Time: Oct 02 00:57:39 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP092023.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Sep 30 04:02:20 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.675	142	547359	24.300	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.804	216	307321	24.019	ng/ul	98
40) Hexachlorocyclopentadiene	12.745	237	102412	14.143	ng/ul	97
41) 2,4,6-Trichlorophenol	13.063	196	197768	25.482	ng/ul	100
42) 2,4,5-Trichlorophenol	13.134	196	209079	25.713	ng/ul	98
43) 1,1'-Biphenyl	13.469	154	707608	23.393	ng/ul	99
44) 2-Chloronaphthalene	13.516	162	570200	23.777	ng/ul	99
45) 2-Nitroaniline	13.757	65	157814	26.788	ng/ul	98
47) Dimethylphthalate	14.104	163	685217	23.817	ng/ul	98
48) 2,6-Dinitrotoluene	14.251	165	138168	27.650	ng/ul	97
50) Acenaphthylene	14.369	152	930143	24.215	ng/ul	99
51) 3-Nitroaniline	14.592	138	135899	25.271	ng/ul	100
52) Acenaphthene	14.704	153	602235	23.740	ng/ul	99
53) 2,4-Dinitrophenol	14.804	184	47660	15.247	ng/ul	95
55) 4-Nitrophenol	14.887	109	90667	23.324	ng/ul	95
56) Dibenzofuran	15.045	168	851031	24.497	ng/ul	99
57) 2,4-Dinitrotoluene	15.045	165	205401	27.905	ng/ul	95
58) 2,3,4,6-Tetrachlorophenol	15.263	232	185399	26.872	ng/ul	96
59) Diethylphthalate	15.463	149	690314	24.819	ng/ul	99
61) Fluorene	15.698	166	693428	24.568	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.692	204	359594	25.138	ng/ul	98
63) 4-Nitroaniline	15.769	138	128401	26.753	ng/ul	100
66) 4,6-Dinitro-2-methylph...	15.798	198	111723	20.527	ng/ul	96
67) N-Nitrosodiphenylamine	15.922	169	578880	23.510	ng/ul	99
68) 4-Bromophenyl-phenylether	16.598	248	221969	24.683	ng/ul	98
69) Hexachlorobenzene	16.681	284	260667	24.979	ng/ul	98
70) Atrazine	16.881	200	213928	24.251	ng/ul	99
71) Pentachlorophenol	17.051	266	151068	23.436	ng/ul	99
72) Phenanthrene	17.475	178	1114117	23.763	ng/ul	99
74) Anthracene	17.563	178	1149169	24.184	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.422	216	315639	22.975	ng/uL	97
76) Pentachlorobenzene	14.934	250	311147	23.975	ng/uL	98
77) Carbazole	17.851	167	1002114	23.710	ng/ul	99
78) Di-n-butylphthalate	18.363	149	1197441	25.765	ng/ul	99
80) Fluoranthene	19.451	202	1345114	24.509	ng/ul	98
82) Pyrene	19.810	202	1411341	24.190	ng/ul	99
83) Butylbenzylphthalate	20.674	149	550430	26.608	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.480	252	421781	22.308	ng/ul	99
85) Benzo(a)anthracene	21.539	228	1417223	24.130	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.416	149	805065	26.948	ng/ul#	98
87) Chrysene	21.598	228	1313955	23.772	ng/ul	99
89) Di-n-octyl phthalate	22.398	149	1389555	26.357	ng/ul	100
90) Benzo(b)fluoranthene	23.327	252	1346989	23.837	ng/ul	99
91) Benzo(k)fluoranthene	23.380	252	1371647	24.060	ng/ul	98
93) Benzo(a)pyrene	24.010	252	1282395	23.719	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.857	276	1529566	22.571	ng/ul	99
95) Dibenzo(a,h)anthracene	26.880	278	1258685	22.300	ng/ul	97
96) Benzo(g,h,i)perylene	27.709	276	1228800	22.485	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP092923\
Data File : BP017475.D
Acq On : 30 Sep 2023 09:10
Operator : MA/JU
Sample : SSTDCCC020
Misc :
ALS Vial : 42 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTD020763

Quant Time: Oct 02 00:57:39 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP092023.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Sep 30 04:02:20 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.819 ng/uL
RT: 3.334 min Scan# 41
Delta R.T. 0.006 min
Lab File: BP017475.D
Acq: 30 Sep 2023 09:10




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion	88	Resp	35654
Ion Ratio	100		
Lower	25.0		
Upper	37.4		
	43	28.5	
	58	61.2	



#5
Pyridine
Concen: 21.903 ng/uL
RT: 3.752 min Scan# 113
Delta R.T. 0.000 min
Lab File: BP017475.D
Acq: 30 Sep 2023 09:10

Tgt Ion	79	Resp	250962
Ion Ratio <td>100</td> <td></td> <td></td>	100		
Lower	51.4		
Upper	77.2		
	52	65.0	
	51	32.8	









#6
Benzaldehyde
Concen: 24.735 ng/ul
RT: 7.105 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP017475.D
Acq: 30 Sep 2023 09:10




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion: 77 Resp: 168108
Ion Ratio Lower Upper
77 100
105 94.7 78.0 117.0
106 92.4 73.2 109.8



#8
Phenol
Concen: 22.640 ng/ul
RT: 7.128 min Scan# 687
Delta R.T. -0.006 min
Lab File: BP017475.D
Acq: 30 Sep 2023 09:10

Tgt Ion: 94 Resp: 303255
Ion Ratio Lower Upper
94 100
65 28.6 21.4 32.2
66 37.5 28.2 42.2









#10
Bis(2-Chloroethyl)ether
Concen: 22.173 ng/ul
RT: 7.387 min Scan# 71
Delta R.T. 0.000 min
Lab File: BP017475.D
Acq: 30 Sep 2023 09:10




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion: 93 Resp: 240085
Ion Ratio Lower Upper
93 100
63 72.0 55.4 83.0
95 32.2 25.1 37.7



#12
2-Chlorophenol
Concen: 23.099 ng/ul
RT: 7.499 min Scan# 750
Delta R.T. -0.006 min
Lab File: BP017475.D
Acq: 30 Sep 2023 09:10

Tgt Ion:128 Resp: 247774
Ion Ratio Lower Upper
128 100
64 45.8 35.8 53.8
130 33.8 25.6 38.4





#13

2-Methylphenol

Concen: 22.917 ng/ul

RT: 8.399 min Scan# 904

Delta R.T. 0.000 min

Lab File: BP017475.D

Acq: 30 Sep 2023 09:10

Instrument :

BNA_P

ClientSampleId :





SSTD020763

Tgt Ion:108 Resp: 227146

Ion Ratio Lower Upper

108 100

107 88.3 73.5 110.3



#14

2,2'-oxybis(1-Chloropropane)

Concen: 13.970 ng/ul

RT: 8.463 min Scan# 914

Delta R.T. -0.006 min

Lab File: BP017475.D

Acq: 30 Sep 2023 09:10




Tgt Ion: 45 Resp: 191312


Ion Ratio Lower Upper

45 100

77 16.3 13.3 19.9

79 12.4 9.4 14.2









#16
Acetophenone
Concen: 23.016 ng/ul
RT: 8.805 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP017475.D
Acq: 30 Sep 2023 09:10




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:105 Resp: 368253
Ion Ratio Lower Upper
105 100
77 80.5 63.6 95.4
51 25.8 19.2 28.8



#17
N-Nitroso-di-n-propylamine
Concen: 22.994 ng/ul
RT: 8.775 min Scan# 967
Delta R.T. 0.000 min
Lab File: BP017475.D
Acq: 30 Sep 2023 09:10

Tgt Ion: 70 Resp: 183608
Ion Ratio Lower Upper
70 100
42 53.7 42.3 63.5
101 8.9 7.5 11.3
130 21.3 17.3 25.9









#18
4-Methylphenol
Concen: 22.840 ng/ul
RT: 8.734 min Scan# 90
Delta R.T. 0.000 min
Lab File: BP017475.D
Acq: 30 Sep 2023 09:10




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:108 Resp: 242566
Ion Ratio Lower Upper
108 100
107 113.3 90.0 135.0



#19
Hexachloroethane
Concen: 23.466 ng/ul
RT: 9.016 min Scan# 1008
Delta R.T. 0.000 min
Lab File: BP017475.D
Acq: 30 Sep 2023 09:10

Tgt Ion:117 Resp: 104491
Ion Ratio Lower Upper
117 100
201 108.4 81.4 122.0
199 67.4 52.4 78.6









#22
Nitrobenzene
Concen: 24.310 ng/ul
RT: 9.199 min Scan# 1039
Delta R.T. 0.000 min
Lab File: BP017475.D
Acq: 30 Sep 2023 09:10




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion: 77 Resp: 289153
Ion Ratio Lower Upper
77 100
123 45.0 37.4 56.0
65 13.7 9.8 14.6



#23
Isophorone
Concen: 23.678 ng/ul
RT: 9.710 min Scan# 1126
Delta R.T. -0.006 min
Lab File: BP017475.D
Acq: 30 Sep 2023 09:10

Tgt Ion: 82 Resp: 504671
Ion Ratio Lower Upper
82 100
95 7.2 5.5 8.3
138 16.5 14.2 21.4









#25
2-Nitrophenol
Concen: 25.213 ng/ul
RT: 9.904 min Scan# 1159
Delta R.T. -0.006 min
Lab File: BP017475.D
Acq: 30 Sep 2023 09:10




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:139 Resp: 139062
Ion Ratio Lower Upper
139 100
65 50.7 40.6 60.8
109 31.0 24.2 36.4



#26
2,4-Dimethylphenol
Concen: 24.410 ng/ul
RT: 9.946 min Scan# 1166
Delta R.T. -0.006 min
Lab File: BP017475.D
Acq: 30 Sep 2023 09:10

Tgt Ion:107 Resp: 268242
Ion Ratio Lower Upper
107 100
121 49.8 41.1 61.7
122 83.9 69.4 104.0









#27
 Bis(2-Chloroethoxy)methane
 Concen: 22.845 ng/ul
 RT: 10.204 min Scan# 1210
 Delta R.T. 0.000 min
 Lab File: BP017475.D
 Acq: 30 Sep 2023 09:10




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020763


Tgt Ion: 93 Resp: 314328
 Ion Ratio Lower Upper
 93 100
 95 32.1 24.9 37.3
 123 11.3 9.4 14.2



#29
 2,4-Dichlorophenol
 Concen: 25.530 ng/ul
 RT: 10.428 min Scan# 1248
 Delta R.T. -0.006 min
 Lab File: BP017475.D
 Acq: 30 Sep 2023 09:10

Tgt Ion: 162 Resp: 242054
 Ion Ratio Lower Upper
 162 100
 164 63.5 52.6 78.8
 98 34.8 27.0 40.4









#30
Naphthalene
Concen: 23.468 ng/ul
RT: 10.840 min Scan# 1319
Delta R.T. 0.000 min
Lab File: BP017475.D
Acq: 30 Sep 2023 09:10




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:128 Resp: 792312
Ion Ratio Lower Upper
128 100
129 10.3 8.6 13.0
127 13.0 10.6 15.8



#32
4-Chloroaniline
Concen: 24.552 ng/ul
RT: 10.981 min Scan# 1342
Delta R.T. 0.000 min
Lab File: BP017475.D
Acq: 30 Sep 2023 09:10

Tgt Ion:127 Resp: 327736
Ion Ratio Lower Upper
127 100
129 31.6 25.1 37.7





#33

Hexachlorobutadiene

Concen: 25.281 ng/ul

RT: 11.063 min Scan# 11

Delta R.T. 0.000 min

Lab File: BP017475.D

Acq: 30 Sep 2023 09:10

Instrument :

BNA_P

ClientSampleId :

SSTD020763





Tgt Ion:225 Resp: 166816

Ion Ratio Lower Upper

225 100

223 60.1 48.2 72.2

227 64.6 50.2 75.4



#34

Caprolactam

Concen: 25.127 ng/ul

RT: 11.804 min Scan# 1482

Delta R.T. -0.006 min

Lab File: BP017475.D

Acq: 30 Sep 2023 09:10




Tgt Ion:113 Resp: 68834


Ion Ratio Lower Upper

113 100

55 155.3 127.4 191.0

56 113.3 94.5 141.7









#35
4-Chloro-3-methylphenol
Concen: 25.819 ng/ul
RT: 12.087 min Scan# 1530
Delta R.T. -0.006 min
Lab File: BP017475.D
Acq: 30 Sep 2023 09:10




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:107 Resp: 246050
Ion Ratio Lower Upper
107 100
144 25.8 20.6 31.0
142 80.0 62.9 94.3



#36
2-Methylnaphthalene
Concen: 24.038 ng/ul
RT: 12.451 min Scan# 1592
Delta R.T. 0.000 min
Lab File: BP017475.D
Acq: 30 Sep 2023 09:10

Tgt Ion:142 Resp: 529693
Ion Ratio Lower Upper
142 100
141 88.8 71.9 107.9





#37

1-Methylnaphthalene

Concen: 24.300 ng/ul

RT: 12.675 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP017475.D

Acq: 30 Sep 2023 09:10

Instrument :

BNA_P

ClientSampleId :

SSTD020763





Tgt Ion:142 Resp: 547359

Ion Ratio Lower Upper

142 100

141 91.2 73.4 110.2

116 3.9 3.1 4.7



#39

1,2,4,5-Tetrachlorobenzene

Concen: 24.019 ng/ul

RT: 12.804 min Scan# 1652

Delta R.T. 0.000 min

Lab File: BP017475.D

Acq: 30 Sep 2023 09:10

Tgt Ion:216 Resp: 307321




Ion Ratio Lower Upper


216 100

214 79.7 62.3 93.5

179 20.3 16.4 24.6

108 16.2 13.8 20.8





#40

Hexachlorocyclopentadiene

Concen: 14.143 ng/ul

RT: 12.745 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP017475.D

Acq: 30 Sep 2023 09:10

Instrument :

BNA_P

ClientSampleId :

SSTD020763





Tgt Ion:237 Resp: 102412

Ion Ratio Lower Upper

237 100

235 64.0 52.8 79.2

272 13.1 9.8 14.6



#41

2,4,6-Trichlorophenol

Concen: 25.482 ng/ul

RT: 13.063 min Scan# 1696

Delta R.T. 0.000 min

Lab File: BP017475.D

Acq: 30 Sep 2023 09:10




Tgt Ion:196 Resp: 197768


Ion Ratio Lower Upper

196 100

198 95.2 76.2 114.2

200 31.0 25.1 37.7





#42

2,4,5-Trichlorophenol

Concen: 25.713 ng/ul

RT: 13.134 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP017475.D

Acq: 30 Sep 2023 09:10

Instrument :

BNA_P

ClientSampleId :

SSTD020763





Tgt Ion:196 Resp: 209079

Ion Ratio Lower Upper

196 100

198 98.9 77.4 116.0

200 30.8 24.3 36.5



#43

1,1'-Biphenyl

Concen: 23.393 ng/ul

RT: 13.469 min Scan# 1765

Delta R.T. 0.000 min

Lab File: BP017475.D

Acq: 30 Sep 2023 09:10




Tgt Ion:154 Resp: 707608


Ion Ratio Lower Upper

154 100

153 39.6 31.5 47.3

76 12.9 11.0 16.6





#44

2-Chloronaphthalene

Concen: 23.777 ng/ul

RT: 13.516 min Scan# 1773

Delta R.T. 0.000 min

Lab File: BP017475.D

Acq: 30 Sep 2023 09:10

Instrument :

BNA_P

ClientSampleId :

SSTD020763





Tgt Ion:162 Resp: 570200

Ion Ratio Lower Upper

162 100

164 32.9 26.8 40.2

127 39.0 30.7 46.1



#45

2-Nitroaniline

Concen: 26.788 ng/ul

RT: 13.757 min Scan# 1814

Delta R.T. 0.000 min

Lab File: BP017475.D

Acq: 30 Sep 2023 09:10




Tgt Ion: 65 Resp: 157814


Ion Ratio Lower Upper

65 100

92 72.4 57.7 86.5

138 108.2 83.5 125.3









#47
Dimethylphthalate
Concen: 23.817 ng/ul
RT: 14.104 min Scan# 1873
Delta R.T. 0.000 min
Lab File: BP017475.D
Acq: 30 Sep 2023 09:10




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:163 Resp: 685217
Ion Ratio Lower Upper
163 100
194 4.0 3.2 4.8
164 10.4 7.5 11.3



#48
2,6-Dinitrotoluene
Concen: 27.650 ng/ul
RT: 14.251 min Scan# 1898
Delta R.T. -0.006 min
Lab File: BP017475.D
Acq: 30 Sep 2023 09:10

Tgt Ion:165 Resp: 138168
Ion Ratio Lower Upper
165 100
89 62.2 51.7 77.5
121 22.3 19.7 29.5





#50

Acenaphthylene

Concen: 24.215 ng/ul

RT: 14.369 min Scan# 1918

Delta R.T. 0.000 min

Lab File: BP017475.D

Acq: 30 Sep 2023 09:10

Instrument :

BNA_P

ClientSampleId :

SSTD020763





Tgt Ion:152 Resp: 930143

Ion Ratio Lower Upper

152 100

151 19.3 16.0 24.0

153 12.9 10.4 15.6



#51

3-Nitroaniline

Concen: 25.271 ng/ul

RT: 14.592 min Scan# 1956

Delta R.T. -0.006 min

Lab File: BP017475.D

Acq: 30 Sep 2023 09:10




Tgt Ion:138 Resp: 135899


Ion Ratio Lower Upper

138 100

108 10.8 8.2 12.4

92 122.1 97.4 146.0









#52
Acenaphthene
Concen: 23.740 ng/ul
RT: 14.704 min Scan# 1975
Delta R.T. -0.006 min
Lab File: BP017475.D
Acq: 30 Sep 2023 09:10




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:153 Resp: 602235
Ion Ratio Lower Upper
153 100
152 48.5 38.0 57.0
154 88.8 70.1 105.1



#53
2,4-Dinitrophenol
Concen: 15.247 ng/ul
RT: 14.804 min Scan# 1992
Delta R.T. 0.000 min
Lab File: BP017475.D
Acq: 30 Sep 2023 09:10

Tgt Ion:184 Resp: 47660
Ion Ratio Lower Upper
184 100
63 67.0 53.2 79.8
154 68.0 48.9 73.3









#55
4-Nitrophenol
Concen: 23.324 ng/ul
RT: 14.887 min Scan# 2006
Delta R.T. -0.006 min
Lab File: BP017475.D
Acq: 30 Sep 2023 09:10




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:109 Resp: 90667
Ion Ratio Lower Upper
109 100
139 123.4 105.7 158.5
65 127.9 105.4 158.0



#56
Dibenzofuran
Concen: 24.497 ng/ul
RT: 15.045 min Scan# 2033
Delta R.T. 0.000 min
Lab File: BP017475.D
Acq: 30 Sep 2023 09:10

Tgt Ion:168 Resp: 851031
Ion Ratio Lower Upper
168 100
139 37.8 30.6 45.8









#57
2,4-Dinitrotoluene
Concen: 27.905 ng/ul
RT: 15.045 min Scan# 2033
Delta R.T. 0.000 min
Lab File: BP017475.D
Acq: 30 Sep 2023 09:10




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:165 Resp: 205401
Ion Ratio Lower Upper
165 100
63 60.8 45.6 68.4
182 2.7 2.0 3.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 26.872 ng/ul
RT: 15.263 min Scan# 2070
Delta R.T. -0.006 min
Lab File: BP017475.D
Acq: 30 Sep 2023 09:10

Tgt Ion:232 Resp: 185399
Ion Ratio Lower Upper
232 100
131 43.7 38.4 57.6
130 2.4 2.0 3.0
166 30.7 24.6 36.8









#59
Diethylphthalate
Concen: 24.819 ng/ul
RT: 15.463 min Scan# 2105
Delta R.T. -0.006 min
Lab File: BP017475.D
Acq: 30 Sep 2023 09:10




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:149 Resp: 690314
Ion Ratio Lower Upper
149 100
177 21.8 17.0 25.6
150 12.1 9.8 14.6



#61
Fluorene
Concen: 24.568 ng/ul
RT: 15.698 min Scan# 2144
Delta R.T. -0.006 min
Lab File: BP017475.D
Acq: 30 Sep 2023 09:10

Tgt Ion:166 Resp: 693428
Ion Ratio Lower Upper
166 100
165 99.0 77.8 116.6
167 13.7 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 25.138 ng/ul
RT: 15.692 min Scan# 2143
Delta R.T. 0.000 min
Lab File: BP017475.D
Acq: 30 Sep 2023 09:10




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:204 Resp: 359594
Ion Ratio Lower Upper
204 100
206 33.1 26.2 39.4
141 55.6 45.9 68.9



#63
4-Nitroaniline
Concen: 26.753 ng/ul
RT: 15.769 min Scan# 2156
Delta R.T. 0.000 min
Lab File: BP017475.D
Acq: 30 Sep 2023 09:10

Tgt Ion:138 Resp: 128401
Ion Ratio Lower Upper
138 100
92 52.7 42.4 63.6
108 91.8 73.0 109.6









#66
4,6-Dinitro-2-methylphenol
Concen: 20.527 ng/ul
RT: 15.798 min Scan# 2161
Delta R.T. -0.006 min
Lab File: BP017475.D
Acq: 30 Sep 2023 09:10




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:198 Resp: 111723
Ion Ratio Lower Upper
198 100
182 4.9 4.3 6.5
77 28.0 20.4 30.6



#67
N-Nitrosodiphenylamine
Concen: 23.510 ng/ul
RT: 15.922 min Scan# 2182
Delta R.T. 0.000 min
Lab File: BP017475.D
Acq: 30 Sep 2023 09:10

Tgt Ion:169 Resp: 578880
Ion Ratio Lower Upper
169 100
168 66.2 53.5 80.3
167 36.5 30.0 45.0









#68
4-Bromophenyl-phenylether
Concen: 24.683 ng/ul
RT: 16.598 min Scan# 2111
Delta R.T. -0.006 min
Lab File: BP017475.D
Acq: 30 Sep 2023 09:10




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:248 Resp: 221969
Ion Ratio Lower Upper
248 100
250 99.1 78.6 117.8
141 73.9 56.3 84.5



#69
Hexachlorobenzene
Concen: 24.979 ng/ul
RT: 16.681 min Scan# 2311
Delta R.T. -0.006 min
Lab File: BP017475.D
Acq: 30 Sep 2023 09:10

Tgt Ion:284 Resp: 260667
Ion Ratio Lower Upper
284 100
142 35.0 29.5 44.3
249 31.3 25.7 38.5





#70

Atrazine

Concen: 24.251 ng/ul

RT: 16.881 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP017475.D

Acq: 30 Sep 2023 09:10

Instrument :

BNA_P

ClientSampleId :

SSTD020763





Tgt Ion:200 Resp: 213928

Ion Ratio Lower Upper

200 100

173 24.8 20.3 30.5

215 46.8 38.2 57.4



#71

Pentachlorophenol

Concen: 23.436 ng/ul

RT: 17.051 min Scan# 2374

Delta R.T. -0.006 min

Lab File: BP017475.D

Acq: 30 Sep 2023 09:10




Tgt Ion:266 Resp: 151068


Ion Ratio Lower Upper

266 100

264 63.0 48.9 73.3

268 64.2 51.6 77.4





#72

Phenanthrene

Concen: 23.763 ng/ul

RT: 17.475 min Scan# 2446

Delta R.T. 0.000 min

Lab File: BP017475.D

Acq: 30 Sep 2023 09:10

Instrument :

BNA_P

ClientSampleId :

SSTD020763





Tgt Ion:178 Resp: 1114117

Ion Ratio Lower Upper

178 100

179 15.4 12.7 19.1

176 19.0 15.8 23.6



#74

Anthracene

Concen: 24.184 ng/ul

RT: 17.563 min Scan# 2461

Delta R.T. -0.006 min

Lab File: BP017475.D

Acq: 30 Sep 2023 09:10




Tgt Ion:178 Resp: 1149169


Ion Ratio Lower Upper

178 100

179 15.4 12.5 18.7

176 18.3 15.1 22.7









#75
1,2,3,4-Tetrachlorobenzene
Concen: 22.975 ng/uL
RT: 13.422 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP017475.D
Acq: 30 Sep 2023 09:10




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:216 Resp: 315639
Ion Ratio Lower Upper
216 100
214 77.7 61.2 91.8
218 48.8 36.8 55.2



#76
Pentachlorobenzene
Concen: 23.975 ng/uL
RT: 14.934 min Scan# 2014
Delta R.T. -0.006 min
Lab File: BP017475.D
Acq: 30 Sep 2023 09:10

Tgt Ion:250 Resp: 311147
Ion Ratio Lower Upper
250 100
248 63.5 50.1 75.1
252 62.9 52.2 78.2









#77
 Carbazole
 Concen: 23.710 ng/ul
 RT: 17.851 min Scan# 2510
 Delta R.T. -0.006 min
 Lab File: BP017475.D
 Acq: 30 Sep 2023 09:10




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020763


Tgt Ion:167 Resp: 1002114
 Ion Ratio Lower Upper
 167 100
 166 20.9 16.5 24.7
 139 12.9 10.5 15.7



#78
 Di-n-butylphthalate
 Concen: 25.765 ng/ul
 RT: 18.363 min Scan# 2597
 Delta R.T. 0.000 min
 Lab File: BP017475.D
 Acq: 30 Sep 2023 09:10

Tgt Ion:149 Resp: 1197441
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.5 11.3
 104 5.4 4.5 6.7





#80

Fluoranthene

Concen: 24.509 ng/ul

RT: 19.451 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP017475.D

Acq: 30 Sep 2023 09:10

Instrument :

BNA_P

ClientSampleId :

SSTD020763





Tgt Ion:202 Resp: 1345114

Ion Ratio Lower Upper

202 100

101 9.5 8.4 12.6

100 7.7 6.6 10.0



#82

Pyrene

Concen: 24.190 ng/ul

RT: 19.810 min Scan# 2843

Delta R.T. 0.000 min

Lab File: BP017475.D

Acq: 30 Sep 2023 09:10




Tgt Ion:202 Resp: 1411341


Ion Ratio Lower Upper

202 100

101 11.7 9.8 14.8

100 9.8 8.1 12.1





#83

Butylbenzylphthalate

Concen: 26.608 ng/ul

RT: 20.674 min Scan# 2991

Delta R.T. 0.000 min

Lab File: BP017475.D

Acq: 30 Sep 2023 09:10

Instrument :

BNA_P

ClientSampleId :

SSTD020763





Tgt Ion:149 Resp: 550430

Ion Ratio Lower Upper

149 100

91 71.9 59.8 89.6

206 21.1 15.7 23.5



#84

3,3'-Dichlorobenzidine

Concen: 22.308 ng/ul

RT: 21.480 min Scan# 3127

Delta R.T. 0.000 min

Lab File: BP017475.D

Acq: 30 Sep 2023 09:10




Tgt Ion:252 Resp: 421781


Ion Ratio Lower Upper

252 100

254 64.6 51.9 77.9

126 10.3 9.1 13.7





#85

Benzo(a)anthracene

Concen: 24.130 ng/ul

RT: 21.539 min Scan# 3139

Delta R.T. 0.000 min

Lab File: BP017475.D

Acq: 30 Sep 2023 09:10

Instrument :

BNA_P

ClientSampleId :

SSTD020763





Tgt Ion:228 Resp: 1417223

Ion Ratio Lower Upper

228 100

229 19.7 15.8 23.6

226 27.6 21.8 32.8



#86

Bis(2-ethylhexyl)phthalate

Concen: 26.948 ng/ul

RT: 21.416 min Scan# 3116

Delta R.T. 0.000 min

Lab File: BP017475.D

Acq: 30 Sep 2023 09:10




Tgt Ion:149 Resp: 805065


Ion Ratio Lower Upper

149 100

167 27.4 21.0 31.6

279 4.6 3.0 4.4#





#87

Chrysene

Concen: 23.772 ng/ul

RT: 21.598 min Scan# 3148

Delta R.T. 0.000 min

Lab File: BP017475.D

Acq: 30 Sep 2023 09:10

Instrument :

BNA_P

ClientSampleId :

SSTD020763





Tgt Ion:228 Resp: 1313955

Ion Ratio Lower Upper

228 100

226 30.4 24.1 36.1

229 19.6 15.0 22.6



#89

Di-n-octyl phthalate

Concen: 26.357 ng/ul




RT: 22.398 min Scan# 3283


Delta R.T. 0.000 min

Lab File: BP017475.D

Acq: 30 Sep 2023 09:10

Tgt Ion:149 Resp: 1389555





#90

Benzo(b)fluoranthene

Concen: 23.837 ng/ul

RT: 23.327 min Scan# 3443

Delta R.T. 0.000 min

Lab File: BP017475.D

Acq: 30 Sep 2023 09:10

Instrument :

BNA_P

ClientSampleId :

SSTD020763





Tgt Ion:252 Resp: 1346989

Ion Ratio Lower Upper

252 100

253 22.2 17.4 26.0

125 9.4 7.8 11.8



#91

Benzo(k)fluoranthene

Concen: 24.060 ng/ul

RT: 23.380 min Scan# 3450

Delta R.T. 0.000 min

Lab File: BP017475.D

Acq: 30 Sep 2023 09:10




Tgt Ion:252 Resp: 1371647


Ion Ratio Lower Upper

252 100

253 21.5 17.9 26.9

125 9.2 7.8 11.6





#93

Benzo(a)pyrene

Concen: 23.719 ng/ul

RT: 24.010 min Scan# 31

Delta R.T. -0.006 min

Lab File: BP017475.D

Acq: 30 Sep 2023 09:10

Instrument :

BNA_P

ClientSampleId :

SSTD020763





Tgt Ion:252 Resp: 1282395

Ion Ratio Lower Upper

252 100

253 21.6 17.4 26.0

125 10.4 8.9 13.3



#94

Indeno(1,2,3-cd)pyrene

Concen: 22.571 ng/ul

RT: 26.857 min Scan# 4041

Delta R.T. 0.000 min

Lab File: BP017475.D

Acq: 30 Sep 2023 09:10




Tgt Ion:276 Resp: 1529566


Ion Ratio Lower Upper

276 100

138 22.1 17.9 26.9

277 24.7 19.5 29.3









#95
 Dibenzo(a,h)anthracene
 Concen: 22.300 ng/ul
 RT: 26.880 min Scan# 4045
 Delta R.T. 0.000 min
 Lab File: BP017475.D
 Acq: 30 Sep 2023 09:10

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020763

Tgt Ion:278 Resp: 1258685

Ion	Ratio	Lower	Upper
278	100		
139	14.7	13.4	20.0
279	23.5	19.5	29.3



#96
 Benzo(g,h,i)perylene
 Concen: 22.485 ng/ul
 RT: 27.709 min Scan# 4186
 Delta R.T. 0.006 min
 Lab File: BP017475.D
 Acq: 30 Sep 2023 09:10

Tgt Ion:276 Resp: 1228800

Ion	Ratio	Lower	Upper
276	100		
138	20.7	17.3	25.9
277	23.3	18.8	28.2

